

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3400</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 08:43</u>
Lab File ID:	<u>VX048252.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.514	0.496		-3.5	20
Chloromethane	0.480	0.468	0.1	-2.5	20
Vinyl Chloride	0.444	0.448		0.9	20
Bromomethane	0.302	0.299		-0.99	20
Chloroethane	0.261	0.261		0	20
Trichlorofluoromethane	0.852	0.813		-4.58	20
1,1,2-Trichlorotrifluoroethane	0.540	0.513		-5	20
1,1-Dichloroethene	0.525	0.498		-5.14	20
Acetone	0.240	0.230		-4.17	20
Carbon Disulfide	1.654	1.439		-13	20
Methyl tert-butyl Ether	1.791	1.576		-12	20
Methyl Acetate	0.502	0.411		-18.13	20
Methylene Chloride	0.586	0.556		-5.12	20
trans-1,2-Dichloroethene	0.577	0.538		-6.76	20
1,1-Dichloroethane	1.051	0.952	0.1	-9.42	20
Cyclohexane	0.870	0.788		-9.43	20
2-Butanone	0.304	0.260		-14.47	20
Carbon Tetrachloride	0.599	0.533		-11.02	20
cis-1,2-Dichloroethene	0.673	0.610		-9.36	20
Bromochloromethane	0.436	0.435		-0.23	20
Chloroform	1.123	1.006		-10.42	20
1,1,1-Trichloroethane	1.035	0.920		-11.11	20
Methylcyclohexane	0.573	0.510		-10.99	20
Benzene	1.409	1.280		-9.15	20
1,2-Dichloroethane	0.520	0.466		-10.39	20
Trichloroethene	0.372	0.337		-9.41	20
1,2-Dichloropropane	0.335	0.305		-8.95	20
Bromodichloromethane	0.563	0.510		-9.41	20
4-Methyl-2-Pentanone	0.381	0.302		-20.74	20
Toluene	0.895	0.795		-11.17	20
t-1,3-Dichloropropene	0.543	0.493		-9.21	20
cis-1,3-Dichloropropene	0.586	0.529		-9.73	20
1,1,2-Trichloroethane	0.333	0.291		-12.61	20
2-Hexanone	0.266	0.227		-14.66	20
Dibromochloromethane	0.447	0.388		-13.2	20
1,2-Dibromoethane	0.355	0.307		-13.52	20
Tetrachloroethene	0.373	0.325		-12.87	20
Chlorobenzene	1.125	1.025	0.3	-8.89	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3400</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 08:43</u>
Lab File ID:	<u>VX048252.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.915	1.738		-9.24	20
m/p-Xylenes	0.712	0.666		-6.46	20
o-Xylene	0.687	0.625		-9.02	20
Styrene	1.170	1.107		-5.39	20
Bromoform	0.330	0.283	0.1	-14.24	20
Isopropylbenzene	3.652	3.379		-7.47	20
1,1,2,2-Tetrachloroethane	0.997	0.857	0.3	-14.04	20
1,3-Dichlorobenzene	1.726	1.555		-9.91	20
1,4-Dichlorobenzene	1.788	1.551		-13.26	20
1,2-Dichlorobenzene	1.621	1.416		-12.65	20
1,2-Dibromo-3-Chloropropane	0.202	0.163		-19.31	20
1,2,4-Trichlorobenzene	1.096	0.885		-19.25	20
1,2,3-Trichlorobenzene	1.006	0.834		-17.1	20
1,2-Dichloroethane-d4	0.701	0.658		-6.13	20
Dibromofluoromethane	0.341	0.340		-0.29	20
Toluene-d8	1.196	1.150		-3.85	20
4-Bromofluorobenzene	0.455	0.419		-7.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.