

LAB CHRONICLE

OrderID:	Q3400	OrderDate:	10/20/2025 1:25:31 PM
Client:	ICE Service Group, Inc.	Project:	NWIRP Northrup Grumman Site – Bethpage, NY
Contact:	Dennis D. Morgan, II	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3400-01	FRAC-TANK-0760450	Water	VOC-TCLVOA-10	8260-Low	10/20/25		10/21/25	10/20/25

Hit Summary Sheet
SW-846

SDG No.: Q3400

Client: ICE Service Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	FRAC-TANK-0760450								
Q3400-01	FRAC-TANK-0760	Water	Acetone	5.10		1.50	3.80	5.00	ug/L
Q3400-01	FRAC-TANK-0760	Water	cis-1,2-Dichloroethene	9.70		0.19	0.75	1.00	ug/L
Q3400-01	FRAC-TANK-0760	Water	Trichloroethene	86.7		0.090	0.75	1.00	ug/L
Q3400-01	FRAC-TANK-0760	Water	1,2-Dichloropropane	17.2		0.20	0.50	1.00	ug/L
Q3400-01	FRAC-TANK-0760	Water	Tetrachloroethene	6.50		0.23	0.50	1.00	ug/L
Total Voc :					125				
Total Concentration:					125				



SAMPLE DATA

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: FRAC-TANK-0760450
Lab Sample ID: Q3400-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/20/25
Date Received: 10/20/25
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-71-8	Dichlorodifluoromethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 16:46	VX102125
74-87-3	Chloromethane	0.50	U	1	0.32	0.50	1.00	ug/L	10/21/25 16:46	VX102125
75-01-4	Vinyl Chloride	0.75	U	1	0.26	0.75	1.00	ug/L	10/21/25 16:46	VX102125
74-83-9	Bromomethane	3.80	U	1	1.40	3.80	5.00	ug/L	10/21/25 16:46	VX102125
75-00-3	Chloroethane	0.75	U	1	0.47	0.75	1.00	ug/L	10/21/25 16:46	VX102125
75-69-4	Trichlorofluoromethane	0.50	U	1	0.33	0.50	1.00	ug/L	10/21/25 16:46	VX102125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	1	0.25	0.50	1.00	ug/L	10/21/25 16:46	VX102125
75-35-4	1,1-Dichloroethene	0.75	U	1	0.23	0.75	1.00	ug/L	10/21/25 16:46	VX102125
67-64-1	Acetone	5.10		1	1.50	3.80	5.00	ug/L	10/21/25 16:46	VX102125
75-15-0	Carbon Disulfide	0.75	U	1	0.21	0.75	1.00	ug/L	10/21/25 16:46	VX102125
1634-04-4	Methyl tert-butyl Ether	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 16:46	VX102125
79-20-9	Methyl Acetate	0.75	U	1	0.27	0.75	1.00	ug/L	10/21/25 16:46	VX102125
75-09-2	Methylene Chloride	0.50	U	1	0.28	0.50	1.00	ug/L	10/21/25 16:46	VX102125
156-60-5	trans-1,2-Dichloroethene	0.50	U	1	0.23	0.50	1.00	ug/L	10/21/25 16:46	VX102125
75-34-3	1,1-Dichloroethane	0.50	U	1	0.23	0.50	1.00	ug/L	10/21/25 16:46	VX102125
110-82-7	Cyclohexane	2.50	U	1	1.50	2.50	5.00	ug/L	10/21/25 16:46	VX102125
78-93-3	2-Butanone	2.50	U	1	0.98	2.50	5.00	ug/L	10/21/25 16:46	VX102125
56-23-5	Carbon Tetrachloride	0.50	U	1	0.25	0.50	1.00	ug/L	10/21/25 16:46	VX102125
156-59-2	cis-1,2-Dichloroethene	9.70		1	0.19	0.75	1.00	ug/L	10/21/25 16:46	VX102125
74-97-5	Bromochloromethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 16:46	VX102125
67-66-3	Chloroform	0.50	U	1	0.25	0.50	1.00	ug/L	10/21/25 16:46	VX102125
71-55-6	1,1,1-Trichloroethane	0.50	U	1	0.20	0.50	1.00	ug/L	10/21/25 16:46	VX102125
108-87-2	Methylcyclohexane	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 16:46	VX102125
71-43-2	Benzene	0.50	U	1	0.15	0.50	1.00	ug/L	10/21/25 16:46	VX102125
107-06-2	1,2-Dichloroethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 16:46	VX102125
79-01-6	Trichloroethene	86.7		1	0.090	0.75	1.00	ug/L	10/21/25 16:46	VX102125
78-87-5	1,2-Dichloropropane	17.2		1	0.20	0.50	1.00	ug/L	10/21/25 16:46	VX102125
75-27-4	Bromodichloromethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 16:46	VX102125
108-10-1	4-Methyl-2-Pentanone	2.50	U	1	0.68	2.50	5.00	ug/L	10/21/25 16:46	VX102125
108-88-3	Toluene	0.50	U	1	0.14	0.50	1.00	ug/L	10/21/25 16:46	VX102125
10061-02-6	t-1,3-Dichloropropene	0.50	U	1	0.17	0.50	1.00	ug/L	10/21/25 16:46	VX102125
10061-01-5	cis-1,3-Dichloropropene	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 16:46	VX102125
79-00-5	1,1,2-Trichloroethane	0.50	U	1	0.21	0.50	1.00	ug/L	10/21/25 16:46	VX102125
591-78-6	2-Hexanone	2.50	U	1	0.89	2.50	5.00	ug/L	10/21/25 16:46	VX102125
124-48-1	Dibromochloromethane	0.50	U	1	0.18	0.50	1.00	ug/L	10/21/25 16:46	VX102125
106-93-4	1,2-Dibromoethane	0.50	U	1	0.15	0.50	1.00	ug/L	10/21/25 16:46	VX102125
127-18-4	Tetrachloroethene	6.50		1	0.23	0.50	1.00	ug/L	10/21/25 16:46	VX102125
108-90-7	Chlorobenzene	0.50	U	1	0.12	0.50	1.00	ug/L	10/21/25 16:46	VX102125
100-41-4	Ethyl Benzene	0.50	U	1	0.13	0.50	1.00	ug/L	10/21/25 16:46	VX102125
179601-23-1	m/p-Xylenes	1.00	U	1	0.24	1.00	2.00	ug/L	10/21/25 16:46	VX102125
95-47-6	o-Xylene	0.50	U	1	0.12	0.50	1.00	ug/L	10/21/25 16:46	VX102125
100-42-5	Styrene	0.50	U	1	0.15	0.50	1.00	ug/L	10/21/25 16:46	VX102125
75-25-2	Bromoform	0.50	U	1	0.19	0.50	1.00	ug/L	10/21/25 16:46	VX102125

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: FRAC-TANK-0760450
Lab Sample ID: Q3400-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/20/25
Date Received: 10/20/25
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.50	U	1	0.12	0.50	1.00	ug/L	10/21/25 16:46	VX102125
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	1	0.26	0.50	1.00	ug/L	10/21/25 16:46	VX102125
541-73-1	1,3-Dichlorobenzene	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 16:46	VX102125
106-46-7	1,4-Dichlorobenzene	0.50	U	1	0.19	0.50	1.00	ug/L	10/21/25 16:46	VX102125
95-50-1	1,2-Dichlorobenzene	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 16:46	VX102125
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	1	0.53	0.75	1.00	ug/L	10/21/25 16:46	VX102125
120-82-1	1,2,4-Trichlorobenzene	0.50	U	1	0.20	0.50	1.00	ug/L	10/21/25 16:46	VX102125
87-61-6	1,2,3-Trichlorobenzene	0.75	U	1	0.20	0.75	1.00	ug/L	10/21/25 16:46	VX102125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	55.7			81 - 118	111%	SPK: 50
1868-53-7	Dibromofluoromethane	47.4			80 - 119	95%	SPK: 50
2037-26-5	Toluene-d8	49.8			89 - 112	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	58.0	*		85 - 114	116%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	118000
540-36-3	1,4-Difluorobenzene	236000
3114-55-4	Chlorobenzene-d5	266000
3855-82-1	1,4-Dichlorobenzene-d4	137000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q3400

Client: ICE Service Group, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3400-01	FRAC-TANK-0760450	1,2-Dichloroethane-d4	50	55.7	111		81	118
		Dibromofluoromethane	50	47.4	95		80	119
		Toluene-d8	50	49.8	100		89	112
		4-Bromofluorobenzene	50	58.0	116	*	85	114
VX1021WBL01	VX1021WBL01	1,2-Dichloroethane-d4	50	52.5	105		81	118
		Dibromofluoromethane	50	47.1	94		80	119
		Toluene-d8	50	46.4	93		89	112
		4-Bromofluorobenzene	50	52.6	105		85	114
VX1021WBS01	VX1021WBS01	1,2-Dichloroethane-d4	50	49.2	98		81	118
		Dibromofluoromethane	50	49.3	99		80	119
		Toluene-d8	50	47.8	96		89	112
		4-Bromofluorobenzene	50	49.7	99		85	114
VX1021WBSD0	VX1021WBSD01	1,2-Dichloroethane-d4	50	46.5	93		81	118
		Dibromofluoromethane	50	48.0	96		80	119
		Toluene-d8	50	47.9	96		89	112
		4-Bromofluorobenzene	50	47.3	95		85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3400</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048255.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1021WBS01	Dichlorodifluoromethane	20	16.8	ug/L	84			32	152	
	Chloromethane	20	17.8	ug/L	89			50	139	
	Vinyl chloride	20	17.6	ug/L	88			58	137	
	Bromomethane	20	19.3	ug/L	97			53	141	
	Chloroethane	20	18.9	ug/L	95			60	138	
	Trichlorofluoromethane	20	17.7	ug/L	89			65	141	
	1,1,2-Trichlorotrifluoroethane	20	17.6	ug/L	88			70	136	
	1,1-Dichloroethene	20	18.0	ug/L	90			71	131	
	Acetone	100	82.3	ug/L	82			39	160	
	Carbon disulfide	20	16.6	ug/L	83			64	133	
	Methyl tert-butyl Ether	20	17.6	ug/L	88			71	124	
	Methyl Acetate	20	17.5	ug/L	88			56	136	
	Methylene Chloride	20	19.1	ug/L	96			74	124	
	trans-1,2-Dichloroethene	20	17.9	ug/L	90			75	124	
	1,1-Dichloroethane	20	18.4	ug/L	92			77	125	
	Cyclohexane	20	19.2	ug/L	96			71	130	
	2-Butanone	100	85.2	ug/L	85			56	143	
	Carbon Tetrachloride	20	17.0	ug/L	85			72	136	
	cis-1,2-Dichloroethene	20	18.9	ug/L	95			78	123	
	Bromochloromethane	20	21.6	ug/L	108			78	123	
	Chloroform	20	18.8	ug/L	94			79	124	
	1,1,1-Trichloroethane	20	18.1	ug/L	91			74	131	
	Methylcyclohexane	20	17.0	ug/L	85			72	132	
	Benzene	20	18.5	ug/L	93			79	120	
	1,2-Dichloroethane	20	19.1	ug/L	96			73	128	
	Trichloroethene	20	18.1	ug/L	91			79	123	
	1,2-Dichloropropane	20	19.3	ug/L	97			78	122	
	Bromodichloromethane	20	18.6	ug/L	93			79	125	
	4-Methyl-2-Pentanone	100	86.9	ug/L	87			67	130	
	Toluene	20	18.8	ug/L	94			80	121	
	t-1,3-Dichloropropene	20	18.4	ug/L	92			73	127	
	cis-1,3-Dichloropropene	20	18.8	ug/L	94			75	124	
	1,1,2-Trichloroethane	20	19.1	ug/L	96			80	119	
	2-Hexanone	100	85.9	ug/L	86			57	139	
	Dibromochloromethane	20	18.1	ug/L	91			74	126	
	1,2-Dibromoethane	20	18.6	ug/L	93			77	121	
	Tetrachloroethene	20	16.9	ug/L	85			74	129	
	Chlorobenzene	20	18.0	ug/L	90			82	118	
	Ethyl Benzene	20	17.8	ug/L	89			79	121	
	m/p-Xylenes	40	36.8	ug/L	92			80	121	
	o-Xylene	20	18.1	ug/L	91			78	122	
	Styrene	20	18.8	ug/L	94			78	123	
	Bromoform	20	17.0	ug/L	85			66	130	
	Isopropylbenzene	20	17.1	ug/L	86			72	131	
	1,1,2,2-Tetrachloroethane	20	17.3	ug/L	86			71	121	
	1,3-Dichlorobenzene	20	17.2	ug/L	86			80	119	
	1,4-Dichlorobenzene	20	16.6	ug/L	83			79	118	
	1,2-Dichlorobenzene	20	17.2	ug/L	86			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3400</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048255.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1021WBS01	1,2-Dibromo-3-Chloropropane	20	14.2	ug/L	71			62	128	
	1,2,4-Trichlorobenzene	20	15.2	ug/L	76			69	130	
	1,2,3-Trichlorobenzene	20	15.5	ug/L	78			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3400</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048256.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1021WBSD01	Dichlorodifluoromethane	20	18.0	ug/L	90	7		32	152	20
	Chloromethane	20	17.3	ug/L	86	3		50	139	20
	Vinyl chloride	20	16.9	ug/L	85	3		58	137	20
	Bromomethane	20	19.0	ug/L	95	2		53	141	20
	Chloroethane	20	19.0	ug/L	95	0		60	138	20
	Trichlorofluoromethane	20	17.9	ug/L	90	1		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	17.7	ug/L	89	1		70	136	20
	1,1-Dichloroethene	20	17.9	ug/L	90	0		71	131	20
	Acetone	100	81.6	ug/L	82	0		39	160	20
	Carbon disulfide	20	16.9	ug/L	85	2		64	133	20
	Methyl tert-butyl Ether	20	18.0	ug/L	90	2		71	124	20
	Methyl Acetate	20	16.9	ug/L	85	3		56	136	20
	Methylene Chloride	20	18.5	ug/L	93	3		74	124	20
	trans-1,2-Dichloroethene	20	18.8	ug/L	94	4		75	124	20
	1,1-Dichloroethane	20	18.8	ug/L	94	2		77	125	20
	Cyclohexane	20	18.4	ug/L	92	4		71	130	20
	2-Butanone	100	85.7	ug/L	86	1		56	143	20
	Carbon Tetrachloride	20	17.2	ug/L	86	1		72	136	20
	cis-1,2-Dichloroethene	20	19.1	ug/L	96	1		78	123	20
	Bromochloromethane	20	21.3	ug/L	106	2		78	123	20
	Chloroform	20	19.0	ug/L	95	1		79	124	20
	1,1,1-Trichloroethane	20	18.3	ug/L	92	1		74	131	20
	Methylcyclohexane	20	18.2	ug/L	91	7		72	132	20
	Benzene	20	18.8	ug/L	94	1		79	120	20
	1,2-Dichloroethane	20	18.5	ug/L	93	3		73	128	20
	Trichloroethene	20	18.3	ug/L	92	1		79	123	20
	1,2-Dichloropropane	20	19.8	ug/L	99	2		78	122	20
	Bromodichloromethane	20	18.8	ug/L	94	1		79	125	20
	4-Methyl-2-Pentanone	100	89.0	ug/L	89	2		67	130	20
	Toluene	20	18.9	ug/L	95	1		80	121	20
	t-1,3-Dichloropropene	20	18.5	ug/L	93	1		73	127	20
	cis-1,3-Dichloropropene	20	18.5	ug/L	93	1		75	124	20
	1,1,2-Trichloroethane	20	19.3	ug/L	97	1		80	119	20
	2-Hexanone	100	87.8	ug/L	88	2		57	139	20
	Dibromochloromethane	20	18.3	ug/L	92	1		74	126	20
	1,2-Dibromoethane	20	18.6	ug/L	93	0		77	121	20
	Tetrachloroethene	20	17.5	ug/L	88	3		74	129	20
	Chlorobenzene	20	18.9	ug/L	95	5		82	118	20
	Ethyl Benzene	20	18.3	ug/L	92	3		79	121	20
	m/p-Xylenes	40	37.5	ug/L	94	2		80	121	20
	o-Xylene	20	18.5	ug/L	93	2		78	122	20
	Styrene	20	19.2	ug/L	96	2		78	123	20
	Bromoform	20	17.7	ug/L	89	5		66	130	20
	Isopropylbenzene	20	18.5	ug/L	93	8		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.4	ug/L	92	7		71	121	20
	1,3-Dichlorobenzene	20	18.2	ug/L	91	6		80	119	20
	1,4-Dichlorobenzene	20	18.2	ug/L	91	9		79	118	20
	1,2-Dichlorobenzene	20	18.6	ug/L	93	8		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3400</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048256.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1021WBSD01	1,2-Dibromo-3-Chloropropane	20	16.9	ug/L	85	18		62	128	20
	1,2,4-Trichlorobenzene	20	16.8	ug/L	84	10		69	130	20
	1,2,3-Trichlorobenzene	20	17.3	ug/L	86	10		69	129	20

VOLATILE METHOD BLANK SUMMARY

Client ID

VX1021WBL01

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Lab File ID: VX048254.D

Lab Sample ID: VX1021WBL01

Date Analyzed: 10/21/2025

Time Analyzed: 09:40

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1021WBS01	VX1021WBS01	VX048255.D	10/21/2025
VX1021WBSD01	VX1021WBSD01	VX048256.D	10/21/2025
FRAC-TANK-0760450	Q3400-01	VX048274.D	10/21/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Lab File ID: VX048226.D

BFB Injection Date: 10/20/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:15

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.7
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.3 (0.4) 1
174	50.0 - 100.0% of mass 95	76.5
175	5.0 - 9.0% of mass 174	5.8 (7.5) 1
176	95.0 - 101.0% of mass 174	73.8 (96.4) 1
177	5.0 - 9.0% of mass 176	4.8 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX048229.D	10/20/2025	10:59
VSTDIC020	VSTDIC020	VX048231.D	10/20/2025	11:40
VSTDIC050	VSTDIC050	VX048232.D	10/20/2025	12:01
VSTDIC100	VSTDIC100	VX048233.D	10/20/2025	12:22
VSTDIC150	VSTDIC150	VX048234.D	10/20/2025	12:43
VSTDIC005	VSTDIC005	VX048236.D	10/20/2025	13:49

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Lab File ID: VX048251.D

BFB Injection Date: 10/21/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:12

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18
75	30.0 - 60.0% of mass 95	50.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.6 (0.7) 1
174	50.0 - 100.0% of mass 95	75.8
175	5.0 - 9.0% of mass 174	5.5 (7.2) 1
176	95.0 - 101.0% of mass 174	72.8 (96.1) 1
177	5.0 - 9.0% of mass 176	5 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048252.D	10/21/2025	08:43
VX1021WBL01	VX1021WBL01	VX048254.D	10/21/2025	09:40
VX1021WBS01	VX1021WBS01	VX048255.D	10/21/2025	10:06
VX1021WBSD01	VX1021WBSD01	VX048256.D	10/21/2025	10:33
FRAC-TANK-0760450	Q3400-01	VX048274.D	10/21/2025	16:46
VSTDCCC050EC	VSTDCCC050	VX048278.D	10/21/2025	18:09

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ICES02
 Lab Code: ACE SDG NO.: Q3400
 Lab File ID: VX048252.D Date Analyzed: 10/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144122	5.53	227103	6.74	197016	10.04
UPPER LIMIT	288244	6.031	454206	7.238	394032	10.537
LOWER LIMIT	72061	5.031	113552	6.238	98508	9.537
EPA SAMPLE NO.						
FRAC-TANK-0760450	117685	5.54	236203	6.75	265928	10.04
VX1021WBL01	121915	5.53	240535	6.75	247603	10.04
VX1021WBS01	141199	5.54	230109	6.74	212342	10.04
VX1021WBSD01	137604	5.53	224103	6.74	203364	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3400</u>
Lab File ID:	<u>VX048252.D</u>	Date Analyzed:	<u>10/21/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:43</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	96991	12.00€				
UPPER LIMIT	193982	12.50€				
LOWER LIMIT	48495.5	11.50€				
EPA SAMPLE NO.						
FRAC-TANK-0760450	136752	12.01				
VX1021WBL01	126551	12.01				
VX1021WBS01	110789	12.01				
VX1021WBSD01	101316	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1021WBL01
Lab Sample ID: VX1021WBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-71-8	Dichlorodifluoromethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 09:40	VX102125
74-87-3	Chloromethane	0.50	U	1	0.32	0.50	1.00	ug/L	10/21/25 09:40	VX102125
75-01-4	Vinyl Chloride	0.75	U	1	0.26	0.75	1.00	ug/L	10/21/25 09:40	VX102125
74-83-9	Bromomethane	3.80	U	1	1.40	3.80	5.00	ug/L	10/21/25 09:40	VX102125
75-00-3	Chloroethane	0.75	U	1	0.47	0.75	1.00	ug/L	10/21/25 09:40	VX102125
75-69-4	Trichlorofluoromethane	0.50	U	1	0.33	0.50	1.00	ug/L	10/21/25 09:40	VX102125
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	1	0.25	0.50	1.00	ug/L	10/21/25 09:40	VX102125
75-35-4	1,1-Dichloroethene	0.75	U	1	0.23	0.75	1.00	ug/L	10/21/25 09:40	VX102125
67-64-1	Acetone	3.80	U	1	1.50	3.80	5.00	ug/L	10/21/25 09:40	VX102125
75-15-0	Carbon Disulfide	0.75	U	1	0.21	0.75	1.00	ug/L	10/21/25 09:40	VX102125
1634-04-4	Methyl tert-butyl Ether	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 09:40	VX102125
79-20-9	Methyl Acetate	0.75	U	1	0.27	0.75	1.00	ug/L	10/21/25 09:40	VX102125
75-09-2	Methylene Chloride	0.50	U	1	0.28	0.50	1.00	ug/L	10/21/25 09:40	VX102125
156-60-5	trans-1,2-Dichloroethene	0.50	U	1	0.23	0.50	1.00	ug/L	10/21/25 09:40	VX102125
75-34-3	1,1-Dichloroethane	0.50	U	1	0.23	0.50	1.00	ug/L	10/21/25 09:40	VX102125
110-82-7	Cyclohexane	2.50	U	1	1.50	2.50	5.00	ug/L	10/21/25 09:40	VX102125
78-93-3	2-Butanone	2.50	U	1	0.98	2.50	5.00	ug/L	10/21/25 09:40	VX102125
56-23-5	Carbon Tetrachloride	0.50	U	1	0.25	0.50	1.00	ug/L	10/21/25 09:40	VX102125
156-59-2	cis-1,2-Dichloroethene	0.75	U	1	0.19	0.75	1.00	ug/L	10/21/25 09:40	VX102125
74-97-5	Bromochloromethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 09:40	VX102125
67-66-3	Chloroform	0.50	U	1	0.25	0.50	1.00	ug/L	10/21/25 09:40	VX102125
71-55-6	1,1,1-Trichloroethane	0.50	U	1	0.20	0.50	1.00	ug/L	10/21/25 09:40	VX102125
108-87-2	Methylcyclohexane	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 09:40	VX102125
71-43-2	Benzene	0.50	U	1	0.15	0.50	1.00	ug/L	10/21/25 09:40	VX102125
107-06-2	1,2-Dichloroethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 09:40	VX102125
79-01-6	Trichloroethene	0.75	U	1	0.090	0.75	1.00	ug/L	10/21/25 09:40	VX102125
78-87-5	1,2-Dichloropropane	0.50	U	1	0.20	0.50	1.00	ug/L	10/21/25 09:40	VX102125
75-27-4	Bromodichloromethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/21/25 09:40	VX102125
108-10-1	4-Methyl-2-Pentanone	2.50	U	1	0.68	2.50	5.00	ug/L	10/21/25 09:40	VX102125
108-88-3	Toluene	0.50	U	1	0.14	0.50	1.00	ug/L	10/21/25 09:40	VX102125
10061-02-6	t-1,3-Dichloropropene	0.50	U	1	0.17	0.50	1.00	ug/L	10/21/25 09:40	VX102125
10061-01-5	cis-1,3-Dichloropropene	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 09:40	VX102125
79-00-5	1,1,2-Trichloroethane	0.50	U	1	0.21	0.50	1.00	ug/L	10/21/25 09:40	VX102125
591-78-6	2-Hexanone	2.50	U	1	0.89	2.50	5.00	ug/L	10/21/25 09:40	VX102125
124-48-1	Dibromochloromethane	0.50	U	1	0.18	0.50	1.00	ug/L	10/21/25 09:40	VX102125
106-93-4	1,2-Dibromoethane	0.50	U	1	0.15	0.50	1.00	ug/L	10/21/25 09:40	VX102125
127-18-4	Tetrachloroethene	0.50	U	1	0.23	0.50	1.00	ug/L	10/21/25 09:40	VX102125
108-90-7	Chlorobenzene	0.50	U	1	0.12	0.50	1.00	ug/L	10/21/25 09:40	VX102125
100-41-4	Ethyl Benzene	0.50	U	1	0.13	0.50	1.00	ug/L	10/21/25 09:40	VX102125
179601-23-1	m/p-Xylenes	1.00	U	1	0.24	1.00	2.00	ug/L	10/21/25 09:40	VX102125
95-47-6	o-Xylene	0.50	U	1	0.12	0.50	1.00	ug/L	10/21/25 09:40	VX102125
100-42-5	Styrene	0.50	U	1	0.15	0.50	1.00	ug/L	10/21/25 09:40	VX102125
75-25-2	Bromoform	0.50	U	1	0.19	0.50	1.00	ug/L	10/21/25 09:40	VX102125

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1021WBL01
Lab Sample ID: VX1021WBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.50	U	1	0.12	0.50	1.00	ug/L	10/21/25 09:40	VX102125
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	1	0.26	0.50	1.00	ug/L	10/21/25 09:40	VX102125
541-73-1	1,3-Dichlorobenzene	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 09:40	VX102125
106-46-7	1,4-Dichlorobenzene	0.50	U	1	0.19	0.50	1.00	ug/L	10/21/25 09:40	VX102125
95-50-1	1,2-Dichlorobenzene	0.50	U	1	0.16	0.50	1.00	ug/L	10/21/25 09:40	VX102125
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	1	0.53	0.75	1.00	ug/L	10/21/25 09:40	VX102125
120-82-1	1,2,4-Trichlorobenzene	0.50	U	1	0.20	0.50	1.00	ug/L	10/21/25 09:40	VX102125
87-61-6	1,2,3-Trichlorobenzene	0.75	U	1	0.20	0.75	1.00	ug/L	10/21/25 09:40	VX102125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	52.5			81 - 118	105%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1			80 - 119	94%	SPK: 50
2037-26-5	Toluene-d8	46.4			89 - 112	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.7			85 - 114	105%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	122000
540-36-3	1,4-Difluorobenzene	241000
3114-55-4	Chlorobenzene-d5	248000
3855-82-1	1,4-Dichlorobenzene-d4	127000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1021WBS01
Lab Sample ID: VX1021WBS01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-71-8	Dichlorodifluoromethane	16.8		1	0.22	0.50	1.00	ug/L	10/21/25 10:06	VX102125
74-87-3	Chloromethane	17.8		1	0.32	0.50	1.00	ug/L	10/21/25 10:06	VX102125
75-01-4	Vinyl Chloride	17.6		1	0.26	0.75	1.00	ug/L	10/21/25 10:06	VX102125
74-83-9	Bromomethane	19.3		1	1.40	3.80	5.00	ug/L	10/21/25 10:06	VX102125
75-00-3	Chloroethane	18.9		1	0.47	0.75	1.00	ug/L	10/21/25 10:06	VX102125
75-69-4	Trichlorofluoromethane	17.7		1	0.33	0.50	1.00	ug/L	10/21/25 10:06	VX102125
76-13-1	1,1,2-Trichlorotrifluoroethane	17.6		1	0.25	0.50	1.00	ug/L	10/21/25 10:06	VX102125
75-35-4	1,1-Dichloroethene	18.0		1	0.23	0.75	1.00	ug/L	10/21/25 10:06	VX102125
67-64-1	Acetone	82.3		1	1.50	3.80	5.00	ug/L	10/21/25 10:06	VX102125
75-15-0	Carbon Disulfide	16.6		1	0.21	0.75	1.00	ug/L	10/21/25 10:06	VX102125
1634-04-4	Methyl tert-butyl Ether	17.6		1	0.16	0.50	1.00	ug/L	10/21/25 10:06	VX102125
79-20-9	Methyl Acetate	17.5		1	0.27	0.75	1.00	ug/L	10/21/25 10:06	VX102125
75-09-2	Methylene Chloride	19.1		1	0.28	0.50	1.00	ug/L	10/21/25 10:06	VX102125
156-60-5	trans-1,2-Dichloroethene	17.9		1	0.23	0.50	1.00	ug/L	10/21/25 10:06	VX102125
75-34-3	1,1-Dichloroethane	18.4		1	0.23	0.50	1.00	ug/L	10/21/25 10:06	VX102125
110-82-7	Cyclohexane	19.2		1	1.50	2.50	5.00	ug/L	10/21/25 10:06	VX102125
78-93-3	2-Butanone	85.2		1	0.98	2.50	5.00	ug/L	10/21/25 10:06	VX102125
56-23-5	Carbon Tetrachloride	17.0		1	0.25	0.50	1.00	ug/L	10/21/25 10:06	VX102125
156-59-2	cis-1,2-Dichloroethene	18.9		1	0.19	0.75	1.00	ug/L	10/21/25 10:06	VX102125
74-97-5	Bromochloromethane	21.6		1	0.22	0.50	1.00	ug/L	10/21/25 10:06	VX102125
67-66-3	Chloroform	18.8		1	0.25	0.50	1.00	ug/L	10/21/25 10:06	VX102125
71-55-6	1,1,1-Trichloroethane	18.1		1	0.20	0.50	1.00	ug/L	10/21/25 10:06	VX102125
108-87-2	Methylcyclohexane	17.0		1	0.16	0.50	1.00	ug/L	10/21/25 10:06	VX102125
71-43-2	Benzene	18.5		1	0.15	0.50	1.00	ug/L	10/21/25 10:06	VX102125
107-06-2	1,2-Dichloroethane	19.1		1	0.22	0.50	1.00	ug/L	10/21/25 10:06	VX102125
79-01-6	Trichloroethene	18.1		1	0.090	0.75	1.00	ug/L	10/21/25 10:06	VX102125
78-87-5	1,2-Dichloropropane	19.3		1	0.20	0.50	1.00	ug/L	10/21/25 10:06	VX102125
75-27-4	Bromodichloromethane	18.6		1	0.22	0.50	1.00	ug/L	10/21/25 10:06	VX102125
108-10-1	4-Methyl-2-Pentanone	86.9		1	0.68	2.50	5.00	ug/L	10/21/25 10:06	VX102125
108-88-3	Toluene	18.8		1	0.14	0.50	1.00	ug/L	10/21/25 10:06	VX102125
10061-02-6	t-1,3-Dichloropropene	18.4		1	0.17	0.50	1.00	ug/L	10/21/25 10:06	VX102125
10061-01-5	cis-1,3-Dichloropropene	18.8		1	0.16	0.50	1.00	ug/L	10/21/25 10:06	VX102125
79-00-5	1,1,2-Trichloroethane	19.1		1	0.21	0.50	1.00	ug/L	10/21/25 10:06	VX102125
591-78-6	2-Hexanone	85.9		1	0.89	2.50	5.00	ug/L	10/21/25 10:06	VX102125
124-48-1	Dibromochloromethane	18.1		1	0.18	0.50	1.00	ug/L	10/21/25 10:06	VX102125
106-93-4	1,2-Dibromoethane	18.6		1	0.15	0.50	1.00	ug/L	10/21/25 10:06	VX102125
127-18-4	Tetrachloroethene	16.9		1	0.23	0.50	1.00	ug/L	10/21/25 10:06	VX102125
108-90-7	Chlorobenzene	18.0		1	0.12	0.50	1.00	ug/L	10/21/25 10:06	VX102125
100-41-4	Ethyl Benzene	17.8		1	0.13	0.50	1.00	ug/L	10/21/25 10:06	VX102125
179601-23-1	m/p-Xylenes	36.8		1	0.24	1.00	2.00	ug/L	10/21/25 10:06	VX102125
95-47-6	o-Xylene	18.1		1	0.12	0.50	1.00	ug/L	10/21/25 10:06	VX102125
100-42-5	Styrene	18.8		1	0.15	0.50	1.00	ug/L	10/21/25 10:06	VX102125
75-25-2	Bromoform	17.0		1	0.19	0.50	1.00	ug/L	10/21/25 10:06	VX102125

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1021WBS01
Lab Sample ID: VX1021WBS01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	17.1		1	0.12	0.50	1.00	ug/L	10/21/25 10:06	VX102125
79-34-5	1,1,2,2-Tetrachloroethane	17.3		1	0.26	0.50	1.00	ug/L	10/21/25 10:06	VX102125
541-73-1	1,3-Dichlorobenzene	17.2		1	0.16	0.50	1.00	ug/L	10/21/25 10:06	VX102125
106-46-7	1,4-Dichlorobenzene	16.6		1	0.19	0.50	1.00	ug/L	10/21/25 10:06	VX102125
95-50-1	1,2-Dichlorobenzene	17.2		1	0.16	0.50	1.00	ug/L	10/21/25 10:06	VX102125
96-12-8	1,2-Dibromo-3-Chloropropane	14.2		1	0.53	0.75	1.00	ug/L	10/21/25 10:06	VX102125
120-82-1	1,2,4-Trichlorobenzene	15.2		1	0.20	0.50	1.00	ug/L	10/21/25 10:06	VX102125
87-61-6	1,2,3-Trichlorobenzene	15.5		1	0.20	0.75	1.00	ug/L	10/21/25 10:06	VX102125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	49.2			81 - 118	98%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3			80 - 119	99%	SPK: 50
2037-26-5	Toluene-d8	47.8			89 - 112	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.7			85 - 114	99%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	141000
540-36-3	1,4-Difluorobenzene	230000
3114-55-4	Chlorobenzene-d5	212000
3855-82-1	1,4-Dichlorobenzene-d4	111000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1021WBSD01
Lab Sample ID: VX1021WBSD01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-71-8	Dichlorodifluoromethane	18.0		1	0.22	0.50	1.00	ug/L	10/21/25 10:33	VX102125
74-87-3	Chloromethane	17.3		1	0.32	0.50	1.00	ug/L	10/21/25 10:33	VX102125
75-01-4	Vinyl Chloride	16.9		1	0.26	0.75	1.00	ug/L	10/21/25 10:33	VX102125
74-83-9	Bromomethane	19.0		1	1.40	3.80	5.00	ug/L	10/21/25 10:33	VX102125
75-00-3	Chloroethane	19.0		1	0.47	0.75	1.00	ug/L	10/21/25 10:33	VX102125
75-69-4	Trichlorofluoromethane	17.9		1	0.33	0.50	1.00	ug/L	10/21/25 10:33	VX102125
76-13-1	1,1,2-Trichlorotrifluoroethane	17.7		1	0.25	0.50	1.00	ug/L	10/21/25 10:33	VX102125
75-35-4	1,1-Dichloroethene	17.9		1	0.23	0.75	1.00	ug/L	10/21/25 10:33	VX102125
67-64-1	Acetone	81.6		1	1.50	3.80	5.00	ug/L	10/21/25 10:33	VX102125
75-15-0	Carbon Disulfide	16.9		1	0.21	0.75	1.00	ug/L	10/21/25 10:33	VX102125
1634-04-4	Methyl tert-butyl Ether	18.0		1	0.16	0.50	1.00	ug/L	10/21/25 10:33	VX102125
79-20-9	Methyl Acetate	16.9		1	0.27	0.75	1.00	ug/L	10/21/25 10:33	VX102125
75-09-2	Methylene Chloride	18.5		1	0.28	0.50	1.00	ug/L	10/21/25 10:33	VX102125
156-60-5	trans-1,2-Dichloroethene	18.8		1	0.23	0.50	1.00	ug/L	10/21/25 10:33	VX102125
75-34-3	1,1-Dichloroethane	18.8		1	0.23	0.50	1.00	ug/L	10/21/25 10:33	VX102125
110-82-7	Cyclohexane	18.4		1	1.50	2.50	5.00	ug/L	10/21/25 10:33	VX102125
78-93-3	2-Butanone	85.7		1	0.98	2.50	5.00	ug/L	10/21/25 10:33	VX102125
56-23-5	Carbon Tetrachloride	17.2		1	0.25	0.50	1.00	ug/L	10/21/25 10:33	VX102125
156-59-2	cis-1,2-Dichloroethene	19.1		1	0.19	0.75	1.00	ug/L	10/21/25 10:33	VX102125
74-97-5	Bromochloromethane	21.3		1	0.22	0.50	1.00	ug/L	10/21/25 10:33	VX102125
67-66-3	Chloroform	19.0		1	0.25	0.50	1.00	ug/L	10/21/25 10:33	VX102125
71-55-6	1,1,1-Trichloroethane	18.3		1	0.20	0.50	1.00	ug/L	10/21/25 10:33	VX102125
108-87-2	Methylcyclohexane	18.2		1	0.16	0.50	1.00	ug/L	10/21/25 10:33	VX102125
71-43-2	Benzene	18.8		1	0.15	0.50	1.00	ug/L	10/21/25 10:33	VX102125
107-06-2	1,2-Dichloroethane	18.5		1	0.22	0.50	1.00	ug/L	10/21/25 10:33	VX102125
79-01-6	Trichloroethene	18.3		1	0.090	0.75	1.00	ug/L	10/21/25 10:33	VX102125
78-87-5	1,2-Dichloropropane	19.8		1	0.20	0.50	1.00	ug/L	10/21/25 10:33	VX102125
75-27-4	Bromodichloromethane	18.8		1	0.22	0.50	1.00	ug/L	10/21/25 10:33	VX102125
108-10-1	4-Methyl-2-Pentanone	89.0		1	0.68	2.50	5.00	ug/L	10/21/25 10:33	VX102125
108-88-3	Toluene	18.9		1	0.14	0.50	1.00	ug/L	10/21/25 10:33	VX102125
10061-02-6	t-1,3-Dichloropropene	18.5		1	0.17	0.50	1.00	ug/L	10/21/25 10:33	VX102125
10061-01-5	cis-1,3-Dichloropropene	18.5		1	0.16	0.50	1.00	ug/L	10/21/25 10:33	VX102125
79-00-5	1,1,2-Trichloroethane	19.3		1	0.21	0.50	1.00	ug/L	10/21/25 10:33	VX102125
591-78-6	2-Hexanone	87.8		1	0.89	2.50	5.00	ug/L	10/21/25 10:33	VX102125
124-48-1	Dibromochloromethane	18.3		1	0.18	0.50	1.00	ug/L	10/21/25 10:33	VX102125
106-93-4	1,2-Dibromoethane	18.6		1	0.15	0.50	1.00	ug/L	10/21/25 10:33	VX102125
127-18-4	Tetrachloroethene	17.5		1	0.23	0.50	1.00	ug/L	10/21/25 10:33	VX102125
108-90-7	Chlorobenzene	18.9		1	0.12	0.50	1.00	ug/L	10/21/25 10:33	VX102125
100-41-4	Ethyl Benzene	18.3		1	0.13	0.50	1.00	ug/L	10/21/25 10:33	VX102125
179601-23-1	m/p-Xylenes	37.5		1	0.24	1.00	2.00	ug/L	10/21/25 10:33	VX102125
95-47-6	o-Xylene	18.5		1	0.12	0.50	1.00	ug/L	10/21/25 10:33	VX102125
100-42-5	Styrene	19.2		1	0.15	0.50	1.00	ug/L	10/21/25 10:33	VX102125
75-25-2	Bromoform	17.7		1	0.19	0.50	1.00	ug/L	10/21/25 10:33	VX102125

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1021WBSD01
Lab Sample ID: VX1021WBSD01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	18.5		1	0.12	0.50	1.00	ug/L	10/21/25 10:33	VX102125
79-34-5	1,1,2,2-Tetrachloroethane	18.4		1	0.26	0.50	1.00	ug/L	10/21/25 10:33	VX102125
541-73-1	1,3-Dichlorobenzene	18.2		1	0.16	0.50	1.00	ug/L	10/21/25 10:33	VX102125
106-46-7	1,4-Dichlorobenzene	18.2		1	0.19	0.50	1.00	ug/L	10/21/25 10:33	VX102125
95-50-1	1,2-Dichlorobenzene	18.6		1	0.16	0.50	1.00	ug/L	10/21/25 10:33	VX102125
96-12-8	1,2-Dibromo-3-Chloropropane	16.9		1	0.53	0.75	1.00	ug/L	10/21/25 10:33	VX102125
120-82-1	1,2,4-Trichlorobenzene	16.8		1	0.20	0.50	1.00	ug/L	10/21/25 10:33	VX102125
87-61-6	1,2,3-Trichlorobenzene	17.3		1	0.20	0.75	1.00	ug/L	10/21/25 10:33	VX102125

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	46.5			81 - 118		93%	SPK: 50
1868-53-7	Dibromofluoromethane	48.0			80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	47.9			89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.3			85 - 114		95%	SPK: 50

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
363-72-4	Pentafluorobenzene	138000
540-36-3	1,4-Difluorobenzene	224000
3114-55-4	Chlorobenzene-d5	203000
3855-82-1	1,4-Dichlorobenzene-d4	101000

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>ICES02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3400</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>10:59</u> <u>13:49</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX048229.D		RRF020 = VX048231.D		RRF050 = VX048232.D		
		RRF100 = VX048233.D		RRF150 = VX048234.D		RRF005 = VX048236.D		
COMPOUND	RRF001	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD
Dichlorodifluoromethane	0.522	0.513	0.483	0.540	0.504	0.522	0.514	3.7
Chloromethane	0.506	0.468	0.449	0.476	0.471	0.508	0.480	4.8
Vinyl Chloride	0.447	0.437	0.416	0.462	0.445	0.458	0.444	3.7
Bromomethane		0.323	0.282	0.280	0.276	0.346	0.302	10.4
Chloroethane	0.233	0.277	0.247	0.282	0.267	0.261	0.261	7.2
Trichlorofluoromethane	0.776	0.874	0.807	0.907	0.869	0.880	0.852	5.8
1,1,2-Trichlorotrifluoroethane	0.512	0.540	0.501	0.583	0.549	0.553	0.540	5.5
1,1-Dichloroethene	0.515	0.548	0.485	0.558	0.531	0.514	0.525	5
Acetone	0.300	0.250	0.202	0.221	0.206	0.262	0.240	15.7
Carbon Disulfide	2.034	1.612	1.442	1.617	1.532	1.689	1.654	12.4
Methyl tert-butyl Ether	1.894	1.909	1.704	1.817	1.764	1.657	1.791	5.7
Methyl Acetate	0.514	0.551	0.466	0.514	0.504	0.464	0.502	6.6
Methylene Chloride	0.555	0.651	0.552	0.597	0.567	0.595	0.586	6.4
trans-1,2-Dichloroethene	0.587	0.585	0.535	0.592	0.558	0.603	0.577	4.4
1,1-Dichloroethane	1.116	1.085	0.977	1.055	1.019	1.054	1.051	4.6
Cyclohexane		0.894	0.821	0.926	0.866	0.843	0.870	4.8
2-Butanone	0.311	0.329	0.281	0.301	0.290	0.315	0.304	5.7
Carbon Tetrachloride	0.650	0.594	0.553	0.604	0.567	0.622	0.599	5.9
cis-1,2-Dichloroethene	0.682	0.700	0.626	0.695	0.661	0.673	0.673	4
Bromochloromethane	0.388	0.404	0.441	0.462	0.452	0.469	0.436	7.5
Chloroform	1.147	1.208	1.046	1.127	1.067	1.143	1.123	5.2
1,1,1-Trichloroethane	1.049	1.085	0.973	1.055	1.007	1.041	1.035	3.8
Methylcyclohexane	0.551	0.548	0.540	0.626	0.582	0.590	0.573	5.7
Benzene	1.493	1.480	1.301	1.426	1.347	1.405	1.409	5.3
1,2-Dichloroethane	0.529	0.566	0.500	0.528	0.496	0.499	0.520	5.2
Trichloroethene	0.400	0.373	0.348	0.380	0.363	0.366	0.372	4.7
1,2-Dichloropropane	0.309	0.362	0.321	0.347	0.333	0.337	0.335	5.5
Bromodichloromethane	0.570	0.592	0.538	0.573	0.538	0.567	0.563	3.8
4-Methyl-2-Pentanone	0.366	0.422	0.377	0.384	0.362	0.374	0.381	5.7
Toluene	0.881	0.944	0.862	0.910	0.849	0.925	0.895	4.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>ICES02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3400</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>10:59</u> <u>13:49</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX048229.D		RRF020 = VX048231.D		RRF050 = VX048232.D		
		RRF100 = VX048233.D		RRF150 = VX048234.D		RRF005 = VX048236.D		
COMPOUND	RRF001	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD
t-1,3-Dichloropropene	0.477	0.582	0.545	0.574	0.553	0.526	0.543	7
cis-1,3-Dichloropropene	0.582	0.604	0.572	0.615	0.585	0.556	0.586	3.7
1,1,2-Trichloroethane	0.318	0.357	0.321	0.333	0.313	0.355	0.333	5.8
2-Hexanone	0.231	0.298	0.267	0.271	0.256	0.271	0.266	8.3
Dibromochloromethane	0.482	0.476	0.425	0.442	0.418	0.441	0.447	5.9
1,2-Dibromoethane	0.346	0.393	0.342	0.359	0.336	0.353	0.355	5.7
Tetrachloroethene	0.418	0.378	0.328	0.370	0.361	0.385	0.373	7.9
Chlorobenzene	1.149	1.159	1.045	1.148	1.102	1.145	1.125	3.9
Ethyl Benzene	1.905	1.954	1.825	2.008	1.922	1.876	1.915	3.3
m/p-Xylenes	0.647	0.750	0.687	0.750	0.707	0.727	0.712	5.6
o-Xylene	0.652	0.712	0.664	0.723	0.686	0.684	0.687	3.9
Styrene	1.054	1.236	1.143	1.247	1.173	1.165	1.170	6
Bromoform	0.338	0.350	0.308	0.335	0.326	0.325	0.330	4.3
Isopropylbenzene	3.618	3.656	3.430	3.875	3.754	3.579	3.652	4.2
1,1,2,2-Tetrachloroethane	0.988	1.065	0.946	1.018	0.992	0.974	0.997	4.1
1,3-Dichlorobenzene	1.863	1.742	1.550	1.747	1.707	1.750	1.726	5.9
1,4-Dichlorobenzene	2.143	1.803	1.577	1.724	1.710	1.772	1.788	10.6
1,2-Dichlorobenzene	1.803	1.631	1.494	1.619	1.588	1.590	1.621	6.3
1,2-Dibromo-3-Chloropropane	0.167	0.219	0.195	0.224	0.223	0.187	0.202	11.5
1,2,4-Trichlorobenzene	1.441	0.975	0.959	1.114	1.094	0.994	1.096	16.5
1,2,3-Trichlorobenzene	1.154	0.942	0.909	1.050	1.033	0.947	1.006	9
1,2-Dichloroethane-d4		0.709	0.682	0.690	0.698	0.725	0.701	2.4
Dibromofluoromethane		0.345	0.330	0.346	0.348	0.338	0.341	2.2
Toluene-d8		1.185	1.185	1.196	1.195	1.221	1.196	1.2
4-Bromofluorobenzene		0.445	0.444	0.446	0.441	0.500	0.455	5.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3400</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 08:43</u>
Lab File ID:	<u>VX048252.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.514	0.496		-3.5	20
Chloromethane	0.480	0.468	0.1	-2.5	20
Vinyl Chloride	0.444	0.448		0.9	20
Bromomethane	0.302	0.299		-0.99	20
Chloroethane	0.261	0.261		0	20
Trichlorofluoromethane	0.852	0.813		-4.58	20
1,1,2-Trichlorotrifluoroethane	0.540	0.513		-5	20
1,1-Dichloroethene	0.525	0.498		-5.14	20
Acetone	0.240	0.230		-4.17	20
Carbon Disulfide	1.654	1.439		-13	20
Methyl tert-butyl Ether	1.791	1.576		-12	20
Methyl Acetate	0.502	0.411		-18.13	20
Methylene Chloride	0.586	0.556		-5.12	20
trans-1,2-Dichloroethene	0.577	0.538		-6.76	20
1,1-Dichloroethane	1.051	0.952	0.1	-9.42	20
Cyclohexane	0.870	0.788		-9.43	20
2-Butanone	0.304	0.260		-14.47	20
Carbon Tetrachloride	0.599	0.533		-11.02	20
cis-1,2-Dichloroethene	0.673	0.610		-9.36	20
Bromochloromethane	0.436	0.435		-0.23	20
Chloroform	1.123	1.006		-10.42	20
1,1,1-Trichloroethane	1.035	0.920		-11.11	20
Methylcyclohexane	0.573	0.510		-10.99	20
Benzene	1.409	1.280		-9.15	20
1,2-Dichloroethane	0.520	0.466		-10.39	20
Trichloroethene	0.372	0.337		-9.41	20
1,2-Dichloropropane	0.335	0.305		-8.95	20
Bromodichloromethane	0.563	0.510		-9.41	20
4-Methyl-2-Pentanone	0.381	0.302		-20.74	20
Toluene	0.895	0.795		-11.17	20
t-1,3-Dichloropropene	0.543	0.493		-9.21	20
cis-1,3-Dichloropropene	0.586	0.529		-9.73	20
1,1,2-Trichloroethane	0.333	0.291		-12.61	20
2-Hexanone	0.266	0.227		-14.66	20
Dibromochloromethane	0.447	0.388		-13.2	20
1,2-Dibromoethane	0.355	0.307		-13.52	20
Tetrachloroethene	0.373	0.325		-12.87	20
Chlorobenzene	1.125	1.025	0.3	-8.89	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3400</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 08:43</u>
Lab File ID:	<u>VX048252.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.915	1.738		-9.24	20
m/p-Xylenes	0.712	0.666		-6.46	20
o-Xylene	0.687	0.625		-9.02	20
Styrene	1.170	1.107		-5.39	20
Bromoform	0.330	0.283	0.1	-14.24	20
Isopropylbenzene	3.652	3.379		-7.47	20
1,1,2,2-Tetrachloroethane	0.997	0.857	0.3	-14.04	20
1,3-Dichlorobenzene	1.726	1.555		-9.91	20
1,4-Dichlorobenzene	1.788	1.551		-13.26	20
1,2-Dichlorobenzene	1.621	1.416		-12.65	20
1,2-Dibromo-3-Chloropropane	0.202	0.163		-19.31	20
1,2,4-Trichlorobenzene	1.096	0.885		-19.25	20
1,2,3-Trichlorobenzene	1.006	0.834		-17.1	20
1,2-Dichloroethane-d4	0.701	0.658		-6.13	20
Dibromofluoromethane	0.341	0.340		-0.29	20
Toluene-d8	1.196	1.150		-3.85	20
4-Bromofluorobenzene	0.455	0.419		-7.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3400</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 18:09</u>
Lab File ID:	<u>VX048278.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.514	0.469		-8.76	50
Chloromethane	0.480	0.395	0.1	-17.71	50
Vinyl Chloride	0.444	0.366		-17.57	50
Bromomethane	0.302	0.258		-14.57	50
Chloroethane	0.261	0.231		-11.49	50
Trichlorofluoromethane	0.852	0.801		-5.99	50
1,1,2-Trichlorotrifluoroethane	0.540	0.479		-11.3	50
1,1-Dichloroethene	0.525	0.490		-6.67	50
Acetone	0.240	0.185		-22.92	50
Carbon Disulfide	1.654	1.375		-16.87	50
Methyl tert-butyl Ether	1.791	1.749		-2.35	50
Methyl Acetate	0.502	0.451		-10.16	50
Methylene Chloride	0.586	0.564		-3.75	50
trans-1,2-Dichloroethene	0.577	0.537		-6.93	50
1,1-Dichloroethane	1.051	0.987	0.1	-6.09	50
Cyclohexane	0.870	0.770		-11.49	50
2-Butanone	0.304	0.262		-13.82	50
Carbon Tetrachloride	0.599	0.547		-8.68	50
cis-1,2-Dichloroethene	0.673	0.652		-3.12	50
Bromochloromethane	0.436	0.416		-4.59	50
Chloroform	1.123	1.104		-1.69	50
1,1,1-Trichloroethane	1.035	1.009		-2.51	50
Methylcyclohexane	0.573	0.489		-14.66	50
Benzene	1.409	1.285		-8.8	50
1,2-Dichloroethane	0.520	0.519		-0.19	50
Trichloroethene	0.372	0.348		-6.45	50
1,2-Dichloropropane	0.335	0.311		-7.16	50
Bromodichloromethane	0.563	0.545		-3.2	50
4-Methyl-2-Pentanone	0.381	0.346		-9.19	50
Toluene	0.895	0.860		-3.91	50
t-1,3-Dichloropropene	0.543	0.540		-0.55	50
cis-1,3-Dichloropropene	0.586	0.561		-4.27	50
1,1,2-Trichloroethane	0.333	0.324		-2.7	50
2-Hexanone	0.266	0.244		-8.27	50
Dibromochloromethane	0.447	0.441		-1.34	50
1,2-Dibromoethane	0.355	0.344		-3.1	50
Tetrachloroethene	0.373	0.316		-15.28	50
Chlorobenzene	1.125	1.027	0.3	-8.71	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3400</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 18:09</u>
Lab File ID:	<u>VX048278.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.915	1.736		-9.35	50
m/p-Xylenes	0.712	0.662		-7.02	50
o-Xylene	0.687	0.645		-6.11	50
Styrene	1.170	1.126		-3.76	50
Bromoform	0.330	0.304	0.1	-7.88	50
Isopropylbenzene	3.652	3.018		-17.36	50
1,1,2,2-Tetrachloroethane	0.997	0.843	0.3	-15.45	50
1,3-Dichlorobenzene	1.726	1.412		-18.19	50
1,4-Dichlorobenzene	1.788	1.440		-19.46	50
1,2-Dichlorobenzene	1.621	1.364		-15.85	50
1,2-Dibromo-3-Chloropropane	0.202	0.166		-17.82	50
1,2,4-Trichlorobenzene	1.096	0.844		-22.99	50
1,2,3-Trichlorobenzene	1.006	0.810		-19.48	50
1,2-Dichloroethane-d4	0.701	0.787		12.27	50
Dibromofluoromethane	0.341	0.360		5.57	50
Toluene-d8	1.196	1.238		3.51	50
4-Bromofluorobenzene	0.455	0.547		20.22	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.