

Continuing Calibration Report for SequenceID : FC102425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC070039.D

Aliphatic C9-C12	8195874.000	60.000	3.278	6.573	136597.900	144785.198	5.655
Aliphatic C12-C16	6186106.000	40.000	6.574	9.974	154652.650	163782.335	5.574
Aliphatic C16-C21	9021881.000	60.000	9.975	13.340	150364.683	161945.502	7.151
Aliphatic C21-C28	10755054.000	80.000	13.341	17.004	134438.175	147142.336	8.634
Aliphatic C28-C40	13211941.000	120.000	17.005	21.957	110099.508	107599.833	-2.323
Aliphatic EPH	47370856.000	360.000	3.278	21.957	131585.711	137884.728	4.568