

Instrument ID: FID_C

Daily Analysis Runlog For Sequence/QC Batch ID # FC100725AL

Review By	yogesh	Review On	10/7/2025 10:33:30 AM
Supervise By	mohammad	Supervise On	10/8/2025 3:35:18 AM
SubDirectory	FC100725AL	HP Acquire Method	HP Processing Method FC100725AL

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773
CCC	PP24771
Internal Standard/PEM	
ICV/I.BLK	PP24769,PP24774
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FC069935.D	06 Oct 2025 14:47		YP/AJ	Ok
2	I.BLK	I.BLK	FC069936.D	06 Oct 2025 15:31		YP/AJ	Ok
3	100 PPM ALIPHATIC HC	100 PPM ALIPHATIC HC	FC069937.D	06 Oct 2025 16:17		YP/AJ	Ok
4	50 PPM ALIPHATIC HC	50 PPM ALIPHATIC HC	FC069938.D	06 Oct 2025 17:03		YP/AJ	Ok
5	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069939.D	06 Oct 2025 17:50		YP/AJ	Ok
6	10 PPM ALIPHATIC HC	10 PPM ALIPHATIC HC	FC069940.D	06 Oct 2025 18:35		YP/AJ	Ok,M
7	5 PPM ALIPHATIC HC	5 PPM ALIPHATIC HC	FC069941.D	06 Oct 2025 19:19		YP/AJ	Ok
8	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069942.D	06 Oct 2025 20:02		YP/AJ	Ok
9	I.BLK	I.BLK	FC069943.D	06 Oct 2025 21:29		YP/AJ	Ok
10	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069944.D	06 Oct 2025 22:12		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_C

Daily Analysis Runlog For Sequence/QC Batch ID # FC102425AL

Review By	yogesh	Review On	10/27/2025 10:56:17 AM
Supervise By		Supervise On	
SubDirectory	FC102425AL	HP Acquire Method	HP Processing Method FC100725AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC	PP24771		
Internal Standard/PEM			
ICV/I.BLK	PP24769,PP24774		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FC070025.D	24 Oct 2025 10:34		YP/AJ	Ok
2	I.BLK	I.BLK	FC070026.D	24 Oct 2025 11:13		YP/AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC070027.D	24 Oct 2025 11:53		YP/AJ	Ok,NS
4	PB170249BL	PB170249BL	FC070028.D	24 Oct 2025 14:00		YP/AJ	Ok
5	PB170249BS	PB170249BS	FC070029.D	24 Oct 2025 14:39		YP/AJ	Ok,NS
6	PB170249BSD	PB170249BSD	FC070030.D	24 Oct 2025 15:20		YP/AJ	Ok,NS
7	Q3410-03	1	FC070031.D	24 Oct 2025 16:02		YP/AJ	Ok,NS
8	Q3410-04	2	FC070032.D	24 Oct 2025 16:44		YP/AJ	Ok
9	Q3410-05	Q3410-05	FC070033.D	24 Oct 2025 17:25	Surrogate fail	YP/AJ	ReRun
10	Q3410-06	Q3410-06	FC070034.D	24 Oct 2025 18:07	Surrogate fail	YP/AJ	ReRun
11	Q3410-07	5	FC070035.D	24 Oct 2025 18:48		YP/AJ	Ok
12	Q3410-07MS	5MS	FC070036.D	24 Oct 2025 19:29	FC070035.D	YP/AJ	Ok,NS
13	Q3410-07MSD	5MSD	FC070037.D	24 Oct 2025 20:10	FC070035.D!FC070036.D	YP/AJ	Ok,NS
14	I.BLK	I.BLK	FC070038.D	24 Oct 2025 21:31		YP/AJ	Ok
15	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC070039.D	24 Oct 2025 22:12		YP/AJ	Ok,NS

M : Manual Integration

Instrument ID: FID_C

Daily Analysis Runlog For Sequence/QC Batch ID # FC102725AL

Review By	yogesh	Review On	10/27/2025 10:57:08 AM
Supervise By		Supervise On	
SubDirectory	FC102725AL	HP Acquire Method	HP Processing Method FC100725AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC	PP24771		
Internal Standard/PEM			
ICV/I.BLK	PP24769,PP24774		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FC070040.D	27 Oct 2025 08:57		YP/AJ	Ok
2	I.BLK	I.BLK	FC070041.D	27 Oct 2025 09:36		YP/AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC070042.D	27 Oct 2025 10:15		YP/AJ	Ok
4	Q3410-05	3	FC070043.D	27 Oct 2025 10:54		YP/AJ	Ok
5	Q3410-06	4	FC070044.D	27 Oct 2025 11:34		YP/AJ	Ok
6	I.BLK	I.BLK	FC070045.D	27 Oct 2025 12:54		YP/AJ	Ok
7	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC070046.D	27 Oct 2025 13:34		YP/AJ	Ok
8	PB170258BL	PB170258BL	FC070047.D	27 Oct 2025 14:15		YP/AJ	Ok
9	PB170258BS	PB170258BS	FC070048.D	27 Oct 2025 14:56		YP/AJ	Ok,NS
10	PB170258BSD	PB170258BSD	FC070049.D	27 Oct 2025 15:37		YP/AJ	Ok
11	Q3451-05	TP-12	FC070050.D	27 Oct 2025 16:17		YP/AJ	Ok
12	Q3451-05D	Q3451-05D	FC070051.D	27 Oct 2025 16:58		YP/AJ	Ok
13	Q3451-05MS	TP-12MS	FC070052.D	27 Oct 2025 17:39	FC070050.D	YP/AJ	Ok,NS
14	Q3451-05MSD	TP-12MSD	FC070053.D	27 Oct 2025 18:20	FC070050.D!FC070052.D	YP/AJ	Ok
15	Q3451-06	TP-12-EPH	FC070054.D	27 Oct 2025 19:01		YP/AJ	Ok
16	I.BLK	I.BLK	FC070055.D	27 Oct 2025 20:23		YP/AJ	Ok
17	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC070056.D	27 Oct 2025 21:03		YP/AJ	Ok

M : Manual Integration