

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

ProjectID: 49 Moore Place Belleville

Lab Code: ACE

SDG No.: Q3410

Calibration Sequence : FG100225		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	187365146	110215	FG016731.D
850	99027342	116503	FG016732.D
340	42723530	125657	FG016733.D
170	22222689	130722	FG016734.D
85	10606021	124777	FG016735.D
AVG RF : 121575		% RSD : 6.699	AVG RT : 15.1602