

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100725

Review By	Semsettin Yesilyurt	Review On	10/10/2025 3:00:22 AM
Supervise By	Maresh Dadoda	Supervise On	10/10/2025 4:38:02 AM
SubDirectory	VY100725	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135766 VP135767,VP135768,VP135769,VP135770,VP135771,VP135772
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133934 VP135773

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023383.D	07 Oct 2025 08:43		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023384.D	07 Oct 2025 09:17		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023385.D	07 Oct 2025 10:02		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023386.D	07 Oct 2025 11:24		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023387.D	07 Oct 2025 11:47		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023388.D	07 Oct 2025 12:09		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023389.D	07 Oct 2025 12:32		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023390.D	07 Oct 2025 13:02		SY/MD	Ok
9	VSTDICV050	ICVVY100725	VY023391.D	07 Oct 2025 14:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102125

Review By	John Carlone	Review On	10/24/2025 2:07:04 PM
Supervise By		Supervise On	
SubDirectory	VY102125	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135871
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135872,VP135873 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023520.D	21 Oct 2025 08:29		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023521.D	21 Oct 2025 09:05		SY/MD	Ok,NS
3	VY1021SBL01	VY1021SBL01	VY023522.D	21 Oct 2025 09:34		SY/MD	Ok
4	VY1021SBS01	VY1021SBS01	VY023523.D	21 Oct 2025 10:04		SY/MD	Ok,NS
5	VY1021SBSD01	VY1021SBSD01	VY023524.D	21 Oct 2025 10:27		SY/MD	Ok,NS
6	Q3383-01	OR-03-101725	VY023525.D	21 Oct 2025 11:03	vial-B	SY/MD	Ok
7	Q3382-01	CENTRAL-AVE-TP	VY023526.D	21 Oct 2025 11:26	vial-B Internal Standard Fail	SY/MD	Ok
8	Q3386-01RE	237RE	VY023527.D	21 Oct 2025 11:50	vial-B Internal Standard Fail	SY/MD	Confirms
9	Q3376-02	PR132-DP01-20251016	VY023528.D	21 Oct 2025 12:14	vial-B	SY/MD	Ok
10	Q3393-03	PR132-S08-013038-20	VY023529.D	21 Oct 2025 12:38	vial-B	SY/MD	Ok
11	Q3402-01	41-LA	VY023530.D	21 Oct 2025 13:04	vial-A	SY/MD	Ok
12	Q3402-04	41-RA	VY023531.D	21 Oct 2025 13:48	vial-A	SY/MD	Ok
13	Q3402-07	41-LB	VY023532.D	21 Oct 2025 14:12	vial-A	SY/MD	Ok
14	Q3402-10	41-RB	VY023533.D	21 Oct 2025 14:35	vial-A	SY/MD	Ok
15	Q3399-03	COMP-ROLLOFF	VY023534.D	21 Oct 2025 14:59	vial-A	SY/MD	Ok
16	Q3410-02	VOC	VY023535.D	21 Oct 2025 15:22	vial-A	SY/MD	Ok
17	Q3403-03	MH-18-VOC	VY023536.D	21 Oct 2025 15:47	vial-A	SY/MD	Ok
18	Q3406-01	ARS20-0037	VY023537.D	21 Oct 2025 16:10	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102125

Review By	John Carlone	Review On	10/24/2025 2:07:04 PM
Supervise By		Supervise On	
SubDirectory	VY102125	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y100725s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135871		
CCC	VP135872,VP135873		
Internal Standard/PEM	VP133934		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3413-01	BP-VPB-184-GW-980-S	VY023538.D	21 Oct 2025 16:34	vial-A End CCC Missing	SY/MD	Not Ok
20	Q3412-03	TP-6-VOC	VY023539.D	21 Oct 2025 16:57	vial-A	SY/MD	Ok
21	Q3404-01	MOO-25-0285-0288	VY023540.D	21 Oct 2025 17:21	vial-A Internal standard fail	SY/MD	ReRun
22	Q3395-03MS	PR132-S07-000102-20	VY023541.D	21 Oct 2025 17:44	vial-B Internal Standard Fail	SY/MD	Not Ok

M : Manual Integration