

Surrogate Summary

SDG No.: Q3410

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3410-02	VOC	1,2-Dichloroethane-d4	50	51.4	103		63	155
		Dibromofluoromethane	50	51.0	102		70	134
		Toluene-d8	50	48.2	96		74	123
		4-Bromofluorobenzene	50	39.1	78		17	146
VY1021SBL01	VY1021SBL01	1,2-Dichloroethane-d4	50	46.0	92		63	155
		Dibromofluoromethane	50	49.7	99		70	134
		Toluene-d8	50	48.1	96		74	123
		4-Bromofluorobenzene	50	36.4	73		17	146
VY1021SBS01	VY1021SBS01	1,2-Dichloroethane-d4	50	46.9	94		63	155
		Dibromofluoromethane	50	50.9	102		70	134
		Toluene-d8	50	51.3	103		74	123
		4-Bromofluorobenzene	50	49.9	100		17	146
VY1021SBSD01	VY1021SBSD01	1,2-Dichloroethane-d4	50	45.6	91		63	155
		Dibromofluoromethane	50	49.6	99		70	134
		Toluene-d8	50	49.7	99		74	123
		4-Bromofluorobenzene	50	47.6	95		17	146