

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Nickel	1910	2000	95	95 - 105	P	10/21/2025	12:40	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Nickel	38.9	40.0	97	80 - 120	P	10/21/2025	12:44	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Nickel	2410	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
CCV02	Nickel	2330	2500	93	90 - 110	P	10/21/2025	15:03	LB137613
CCV03	Nickel	2350	2500	94	90 - 110	P	10/21/2025	15:56	LB137613
CCV04	Nickel	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
CCV05	Nickel	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
CCV06	Nickel	2300	2500	92	90 - 110	P	10/21/2025	18:27	LB137613
CCV07	Nickel	2350	2500	94	90 - 110	P	10/21/2025	19:18	LB137613
CCV08	Nickel	2350	2500	94	90 - 110	P	10/21/2025	20:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Nickel	1940	2000	97	95 - 105	P	10/23/2025	13:05	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Nickel	39.8	40.0	100	80 - 120	P	10/23/2025	13:14	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Nickel	2580	2500	103	90 - 110	P	10/23/2025	13:52	LB137630
CCV02	Nickel	2610	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
CCV03	Nickel	2490	2500	100	90 - 110	P	10/23/2025	15:56	LB137630
CCV04	Nickel	2490	2500	100	90 - 110	P	10/23/2025	16:43	LB137630
CCV05	Nickel	2610	2500	104	90 - 110	P	10/23/2025	17:21	LB137630
CCV06	Nickel	2630	2500	105	90 - 110	P	10/23/2025	18:11	LB137630
CCV07	Nickel	2490	2500	100	90 - 110	P	10/23/2025	18:59	LB137630
CCV08	Nickel	2620	2500	105	90 - 110	P	10/23/2025	19:48	LB137630
CCV09	Nickel	2650	2500	106	90 - 110	P	10/23/2025	20:39	LB137630
CCV10	Nickel	2510	2500	100	90 - 110	P	10/23/2025	21:30	LB137630
CCV11	Nickel	2530	2500	101	90 - 110	P	10/23/2025	22:20	LB137630
CCV12	Nickel	2470	2500	99	90 - 110	P	10/23/2025	22:33	LB137630