



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Nickel	38.6	40.0	96	65 - 135	P	10/21/2025	12:54	LB137613
CRI01	Nickel	38.9	40.0	97	65 - 135	P	10/23/2025	13:24	LB137630