



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	12:50	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	13:54	LB137613
CCB02	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	15:07	LB137613
CCB03	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	16:00	LB137613
CCB04	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	16:52	LB137613
CCB05	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	17:41	LB137613
CCB06	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	18:31	LB137613
CCB07	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	19:22	LB137613
CCB08	Nickel	3.06	+/-20	U	40.0	P	10/21/2025	20:07	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Nickel	6.19	+/-20	J	40.0	P	10/23/2025	13:20	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	13:57	LB137630
CCB02	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	15:08	LB137630
CCB03	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	16:11	LB137630
CCB04	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	16:47	LB137630
CCB05	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	17:25	LB137630
CCB06	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	18:15	LB137630
CCB07	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	19:04	LB137630
CCB08	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	19:52	LB137630
CCB09	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	20:43	LB137630
CCB10	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	21:34	LB137630
CCB11	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	22:24	LB137630
CCB12	Nickel	3.06	+/-20	U	40.0	P	10/23/2025	22:37	LB137630