

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: METEM A GE POWER Business

SDG No.: Q3411

Contract: METE01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Nickel	3.80	2.0	190	-38	42	10/21/2025	12:59	LB137613
ICSAB01	Nickel	973	954	102	810	1100	10/21/2025	13:03	LB137613
ICSA	Nickel	4.75	2.0	238	-38	42	10/21/2025	13:07	LB137613
ICSAB	Nickel	914	954	96	810	1100	10/21/2025	13:11	LB137613
ICSA01	Nickel	3.96	2.0	198	-38	42	10/23/2025	13:28	LB137630
ICSAB01	Nickel	993	954	104	810	1100	10/23/2025	13:39	LB137630