

metals
- 5a -
MATRIX SPIKE SUMMARY

client: METEM A GE POWER Business **level:** low **sdg no.:** Q3411
contract: METE01 **lab code:** ACE
matrix: Water **sample id:** Q3385-04 **client id:** EFF-WWMS
Percent Solids for Sample: NA **Spiked ID:** Q3385-05 **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Nickel	ug/L	75 - 125	257		2.75	J	250	102		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: METEM A GE POWER Business **level:** low **sdg no.:** Q3411
contract: METE01 **lab code:** ACE
matrix: Water **sample id:** Q3385-04 **client id:** EFF-WWMSD
Percent Solids for Sample: NA **Spiked ID:** Q3385-06 **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Nickel	ug/L	75 - 125	275		2.75	J	250	109		P