

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>METEM A GE POWER Business</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3411</u>
Contract:	<u>METE01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3385-04</u>	Client ID:	<u>EFF-WWDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3385-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Nickel	ug/L	20	2.75	J	2.70	J	2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>METEM A GE POWER Business</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3411</u>
Contract:	<u>METE01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3385-05</u>	Client ID:	<u>EFF-WWMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3385-06	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Nickel	ug/L	20	257		275		7		P