

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3413</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 18:09</u>
Lab File ID:	<u>VX048278.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.514	0.469		-8.76	50
Chloromethane	0.480	0.395	0.1	-17.71	50
Vinyl Chloride	0.444	0.366		-17.57	50
Bromomethane	0.302	0.258		-14.57	50
Chloroethane	0.261	0.231		-11.49	50
Trichlorofluoromethane	0.852	0.801		-5.99	50
1,1,2-Trichlorotrifluoroethane	0.540	0.479		-11.3	50
1,1-Dichloroethene	0.525	0.490		-6.67	50
Acetone	0.240	0.185		-22.92	50
Carbon Disulfide	1.654	1.375		-16.87	50
Methyl tert-butyl Ether	1.791	1.749		-2.35	50
Methyl Acetate	0.502	0.451		-10.16	50
Methylene Chloride	0.586	0.564		-3.75	50
trans-1,2-Dichloroethene	0.577	0.537		-6.93	50
1,1-Dichloroethane	1.051	0.987	0.1	-6.09	50
Cyclohexane	0.870	0.770		-11.49	50
2-Butanone	0.304	0.262		-13.82	50
Carbon Tetrachloride	0.599	0.547		-8.68	50
cis-1,2-Dichloroethene	0.673	0.652		-3.12	50
Bromochloromethane	0.436	0.416		-4.59	50
Chloroform	1.123	1.104		-1.69	50
1,1,1-Trichloroethane	1.035	1.009		-2.51	50
Methylcyclohexane	0.573	0.489		-14.66	50
Benzene	1.409	1.285		-8.8	50
1,2-Dichloroethane	0.520	0.519		-0.19	50
Trichloroethene	0.372	0.348		-6.45	50
1,2-Dichloropropane	0.335	0.311		-7.16	50
Bromodichloromethane	0.563	0.545		-3.2	50
4-Methyl-2-Pentanone	0.381	0.346		-9.19	50
Toluene	0.895	0.860		-3.91	50
t-1,3-Dichloropropene	0.543	0.540		-0.55	50
cis-1,3-Dichloropropene	0.586	0.561		-4.27	50
1,1,2-Trichloroethane	0.333	0.324		-2.7	50
2-Hexanone	0.266	0.244		-8.27	50
Dibromochloromethane	0.447	0.441		-1.34	50
1,2-Dibromoethane	0.355	0.344		-3.1	50
Tetrachloroethene	0.373	0.316		-15.28	50
Chlorobenzene	1.125	1.027	0.3	-8.71	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3413</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/21/2025 18:09</u>
Lab File ID:	<u>VX048278.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.915	1.736		-9.35	50
m/p-Xylenes	0.712	0.662		-7.02	50
o-Xylene	0.687	0.645		-6.11	50
Styrene	1.170	1.126		-3.76	50
Bromoform	0.330	0.304	0.1	-7.88	50
Isopropylbenzene	3.652	3.018		-17.36	50
1,1,2,2-Tetrachloroethane	0.997	0.843	0.3	-15.45	50
1,3-Dichlorobenzene	1.726	1.412		-18.19	50
1,4-Dichlorobenzene	1.788	1.440		-19.46	50
1,2-Dichlorobenzene	1.621	1.364		-15.85	50
1,2-Dibromo-3-Chloropropane	0.202	0.166		-17.82	50
1,2,4-Trichlorobenzene	1.096	0.844		-22.99	50
1,2,3-Trichlorobenzene	1.006	0.810		-19.48	50
1,2-Dichloroethane-d4	0.701	0.787		12.27	50
Dibromofluoromethane	0.341	0.360		5.57	50
Toluene-d8	1.196	1.238		3.51	50
4-Bromofluorobenzene	0.455	0.547		20.22	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.