

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3413
Lab File ID: VX048252.D Date Analyzed: 10/21/2025
Instrument ID: MSVOA_X Time Analyzed: 08:43
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144122	5.53	227103	6.74	197016	10.04
UPPER LIMIT	288244	6.031	454206	7.238	394032	10.537
LOWER LIMIT	72061	5.031	113552	6.238	98508	9.537
EPA SAMPLE NO.						
BP-VPB-184-TB-20251017	110202	5.54	219240	6.75	229169	10.04
VX1021WBL01	121915	5.53	240535	6.75	247603	10.04
VX1021WBS01	141199	5.54	230109	6.74	212342	10.04
VX1021WBSD01	137604	5.53	224103	6.74	203364	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.