

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06  
 Lab Code: ACE SDG NO.: Q3413  
 Lab File ID: VY023559.D Date Analyzed: 10/24/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 08:56  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                       | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|-----------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD           | 558901        | 7.71  | 809883        | 8.62  | 712621        | 11.41  |
| UPPER LIMIT           | 1117800       | 8.213 | 1619770       | 9.115 | 1425240       | 11.914 |
| LOWER LIMIT           | 279451        | 7.213 | 404942        | 8.115 | 356311        | 10.914 |
| EPA SAMPLE NO.        |               |       |               |       |               |        |
| BP-VPB-184-GW-980-982 | 812743        | 7.71  | 1310630       | 8.62  | 1133524       | 11.41  |
| VY1024SBL01           | 898566        | 7.71  | 1421897       | 8.62  | 1164716       | 11.41  |
| VY1024SBS01           | 551537        | 7.71  | 799010        | 8.62  | 709717        | 11.41  |
| VY1024SBSD01          | 536786        | 7.71  | 783587        | 8.62  | 686581        | 11.41  |

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.