

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06  
 Lab Code: ACE SDG NO.: Q3413  
 Lab File ID: VY023559.D Date Analyzed: 10/24/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 08:56  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                       | IS4<br>AREA # | RT #   |  |  |  |  |
|-----------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD           | 380279        | 13.346 |  |  |  |  |
| UPPER LIMIT           | 760558        | 13.846 |  |  |  |  |
| LOWER LIMIT           | 190140        | 12.846 |  |  |  |  |
| EPA SAMPLE NO.        |               |        |  |  |  |  |
| BP-VPB-184-GW-980-982 | 486654        | 13.35  |  |  |  |  |
| VY1024SBL01           | 469443        | 13.35  |  |  |  |  |
| VY1024SBS01           | 380713        | 13.35  |  |  |  |  |
| VY1024SBSD01          | 360721        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.