

**DATA PACKAGE
VOLATILE ORGANICS**

PROJECT NAME : AE1-CTY 3.2.Z-PR-DOCK-DES-ODC - 60693795-1719206

AECOM

605 3rd Avenue

29th Floor

New York, NY - 10158

Phone No: 212-973-2900

ORDER ID : Q3419

ATTENTION : Rob Forstner



Laboratory Certification ID # 20012



1) VOLATILES DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For VOC-TCLVOA-10	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) MS/MSD Summary	15
9.3) LCS/LCSD Summary	19
9.4) Method Blank Summary	23
9.5) GS/MS Tune Summary	25
9.6) Internal Standard Area and RT Summary	28
10) Sample Data	32
10.1) PR-132-S16-015094-20251021	33
10.2) PR132-S16-119133-20251021	48
10.3) PR132-S16-094119-20251021	67
11) Calibration Data Summary	89
11.1) Initial Calibration Data	90
11.1.1) VY100725	90
11.2) Continued Calibration Data	464
11.2.1) VY023543.D	464
11.2.2) VY023559.D	524
12) QC Sample Data	584
12.1) Tune Raw Data	585
12.2) Method Blank Data	588
12.3) LCS Data	612
12.4) MS Data	720
12.5) MSD Data	774
13) Manual Integration	828
14) Analytical Runlogs	832
15) Percent Solid	840
16) Standard Prep Logs	842

Table Of Contents for Q3419

17) VOC Preservation Log	921
18) Shipping Document	922
18.1) Chain Of Custody	923
18.2) Lab Certificate	924
18.3) Internal COC	925

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q3419

Project ID : AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Client : AECOM

Lab Sample Number

Q3419-01
Q3419-02
Q3419-03
Q3419-04
Q3419-05

Client Sample Number

PR-132-S16-015094-20251021
PR-132-S16-015094-20251021MS
PR-132-S16-015094-20251021MSD
PR132-S16-119133-20251021
PR132-S16-094119-20251021

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:54 am, Nov 03, 2025

Date: 10/31/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

AECOM

Project Name: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

Project # N/A

Order ID # Q3419

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

5 Solid samples were received on 10/21/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for PR-132-S16-015094-20251021MSD [Toluene-d8 - 69%] due to matrix interference.

The Internal Standards Areas were met for all analysis except for PR-132-S16-015094-20251021MS, PR-132-S16-015094-20251021MSD samples were re-analyzed but still we have internal standard recovery failure due to matrix interference, lab will report first analysis with failure.

The Retention Times were met for all analysis.

The MS {Q3419-02MS} with File ID: VY023580.D recoveries met the requirements for all compounds except for 1,1,2,2-Tetrachloroethane[290%], 1,1,2-Trichloroethane[150%], 1,2-Dibromo-3-Chloropropane[240%], 4-Methyl-2-Pentanone[175%], Acetone[179%], Bromoform[150%], cis-1,3-Dichloropropene[53%] and Methyl Acetate[360%] due to matrix interference.

The MSD {Q3419-03MSD} with File ID: VY023581.D recoveries met the requirements for all compounds except for 1,1,2,2-Tetrachloroethane[354%], 1,1,2-Trichloroethane[177%], 1,2-Dibromo-3-Chloropropane[316%], 1,2-Dibromoethane[152%], 2-Butanone[249%], 2-Hexanone[238%], 4-Methyl-2-Pentanone[251%], Acetone[298%], Bromoform[190%], Carbon disulfide[142%], cis-1,3-

Dichloropropene[30%], Methyl Acetate[544%], Methyl tert-butyl Ether[190%] and t-1,3-Dichloropropene[43%] due to matrix interference.

The RPD for {Q3419-03MSD} with File ID: VY023581.D met criteria except for 1,2,3-Trichlorobenzene[29%], 1,2,4-Trichlorobenzene[22%], 1,2-Dibromo-3-Chloropropane[27%], 2-Butanone[37%], 2-Hexanone[38%], 4-Methyl-2-Pentanone[36%], Acetone[50%], Bromochloromethane[33%], Bromoform[23%], Bromomethane[29%], Carbon disulfide[45%], cis-1,3-Dichloropropene[54%], Methyl Acetate[41%], Methyl tert-butyl Ether[23%] and t-1,3-Dichloropropene[39%] due to difference in results of MS and MSD.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

The Sample #PR-132-S16-015094-20251021, PR132-S16-119133-20251021 and PR132-S16-094119-20251021 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:55 am, Nov 03, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3419

MATRIX: Solid

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

The Surrogate recoveries were met for all analysis except for PR-132-S16-015094-20251021MSD [Toluene-d8 - 69%] due to matrix interference.


GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS {Q3419-02MS} with File ID: VY023580.D recoveries met the requirements for all compounds except for 1,1,2,2-Tetrachloroethane[290%], 1,1,2-Trichloroethane[150%], 1,2-Dibromo-3-Chloropropane[240%], 4-Methyl-2-Pentanone[175%], Acetone[179%], Bromoform[150%], cis-1,3-Dichloropropene[53%] and Methyl Acetate[360%] due to matrix interference.			
The MSD {Q3419-03MSD} with File ID: VY023581.D recoveries met the requirements for all compounds except for 1,1,2,2-Tetrachloroethane[354%], 1,1,2-Trichloroethane[177%], 1,2-Dibromo-3-Chloropropane[316%], 1,2-Dibromoethane[152%], 2-Butanone[249%], 2-Hexanone[238%], 4-Methyl-2-Pentanone[251%], Acetone[298%], Bromoform[190%], Carbon disulfide[142%], cis-1,3-Dichloropropene[30%], Methyl Acetate[544%], Methyl tert-butyl Ether[190%] and t-1,3-Dichloropropene[43%] due to matrix interference.			
The RPD for {Q3419-03MSD} with File ID: VY023581.D met criteria except for 1,2,3-Trichlorobenzene[29%], 1,2,4-Trichlorobenzene[22%], 1,2-Dibromo-3-Chloropropane[27%], 2-Butanone[37%], 2-Hexanone[38%], 4-Methyl-2-Pentanone[36%], Acetone[49%], Bromochloromethane[33%], Bromoform[23%], Bromomethane[29%], Carbon disulfide[45%], cis-1,3-Dichloropropene[54%], Methyl Acetate[50%], Methyl tert-butyl Ether[23%] and t-1,3-Dichloropropene[39%] due to difference in results of MS and MSD..			
9. Internal Standard Area/Retention Time Shift Meet Criteria		✓	
Comments: The Internal Standards Areas were met for all analysis except for PR-132-S16-015094-20251021MS, PR-132-S16-015094-20251021MSD samples were re-analyzed but still we have internal standard recovery failure due to matrix interference, lab will report first analysis with failure.			
10. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The Sample #PR-132-S16-015094-20251021, PR132-S16-119133-20251021 and PR132-S16-094119-20251021 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.
Trip Blank was not provided with this set of samples.
The soil samples results are based on a dry weight basis.

QA REVIEW



APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3419

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/31/2025

LAB CHRONICLE

OrderID:	Q3419	OrderDate:	10/22/2025 11:01:00 AM
Client:	AECOM	Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Contact:	Rob Forstner	Location:	D41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3419-01	PR-132-S16-015094-2 0251021	SOIL			10/21/25			10/21/25
			VOC-TCLVOA-10	8260D			10/22/25	
Q3419-04	PR132-S16-119133-2 0251021	SOIL			10/21/25			10/21/25
			VOC-TCLVOA-10	8260D			10/22/25	
Q3419-05	PR132-S16-094119-2 0251021	SOIL			10/21/25			10/21/25
			VOC-TCLVOA-10	8260D			10/22/25	

Hit Summary Sheet
SW-846

SDG No.: Q3419
Client: AECOM

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: PR-132-S16-015094-20251021								
Q3419-01	PR-132-S16-015094	SOIL	Acetone	36.3	J	7.60	39.9	ug/Kg
Q3419-01	PR-132-S16-015094	SOIL	Carbon Disulfide	7.80	J	1.70	8.00	ug/Kg
			Total Voc :	44.1				
			Total Concentration:	44.1				
Client ID: PR132-S16-119133-20251021								
Q3419-04	PR132-S16-119133	SOIL	Acetone	63.1		8.70	46.1	ug/Kg
Q3419-04	PR132-S16-119133	SOIL	Carbon Disulfide	3.80	J	2.00	9.20	ug/Kg
Q3419-04	PR132-S16-119133	SOIL	2-Butanone	20.1	J	12.0	46.1	ug/Kg
Q3419-04	PR132-S16-119133	SOIL	Methylcyclohexane	16.8		1.70	9.20	ug/Kg
			Total Voc :	104				
Q3419-04	PR132-S16-119133	SOIL	Cyclohexane, 1,1,3-trimethyl-	* 140	J	0	0	ug/Kg
Q3419-04	PR132-S16-119133	SOIL	Octane, 3,6-dimethyl-	* 200	J	0	0	ug/Kg
Q3419-04	PR132-S16-119133	SOIL	Undecane, 2,6-dimethyl-	* 150	J	0	0	ug/Kg
Q3419-04	PR132-S16-119133	SOIL	1,2,4-Trimethylbenzene	* 2.30	J	1.20	9.20	ug/Kg
			Total Tics :	492				
			Total Concentration:	596				
Client ID: PR132-S16-094119-20251021								
Q3419-05	PR132-S16-094119	SOIL	Acetone	180		10.1	53.5	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	Carbon Disulfide	11.4		2.30	10.7	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	2-Butanone	53.2	J	14.0	53.5	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	Methylcyclohexane	16.9		1.90	10.7	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	Benzene	5.70	J	1.70	10.7	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	m/p-Xylenes	7.30	J	2.70	21.4	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	o-Xylene	18.2		1.80	10.7	ug/Kg
			Total Voc :	293				
Q3419-05	PR132-S16-094119	SOIL	Naphthalene, decahydro-	* 64.4	J	0	0	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	3-Heptene, 2,2,4,6,6-pentameth	* 98.3	J	0	0	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	Cyclohexanone, 5-methyl-2-(1-	* 64.6	J	0	0	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	Undecane, 2,6-dimethyl-	* 58.4	J	0	0	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	1,3,5-Trimethylbenzene	* 7.00	J	1.80	10.7	ug/Kg
Q3419-05	PR132-S16-094119	SOIL	1,2,4-Trimethylbenzene	* 12.0	J	1.40	10.7	ug/Kg
			Total Tics :	305				
			Total Concentration:	597				



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Surrogate Summary

SDG No.: Q3419

Client: AECOM

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3419-01	PR-132-S16-015094-20251021	1,2-Dichloroethane-d4	50	44.2	88		63	155
		Dibromofluoromethane	50	47.8	96		70	134
		Toluene-d8	50	47.7	95		74	123
		4-Bromofluorobenzene	50	39.4	79		17	146
Q3419-02MS	PR-132-S16-015094-20251021MS	1,2-Dichloroethane-d4	50	53.2	106		63	155
		Dibromofluoromethane	50	52.5	105		70	134
		Toluene-d8	50	42.6	85		74	123
		4-Bromofluorobenzene	50	28.2	56		17	146
Q3419-03MSD	PR-132-S16-015094-20251021MSD	1,2-Dichloroethane-d4	50	54.7	109		63	155
		Dibromofluoromethane	50	46.4	93		70	134
		Toluene-d8	50	34.6	69	*	74	123
		4-Bromofluorobenzene	50	25.0	50		17	146
Q3419-04	PR132-S16-119133-20251021	1,2-Dichloroethane-d4	50	53.5	107		63	155
		Dibromofluoromethane	50	51.4	103		70	134
		Toluene-d8	50	50.6	101		74	123
		4-Bromofluorobenzene	50	47.1	94		17	146
Q3419-05	PR132-S16-094119-20251021	1,2-Dichloroethane-d4	50	52.0	104		63	155
		Dibromofluoromethane	50	50.8	102		70	134
		Toluene-d8	50	49.8	100		74	123
		4-Bromofluorobenzene	50	46.3	93		17	146
VY1022SBL01	VY1022SBL01	1,2-Dichloroethane-d4	50	48.3	97		63	155
		Dibromofluoromethane	50	49.8	100		70	134
		Toluene-d8	50	48.6	97		74	123
		4-Bromofluorobenzene	50	41.1	82		17	146
VY1022SBS01	VY1022SBS01	1,2-Dichloroethane-d4	50	45.7	91		63	155
		Dibromofluoromethane	50	49.3	99		70	134
		Toluene-d8	50	49.8	100		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146
VY1024SBL01	VY1024SBL01	1,2-Dichloroethane-d4	50	44.4	89		63	155
		Dibromofluoromethane	50	48.7	97		70	134
		Toluene-d8	50	47.8	95		74	123
		4-Bromofluorobenzene	50	40.6	81		17	146
VY1024SBS01	VY1024SBS01	1,2-Dichloroethane-d4	50	44.2	88		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	48.4	97		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3419

Analytical Method:

SW8260D

Client: AECOM

Datafile :

VY023580.D

		Sample				Rec	RPD	Limits			
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	Q3419-02MS	Client Sample ID :	PR-132-S16-015094-20251021MS								
Dichlorodifluoromethane	100	0	130	ug/Kg	130				40	150	
Chloromethane	100	0	110	ug/Kg	110				24	175	
Vinyl chloride	100	0	120	ug/Kg	120				21	175	
Bromomethane	100	0	76.6	ug/Kg	77				46	150	
Chloroethane	100	0	130	ug/Kg	130				30	175	
Trichlorofluoromethane	100	0	130	ug/Kg	130				53	150	
1,1,2-Trichlorotrifluoroethane	100	0	120	ug/Kg	120				60	135	
1,1-Dichloroethene	100	0	130	ug/Kg	130				63	136	
Acetone	510	36.3	950	ug/Kg	179	*			41	175	
Carbon disulfide	100	7.80	97.7	ug/Kg	90				45	126	
Methyl tert-butyl Ether	100	0	150	ug/Kg	150				56	175	
Methyl Acetate	100	0	360	ug/Kg	360	*			32	175	
Methylene Chloride	100	0	130	ug/Kg	130				59	175	
trans-1,2-Dichloroethene	100	0	110	ug/Kg	110				63	137	
1,1-Dichloroethane	100	0	130	ug/Kg	130				62	154	
Cyclohexane	100	0	62.0	ug/Kg	62				41	131	
2-Butanone	510	0	870	ug/Kg	171				48	174	
Carbon Tetrachloride	100	0	70.9	ug/Kg	71				48	149	
cis-1,2-Dichloroethene	100	0	120	ug/Kg	120				59	153	
Bromochloromethane	100	0	87.2	ug/Kg	87				54	172	
Chloroform	100	0	130	ug/Kg	130				60	159	
1,1,1-Trichloroethane	100	0	130	ug/Kg	130				60	149	
Methylcyclohexane	100	0	33.3	ug/Kg	33				19	135	
Benzene	100	0	130	ug/Kg	130				59	140	
1,2-Dichloroethane	100	0	140	ug/Kg	140				63	153	
Trichloroethene	100	0	130	ug/Kg	130				45	154	
1,2-Dichloropropane	100	0	130	ug/Kg	130				67	141	
Bromodichloromethane	100	0	120	ug/Kg	120				54	160	
4-Methyl-2-Pentanone	510	0	890	ug/Kg	175	*			54	164	
Toluene	100	0	110	ug/Kg	110				61	134	
t-1,3-Dichloropropene	100	0	63.4	ug/Kg	63				47	155	
cis-1,3-Dichloropropene	100	0	52.7	ug/Kg	53	*			56	141	
1,1,2-Trichloroethane	100	0	150	ug/Kg	150	*			69	148	
2-Hexanone	510	0	830	ug/Kg	163				43	164	
Dibromochloromethane	100	0	120	ug/Kg	120				65	143	
1,2-Dibromoethane	100	0	130	ug/Kg	130				63	144	
Tetrachloroethene	100	0	100	ug/Kg	100				27	175	
Chlorobenzene	100	0	120	ug/Kg	120				66	127	
Ethyl Benzene	100	0	96.1	ug/Kg	96				54	134	
m/p-Xylenes	200	0	190	ug/Kg	95				51	137	
o-Xylene	100	0	95.7	ug/Kg	96				57	139	
Styrene	100	0	87.0	ug/Kg	87				47	136	
Bromoform	100	0	150	ug/Kg	150	*			64	144	
Isopropylbenzene	100	0	100	ug/Kg	100				44	160	
1,1,2,2-Tetrachloroethane	100	0	290	ug/Kg	290	*			18	175	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3419

Analytical Method: SW8260D

Client: AECOM

Datafile : VY023580.D

Parameter	Spike	Sample	Result	Units	Rec		RPD	Limits		
		Result			Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	100	0	77.7	ug/Kg	78				55	131
1,4-Dichlorobenzene	100	0	82.3	ug/Kg	82				57	129
1,2-Dichlorobenzene	100	0	86.7	ug/Kg	87				54	140
1,2-Dibromo-3-Chloropropane	100	0	240	ug/Kg	240	*			46	175
1,2,4-Trichlorobenzene	100	0	26.5	ug/Kg	26				12	127
1,2,3-Trichlorobenzene	100	0	28.0	ug/Kg	28				10	160

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3419

Analytical Method:

SW8260D

Client: AECOM

Datafile :

VY023581.D

		Sample				Rec	RPD		Limits		
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	Q3419-03MSD	Client Sample ID :	PR-132-S16-015094-20251021MSD								
Dichlorodifluoromethane	79	0	99.9	ug/Kg	126		3		40	150	20
Chloromethane	79	0	89.3	ug/Kg	113		3		24	175	20
Vinyl chloride	79	0	95.4	ug/Kg	121		1		21	175	20
Bromomethane	79	0	45.2	ug/Kg	57		29	*	46	150	20
Chloroethane	79	0	100	ug/Kg	127		3		30	175	20
Trichlorofluoromethane	79	0	100	ug/Kg	127		3		53	150	20
1,1,2-Trichlorotrifluoroethane	79	0	83.0	ug/Kg	105		13		60	135	20
1,1-Dichloroethene	79	0	100	ug/Kg	127		3		63	136	20
Acetone	390	36.3	1200	ug/Kg	298	*	50	*	41	175	20
Carbon disulfide	79	7.80	120	ug/Kg	142	*	45	*	45	126	20
Methyl tert-butyl Ether	79	0	150	ug/Kg	190	*	23	*	56	175	20
Methyl Acetate	79	0	430	ug/Kg	544	*	41	*	32	175	20
Methylene Chloride	79	0	100	ug/Kg	127		3		59	175	20
trans-1,2-Dichloroethene	79	0	82.6	ug/Kg	105		5		63	137	20
1,1-Dichloroethane	79	0	100	ug/Kg	127		3		62	154	20
Cyclohexane	79	0	45.2	ug/Kg	57		8		41	131	20
2-Butanone	390	0	970	ug/Kg	249	*	37	*	48	174	20
Carbon Tetrachloride	79	0	58.9	ug/Kg	75		5		48	149	20
cis-1,2-Dichloroethene	79	0	95.3	ug/Kg	121		1		59	153	20
Bromochloromethane	79	0	96.1	ug/Kg	122		33	*	54	172	20
Chloroform	79	0	110	ug/Kg	139		7		60	159	20
1,1,1-Trichloroethane	79	0	98.5	ug/Kg	125		4		60	149	20
Methylcyclohexane	79	0	29.4	ug/Kg	37		11		19	135	20
Benzene	79	0	100	ug/Kg	127		3		59	140	20
1,2-Dichloroethane	79	0	120	ug/Kg	152		8		63	153	20
Trichloroethene	79	0	89.8	ug/Kg	114		13		45	154	20
1,2-Dichloropropane	79	0	110	ug/Kg	139		7		67	141	20
Bromodichloromethane	79	0	100	ug/Kg	127		5		54	160	20
4-Methyl-2-Pentanone	390	0	980	ug/Kg	251	*	36	*	54	164	20
Toluene	79	0	80.6	ug/Kg	102		8		61	134	20
t-1,3-Dichloropropene	79	0	33.6	ug/Kg	43	*	39	*	47	155	20
cis-1,3-Dichloropropene	79	0	24.0	ug/Kg	30	*	54	*	56	141	20
1,1,2-Trichloroethane	79	0	140	ug/Kg	177	*	17		69	148	20
2-Hexanone	390	0	930	ug/Kg	238	*	38	*	43	164	20
Dibromochloromethane	79	0	110	ug/Kg	139		15		65	143	20
1,2-Dibromoethane	79	0	120	ug/Kg	152	*	16		63	144	20
Tetrachloroethene	79	0	85.4	ug/Kg	108		8		27	175	20
Chlorobenzene	79	0	87.9	ug/Kg	111		8		66	127	20
Ethyl Benzene	79	0	69.7	ug/Kg	88		9		54	134	20
m/p-Xylenes	160	0	130	ug/Kg	81		16		51	137	20
o-Xylene	79	0	70.9	ug/Kg	90		6		57	139	20
Styrene	79	0	56.7	ug/Kg	72		19		47	136	20
Bromoform	79	0	150	ug/Kg	190	*	23	*	64	144	20
Isopropylbenzene	79	0	77.2	ug/Kg	98		2		44	160	20
1,1,2,2-Tetrachloroethane	79	0	280	ug/Kg	354	*	20		18	175	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3419

Analytical Method: SW8260D

Client: AECOM

Datafile : VY023581.D

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
1,3-Dichlorobenzene	79	0	62.5	ug/Kg	79		2		55	131	20
1,4-Dichlorobenzene	79	0	65.6	ug/Kg	83		1		57	129	20
1,2-Dichlorobenzene	79	0	72.3	ug/Kg	92		5		54	140	20
1,2-Dibromo-3-Chloropropane	79	0	250	ug/Kg	316	*	27	*	46	175	20
1,2,4-Trichlorobenzene	79	0	26.2	ug/Kg	33		22	*	12	127	20
1,2,3-Trichlorobenzene	79	0	29.6	ug/Kg	37		29	*	10	160	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q3419	SW-846	Analytical Method:	SW8260D
Client:	AECOM		Datafile :	VY023545.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1022SBS01	Dichlorodifluoromethane	20	20.0	ug/Kg	100			64	136	
	Chloromethane	20	18.0	ug/Kg	90			52	151	
	Vinyl chloride	20	18.7	ug/Kg	94			56	148	
	Bromomethane	20	20.0	ug/Kg	100			58	141	
	Chloroethane	20	19.8	ug/Kg	99			69	130	
	Trichlorofluoromethane	20	20.2	ug/Kg	101			69	134	
	1,1,2-Trichlorotrifluoroethane	20	21.3	ug/Kg	106			81	123	
	1,1-Dichloroethene	20	20.7	ug/Kg	104			79	121	
	Acetone	100	130	ug/Kg	130			40	171	
	Carbon disulfide	20	19.3	ug/Kg	97			59	130	
	Methyl tert-butyl Ether	20	19.8	ug/Kg	99			77	129	
	Methyl Acetate	20	17.3	ug/Kg	86			69	149	
	Methylene Chloride	20	18.4	ug/Kg	92			72	131	
	trans-1,2-Dichloroethene	20	20.4	ug/Kg	102			80	123	
	1,1-Dichloroethane	20	20.5	ug/Kg	103			82	123	
	Cyclohexane	20	19.3	ug/Kg	97			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	21.4	ug/Kg	107			76	129	
	cis-1,2-Dichloroethene	20	21.2	ug/Kg	106			82	123	
	Bromochloromethane	20	19.3	ug/Kg	97			80	127	
	Chloroform	20	21.2	ug/Kg	106			82	125	
	1,1,1-Trichloroethane	20	21.3	ug/Kg	106			80	126	
	Methylcyclohexane	20	20.7	ug/Kg	104			77	123	
	Benzene	20	21.2	ug/Kg	106			84	121	
	1,2-Dichloroethane	20	19.8	ug/Kg	99			81	126	
	Trichloroethene	20	21.9	ug/Kg	110			83	122	
	1,2-Dichloropropane	20	21.0	ug/Kg	105			83	122	
	Bromodichloromethane	20	21.2	ug/Kg	106			82	123	
	4-Methyl-2-Pentanone	100	96.0	ug/Kg	96			70	135	
	Toluene	20	22.0	ug/Kg	110			83	122	
	t-1,3-Dichloropropene	20	21.0	ug/Kg	105			78	124	
	cis-1,3-Dichloropropene	20	20.9	ug/Kg	104			81	122	
	1,1,2-Trichloroethane	20	21.6	ug/Kg	108			82	125	
	2-Hexanone	100	110	ug/Kg	110			66	138	
	Dibromochloromethane	20	21.4	ug/Kg	107			79	125	
	1,2-Dibromoethane	20	21.2	ug/Kg	106			80	125	
	Tetrachloroethene	20	22.9	ug/Kg	115			83	125	
	Chlorobenzene	20	21.4	ug/Kg	107			84	122	
	Ethyl Benzene	20	21.2	ug/Kg	106			82	124	
	m/p-Xylenes	40	43.3	ug/Kg	108			83	124	
	o-Xylene	20	21.6	ug/Kg	108			83	123	
	Styrene	20	21.3	ug/Kg	106			82	124	
	Bromoform	20	20.5	ug/Kg	103			75	127	
	Isopropylbenzene	20	21.7	ug/Kg	109			82	124	
	1,1,2,2-Tetrachloroethane	20	19.2	ug/Kg	96			77	127	
	1,3-Dichlorobenzene	20	21.3	ug/Kg	106			83	122	
	1,4-Dichlorobenzene	20	21.1	ug/Kg	106			84	121	
	1,2-Dichlorobenzene	20	21.3	ug/Kg	106			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3419</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>AECOM</u>	Datafile :	<u>VY023545.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1022SBS01	1,2-Dibromo-3-Chloropropane	20	18.8	ug/Kg	94			66	134	
	1,2,4-Trichlorobenzene	20	20.4	ug/Kg	102			78	127	
	1,2,3-Trichlorobenzene	20	19.8	ug/Kg	99			70	137	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q3419	SW-846	Analytical Method:	SW8260D
Client:	AECOM		Datafile :	VY023561.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1024SBS01	Dichlorodifluoromethane	20	18.6	ug/Kg	93			64	136	
	Chloromethane	20	17.3	ug/Kg	86			52	151	
	Vinyl chloride	20	18.0	ug/Kg	90			56	148	
	Bromomethane	20	20.6	ug/Kg	103			58	141	
	Chloroethane	20	18.5	ug/Kg	93			69	130	
	Trichlorofluoromethane	20	19.0	ug/Kg	95			69	134	
	1,1,2-Trichlorotrifluoroethane	20	20.0	ug/Kg	100			81	123	
	1,1-Dichloroethene	20	19.8	ug/Kg	99			79	121	
	Acetone	100	130	ug/Kg	130			40	171	
	Carbon disulfide	20	18.4	ug/Kg	92			59	130	
	Methyl tert-butyl Ether	20	20.0	ug/Kg	100			77	129	
	Methyl Acetate	20	17.2	ug/Kg	86			69	149	
	Methylene Chloride	20	18.5	ug/Kg	93			72	131	
	trans-1,2-Dichloroethene	20	20.1	ug/Kg	101			80	123	
	1,1-Dichloroethane	20	19.8	ug/Kg	99			82	123	
	Cyclohexane	20	18.3	ug/Kg	92			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	20.6	ug/Kg	103			76	129	
	cis-1,2-Dichloroethene	20	20.9	ug/Kg	104			82	123	
	Bromochloromethane	20	18.9	ug/Kg	95			80	127	
	Chloroform	20	20.4	ug/Kg	102			82	125	
	1,1,1-Trichloroethane	20	20.5	ug/Kg	103			80	126	
	Methylcyclohexane	20	20.0	ug/Kg	100			77	123	
	Benzene	20	20.7	ug/Kg	104			84	121	
	1,2-Dichloroethane	20	19.6	ug/Kg	98			81	126	
	Trichloroethene	20	21.4	ug/Kg	107			83	122	
	1,2-Dichloropropane	20	20.3	ug/Kg	102			83	122	
	Bromodichloromethane	20	20.8	ug/Kg	104			82	123	
	4-Methyl-2-Pentanone	100	96.4	ug/Kg	96			70	135	
	Toluene	20	21.3	ug/Kg	106			83	122	
	t-1,3-Dichloropropene	20	20.3	ug/Kg	102			78	124	
	cis-1,3-Dichloropropene	20	20.6	ug/Kg	103			81	122	
	1,1,2-Trichloroethane	20	21.6	ug/Kg	108			82	125	
	2-Hexanone	100	110	ug/Kg	110			66	138	
	Dibromochloromethane	20	21.3	ug/Kg	106			79	125	
	1,2-Dibromoethane	20	21.4	ug/Kg	107			80	125	
	Tetrachloroethene	20	20.6	ug/Kg	103			83	125	
	Chlorobenzene	20	20.9	ug/Kg	104			84	122	
	Ethyl Benzene	20	20.5	ug/Kg	103			82	124	
	m/p-Xylenes	40	42.3	ug/Kg	106			83	124	
	o-Xylene	20	21.1	ug/Kg	106			83	123	
	Styrene	20	20.9	ug/Kg	104			82	124	
	Bromoform	20	20.9	ug/Kg	104			75	127	
	Isopropylbenzene	20	20.6	ug/Kg	103			82	124	
	1,1,2,2-Tetrachloroethane	20	20.3	ug/Kg	102			77	127	
	1,3-Dichlorobenzene	20	20.4	ug/Kg	102			83	122	
	1,4-Dichlorobenzene	20	20.6	ug/Kg	103			84	121	
	1,2-Dichlorobenzene	20	20.4	ug/Kg	102			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3419</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>AECOM</u>	Datafile :	<u>VY023561.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY1024SBS01	1,2-Dibromo-3-Chloropropane	20	19.1	ug/Kg	96			66	134	
	1,2,4-Trichlorobenzene	20	19.9	ug/Kg	100			78	127	
	1,2,3-Trichlorobenzene	20	19.6	ug/Kg	98			70	137	

VOLATILE METHOD BLANK SUMMARY

Client ID

VY1022SBL01

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3419

Lab File ID: VY023544.D

Lab Sample ID: VY1022SBL01

Date Analyzed: 10/22/2025

Time Analyzed: 10:00

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY1022SBS01	VY1022SBS01	VY023545.D	10/22/2025
PR-132-S16-015094-20251021	Q3419-01	VY023551.D	10/22/2025
PR132-S16-119133-20251021	Q3419-04	VY023552.D	10/22/2025
PR132-S16-094119-20251021	Q3419-05	VY023553.D	10/22/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

Client ID

VY1024SBL01

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3419

Lab File ID: VY023560.D

Lab Sample ID: VY1024SBL01

Date Analyzed: 10/24/2025

Time Analyzed: 09:24

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY1024SBS01	VY1024SBS01	VY023561.D	10/24/2025
PR-132-S16-015094-20251021MS	Q3419-02MS	VY023580.D	10/24/2025
PR-132-S16-015094-20251021MSD	Q3419-03MSD	VY023581.D	10/24/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
 BROMOFLUOROBENZENE (BFB)
Lab Name: AllianceContract: AECO02Lab Code: ACESDG NO.: Q3419Lab File ID: VY023383.DBFB Injection Date: 10/07/2025Instrument ID: MSVOA_YBFB Injection Time: 08:43GC Column: RXI-624 ID: 0.25 (mm)Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.9
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.9 (0.9) 1
174	50.0 - 100.0% of mass 95	99.1
175	5.0 - 9.0% of mass 174	7.3 (7.4) 1
176	95.0 - 101.0% of mass 174	95.4 (96.2) 1
177	5.0 - 9.0% of mass 176	6.3 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY023384.D	10/07/2025	09:17
VSTDIC010	VSTDIC010	VY023385.D	10/07/2025	10:02
VSTDIC020	VSTDIC020	VY023386.D	10/07/2025	11:24
VSTDIC050	VSTDIC050	VY023387.D	10/07/2025	11:47
VSTDIC100	VSTDIC100	VY023388.D	10/07/2025	12:09
VSTDIC150	VSTDIC150	VY023389.D	10/07/2025	12:32

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
 BROMOFLUOROBENZENE (BFB)
Lab Name: AllianceContract: AECO02Lab Code: ACESDG NO.: Q3419Lab File ID: VY023542.DBFB Injection Date: 10/22/2025Instrument ID: MSVOA_YBFB Injection Time: 09:00GC Column: RXI-624 ID: 0.25 (mm)Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.8
75	30.0 - 60.0% of mass 95	47.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1 (1) 1
174	50.0 - 100.0% of mass 95	99.3
175	5.0 - 9.0% of mass 174	7.2 (7.2) 1
176	95.0 - 101.0% of mass 174	97.5 (98.2) 1
177	5.0 - 9.0% of mass 176	6.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023543.D	10/22/2025	09:30
VY1022SBL01	VY1022SBL01	VY023544.D	10/22/2025	10:00
VY1022SBS01	VY1022SBS01	VY023545.D	10/22/2025	10:28
PR-132-S16-015094-20251021	Q3419-01	VY023551.D	10/22/2025	14:07
PR132-S16-119133-20251021	Q3419-04	VY023552.D	10/22/2025	15:34
PR132-S16-094119-20251021	Q3419-05	VY023553.D	10/22/2025	15:58

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
 BROMOFLUOROBENZENE (BFB)
Lab Name: AllianceContract: AECO02Lab Code: ACESDG NO.: Q3419Lab File ID: VY023558.DBFB Injection Date: 10/24/2025Instrument ID: MSVOA_YBFB Injection Time: 07:47GC Column: RXI-624 ID: 0.25 (mm)Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.7
75	30.0 - 60.0% of mass 95	47.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.1 (1.1) 1
174	50.0 - 100.0% of mass 95	98.2
175	5.0 - 9.0% of mass 174	7.3 (7.4) 1
176	95.0 - 101.0% of mass 174	96.3 (98) 1
177	5.0 - 9.0% of mass 176	6.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023559.D	10/24/2025	08:56
VY1024SBL01	VY1024SBL01	VY023560.D	10/24/2025	09:24
VY1024SBS01	VY1024SBS01	VY023561.D	10/24/2025	09:52
PR-132-S16-015094-20251021MS	Q3419-02MS	VY023580.D	10/24/2025	17:58
PR-132-S16-015094-20251021MSD	Q3419-03MSD	VY023581.D	10/24/2025	18:20

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
Lab Code: ACE SDG NO.: Q3419
Lab File ID: VY023543.D Date Analyzed: 10/22/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:30
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	557873	7.71	806690	8.62	720008	11.42
UPPER LIMIT	1115750	8.213	1613380	9.116	1440020	11.92
LOWER LIMIT	278937	7.213	403345	8.116	360004	10.92
EPA SAMPLE NO.						
PR-132-S16-015094-20251021	785988	7.71	1180613	8.62	922887	11.42
PR132-S16-119133-20251021	900343	7.71	1446852	8.62	1218442	11.42
PR132-S16-094119-20251021	942425	7.71	1514027	8.62	1286219	11.41
VY1022SBL01	890059	7.71	1434768	8.62	1180185	11.42
VY1022SBS01	546555	7.71	794303	8.62	700136	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3419
 Lab File ID: VY023543.D Date Analyzed: 10/22/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	373234	13.346				
UPPER LIMIT	746468	13.846				
LOWER LIMIT	186617	12.846				
EPA SAMPLE NO.						
PR-132-S16-015094-20251021	368413	13.35				
PR132-S16-119133-20251021	499022	13.35				
PR132-S16-094119-20251021	504876	13.35				
VY1022SBL01	484557	13.35				
VY1022SBS01	365280	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3419
 Lab File ID: VY023559.D Date Analyzed: 10/24/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	558901	7.71	809883	8.62	712621	11.41
UPPER LIMIT	1117800	8.213	1619770	9.115	1425240	11.914
LOWER LIMIT	279451	7.213	404942	8.115	356311	10.914
EPA SAMPLE NO.						
PR-132-S16-015094-20251021MS	271890 *	7.71	386012 *	8.62	249668 *	11.42
PR-132-S16-015094-20251021MSD	173647 *	7.71	245581 *	8.62	157959 *	11.41
VY1024SBL01	898566	7.71	1421897	8.62	1164716	11.41
VY1024SBS01	551537	7.71	799010	8.62	709717	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
Lab Code: ACE SDG NO.: Q3419
Lab File ID: VY023559.D Date Analyzed: 10/24/2025
Instrument ID: MSVOA_Y Time Analyzed: 08:56
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	380279	13.346				
UPPER LIMIT	760558	13.846				
LOWER LIMIT	190140	12.846				
EPA SAMPLE NO.						
PR-132-S16-015094-20251021MS	80067 *	13.35				
PR-132-S16-015094-20251021MSI	48940 *	13.35				
VY1024SBL01	469443	13.35				
VY1024SBS01	380713	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: PR-132-S16-015094-20251C
 Lab Sample ID: Q3419-01
 Analytical Method: 8260D
 Sample Wt/Vol: 5.76 g

Level : LOW
 Final Vol: 5000 uL

Date Collected: 10/21/25
 Date Received: 10/21/25
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 54.4
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	8.00	U	1	1.80	8.00	ug/Kg	10/22/25 14:07	VY102225
74-87-3	Chloromethane	8.00	U	1	1.80	8.00	ug/Kg	10/22/25 14:07	VY102225
75-01-4	Vinyl Chloride	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
74-83-9	Bromomethane	8.00	U	1	1.70	8.00	ug/Kg	10/22/25 14:07	VY102225
75-00-3	Chloroethane	8.00	U	1	2.00	8.00	ug/Kg	10/22/25 14:07	VY102225
75-69-4	Trichlorofluoromethane	8.00	U	1	1.90	8.00	ug/Kg	10/22/25 14:07	VY102225
76-13-1	1,1,2-Trichlorotrifluoroethane	8.00	U	1	1.70	8.00	ug/Kg	10/22/25 14:07	VY102225
75-35-4	1,1-Dichloroethene	8.00	U	1	1.60	8.00	ug/Kg	10/22/25 14:07	VY102225
67-64-1	Acetone	36.3	J	1	7.60	39.9	ug/Kg	10/22/25 14:07	VY102225
75-15-0	Carbon Disulfide	7.80	J	1	1.70	8.00	ug/Kg	10/22/25 14:07	VY102225
1634-04-4	Methyl tert-butyl Ether	8.00	U	1	1.20	8.00	ug/Kg	10/22/25 14:07	VY102225
79-20-9	Methyl Acetate	8.00	U	1	2.50	8.00	ug/Kg	10/22/25 14:07	VY102225
75-09-2	Methylene Chloride	16.0	U	1	5.60	16.0	ug/Kg	10/22/25 14:07	VY102225
156-60-5	trans-1,2-Dichloroethene	8.00	U	1	1.40	8.00	ug/Kg	10/22/25 14:07	VY102225
75-34-3	1,1-Dichloroethane	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
110-82-7	Cyclohexane	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
78-93-3	2-Butanone	39.9	U	1	10.4	39.9	ug/Kg	10/22/25 14:07	VY102225
56-23-5	Carbon Tetrachloride	8.00	U	1	1.50	8.00	ug/Kg	10/22/25 14:07	VY102225
156-59-2	cis-1,2-Dichloroethene	8.00	U	1	1.20	8.00	ug/Kg	10/22/25 14:07	VY102225
74-97-5	Bromochloromethane	8.00	U	1	1.80	8.00	ug/Kg	10/22/25 14:07	VY102225
67-66-3	Chloroform	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
71-55-6	1,1,1-Trichloroethane	8.00	U	1	1.50	8.00	ug/Kg	10/22/25 14:07	VY102225
108-87-2	Methylcyclohexane	8.00	U	1	1.50	8.00	ug/Kg	10/22/25 14:07	VY102225
71-43-2	Benzene	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
107-06-2	1,2-Dichloroethane	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
79-01-6	Trichloroethene	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
78-87-5	1,2-Dichloropropane	8.00	U	1	1.50	8.00	ug/Kg	10/22/25 14:07	VY102225
75-27-4	Bromodichloromethane	8.00	U	1	1.20	8.00	ug/Kg	10/22/25 14:07	VY102225
108-10-1	4-Methyl-2-Pentanone	39.9	U	1	5.70	39.9	ug/Kg	10/22/25 14:07	VY102225
108-88-3	Toluene	8.00	U	1	1.20	8.00	ug/Kg	10/22/25 14:07	VY102225
10061-02-6	t-1,3-Dichloropropene	8.00	U	1	1.00	8.00	ug/Kg	10/22/25 14:07	VY102225
10061-01-5	cis-1,3-Dichloropropene	8.00	U	1	0.99	8.00	ug/Kg	10/22/25 14:07	VY102225
79-00-5	1,1,2-Trichloroethane	8.00	U	1	1.50	8.00	ug/Kg	10/22/25 14:07	VY102225
591-78-6	2-Hexanone	39.9	U	1	5.90	39.9	ug/Kg	10/22/25 14:07	VY102225
124-48-1	Dibromochloromethane	8.00	U	1	1.40	8.00	ug/Kg	10/22/25 14:07	VY102225
106-93-4	1,2-Dibromoethane	8.00	U	1	1.40	8.00	ug/Kg	10/22/25 14:07	VY102225
127-18-4	Tetrachloroethene	8.00	U	1	1.70	8.00	ug/Kg	10/22/25 14:07	VY102225
108-90-7	Chlorobenzene	8.00	U	1	1.50	8.00	ug/Kg	10/22/25 14:07	VY102225
100-41-4	Ethyl Benzene	8.00	U	1	1.10	8.00	ug/Kg	10/22/25 14:07	VY102225
179601-23-1	m/p-Xylenes	16.0	U	1	2.00	16.0	ug/Kg	10/22/25 14:07	VY102225
95-47-6	o-Xylene	8.00	U	1	1.30	8.00	ug/Kg	10/22/25 14:07	VY102225
100-42-5	Styrene	8.00	U	1	1.10	8.00	ug/Kg	10/22/25 14:07	VY102225
75-25-2	Bromoform	8.00	U	1	1.40	8.00	ug/Kg	10/22/25 14:07	VY102225

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: PR-132-S16-015094-202510
 Lab Sample ID: Q3419-01
 Analytical Method: 8260D
 Sample Wt/Vol: 5.76 g

Level : LOW
 Final Vol: 5000 uL

Date Collected: 10/21/25
 Date Received: 10/21/25
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 54.4
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	8.00	U	1	1.20	8.00	ug/Kg	10/22/25 14:07	VY102225
79-34-5	1,1,2,2-Tetrachloroethane	8.00	U	1	1.90	8.00	ug/Kg	10/22/25 14:07	VY102225
541-73-1	1,3-Dichlorobenzene	8.00	U	1	2.70	8.00	ug/Kg	10/22/25 14:07	VY102225
106-46-7	1,4-Dichlorobenzene	8.00	U	1	2.50	8.00	ug/Kg	10/22/25 14:07	VY102225
95-50-1	1,2-Dichlorobenzene	8.00	U	1	2.30	8.00	ug/Kg	10/22/25 14:07	VY102225
96-12-8	1,2-Dibromo-3-Chloropropane	8.00	U	1	2.90	8.00	ug/Kg	10/22/25 14:07	VY102225
120-82-1	1,2,4-Trichlorobenzene	8.00	U	1	4.70	8.00	ug/Kg	10/22/25 14:07	VY102225
87-61-6	1,2,3-Trichlorobenzene	8.00	U	1	5.10	8.00	ug/Kg	10/22/25 14:07	VY102225
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	44.2			63 - 155	88%	SPK: 50		
1868-53-7	Dibromofluoromethane	47.8			70 - 134	96%	SPK: 50		
2037-26-5	Toluene-d8	47.7			74 - 123	95%	SPK: 50		
460-00-4	4-Bromofluorobenzene	39.4			17 - 146	79%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	786000							
540-36-3	1,4-Difluorobenzene	1180000							
3114-55-4	Chlorobenzene-d5	923000							
3855-82-1	1,4-Dichlorobenzene-d4	368000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Quant Time: Oct 24 02:35:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	785988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1180613	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	922887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	368413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	290223	44.157	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.320%
35) Dibromofluoromethane	7.640	113	346742	47.798	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.600%
50) Toluene-d8	10.109	98	1289709	47.739	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.480%
62) 4-Bromofluorobenzene	12.408	95	351404	39.398	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.800%
Target Compounds						
					Qvalue	
16) Acetone	3.873	43	35575	22.736	ug/l	99
17) Carbon Disulfide	4.116	76	98634	4.915	ug/l	99
20) Methylene Chloride	4.616	84	15693	1.855	ug/l	96
25) 2-Butanone	6.903	43	12134	6.352	ug/l	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Quant Time: Oct 24 02:35:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023551.D
Acq: 22 Oct 2025 14:07




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021


Tgt Ion:168 Resp: 785988
Ion Ratio Lower Upper
168 100
99 47.8 39.7 59.5



#16
Acetone
Concen: 22.736 ug/l
RT: 3.873 min Scan# 353
Delta R.T. 0.006 min
Lab File: VY023551.D
Acq: 22 Oct 2025 14:07

Tgt Ion: 43 Resp: 35575
Ion Ratio Lower Upper
43 100
58 35.4 27.7 41.5









#17
Carbon Disulfide
Concen: 4.915 ug/l
RT: 4.116 min Scan# 391
Delta R.T. 0.012 min
Lab File: VY023551.D
Acq: 22 Oct 2025 14:07




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021


Tgt Ion: 76 Resp: 98634
Ion Ratio Lower Upper
76 100
78 9.3 7.3 10.9



#20
Methylene Chloride
Concen: 1.855 ug/l
RT: 4.616 min Scan# 475
Delta R.T. 0.000 min
Lab File: VY023551.D
Acq: 22 Oct 2025 14:07

Tgt Ion: 84 Resp: 15693
Ion Ratio Lower Upper
84 100
49 103.2 80.2 120.4
51 33.5 25.9 38.9
86 68.6 50.6 75.8





#25

2-Butanone

Concen: 6.352 ug/l

RT: 6.903 min Scan# 848

Delta R.T. 0.012 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07

Instrument :

MSVOA_Y

ClientSampleId :





PR-132-S16-015094-20251021

Tgt Ion: 43 Resp: 12134

Ion Ratio Lower Upper

43 100

72 28.2 27.4 41.2



#33

1,2-Dichloroethane-d4

Concen: 44.157 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VY023551.D




Acq: 22 Oct 2025 14:07


Tgt Ion: 65 Resp: 290223

Ion Ratio Lower Upper

65 100

67 53.8 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021





Tgt Ion:114 Resp: 1180613

Ion Ratio Lower Upper

114 100

63 16.2 0.0 34.2

88 13.9 0.0 28.4



#35

Dibromofluoromethane

Concen: 47.798 ug/l

RT: 7.640 min Scan# 971

Delta R.T. 0.006 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07




Tgt Ion:113 Resp: 346742


Ion Ratio Lower Upper

113 100

111 103.8 81.8 122.6

192 23.0 18.0 27.0





#50

Toluene-d8

Concen: 47.739 ug/l

RT: 10.109 min Scan# 1375

Delta R.T. 0.006 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07

Instrument :

MSVOA_Y

ClientSampleId :





PR-132-S16-015094-20251021

Tgt Ion: 98 Resp: 1289709

Ion Ratio Lower Upper

98 100

100 65.7 53.0 79.4



#62

4-Bromofluorobenzene

Concen: 39.398 ug/l

RT: 12.408 min Scan# 1753

Delta R.T. 0.006 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07




Tgt Ion: 95 Resp: 351404


Ion Ratio Lower Upper

95 100

174 98.6 0.0 192.8

176 95.3 0.0 187.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021





Tgt Ion:117 Resp: 922887

Ion Ratio Lower Upper

117 100

82 48.1 41.0 61.4

119 32.0 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY023551.D

Acq: 22 Oct 2025 14:07




Tgt Ion:152 Resp: 368413

Ion Ratio Lower Upper

152 100

115 54.5 27.1 81.2

150 156.4 0.0 347.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Title : SW846 8260

Signal : TIC: VY023551.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	3.867	344	352	362	rBV	20581	59271	1.83%	0.288%
2	4.110	382	392	404	rBV2	46219	129302	3.99%	0.629%
3	4.616	464	475	487	rBV2	20044	69471	2.14%	0.338%
4	6.896	838	849	861	rBV3	13832	46886	1.45%	0.228%
5	7.640	957	971	976	rBV	480239	1161461	35.84%	5.653%
6	7.713	976	983	994	rVB	900182	2088320	64.44%	10.164%
7	8.067	1031	1041	1054	rBV	373407	842350	25.99%	4.100%
8	8.250	1065	1071	1080	rVB2	18164	46787	1.44%	0.228%
9	8.616	1121	1131	1146	rBV	1314971	2502828	77.23%	12.181%
10	8.853	1161	1170	1179	rBV	16334	37477	1.16%	0.182%
11	9.116	1200	1213	1218	rBV4	18450	51109	1.58%	0.249%
12	9.676	1295	1305	1312	rBV2	25661	60777	1.88%	0.296%
13	10.109	1369	1376	1390	rVB	1901469	3240870	100.00%	15.773%
14	10.512	1436	1442	1447	rBV2	17050	35908	1.11%	0.175%
15	11.414	1583	1590	1600	rBV	1547409	2598627	80.18%	12.648%
16	11.609	1616	1622	1627	rBV6	14549	34699	1.07%	0.169%
17	11.670	1628	1632	1642	rVB5	17597	45386	1.40%	0.221%
18	11.914	1667	1672	1682	rBV8	12990	43290	1.34%	0.211%
19	12.127	1698	1707	1709	rBV8	23571	43593	1.35%	0.212%
20	12.164	1709	1713	1718	rVB4	21664	39977	1.23%	0.195%
21	12.231	1718	1724	1730	rBV10	14597	41250	1.27%	0.201%
22	12.408	1745	1753	1761	rBV2	1291974	2214473	68.33%	10.778%
23	12.493	1763	1767	1772	rVV3	21525	39734	1.23%	0.193%
24	12.560	1772	1778	1785	rVB5	39402	104948	3.24%	0.511%
25	12.658	1786	1794	1800	rBV7	33412	95149	2.94%	0.463%
26	12.761	1800	1811	1820	rBV8	60640	205163	6.33%	0.999%
27	12.853	1821	1826	1833	rVV	92170	168062	5.19%	0.818%
28	12.920	1833	1837	1842	rVV	115327	178086	5.50%	0.867%
29	12.993	1842	1849	1854	rVB	117886	207096	6.39%	1.008%
30	13.097	1862	1866	1874	rVB3	49444	100251	3.09%	0.488%
31	13.206	1875	1884	1892	rBV	211990	423366	13.06%	2.061%
32	13.273	1892	1895	1899	rVV2	29877	48549	1.50%	0.236%
33	13.347	1899	1907	1912	rVV	1345721	2103477	64.90%	10.238%
34	13.401	1912	1916	1926	rVV2	143153	307700	9.49%	1.498%
35	13.487	1927	1930	1934	rVB5	29464	40627	1.25%	0.198%
36	13.566	1934	1943	1947	rBV4	55437	125107	3.86%	0.609%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Integration Parameters: RTEINT.P
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260

37	13.883	1990	1995	2006	rVB2	264999	475551	14.67%	2.315%
38	14.066	2020	2025	2030	rVB9	25373	57072	1.76%	0.278%
39	14.176	2038	2043	2046	rBV4	44651	74241	2.29%	0.361%
40	14.304	2060	2064	2067	rBV3	36862	61479	1.90%	0.299%
41	14.645	2117	2120	2126	rVB6	32313	53830	1.66%	0.262%
42	14.767	2136	2140	2143	rBV5	21873	35599	1.10%	0.173%
43	14.852	2152	2154	2159	rVB4	26645	37044	1.14%	0.180%
44	14.913	2159	2164	2166	rBV3	24621	43677	1.35%	0.213%
45	15.029	2180	2183	2191	rVB7	33432	63146	1.95%	0.307%
46	15.505	2258	2261	2266	rVB4	36008	63375	1.96%	0.308%




Sum of corrected areas: 20546441

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023551.D
Acq On : 22 Oct 2025 14:07
Operator : SY/MD
Sample : Q3419-01
Misc : 5.76g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: PR132-S16-119133-202510
 Lab Sample ID: Q3419-04
 Analytical Method: 8260D
 Sample Wt/Vol: 5.25 g

Level : LOW
 Final Vol: 5000 uL

Date Collected: 10/21/25
 Date Received: 10/21/25
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 51.7
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	9.20	U	1	2.10	9.20	ug/Kg	10/22/25 15:34	VY102225
74-87-3	Chloromethane	9.20	U	1	2.10	9.20	ug/Kg	10/22/25 15:34	VY102225
75-01-4	Vinyl Chloride	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
74-83-9	Bromomethane	9.20	U	1	2.00	9.20	ug/Kg	10/22/25 15:34	VY102225
75-00-3	Chloroethane	9.20	U	1	2.30	9.20	ug/Kg	10/22/25 15:34	VY102225
75-69-4	Trichlorofluoromethane	9.20	U	1	2.20	9.20	ug/Kg	10/22/25 15:34	VY102225
76-13-1	1,1,2-Trichlorotrifluoroethane	9.20	U	1	2.00	9.20	ug/Kg	10/22/25 15:34	VY102225
75-35-4	1,1-Dichloroethene	9.20	U	1	1.80	9.20	ug/Kg	10/22/25 15:34	VY102225
67-64-1	Acetone	63.1		1	8.70	46.1	ug/Kg	10/22/25 15:34	VY102225
75-15-0	Carbon Disulfide	3.80	J	1	2.00	9.20	ug/Kg	10/22/25 15:34	VY102225
1634-04-4	Methyl tert-butyl Ether	9.20	U	1	1.30	9.20	ug/Kg	10/22/25 15:34	VY102225
79-20-9	Methyl Acetate	9.20	U	1	2.80	9.20	ug/Kg	10/22/25 15:34	VY102225
75-09-2	Methylene Chloride	18.4	U	1	6.50	18.4	ug/Kg	10/22/25 15:34	VY102225
156-60-5	trans-1,2-Dichloroethene	9.20	U	1	1.60	9.20	ug/Kg	10/22/25 15:34	VY102225
75-34-3	1,1-Dichloroethane	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
110-82-7	Cyclohexane	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
78-93-3	2-Butanone	20.1	J	1	12.0	46.1	ug/Kg	10/22/25 15:34	VY102225
56-23-5	Carbon Tetrachloride	9.20	U	1	1.80	9.20	ug/Kg	10/22/25 15:34	VY102225
156-59-2	cis-1,2-Dichloroethene	9.20	U	1	1.40	9.20	ug/Kg	10/22/25 15:34	VY102225
74-97-5	Bromochloromethane	9.20	U	1	2.10	9.20	ug/Kg	10/22/25 15:34	VY102225
67-66-3	Chloroform	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
71-55-6	1,1,1-Trichloroethane	9.20	U	1	1.70	9.20	ug/Kg	10/22/25 15:34	VY102225
108-87-2	Methylcyclohexane	16.8		1	1.70	9.20	ug/Kg	10/22/25 15:34	VY102225
71-43-2	Benzene	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
107-06-2	1,2-Dichloroethane	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
79-01-6	Trichloroethene	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
78-87-5	1,2-Dichloropropane	9.20	U	1	1.70	9.20	ug/Kg	10/22/25 15:34	VY102225
75-27-4	Bromodichloromethane	9.20	U	1	1.40	9.20	ug/Kg	10/22/25 15:34	VY102225
108-10-1	4-Methyl-2-Pentanone	46.1	U	1	6.60	46.1	ug/Kg	10/22/25 15:34	VY102225
108-88-3	Toluene	9.20	U	1	1.40	9.20	ug/Kg	10/22/25 15:34	VY102225
10061-02-6	t-1,3-Dichloropropene	9.20	U	1	1.20	9.20	ug/Kg	10/22/25 15:34	VY102225
10061-01-5	cis-1,3-Dichloropropene	9.20	U	1	1.10	9.20	ug/Kg	10/22/25 15:34	VY102225
79-00-5	1,1,2-Trichloroethane	9.20	U	1	1.70	9.20	ug/Kg	10/22/25 15:34	VY102225
591-78-6	2-Hexanone	46.1	U	1	6.80	46.1	ug/Kg	10/22/25 15:34	VY102225
124-48-1	Dibromochloromethane	9.20	U	1	1.60	9.20	ug/Kg	10/22/25 15:34	VY102225
106-93-4	1,2-Dibromoethane	9.20	U	1	1.60	9.20	ug/Kg	10/22/25 15:34	VY102225
127-18-4	Tetrachloroethene	9.20	U	1	1.90	9.20	ug/Kg	10/22/25 15:34	VY102225
108-90-7	Chlorobenzene	9.20	U	1	1.70	9.20	ug/Kg	10/22/25 15:34	VY102225
100-41-4	Ethyl Benzene	9.20	U	1	1.20	9.20	ug/Kg	10/22/25 15:34	VY102225
179601-23-1	m/p-Xylenes	18.4	U	1	2.30	18.4	ug/Kg	10/22/25 15:34	VY102225
95-47-6	o-Xylene	9.20	U	1	1.50	9.20	ug/Kg	10/22/25 15:34	VY102225
100-42-5	Styrene	9.20	U	1	1.30	9.20	ug/Kg	10/22/25 15:34	VY102225
75-25-2	Bromoform	9.20	U	1	1.60	9.20	ug/Kg	10/22/25 15:34	VY102225

Report of Analysis

Client: AECOM	Date Collected: 10/21/25	
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received: 10/21/25	
Client Sample ID: PR132-S16-119133-202510	SDG No.: Q3419	
Lab Sample ID: Q3419-04	Matrix: SOIL	
Analytical Method: 8260D	% Solid: 51.7	
Sample Wt/Vol: 5.25 g	Test: VOC-TCLVOA-10	
Level: LOW		
Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	9.20	U	1	1.40	9.20	ug/Kg	10/22/25 15:34	VY102225
79-34-5	1,1,2,2-Tetrachloroethane	9.20	U	1	2.20	9.20	ug/Kg	10/22/25 15:34	VY102225
541-73-1	1,3-Dichlorobenzene	9.20	U	1	3.20	9.20	ug/Kg	10/22/25 15:34	VY102225
106-46-7	1,4-Dichlorobenzene	9.20	U	1	2.90	9.20	ug/Kg	10/22/25 15:34	VY102225
95-50-1	1,2-Dichlorobenzene	9.20	U	1	2.70	9.20	ug/Kg	10/22/25 15:34	VY102225
96-12-8	1,2-Dibromo-3-Chloropropane	9.20	U	1	3.40	9.20	ug/Kg	10/22/25 15:34	VY102225
120-82-1	1,2,4-Trichlorobenzene	9.20	U	1	5.50	9.20	ug/Kg	10/22/25 15:34	VY102225
87-61-6	1,2,3-Trichlorobenzene	9.20	U	1	5.90	9.20	ug/Kg	10/22/25 15:34	VY102225
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	53.5			63 - 155	107%	SPK: 50		
1868-53-7	Dibromofluoromethane	51.4			70 - 134	103%	SPK: 50		
2037-26-5	Toluene-d8	50.6			74 - 123	101%	SPK: 50		
460-00-4	4-Bromofluorobenzene	47.1			17 - 146	94%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	900000							
540-36-3	1,4-Difluorobenzene	1450000							
3114-55-4	Chlorobenzene-d5	1220000							
3855-82-1	1,4-Dichlorobenzene-d4	499000							
TENTATIVE IDENTIFIED COMPOUNDS									
003073-66-3	Cyclohexane, 1,1,3-trimethyl-	140	J			11.0	ug/Kg		
015869-94-0	Octane, 3,6-dimethyl-	200	J			12.1	ug/Kg		
95-63-6	1,2,4-Trimethylbenzene	2.30	J			13.0	ug/Kg		
017301-23-4	Undecane, 2,6-dimethyl-	150	J			14.7	ug/Kg		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023552.D
 Acq On : 22 Oct 2025 15:34
 Operator : SY/MD
 Sample : Q3419-04
 Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PR132-S16-119133-20251021

Quant Time: Oct 24 02:35:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.713	168	900343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1446852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1218442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	499022	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	402685	53.486	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.980%	
35) Dibromofluoromethane	7.634	113	456664	51.367	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.740%	
50) Toluene-d8	10.109	98	1676161	50.627	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.260%	
62) 4-Bromofluorobenzene	12.408	95	515265	47.140	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.280%	
Target Compounds						Qvalue
16) Acetone	3.872	43	61394	34.253	ug/l	97
17) Carbon Disulfide	4.116	76	46881	2.040	ug/l	98
20) Methylene Chloride	4.616	84	19228	1.984	ug/l	94
25) 2-Butanone	6.902	43	23901	10.923	ug/l	98
39) Methylcyclohexane	9.109	83	140047	9.139	ug/l #	90
84) 1,2,4-Trimethylbenzene	13.048	105	33222	1.253	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021

Quant Time: Oct 24 02:35:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023552.D
Acq: 22 Oct 2025 15:34




Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021


Tgt Ion:168 Resp: 900343
Ion Ratio Lower Upper
168 100
99 47.5 39.7 59.5



#16
Acetone
Concen: 34.253 ug/l
RT: 3.872 min Scan# 353
Delta R.T. 0.006 min
Lab File: VY023552.D
Acq: 22 Oct 2025 15:34

Tgt Ion: 43 Resp: 61394
Ion Ratio Lower Upper
43 100
58 36.4 27.7 41.5









#17
Carbon Disulfide
Concen: 2.040 ug/l
RT: 4.116 min Scan# 391
Delta R.T. 0.012 min
Lab File: VY023552.D
Acq: 22 Oct 2025 15:34




Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021


Tgt Ion: 76 Resp: 46881
Ion Ratio Lower Upper
76 100
78 9.9 7.3 10.9



#20
Methylene Chloride
Concen: 1.984 ug/l
RT: 4.616 min Scan# 475
Delta R.T. -0.000 min
Lab File: VY023552.D
Acq: 22 Oct 2025 15:34

Tgt Ion: 84 Resp: 19228
Ion Ratio Lower Upper
84 100
49 102.8 80.2 120.4
51 26.9 25.9 38.9
86 69.8 50.6 75.8





#25

2-Butanone

Concen: 10.923 ug/l

RT: 6.902 min Scan# 848

Delta R.T. 0.012 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34

Instrument :

MSVOA_Y

ClientSampleId :





PR132-S16-119133-20251021

Tgt Ion: 43 Resp: 23901

Ion Ratio Lower Upper

43 100

72 32.9 27.4 41.2



#33

1,2-Dichloroethane-d4

Concen: 53.486 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VY023552.D




Acq: 22 Oct 2025 15:34


Tgt Ion: 65 Resp: 402685

Ion Ratio Lower Upper

65 100

67 53.5 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-119133-20251021





Tgt Ion:114 Resp: 1446852

Ion Ratio Lower Upper

114 100

63 17.0 0.0 34.2

88 14.1 0.0 28.4



#35

Dibromofluoromethane

Concen: 51.367 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.000 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34




Tgt Ion:113 Resp: 456664


Ion Ratio Lower Upper

113 100

111 102.2 81.8 122.6

192 22.9 18.0 27.0





#39

Methylcyclohexane

Concen: 9.139 ug/l

RT: 9.109 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-119133-20251021





Tgt Ion: 83 Resp: 140047

Ion Ratio Lower Upper

83 100

55 78.4 51.4 77.2#

98 50.7 40.4 60.6



#50

Toluene-d8

Concen: 50.627 ug/l

RT: 10.109 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VY023552.D




Acq: 22 Oct 2025 15:34


Tgt Ion: 98 Resp: 1676161

Ion Ratio Lower Upper

98 100

100 64.5 53.0 79.4





#62

4-Bromofluorobenzene

Concen: 47.140 ug/l

RT: 12.408 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-119133-20251021





Tgt Ion: 95 Resp: 515265

Ion Ratio Lower Upper

95 100

174 92.8 0.0 192.8

176 90.0 0.0 187.6



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 1591

Delta R.T. 0.006 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34




Tgt Ion: 117 Resp: 1218442


Ion Ratio Lower Upper

117 100

82 47.1 41.0 61.4

119 32.8 25.6 38.4





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY023552.D

Acq: 22 Oct 2025 15:34

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-119133-20251021





Tgt Ion:152 Resp: 499022

Ion Ratio Lower Upper

152 100

115 54.3 27.1 81.2

150 150.9 0.0 347.4



#84

1,2,4-Trimethylbenzene

Concen: 1.253 ug/l

RT: 13.048 min Scan# 1858

Delta R.T. 0.006 min

Lab File: VY023552.D




Acq: 22 Oct 2025 15:34

Tgt Ion:105 Resp: 33222

Ion Ratio Lower Upper

105 100

120 47.7 23.8 71.2



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Title : SW846 8260

Signal : TIC: VY023552.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.830	13	18	30	rVB2	92206	182484	2.19%	0.100%
2	2.202	71	79	91	rBV	79183	155002	1.86%	0.085%
3	2.848	178	185	197	rVB	148807	329065	3.95%	0.181%
4	5.141	549	561	578	rVB4	60675	225661	2.71%	0.124%
5	7.640	956	971	976	rBV	633059	1570964	18.88%	0.864%
6	7.713	976	983	991	rVV	1048123	2463023	29.59%	1.354%
7	7.799	991	997	1008	rVB	95450	249543	3.00%	0.137%
8	7.921	1008	1017	1024	rBV	80000	199135	2.39%	0.109%
9	8.067	1032	1041	1053	rVV	522024	1141939	13.72%	0.628%
10	8.195	1054	1062	1070	rVV3	70358	180246	2.17%	0.099%
11	8.341	1080	1086	1098	rVB	109461	253253	3.04%	0.139%
12	8.616	1122	1131	1143	rBV	1580672	3133713	37.65%	1.723%
13	8.933	1174	1183	1192	rVB2	44045	109821	1.32%	0.060%
14	9.085	1198	1208	1209	rBV3	254598	489200	5.88%	0.269%
15	9.115	1209	1213	1218	rVV2	410563	847533	10.18%	0.466%
16	9.170	1218	1222	1231	rVB	285545	537727	6.46%	0.296%
17	9.390	1248	1258	1266	rVB	441732	953433	11.46%	0.524%
18	9.536	1275	1282	1290	rVB2	506375	959835	11.53%	0.528%
19	9.688	1301	1307	1311	rBV	199211	359093	4.31%	0.197%
20	9.731	1311	1314	1321	rVB2	157772	277430	3.33%	0.153%
21	9.865	1329	1336	1341	rBV	273437	548579	6.59%	0.302%
22	9.920	1341	1345	1351	rVB	210084	363944	4.37%	0.200%
23	9.999	1351	1358	1369	rBV5	210552	630265	7.57%	0.346%
24	10.109	1369	1376	1389	rVB	3117260	5877328	70.62%	3.231%
25	10.237	1389	1397	1399	rBV2	116982	219244	2.63%	0.121%
26	10.298	1399	1407	1414	rVB2	459961	1174305	14.11%	0.646%
27	10.414	1414	1426	1429	rBV	659048	1266406	15.22%	0.696%
28	10.451	1429	1432	1439	rVB	499247	778433	9.35%	0.428%
29	10.548	1440	1448	1452	rBV4	312857	758297	9.11%	0.417%
30	10.597	1452	1456	1459	rVB	257727	370580	4.45%	0.204%
31	10.640	1459	1463	1468	rVB	611374	957869	11.51%	0.527%
32	10.731	1473	1478	1484	rVB	2646311	4108834	49.37%	2.259%
33	10.835	1484	1495	1502	rBV	1317701	3002661	36.08%	1.651%
34	10.902	1502	1506	1509	rBV2	401283	586837	7.05%	0.323%
35	10.932	1509	1511	1513	rVV	141151	161041	1.93%	0.089%
36	10.963	1513	1516	1520	rVV	386503	528580	6.35%	0.291%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260

37	11.018	1520	1525	1531	rVV	2923322	5137181	61.73%	2.824%
38	11.072	1531	1534	1537	rVB	339176	406773	4.89%	0.224%
39	11.133	1537	1544	1548	rBV2	1653595	3228583	38.79%	1.775%
40	11.176	1548	1551	1555	rVV2	548184	924001	11.10%	0.508%
41	11.225	1555	1559	1567	rVB	1314609	2396093	28.79%	1.317%
42	11.304	1567	1572	1576	rBV2	607959	1040103	12.50%	0.572%
43	11.347	1576	1579	1585	rBV5	162535	241060	2.90%	0.133%
44	11.420	1585	1591	1595	rBV	2036081	3377732	40.59%	1.857%
45	11.456	1595	1597	1601	rVB	317105	366305	4.40%	0.201%
46	11.554	1608	1613	1616	rBV	740807	1068264	12.84%	0.587%
47	11.603	1616	1621	1626	rBV4	908343	1737523	20.88%	0.955%
48	11.664	1626	1631	1635	rBV	1455590	2490183	29.92%	1.369%
49	11.706	1635	1638	1643	rVB2	1281279	2016848	24.23%	1.109%
50	11.767	1643	1648	1652	rBV	841922	1549442	18.62%	0.852%
51	11.859	1659	1663	1668	rVB2	731117	1105829	13.29%	0.608%
52	11.932	1668	1675	1681	rBV4	2415026	5454095	65.53%	2.998%
53	11.987	1681	1684	1686	rVV3	641991	903917	10.86%	0.497%
54	12.017	1686	1689	1690	rVV	736012	906834	10.90%	0.498%
55	12.054	1690	1695	1699	rVV	4699514	7275143	87.41%	3.999%
56	12.127	1699	1707	1709	rVV3	2330471	4409456	52.98%	2.424%
57	12.164	1709	1713	1719	rVV3	4075813	8322563	100.00%	4.575%
58	12.249	1723	1727	1730	rVV3	955475	1916424	23.03%	1.053%
59	12.298	1730	1735	1742	rVB6	1803983	4076078	48.98%	2.241%
60	12.408	1748	1753	1759	rBV3	2164101	3822644	45.93%	2.101%
61	12.499	1759	1768	1771	rBV5	800933	2049195	24.62%	1.126%
62	12.566	1775	1779	1786	rVB2	2704128	4858591	58.38%	2.671%
63	12.651	1786	1793	1798	rBV5	1331338	3206272	38.53%	1.763%
64	12.731	1798	1806	1811	rVV3	2593470	7030116	84.47%	3.865%
65	12.785	1812	1815	1820	rVB3	1556823	2418197	29.06%	1.329%
66	12.877	1820	1830	1833	rBV4	1390315	2952566	35.48%	1.623%
67	12.932	1833	1839	1841	rBV3	844778	1611722	19.37%	0.886%
68	12.968	1841	1845	1851	rVB	3434058	5347034	64.25%	2.939%
69	13.029	1851	1855	1857	rBV3	456441	665868	8.00%	0.366%
70	13.090	1862	1865	1870	rVB2	798266	1161825	13.96%	0.639%
71	13.145	1870	1874	1877	rBV	1180082	1654722	19.88%	0.910%
72	13.243	1882	1890	1894	rBV3	2554978	5427781	65.22%	2.984%
73	13.346	1902	1907	1913	rBV	1654788	2745425	32.99%	1.509%
74	13.493	1926	1931	1934	rBV3	813786	1199974	14.42%	0.660%
75	13.529	1934	1937	1945	rVB8	501831	938719	11.28%	0.516%
76	13.621	1945	1952	1953	rBV4	370360	832640	10.00%	0.458%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260

77	13.669	1954	1960	1965	rBV5	1965197	3913741	47.03%	2.151%
78	13.718	1965	1968	1972	rVB2	864375	1126433	13.53%	0.619%
79	13.779	1975	1978	1984	rVV7	479725	908602	10.92%	0.499%
80	13.840	1984	1988	1993	rVV5	793267	1382577	16.61%	0.760%
81	13.938	2001	2004	2009	rVB	1206323	1586494	19.06%	0.872%
82	13.999	2010	2014	2019	rVB5	530014	776154	9.33%	0.427%
83	14.066	2020	2025	2030	rVB4	823042	1458657	17.53%	0.802%
84	14.121	2031	2034	2037	rVB5	416519	514021	6.18%	0.283%
85	14.182	2037	2044	2050	rBV5	2077886	4700218	56.48%	2.584%
86	14.340	2066	2070	2074	rBV2	1160242	1590328	19.11%	0.874%
87	14.468	2088	2091	2094	rBV3	265685	385600	4.63%	0.212%
88	14.535	2099	2102	2105	rBV4	491924	740551	8.90%	0.407%
89	14.602	2109	2113	2117	rVV2	1091823	1924988	23.13%	1.058%
90	14.651	2117	2121	2128	rVB	2862466	4436333	53.30%	2.439%
91	14.724	2131	2133	2137	rVB5	443651	528508	6.35%	0.291%
92	14.767	2137	2140	2142	rBV2	285614	344037	4.13%	0.189%
93	14.858	2152	2155	2159	rVB3	632367	890744	10.70%	0.490%
94	15.127	2194	2199	2205	rBV	2363686	3588405	43.12%	1.973%
95	15.352	2232	2236	2242	rBV7	379381	882241	10.60%	0.485%
96	15.413	2243	2246	2248	rBV4	308947	438942	5.27%	0.241%
97	15.511	2257	2262	2268	rVB2	1307890	2348412	28.22%	1.291%
98	15.974	2335	2338	2345	rVB3	359673	777026	9.34%	0.427%
99	16.047	2346	2350	2357	rVB	1076998	1828741	21.97%	1.005%
100	16.358	2397	2401	2407	rBV4	324671	616010	7.40%	0.339%




Sum of corrected areas: 181912787

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 Cyclohexane, 1,1,3-trimethyl- Concentration Rank 10

R.T.	EstConc	Area	Relative to ISTD	R.T.
11.018	76.04 ug/l	5137180	Chlorobenzene-d5	11.420

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Cyclohexane, 1,1,3-trimethyl-	126	C9H18	003073-66-3	94
2			4-Amino-6-hydroxypyrimidine	111	C4H5N3O	001193-22-2	64
3			Cyclohexane, 1-ethyl-2,4-dimethyl-	140	C10H20	061142-69-6	53
4			Benzenamine, 2-fluoro-	111	C6H6FN	000348-54-9	52
5			Cyclohexane, 1,3,5-trimethyl-	126	C9H18	001839-63-0	50



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 3 Octane, 3,6-dimethyl- Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
12.054	107.69 ug/l	7275140	Chlorobenzene-d5	11.420

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Octane, 3,6-dimethyl-	142	C10H22	015869-94-0	91
2			Octane, 2,6-dimethyl-	142	C10H22	002051-30-1	91
3			Nonane, 3-methyl-	142	C10H22	005911-04-6	91
4			Undecane, 6,6-dimethyl-	184	C13H28	017312-76-4	80
5			Sulfurous acid, di(2-ethylhexyl)...	306	C16H34O3S	1000309-19-1	72



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 10 Undecane, 2,6-dimethyl- Concentration Rank 8

R.T.	EstConc	Area	Relative to ISTD	R.T.
14.651	80.80 ug/l	4436330	1,4-Dichlorobenzene-d4	13.346

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Undecane, 2,6-dimethyl-	184	C13H28	017301-23-4	94
2			Nonane, 3-methyl-	142	C10H22	005911-04-6	70
3			Sulfurous acid, butyl octyl ester	250	C12H26O3S	1000309-17-5	59
4			Undecane, 2,5-dimethyl-	184	C13H28	017301-22-3	55
5			Sulfurous acid, di(2-ethylhexyl)...	306	C16H34O3S	1000309-19-1	50



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023552.D
Acq On : 22 Oct 2025 15:34
Operator : SY/MD
Sample : Q3419-04
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-119133-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
Cyclohexane, 1,...	11.018	76.0	ug/l	5137180	3	11.420	3377730	50.0
Octane, 3,6-dim...	12.054	107.7	ug/l	7275140	3	11.420	3377730	50.0
Undecane, 2,6-d...	14.651	80.8	ug/l	4436330	4	13.346	2745430	50.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: AECOM
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Client Sample ID: PR132-S16-094119-202510
Lab Sample ID: Q3419-05
Analytical Method: 8260D
Sample Wt/Vol: 4.98 g

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/21/25
Date Received: 10/21/25
SDG No.: Q3419
Matrix: SOIL
% Solid: 46.9
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	10.7	U	1	2.40	10.7	ug/Kg	10/22/25 15:58	VY102225
74-87-3	Chloromethane	10.7	U	1	2.40	10.7	ug/Kg	10/22/25 15:58	VY102225
75-01-4	Vinyl Chloride	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
74-83-9	Bromomethane	10.7	U	1	2.30	10.7	ug/Kg	10/22/25 15:58	VY102225
75-00-3	Chloroethane	10.7	U	1	2.70	10.7	ug/Kg	10/22/25 15:58	VY102225
75-69-4	Trichlorofluoromethane	10.7	U	1	2.60	10.7	ug/Kg	10/22/25 15:58	VY102225
76-13-1	1,1,2-Trichlorotrifluoroethane	10.7	U	1	2.30	10.7	ug/Kg	10/22/25 15:58	VY102225
75-35-4	1,1-Dichloroethene	10.7	U	1	2.10	10.7	ug/Kg	10/22/25 15:58	VY102225
67-64-1	Acetone	180		1	10.1	53.5	ug/Kg	10/22/25 15:58	VY102225
75-15-0	Carbon Disulfide	11.4		1	2.30	10.7	ug/Kg	10/22/25 15:58	VY102225
1634-04-4	Methyl tert-butyl Ether	10.7	U	1	1.60	10.7	ug/Kg	10/22/25 15:58	VY102225
79-20-9	Methyl Acetate	10.7	U	1	3.30	10.7	ug/Kg	10/22/25 15:58	VY102225
75-09-2	Methylene Chloride	21.4	U	1	7.60	21.4	ug/Kg	10/22/25 15:58	VY102225
156-60-5	trans-1,2-Dichloroethene	10.7	U	1	1.80	10.7	ug/Kg	10/22/25 15:58	VY102225
75-34-3	1,1-Dichloroethane	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
110-82-7	Cyclohexane	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
78-93-3	2-Butanone	53.2	J	1	14.0	53.5	ug/Kg	10/22/25 15:58	VY102225
56-23-5	Carbon Tetrachloride	10.7	U	1	2.10	10.7	ug/Kg	10/22/25 15:58	VY102225
156-59-2	cis-1,2-Dichloroethene	10.7	U	1	1.60	10.7	ug/Kg	10/22/25 15:58	VY102225
74-97-5	Bromochloromethane	10.7	U	1	2.50	10.7	ug/Kg	10/22/25 15:58	VY102225
67-66-3	Chloroform	10.7	U	1	1.80	10.7	ug/Kg	10/22/25 15:58	VY102225
71-55-6	1,1,1-Trichloroethane	10.7	U	1	2.00	10.7	ug/Kg	10/22/25 15:58	VY102225
108-87-2	Methylcyclohexane	16.9		1	1.90	10.7	ug/Kg	10/22/25 15:58	VY102225
71-43-2	Benzene	5.70	J	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
107-06-2	1,2-Dichloroethane	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
79-01-6	Trichloroethene	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
78-87-5	1,2-Dichloropropane	10.7	U	1	1.90	10.7	ug/Kg	10/22/25 15:58	VY102225
75-27-4	Bromodichloromethane	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
108-10-1	4-Methyl-2-Pentanone	53.5	U	1	7.70	53.5	ug/Kg	10/22/25 15:58	VY102225
108-88-3	Toluene	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
10061-02-6	t-1,3-Dichloropropene	10.7	U	1	1.40	10.7	ug/Kg	10/22/25 15:58	VY102225
10061-01-5	cis-1,3-Dichloropropene	10.7	U	1	1.30	10.7	ug/Kg	10/22/25 15:58	VY102225
79-00-5	1,1,2-Trichloroethane	10.7	U	1	2.00	10.7	ug/Kg	10/22/25 15:58	VY102225
591-78-6	2-Hexanone	53.5	U	1	7.90	53.5	ug/Kg	10/22/25 15:58	VY102225
124-48-1	Dibromochloromethane	10.7	U	1	1.90	10.7	ug/Kg	10/22/25 15:58	VY102225
106-93-4	1,2-Dibromoethane	10.7	U	1	1.90	10.7	ug/Kg	10/22/25 15:58	VY102225
127-18-4	Tetrachloroethene	10.7	U	1	2.20	10.7	ug/Kg	10/22/25 15:58	VY102225
108-90-7	Chlorobenzene	10.7	U	1	1.90	10.7	ug/Kg	10/22/25 15:58	VY102225
100-41-4	Ethyl Benzene	10.7	U	1	1.40	10.7	ug/Kg	10/22/25 15:58	VY102225
179601-23-1	m/p-Xylenes	7.30	J	1	2.70	21.4	ug/Kg	10/22/25 15:58	VY102225
95-47-6	o-Xylene	18.2		1	1.80	10.7	ug/Kg	10/22/25 15:58	VY102225
100-42-5	Styrene	10.7	U	1	1.50	10.7	ug/Kg	10/22/25 15:58	VY102225
75-25-2	Bromoform	10.7	U	1	1.80	10.7	ug/Kg	10/22/25 15:58	VY102225

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: PR132-S16-094119-202510
 Lab Sample ID: Q3419-05
 Analytical Method: 8260D
 Sample Wt/Vol: 4.98 g

Level : LOW
 Final Vol: 5000 uL

Date Collected: 10/21/25
 Date Received: 10/21/25
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 46.9
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	10.7	U	1	1.70	10.7	ug/Kg	10/22/25 15:58	VY102225
79-34-5	1,1,2,2-Tetrachloroethane	10.7	U	1	2.60	10.7	ug/Kg	10/22/25 15:58	VY102225
541-73-1	1,3-Dichlorobenzene	10.7	U	1	3.70	10.7	ug/Kg	10/22/25 15:58	VY102225
106-46-7	1,4-Dichlorobenzene	10.7	U	1	3.30	10.7	ug/Kg	10/22/25 15:58	VY102225
95-50-1	1,2-Dichlorobenzene	10.7	U	1	3.10	10.7	ug/Kg	10/22/25 15:58	VY102225
96-12-8	1,2-Dibromo-3-Chloropropane	10.7	U	1	3.90	10.7	ug/Kg	10/22/25 15:58	VY102225
120-82-1	1,2,4-Trichlorobenzene	10.7	U	1	6.40	10.7	ug/Kg	10/22/25 15:58	VY102225
87-61-6	1,2,3-Trichlorobenzene	10.7	U	1	6.80	10.7	ug/Kg	10/22/25 15:58	VY102225
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	52.0			63 - 155	104%	SPK: 50		
1868-53-7	Dibromofluoromethane	50.8			70 - 134	102%	SPK: 50		
2037-26-5	Toluene-d8	49.8			74 - 123	100%	SPK: 50		
460-00-4	4-Bromofluorobenzene	46.3			17 - 146	93%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	942000							
540-36-3	1,4-Difluorobenzene	1510000							
3114-55-4	Chlorobenzene-d5	1290000							
3855-82-1	1,4-Dichlorobenzene-d4	505000							
TENTATIVE IDENTIFIED COMPOUNDS									
108-67-8	1,3,5-Trimethylbenzene	7.00	J			12.7	ug/Kg		
000123-48-8	3-Heptene, 2,2,4,6,6-pentamethyl-	98.3	J			12.9	ug/Kg		
95-63-6	1,2,4-Trimethylbenzene	12.0	J			13.0	ug/Kg		
000091-17-8	Naphthalene, decahydro-	64.4	J			13.7	ug/Kg		
015932-80-6	Cyclohexanone, 5-methyl-2-(1-meth	64.6	J			14.2	ug/Kg		
017301-23-4	Undecane, 2,6-dimethyl-	58.4	J			14.6	ug/Kg		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023553.D
 Acq On : 22 Oct 2025 15:58
 Operator : SY/MD
 Sample : Q3419-05
 Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PR132-S16-094119-20251021

Quant Time: Oct 24 02:36:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.713	168	942425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1514027	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1286219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	504876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	409580	51.973	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.940%	
35) Dibromofluoromethane	7.634	113	472529	50.793	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.580%	
50) Toluene-d8	10.109	98	1725302	49.799	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.600%	
62) 4-Bromofluorobenzene	12.408	95	529575	46.299	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 92.600%	
Target Compounds						
					Qvalue	
16) Acetone	3.873	43	161831	86.257	ug/l	94
17) Carbon Disulfide	4.110	76	128146	5.326	ug/l	99
20) Methylene Chloride	4.622	84	16514	1.628	ug/l #	95
25) 2-Butanone	6.903	43	56883	24.835	ug/l	97
39) Methylcyclohexane	9.110	83	126416	7.883	ug/l #	79
40) Benzene	8.079	78	104295	2.674	ug/l	97
68) m/p-Xylenes	11.627	106	61705	3.397	ug/l	92
69) o-Xylene	11.957	106	145222	8.524	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	87778	3.265	ug/l	96
84) 1,2,4-Trimethylbenzene	13.042	105	150406	5.605	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Quant Time: Oct 24 02:36:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023553.D
Acq: 22 Oct 2025 15:58




Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021


Tgt Ion:168 Resp: 942425
Ion Ratio Lower Upper
168 100
99 45.9 39.7 59.5



#16
Acetone
Concen: 86.257 ug/l
RT: 3.873 min Scan# 353
Delta R.T. 0.006 min
Lab File: VY023553.D
Acq: 22 Oct 2025 15:58

Tgt Ion: 43 Resp: 161831
Ion Ratio Lower Upper
43 100
58 37.8 27.7 41.5





#17

Carbon Disulfide

Concen: 5.326 ug/l

RT: 4.110 min Scan# 391

Delta R.T. 0.006 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :





PR132-S16-094119-20251021

Tgt Ion: 76 Resp: 128146

Ion Ratio Lower Upper

76 100

78 9.3 7.3 10.9



#20

Methylene Chloride

Concen: 1.628 ug/l

RT: 4.622 min Scan# 476

Delta R.T. 0.006 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Tgt Ion: 84 Resp: 16514




Ion Ratio Lower Upper


84 100

49 104.2 80.2 120.4

51 24.7 25.9 38.9

86 65.9 50.6 75.8





#25

2-Butanone

Concen: 24.835 ug/l

RT: 6.903 min Scan# 848

Delta R.T. 0.012 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :





PR132-S16-094119-20251021

Tgt Ion: 43 Resp: 56883

Ion Ratio Lower Upper

43 100

72 32.7 27.4 41.2



#33

1,2-Dichloroethane-d4

Concen: 51.973 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VY023553.D




Acq: 22 Oct 2025 15:58


Tgt Ion: 65 Resp: 409580

Ion Ratio Lower Upper

65 100

67 53.6 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-094119-20251021





Tgt Ion:114 Resp: 1514027

Ion Ratio Lower Upper

114 100

63 16.7 0.0 34.2

88 14.3 0.0 28.4



#35

Dibromofluoromethane

Concen: 50.793 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.000 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58




Tgt Ion:113 Resp: 472529


Ion Ratio Lower Upper

113 100

111 101.9 81.8 122.6

192 22.7 18.0 27.0





#39

Methylcyclohexane

Concen: 7.883 ug/l

RT: 9.110 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-094119-20251021





Tgt Ion: 83 Resp: 126416

Ion Ratio Lower Upper

83 100

55 90.1 51.4 77.2#

98 54.6 40.4 60.6



#40

Benzene

Concen: 2.674 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. 0.000 min

Lab File: VY023553.D




Acq: 22 Oct 2025 15:58


Tgt Ion: 78 Resp: 104295

Ion Ratio Lower Upper

78 100

77 24.4 18.3 27.5





#50

Toluene-d8

Concen: 49.799 ug/l

RT: 10.109 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :





PR132-S16-094119-20251021

Tgt Ion: 98 Resp: 1725302

Ion Ratio Lower Upper

98 100

100 65.6 53.0 79.4



#62

4-Bromofluorobenzene

Concen: 46.299 ug/l

RT: 12.408 min Scan# 1753

Delta R.T. 0.006 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58




Tgt Ion: 95 Resp: 529575


Ion Ratio Lower Upper

95 100

174 93.6 0.0 192.8

176 89.7 0.0 187.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :

PR132-S16-094119-20251021





Tgt Ion:117 Resp: 1286219

Ion Ratio Lower Upper

117 100

82 50.2 41.0 61.4

119 31.9 25.6 38.4



#68

m/p-Xylenes

Concen: 3.397 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. 0.000 min

Lab File: VY023553.D




Acq: 22 Oct 2025 15:58


Tgt Ion:106 Resp: 61705

Ion Ratio Lower Upper

106 100

91 202.3 152.6 229.0





#69

o-Xylene

Concen: 8.524 ug/l

RT: 11.957 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :





PR132-S16-094119-20251021

Tgt Ion:106 Resp: 145222

Ion Ratio Lower Upper

106 100

91 204.6 101.4 304.1



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58




Tgt Ion:152 Resp: 504876


Ion Ratio Lower Upper

152 100

115 52.6 27.1 81.2

150 149.7 0.0 347.4





#80

1,3,5-Trimethylbenzene

Concen: 3.265 ug/l

RT: 12.737 min Scan# 1806

Delta R.T. 0.006 min

Lab File: VY023553.D

Acq: 22 Oct 2025 15:58

Instrument :

MSVOA_Y

ClientSampleId :





PR132-S16-094119-20251021

Tgt Ion:105 Resp: 87778

Ion Ratio Lower Upper

105 100

120 53.7 25.4 76.4



#84

1,2,4-Trimethylbenzene

Concen: 5.605 ug/l

RT: 13.042 min Scan# 1857

Delta R.T. 0.000 min

Lab File: VY023553.D




Acq: 22 Oct 2025 15:58

Tgt Ion:105 Resp: 150406

Ion Ratio Lower Upper

105 100

120 45.4 23.8 71.2



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y1007255.M

Title : SW846 8260

Signal : TIC: VY023553.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.830	13	18	28	rVB2	70864	131882	2.66%	0.154%
2	2.336	96	101	110	rVB	69272	122027	2.46%	0.143%
3	2.842	178	184	192	rVB	52232	108568	2.19%	0.127%
4	3.867	342	352	362	rBV	96039	264245	5.33%	0.309%
5	4.110	383	392	405	rBV	62614	166433	3.36%	0.194%
6	4.623	464	476	480	rBV3	28341	94772	1.91%	0.111%
7	5.141	547	561	576	rVB2	32570	113335	2.29%	0.132%
8	6.592	787	799	809	rBV3	30859	109452	2.21%	0.128%
9	6.896	838	849	863	rBV2	40694	123079	2.48%	0.144%
10	7.634	956	970	976	rBV	645137	1583266	31.93%	1.849%
11	7.713	976	983	991	rVV	1109449	2635577	53.15%	3.079%
12	7.793	991	996	1006	rVV2	112298	273712	5.52%	0.320%
13	7.921	1009	1017	1023	rVV	49587	123862	2.50%	0.145%
14	8.067	1032	1041	1053	rVV2	580408	1407823	28.39%	1.645%
15	8.195	1053	1062	1064	rVV2	62650	132854	2.68%	0.155%
16	8.250	1064	1071	1080	rVV	404672	1052120	21.22%	1.229%
17	8.341	1080	1086	1097	rVB2	70915	170814	3.44%	0.200%
18	8.616	1122	1131	1147	rBV	1615873	3230564	65.15%	3.774%
19	8.853	1162	1170	1177	rBV	170382	374762	7.56%	0.438%
20	9.110	1199	1212	1218	rBV4	421004	1166438	23.52%	1.363%
21	9.170	1218	1222	1229	rVV	221877	415755	8.38%	0.486%
22	9.250	1230	1235	1239	rVV	41513	88354	1.78%	0.103%
23	9.305	1239	1244	1249	rVV2	60934	150007	3.03%	0.175%
24	9.390	1249	1258	1265	rVV3	175642	455395	9.18%	0.532%
25	9.463	1265	1270	1276	rVV4	37641	81684	1.65%	0.095%
26	9.542	1276	1283	1290	rVB2	409118	806093	16.26%	0.942%
27	9.683	1296	1306	1311	rBV3	390566	873023	17.61%	1.020%
28	9.731	1311	1314	1319	rVB	114269	176658	3.56%	0.206%
29	9.774	1319	1321	1328	rVB4	34352	67678	1.36%	0.079%
30	9.865	1329	1336	1342	rBV2	148192	309773	6.25%	0.362%
31	9.926	1342	1346	1352	rVB5	53843	113870	2.30%	0.133%
32	10.000	1352	1358	1360	rBV3	59902	107576	2.17%	0.126%
33	10.042	1360	1365	1370	rVB	327895	516450	10.42%	0.603%
34	10.109	1370	1376	1389	rVB	2755460	4958386	100.00%	5.792%
35	10.237	1391	1397	1398	rBV2	50854	78591	1.59%	0.092%
36	10.274	1398	1403	1416	rVB6	150346	553839	11.17%	0.647%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260

37	10.408	1416	1425	1428	rBV4	153099	328773	6.63%	0.384%
38	10.451	1428	1432	1439	rVB2	177473	290342	5.86%	0.339%
39	10.548	1440	1448	1454	rBV6	128417	412944	8.33%	0.482%
40	10.640	1460	1463	1468	rVB2	131586	198636	4.01%	0.232%
41	10.731	1468	1478	1484	rBV3	384452	767335	15.48%	0.896%
42	10.829	1484	1494	1502	rBV4	403593	1039845	20.97%	1.215%
43	10.902	1502	1506	1509	rBV2	108568	186440	3.76%	0.218%
44	10.963	1513	1516	1520	rBV	138900	201284	4.06%	0.235%
45	11.018	1520	1525	1530	rVB	901193	1489604	30.04%	1.740%
46	11.073	1530	1534	1538	rVB	223558	313169	6.32%	0.366%
47	11.127	1538	1543	1547	rVB4	195304	358814	7.24%	0.419%
48	11.195	1547	1554	1555	rBV3	238387	456972	9.22%	0.534%
49	11.219	1556	1558	1568	rVB3	408711	721547	14.55%	0.843%
50	11.310	1568	1573	1576	rBV3	112256	226146	4.56%	0.264%
51	11.341	1576	1578	1582	rVB4	68204	86703	1.75%	0.101%
52	11.414	1584	1590	1600	rBV	2216005	3813645	76.91%	4.455%
53	11.493	1600	1603	1608	rVB2	91901	137372	2.77%	0.160%
54	11.554	1608	1613	1617	rBV	192604	284653	5.74%	0.333%
55	11.615	1617	1623	1627	rBV2	458429	960339	19.37%	1.122%
56	11.664	1627	1631	1636	rBV3	323176	552501	11.14%	0.645%
57	11.768	1643	1648	1651	rBV	216704	393617	7.94%	0.460%
58	11.816	1651	1656	1660	rVV3	156835	353467	7.13%	0.413%
59	11.865	1660	1664	1667	rVV2	216324	306003	6.17%	0.357%
60	11.932	1667	1675	1682	rVV4	816376	2809601	56.66%	3.282%
61	12.018	1686	1689	1690	rVV2	261462	317909	6.41%	0.371%
62	12.048	1690	1694	1698	rVB	833963	1212343	24.45%	1.416%
63	12.091	1698	1701	1702	rBV	111231	124537	2.51%	0.145%
64	12.127	1702	1707	1709	rVV3	743477	1222824	24.66%	1.428%
65	12.158	1709	1712	1717	rVB	1350630	2101770	42.39%	2.455%
66	12.243	1717	1726	1730	rBV8	429190	1080239	21.79%	1.262%
67	12.408	1746	1753	1758	rVB2	2005093	3645942	73.53%	4.259%
68	12.499	1758	1768	1771	rBV5	486458	1214053	24.48%	1.418%
69	12.566	1774	1779	1786	rVB3	1243138	2433297	49.07%	2.842%
70	12.652	1786	1793	1800	rBV5	745008	2195526	44.28%	2.565%
71	12.737	1800	1807	1810	rBV6	722226	1805497	36.41%	2.109%
72	12.853	1821	1826	1832	rBV2	1186254	2571637	51.86%	3.004%
73	12.920	1832	1837	1841	rBV	1504499	2337329	47.14%	2.730%
74	12.962	1841	1844	1847	rBV2	613621	874060	17.63%	1.021%
75	13.042	1853	1857	1861	rVB5	367061	637907	12.87%	0.745%
76	13.091	1861	1865	1871	rBV3	525911	983033	19.83%	1.148%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260

77	13.206	1880	1884	1892	rVB3	930813	1832798	36.96%	2.141%
78	13.273	1892	1895	1900	rVB4	327148	459874	9.27%	0.537%
79	13.347	1901	1907	1911	rBV	1610623	2543990	51.31%	2.972%
80	13.401	1913	1916	1924	rVB3	580286	1107536	22.34%	1.294%
81	13.487	1926	1930	1934	rVB2	617461	878557	17.72%	1.026%
82	13.536	1934	1938	1939	rBV3	223864	284818	5.74%	0.333%
83	13.664	1954	1959	1964	rBV3	899373	1529570	30.85%	1.787%
84	13.883	1993	1995	2001	rVB4	422604	609715	12.30%	0.712%
85	13.993	2009	2013	2018	rBV4	372360	578390	11.66%	0.676%
86	14.060	2019	2024	2030	rVB5	387575	826368	16.67%	0.965%
87	14.170	2038	2042	2051	rBV5	653435	1534846	30.95%	1.793%
88	14.304	2060	2064	2066	rBV3	295518	435919	8.79%	0.509%
89	14.340	2066	2070	2074	rVB2	672227	971932	19.60%	1.135%
90	14.535	2099	2102	2106	rBV5	163468	229865	4.64%	0.269%
91	14.602	2109	2113	2116	rVV4	346309	632641	12.76%	0.739%
92	14.645	2116	2120	2127	rVB3	863997	1387556	27.98%	1.621%
93	14.761	2136	2139	2142	rBV2	214522	306306	6.18%	0.358%
94	14.858	2152	2155	2159	rVB5	309431	412910	8.33%	0.482%
95	15.029	2180	2183	2191	rVB4	418145	915236	18.46%	1.069%
96	15.127	2194	2199	2205	rBV2	535391	981175	19.79%	1.146%
97	15.346	2231	2235	2242	rBV7	190939	420020	8.47%	0.491%
98	15.505	2256	2261	2267	rVB2	619821	1130843	22.81%	1.321%
99	16.047	2346	2350	2357	rVB3	355713	619656	12.50%	0.724%
100	16.358	2397	2401	2407	rVB5	216294	389841	7.86%	0.455%




Sum of corrected areas: 85606734

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 6 3-Heptene, 2,2,4,6,6-pentam... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
12.920	45.94 ug/l	2337330	1,4-Dichlorobenzene-d4	13.347

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			3-Heptene, 2,2,4,6,6-pentamethyl-	168	C12H24	000123-48-8	95
2			2-Pentene, 3,4,4-trimethyl-	112	C8H16	000598-96-9	64
3			2-Pentene, 2,4,4-trimethyl-	112	C8H16	000107-40-4	58
4			2-Pyrazoline, 1,4,5-trimethyl-	112	C6H12N2	007423-11-2	53
5			1H-Pyrazole, 4,5-dihydro-3,5,5-t...	112	C6H12N2	003975-85-7	53



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 8 Naphthalene, decahydro- Concentration Rank 7

R.T.	EstConc	Area	Relative to ISTD	R.T.
13.664	30.06 ug/l	1529570	1,4-Dichlorobenzene-d4	13.347

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Naphthalene, decahydro-	138	C10H18	000091-17-8	87
2			Bicyclo[4.1.0]heptane, 3,7,7-tri...	138	C10H18	018968-23-5	60
3			Spiro[4.5]decane	138	C10H18	000176-63-6	58
4			Naphthalene, decahydro-, cis-	138	C10H18	000493-01-6	55
5			Cyclohexane, 1-methyl-4-(1-methy...	138	C10H18	001124-25-0	55



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 9 Cyclohexanone, 5-methyl-2-(... Concentration Rank 6

R.T.	EstConc	Area	Relative to ISTD	R.T.
14.170	30.17 ug/l	1534850	1,4-Dichlorobenzene-d4	13.347

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Cyclohexanone, 5-methyl-2-(1-met...	152	C10H16O	015932-80-6	86
2			Bicyclo[3.3.2]decan-9-one	152	C10H16O	028054-91-3	76
3			Naphthalene, decahydro-2-methyl-	152	C11H20	002958-76-1	68
4			Pulegone	152	C10H16O	000089-82-7	58
5			trans-Decalin, 2-methyl-	152	C11H20	1000152-47-3	58



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1










Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 10 Undecane, 2,6-dimethyl- Concentration Rank 9

R.T.	EstConc	Area	Relative to ISTD		R.T.	
14.645	27.27 ug/l	1387560	1,4-Dichlorobenzene-d4		13.347	
Hit# of 5	Tentative ID		MW	MolForm	CAS#	Qual
1	Undecane, 2,6-dimethyl-		184	C13H28	017301-23-4	92
2	Decane, 2,6,8-trimethyl-		184	C13H28	062108-26-3	49
3	Undecane, 2,5-dimethyl-		184	C13H28	017301-22-3	46
4	Sulfurous acid, decyl 2-ethylhex...		334	C18H38O3S	1000309-19-3	46
5	Pentadecane, 7-methyl-		226	C16H34	006165-40-8	43



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023553.D
Acq On : 22 Oct 2025 15:58
Operator : SY/MD
Sample : Q3419-05
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PR132-S16-094119-20251021

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
3-Heptene, 2,2,...	12.920	45.9	ug/l	2337330	4	13.347	2543990	50.0
Naphthalene, de...	13.664	30.1	ug/l	1529570	4	13.347	2543990	50.0
Cyclohexanone, ...	14.170	30.2	ug/l	1534850	4	13.347	2543990	50.0
Undecane, 2,6-d...	14.645	27.3	ug/l	1387560	4	13.347	2543990	50.0



CALIBRATION SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG No.: Q3419
 Instrument ID: MSVOA_Y Calibration Date(s): 10/07/2025 10/07/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 09:17 12:32
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY023384.D			RRF010 = VY023385.D		RRF020 = VY023386.D		RRF050 = VY023387.D		RRF100 = VY023388.D		RRF150 = VY023389.D	
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD					
Dichlorodifluoromethane	0.397	0.399	0.400	0.321	0.310	0.333	0.360	12					
Chloromethane	0.518	0.557	0.512	0.449	0.438	0.467	0.490	9.4					
Vinyl Chloride	0.660	0.684	0.659	0.598	0.597	0.633	0.639	5.6					
Bromomethane	0.601	0.596	0.548	0.469	0.483	0.518	0.536	10.4					
Chloroethane	0.441	0.462	0.453	0.414	0.413	0.429	0.435	4.7					
Trichlorofluoromethane	0.884	0.920	0.946	0.842	0.850	0.877	0.887	4.5					
1,1,2-Trichlorotrifluoroethane	0.484	0.486	0.479	0.437	0.429	0.445	0.460	5.6					
1,1-Dichloroethene	0.449	0.475	0.456	0.427	0.426	0.443	0.446	4.1					
Acetone	0.128	0.117	0.094	0.084	0.085	0.089	0.100	18.7					
Carbon Disulfide	1.291	1.367	1.320	1.225	1.210	1.246	1.277	4.7					
Methyl tert-butyl Ether	1.006	1.075	1.075	1.020	1.091	1.165	1.072	5.3					
Methyl Acetate	0.244	0.265	0.260	0.245	0.266	0.286	0.261	6					
Methylene Chloride	0.625	0.634	0.576	0.458	0.463	0.473	0.538	15.4					
trans-1,2-Dichloroethene	0.494	0.504	0.507	0.470	0.478	0.494	0.491	2.9					
1,1-Dichloroethane	0.800	0.835	0.827	0.759	0.765	0.792	0.796	3.9					
Cyclohexane	0.797	0.717	0.697	0.643	0.644	0.676	0.696	8.3					
2-Butanone	0.133	0.134	0.115	0.105	0.115	0.126	0.122	9.6					
Carbon Tetrachloride	0.483	0.501	0.504	0.481	0.487	0.503	0.493	2.1					
cis-1,2-Dichloroethene	0.542	0.584	0.582	0.542	0.563	0.591	0.567	3.8					
Bromochloromethane	0.332	0.323	0.320	0.267	0.270	0.273	0.297	10.2					
Chloroform	0.886	0.912	0.901	0.835	0.842	0.874	0.875	3.5					
1,1,1-Trichloroethane	0.808	0.814	0.822	0.767	0.764	0.796	0.795	3.1					
Methylcyclohexane	0.491	0.486	0.537	0.527	0.554	0.583	0.530	7					
Benzene	1.253	1.302	1.322	1.259	1.281	1.312	1.288	2.2					
1,2-Dichloroethane	0.328	0.346	0.343	0.315	0.324	0.338	0.332	3.6					
Trichloroethene	0.380	0.381	0.386	0.367	0.372	0.376	0.377	1.8					
1,2-Dichloropropane	0.276	0.289	0.292	0.274	0.283	0.293	0.285	2.9					
Bromodichloromethane	0.440	0.463	0.459	0.439	0.449	0.468	0.453	2.7					
4-Methyl-2-Pentanone	0.141	0.159	0.161	0.159	0.180	0.196	0.166	11.6					
Toluene	0.777	0.818	0.855	0.827	0.855	0.885	0.836	4.5					

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG No.: Q3419
 Instrument ID: MSVOA_Y Calibration Date(s): 10/07/2025 10/07/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 09:17 12:32
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY023384.D		RRF010 = VY023385.D		RRF020 = VY023386.D			
		RRF050 = VY023387.D		RRF100 = VY023388.D		RRF150 = VY023389.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.357	0.380	0.390	0.385	0.415	0.438	0.394	7.2	
cis-1,3-Dichloropropene	0.424	0.460	0.465	0.455	0.481	0.505	0.465	5.9	
1,1,2-Trichloroethane	0.234	0.247	0.241	0.231	0.244	0.254	0.242	3.5	
2-Hexanone	0.108	0.117	0.110	0.112	0.126	0.138	0.119	9.9	
Dibromochloromethane	0.319	0.337	0.340	0.327	0.347	0.365	0.339	4.7	
1,2-Dibromoethane	0.215	0.229	0.231	0.223	0.237	0.249	0.231	5.2	
Tetrachloroethene	0.538	0.517	0.561	0.523	0.502	0.456	0.516	6.9	
Chlorobenzene	1.086	1.095	1.092	1.053	1.083	1.116	1.087	1.9	
Ethyl Benzene	1.606	1.711	1.753	1.777	1.824	1.891	1.760	5.5	
m/p-Xylenes	0.632	0.675	0.715	0.719	0.734	0.763	0.706	6.5	
o-Xylene	0.588	0.623	0.653	0.670	0.701	0.738	0.662	8.1	
Styrene	0.902	1.026	1.098	1.109	1.166	1.233	1.089	10.5	
Bromoform	0.225	0.242	0.238	0.230	0.244	0.262	0.240	5.4	
Isopropylbenzene	2.967	3.153	3.353	3.317	3.425	3.555	3.295	6.3	
1,1,2,2-Tetrachloroethane	0.530	0.548	0.502	0.499	0.543	0.611	0.539	7.6	
1,3-Dichlorobenzene	1.667	1.698	1.718	1.657	1.701	1.794	1.706	2.9	
1,4-Dichlorobenzene	1.703	1.703	1.694	1.624	1.653	1.732	1.685	2.3	
1,2-Dichlorobenzene	1.485	1.498	1.511	1.446	1.490	1.543	1.496	2.1	
1,2-Dibromo-3-Chloropropane	0.087	0.093	0.083	0.083	0.088	0.096	0.088	6.2	
1,2,4-Trichlorobenzene	0.846	0.867	0.901	0.918	1.007	1.054	0.932	8.7	
1,2,3-Trichlorobenzene	0.719	0.753	0.795	0.793	0.866	0.920	0.808	9.1	
1,2-Dichloroethane-d4	0.455	0.436	0.429	0.392	0.392	0.405	0.418	6.2	
Dibromofluoromethane	0.311	0.314	0.317	0.298	0.299	0.303	0.307	2.6	
Toluene-d8	1.132	1.170	1.177	1.129	1.122	1.135	1.144	2	
4-Bromofluorobenzene	0.401	0.370	0.373	0.365	0.372	0.386	0.378	3.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\
 Method File : 82Y100725S.M
 Title : SW846 8260
 Last Update : Wed Oct 08 05:12:50 2025
 Response Via : Initial Calibration

Calibration Files

5 =VY023384.D 10 =VY023385.D 20 =VY023386.D 50 =VY023387.D 100 =VY023388.D 150 =VY023389.D

	Compound	5	10	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.397	0.399	0.400	0.321	0.310	0.333	0.360	11.97
3) P	Chloromethane	0.518	0.557	0.512	0.449	0.438	0.467	0.490	9.45
4) C	Vinyl Chloride	0.660	0.684	0.659	0.598	0.597	0.633	0.639	5.59#
5) T	Bromomethane	0.601	0.596	0.548	0.469	0.483	0.518	0.536	10.42
6) T	Chloroethane	0.441	0.462	0.453	0.414	0.413	0.429	0.435	4.67
7) T	Trichlorofluor...	0.884	0.920	0.946	0.842	0.850	0.877	0.887	4.54
8) T	Diethyl Ether	0.240	0.239	0.236	0.215	0.224	0.237	0.232	4.42
9) T	1,1,2-Trichlor...	0.484	0.486	0.479	0.437	0.429	0.445	0.460	5.57
10) T	Methyl Iodide	0.404	0.503	0.518	0.538	0.547	0.562	0.512	11.09
11) T	Tert butyl alc...	0.033	0.037	0.030	0.025	0.029	0.033	0.031	13.08
12) CM	1,1-Dichloroet...	0.449	0.475	0.456	0.427	0.426	0.443	0.446	4.14#
13) T	Acrolein	0.045	0.043	0.042	0.042	0.044	0.048	0.044	5.27
14) T	Allyl chloride	0.540	0.555	0.546	0.526	0.535	0.560	0.544	2.33
15) T	Acrylonitrile	0.088	0.093	0.086	0.082	0.088	0.095	0.089	5.52
16) T	Acetone	0.128	0.117	0.094	0.084	0.085	0.089	0.100	18.67
17) T	Carbon Disulfide	1.291	1.367	1.320	1.225	1.210	1.246	1.277	4.74
18) T	Methyl Acetate	0.244	0.265	0.260	0.245	0.266	0.286	0.261	6.00
19) T	Methyl tert-bu...	1.006	1.075	1.075	1.020	1.091	1.165	1.072	5.30
20) T	Methylene Chlo...	0.625	0.634	0.576	0.458	0.463	0.473	0.538	15.41
21) T	trans-1,2-Dich...	0.494	0.504	0.507	0.470	0.478	0.494	0.491	2.95
22) T	Diisopropyl ether	1.107	1.238	1.255	1.163	1.195	1.257	1.202	4.95
23) T	Vinyl Acetate	0.594	0.646	0.651	0.625	0.667	0.722	0.651	6.65
24) P	1,1-Dichloroet...	0.800	0.835	0.827	0.759	0.765	0.792	0.796	3.90
25) T	2-Butanone	0.133	0.134	0.115	0.105	0.115	0.126	0.122	9.56
26) T	2,2-Dichloropr...	0.752	0.771	0.731	0.705	0.699	0.726	0.731	3.74
27) T	cis-1,2-Dichlo...	0.542	0.584	0.582	0.542	0.563	0.591	0.567	3.81
28) T	Bromochloromet...	0.332	0.323	0.320	0.267	0.270	0.273	0.297	10.22
29) T	Tetrahydrofuran	0.066	0.070	0.067	0.064	0.070	0.078	0.069	7.25
30) C	Chloroform	0.886	0.912	0.901	0.835	0.842	0.874	0.875	3.54#
31) T	Cyclohexane	0.797	0.717	0.697	0.643	0.644	0.676	0.696	8.27
32) T	1,1,1-Trichlor...	0.808	0.814	0.822	0.767	0.764	0.796	0.795	3.08
33) S	1,2-Dichloroet...	0.455	0.436	0.429	0.392	0.392	0.405	0.418	6.21
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...	0.311	0.314	0.317	0.298	0.299	0.303	0.307	2.62
36) T	1,1-Dichloropr...	0.395	0.413	0.413	0.397	0.405	0.417	0.407	2.20
37) T	Ethyl Acetate	0.148	0.162	0.163	0.150	0.164	0.179	0.161	6.94
38) T	Carbon Tetrach...	0.483	0.501	0.504	0.481	0.487	0.503	0.493	2.14
39) T	Methylcyclohexane	0.491	0.486	0.537	0.527	0.554	0.583	0.530	7.03
40) TM	Benzene	1.253	1.302	1.322	1.259	1.281	1.312	1.288	2.22
41) T	Methacrylonitrile	0.082	0.085	0.091	0.080	0.089	0.096	0.087	7.09
42) TM	1,2-Dichloroet...	0.328	0.346	0.343	0.315	0.324	0.338	0.332	3.59
43) T	Isopropyl Acetate	0.297	0.313	0.313	0.302	0.335	0.369	0.321	8.32
44) TM	Trichloroethene	0.380	0.381	0.386	0.367	0.372	0.376	0.377	1.81
45) C	1,2-Dichloropr...	0.276	0.289	0.292	0.274	0.283	0.293	0.285	2.89#
46) T	Dibromomethane	0.172	0.182	0.183	0.173	0.181	0.188	0.180	3.43
47) T	Bromodichlorom...	0.440	0.463	0.459	0.439	0.449	0.468	0.453	2.70
48) T	Methyl methacr...	0.125	0.139	0.149	0.148	0.165	0.181	0.151	12.98
49) T	1,4-Dioxane	0.002	0.002	0.002	0.002	0.002	0.002	0.002	10.06
50) S	Toluene-d8	1.132	1.170	1.177	1.129	1.122	1.135	1.144	2.03
51) T	4-Methyl-2-Pen...	0.141	0.159	0.161	0.159	0.180	0.196	0.166	11.55
52) CM	Toluene	0.777	0.818	0.855	0.827	0.855	0.885	0.836	4.49#
53) T	t-1,3-Dichloro...	0.357	0.380	0.390	0.385	0.415	0.438	0.394	7.24
54) T	cis-1,3-Dichlo...	0.424	0.460	0.465	0.455	0.481	0.505	0.465	5.86
55) T	1,1,2-Trichlor...	0.234	0.247	0.241	0.231	0.244	0.254	0.242	3.52
56) T	Ethyl methacry...	0.219	0.248	0.262	0.279	0.321	0.345	0.279	16.80

Method Path : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\

Method File : 82Y100725S.M

57)	T	1,3-Dichloropr...	0.361	0.381	0.383	0.374	0.392	0.409	0.383	4.26
58)	T	2-Chloroethyl ...	0.093	0.112	0.112	0.129	0.145	0.159	0.125	19.35
59)	T	2-Hexanone	0.108	0.117	0.110	0.112	0.126	0.138	0.119	9.86
60)	T	Dibromochlorom...	0.319	0.337	0.340	0.327	0.347	0.365	0.339	4.69
61)	T	1,2-Dibromoethane	0.215	0.229	0.231	0.223	0.237	0.249	0.231	5.16
62)	S	4-Bromofluorob...	0.401	0.370	0.373	0.365	0.372	0.386	0.378	3.51
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.538	0.517	0.561	0.523	0.502	0.456	0.516	6.92
65)	PM	Chlorobenzene	1.086	1.095	1.092	1.053	1.083	1.116	1.087	1.89
66)	T	1,1,1,2-Tetrac...	0.388	0.389	0.390	0.373	0.384	0.402	0.388	2.44
67)	C	Ethyl Benzene	1.606	1.711	1.753	1.777	1.824	1.891	1.760	5.54#
68)	T	m/p-Xylenes	0.632	0.675	0.715	0.719	0.734	0.763	0.706	6.54
69)	T	o-Xylene	0.588	0.623	0.653	0.670	0.701	0.738	0.662	8.15
70)	T	Styrene	0.902	1.026	1.098	1.109	1.166	1.233	1.089	10.55
71)	P	Bromoform	0.225	0.242	0.238	0.230	0.244	0.262	0.240	5.41
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	2.967	3.153	3.353	3.317	3.425	3.555	3.295	6.30
74)	T	N-amyl acetate	0.477	0.535	0.556	0.581	0.665	0.738	0.592	15.95
75)	P	1,1,2,2-Tetrac...	0.530	0.548	0.502	0.499	0.543	0.611	0.539	7.60
76)	T	1,2,3-Trichlor...	0.406	0.399	0.484	0.433	0.455	0.469	0.441	7.76
77)	T	Bromobenzene	0.849	0.880	0.885	0.845	0.877	0.921	0.876	3.14
78)	T	n-propylbenzene	3.373	3.676	3.906	3.883	3.955	4.086	3.813	6.64
79)	T	2-Chlorotoluene	2.095	2.208	2.282	2.227	2.249	2.338	2.233	3.66
80)	T	1,3,5-Trimethy...	2.309	2.557	2.736	2.706	2.772	2.896	2.663	7.70
81)	T	trans-1,4-Dich...	0.171	0.175	0.176	0.169	0.185	0.203	0.180	7.04
82)	T	4-Chlorotoluene	2.173	2.266	2.362	2.276	2.316	2.404	2.300	3.52
83)	T	tert-Butylbenzene	2.199	2.363	2.465	2.522	2.533	2.668	2.458	6.55
84)	T	1,2,4-Trimethy...	2.301	2.552	2.701	2.706	2.782	2.903	2.657	7.86
85)	T	sec-Butylbenzene	3.175	3.388	3.571	3.583	3.664	3.778	3.527	6.08
86)	T	p-Isopropyltol...	2.575	2.804	3.032	3.087	3.183	3.334	3.003	9.09
87)	T	1,3-Dichlorobe...	1.667	1.698	1.718	1.657	1.701	1.794	1.706	2.85
88)	T	1,4-Dichlorobe...	1.703	1.703	1.694	1.624	1.653	1.732	1.685	2.33
89)	T	n-Butylbenzene	2.279	2.428	2.626	2.682	2.778	2.904	2.616	8.76
90)	T	Hexachloroethane	0.667	0.631	0.647	0.625	0.643	0.662	0.646	2.54
91)	T	1,2-Dichlorobe...	1.485	1.498	1.511	1.446	1.490	1.543	1.496	2.13
92)	T	1,2-Dibromo-3-...	0.087	0.093	0.083	0.083	0.088	0.096	0.088	6.18
93)	T	1,2,4-Trichlor...	0.846	0.867	0.901	0.918	1.007	1.054	0.932	8.75
94)	T	Hexachlorobuta...	0.590	0.571	0.572	0.577	0.587	0.590	0.581	1.52
95)	T	Naphthalene	1.206	1.281	1.346	1.466	1.722	1.887	1.485	17.98
96)	T	1,2,3-Trichlor...	0.719	0.753	0.795	0.793	0.866	0.920	0.808	9.14

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023384.D
 Acq On : 07 Oct 2025 09:17
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:39:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	585717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	882488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	753398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	378104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	26670	5.258	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	10.520%#		
35) Dibromofluoromethane	7.634	113	27455	5.126	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	10.260%#		
50) Toluene-d8	10.103	98	99905	4.946	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	9.900%#		
62) 4-Bromofluorobenzene	12.401	95	35355	5.548	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	11.100%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	23245	5.038	ug/l	92
3) Chloromethane	2.074	50	30346	4.627	ug/l	94
4) Vinyl Chloride	2.202	62	38657	4.668	ug/l	95
5) Bromomethane	2.598	94	35214	5.184	ug/l	94
6) Chloroethane	2.732	64	25848	4.597	ug/l	98
7) Trichlorofluoromethane	3.055	101	51797	4.869	ug/l	98
8) Diethyl Ether	3.452	74	14079	5.060	ug/l	96
9) 1,1,2-Trichlorotrifluo...	3.818	101	28368	4.989	ug/l	99
10) Methyl Iodide	4.007	142	23692	3.606	ug/l	99
11) Tert butyl alcohol	4.854	59	9774	32.746	ug/l #	80
12) 1,1-Dichloroethene	3.787	96	26290	4.806	ug/l	94
13) Acrolein	3.653	56	13176	26.054	ug/l	100
14) Allyl chloride	4.385	41	31614	4.616	ug/l	98
15) Acrylonitrile	5.055	53	25891	23.754	ug/l	99
16) Acetone	3.866	43	37573	32.940	ug/l	96
17) Carbon Disulfide	4.104	76	75645	4.432	ug/l	99
18) Methyl Acetate	4.391	43	14279	4.926	ug/l	98
19) Methyl tert-butyl Ether	5.116	73	58919	4.600	ug/l	95
20) Methylene Chloride	4.616	84	36598	5.917	ug/l	100
21) trans-1,2-Dichloroethene	5.110	96	28923	4.754	ug/l	99
22) Diisopropyl ether	6.018	45	64833	4.269	ug/l	94
23) Vinyl Acetate	5.964	43	173851	20.824	ug/l	99
24) 1,1-Dichloroethane	5.909	63	46850	4.679	ug/l	100
25) 2-Butanone	6.890	43	38990	27.656	ug/l	93
26) 2,2-Dichloropropane	6.884	77	44062	4.953	ug/l	98
27) cis-1,2-Dichloroethene	6.884	96	31743	4.600	ug/l	98
28) Bromochloromethane	7.244	49	19445	5.006	ug/l	99
29) Tetrahydrofuran	7.262	42	19296	23.059	ug/l	98
30) Chloroform	7.421	83	51887	4.806	ug/l	97
31) Cyclohexane	7.707	56	46675	5.279	ug/l #	79
32) 1,1,1-Trichloroethane	7.616	97	47305	4.881	ug/l	99
36) 1,1-Dichloropropene	7.835	75	34888	4.617	ug/l	98
37) Ethyl Acetate	6.982	43	13082	4.379	ug/l #	93
38) Carbon Tetrachloride	7.817	117	42654	4.859	ug/l	95
39) Methylcyclohexane	9.109	83	43329	4.479	ug/l	98
40) Benzene	8.079	78	110551	4.693	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023384.D
 Acq On : 07 Oct 2025 09:17
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:39:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	7226m	4.658	ug/l	
42) 1,2-Dichloroethane	8.158	62	28940	4.828	ug/l	100
43) Isopropyl Acetate	8.201	43	26166	4.503	ug/l	99
44) Trichloroethene	8.859	130	33564	5.077	ug/l	99
45) 1,2-Dichloropropane	9.140	63	24358	4.595	ug/l	92
46) Dibromomethane	9.231	93	15179	4.682	ug/l	97
47) Bromodichloromethane	9.420	83	38813	4.756	ug/l	93
48) Methyl methacrylate	9.219	41	11038	4.071	ug/l	98
49) 1,4-Dioxane	9.219	88	3255	96.953	ug/l #	86
51) 4-Methyl-2-Pentanone	9.999	43	62279	20.925	ug/l	98
52) Toluene	10.170	92	68550	4.585	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	31477	4.491	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	37392	4.463	ug/l	96
55) 1,1,2-Trichloroethane	10.572	97	20641	4.786	ug/l	97
56) Ethyl methacrylate	10.438	69	19324	3.891	ug/l	94
57) 1,3-Dichloropropane	10.719	76	31831	4.560	ug/l	96
58) 2-Chloroethyl Vinyl ether	9.713	63	41054m	18.894	ug/l	
59) 2-Hexanone	10.761	43	47778	23.185	ug/l	97
60) Dibromochloromethane	10.908	129	28135	4.783	ug/l	97
61) 1,2-Dibromoethane	11.017	107	18957	4.673	ug/l	97
64) Tetrachloroethene	10.646	164	40570	5.383	ug/l	98
65) Chlorobenzene	11.444	112	81802	4.997	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.517	131	29222	5.075	ug/l	97
67) Ethyl Benzene	11.517	91	121005	4.529	ug/l	99
68) m/p-Xylenes	11.633	106	95176	8.887	ug/l	100
69) o-Xylene	11.956	106	44280	4.466	ug/l	98
70) Styrene	11.969	104	67960	4.152	ug/l	99
71) Bromoform	12.133	173	16988	4.929	ug/l #	99
73) Isopropylbenzene	12.255	105	112194	4.455	ug/l	99
74) N-amyl acetate	12.066	43	18041	3.833	ug/l #	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	20052	4.773	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	15359m	4.118	ug/l	
77) Bromobenzene	12.529	156	32106	4.967	ug/l	99
78) n-propylbenzene	12.596	91	127521	4.305	ug/l	99
79) 2-Chlorotoluene	12.676	91	79199	4.631	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	87297	4.264	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	6474	4.680	ug/l	98
82) 4-Chlorotoluene	12.779	91	82155	4.600	ug/l	100
83) tert-Butylbenzene	12.999	119	83151	4.457	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	86988	4.294	ug/l	99
85) sec-Butylbenzene	13.176	105	120061	4.443	ug/l	99
86) p-Isopropyltoluene	13.291	119	97377	4.264	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	63027	4.936	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	64389	5.085	ug/l	95
89) n-Butylbenzene	13.615	91	86179	4.274	ug/l	99
90) Hexachloroethane	13.877	117	25204	5.084	ug/l	97
91) 1,2-Dichlorobenzene	13.657	146	56167	5.041	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	3305	4.964	ug/l	95
93) 1,2,4-Trichlorobenzene	14.919	180	32000	4.875	ug/l	99
94) Hexachlorobutadiene	15.023	225	22316	5.456	ug/l	99
95) Naphthalene	15.145	128	45598	4.360	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	27176	4.772	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023384.D
Acq On : 07 Oct 2025 09:17
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:39:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 04:39:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023384.D
Acq On : 07 Oct 2025 09:17
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Manual Integrations

APPROVED

Reviewed By : Semsettin Yesilyurt 10/10/2025

Supervised By : Mahesh Dadoda 10/10/2025


Quant Time: Oct 08 04:39:47 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 04:39:16 2025


Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:168 Resp: 58571
Ion Ratio Lower Upper
168 100
99 49.8 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 5.038 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 85 Resp: 23245
Ion Ratio Lower Upper
85 100
87 36.2 16.0 47.9





#3

Chloromethane

Concen: 4.627 ug/l

RT: 2.074 min Scan# 57

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 50 Resp: 30340

Ion Ratio Lower Upper

50 100





52 35.8 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#4

Vinyl Chloride

Concen: 4.668 ug/l

RT: 2.202 min Scan# 79

Delta R.T. -0.000 min

Lab File: VY023384.D




Acq: 07 Oct 2025 09:17


Tgt Ion: 62 Resp: 38657

Ion Ratio Lower Upper

62 100

64 35.0 25.8 38.8









#5
Bromomethane
Concen: 5.184 ug/l
RT: 2.598 min Scan# 144
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 94 Resp: 35214
Ion Ratio Lower Upper
94 100
96 99.0 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#6
Chloroethane
Concen: 4.597 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 64 Resp: 25848
Ion Ratio Lower Upper
64 100
66 31.9 26.2 39.4









#7
Trichlorofluoromethane
Concen: 4.869 ug/l
RT: 3.055 min Scan# 219
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:101 Resp: 5179
Ion Ratio Lower Upper
101 100
103 62.4 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#8
Diethyl Ether
Concen: 5.060 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 74 Resp: 14079
Ion Ratio Lower Upper
74 100
45 82.9 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.989 ug/l
RT: 3.818 min Scan# 343
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:101 Resp: 2836
Ion Ratio Lower Upper
101 100
85 43.5 34.6 51.8
151 82.3 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#10
Methyl Iodide
Concen: 3.606 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion:142 Resp: 23692
Ion Ratio Lower Upper
142 100
127 43.3 34.8 52.2
141 16.2 11.6 17.4





#11

Tert butyl alcohol

Concen: 32.746 ug/l

RT: 4.854 min Scan# 515

Delta R.T. -0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 9774

Ion Ratio Lower Upper

59 100





57 3.0 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 4.806 ug/l

RT: 3.787 min Scan# 339

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion: 96 Resp: 26290


Ion Ratio Lower Upper

96 100

61 141.0 106.7 160.1

98 59.1 50.2 75.4





#13

Acrolein

Concen: 26.054 ug/l

RT: 3.653 min Scan# 316

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 13170

Ion Ratio Lower Upper

56 100





55 70.1 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#14

Allyl chloride

Concen: 4.616 ug/l

RT: 4.385 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion: 41 Resp: 31614


Ion Ratio Lower Upper

41 100

39 76.6 59.9 89.9

76 46.2 38.0 57.0





#15

Acrylonitrile

Concen: 23.754 ug/l

RT: 5.055 min Scan# 547

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y





ClientSampleId :

VSTDICC005

Tgt Ion: 53	Resp: 2589
Ion Ratio	Lower Upper
53 100	
52 79.7	64.6 97.0
51 36.3	29.5 44.3

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 32.940 ug/l




RT: 3.866 min Scan# 352


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion: 43	Resp: 37573
Ion Ratio	Lower Upper
43 100	
58 32.1	27.7 41.5





#17

Carbon Disulfide

Concen: 4.432 ug/l

RT: 4.104 min Scan# 391

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 7564

Ion Ratio Lower Upper

76 100





78 8.8 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#18

Methyl Acetate

Concen: 4.926 ug/l

RT: 4.391 min Scan# 438

Delta R.T. 0.006 min

Lab File: VY023384.D




Acq: 07 Oct 2025 09:17


Tgt Ion: 43 Resp: 14279

Ion Ratio Lower Upper

43 100

74 31.4 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 4.600 ug/l

RT: 5.116 min Scan# 5

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 73 Resp: 58919

Ion Ratio Lower Upper

73 100





57 19.4 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 5.917 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion: 84 Resp: 36598




Ion Ratio Lower Upper


84 100

49 100.3 80.2 120.4

51 31.6 25.9 38.9

86 63.2 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 4.754 ug/l

RT: 5.110 min Scan# 511

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:	96	Resp:	2892
Ion	Ratio	Lower	Upper
96	100		
61	114.7	92.7	139.1
98	61.0	49.7	74.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#22

Diisopropyl ether

Concen: 4.269 ug/l




RT: 6.018 min Scan# 705


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	45	Resp:	64833
Ion	Ratio	Lower	Upper
45	100		
43	63.0	45.6	68.4
87	32.9	28.1	42.1
59	13.5	10.2	15.2





#23

Vinyl Acetate

Concen: 20.824 ug/l

RT: 5.964 min Scan# 64

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 17385

Ion Ratio Lower Upper

43

100

86

13.6

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 4.679 ug/l

RT: 5.909 min Scan# 687

Delta R.T. -0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion: 63 Resp: 46850

Ion Ratio Lower Upper

63

100

98

8.8

4.3




13.1


100

5.8

2.8

8.4





#25

2-Butanone

Concen: 27.656 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 38990

Ion Ratio Lower Upper

43 100





72 30.0 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 4.953 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY023384.D




Acq: 07 Oct 2025 09:17


Tgt Ion: 77 Resp: 44062

Ion Ratio Lower Upper

77 100

97 23.4 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 4.600 ug/l

RT: 6.884 min Scan# 848

Delta R.T. -0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC005

Tgt Ion:	96	Resp:	3174
Ion	Ratio	Lower	Upper
96	100		
61	133.3	0.0	258.2
98	66.1	0.0	132.0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane

Concen: 5.006 ug/l




RT: 7.244 min Scan# 906


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	49	Resp:	19445
Ion	Ratio	Lower	Upper
49	100		
129	3.1	0.0	5.6
130	101.5	81.9	122.9





#29

Tetrahydrofuran

Concen: 23.059 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC005

Tgt Ion:	42	Resp:	19290
Ion	Ratio	Lower	Upper
42	100		
72	51.0	43.0	64.6
71	50.4	40.1	60.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 4.806 ug/l




RT: 7.421 min Scan# 935


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	83	Resp:	51887
Ion	Ratio	Lower	Upper
83	100		
85	63.0	52.5	78.7





#31

Cyclohexane

Concen: 5.279 ug/l

RT: 7.707 min Scan# 981

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 46675

Ion Ratio Lower Upper

56 100

69 62.0 30.0 45.0





84 115.8 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#32

1,1,1-Trichloroethane

Concen: 4.881 ug/l

RT: 7.616 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion: 97 Resp: 47305


Ion Ratio Lower Upper

97 100

99 63.1 51.4 77.2

61 39.4 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 5.258 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 26670

Ion Ratio Lower Upper

65 100





67 53.7 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion:114 Resp: 882488


Ion Ratio Lower Upper

114 100

63 16.7 0.0 34.2

88 14.9 0.0 28.4





#35

Dibromofluoromethane

Concen: 5.126 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion:113 Resp: 2745

Ion Ratio Lower Upper

113 100

111 99.0 81.8 122.6





192 23.4 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 4.617 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion: 75 Resp: 34888


Ion Ratio Lower Upper

75 100

110 38.3 19.9 59.7

77 32.1 25.0 37.6





#37

Ethyl Acetate

Concen: 4.379 ug/l

RT: 6.982 min Scan# 80

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	43	Resp:	1308
Ion	Ratio	Lower	Upper
43	100		
61	18.2	11.8	17.8
70	15.0	10.5	15.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 4.859 ug/l




RT: 7.817 min Scan# 1000


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	117	Resp:	42654
Ion	Ratio	Lower	Upper
117	100		
119	91.0	76.8	115.2
121	28.0	24.5	36.7





#39

Methylcyclohexane

Concen: 4.479 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:	83	Resp:	43329
Ion	Ratio	Lower	Upper
83	100		
55	67.2	51.4	77.2
98	50.8	40.4	60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 4.693 ug/l




RT: 8.079 min Scan# 1043


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17


Tgt Ion:	78	Resp:	110551
Ion	Ratio	Lower	Upper
78	100		
77	23.0	18.3	27.5





#41
Methacrylonitrile
Concen: 4.658 ug/l m
RT: 7.225 min Scan# 902
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 41 Resp: 7220
Ion Ratio Lower Upper
41 100
39 0.0 52.2 78.4
67 95.1 74.9 112.3
52 36.3 29.2 43.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#42
1,2-Dichloroethane
Concen: 4.828 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17



Tgt Ion: 62 Resp: 28940
Ion Ratio Lower Upper
62 100
98 10.6 0.0 21.6





#43

Isopropyl Acetate

Concen: 4.503 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion: 43	Resp: 2616
Ion Ratio	Lower Upper
43 100	
61 22.6	18.1 27.1
87 15.9	13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 5.077 ug/l




RT: 8.859 min Scan# 1171


Delta R.T. -0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:130	Resp: 33564
Ion Ratio	Lower Upper
130 100	
95 89.2	0.0 180.6





#45

1,2-Dichloropropane

Concen: 4.595 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 2435

Ion Ratio Lower Upper

63 100





65 35.7 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 4.682 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion: 93 Resp: 15179


Ion Ratio Lower Upper

93 100

95 87.6 65.4 98.0

174 116.2 92.9 139.3





#47

Bromodichloromethane

Concen: 4.756 ug/l

RT: 9.420 min Scan# 1263

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:	83	Resp:	3881
Ion	Ratio	Lower	Upper
83	100		
85	57.9	50.8	76.2
127	8.4	7.4	11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#48

Methyl methacrylate

Concen: 4.071 ug/l




RT: 9.219 min Scan# 1230


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	41	Resp:	11038
Ion	Ratio	Lower	Upper
41	100		
69	102.5	81.0	121.4
39	62.4	48.2	72.4





#49

1,4-Dioxane

Concen: 96.953 ug/l

RT: 9.219 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:	88	Resp:	325
Ion	Ratio	Lower	Upper
88	100		
43	43.1	21.8	32.8
58	63.4	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 4.946 ug/l




RT: 10.103 min Scan# 1375


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	98	Resp:	99905
Ion	Ratio	Lower	Upper
98	100		
100	65.3	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 20.925 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 62279
Ion Ratio Lower Upper
43 100
58 42.7 35.3 52.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#52
Toluene
Concen: 4.585 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 92 Resp: 68550
Ion Ratio Lower Upper
92 100
91 171.2 136.4 204.6









#53
t-1,3-Dichloropropene
Concen: 4.491 ug/l
RT: 10.396 min Scan# 1423
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 75 Resp: 3147
Ion Ratio Lower Upper
75 100
77 32.1 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#54
cis-1,3-Dichloropropene
Concen: 4.463 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 75 Resp: 37392
Ion Ratio Lower Upper
75 100
77 34.2 25.4 38.0
39 43.0 32.8 49.2





#55

1,1,2-Trichloroethane

Concen: 4.786 ug/l

RT: 10.572 min Scan# 1452

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 2064

Ion Ratio Lower Upper

97 100

83 80.3 61.8 92.6

85 52.8 41.3 61.9





99 57.9 48.7 73.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#56

Ethyl methacrylate

Concen: 3.891 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion: 69 Resp: 19324


Ion Ratio Lower Upper

69 100

41 60.6 44.6 67.0

39 36.2 27.0 40.4









#57
1,3-Dichloropropane
Concen: 4.560 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 76 Resp: 3183
Ion Ratio Lower Upper
76 100
78 34.4 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#58
2-Chloroethyl Vinyl ether
Concen: 18.894 ug/l m
RT: 9.713 min Scan# 1311
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 63 Resp: 41054
Ion Ratio Lower Upper
63 100
106 34.1 26.6 39.8





#59

2-Hexanone

Concen: 23.185 ug/l

RT: 10.761 min Scan# 1482

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 47778

Ion Ratio Lower Upper

43 100





58 58.4 30.5 91.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#60

Dibromochloromethane

Concen: 4.783 ug/l

RT: 10.908 min Scan# 1507

Delta R.T. -0.000 min

Lab File: VY023384.D




Acq: 07 Oct 2025 09:17


Tgt Ion:129 Resp: 28135

Ion Ratio Lower Upper

129 100


127 79.7 38.6 115.7





#61
1,2-Dibromoethane
Concen: 4.673 ug/l
RT: 11.017 min Scan# 1107
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:107 Resp: 1895
Ion Ratio Lower Upper
107 100
109 97.6 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 5.548 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17


Tgt Ion: 95 Resp: 35355
Ion Ratio Lower Upper
95 100
174 97.9 0.0 192.8
176 94.0 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:117 Resp: 753398
Ion Ratio Lower Upper
117 100
82 50.3 41.0 61.4
119 32.1 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#64
Tetrachloroethene
Concen: 5.383 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17


Tgt Ion:164 Resp: 40570
Ion Ratio Lower Upper
164 100
166 128.1 99.4 149.0
129 87.4 69.4 104.2
131 83.1 67.5 101.3





#65
Chlorobenzene
Concen: 4.997 ug/l
RT: 11.444 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:112 Resp: 8180
Ion Ratio Lower Upper
112 100
114 30.8 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 5.075 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion:131 Resp: 29222
Ion Ratio Lower Upper
131 100
133 91.7 48.4 145.2
119 64.1 32.4 97.2





#67

Ethyl Benzene

Concen: 4.529 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 12100

Ion Ratio Lower Upper

91 100





106 33.0 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#68

m/p-Xylenes

Concen: 8.887 ug/l

RT: 11.633 min Scan# 1626

Delta R.T. 0.006 min

Lab File: VY023384.D




Acq: 07 Oct 2025 09:17


Tgt Ion:106 Resp: 95176

Ion Ratio Lower Upper

106 100

91 190.5 152.6 229.0





#69

o-Xylene

Concen: 4.466 ug/l

RT: 11.956 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 44280

Ion Ratio Lower Upper

106 100





91 199.7 101.4 304.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#70

Styrene

Concen: 4.152 ug/l

RT: 11.969 min Scan# 1681

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17




Tgt Ion:104 Resp: 67960


Ion Ratio Lower Upper

104 100

78 45.1 36.1 54.1


103 55.9 43.2 64.8





#71
Bromoform
Concen: 4.929 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion:173 Resp: 16988
Ion Ratio Lower Upper
173 100
175 48.5 24.8 74.3
254 0.0 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Tgt Ion:152 Resp: 378104
Ion Ratio Lower Upper
152 100
115 53.5 27.1 81.2
150 156.6 0.0 347.4





#73
Isopropylbenzene
Concen: 4.455 ug/l
RT: 12.255 min Scan# 1727
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:105 Resp: 112194
Ion Ratio Lower Upper
105 100
120 27.2 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 3.833 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 43 Resp: 18041
Ion Ratio Lower Upper
43 100
70 53.0 41.1 61.7
55 35.3 23.0 34.6#
61 26.5 21.4 32.2





#75

1,1,2,2-Tetrachloroethane

Concen: 4.773 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	83	Resp:	2005
Ion	Ratio	Lower	Upper
83	100		
131	11.6	5.9	17.6
85	63.4	31.8	95.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#76

1,2,3-Trichloropropane

Concen: 4.118 ug/l m




RT: 12.554 min Scan# 1777


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17


Tgt Ion:	75	Resp:	15359
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0





#77
Bromobenzene
Concen: 4.967 ug/l
RT: 12.529 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion: 156 Resp: 32100
Ion Ratio Lower Upper
156 100
77 154.2 76.8 230.5
158 95.0 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#78
n-propylbenzene
Concen: 4.305 ug/l
RT: 12.596 min Scan# 1784
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 91 Resp: 127521
Ion Ratio Lower Upper
91 100
120 25.1 12.3 36.8





#79

2-Chlorotoluene

Concen: 4.631 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 79199

Ion Ratio Lower Upper

91 100





126 36.4 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 4.264 ug/l

RT: 12.737 min Scan# 1807

Delta R.T. 0.006 min

Lab File: VY023384.D




Acq: 07 Oct 2025 09:17


Tgt Ion:105 Resp: 87297

Ion Ratio Lower Upper

105 100


120 50.6 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 4.680 ug/l
RT: 12.304 min Scan# 1735
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion: 75 Resp: 6474
Ion Ratio Lower Upper
75 100
53 85.6 68.6 102.8
89 44.4 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#82
4-Chlorotoluene
Concen: 4.600 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 91 Resp: 82155
Ion Ratio Lower Upper
91 100
126 36.9 18.4 55.4





#83

tert-Butylbenzene

Concen: 4.457 ug/l

RT: 12.999 min Scan# 1850

Delta R.T. 0.006 min

Lab File: VY023384.D


Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005



Tgt Ion:119 Resp: 8315

Ion Ratio Lower Upper

119 100

91 60.0 29.9 89.8




134 26.2 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#84

1,2,4-Trimethylbenzene


Concen: 4.294 ug/l

RT: 13.042 min Scan# 1857

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17






Tgt Ion:105 Resp: 86988

Ion Ratio Lower Upper

105 100


120 46.8 23.8 71.2





#85
sec-Butylbenzene
Concen: 4.443 ug/l
RT: 13.176 min Scan# 1878
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion:105 Resp: 12006
Ion Ratio Lower Upper
105 100
134 20.9 10.7 32.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#86
p-Isopropyltoluene
Concen: 4.264 ug/l
RT: 13.291 min Scan# 1898
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion:119 Resp: 97377
Ion Ratio Lower Upper
119 100
134 27.5 13.6 40.7
91 22.7 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 4.936 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC005

Tgt Ion:	146	Resp:	6302
Ion	Ratio	Lower	Upper
146	100		
111	37.5	18.7	56.1
148	64.1	31.9	95.9

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 5.085 ug/l




RT: 13.365 min Scan# 1910


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:	146	Resp:	64389
Ion	Ratio	Lower	Upper
146	100		
111	43.0	18.0	54.0
148	65.1	31.9	95.7





#89

n-Butylbenzene

Concen: 4.274 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	91	Resp:	86179
Ion	Ratio	Lower	Upper
91	100		
92	53.3	27.0	80.8
134	29.2	14.1	42.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#90

Hexachloroethane

Concen: 5.084 ug/l




RT: 13.877 min Scan# 1994


Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17


Tgt Ion:	117	Resp:	25204
Ion	Ratio	Lower	Upper
117	100		
201	84.7	43.8	131.4





#91
1,2-Dichlorobenzene
Concen: 5.041 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:146 Resp: 5616
Ion Ratio Lower Upper
146 100
111 37.8 19.0 57.0
148 63.0 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 4.964 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion: 75 Resp: 3305
Ion Ratio Lower Upper
75 100
155 91.7 50.0 150.0
157 124.4 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 4.875 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:180	Resp:	32000
Ion Ratio	Lower	Upper
180	100	
182	97.1	47.9 143.6
145	29.9	14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene

Concen: 5.456 ug/l




RT: 15.023 min Scan# 2182


Delta R.T. 0.006 min

Lab File: VY023384.D

Acq: 07 Oct 2025 09:17

Tgt Ion:225	Resp:	22316
Ion Ratio	Lower	Upper
225	100	
223	62.8	30.8 92.3
227	63.1	31.9 95.7









#95
Naphthalene
Concen: 4.360 ug/l
RT: 15.145 min Scan# 2201
Delta R.T. 0.006 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:128 Resp: 45598
Ion Ratio Lower Upper
128 100
127 12.7 10.4 15.6
129 10.7 8.9 13.3




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#96
1,2,3-Trichlorobenzene
Concen: 4.772 ug/l
RT: 15.328 min Scan# 2232
Delta R.T. -0.000 min
Lab File: VY023384.D
Acq: 07 Oct 2025 09:17

Tgt Ion:180 Resp: 27176
Ion Ratio Lower Upper
180 100
182 96.3 47.9 143.8
145 32.7 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023385.D
 Acq On : 07 Oct 2025 10:02
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC010

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:40:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	559828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	847077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	736152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	374998	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	48826	10.071	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	20.140%#		
35) Dibromofluoromethane	7.634	113	53279	10.364	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	20.720%#		
50) Toluene-d8	10.109	98	198242	10.225	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	20.460%#		
62) 4-Bromofluorobenzene	12.401	95	62601	10.235	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	20.460%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	44666	10.129	ug/l	97
3) Chloromethane	2.068	50	62413	9.957	ug/l	97
4) Vinyl Chloride	2.202	62	76587	9.676	ug/l	99
5) Bromomethane	2.592	94	66689	10.271	ug/l	90
6) Chloroethane	2.732	64	51757	9.632	ug/l	92
7) Trichlorofluoromethane	3.055	101	102972	10.128	ug/l	99
8) Diethyl Ether	3.458	74	26719	10.047	ug/l	96
9) 1,1,2-Trichlorotrifluo...	3.818	101	54405	10.011	ug/l	99
10) Methyl Iodide	4.007	142	56313	8.967	ug/l	99
11) Tert butyl alcohol	4.854	59	20540	71.998	ug/l #	89
12) 1,1-Dichloroethene	3.793	96	53191	10.173	ug/l	98
13) Acrolein	3.653	56	24014	49.681	ug/l	94
14) Allyl chloride	4.385	41	62154	9.494	ug/l	99
15) Acrylonitrile	5.061	53	51937	49.854	ug/l	99
16) Acetone	3.872	43	65410	59.997	ug/l	94
17) Carbon Disulfide	4.104	76	153055	9.382	ug/l	100
18) Methyl Acetate	4.391	43	29725	10.728	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	120347	9.830	ug/l	100
20) Methylene Chloride	4.616	84	71034	12.016	ug/l	96
21) trans-1,2-Dichloroethene	5.116	96	56465	9.710	ug/l	97
22) Diisopropyl ether	6.024	45	138589	9.547	ug/l	98
23) Vinyl Acetate	5.963	43	361478	45.300	ug/l	99
24) 1,1-Dichloroethane	5.915	63	93512	9.772	ug/l	98
25) 2-Butanone	6.890	43	75171	55.784	ug/l	92
26) 2,2-Dichloropropane	6.890	77	86325	10.153	ug/l	99
27) cis-1,2-Dichloroethene	6.896	96	65335	9.905	ug/l	99
28) Bromochloromethane	7.244	49	36132	9.733	ug/l	99
29) Tetrahydrofuran	7.262	42	38990	48.749	ug/l	99
30) Chloroform	7.421	83	102058	9.890	ug/l	99
31) Cyclohexane	7.701	56	80245	9.495	ug/l #	88
32) 1,1,1-Trichloroethane	7.622	97	91186	9.844	ug/l	99
36) 1,1-Dichloropropene	7.835	75	69913	9.639	ug/l	99
37) Ethyl Acetate	6.988	43	27388	9.551	ug/l	96
38) Carbon Tetrachloride	7.817	117	84799	10.065	ug/l	99
39) Methylcyclohexane	9.109	83	82275	8.860	ug/l	95
40) Benzene	8.079	78	220629	9.758	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023385.D
 Acq On : 07 Oct 2025 10:02
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:40:53 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 04:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	14344	9.632	ug/l #	92
42) 1,2-Dichloroethane	8.158	62	58541	10.174	ug/l	99
43) Isopropyl Acetate	8.201	43	53042	9.510	ug/l	98
44) Trichloroethene	8.865	130	64537	10.171	ug/l	96
45) 1,2-Dichloropropane	9.140	63	49009	9.632	ug/l	100
46) Dibromomethane	9.231	93	30780	9.890	ug/l	99
47) Bromodichloromethane	9.426	83	78392	10.007	ug/l	99
48) Methyl methacrylate	9.225	41	23484	9.023	ug/l	97
49) 1,4-Dioxane	9.231	88	6822	211.693	ug/l	94
51) 4-Methyl-2-Pentanone	9.999	43	134746	47.165	ug/l	99
52) Toluene	10.170	92	138662	9.661	ug/l	100
53) t-1,3-Dichloropropene	10.396	75	64315	9.559	ug/l	96
54) cis-1,3-Dichloropropene	9.853	75	77905	9.687	ug/l	98
55) 1,1,2-Trichloroethane	10.572	97	41881	10.118	ug/l	92
56) Ethyl methacrylate	10.438	69	42096	8.830	ug/l	98
57) 1,3-Dichloropropane	10.719	76	64537	9.633	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	94970m	45.535	ug/l	
59) 2-Hexanone	10.761	43	99175	50.138	ug/l	96
60) Dibromochloromethane	10.908	129	57151	10.123	ug/l	97
61) 1,2-Dibromoethane	11.011	107	38861	9.979	ug/l	99
64) Tetrachloroethene	10.646	164	76110	10.335	ug/l	98
65) Chlorobenzene	11.438	112	161155	10.074	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.517	131	57233	10.173	ug/l	98
67) Ethyl Benzene	11.517	91	251958	9.651	ug/l	98
68) m/p-Xylenes	11.627	106	198866	19.004	ug/l	99
69) o-Xylene	11.950	106	91710	9.467	ug/l	100
70) Styrene	11.969	104	151094	9.446	ug/l	99
71) Bromoform	12.133	173	35681	10.595	ug/l #	94
73) Isopropylbenzene	12.249	105	236489	9.467	ug/l	99
74) N-amyl acetate	12.066	43	40098	8.590	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	41130	9.871	ug/l	97
76) 1,2,3-Trichloropropane	12.554	75	29944m	8.095	ug/l	
77) Bromobenzene	12.529	156	66002	10.295	ug/l	99
78) n-propylbenzene	12.590	91	275671	9.384	ug/l	100
79) 2-Chlorotoluene	12.676	91	165582	9.762	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	191799	9.445	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	13110	9.556	ug/l	99
82) 4-Chlorotoluene	12.773	91	169968	9.595	ug/l	100
83) tert-Butylbenzene	12.993	119	177250	9.579	ug/l	99
84) 1,2,4-Trimethylbenzene	13.041	105	191423	9.528	ug/l	98
85) sec-Butylbenzene	13.170	105	254134	9.483	ug/l	100
86) p-Isopropyltoluene	13.285	119	210320	9.287	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	127377	10.058	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	127757	10.172	ug/l	98
89) n-Butylbenzene	13.615	91	182116	9.107	ug/l	98
90) Hexachloroethane	13.877	117	47355	9.630	ug/l	97
91) 1,2-Dichlorobenzene	13.657	146	112325	10.165	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	7005	10.608	ug/l	94
93) 1,2,4-Trichlorobenzene	14.919	180	65050	9.992	ug/l	98
94) Hexachlorobutadiene	15.023	225	42813	10.554	ug/l	97
95) Naphthalene	15.139	128	96092	9.264	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	56499	10.004	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023385.D
Acq On : 07 Oct 2025 10:02
Operator : SY/MD
Sample : VSTDICC010
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:40:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 04:39:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Manual Integrations APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 9
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010



Tgt Ion:168 Resp: 559828
Ion Ratio Lower Upper
168 100
99 48.2 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 10.129 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 85 Resp: 44666
Ion Ratio Lower Upper
85 100
87 33.8 16.0 47.9









#3
Chloromethane
Concen: 9.957 ug/l
RT: 2.068 min Scan# 57
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Tgt Ion: 50 Resp: 6241
Ion Ratio Lower Upper
50 100
52 31.2 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#4
Vinyl Chloride
Concen: 9.676 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 62 Resp: 76587
Ion Ratio Lower Upper
62 100
64 32.7 25.8 38.8









#5
Bromomethane
Concen: 10.271 ug/l
RT: 2.592 min Scan# 143
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion: 94 Resp: 66689
Ion Ratio Lower Upper
94 100
96 83.4 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#6
Chloroethane
Concen: 9.632 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02


Tgt Ion: 64 Resp: 51757
Ion Ratio Lower Upper
64 100
66 37.2 26.2 39.4





#7
Trichlorofluoromethane
Concen: 10.128 ug/l
RT: 3.055 min Scan# 219
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion:101 Resp: 102972
Ion Ratio Lower Upper
101 100
103 65.1 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#8
Diethyl Ether
Concen: 10.047 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 74 Resp: 26719
Ion Ratio Lower Upper
74 100
45 83.5 39.9 119.7





#9

1,1,2-Trichlorotrifluoroethane

Concen: 10.011 ug/l

RT: 3.818 min Scan# 34

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion:	101	Resp:	5440
Ion	Ratio	Lower	Upper
101	100		
85	42.2	34.6	51.8
151	82.7	67.0	100.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#10

Methyl Iodide

Concen: 8.967 ug/l




RT: 4.007 min Scan# 375


Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	142	Resp:	56313
Ion	Ratio	Lower	Upper
142	100		
127	42.9	34.8	52.2
141	14.6	11.6	17.4





#11

Tert butyl alcohol

Concen: 71.998 ug/l

RT: 4.854 min Scan# 51

Delta R.T. -0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 59 Resp: 20540

Ion Ratio Lower Upper

59 100





57 6.3 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 10.173 ug/l

RT: 3.793 min Scan# 340

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion: 96 Resp: 53191


Ion Ratio Lower Upper

96 100

61 135.7 106.7 160.1

98 63.4 50.2 75.4





#13

Acrolein

Concen: 49.681 ug/l

RT: 3.653 min Scan# 316

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 56 Resp: 24014

Ion Ratio Lower Upper

56 100





55 75.2 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#14

Allyl chloride

Concen: 9.494 ug/l

RT: 4.385 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion: 41 Resp: 62154


Ion Ratio Lower Upper

41 100

39 74.2 59.9 89.9

76 45.9 38.0 57.0





#15

Acrylonitrile

Concen: 49.854 ug/l

RT: 5.061 min Scan# 54

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion: 53	Resp: 5193
Ion Ratio	Lower Upper
53	100
52	79.5 64.6 97.0
51	37.4 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 59.997 ug/l




RT: 3.872 min Scan# 353


Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion: 43	Resp: 65410
Ion Ratio	Lower Upper
43	100
58	37.9 27.7 41.5





#17

Carbon Disulfide

Concen: 9.382 ug/l

RT: 4.104 min Scan# 391

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 76 Resp: 15305

Ion Ratio Lower Upper

76 100





78 9.2 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#18

Methyl Acetate

Concen: 10.728 ug/l

RT: 4.391 min Scan# 438

Delta R.T. 0.006 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion: 43 Resp: 29725

Ion Ratio Lower Upper

43 100

74 30.3 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 9.830 ug/l

RT: 5.116 min Scan# 51

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 73 Resp: 12034

Ion Ratio Lower Upper

73 100





57 21.5 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 12.016 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion: 84 Resp: 71034




Ion Ratio Lower Upper


84 100

49 102.9 80.2 120.4

51 32.6 25.9 38.9


86 68.7 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 9.710 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion: 96 Resp: 5646
Ion Ratio Lower Upper
96 100
61 121.0 92.7 139.1
98 62.0 49.7 74.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#22
Diisopropyl ether
Concen: 9.547 ug/l
RT: 6.024 min Scan# 706
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02



Tgt Ion: 45 Resp: 138589
Ion Ratio Lower Upper
45 100
43 58.2 45.6 68.4
87 33.7 28.1 42.1
59 12.2 10.2 15.2





#23

Vinyl Acetate

Concen: 45.300 ug/l

RT: 5.963 min Scan# 695

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 361478

Ion Ratio Lower Upper

43

100

86

13.5

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 9.772 ug/l

RT: 5.915 min Scan# 688

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion: 63 Resp: 93512

Ion Ratio Lower Upper

63

100

98

8.4

4.3




13.1


100

4.3

2.8

8.4





#25

2-Butanone

Concen: 55.784 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 7517

Ion Ratio Lower Upper

43 100





72 29.4 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 10.153 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion: 77 Resp: 86325

Ion Ratio Lower Upper

77 100

97 23.9 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 9.905 ug/l

RT: 6.896 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion: 96	Resp: 6533
Ion Ratio	Lower Upper
96 100	
61 129.1	0.0 258.2
98 64.5	0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane

Concen: 9.733 ug/l




RT: 7.244 min Scan# 906


Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion: 49	Resp: 36132
Ion Ratio	Lower Upper
49 100	
129 3.1	0.0 5.6
130 103.0	81.9 122.9





#29

Tetrahydrofuran

Concen: 48.749 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC010

Tgt Ion:	42	Resp:	38990
Ion	Ratio	Lower	Upper
42	100		
72	52.6	43.0	64.6
71	49.8	40.1	60.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 9.890 ug/l




RT: 7.421 min Scan# 935


Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	83	Resp:	102058
Ion	Ratio	Lower	Upper
83	100		
85	64.9	52.5	78.7





#31

Cyclohexane

Concen: 9.495 ug/l

RT: 7.701 min Scan# 981

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y





ClientSampleId :

VSTDICC010

Tgt Ion:	56	Resp:	8024
Ion	Ratio	Lower	Upper
56	100		
69	50.5	30.0	45.0
84	109.7	81.2	121.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#32

1,1,1-Trichloroethane

Concen: 9.844 ug/l




RT: 7.622 min Scan# 968


Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	97	Resp:	91186
Ion	Ratio	Lower	Upper
97	100		
99	64.2	51.4	77.2
61	41.0	31.9	47.9





#33

1,2-Dichloroethane-d4

Concen: 10.071 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 65 Resp: 48820

Ion Ratio Lower Upper

65 100





67 55.4 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion:114 Resp: 847077


Ion Ratio Lower Upper

114 100

63 16.9 0.0 34.2

88 14.7 0.0 28.4





#35

Dibromofluoromethane

Concen: 10.364 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion:113 Resp: 53279

Ion Ratio Lower Upper

113 100

111 102.1 81.8 122.6





192 23.3 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 9.639 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion: 75 Resp: 69913


Ion Ratio Lower Upper

75 100

110 39.2 19.9 59.7

77 30.4 25.0 37.6





#37

Ethyl Acetate

Concen: 9.551 ug/l

RT: 6.988 min Scan# 80

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 2738

Ion Ratio Lower Upper

43 100

61 17.3 11.8 17.8





70 12.8 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 10.065 ug/l

RT: 7.817 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion:117 Resp: 84799


Ion Ratio Lower Upper

117 100

119 96.5 76.8 115.2

121 29.9 24.5 36.7





#39

Methylcyclohexane

Concen: 8.860 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion:	83	Resp:	82275
Ion	Ratio	Lower	Upper
83	100		
55	69.7	51.4	77.2
98	48.6	40.4	60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 9.758 ug/l




RT: 8.079 min Scan# 1043


Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	78	Resp:	220629
Ion	Ratio	Lower	Upper
78	100		
77	22.8	18.3	27.5





#41

Methacrylonitrile

Concen: 9.632 ug/l

RT: 7.225 min Scan# 902

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 41 Resp: 14344

Ion Ratio Lower Upper

41 100

39 49.9 52.2 78.4

67 90.2 74.9 112.3





52 36.9 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#42

1,2-Dichloroethane

Concen: 10.174 ug/l

RT: 8.158 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion: 62 Resp: 58541

Ion Ratio Lower Upper

62 100

98 11.0 0.0 21.6





#43

Isopropyl Acetate

Concen: 9.510 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion: 43	Resp: 5304
Ion Ratio	Lower Upper
43	100
61	21.4 18.1 27.1
87	17.0 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 10.171 ug/l




RT: 8.865 min Scan# 1172


Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:130	Resp: 64537
Ion Ratio	Lower Upper
130	100
95	94.2 0.0 180.6





#45

1,2-Dichloropropane

Concen: 9.632 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 63 Resp: 49009

Ion Ratio Lower Upper

63 100





65 31.0 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 9.890 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion: 93 Resp: 30780


Ion Ratio Lower Upper

93 100

95 83.7 65.4 98.0

174 116.7 92.9 139.3





#47

Bromodichloromethane

Concen: 10.007 ug/l

RT: 9.426 min Scan# 1263

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion:	83	Resp:	7839
Ion	Ratio	Lower	Upper
83	100		
85	64.5	50.8	76.2
127	8.6	7.4	11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#48

Methyl methacrylate

Concen: 9.023 ug/l




RT: 9.225 min Scan# 1231


Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	41	Resp:	23484
Ion	Ratio	Lower	Upper
41	100		
69	105.8	81.0	121.4
39	61.4	48.2	72.4





#49

1,4-Dioxane

Concen: 211.693 ug/l

RT: 9.231 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion:	88	Resp:	682
Ion	Ratio	Lower	Upper
88	100		
43	32.1	21.8	32.8
58	62.0	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 10.225 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	98	Resp:	198242
Ion	Ratio	Lower	Upper
98	100		
100	65.6	53.0	79.4





#51

4-Methyl-2-Pentanone

Concen: 47.165 ug/l

RT: 9.999 min Scan# 1357

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 134740

Ion Ratio Lower Upper

43 100





58 43.5 35.3 52.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#52

Toluene

Concen: 9.661 ug/l

RT: 10.170 min Scan# 1386

Delta R.T. -0.000 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion: 92 Resp: 138662

Ion Ratio Lower Upper

92 100

91 170.6 136.4 204.6





#53

t-1,3-Dichloropropene

Concen: 9.559 ug/l

RT: 10.396 min Scan# 1422

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 75 Resp: 6431

Ion Ratio Lower Upper

75 100





77 33.4 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#54

cis-1,3-Dichloropropene

Concen: 9.687 ug/l

RT: 9.853 min Scan# 1334

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion: 75 Resp: 77905


Ion Ratio Lower Upper

75 100

77 30.0 25.4 38.0

39 40.2 32.8 49.2





#55

1,1,2-Trichloroethane

Concen: 10.118 ug/l

RT: 10.572 min Scan# 1452

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 97 Resp: 4188

Ion Ratio Lower Upper

97 100

83 90.6 61.8 92.6

85 53.8 41.3 61.9





99 64.0 48.7 73.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#56

Ethyl methacrylate

Concen: 8.830 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02




Tgt Ion: 69 Resp: 42096


Ion Ratio Lower Upper

69 100

41 58.1 44.6 67.0

39 34.5 27.0 40.4





#57

1,3-Dichloropropane

Concen: 9.633 ug/l

RT: 10.719 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 76 Resp: 6453

Ion Ratio Lower Upper

76 100





78 33.6 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 45.535 ug/l m

RT: 9.713 min Scan# 1311

Delta R.T. 0.006 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion: 63 Resp: 94970

Ion Ratio Lower Upper

63 100


106 33.4 26.6 39.8





#59
2-Hexanone
Concen: 50.138 ug/l
RT: 10.761 min Scan# 1482
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 43 Resp: 9917
Ion Ratio Lower Upper
43 100
58 58.0 30.5 91.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#60
Dibromochloromethane
Concen: 10.123 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Tgt Ion:129 Resp: 57151
Ion Ratio Lower Upper
129 100
127 74.9 38.6 115.7





#61
1,2-Dibromoethane
Concen: 9.979 ug/l
RT: 11.011 min Scan# 1524
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion: 107 Resp: 3886
Ion Ratio Lower Upper
107 100
109 93.8 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 10.235 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02


Tgt Ion: 95 Resp: 62601
Ion Ratio Lower Upper
95 100
174 94.6 0.0 192.8
176 92.1 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion:117 Resp: 736152
Ion Ratio Lower Upper
117 100
82 51.8 41.0 61.4
119 32.3 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#64
Tetrachloroethene
Concen: 10.335 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Tgt Ion:164 Resp: 76110
Ion Ratio Lower Upper
164 100
166 126.1 99.4 149.0
129 89.0 69.4 104.2
131 87.6 67.5 101.3





#65
Chlorobenzene
Concen: 10.074 ug/l
RT: 11.438 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion:112 Resp: 16115
Ion Ratio Lower Upper
112 100
114 29.5 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 10.173 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion:131 Resp: 57233
Ion Ratio Lower Upper
131 100
133 95.8 48.4 145.2
119 66.9 32.4 97.2





#67

Ethyl Benzene

Concen: 9.651 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 91 Resp: 251958

Ion Ratio Lower Upper

91 100





106 33.1 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#68

m/p-Xylenes

Concen: 19.004 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion:106 Resp: 198866

Ion Ratio Lower Upper

106 100

91 192.3 152.6 229.0









#69
o-Xylene
Concen: 9.467 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion:106 Resp: 91710
Ion Ratio Lower Upper
106 100
91 202.0 101.4 304.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#70
Styrene
Concen: 9.446 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02


Tgt Ion:104 Resp: 151094
Ion Ratio Lower Upper
104 100
78 45.4 36.1 54.1
103 55.0 43.2 64.8





#71
Bromoform
Concen: 10.595 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.006 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion:173 Resp: 3568
Ion Ratio Lower Upper
173 100
175 45.5 24.8 74.3
254 0.1 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Tgt Ion:152 Resp: 374998
Ion Ratio Lower Upper
152 100
115 54.2 27.1 81.2
150 160.7 0.0 347.4





#73
Isopropylbenzene
Concen: 9.467 ug/l
RT: 12.249 min Scan# 1727
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010



Tgt Ion:105 Resp: 236489
Ion Ratio Lower Upper
105 100
120 27.3 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 8.590 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 43 Resp: 40098
Ion Ratio Lower Upper
43 100
70 52.7 41.1 61.7
55 31.6 23.0 34.6
61 28.2 21.4 32.2









#75
1,1,2,2-Tetrachloroethane
Concen: 9.871 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Tgt Ion: 83 Resp: 41130
Ion Ratio Lower Upper
83 100
131 12.0 5.9 17.6
85 66.1 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#76
1,2,3-Trichloropropane
Concen: 8.095 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 75 Resp: 29944
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 10.295 ug/l
RT: 12.529 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Tgt Ion:156 Resp: 6600
Ion Ratio Lower Upper
156 100
77 151.8 76.8 230.5
158 97.0 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#78
n-propylbenzene
Concen: 9.384 ug/l
RT: 12.590 min Scan# 1783
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 91 Resp: 275671
Ion Ratio Lower Upper
91 100
120 24.4 12.3 36.8





#79

2-Chlorotoluene

Concen: 9.762 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 91 Resp: 16558

Ion Ratio Lower Upper

91 100





126 36.9 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 9.445 ug/l

RT: 12.731 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion:105 Resp: 191799

Ion Ratio Lower Upper

105 100

120 50.9 25.4 76.4





#81

trans-1,4-Dichloro-2-butene

Concen: 9.556 ug/l

RT: 12.298 min Scan# 1735

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010

Tgt Ion: 75 Resp: 13110

Ion Ratio Lower Upper

75 100

53 84.6 68.6 102.8





89 48.5 39.0 58.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#82

4-Chlorotoluene

Concen: 9.595 ug/l

RT: 12.773 min Scan# 1813

Delta R.T. -0.000 min

Lab File: VY023385.D




Acq: 07 Oct 2025 10:02


Tgt Ion: 91 Resp: 169968

Ion Ratio Lower Upper

91 100

126 36.8 18.4 55.4





#83

tert-Butylbenzene

Concen: 9.579 ug/l

RT: 12.993 min Scan# 1849

Delta R.T. -0.000 min

Lab File: VY023385.D


Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010



Tgt Ion:119 Resp: 177250

Ion Ratio Lower Upper

119 100

91 60.5 29.9 89.8




134 25.6 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#84

1,2,4-Trimethylbenzene


Concen: 9.528 ug/l

RT: 13.041 min Scan# 1857

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02






Tgt Ion:105 Resp: 191423

Ion Ratio Lower Upper

105 100


120 46.4 23.8 71.2





#85
sec-Butylbenzene
Concen: 9.483 ug/l
RT: 13.170 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010



Tgt Ion:105 Resp: 254134
Ion Ratio Lower Upper
105 100
134 21.4 10.7 32.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#86
p-Isopropyltoluene
Concen: 9.287 ug/l
RT: 13.285 min Scan# 1897
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion:119 Resp: 210320
Ion Ratio Lower Upper
119 100
134 27.0 13.6 40.7
91 22.3 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 10.058 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC010

Tgt Ion:	146	Resp:	12737
Ion Ratio	Lower	Upper	
146	100		
111	37.1	18.7	56.1
148	63.2	31.9	95.9

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 10.172 ug/l




RT: 13.365 min Scan# 1910


Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:	146	Resp:	127757
Ion Ratio	Lower	Upper	
146	100		
111	38.6	18.0	54.0
148	64.2	31.9	95.7





#89

n-Butylbenzene

Concen: 9.107 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion:	91	Resp:	182110
Ion	Ratio	Lower	Upper
91	100		
92	52.2	27.0	80.8
134	28.0	14.1	42.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#90

Hexachloroethane

Concen: 9.630 ug/l




RT: 13.877 min Scan# 1994


Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02


Tgt Ion:	117	Resp:	47355
Ion	Ratio	Lower	Upper
117	100		
201	85.0	43.8	131.4





#91
1,2-Dichlorobenzene
Concen: 10.165 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010



Tgt Ion:146 Resp: 11232
Ion Ratio Lower Upper
146 100
111 39.4 19.0 57.0
148 63.4 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 10.608 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023385.D
Acq: 07 Oct 2025 10:02

Tgt Ion: 75 Resp: 7005
Ion Ratio Lower Upper
75 100
155 94.2 50.0 150.0
157 121.7 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 9.992 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023385.D


Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010






Tgt Ion:	180	Resp:	65050
Ion Ratio	Lower	Upper	
180	100		
182	93.5	47.9	143.6
145	27.9	14.7	44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene


Concen: 10.554 ug/l

RT: 15.023 min Scan# 2182



Delta R.T. 0.006 min


Lab File: VY023385.D

Acq: 07 Oct 2025 10:02



Tgt Ion:	225	Resp:	42813
Ion Ratio	Lower	Upper	
225	100		
223	63.4	30.8	92.3
227	66.0	31.9	95.7





#95

Naphthalene

Concen: 9.264 ug/l

RT: 15.139 min Scan# 2201

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion:128	Resp:	9609
Ion Ratio	Lower	Upper
128	100	
127	13.0	10.4 15.6
129	10.4	8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#96

1,2,3-Trichlorobenzene

Concen: 10.004 ug/l




RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023385.D

Acq: 07 Oct 2025 10:02

Tgt Ion:180	Resp:	56499
Ion Ratio	Lower	Upper
180	100	
182	95.1	47.9 143.8
145	30.8	15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023386.D
 Acq On : 07 Oct 2025 11:24
 Operator : SY/MD
 Sample : VSTDICC020
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC020

Quant Time: Oct 08 04:41:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	574363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	851961	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	746872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	382604	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.061	65	98481	19.798	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	39.600%#
35) Dibromofluoromethane	7.634	113	108034	20.895	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	41.800%#
50) Toluene-d8	10.103	98	401067	20.568	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	41.140%#
62) 4-Bromofluorobenzene	12.401	95	127216	20.680	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	41.360%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	91924	20.318	ug/l	98
3) Chloromethane	2.074	50	117585	18.284	ug/l	100
4) Vinyl Chloride	2.202	62	151460	18.652	ug/l	99
5) Bromomethane	2.586	94	125948	18.907	ug/l	97
6) Chloroethane	2.732	64	103973	18.859	ug/l	94
7) Trichlorofluoromethane	3.049	101	217452	20.846	ug/l	99
8) Diethyl Ether	3.452	74	54124	19.837	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.811	101	109936	19.718	ug/l	99
10) Methyl Iodide	4.007	142	119083	18.482	ug/l	99
11) Tert butyl alcohol	4.854	59	34346	117.345	ug/l	98
12) 1,1-Dichloroethene	3.787	96	104778	19.531	ug/l	98
13) Acrolein	3.653	56	48115	97.023	ug/l	98
14) Allyl chloride	4.384	41	125424	18.674	ug/l	98
15) Acrylonitrile	5.055	53	99018	92.641	ug/l	100
16) Acetone	3.866	43	108209	96.742	ug/l	96
17) Carbon Disulfide	4.104	76	303284	18.121	ug/l	99
18) Methyl Acetate	4.384	43	59704	21.002	ug/l	97
19) Methyl tert-butyl Ether	5.116	73	247030	19.667	ug/l	97
20) Methylene Chloride	4.616	84	132229	21.801	ug/l	98
21) trans-1,2-Dichloroethene	5.110	96	116410	19.512	ug/l	95
22) Diisopropyl ether	6.018	45	288429	19.366	ug/l	98
23) Vinyl Acetate	5.957	43	748048	91.373	ug/l	99
24) 1,1-Dichloroethane	5.915	63	190012	19.354	ug/l	99
25) 2-Butanone	6.896	43	132159	95.593	ug/l	93
26) 2,2-Dichloropropane	6.884	77	167903	19.247	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	133634	19.747	ug/l	99
28) Bromochloromethane	7.244	49	73407	19.273	ug/l	98
29) Tetrahydrofuran	7.262	42	76502	93.229	ug/l	99
30) Chloroform	7.421	83	207096	19.562	ug/l	100
31) Cyclohexane	7.701	56	160159	18.472	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	188861	19.873	ug/l	100
36) 1,1-Dichloropropene	7.835	75	140912	19.317	ug/l	100
37) Ethyl Acetate	6.982	43	55705	19.315	ug/l	98
38) Carbon Tetrachloride	7.817	117	171925	20.289	ug/l	99
39) Methylcyclohexane	9.109	83	182831	19.576	ug/l	99
40) Benzene	8.079	78	450471	19.809	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023386.D
 Acq On : 07 Oct 2025 11:24
 Operator : SY/MD
 Sample : VSTDICC020
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:41:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 04:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	31129m	20.783	ug/l	
42) 1,2-Dichloroethane	8.158	62	116941	20.206	ug/l	99
43) Isopropyl Acetate	8.195	43	106595	19.002	ug/l #	87
44) Trichloroethene	8.865	130	131459	20.598	ug/l	98
45) 1,2-Dichloropropane	9.140	63	99363	19.417	ug/l	98
46) Dibromomethane	9.231	93	62277	19.896	ug/l	98
47) Bromodichloromethane	9.420	83	156426	19.854	ug/l	98
48) Methyl methacrylate	9.219	41	50661	19.352	ug/l	99
49) 1,4-Dioxane	9.225	88	13636	420.712	ug/l	99
51) 4-Methyl-2-Pentanone	9.999	43	274059	95.378	ug/l	100
52) Toluene	10.170	92	291453	20.191	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	132848	19.632	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	158580	19.606	ug/l	95
55) 1,1,2-Trichloroethane	10.572	97	82183	19.740	ug/l	96
56) Ethyl methacrylate	10.438	69	89244	18.613	ug/l	99
57) 1,3-Dichloropropane	10.719	76	130650	19.389	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	190850m	90.982	ug/l	
59) 2-Hexanone	10.761	43	187387	94.190	ug/l	99
60) Dibromochloromethane	10.908	129	115907	20.412	ug/l	100
61) 1,2-Dibromoethane	11.011	107	78579	20.062	ug/l	100
64) Tetrachloroethene	10.646	164	167648	22.438	ug/l	97
65) Chlorobenzene	11.438	112	326360	20.109	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.517	131	116412	20.396	ug/l	99
67) Ethyl Benzene	11.517	91	523606	19.768	ug/l	99
68) m/p-Xylenes	11.627	106	426913	40.211	ug/l	100
69) o-Xylene	11.956	106	195021	19.843	ug/l	96
70) Styrene	11.968	104	327934	20.208	ug/l	99
71) Bromoform	12.133	173	71119	20.814	ug/l #	99
73) Isopropylbenzene	12.255	105	513183	20.136	ug/l	99
74) N-amyl acetate	12.072	43	85063	17.860	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	76821	18.069	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	74100m	19.634	ug/l	
77) Bromobenzene	12.529	156	135402	20.700	ug/l	98
78) n-propylbenzene	12.596	91	597829	19.945	ug/l	100
79) 2-Chlorotoluene	12.676	91	349203	20.179	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	418684	20.208	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	26922	19.234	ug/l	98
82) 4-Chlorotoluene	12.773	91	361458	20.000	ug/l	99
83) tert-Butylbenzene	12.993	119	377277	19.983	ug/l	99
84) 1,2,4-Trimethylbenzene	13.041	105	413292	20.162	ug/l	100
85) sec-Butylbenzene	13.176	105	546582	19.990	ug/l	100
86) p-Isopropyltoluene	13.291	119	464005	20.081	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	262901	20.347	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	259217	20.230	ug/l	99
89) n-Butylbenzene	13.615	91	401893	19.697	ug/l	98
90) Hexachloroethane	13.877	117	98983	19.730	ug/l	97
91) 1,2-Dichlorobenzene	13.657	146	231199	20.507	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	12668	18.802	ug/l	99
93) 1,2,4-Trichlorobenzene	14.919	180	137963	20.770	ug/l	99
94) Hexachlorobutadiene	15.023	225	87575	21.159	ug/l	99
95) Naphthalene	15.145	128	206017	19.467	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	121656	21.113	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023386.D
Acq On : 07 Oct 2025 11:24
Operator : SY/MD
Sample : VSTDICC020
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:41:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 04:39:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023386.D
Acq On : 07 Oct 2025 11:24
Operator : SY/MD
Sample : VSTDIC020
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Quant Time: Oct 08 04:41:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 04:39:16 2025


Response via : Initial Calibration


Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025


Supervised By :Mahesh Dadoda 10/10/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 57436
Ion Ratio Lower Upper
168 100
99 49.2 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 20.318 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion: 85 Resp: 91924
Ion Ratio Lower Upper
85 100
87 31.0 16.0 47.9









#3
Chloromethane
Concen: 18.284 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 11758
Ion Ratio Lower Upper
50 100
52 32.7 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#4
Vinyl Chloride
Concen: 18.652 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24


Tgt Ion: 62 Resp: 151460
Ion Ratio Lower Upper
62 100
64 31.7 25.8 38.8





#5
Bromomethane
Concen: 18.907 ug/l
RT: 2.586 min Scan# 143
Delta R.T. -0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020




Tgt Ion: 94 Resp: 125948
Ion Ratio Lower Upper
94 100
96 95.9 74.5 111.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#6
Chloroethane
Concen: 18.859 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24



Tgt Ion: 64 Resp: 103973
Ion Ratio Lower Upper
64 100
66 29.3 26.2 39.4





#7

Trichlorofluoromethane

Concen: 20.846 ug/l

RT: 3.049 min Scan# 219

Delta R.T. -0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion:101 Resp: 21745

Ion Ratio Lower Upper

101 100





103 65.2 51.3 76.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#8

Diethyl Ether

Concen: 19.837 ug/l

RT: 3.452 min Scan# 284

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 74 Resp: 54124

Ion Ratio Lower Upper

74 100

45 80.6 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.718 ug/l
RT: 3.811 min Scan# 343
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 109930
Ion Ratio Lower Upper
101 100
85 43.1 34.6 51.8
151 82.2 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#10
Methyl Iodide
Concen: 18.482 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion:142 Resp: 119083
Ion Ratio Lower Upper
142 100
127 42.9 34.8 52.2
141 14.5 11.6 17.4





#11

Tert butyl alcohol

Concen: 117.345 ug/l

RT: 4.854 min Scan# 515

Delta R.T. -0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 59 Resp: 3434

Ion Ratio Lower Upper

59 100





57 9.8 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 19.531 ug/l

RT: 3.787 min Scan# 339

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 96 Resp: 104778


Ion Ratio Lower Upper

96 100

61 136.2 106.7 160.1

98 64.9 50.2 75.4





#13

Acrolein

Concen: 97.023 ug/l

RT: 3.653 min Scan# 31

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 4811

Ion Ratio Lower Upper

56 100





55 72.0 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#14

Allyl chloride

Concen: 18.674 ug/l

RT: 4.384 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 41 Resp: 125424


Ion Ratio Lower Upper

41 100

39 77.6 59.9 89.9

76 47.9 38.0 57.0





#15

Acrylonitrile

Concen: 92.641 ug/l

RT: 5.055 min Scan# 547

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 53 Resp: 99018

Ion Ratio Lower Upper

53 100

52 80.7 64.6 97.0





51 36.6 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 96.742 ug/l

RT: 3.866 min Scan# 352

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 43 Resp: 108209

Ion Ratio Lower Upper

43 100


58 32.1 27.7 41.5





#17
Carbon Disulfide
Concen: 18.121 ug/l
RT: 4.104 min Scan# 391
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion: 76 Resp: 303284
Ion Ratio Lower Upper
76 100
78 8.8 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#18
Methyl Acetate
Concen: 21.002 ug/l
RT: 4.384 min Scan# 437
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion: 43 Resp: 59704
Ion Ratio Lower Upper
43 100
74 28.7 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 19.667 ug/l

RT: 5.116 min Scan# 51

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 73 Resp: 247030

Ion Ratio Lower Upper

73 100





57 20.1 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 21.801 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion: 84 Resp: 132229




Ion Ratio Lower Upper


84 100

49 104.2 80.2 120.4

51 31.7 25.9 38.9

86 63.9 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 19.512 ug/l

RT: 5.110 min Scan# 511

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 96 Resp: 116410

Ion Ratio Lower Upper

96 100

61 120.5 92.7 139.1





98 66.7 49.7 74.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#22

Diisopropyl ether

Concen: 19.366 ug/l

RT: 6.018 min Scan# 705

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion: 45 Resp: 288429




Ion Ratio Lower Upper


45 100

43 58.8 45.6 68.4

87 34.5 28.1 42.1

59 13.0 10.2 15.2





#23

Vinyl Acetate

Concen: 91.373 ug/l

RT: 5.957 min Scan# 695

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 74804

Ion Ratio Lower Upper

43 100





86 13.5 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 19.354 ug/l

RT: 5.915 min Scan# 688

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 63 Resp: 190012


Ion Ratio Lower Upper

63 100

98 7.9 4.3 13.1

100 5.5 2.8 8.4





#25

2-Butanone

Concen: 95.593 ug/l

RT: 6.896 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 132159

Ion Ratio Lower Upper

43 100





72 30.5 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 19.247 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 77 Resp: 167903

Ion Ratio Lower Upper

77 100

97 23.9 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 19.747 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC020

Tgt Ion: 96	Resp: 133634
Ion Ratio	Lower Upper
96 100	
61 128.1	0.0 258.2
98 65.3	0.0 132.0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane

Concen: 19.273 ug/l




RT: 7.244 min Scan# 906


Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion: 49	Resp: 73407
Ion Ratio	Lower Upper
49 100	
129 2.7	0.0 5.6
130 100.8	81.9 122.9





#29

Tetrahydrofuran

Concen: 93.229 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC020

Tgt Ion:	42	Resp:	76502
Ion	Ratio	Lower	Upper
42	100		
72	52.7	43.0	64.6
71	50.6	40.1	60.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 19.562 ug/l




RT: 7.421 min Scan# 935


Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24


Tgt Ion:	83	Resp:	207096
Ion	Ratio	Lower	Upper
83	100		
85	65.5	52.5	78.7





#31
Cyclohexane
Concen: 18.472 ug/l
RT: 7.701 min Scan# 981
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion: 56 Resp: 160159
Ion Ratio Lower Upper
56 100
69 41.2 30.0 45.0
84 103.3 81.2 121.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#32
1,1,1-Trichloroethane
Concen: 19.873 ug/l
RT: 7.616 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion: 97 Resp: 188861
Ion Ratio Lower Upper
97 100
99 64.4 51.4 77.2
61 40.1 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 19.798 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 65 Resp: 9848

Ion Ratio Lower Upper

65 100





67 53.3 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion:114 Resp: 851961


Ion Ratio Lower Upper

114 100

63 17.4 0.0 34.2

88 14.3 0.0 28.4





#35

Dibromofluoromethane

Concen: 20.895 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion:113 Resp: 108034

Ion Ratio Lower Upper

113 100

111 100.9 81.8 122.6





192 22.3 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 19.317 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 75 Resp: 140912


Ion Ratio Lower Upper

75 100

110 39.7 19.9 59.7

77 31.0 25.0 37.6





#37

Ethyl Acetate

Concen: 19.315 ug/l

RT: 6.982 min Scan# 863

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y





ClientSampleId :

VSTDIC020

Tgt Ion:	43	Resp:	5570
Ion	Ratio	Lower	Upper
43	100		
61	13.9	11.8	17.8
70	12.7	10.5	15.7

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 20.289 ug/l




RT: 7.817 min Scan# 1000


Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion:	117	Resp:	171925
Ion	Ratio	Lower	Upper
117	100		
119	95.8	76.8	115.2
121	29.4	24.5	36.7





#39

Methylcyclohexane

Concen: 19.576 ug/l

RT: 9.109 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 83 Resp: 18283

Ion Ratio Lower Upper

83 100

55 64.8 51.4 77.2





98 49.9 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 19.809 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 78 Resp: 450471

Ion Ratio Lower Upper

78 100

77 23.2 18.3 27.5









#41
Methacrylonitrile
Concen: 20.783 ug/l m
RT: 7.225 min Scan# 902
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Tgt Ion: 41 Resp: 31129
Ion Ratio Lower Upper
41 100
39 68.6 52.2 78.4
67 0.0 74.9 112.3
52 0.0 29.2 43.8#




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#42
1,2-Dichloroethane
Concen: 20.206 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion: 62 Resp: 116941
Ion Ratio Lower Upper
62 100
98 11.2 0.0 21.6





#43

Isopropyl Acetate

Concen: 19.002 ug/l

RT: 8.195 min Scan# 1062

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020





Tgt Ion: 43 Resp: 10659

Ion	Ratio	Lower	Upper
43	100		
61	33.1	18.1	27.1
87	16.9	13.5	20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 20.598 ug/l

RT: 8.865 min Scan# 1172




Delta R.T. -0.000 min


Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion:130 Resp: 131459

Ion	Ratio	Lower	Upper
130	100		
95	92.0	0.0	180.6





#45

1,2-Dichloropropane

Concen: 19.417 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 9936

Ion Ratio Lower Upper

63 100





65 32.2 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 19.896 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 93 Resp: 62277


Ion Ratio Lower Upper

93 100

95 81.3 65.4 98.0

174 113.6 92.9 139.3





#47

Bromodichloromethane

Concen: 19.854 ug/l

RT: 9.420 min Scan# 1263

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC020

Tgt Ion:	83	Resp:	156420
Ion	Ratio	Lower	Upper
83	100		
85	64.9	50.8	76.2
127	9.4	7.4	11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#48

Methyl methacrylate

Concen: 19.352 ug/l




RT: 9.219 min Scan# 1230


Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion:	41	Resp:	50661
Ion	Ratio	Lower	Upper
41	100		
69	102.0	81.0	121.4
39	58.8	48.2	72.4





#49

1,4-Dioxane

Concen: 420.712 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 88 Resp: 1363

Ion Ratio Lower Upper

88 100

43 28.9 21.8 32.8





58 58.1 46.8 70.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 20.568 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 98 Resp: 401067

Ion Ratio Lower Upper

98 100

100 65.3 53.0 79.4





#51

4-Methyl-2-Pentanone

Concen: 95.378 ug/l

RT: 9.999 min Scan# 1357

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 274059

Ion Ratio Lower Upper

43 100





58 43.8 35.3 52.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#52

Toluene

Concen: 20.191 ug/l

RT: 10.170 min Scan# 1386

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 92 Resp: 291453

Ion Ratio Lower Upper

92 100

91 170.0 136.4 204.6





#53

t-1,3-Dichloropropene

Concen: 19.632 ug/l

RT: 10.390 min Scan# 1422

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 75 Resp: 13284

Ion Ratio Lower Upper

75 100





77 31.9 25.1 37.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#54

cis-1,3-Dichloropropene

Concen: 19.606 ug/l

RT: 9.853 min Scan# 1334

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 75 Resp: 158580


Ion Ratio Lower Upper

75 100

77 32.8 25.4 38.0

39 45.1 32.8 49.2





#55

1,1,2-Trichloroethane

Concen: 19.740 ug/l

RT: 10.572 min Scan# 1452

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 97 Resp: 8218

Ion Ratio Lower Upper

97 100

83 80.9 61.8 92.6

85 54.8 41.3 61.9





99 62.5 48.7 73.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#56

Ethyl methacrylate

Concen: 18.613 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion: 69 Resp: 89244


Ion Ratio Lower Upper

69 100

41 56.5 44.6 67.0

39 34.5 27.0 40.4





#57

1,3-Dichloropropane

Concen: 19.389 ug/l

RT: 10.719 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 130650

Ion Ratio Lower Upper

76 100





78 33.5 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 90.982 ug/l m

RT: 9.713 min Scan# 1311

Delta R.T. 0.006 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion: 63 Resp: 190850

Ion Ratio Lower Upper

63 100

106 32.9 26.6 39.8









#59
2-Hexanone
Concen: 94.190 ug/l
RT: 10.761 min Scan# 1482
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 18738
Ion Ratio Lower Upper
43 100
58 60.2 30.5 91.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#60
Dibromochloromethane
Concen: 20.412 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24


Tgt Ion:129 Resp: 115907
Ion Ratio Lower Upper
129 100
127 77.1 38.6 115.7





#61
1,2-Dibromoethane
Concen: 20.062 ug/l
RT: 11.011 min Scan# 1524
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion:107 Resp: 78579
Ion Ratio Lower Upper
107 100
109 94.7 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 20.680 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24


Tgt Ion: 95 Resp: 127216
Ion Ratio Lower Upper
95 100
174 95.8 0.0 192.8
176 92.7 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020




Tgt Ion:117 Resp: 746872
Ion Ratio Lower Upper
117 100
82 50.3 41.0 61.4
119 31.4 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#64
Tetrachloroethene
Concen: 22.438 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24



Tgt Ion:164 Resp: 167648
Ion Ratio Lower Upper
164 100
166 128.9 99.4 149.0
129 90.1 69.4 104.2
131 82.5 67.5 101.3





#65

Chlorobenzene

Concen: 20.109 ug/l

RT: 11.438 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion:112 Resp: 326360

Ion Ratio Lower Upper

112 100





114 31.6 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 20.396 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion:131 Resp: 116412


Ion Ratio Lower Upper

131 100

133 95.9 48.4 145.2

119 64.4 32.4 97.2





#67

Ethyl Benzene

Concen: 19.768 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 523600

Ion Ratio Lower Upper

91 100





106 32.8 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#68

m/p-Xylenes

Concen: 40.211 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion:106 Resp: 426913

Ion Ratio Lower Upper

106 100

91 190.8 152.6 229.0





#69

o-Xylene

Concen: 19.843 ug/l

RT: 11.956 min Scan# 1678

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 19502

Ion Ratio Lower Upper

106 100





91 208.5 101.4 304.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#70

Styrene

Concen: 20.208 ug/l

RT: 11.968 min Scan# 1681

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion:104 Resp: 327934


Ion Ratio Lower Upper

104 100

78 45.4 36.1 54.1


103 53.3 43.2 64.8





#71
Bromoform
Concen: 20.814 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020




Tgt Ion:173 Resp: 71119
Ion Ratio Lower Upper
173 100
175 48.7 24.8 74.3
254 0.0 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Tgt Ion:152 Resp: 382604
Ion Ratio Lower Upper
152 100
115 54.6 27.1 81.2
150 163.2 0.0 347.4





#73
Isopropylbenzene
Concen: 20.136 ug/l
RT: 12.255 min Scan# 1727
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion: 105 Resp: 51318
Ion Ratio Lower Upper
105 100
120 27.2 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 17.860 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion: 43 Resp: 85063
Ion Ratio Lower Upper
43 100
70 51.8 41.1 61.7
55 30.6 23.0 34.6
61 27.9 21.4 32.2





#75

1,1,2-Tetrachloroethane

Concen: 18.069 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC020





Tgt Ion: 83 Resp: 7682

Ion	Ratio	Lower	Upper
83	100		
131	11.8	5.9	17.6
85	64.1	31.8	95.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#76

1,2,3-Trichloropropane

Concen: 19.634 ug/l m

RT: 12.554 min Scan# 1777




Delta R.T. -0.000 min


Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion: 75 Resp: 74100

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0





#77

Bromobenzene

Concen: 20.700 ug/l

RT: 12.529 min Scan# 1773

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC020

Tgt Ion:	156	Resp:	13540
Ion Ratio	Lower	Upper	
156	100		
77	156.2	76.8	230.5
158	96.1	49.0	147.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#78

n-propylbenzene

Concen: 19.945 ug/l




RT: 12.596 min Scan# 1784


Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Tgt Ion:	91	Resp:	597829
Ion Ratio	Lower	Upper	
91	100		
120	24.6	12.3	36.8





#79

2-Chlorotoluene

Concen: 20.179 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 34920

Ion Ratio Lower Upper

91 100





126 36.8 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 20.208 ug/l

RT: 12.737 min Scan# 1807

Delta R.T. 0.006 min

Lab File: VY023386.D




Acq: 07 Oct 2025 11:24


Tgt Ion:105 Resp: 418684

Ion Ratio Lower Upper

105 100


120 51.1 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 19.234 ug/l
RT: 12.304 min Scan# 1735
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion: 75 Resp: 2692
Ion Ratio Lower Upper
75 100
53 84.4 68.6 102.8
89 46.4 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#82
4-Chlorotoluene
Concen: 20.000 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24


Tgt Ion: 91 Resp: 361458
Ion Ratio Lower Upper
91 100
126 36.1 18.4 55.4





#83
tert-Butylbenzene
Concen: 19.983 ug/l
RT: 12.993 min Scan# 1849
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 37727
Ion Ratio Lower Upper
119 100
91 60.4 29.9 89.8
134 26.1 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.162 ug/l
RT: 13.041 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion:105 Resp: 413292
Ion Ratio Lower Upper
105 100
120 47.6 23.8 71.2





#85

sec-Butylbenzene

Concen: 19.990 ug/l

RT: 13.176 min Scan# 1878

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion:105 Resp: 54658

Ion Ratio Lower Upper

105 100





134 21.3 10.7 32.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#86

p-Isopropyltoluene

Concen: 20.081 ug/l

RT: 13.291 min Scan# 1898

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion:119 Resp: 464005


Ion Ratio Lower Upper

119 100

134 27.2 13.6 40.7

91 21.6 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 20.347 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 26290

Ion Ratio Lower Upper

146 100

111 37.2 18.7 56.1





148 63.1 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 20.230 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion:146 Resp: 259217


Ion Ratio Lower Upper

146 100

111 37.3 18.0 54.0

148 63.6 31.9 95.7





#89

n-Butylbenzene

Concen: 19.697 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC020

Tgt Ion:	91	Resp:	40189
Ion	Ratio	Lower	Upper
91	100		
92	52.6	27.0	80.8
134	27.3	14.1	42.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#90

Hexachloroethane

Concen: 19.730 ug/l




RT: 13.877 min Scan# 1994


Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24


Tgt Ion:	117	Resp:	98983
Ion	Ratio	Lower	Upper
117	100		
201	84.8	43.8	131.4





#91
1,2-Dichlorobenzene
Concen: 20.507 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020



Tgt Ion:146 Resp: 231199
Ion Ratio Lower Upper
146 100
111 38.6 19.0 57.0
148 64.1 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.802 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion: 75 Resp: 12668
Ion Ratio Lower Upper
75 100
155 99.1 50.0 150.0
157 127.9 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 20.770 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC020

Tgt Ion:180 Resp: 13796

Ion Ratio Lower Upper

180 100

182 95.3 47.9 143.6





145 28.3 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene

Concen: 21.159 ug/l

RT: 15.023 min Scan# 2182

Delta R.T. 0.006 min

Lab File: VY023386.D

Acq: 07 Oct 2025 11:24




Tgt Ion:225 Resp: 87575


Ion Ratio Lower Upper

225 100

223 63.1 30.8 92.3

227 63.4 31.9 95.7









#95
Naphthalene
Concen: 19.467 ug/l
RT: 15.145 min Scan# 2201
Delta R.T. 0.006 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC020

Tgt Ion:128 Resp: 20601
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.1 8.9 13.3




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#96
1,2,3-Trichlorobenzene
Concen: 21.113 ug/l
RT: 15.328 min Scan# 2232
Delta R.T. -0.000 min
Lab File: VY023386.D
Acq: 07 Oct 2025 11:24

Tgt Ion:180 Resp: 121656
Ion Ratio Lower Upper
180 100
182 93.3 47.9 143.8
145 29.8 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023387.D
 Acq On : 07 Oct 2025 11:47
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:43:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	590340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	858767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	749869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	392169	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	231295	45.240	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	90.480%		
35) Dibromofluoromethane	7.634	113	256331	49.185	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.360%		
50) Toluene-d8	10.103	98	969649	49.334	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	98.660%		
62) 4-Bromofluorobenzene	12.402	95	313426	50.546	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	101.100%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	189467	40.745	ug/l	100
3) Chloromethane	2.068	50	265142	40.113	ug/l	100
4) Vinyl Chloride	2.202	62	352993	42.293	ug/l	100
5) Bromomethane	2.592	94	277005	40.458	ug/l	100
6) Chloroethane	2.733	64	244551	43.157	ug/l	100
7) Trichlorofluoromethane	3.056	101	496981	46.354	ug/l	100
8) Diethyl Ether	3.452	74	126634	45.158	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.812	101	258100	45.039	ug/l	100
10) Methyl Iodide	4.001	142	317731	47.979	ug/l	100
11) Tert butyl alcohol	4.860	59	73642	244.793	ug/l	100
12) 1,1-Dichloroethene	3.787	96	252190	45.738	ug/l	100
13) Acrolein	3.647	56	123642	242.575	ug/l	100
14) Allyl chloride	4.379	41	310569	44.989	ug/l	100
15) Acrylonitrile	5.055	53	240668	219.075	ug/l	100
16) Acetone	3.866	43	247122	214.955	ug/l	100
17) Carbon Disulfide	4.104	76	722982	42.028	ug/l	100
18) Methyl Acetate	4.385	43	144816	49.563	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	601934	46.626	ug/l	100
20) Methylene Chloride	4.616	84	270612	43.409	ug/l	100
21) trans-1,2-Dichloroethene	5.110	96	277516	45.256	ug/l	100
22) Diisopropyl ether	6.019	45	686481	44.844	ug/l	100
23) Vinyl Acetate	5.958	43	1843784	219.119	ug/l	100
24) 1,1-Dichloroethane	5.915	63	448267	44.423	ug/l	100
25) 2-Butanone	6.890	43	310237	218.327	ug/l	100
26) 2,2-Dichloropropane	6.884	77	416391	46.440	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	320028	46.010	ug/l	100
28) Bromochloromethane	7.244	49	157414	40.211	ug/l	100
29) Tetrahydrofuran	7.262	42	187787	222.653	ug/l	100
30) Chloroform	7.421	83	493098	45.316	ug/l	100
31) Cyclohexane	7.701	56	379644	42.602	ug/l	100
32) 1,1,1-Trichloroethane	7.616	97	452713	46.347	ug/l	100
36) 1,1-Dichloropropene	7.835	75	341353	46.424	ug/l	100
37) Ethyl Acetate	6.982	43	129053	44.394	ug/l	100
38) Carbon Tetrachloride	7.817	117	413187	48.374	ug/l	100
39) Methylcyclohexane	9.109	83	452959	48.113	ug/l	100
40) Benzene	8.079	78	1080883	47.154	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023387.D
 Acq On : 07 Oct 2025 11:47
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:43:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	68542	45.399	ug/l	100
42) 1,2-Dichloroethane	8.152	62	270387	46.350	ug/l	100
43) Isopropyl Acetate	8.195	43	259041	45.811	ug/l	100
44) Trichloroethene	8.866	130	315104	48.982	ug/l	100
45) 1,2-Dichloropropane	9.140	63	234978	45.555	ug/l	100
46) Dibromomethane	9.231	93	148914	47.198	ug/l	100
47) Bromodichloromethane	9.420	83	377084	47.482	ug/l	100
48) Methyl methacrylate	9.219	41	126732	48.027	ug/l	100
49) 1,4-Dioxane	9.225	88	34053	1042.312	ug/l	100
51) 4-Methyl-2-Pentanone	9.993	43	682879	235.772	ug/l	100
52) Toluene	10.170	92	710404	48.824	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	330392	48.438	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	390619	47.910	ug/l	100
55) 1,1,2-Trichloroethane	10.573	97	198478	47.296	ug/l	100
56) Ethyl methacrylate	10.438	69	239407	49.535	ug/l	100
57) 1,3-Dichloropropane	10.713	76	321276	47.301	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	553816	261.921	ug/l	100
59) 2-Hexanone	10.756	43	479606	239.162	ug/l	100
60) Dibromochloromethane	10.908	129	281047	49.103	ug/l	100
61) 1,2-Dibromoethane	11.012	107	191356	48.468	ug/l	100
64) Tetrachloroethene	10.646	164	392056	52.264	ug/l	100
65) Chlorobenzene	11.438	112	789650	48.460	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.511	131	279583	48.788	ug/l	100
67) Ethyl Benzene	11.518	91	1332350	50.099	ug/l	100
68) m/p-Xylenes	11.627	106	1077583	101.092	ug/l	100
69) o-Xylene	11.950	106	502722	50.946	ug/l	100
70) Styrene	11.969	104	831908	51.059	ug/l	100
71) Bromoform	12.127	173	172315	50.229	ug/l #	100
73) Isopropylbenzene	12.249	105	1301009	49.803	ug/l	100
74) N-amyl acetate	12.066	43	227684	46.640	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.505	83	195565	44.878	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	169814m	43.898	ug/l	100
77) Bromobenzene	12.530	156	331331	49.419	ug/l	100
78) n-propylbenzene	12.591	91	1522965	49.572	ug/l	100
79) 2-Chlorotoluene	12.676	91	873335	49.235	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	1061112	49.966	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	66299	46.212	ug/l	100
82) 4-Chlorotoluene	12.773	91	892764	48.192	ug/l	100
83) tert-Butylbenzene	12.993	119	988936	51.104	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	1061159	50.505	ug/l	100
85) sec-Butylbenzene	13.170	105	1405003	50.133	ug/l	100
86) p-Isopropyltoluene	13.286	119	1210644	51.115	ug/l	100
87) 1,3-Dichlorobenzene	13.286	146	649899	49.072	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	636694	48.476	ug/l	100
89) n-Butylbenzene	13.615	91	1051949	50.299	ug/l	100
90) Hexachloroethane	13.877	117	245108	47.664	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	567114	49.075	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	32585	47.184	ug/l	100
93) 1,2,4-Trichlorobenzene	14.919	180	359833	52.852	ug/l	100
94) Hexachlorobutadiene	15.017	225	226316	53.346	ug/l	100
95) Naphthalene	15.139	128	575044	53.012	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	310831	52.628	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023387.D
Acq On : 07 Oct 2025 11:47
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:43:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 04:39:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023387.D
 Acq On : 07 Oct 2025 11:47
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Quant Time: Oct 08 04:43:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By : Semsettin Yesilyurt 10/10/2025
 Supervised By : Mahesh Dadoda 10/10/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion:168 Resp: 590340
Ion Ratio Lower Upper
168 100
99 49.6 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 40.745 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47


Tgt Ion: 85 Resp: 189467
Ion Ratio Lower Upper
85 100
87 31.9 16.0 47.9





#3
Chloromethane
Concen: 40.113 ug/l
RT: 2.068 min Scan# 57
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion: 50 Resp: 26514
Ion Ratio Lower Upper
50 100
52 32.7 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#4
Vinyl Chloride
Concen: 42.293 ug/l
RT: 2.202 min Scan# 79
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 62 Resp: 352993
Ion Ratio Lower Upper
62 100
64 32.3 25.8 38.8









#5
Bromomethane
Concen: 40.458 ug/l
RT: 2.592 min Scan# 143
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Tgt Ion: 94 Resp: 277000
Ion Ratio Lower Upper
94 100
96 93.1 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#6
Chloroethane
Concen: 43.157 ug/l
RT: 2.733 min Scan# 166
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 64 Resp: 244551
Ion Ratio Lower Upper
64 100
66 32.8 26.2 39.4









#7
Trichlorofluoromethane
Concen: 46.354 ug/l
RT: 3.056 min Scan# 219
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 49698
Ion Ratio Lower Upper
101 100
103 64.1 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#8
Diethyl Ether
Concen: 45.158 ug/l
RT: 3.452 min Scan# 284
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 74 Resp: 126634
Ion Ratio Lower Upper
74 100
45 79.8 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.039 ug/l
RT: 3.812 min Scan# 343
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 258100
Ion Ratio Lower Upper
101 100
85 43.2 34.6 51.8
151 83.8 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#10
Methyl Iodide
Concen: 47.979 ug/l
RT: 4.001 min Scan# 374
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion:142 Resp: 317731
Ion Ratio Lower Upper
142 100
127 43.5 34.8 52.2
141 14.5 11.6 17.4





#11

Tert butyl alcohol

Concen: 244.793 ug/l

RT: 4.860 min Scan# 515

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 7364

Ion Ratio Lower Upper

59 100





57 10.4 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 45.738 ug/l

RT: 3.787 min Scan# 339

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion: 96 Resp: 252190


Ion Ratio Lower Upper

96 100

61 133.4 106.7 160.1


98 62.8 50.2 75.4





#13
Acrolein
Concen: 242.575 ug/l
RT: 3.647 min Scan# 316
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 123642
Ion Ratio Lower Upper
56 100
55 70.1 56.1 84.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#14
Allyl chloride
Concen: 44.989 ug/l
RT: 4.379 min Scan# 436
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 41 Resp: 310569
Ion Ratio Lower Upper
41 100
39 74.9 59.9 89.9
76 47.5 38.0 57.0





#15

Acrylonitrile

Concen: 219.075 ug/l

RT: 5.055 min Scan# 547

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 53 Resp: 240668

Ion Ratio Lower Upper

53 100

52 80.8 64.6 97.0





51 36.9 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 214.955 ug/l

RT: 3.866 min Scan# 352

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 43 Resp: 247122

Ion Ratio Lower Upper

43 100

58 34.6 27.7 41.5





#17

Carbon Disulfide

Concen: 42.028 ug/l

RT: 4.104 min Scan# 391

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 72298

Ion Ratio Lower Upper

76 100





78 9.1 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#18

Methyl Acetate

Concen: 49.563 ug/l

RT: 4.385 min Scan# 437

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 43 Resp: 144816

Ion Ratio Lower Upper

43 100

74 30.3 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 46.626 ug/l

RT: 5.116 min Scan# 5

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 601934

Ion Ratio Lower Upper

73 100





57 21.7 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 43.409 ug/l

RT: 4.616 min Scan# 475

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion: 84 Resp: 270612




Ion Ratio Lower Upper


84 100

49 100.3 80.2 120.4

51 32.4 25.9 38.9


86 63.2 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 45.256 ug/l
RT: 5.110 min Scan# 556
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 277510
Ion Ratio Lower Upper
96 100
61 115.9 92.7 139.1
98 62.1 49.7 74.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#22
Diisopropyl ether
Concen: 44.844 ug/l
RT: 6.019 min Scan# 705
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 45 Resp: 686481
Ion Ratio Lower Upper
45 100
43 57.0 45.6 68.4
87 35.1 28.1 42.1
59 12.7 10.2 15.2





#23

Vinyl Acetate

Concen: 219.119 ug/l

RT: 5.958 min Scan# 695

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 1843784

Ion Ratio Lower Upper

43

100

86

13.2

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 44.423 ug/l

RT: 5.915 min Scan# 688

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion: 63 Resp: 448267

Ion Ratio Lower Upper

63

100

98




8.7


100

5.6

2.8

8.4





#25

2-Butanone

Concen: 218.327 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 31023

Ion Ratio Lower Upper

43 100





72 34.3 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 46.440 ug/l

RT: 6.884 min Scan# 847

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 77 Resp: 416391

Ion Ratio Lower Upper

77 100

97 24.2 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 46.010 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 320028

Ion Ratio Lower Upper

96 100

61 129.1 0.0 258.2





98 66.0 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane

Concen: 40.211 ug/l

RT: 7.244 min Scan# 906

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion: 49 Resp: 157414


Ion Ratio Lower Upper

49 100

129 2.8 0.0 5.6

130 102.4 81.9 122.9





#29

Tetrahydrofuran

Concen: 222.653 ug/l

RT: 7.262 min Scan# 909

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 18778

Ion Ratio Lower Upper

42 100

72 53.8 43.0 64.6





71 50.1 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 45.316 ug/l

RT: 7.421 min Scan# 935

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 83 Resp: 493098

Ion Ratio Lower Upper

83 100

85 65.6 52.5 78.7





#31

Cyclohexane

Concen: 42.602 ug/l

RT: 7.701 min Scan# 981

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 379644

Ion Ratio Lower Upper

56 100

69 37.5 30.0 45.0





84 101.5 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#32

1,1,1-Trichloroethane

Concen: 46.347 ug/l

RT: 7.616 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion: 97 Resp: 452713


Ion Ratio Lower Upper

97 100

99 64.3 51.4 77.2

61 39.9 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 45.240 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 23129

Ion Ratio Lower Upper

65 100





67 52.9 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion:114 Resp: 858767


Ion Ratio Lower Upper

114 100

63 17.1 0.0 34.2

88 14.2 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.185 ug/l

RT: 7.634 min Scan# 91

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion:113 Resp: 25633

Ion Ratio Lower Upper

113 100

111 102.2 81.8 122.6





192 22.5 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 46.424 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion: 75 Resp: 341353


Ion Ratio Lower Upper

75 100

110 39.8 19.9 59.7

77 31.3 25.0 37.6





#37

Ethyl Acetate

Concen: 44.394 ug/l

RT: 6.982 min Scan# 863

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 12905

Ion Ratio Lower Upper

43 100

61 14.8 11.8 17.8





70 13.1 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 48.374 ug/l

RT: 7.817 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion:117 Resp: 413187


Ion Ratio Lower Upper

117 100

119 96.0 76.8 115.2

121 30.6 24.5 36.7





#39

Methylcyclohexane

Concen: 48.113 ug/l

RT: 9.109 min Scan# 1043

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 452959

Ion Ratio Lower Upper

83 100

55 64.3 51.4 77.2





98 50.5 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 47.154 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 78 Resp: 1080883

Ion Ratio Lower Upper

78 100

77 22.9 18.3 27.5





#41

Methacrylonitrile

Concen: 45.399 ug/l

RT: 7.220 min Scan# 902

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 41 Resp: 6854

Ion Ratio Lower Upper

41 100

39 65.3 52.2 78.4

67 93.6 74.9 112.3





52 36.5 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#42

1,2-Dichloroethane

Concen: 46.350 ug/l

RT: 8.152 min Scan# 1055

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 62 Resp: 270387

Ion Ratio Lower Upper

62 100

98 10.8 0.0 21.6





#43

Isopropyl Acetate

Concen: 45.811 ug/l

RT: 8.195 min Scan# 1062

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 25904

Ion Ratio Lower Upper

43 100

61 22.6 18.1 27.1





87 16.9 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 48.982 ug/l

RT: 8.866 min Scan# 1172

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion:130 Resp: 315104

Ion Ratio Lower Upper

130 100

95 90.3 0.0 180.6





#45

1,2-Dichloropropane

Concen: 45.555 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 234978

Ion Ratio Lower Upper

63 100





65 31.1 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 47.198 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion: 93 Resp: 148914


Ion Ratio Lower Upper

93 100

95 81.7 65.4 98.0

174 116.1 92.9 139.3





#47

Bromodichloromethane

Concen: 47.482 ug/l

RT: 9.420 min Scan# 1263

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICCC050

Tgt Ion:	83	Resp:	377084
Ion	Ratio	Lower	Upper
83	100		
85	63.5	50.8	76.2
127	9.3	7.4	11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#48

Methyl methacrylate

Concen: 48.027 ug/l




RT: 9.219 min Scan# 1230


Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion:	41	Resp:	126732
Ion	Ratio	Lower	Upper
41	100		
69	101.2	81.0	121.4
39	60.3	48.2	72.4





#49

1,4-Dioxane

Concen: 1042.312 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICCC050

Tgt Ion:	88	Resp:	3405
Ion	Ratio	Lower	Upper
88	100		
43	27.3	21.8	32.8
58	58.5	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 49.334 ug/l




RT: 10.103 min Scan# 1375


Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion:	98	Resp:	969649
Ion	Ratio	Lower	Upper
98	100		
100	66.2	53.0	79.4





#51

4-Methyl-2-Pentanone

Concen: 235.772 ug/l

RT: 9.993 min Scan# 1357

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 682879

Ion Ratio Lower Upper

43 100





58 44.1 35.3 52.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#52

Toluene

Concen: 48.824 ug/l

RT: 10.170 min Scan# 1386

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 92 Resp: 710404

Ion Ratio Lower Upper

92 100

91 170.5 136.4 204.6









#53
t-1,3-Dichloropropene
Concen: 48.438 ug/l
RT: 10.390 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 330392
Ion Ratio Lower Upper
75 100
77 31.4 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#54
cis-1,3-Dichloropropene
Concen: 47.910 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47


Tgt Ion: 75 Resp: 390619
Ion Ratio Lower Upper
75 100
77 31.7 25.4 38.0
39 41.0 32.8 49.2





#55
1,1,2-Trichloroethane
Concen: 47.296 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion: 97 Resp: 198478
Ion Ratio Lower Upper
97 100
83 77.2 61.8 92.6
85 51.6 41.3 61.9
99 60.9 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#56
Ethyl methacrylate
Concen: 49.535 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 69 Resp: 239407
Ion Ratio Lower Upper
69 100
41 55.8 44.6 67.0
39 33.7 27.0 40.4





#57

1,3-Dichloropropane

Concen: 47.301 ug/l

RT: 10.713 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 321270

Ion Ratio Lower Upper

76 100





78 32.4 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 261.921 ug/l

RT: 9.707 min Scan# 1310

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 63 Resp: 553816

Ion Ratio Lower Upper

63 100

106 33.2 26.6 39.8





#59

2-Hexanone

Concen: 239.162 ug/l

RT: 10.756 min Scan# 1482

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 479600

Ion Ratio Lower Upper

43 100





58 61.0 30.5 91.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#60

Dibromochloromethane

Concen: 49.103 ug/l

RT: 10.908 min Scan# 1507

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion:129 Resp: 281047

Ion Ratio Lower Upper

129 100


127 77.1 38.6 115.7





#61
1,2-Dibromoethane
Concen: 48.468 ug/l
RT: 11.012 min Scan# 1524
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion: 107 Resp: 191350
Ion Ratio Lower Upper
107 100
109 95.1 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 50.546 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 95 Resp: 313426
Ion Ratio Lower Upper
95 100
174 96.4 0.0 192.8
176 93.8 0.0 187.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion:117 Resp: 749869

Ion Ratio Lower Upper

117 100

82 51.2 41.0 61.4





119 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#64

Tetrachloroethene

Concen: 52.264 ug/l

RT: 10.646 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion:164 Resp: 392056




Ion Ratio Lower Upper


164 100

166 124.2 99.4 149.0

129 86.8 69.4 104.2

131 84.4 67.5 101.3





#65

Chlorobenzene

Concen: 48.460 ug/l

RT: 11.438 min Scan# 1594

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion:112 Resp: 789650

Ion Ratio Lower Upper

112 100





114 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 48.788 ug/l

RT: 11.511 min Scan# 1606

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion:131 Resp: 279583


Ion Ratio Lower Upper

131 100

133 96.8 48.4 145.2

119 64.8 32.4 97.2





#67

Ethyl Benzene

Concen: 50.099 ug/l

RT: 11.518 min Scan# 1607

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 1332350

Ion Ratio Lower Upper

91 100





106 32.3 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#68

m/p-Xylenes

Concen: 101.092 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion:106 Resp: 1077583

Ion Ratio Lower Upper

106 100

91 190.8 152.6 229.0









#69
o-Xylene
Concen: 50.946 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 502722
Ion Ratio Lower Upper
106 100
91 202.7 101.4 304.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#70
Styrene
Concen: 51.059 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47


Tgt Ion:104 Resp: 831908
Ion Ratio Lower Upper
104 100
78 45.1 36.1 54.1
103 54.0 43.2 64.8





#71
Bromoform
Concen: 50.229 ug/l
RT: 12.127 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 172315
Ion Ratio Lower Upper
173 100
175 49.5 24.8 74.3
254 0.1 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47


Tgt Ion:152 Resp: 392169
Ion Ratio Lower Upper
152 100
115 54.1 27.1 81.2
150 173.7 0.0 347.4





#73
Isopropylbenzene
Concen: 49.803 ug/l
RT: 12.249 min Scan# 1727
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion:105 Resp: 1301009
Ion Ratio Lower Upper
105 100
120 27.8 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 46.640 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 43 Resp: 227684
Ion Ratio Lower Upper
43 100
70 51.4 41.1 61.7
55 28.8 23.0 34.6
61 26.8 21.4 32.2





#75

1,1,2-Tetrachloroethane

Concen: 44.878 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 19556

Ion Ratio Lower Upper

83 100

131 11.7 5.9 17.6





85 63.5 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#76

1,2,3-Trichloropropane

Concen: 43.898 ug/l m

RT: 12.554 min Scan# 1777

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion: 75 Resp: 169814

Ion Ratio Lower Upper

75 100


77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 49.419 ug/l
RT: 12.530 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050




Tgt Ion:156 Resp: 33133
Ion Ratio Lower Upper
156 100
77 153.7 76.8 230.5
158 98.1 49.0 147.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#78
n-propylbenzene
Concen: 49.572 ug/l
RT: 12.591 min Scan# 1783
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47



Tgt Ion: 91 Resp: 1522965
Ion Ratio Lower Upper
91 100
120 24.5 12.3 36.8





#79

2-Chlorotoluene

Concen: 49.235 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 87333

Ion Ratio Lower Upper

91 100





126 37.0 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 49.966 ug/l

RT: 12.731 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion:105 Resp: 1061112

Ion Ratio Lower Upper

105 100


120 50.9 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 46.212 ug/l
RT: 12.298 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 66295
Ion Ratio Lower Upper
75 100
53 85.7 68.6 102.8
89 48.7 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#82
4-Chlorotoluene
Concen: 48.192 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 91 Resp: 892764
Ion Ratio Lower Upper
91 100
126 36.9 18.4 55.4





#83

tert-Butylbenzene

Concen: 51.104 ug/l

RT: 12.993 min Scan# 1849

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICCC050

Tgt Ion	Ratio	Lower	Upper
119	100		
91	59.9	29.9	89.8
134	25.8	12.9	38.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#84

1,2,4-Trimethylbenzene

Concen: 50.505 ug/l




RT: 13.042 min Scan# 1857


Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion	Ratio	Lower	Upper
105	100		
120	47.5	23.8	71.2





#85

sec-Butylbenzene

Concen: 50.133 ug/l

RT: 13.170 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion:105 Resp: 140500

Ion Ratio Lower Upper

105 100





134 21.4 10.7 32.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#86

p-Isopropyltoluene

Concen: 51.115 ug/l

RT: 13.286 min Scan# 1897

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion:119 Resp: 1210644


Ion Ratio Lower Upper

119 100

134 27.1 13.6 40.7

91 21.9 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 49.072 ug/l

RT: 13.286 min Scan# 1897

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 649899

Ion Ratio Lower Upper

146 100

111 37.4 18.7 56.1





148 63.9 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 48.476 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion:146 Resp: 636694


Ion Ratio Lower Upper

146 100

111 36.0 18.0 54.0

148 63.8 31.9 95.7





#89

n-Butylbenzene

Concen: 50.299 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 1051949

Ion Ratio Lower Upper

91 100

92 53.9 27.0 80.8





134 28.1 14.1 42.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#90

Hexachloroethane

Concen: 47.664 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. 0.000 min

Lab File: VY023387.D




Acq: 07 Oct 2025 11:47


Tgt Ion:117 Resp: 245108

Ion Ratio Lower Upper

117 100


201 87.6 43.8 131.4





#91
1,2-Dichlorobenzene
Concen: 49.075 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050



Tgt Ion:146 Resp: 567114
Ion Ratio Lower Upper
146 100
111 38.0 19.0 57.0
148 64.2 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.184 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. 0.000 min
Lab File: VY023387.D
Acq: 07 Oct 2025 11:47

Tgt Ion: 75 Resp: 32585
Ion Ratio Lower Upper
75 100
155 100.0 50.0 150.0
157 127.7 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 52.852 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICCC050

Tgt Ion:180 Resp: 35983

Ion Ratio Lower Upper

180 100

182 95.7 47.9 143.6





145 29.4 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene

Concen: 53.346 ug/l

RT: 15.017 min Scan# 2181

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47




Tgt Ion:225 Resp: 226316


Ion Ratio Lower Upper

225 100

223 61.5 30.8 92.3

227 63.8 31.9 95.7





#95

Naphthalene

Concen: 53.012 ug/l

RT: 15.139 min Scan# 2201

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Instrument :

MSVOA_Y





ClientSampleId :

VSTDICCC050

Tgt Ion:	128	Resp:	57504
Ion Ratio	Lower	Upper	
128	100		
127	13.0	10.4	15.6
129	11.1	8.9	13.3

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#96

1,2,3-Trichlorobenzene

Concen: 52.628 ug/l




RT: 15.328 min Scan# 2232

Delta R.T. 0.000 min

Lab File: VY023387.D

Acq: 07 Oct 2025 11:47

Tgt Ion:	180	Resp:	310831
Ion Ratio	Lower	Upper	
180	100		
182	95.9	47.9	143.8
145	30.6	15.3	45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023388.D
 Acq On : 07 Oct 2025 12:09
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Quant Time: Oct 08 04:44:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/10/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	595770	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	855682	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	764361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	400285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	467005	90.510	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 181.020%#			
35) Dibromofluoromethane	7.634	113	511478	98.497	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 197.000%#			
50) Toluene-d8	10.103	98	1919872	98.031	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 196.060%#			
62) 4-Bromofluorobenzene	12.401	95	636435	103.007	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 206.020%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	368985	78.627	ug/l	96
3) Chloromethane	2.068	50	521606	78.195	ug/l	96
4) Vinyl Chloride	2.202	62	711235	84.438	ug/l	100
5) Bromomethane	2.586	94	574957	83.210	ug/l	99
6) Chloroethane	2.732	64	491625	85.968	ug/l	99
7) Trichlorofluoromethane	3.049	101	1012851	93.610	ug/l	99
8) Diethyl Ether	3.452	74	266779	94.266	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.812	101	511482	88.441	ug/l	100
10) Methyl Iodide	4.001	142	651575	97.495	ug/l	99
11) Tert butyl alcohol	4.866	59	172718	568.898	ug/l	99
12) 1,1-Dichloroethene	3.787	96	508059	91.303	ug/l	96
13) Acrolein	3.647	56	262636	510.573	ug/l	99
14) Allyl chloride	4.378	41	637073	91.446	ug/l	99
15) Acrylonitrile	5.055	53	522511	471.296	ug/l	100
16) Acetone	3.866	43	504700	435.004	ug/l	98
17) Carbon Disulfide	4.104	76	1442005	83.061	ug/l	99
18) Methyl Acetate	4.385	43	316911	107.474	ug/l	98
19) Methyl tert-butyl Ether	5.116	73	1300206	99.797	ug/l	97
20) Methylene Chloride	4.610	84	552018	87.742	ug/l	98
21) trans-1,2-Dichloroethene	5.110	96	569334	91.999	ug/l	96
22) Diisopropyl ether	6.018	45	1423347	92.133	ug/l	97
23) Vinyl Acetate	5.957	43	3975052	468.098	ug/l	99
24) 1,1-Dichloroethane	5.915	63	911619	89.517	ug/l	99
25) 2-Butanone	6.890	43	685693	478.154	ug/l	99
26) 2,2-Dichloropropane	6.884	77	833452	92.108	ug/l	98
27) cis-1,2-Dichloroethene	6.890	96	670773	95.557	ug/l	98
28) Bromochloromethane	7.244	49	321657	81.418	ug/l	99
29) Tetrahydrofuran	7.262	42	415867	488.587	ug/l	99
30) Chloroform	7.421	83	1003722	91.401	ug/l	100
31) Cyclohexane	7.701	56	767083	85.293	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	910837	92.399	ug/l	100
36) 1,1-Dichloropropene	7.835	75	692477	94.516	ug/l	99
37) Ethyl Acetate	6.982	43	280560	96.859	ug/l	98
38) Carbon Tetrachloride	7.817	117	833170	97.894	ug/l	99
39) Methylcyclohexane	9.109	83	947664	101.024	ug/l	99
40) Benzene	8.079	78	2192645	96.000	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023388.D
 Acq On : 07 Oct 2025 12:09
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:44:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	152239	101.200	ug/l	98
42) 1,2-Dichloroethane	8.158	62	554810	95.449	ug/l	100
43) Isopropyl Acetate	8.195	43	573653	101.815	ug/l	99
44) Trichloroethene	8.866	130	636149	99.245	ug/l	100
45) 1,2-Dichloropropane	9.140	63	485084	94.381	ug/l	99
46) Dibromomethane	9.231	93	309964	98.596	ug/l	99
47) Bromodichloromethane	9.420	83	769238	97.210	ug/l	99
48) Methyl methacrylate	9.219	41	282582	107.476	ug/l	99
49) 1,4-Dioxane	9.225	88	77305	2374.722	ug/l	99
51) 4-Methyl-2-Pentanone	9.993	43	1539324	533.386	ug/l	100
52) Toluene	10.170	92	1462927	100.905	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	710475	104.537	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	822967	101.302	ug/l	98
55) 1,1,2-Trichloroethane	10.573	97	417623	99.875	ug/l	96
56) Ethyl methacrylate	10.438	69	549366	114.077	ug/l	99
57) 1,3-Dichloropropane	10.713	76	670961	99.140	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	1240361	588.731	ug/l	99
59) 2-Hexanone	10.755	43	1081568	541.284	ug/l	99
60) Dibromochloromethane	10.908	129	593878	104.133	ug/l	100
61) 1,2-Dibromoethane	11.011	107	406293	103.280	ug/l	99
64) Tetrachloroethene	10.646	164	767753	100.406	ug/l	99
65) Chlorobenzene	11.438	112	1655143	99.649	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	586878	100.470	ug/l	99
67) Ethyl Benzene	11.517	91	2788252	102.856	ug/l	98
68) m/p-Xylenes	11.627	106	2243033	206.437	ug/l	100
69) o-Xylene	11.950	106	1072335	106.611	ug/l	99
70) Styrene	11.969	104	1782757	107.344	ug/l	100
71) Bromoform	12.127	173	373208	106.727	ug/l #	100
73) Isopropylbenzene	12.249	105	2741698	102.825	ug/l	100
74) N-amyl acetate	12.066	43	532255	106.818	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	434720	97.735	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	364622m	92.345	ug/l	
77) Bromobenzene	12.529	156	701741	102.545	ug/l	100
78) n-propylbenzene	12.590	91	3165894	100.958	ug/l	99
79) 2-Chlorotoluene	12.676	91	1800124	99.426	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	2219080	102.374	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.298	75	148302	101.273	ug/l	98
82) 4-Chlorotoluene	12.773	91	1854165	98.060	ug/l	99
83) tert-Butylbenzene	12.993	119	2028198	102.684	ug/l	97
84) 1,2,4-Trimethylbenzene	13.042	105	2227197	103.853	ug/l	100
85) sec-Butylbenzene	13.170	105	2932949	102.530	ug/l	100
86) p-Isopropyltoluene	13.285	119	2548220	105.408	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	1361716	100.735	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	1323618	98.734	ug/l	100
89) n-Butylbenzene	13.615	91	2223744	104.174	ug/l	99
90) Hexachloroethane	13.877	117	514746	98.069	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	1193043	101.146	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	70107	99.458	ug/l	94
93) 1,2,4-Trichlorobenzene	14.919	180	806328	116.031	ug/l	99
94) Hexachlorobutadiene	15.017	225	470072	108.557	ug/l	99
95) Naphthalene	15.139	128	1378291	124.485	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	693477	115.034	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023388.D
Acq On : 07 Oct 2025 12:09
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:44:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 04:39:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023388.D
 Acq On : 07 Oct 2025 12:09
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Quant Time: Oct 08 04:44:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/10/2025
 Supervised By : Mahesh Dadoda 10/10/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Tgt Ion:168 Resp: 595770
Ion Ratio Lower Upper
168 100
99 49.3 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 78.627 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion: 85 Resp: 368985
Ion Ratio Lower Upper
85 100
87 34.0 16.0 47.9









#3
Chloromethane
Concen: 78.195 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Tgt Ion: 50 Resp: 521600
Ion Ratio Lower Upper
50 100
52 34.7 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#4
Vinyl Chloride
Concen: 84.438 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion: 62 Resp: 711235
Ion Ratio Lower Upper
62 100
64 32.5 25.8 38.8









#5
Bromomethane
Concen: 83.210 ug/l
RT: 2.586 min Scan# 143
Delta R.T. -0.006 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Tgt Ion: 94 Resp: 57495
Ion Ratio Lower Upper
94 100
96 93.7 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#6
Chloroethane
Concen: 85.968 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion: 64 Resp: 491625
Ion Ratio Lower Upper
64 100
66 32.0 26.2 39.4









#7
Trichlorofluoromethane
Concen: 93.610 ug/l
RT: 3.049 min Scan# 219
Delta R.T. -0.006 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC100

Tgt Ion:101 Resp: 101285
Ion Ratio Lower Upper
101 100
103 65.2 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#8
Diethyl Ether
Concen: 94.266 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion: 74 Resp: 266779
Ion Ratio Lower Upper
74 100
45 79.9 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 88.441 ug/l
RT: 3.812 min Scan# 343
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Tgt Ion:101 Resp: 51148
Ion Ratio Lower Upper
101 100
85 43.6 34.6 51.8
151 84.1 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#10
Methyl Iodide
Concen: 97.495 ug/l
RT: 4.001 min Scan# 374
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion:142 Resp: 651575
Ion Ratio Lower Upper
142 100
127 42.7 34.8 52.2
141 14.3 11.6 17.4





#11

Tert butyl alcohol

Concen: 568.898 ug/l

RT: 4.866 min Scan# 515

Delta R.T. 0.006 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 172718

Ion Ratio Lower Upper

59 100





57 10.1 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 91.303 ug/l

RT: 3.787 min Scan# 339

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 96 Resp: 508059


Ion Ratio Lower Upper

96 100

61 139.9 106.7 160.1

98 64.6 50.2 75.4





#13

Acrolein

Concen: 510.573 ug/l

RT: 3.647 min Scan# 316

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 262630

Ion Ratio Lower Upper

56 100





55 69.6 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#14

Allyl chloride

Concen: 91.446 ug/l

RT: 4.378 min Scan# 436

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 41 Resp: 637073


Ion Ratio Lower Upper

41 100

39 75.8 59.9 89.9

76 47.5 38.0 57.0





#15

Acrylonitrile

Concen: 471.296 ug/l

RT: 5.055 min Scan# 547

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 53 Resp: 52251

Ion Ratio Lower Upper

53 100

52 81.2 64.6 97.0





51 36.7 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 435.004 ug/l

RT: 3.866 min Scan# 352

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 43 Resp: 504700

Ion Ratio Lower Upper

43 100

58 33.5 27.7 41.5





#17

Carbon Disulfide

Concen: 83.061 ug/l

RT: 4.104 min Scan# 391

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 144200

Ion Ratio Lower Upper

76 100





78 9.6 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#18

Methyl Acetate

Concen: 107.474 ug/l

RT: 4.385 min Scan# 437

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 43 Resp: 316911

Ion Ratio Lower Upper

43 100

74 29.1 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 99.797 ug/l

RT: 5.116 min Scan# 51

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 73 Resp: 1300200

Ion Ratio Lower Upper

73 100





57 20.1 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 87.742 ug/l

RT: 4.610 min Scan# 474

Delta R.T. -0.006 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Tgt Ion: 84 Resp: 552018




Ion Ratio Lower Upper


84 100

49 102.4 80.2 120.4

51 33.3 25.9 38.9

86 62.4 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 91.999 ug/l

RT: 5.110 min Scan# 556

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 96 Resp: 569334

Ion Ratio Lower Upper

96 100

61 120.4 92.7 139.1





98 64.2 49.7 74.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#22

Diisopropyl ether

Concen: 92.133 ug/l

RT: 6.018 min Scan# 705

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Tgt Ion: 45 Resp: 1423347




Ion Ratio Lower Upper


45 100

43 54.4 45.6 68.4

87 33.8 28.1 42.1

59 12.9 10.2 15.2





#23

Vinyl Acetate

Concen: 468.098 ug/l

RT: 5.957 min Scan# 695

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 397505

Ion Ratio Lower Upper

43

100

86

13.5

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 89.517 ug/l

RT: 5.915 min Scan# 688

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Tgt Ion: 63 Resp: 911619

Ion Ratio Lower Upper

63

100

98




8.3


100

5.1

2.8

8.4





#25

2-Butanone

Concen: 478.154 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 68569

Ion Ratio Lower Upper

43 100





72 33.9 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 92.108 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 77 Resp: 833452

Ion Ratio Lower Upper

77 100

97 25.0 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 95.557 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 96 Resp: 67077

Ion Ratio Lower Upper

96 100

61 125.9 0.0 258.2





98 64.7 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane

Concen: 81.418 ug/l

RT: 7.244 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 49 Resp: 321657


Ion Ratio Lower Upper

49 100

129 2.5 0.0 5.6

130 100.9 81.9 122.9





#29

Tetrahydrofuran

Concen: 488.587 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 41586

Ion Ratio Lower Upper

42 100

72 54.1 43.0 64.6





71 50.9 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 91.401 ug/l

RT: 7.421 min Scan# 935

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 83 Resp: 1003722

Ion Ratio Lower Upper

83 100

85 65.4 52.5 78.7





#31

Cyclohexane

Concen: 85.293 ug/l

RT: 7.701 min Scan# 981

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 56 Resp: 767083

Ion Ratio Lower Upper

56 100

69 34.9 30.0 45.0





84 100.2 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#32

1,1,1-Trichloroethane

Concen: 92.399 ug/l

RT: 7.616 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 97 Resp: 910837


Ion Ratio Lower Upper

97 100

99 64.3 51.4 77.2

61 39.7 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 90.510 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 467000

Ion Ratio Lower Upper

65 100





67 53.0 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion:114 Resp: 855682


Ion Ratio Lower Upper

114 100

63 17.0 0.0 34.2

88 13.9 0.0 28.4





#35

Dibromofluoromethane

Concen: 98.497 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion:113 Resp: 511478

Ion Ratio Lower Upper

113 100

111 102.4 81.8 122.6





192 22.5 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 94.516 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 75 Resp: 692477


Ion Ratio Lower Upper

75 100

110 39.2 19.9 59.7

77 31.4 25.0 37.6





#37

Ethyl Acetate

Concen: 96.859 ug/l

RT: 6.982 min Scan# 80

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 280560

Ion Ratio Lower Upper

43 100

61 15.9 11.8 17.8





70 12.4 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 97.894 ug/l

RT: 7.817 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion:117 Resp: 833170


Ion Ratio Lower Upper

117 100

119 96.7 76.8 115.2

121 30.9 24.5 36.7





#39

Methylcyclohexane

Concen: 101.024 ug/l

RT: 9.109 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 947664

Ion Ratio Lower Upper

83 100

55 65.4 51.4 77.2





98 51.3 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 96.000 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 78 Resp: 2192645

Ion Ratio Lower Upper

78 100


77 23.8 18.3 27.5





#41
Methacrylonitrile
Concen: 101.200 ug/l
RT: 7.219 min Scan# 902
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100




Tgt Ion: 41 Resp: 152239
Ion Ratio Lower Upper
41 100
39 62.6 52.2 78.4
67 93.9 74.9 112.3
52 34.9 29.2 43.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#42
1,2-Dichloroethane
Concen: 95.449 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09



Tgt Ion: 62 Resp: 554810
Ion Ratio Lower Upper
62 100
98 10.8 0.0 21.6





#43

Isopropyl Acetate

Concen: 101.815 ug/l

RT: 8.195 min Scan# 1062

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 57365

Ion Ratio Lower Upper

43 100

61 23.0 18.1 27.1





87 16.6 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 99.245 ug/l

RT: 8.866 min Scan# 1172

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion:130 Resp: 636149

Ion Ratio Lower Upper

130 100

95 90.1 0.0 180.6





#45

1,2-Dichloropropane

Concen: 94.381 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 485084

Ion Ratio Lower Upper

63 100





65 30.3 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 98.596 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 93 Resp: 309964


Ion Ratio Lower Upper

93 100

95 83.8 65.4 98.0

174 115.6 92.9 139.3





#47

Bromodichloromethane

Concen: 97.210 ug/l

RT: 9.420 min Scan# 1263

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 769238

Ion Ratio Lower Upper

83 100

85 64.1 50.8 76.2





127 9.0 7.4 11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#48

Methyl methacrylate

Concen: 107.476 ug/l

RT: 9.219 min Scan# 1230

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 41 Resp: 282582


Ion Ratio Lower Upper

41 100

69 100.8 81.0 121.4

39 60.9 48.2 72.4





#49

1,4-Dioxane

Concen: 2374.722 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y





ClientSampleId :

VSTDICC100

Tgt Ion:	88	Resp:	77308
Ion	Ratio	Lower	Upper
88	100		
43	27.5	21.8	32.8
58	58.0	46.8	70.2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 98.031 ug/l




RT: 10.103 min Scan# 1375


Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Tgt Ion:	98	Resp:	1919872
Ion	Ratio	Lower	Upper
98	100		
100	66.3	53.0	79.4





#51

4-Methyl-2-Pentanone

Concen: 533.386 ug/l

RT: 9.993 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 1539324

Ion Ratio Lower Upper

43 100





58 44.0 35.3 52.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#52

Toluene

Concen: 100.905 ug/l

RT: 10.170 min Scan# 1386

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 92 Resp: 1462927

Ion Ratio Lower Upper

92 100

91 169.6 136.4 204.6









#53
t-1,3-Dichloropropene
Concen: 104.537 ug/l
RT: 10.390 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 710475
Ion Ratio Lower Upper
75 100
77 32.3 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#54
cis-1,3-Dichloropropene
Concen: 101.302 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion: 75 Resp: 822967
Ion Ratio Lower Upper
75 100
77 32.2 25.4 38.0
39 42.5 32.8 49.2





#55

1,1,2-Trichloroethane

Concen: 99.875 ug/l

RT: 10.573 min Scan# 1452

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 41762

Ion Ratio Lower Upper

97 100

83 81.3 61.8 92.6

85 53.6 41.3 61.9





99 63.8 48.7 73.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#56

Ethyl methacrylate

Concen: 114.077 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 69 Resp: 549366


Ion Ratio Lower Upper

69 100

41 55.1 44.6 67.0

39 33.3 27.0 40.4





#57

1,3-Dichloropropane

Concen: 99.140 ug/l

RT: 10.713 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 67096

Ion Ratio Lower Upper

76 100





78 32.6 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 588.731 ug/l

RT: 9.707 min Scan# 1310

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion: 63 Resp: 1240361

Ion Ratio Lower Upper

63 100

106 32.8 26.6 39.8





#59

2-Hexanone

Concen: 541.284 ug/l

RT: 10.755 min Scan# 1482

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 108156

Ion Ratio Lower Upper

43 100





58 61.5 30.5 91.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#60

Dibromochloromethane

Concen: 104.133 ug/l

RT: 10.908 min Scan# 1507

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion:129 Resp: 593878

Ion Ratio Lower Upper

129 100


127 76.9 38.6 115.7





#61
1,2-Dibromoethane
Concen: 103.280 ug/l
RT: 11.011 min Scan# 1524
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion:107 Resp: 406293
Ion Ratio Lower Upper
107 100
109 94.0 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 103.007 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09


Tgt Ion: 95 Resp: 636435
Ion Ratio Lower Upper
95 100
174 96.7 0.0 192.8
176 94.1 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion:117 Resp: 76436
Ion Ratio Lower Upper
117 100
82 49.3 41.0 61.4
119 32.2 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#64
Tetrachloroethene
Concen: 100.406 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion:164 Resp: 767753
Ion Ratio Lower Upper
164 100
166 126.9 99.4 149.0
129 86.5 69.4 104.2
131 83.4 67.5 101.3





#65

Chlorobenzene

Concen: 99.649 ug/l

RT: 11.438 min Scan# 1594

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion:112 Resp: 165514

Ion Ratio Lower Upper

112 100





114 32.6 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 100.470 ug/l

RT: 11.511 min Scan# 1606

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion:131 Resp: 586878


Ion Ratio Lower Upper

131 100

133 96.0 48.4 145.2

119 65.7 32.4 97.2





#67

Ethyl Benzene

Concen: 102.856 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 278825

Ion Ratio Lower Upper

91 100





106 33.4 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#68

m/p-Xylenes

Concen: 206.437 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion:106 Resp: 2243033

Ion Ratio Lower Upper

106 100


91 190.5 152.6 229.0





#69
o-Xylene
Concen: 106.611 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion:106 Resp: 1072335
Ion Ratio Lower Upper
106 100
91 201.8 101.4 304.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#70
Styrene
Concen: 107.344 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09


Tgt Ion:104 Resp: 1782757
Ion Ratio Lower Upper
104 100
78 45.3 36.1 54.1
103 53.9 43.2 64.8





#71
Bromoform
Concen: 106.727 ug/l
RT: 12.127 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100




Tgt Ion:173 Resp: 373208
Ion Ratio Lower Upper
173 100
175 49.4 24.8 74.3
254 0.2 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.340 min Scan# 1906
Delta R.T. -0.006 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Tgt Ion:152 Resp: 400285
Ion Ratio Lower Upper
152 100
115 55.0 27.1 81.2
150 190.0 0.0 347.4





#73
Isopropylbenzene
Concen: 102.825 ug/l
RT: 12.249 min Scan# 1727
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion:105 Resp: 2741698
Ion Ratio Lower Upper
105 100
120 27.7 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 106.818 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09


Tgt Ion: 43 Resp: 532255
Ion Ratio Lower Upper
43 100
70 51.0 41.1 61.7
55 26.8 23.0 34.6
61 27.1 21.4 32.2





#75
1,1,2,2-Tetrachloroethane
Concen: 97.735 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion: 83 Resp: 434720
Ion Ratio Lower Upper
83 100
131 11.5 5.9 17.6
85 64.9 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#76
1,2,3-Trichloropropane
Concen: 92.345 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09


Tgt Ion: 75 Resp: 364622
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 102.545 ug/l
RT: 12.529 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion: 156 Resp: 70174
Ion Ratio Lower Upper
156 100
77 154.5 76.8 230.5
158 98.1 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#78
n-propylbenzene
Concen: 100.958 ug/l
RT: 12.590 min Scan# 1783
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion: 91 Resp: 3165894
Ion Ratio Lower Upper
91 100
120 24.8 12.3 36.8





#79

2-Chlorotoluene

Concen: 99.426 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 1800124

Ion Ratio Lower Upper

91 100





126 37.2 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 102.374 ug/l

RT: 12.731 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion:105 Resp: 2219080

Ion Ratio Lower Upper

105 100


120 51.6 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 101.273 ug/l
RT: 12.298 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 148302
Ion Ratio Lower Upper
75 100
53 84.5 68.6 102.8
89 46.6 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#82
4-Chlorotoluene
Concen: 98.060 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09


Tgt Ion: 91 Resp: 1854165
Ion Ratio Lower Upper
91 100
126 37.4 18.4 55.4





#83
tert-Butylbenzene
Concen: 102.684 ug/l
RT: 12.993 min Scan# 1849
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100



Tgt Ion:119 Resp: 2028198
Ion Ratio Lower Upper
119 100
91 61.9 29.9 89.8
134 26.9 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#84
1,2,4-Trimethylbenzene
Concen: 103.853 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09

Tgt Ion:105 Resp: 2227197
Ion Ratio Lower Upper
105 100
120 47.4 23.8 71.2





#85

sec-Butylbenzene

Concen: 102.530 ug/l

RT: 13.170 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion:105 Resp: 2932949

Ion Ratio Lower Upper

105 100





134 21.5 10.7 32.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#86

p-Isopropyltoluene

Concen: 105.408 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion:119 Resp: 2548220


Ion Ratio Lower Upper

119 100

134 27.0 13.6 40.7

91 22.1 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 100.735 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 1361710

Ion Ratio Lower Upper

146 100

111 37.4 18.7 56.1





148 64.1 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 98.734 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion:146 Resp: 1323618


Ion Ratio Lower Upper

146 100

111 36.0 18.0 54.0

148 63.8 31.9 95.7





#89

n-Butylbenzene

Concen: 104.174 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 2223744

Ion Ratio Lower Upper

91 100

92 54.2 27.0 80.8





134 28.6 14.1 42.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#90

Hexachloroethane

Concen: 98.069 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. -0.000 min

Lab File: VY023388.D




Acq: 07 Oct 2025 12:09


Tgt Ion:117 Resp: 514746

Ion Ratio Lower Upper

117 100

201 86.8 43.8 131.4





#91

1,2-Dichlorobenzene

Concen: 101.146 ug/l

RT: 13.657 min Scan# 1958

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC100

Tgt Ion:146 Resp: 119304

Ion Ratio Lower Upper

146 100

111 38.3 19.0 57.0





148 64.1 32.1 96.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 99.458 ug/l

RT: 14.267 min Scan# 2058

Delta R.T. -0.006 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion: 75 Resp: 70107


Ion Ratio Lower Upper

75 100

155 105.2 50.0 150.0

157 135.9 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 116.031 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC100

Tgt Ion:180 Resp: 806328

Ion Ratio Lower Upper

180 100

182 94.9 47.9 143.6





145 28.7 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene

Concen: 108.557 ug/l

RT: 15.017 min Scan# 2181

Delta R.T. -0.000 min

Lab File: VY023388.D

Acq: 07 Oct 2025 12:09




Tgt Ion:225 Resp: 470072


Ion Ratio Lower Upper

225 100

223 62.9 30.8 92.3


227 64.6 31.9 95.7





#95
Naphthalene
Concen: 124.485 ug/l
RT: 15.139 min Scan# 2201
Delta R.T. -0.000 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100




Tgt Ion:128 Resp: 137829
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 11.1 8.9 13.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#96
1,2,3-Trichlorobenzene
Concen: 115.034 ug/l
RT: 15.322 min Scan# 2231
Delta R.T. -0.006 min
Lab File: VY023388.D
Acq: 07 Oct 2025 12:09



Tgt Ion:180 Resp: 693477
Ion Ratio Lower Upper
180 100
182 95.6 47.9 143.8
145 30.3 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023389.D
 Acq On : 07 Oct 2025 12:32
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Quant Time: Oct 08 04:45:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 04:39:16 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/10/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	588489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	855488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	765357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	402368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	714730	140.236	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 280.480%#			
35) Dibromofluoromethane	7.634	113	778632	149.977	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 299.960%#			
50) Toluene-d8	10.103	98	2912319	148.741	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 297.480%#			
62) 4-Bromofluorobenzene	12.402	95	990964	160.423	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 320.840%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	587619	126.766	ug/l	97
3) Chloromethane	2.068	50	825167	125.233	ug/l	100
4) Vinyl Chloride	2.208	62	1118134	134.387	ug/l	100
5) Bromomethane	2.580	94	915389	134.118	ug/l	98
6) Chloroethane	2.726	64	756912	133.994	ug/l	99
7) Trichlorofluoromethane	3.050	101	1549160	144.948	ug/l	100
8) Diethyl Ether	3.452	74	418615	149.747	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.812	101	786278	137.639	ug/l	99
10) Methyl Iodide	4.007	142	992925	150.409	ug/l	99
11) Tert butyl alcohol	4.866	59	288878	963.279	ug/l	99
12) 1,1-Dichloroethene	3.787	96	781235	142.133	ug/l	97
13) Acrolein	3.653	56	423418	833.323	ug/l	100
14) Allyl chloride	4.385	41	988378	143.628	ug/l	99
15) Acrylonitrile	5.055	53	841857	768.735	ug/l	99
16) Acetone	3.866	43	789668	689.040	ug/l	99
17) Carbon Disulfide	4.104	76	2199339	128.252	ug/l	100
18) Methyl Acetate	4.385	43	505344	173.497	ug/l	99
19) Methyl tert-butyl Ether	5.116	73	2056533	159.802	ug/l	97
20) Methylene Chloride	4.616	84	835005	134.365	ug/l	98
21) trans-1,2-Dichloroethene	5.116	96	871380	142.549	ug/l	98
22) Diisopropyl ether	6.019	45	2218430	145.375	ug/l	99
23) Vinyl Acetate	5.958	43	6377047	760.246	ug/l	99
24) 1,1-Dichloroethane	5.915	63	1398477	139.024	ug/l	98
25) 2-Butanone	6.890	43	1116343	788.091	ug/l	97
26) 2,2-Dichloropropane	6.884	77	1281633	143.391	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	1043574	150.505	ug/l	98
28) Bromochloromethane	7.244	49	482210	123.567	ug/l	99
29) Tetrahydrofuran	7.262	42	687719	817.972	ug/l	100
30) Chloroform	7.421	83	1542715	142.221	ug/l	99
31) Cyclohexane	7.701	56	1192778	134.268	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	1405131	144.305	ug/l	100
36) 1,1-Dichloropropene	7.835	75	1069665	146.031	ug/l	100
37) Ethyl Acetate	6.982	43	460027	158.853	ug/l	98
38) Carbon Tetrachloride	7.817	117	1290353	151.646	ug/l	99
39) Methylcyclohexane	9.109	83	1496612	159.580	ug/l	99
40) Benzene	8.079	78	3366812	147.442	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023389.D
 Acq On : 07 Oct 2025 12:32
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:45:15 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 04:39:16 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	246974	164.212	ug/l	98
42) 1,2-Dichloroethane	8.158	62	866250	149.063	ug/l	100
43) Isopropyl Acetate	8.195	43	946087	167.955	ug/l	100
44) Trichloroethene	8.866	130	966011	150.740	ug/l	99
45) 1,2-Dichloropropane	9.140	63	752286	146.403	ug/l	99
46) Dibromomethane	9.231	93	483702	153.895	ug/l	99
47) Bromodichloromethane	9.420	83	1202247	151.965	ug/l	98
48) Methyl methacrylate	9.219	41	463614	176.368	ug/l	100
49) 1,4-Dioxane	9.225	88	123991	3809.726	ug/l	97
51) 4-Methyl-2-Pentanone	9.993	43	2516107	872.045	ug/l	100
52) Toluene	10.170	92	2272574	156.786	ug/l	98
53) t-1,3-Dichloropropene	10.390	75	1123644	165.367	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	1296713	159.654	ug/l	99
55) 1,1,2-Trichloroethane	10.573	97	652242	156.020	ug/l	96
56) Ethyl methacrylate	10.438	69	886127	184.048	ug/l	100
57) 1,3-Dichloropropane	10.713	76	1049593	155.122	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	2040136	968.559	ug/l	100
59) 2-Hexanone	10.755	43	1776032	889.038	ug/l	99
60) Dibromochloromethane	10.908	129	935577	164.084	ug/l	99
61) 1,2-Dibromoethane	11.012	107	639895	162.699	ug/l	99
64) Tetrachloroethene	10.646	164	1046931	136.739	ug/l	98
65) Chlorobenzene	11.438	112	2563176	154.117	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.511	131	923367	157.870	ug/l	99
67) Ethyl Benzene	11.518	91	4341801	159.957	ug/l	97
68) m/p-Xylenes	11.627	106	3502137	321.899	ug/l	99
69) o-Xylene	11.950	106	1694987	168.295	ug/l	98
70) Styrene	11.969	104	2830398	170.203	ug/l	99
71) Bromoform	12.127	173	602644	172.115	ug/l #	99
73) Isopropylbenzene	12.249	105	4290662	160.084	ug/l	100
74) N-amyl acetate	12.066	43	890882	177.866	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	737773	165.010	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	565754m	142.542	ug/l	
77) Bromobenzene	12.530	156	1111318	161.555	ug/l	100
78) n-propylbenzene	12.591	91	4932015	156.465	ug/l	99
79) 2-Chlorotoluene	12.676	91	2822049	155.063	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	3496249	160.460	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.298	75	245190	166.570	ug/l	98
82) 4-Chlorotoluene	12.773	91	2902371	152.701	ug/l	99
83) tert-Butylbenzene	12.993	119	3220384	162.197	ug/l	97
84) 1,2,4-Trimethylbenzene	13.042	105	3503639	162.527	ug/l	99
85) sec-Butylbenzene	13.170	105	4560715	158.609	ug/l	99
86) p-Isopropyltoluene	13.286	119	4024712	165.622	ug/l	100
87) 1,3-Dichlorobenzene	13.286	146	2165278	159.350	ug/l	100
88) 1,4-Dichlorobenzene	13.365	146	2090133	155.104	ug/l	100
89) n-Butylbenzene	13.615	91	3505715	163.379	ug/l	99
90) Hexachloroethane	13.877	117	799264	151.487	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	1862691	157.101	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.267	75	116273	164.099	ug/l	96
93) 1,2,4-Trichlorobenzene	14.913	180	1272621	182.183	ug/l	100
94) Hexachlorobutadiene	15.017	225	711748	163.518	ug/l	98
95) Naphthalene	15.139	128	2277549	204.639	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	1110193	183.206	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023389.D
Acq On : 07 Oct 2025 12:32
Operator : SY/MD
Sample : VSTDICC150
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 04:45:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 04:39:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion:168 Resp: 588489
Ion Ratio Lower Upper
168 100
99 48.3 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 126.766 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 85 Resp: 587619
Ion Ratio Lower Upper
85 100
87 33.7 16.0 47.9









#3
Chloromethane
Concen: 125.233 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion: 50 Resp: 82516
Ion Ratio Lower Upper
50 100
52 32.9 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#4
Vinyl Chloride
Concen: 134.387 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 62 Resp: 1118134
Ion Ratio Lower Upper
62 100
64 32.5 25.8 38.8









#5
Bromomethane
Concen: 134.118 ug/l
RT: 2.580 min Scan# 141
Delta R.T. -0.012 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion: 94 Resp: 915389
Ion Ratio Lower Upper
94 100
96 94.8 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#6
Chloroethane
Concen: 133.994 ug/l
RT: 2.726 min Scan# 165
Delta R.T. -0.006 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 64 Resp: 756912
Ion Ratio Lower Upper
64 100
66 32.0 26.2 39.4









#7
Trichlorofluoromethane
Concen: 144.948 ug/l
RT: 3.050 min Scan# 219
Delta R.T. -0.006 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 154916
Ion Ratio Lower Upper
101 100
103 64.4 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#8
Diethyl Ether
Concen: 149.747 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 74 Resp: 418615
Ion Ratio Lower Upper
74 100
45 79.7 39.9 119.7





#9

1,1,2-Trichlorotrifluoroethane

Concen: 137.639 ug/l

RT: 3.812 min Scan# 343

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion:101 Resp: 786278

Ion Ratio Lower Upper

101 100

85 43.5 34.6 51.8





151 85.2 67.0 100.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#10

Methyl Iodide

Concen: 150.409 ug/l

RT: 4.007 min Scan# 375

Delta R.T. 0.006 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:142 Resp: 992925


Ion Ratio Lower Upper

142 100

127 42.8 34.8 52.2

141 14.1 11.6 17.4





#11

Tert butyl alcohol

Concen: 963.279 ug/l

RT: 4.866 min Scan# 515

Delta R.T. 0.006 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion: 59 Resp: 288878

Ion Ratio Lower Upper

59 100





57 10.1 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 142.133 ug/l

RT: 3.787 min Scan# 339

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 96 Resp: 781235


Ion Ratio Lower Upper

96 100

61 139.3 106.7 160.1

98 62.9 50.2 75.4





#13

Acrolein

Concen: 833.323 ug/l

RT: 3.653 min Scan# 316

Delta R.T. 0.006 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 56 Resp: 423418

Ion Ratio Lower Upper

56 100





55 69.7 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#14

Allyl chloride

Concen: 143.628 ug/l

RT: 4.385 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 41 Resp: 988378


Ion Ratio Lower Upper

41 100

39 75.6 59.9 89.9

76 47.2 38.0 57.0





#15

Acrylonitrile

Concen: 768.735 ug/l

RT: 5.055 min Scan# 547

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 53 Resp: 84185

Ion Ratio Lower Upper

53 100

52 80.9 64.6 97.0





51 36.0 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 689.040 ug/l

RT: 3.866 min Scan# 352

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion: 43 Resp: 789668

Ion Ratio Lower Upper

43 100


58 34.0 27.7 41.5





#17
Carbon Disulfide
Concen: 128.252 ug/l
RT: 4.104 min Scan# 391
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC150



Tgt Ion: 76 Resp: 2199339
Ion Ratio Lower Upper
76 100
78 9.2 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#18
Methyl Acetate
Concen: 173.497 ug/l
RT: 4.385 min Scan# 437
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 43 Resp: 505344
Ion Ratio Lower Upper
43 100
74 29.7 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 159.802 ug/l

RT: 5.116 min Scan# 5

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 73 Resp: 205653

Ion Ratio Lower Upper

73 100





57 20.3 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 134.365 ug/l

RT: 4.616 min Scan# 475

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Tgt Ion: 84 Resp: 835005




Ion Ratio Lower Upper


84 100

49 102.7 80.2 120.4

51 31.9 25.9 38.9


86 65.3 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 142.549 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 871380
Ion Ratio Lower Upper
96 100
61 117.5 92.7 139.1
98 64.1 49.7 74.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#22
Diisopropyl ether
Concen: 145.375 ug/l
RT: 6.019 min Scan# 705
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 45 Resp: 2218430
Ion Ratio Lower Upper
45 100
43 57.6 45.6 68.4
87 34.7 28.1 42.1
59 12.5 10.2 15.2





#23

Vinyl Acetate

Concen: 760.246 ug/l

RT: 5.958 min Scan# 695

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 637704

Ion Ratio Lower Upper

43

100

86

13.6

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 139.024 ug/l

RT: 5.915 min Scan# 688

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Tgt Ion: 63 Resp: 1398477

Ion Ratio Lower Upper

63

100

98




8.0


100

5.0

2.8

8.4





#25

2-Butanone

Concen: 788.091 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 111634

Ion Ratio Lower Upper

43 100





72 32.3 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 143.391 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion: 77 Resp: 1281633

Ion Ratio Lower Upper

77 100

97 24.6 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 150.505 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion: 96 Resp: 1043574

Ion Ratio Lower Upper

96 100

61 125.7 0.0 258.2





98 65.0 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane

Concen: 123.567 ug/l

RT: 7.244 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 49 Resp: 482210


Ion Ratio Lower Upper

49 100

129 2.5 0.0 5.6

130 101.2 81.9 122.9





#29

Tetrahydrofuran

Concen: 817.972 ug/l

RT: 7.262 min Scan# 909

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 687715

Ion Ratio Lower Upper

42 100

72 53.9 43.0 64.6





71 50.4 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 142.221 ug/l

RT: 7.421 min Scan# 935

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion: 83 Resp: 1542715

Ion Ratio Lower Upper

83 100

85 65.0 52.5 78.7





#31

Cyclohexane

Concen: 134.268 ug/l

RT: 7.701 min Scan# 981

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 56 Resp: 1192778

Ion Ratio Lower Upper

56 100

69 34.2 30.0 45.0





84 101.7 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#32

1,1,1-Trichloroethane

Concen: 144.305 ug/l

RT: 7.616 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 97 Resp: 1405131


Ion Ratio Lower Upper

97 100

99 64.3 51.4 77.2

61 39.5 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 140.236 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion: 65 Resp: 714730

Ion Ratio Lower Upper

65 100





67 52.8 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:114 Resp: 855488


Ion Ratio Lower Upper

114 100

63 16.4 0.0 34.2

88 14.6 0.0 28.4





#35

Dibromofluoromethane

Concen: 149.977 ug/l

RT: 7.634 min Scan# 91

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion:113 Resp: 778632

Ion Ratio Lower Upper

113 100

111 102.2 81.8 122.6





192 22.7 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 146.031 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 75 Resp: 1069665


Ion Ratio Lower Upper

75 100

110 39.9 19.9 59.7

77 31.4 25.0 37.6





#37

Ethyl Acetate

Concen: 158.853 ug/l

RT: 6.982 min Scan# 80

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 46002

Ion Ratio Lower Upper

43 100

61 15.9 11.8 17.8





70 12.7 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 151.646 ug/l

RT: 7.817 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:117 Resp: 1290353


Ion Ratio Lower Upper

117 100

119 95.8 76.8 115.2

121 31.3 24.5 36.7





#39

Methylcyclohexane

Concen: 159.580 ug/l

RT: 9.109 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 83 Resp: 1496612

Ion Ratio Lower Upper

83 100

55 65.2 51.4 77.2





98 51.2 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 147.442 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion: 78 Resp: 3366812

Ion Ratio Lower Upper

78 100


77 23.7 18.3 27.5





#41
Methacrylonitrile
Concen: 164.212 ug/l
RT: 7.220 min Scan# 902
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150




Tgt Ion: 41 Resp: 246974
Ion Ratio Lower Upper
41 100
39 61.7 52.2 78.4
67 93.3 74.9 112.3
52 35.5 29.2 43.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#42
1,2-Dichloroethane
Concen: 149.063 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32



Tgt Ion: 62 Resp: 866250
Ion Ratio Lower Upper
62 100
98 11.0 0.0 21.6





#43

Isopropyl Acetate

Concen: 167.955 ug/l

RT: 8.195 min Scan# 1062

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 94608

Ion Ratio Lower Upper

43 100

61 22.6 18.1 27.1





87 16.7 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 150.740 ug/l

RT: 8.866 min Scan# 1172

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion:130 Resp: 966011

Ion Ratio Lower Upper

130 100

95 91.3 0.0 180.6





#45

1,2-Dichloropropane

Concen: 146.403 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 63 Resp: 752280

Ion Ratio Lower Upper

63 100





65 31.7 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 153.895 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 93 Resp: 483702


Ion Ratio Lower Upper

93 100

95 83.7 65.4 98.0


174 116.7 92.9 139.3





#47
Bromodichloromethane
Concen: 151.965 ug/l
RT: 9.420 min Scan# 1263
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion: 83 Resp: 120224
Ion Ratio Lower Upper
83 100
85 65.6 50.8 76.2
127 9.0 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#48
Methyl methacrylate
Concen: 176.368 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 41 Resp: 463614
Ion Ratio Lower Upper
41 100
69 100.4 81.0 121.4
39 60.3 48.2 72.4





#49

1,4-Dioxane

Concen: 3809.726 ug/l

RT: 9.225 min Scan# 12399

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 88 Resp: 12399

Ion Ratio Lower Upper

88 100

43 27.5 21.8 32.8





58 62.1 46.8 70.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 148.741 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion: 98 Resp: 2912319

Ion Ratio Lower Upper

98 100


100 66.2 53.0 79.4





#51
4-Methyl-2-Pentanone
Concen: 872.045 ug/l
RT: 9.993 min Scan# 1357
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150




Tgt Ion: 43 Resp: 251610
Ion Ratio Lower Upper
43 100
58 44.1 35.3 52.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#52
Toluene
Concen: 156.786 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Tgt Ion: 92 Resp: 2272574
Ion Ratio Lower Upper
92 100
91 168.3 136.4 204.6





#53
t-1,3-Dichloropropene
Concen: 165.367 ug/l
RT: 10.390 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion: 75 Resp: 1123644
Ion Ratio Lower Upper
75 100
77 31.6 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#54
cis-1,3-Dichloropropene
Concen: 159.654 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 75 Resp: 1296713
Ion Ratio Lower Upper
75 100
77 31.4 25.4 38.0
39 41.7 32.8 49.2





#55

1,1,2-Trichloroethane

Concen: 156.020 ug/l

RT: 10.573 min Scan# 1452

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 97 Resp: 65224

Ion Ratio Lower Upper

97 100

83 83.2 61.8 92.6

85 53.7 41.3 61.9





99 62.8 48.7 73.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#56

Ethyl methacrylate

Concen: 184.048 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion: 69 Resp: 886127


Ion Ratio Lower Upper

69 100

41 56.0 44.6 67.0

39 33.6 27.0 40.4





#57

1,3-Dichloropropane

Concen: 155.122 ug/l

RT: 10.713 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 104959

Ion Ratio Lower Upper

76 100





78 32.6 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 968.559 ug/l

RT: 9.707 min Scan# 1310

Delta R.T. 0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion: 63 Resp: 2040136

Ion Ratio Lower Upper

63 100

106 33.2 26.6 39.8





#59

2-Hexanone

Concen: 889.038 ug/l

RT: 10.755 min Scan# 1482

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 177603

Ion Ratio Lower Upper

43 100





58 61.4 30.5 91.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#60

Dibromochloromethane

Concen: 164.084 ug/l

RT: 10.908 min Scan# 1507

Delta R.T. -0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion:129 Resp: 935577

Ion Ratio Lower Upper

129 100


127 76.4 38.6 115.7





#61
1,2-Dibromoethane
Concen: 162.699 ug/l
RT: 11.012 min Scan# 1524
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:107 Resp: 639895
Ion Ratio Lower Upper
107 100
109 94.4 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 160.423 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32


Tgt Ion: 95 Resp: 990964
Ion Ratio Lower Upper
95 100
174 95.8 0.0 192.8
176 92.8 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:117 Resp: 76535
Ion Ratio Lower Upper
117 100
82 51.3 41.0 61.4
119 33.3 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#64
Tetrachloroethene
Concen: 136.739 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion:164 Resp: 1046931
Ion Ratio Lower Upper
164 100
166 127.5 99.4 149.0
129 88.4 69.4 104.2
131 84.9 67.5 101.3





#65

Chlorobenzene

Concen: 154.117 ug/l

RT: 11.438 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion:112 Resp: 2563170

Ion Ratio Lower Upper

112 100





114 32.4 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 157.870 ug/l

RT: 11.511 min Scan# 1606

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:131 Resp: 923367


Ion Ratio Lower Upper

131 100

133 95.4 48.4 145.2

119 65.2 32.4 97.2









#67
Ethyl Benzene
Concen: 159.957 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion: 91 Resp: 434180
Ion Ratio Lower Upper
91 100
106 33.8 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#68
m/p-Xylenes
Concen: 321.899 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion:106 Resp: 3502137
Ion Ratio Lower Upper
106 100
91 189.2 152.6 229.0





#69

o-Xylene

Concen: 168.295 ug/l

RT: 11.950 min Scan# 1678

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion:106 Resp: 169498

Ion Ratio Lower Upper

106 100





91 200.0 101.4 304.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#70

Styrene

Concen: 170.203 ug/l

RT: 11.969 min Scan# 1681

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:104 Resp: 2830398


Ion Ratio Lower Upper

104 100

78 45.5 36.1 54.1


103 54.3 43.2 64.8





#71
Bromoform
Concen: 172.115 ug/l
RT: 12.127 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:173 Resp: 602644
Ion Ratio Lower Upper
173 100
175 48.9 24.8 74.3
254 0.2 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32


Tgt Ion:152 Resp: 402368
Ion Ratio Lower Upper
152 100
115 54.1 27.1 81.2
150 210.4 0.0 347.4





#73
Isopropylbenzene
Concen: 160.084 ug/l
RT: 12.249 min Scan# 1727
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:105 Resp: 4290662
Ion Ratio Lower Upper
105 100
120 27.9 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 177.866 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 43 Resp: 890882
Ion Ratio Lower Upper
43 100
70 50.6 41.1 61.7
55 26.4 23.0 34.6
61 27.2 21.4 32.2









#75
1,1,2-Tetrachloroethane
Concen: 165.010 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion: 83 Resp: 73777
Ion Ratio Lower Upper
83 100
131 11.4 5.9 17.6
85 64.4 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#76
1,2,3-Trichloropropane
Concen: 142.542 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 75 Resp: 565754
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 161.555 ug/l
RT: 12.530 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Tgt Ion:156 Resp: 1111318
Ion Ratio Lower Upper
156 100
77 153.9 76.8 230.5
158 97.4 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#78
n-propylbenzene
Concen: 156.465 ug/l
RT: 12.591 min Scan# 1783
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 91 Resp: 4932015
Ion Ratio Lower Upper
91 100
120 25.1 12.3 36.8





#79

2-Chlorotoluene

Concen: 155.063 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC150

Tgt Ion: 91 Resp: 2822049

Ion Ratio Lower Upper

91 100





126 37.1 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 160.460 ug/l

RT: 12.731 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VY023389.D




Acq: 07 Oct 2025 12:32


Tgt Ion:105 Resp: 3496249

Ion Ratio Lower Upper

105 100


120 51.4 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 166.570 ug/l
RT: 12.298 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion: 75 Resp: 245190
Ion Ratio Lower Upper
75 100
53 83.6 68.6 102.8
89 46.8 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#82
4-Chlorotoluene
Concen: 152.701 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32


Tgt Ion: 91 Resp: 2902371
Ion Ratio Lower Upper
91 100
126 37.3 18.4 55.4





#83
tert-Butylbenzene
Concen: 162.197 ug/l
RT: 12.993 min Scan# 1849
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:119 Resp: 3220384
Ion Ratio Lower Upper
119 100
91 61.6 29.9 89.8
134 27.4 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#84
1,2,4-Trimethylbenzene
Concen: 162.527 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. 0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32


Tgt Ion:105 Resp: 3503639
Ion Ratio Lower Upper
105 100
120 47.9 23.8 71.2





#85
sec-Butylbenzene
Concen: 158.609 ug/l
RT: 13.170 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:105 Resp: 4560712
Ion Ratio Lower Upper
105 100
134 21.6 10.7 32.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#86
p-Isopropyltoluene
Concen: 165.622 ug/l
RT: 13.286 min Scan# 1897
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion:119 Resp: 4024712
Ion Ratio Lower Upper
119 100
134 27.4 13.6 40.7
91 22.1 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 159.350 ug/l

RT: 13.286 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion:146 Resp: 2165278

Ion Ratio Lower Upper

146 100

111 37.4 18.7 56.1





148 64.0 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 155.104 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:146 Resp: 2090133


Ion Ratio Lower Upper

146 100

111 35.8 18.0 54.0

148 63.6 31.9 95.7





#89

n-Butylbenzene

Concen: 163.379 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. 0.000 min

Lab File: VY023389.D


Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150



Tgt Ion: 91 Resp: 350571

Ion Ratio Lower Upper

91 100

92 54.6 27.0 80.8




134 28.6 14.1 42.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#90

Hexachloroethane


Concen: 151.487 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32






Tgt Ion:117 Resp: 799264

Ion Ratio Lower Upper

117 100


201 86.8 43.8 131.4





#91
1,2-Dichlorobenzene
Concen: 157.101 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150



Tgt Ion:146 Resp: 186269
Ion Ratio Lower Upper
146 100
111 38.7 19.0 57.0
148 64.5 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 164.099 ug/l
RT: 14.267 min Scan# 2058
Delta R.T. -0.006 min
Lab File: VY023389.D
Acq: 07 Oct 2025 12:32

Tgt Ion: 75 Resp: 116273
Ion Ratio Lower Upper
75 100
155 103.6 50.0 150.0
157 133.4 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 182.183 ug/l

RT: 14.913 min Scan# 2165

Delta R.T. -0.006 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion:180 Resp: 127262

Ion Ratio Lower Upper

180 100

182 95.6 47.9 143.6





145 28.9 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene

Concen: 163.518 ug/l

RT: 15.017 min Scan# 2181

Delta R.T. -0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:225 Resp: 711748


Ion Ratio Lower Upper

225 100

223 62.9 30.8 92.3

227 65.0 31.9 95.7





#95

Naphthalene

Concen: 204.639 ug/l

RT: 15.139 min Scan# 2201

Delta R.T. 0.000 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC150

Tgt Ion:128 Resp: 2277549

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6





129 11.1 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#96

1,2,3-Trichlorobenzene

Concen: 183.206 ug/l

RT: 15.322 min Scan# 2231

Delta R.T. -0.006 min

Lab File: VY023389.D

Acq: 07 Oct 2025 12:32




Tgt Ion:180 Resp: 1110193

Ion Ratio Lower Upper

180 100

182 95.6 47.9 143.8

145 30.5 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023391.D
 Acq On : 07 Oct 2025 14:12
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY100725

Quant Time: Oct 08 05:15:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/10/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	574332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	861639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	771372	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	405636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	237493	49.450	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.900%		
35) Dibromofluoromethane	7.634	113	260038	49.116	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.240%		
50) Toluene-d8	10.103	98	970614	49.228	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	98.460%		
62) 4-Bromofluorobenzene	12.402	95	324700	49.881	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	99.760%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	185683	44.916	ug/l	99
3) Chloromethane	2.068	50	295766	52.519	ug/l	98
4) Vinyl Chloride	2.208	62	362714	49.449	ug/l	100
5) Bromomethane	2.598	94	300735	48.856	ug/l	98
6) Chloroethane	2.739	64	256831	51.367	ug/l	99
7) Trichlorofluoromethane	3.056	101	494789	48.582	ug/l	99
8) Diethyl Ether	3.458	74	135857	51.050	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.812	101	257524	48.728	ug/l	100
10) Methyl Iodide	4.007	142	305209	51.875	ug/l	98
11) Tert butyl alcohol	4.860	59	82562	231.079	ug/l	100
12) 1,1-Dichloroethene	3.787	96	254382	49.653	ug/l	97
13) Acrolein	3.653	56	124288	246.184	ug/l	99
14) Allyl chloride	4.385	41	316213	50.645	ug/l	99
15) Acrylonitrile	5.055	53	256219	251.575	ug/l	99
16) Acetone	3.866	43	243433	212.910	ug/l	98
17) Carbon Disulfide	4.104	76	733457	50.021	ug/l	100
18) Methyl Acetate	4.385	43	151281	50.439	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	643111	52.230	ug/l	97
20) Methylene Chloride	4.616	84	324604	52.503	ug/l	97
21) trans-1,2-Dichloroethene	5.116	96	290384	51.482	ug/l	98
22) Diisopropyl ether	6.019	45	709381	51.364	ug/l	99
23) Vinyl Acetate	5.958	43	1950969	260.981	ug/l	99
24) 1,1-Dichloroethane	5.915	63	464207	50.742	ug/l	98
25) 2-Butanone	6.890	43	325351	233.083	ug/l	99
26) 2,2-Dichloropropane	6.884	77	416585	49.626	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	341964	52.485	ug/l	97
28) Bromochloromethane	7.244	49	155146	45.427	ug/l	100
29) Tetrahydrofuran	7.262	42	200513	253.317	ug/l	99
30) Chloroform	7.421	83	513997	51.137	ug/l	98
31) Cyclohexane	7.701	56	378819	47.416	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	456112	49.934	ug/l	100
36) 1,1-Dichloropropene	7.835	75	348786	49.761	ug/l	99
37) Ethyl Acetate	6.982	43	138119	49.740	ug/l	98
38) Carbon Tetrachloride	7.817	117	413653	48.671	ug/l	99
39) Methylcyclohexane	9.109	83	456031	49.970	ug/l	99
40) Benzene	8.079	78	1107790	49.906	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023391.D
 Acq On : 07 Oct 2025 14:12
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY100725

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 10/10/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 05:15:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	72100	48.008	ug/l	96
42) 1,2-Dichloroethane	8.158	62	282332	49.318	ug/l	99
43) Isopropyl Acetate	8.195	43	279887	50.548	ug/l	100
44) Trichloroethene	8.866	130	319131	49.120	ug/l	98
45) 1,2-Dichloropropane	9.140	63	246887	50.355	ug/l	100
46) Dibromomethane	9.231	93	157758	50.885	ug/l	98
47) Bromodichloromethane	9.420	83	393918	50.450	ug/l	99
48) Methyl methacrylate	9.219	41	128606	49.440	ug/l	95
49) 1,4-Dioxane	9.231	88	37356	1039.252	ug/l	97
51) 4-Methyl-2-Pentanone	10.000	43	724992	253.421	ug/l	100
52) Toluene	10.170	92	743055	51.556	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	347336	51.160	ug/l	99
54) cis-1,3-Dichloropropene	9.853	75	413860	51.649	ug/l	99
55) 1,1,2-Trichloroethane	10.573	97	210245	50.430	ug/l	97
56) Ethyl methacrylate	10.438	69	256731	53.385	ug/l	99
57) 1,3-Dichloropropane	10.713	76	337673	51.113	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	586266	272.133	ug/l	100
59) 2-Hexanone	10.762	43	501877	245.480	ug/l	99
60) Dibromochloromethane	10.908	129	297946	50.974	ug/l	99
61) 1,2-Dibromoethane	11.018	107	201415	50.658	ug/l	100
64) Tetrachloroethene	10.646	164	382998	48.087	ug/l	99
65) Chlorobenzene	11.438	112	828466	49.381	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.518	131	296157	49.537	ug/l	99
67) Ethyl Benzene	11.518	91	1380691	50.841	ug/l	100
68) m/p-Xylenes	11.627	106	1108386	101.757	ug/l	100
69) o-Xylene	11.950	106	526954	51.577	ug/l	100
70) Styrene	11.969	104	875709	52.122	ug/l	100
71) Bromoform	12.133	173	185682	50.070	ug/l #	98
73) Isopropylbenzene	12.255	105	1345109	50.318	ug/l	100
74) N-amyl acetate	12.066	43	245291	51.087	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.505	83	217245	49.688	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	153438m	42.874	ug/l	
77) Bromobenzene	12.530	156	352420	49.590	ug/l	99
78) n-propylbenzene	12.591	91	1568180	50.694	ug/l	100
79) 2-Chlorotoluene	12.676	91	906610	50.047	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	1106664	51.231	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	67959	46.566	ug/l	97
82) 4-Chlorotoluene	12.773	91	941314	50.455	ug/l	100
83) tert-Butylbenzene	12.993	119	1037510	52.019	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	1110672	51.520	ug/l	99
85) sec-Butylbenzene	13.176	105	1454774	50.848	ug/l	100
86) p-Isopropyltoluene	13.292	119	1263377	51.864	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	686854	49.632	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	671644	49.140	ug/l	100
89) n-Butylbenzene	13.615	91	1100086	51.829	ug/l	99
90) Hexachloroethane	13.877	117	257335	49.116	ug/l	98
91) 1,2-Dichlorobenzene	13.657	146	602989	49.698	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	34928	48.687	ug/l	99
93) 1,2,4-Trichlorobenzene	14.919	180	386806	51.138	ug/l	100
94) Hexachlorobutadiene	15.023	225	236040	50.061	ug/l	99
95) Naphthalene	15.139	128	628190	52.154	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	335267	51.172	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
 Data File : VY023391.D
 Acq On : 07 Oct 2025 14:12
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :


ICVVY100725


Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
 Supervised By :Mahesh Dadoda 10/10/2025


Quant Time: Oct 08 05:15:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:168 Resp: 57433
Ion Ratio Lower Upper
168 100
99 50.4 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#2
Dichlorodifluoromethane
Concen: 44.916 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion: 85 Resp: 185683
Ion Ratio Lower Upper
85 100
87 32.6 16.0 47.9





#3
Chloromethane
Concen: 52.519 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725




Tgt Ion: 50 Resp: 295760
Ion Ratio Lower Upper
50 100
52 33.7 26.2 39.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#4
Vinyl Chloride
Concen: 49.449 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12



Tgt Ion: 62 Resp: 362714
Ion Ratio Lower Upper
62 100
64 32.3 25.8 38.8









#5
Bromomethane
Concen: 48.856 ug/l
RT: 2.598 min Scan# 143
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Tgt Ion: 94 Resp: 300735
Ion Ratio Lower Upper
94 100
96 91.5 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#6
Chloroethane
Concen: 51.367 ug/l
RT: 2.739 min Scan# 167
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion: 64 Resp: 256831
Ion Ratio Lower Upper
64 100
66 32.1 26.2 39.4





#7
Trichlorofluoromethane
Concen: 48.582 ug/l
RT: 3.056 min Scan# 219
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 101 Resp: 494789
Ion Ratio Lower Upper
101 100
103 65.3 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#8
Diethyl Ether
Concen: 51.050 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 74 Resp: 135857
Ion Ratio Lower Upper
74 100
45 80.1 39.9 119.7





#9

1,1,2-Trichlorotrifluoroethane

Concen: 48.728 ug/l

RT: 3.812 min Scan# 343

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :





ICVY100725

Tgt Ion:	101	Resp:	257524
Ion	Ratio	Lower	Upper
101	100		
85	43.2	34.6	51.8
151	83.2	67.0	100.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#10

Methyl Iodide

Concen: 51.875 ug/l




RT: 4.007 min Scan# 375


Delta R.T. 0.006 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Tgt Ion:	142	Resp:	305209
Ion	Ratio	Lower	Upper
142	100		
127	41.8	34.8	52.2
141	14.1	11.6	17.4





#11

Tert butyl alcohol

Concen: 231.079 ug/l

RT: 4.860 min Scan# 515

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 59 Resp: 82562

Ion Ratio Lower Upper

59 100





57 10.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#12

1,1-Dichloroethene

Concen: 49.653 ug/l

RT: 3.787 min Scan# 339

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion: 96 Resp: 254382


Ion Ratio Lower Upper

96 100

61 139.1 106.7 160.1


98 62.3 50.2 75.4





#13
Acrolein
Concen: 246.184 ug/l
RT: 3.653 min Scan# 316
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 56 Resp: 124288
Ion Ratio Lower Upper
56 100
55 71.1 56.1 84.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#14
Allyl chloride
Concen: 50.645 ug/l
RT: 4.385 min Scan# 437
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 41 Resp: 316213
Ion Ratio Lower Upper
41 100
39 76.2 59.9 89.9
76 48.1 38.0 57.0





#15

Acrylonitrile

Concen: 251.575 ug/l

RT: 5.055 min Scan# 547

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 53 Resp: 256219

Ion Ratio Lower Upper

53 100

52 81.7 64.6 97.0





51 36.4 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#16

Acetone

Concen: 212.910 ug/l

RT: 3.866 min Scan# 352

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion: 43 Resp: 243433

Ion Ratio Lower Upper

43 100

58 35.5 27.7 41.5





#17

Carbon Disulfide

Concen: 50.021 ug/l

RT: 4.104 min Scan# 391

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12

Tgt Ion: 76 Resp: 73345

Ion Ratio Lower Upper

76 100

78 9.1 7.3 10.9



Instrument :

MSVOA_Y

ClientSampleId :


ICVVY100725

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#18

Methyl Acetate

Concen: 50.439 ug/l

RT: 4.385 min Scan# 437

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion: 43 Resp: 151281

Ion Ratio Lower Upper

43 100

74 30.5 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 52.230 ug/l

RT: 5.116 min Scan# 5

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 73 Resp: 64311

Ion Ratio Lower Upper

73 100





57 20.3 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#20

Methylene Chloride

Concen: 52.503 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Tgt Ion: 84 Resp: 324604




Ion Ratio Lower Upper


84 100

49 103.5 80.2 120.4

51 34.1 25.9 38.9

86 66.3 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 51.482 ug/l

RT: 5.116 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023391.D


Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725






Tgt Ion:	96	Resp:	290384
Ion	Ratio	Lower	Upper
96	100		
61	117.9	92.7	139.1
98	63.8	49.7	74.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#22

Diisopropyl ether


Concen: 51.364 ug/l

RT: 6.019 min Scan# 705



Delta R.T. -0.000 min


Lab File: VY023391.D

Acq: 07 Oct 2025 14:12



Tgt Ion:	45	Resp:	709381
Ion	Ratio	Lower	Upper
45	100		
43	57.4	45.6	68.4
87	34.6	28.1	42.1
59	13.3	10.2	15.2





#23

Vinyl Acetate

Concen: 260.981 ug/l

RT: 5.958 min Scan# 695

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 43 Resp: 195096

Ion Ratio Lower Upper

43

100

86

13.7

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#24

1,1-Dichloroethane

Concen: 50.742 ug/l

RT: 5.915 min Scan# 688

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Tgt Ion: 63 Resp: 464207

Ion Ratio Lower Upper

63

100

98

8.1

4.3




13.1


100

5.0

2.8

8.4





#25

2-Butanone

Concen: 233.083 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 43 Resp: 32535

Ion Ratio Lower Upper

43 100





72 33.8 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#26

2,2-Dichloropropane

Concen: 49.626 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion: 77 Resp: 416585

Ion Ratio Lower Upper

77 100

97 24.6 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 52.485 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023391.D


Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725



Tgt Ion: 96 Resp: 341964

Ion Ratio Lower Upper

96 100

61 125.5 0.0 258.2




98 64.3 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#28

Bromochloromethane


Concen: 45.427 ug/l

RT: 7.244 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12





Tgt Ion: 49 Resp: 155146


Ion Ratio Lower Upper

49 100

129 2.8 0.0 5.6

130 102.6 81.9 122.9





#29

Tetrahydrofuran

Concen: 253.317 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 42 Resp: 20051

Ion Ratio Lower Upper

42 100

72 54.6 43.0 64.6





71 51.1 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#30

Chloroform

Concen: 51.137 ug/l

RT: 7.421 min Scan# 935

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion: 83 Resp: 513997

Ion Ratio Lower Upper

83 100

85 67.5 52.5 78.7





#31

Cyclohexane

Concen: 47.416 ug/l

RT: 7.701 min Scan# 981

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 56 Resp: 378819

Ion Ratio Lower Upper

56 100

69 37.1 30.0 45.0





84 99.4 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#32

1,1,1-Trichloroethane

Concen: 49.934 ug/l

RT: 7.616 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion: 97 Resp: 456112


Ion Ratio Lower Upper

97 100

99 64.7 51.4 77.2

61 39.8 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 49.450 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 65 Resp: 23749

Ion Ratio Lower Upper

65 100





67 52.8 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion:114 Resp: 861639


Ion Ratio Lower Upper

114 100

63 17.1 0.0 34.2

88 14.9 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.116 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion:113 Resp: 260038

Ion Ratio Lower Upper

113 100

111 103.5 81.8 122.6





192 22.5 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#36

1,1-Dichloropropene

Concen: 49.761 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion: 75 Resp: 348786


Ion Ratio Lower Upper

75 100

110 38.9 19.9 59.7

77 31.0 25.0 37.6





#37

Ethyl Acetate

Concen: 49.740 ug/l

RT: 6.982 min Scan# 80

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 43 Resp: 138119

Ion Ratio Lower Upper

43 100

61 15.8 11.8 17.8





70 12.7 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#38

Carbon Tetrachloride

Concen: 48.671 ug/l

RT: 7.817 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion:117 Resp: 413653


Ion Ratio Lower Upper

117 100

119 94.5 76.8 115.2

121 30.8 24.5 36.7





#39

Methylcyclohexane

Concen: 49.970 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 83 Resp: 45603

Ion Ratio Lower Upper

83 100

55 63.8 51.4 77.2





98 49.9 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#40

Benzene

Concen: 49.906 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion: 78 Resp: 1107790

Ion Ratio Lower Upper

78 100


77 24.2 18.3 27.5





#41
Methacrylonitrile
Concen: 48.008 ug/l
RT: 7.220 min Scan# 902
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725




Tgt Ion: 41 Resp: 72100
Ion Ratio Lower Upper
41 100
39 62.2 52.2 78.4
67 97.9 74.9 112.3
52 37.2 29.2 43.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#42
1,2-Dichloroethane
Concen: 49.318 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12



Tgt Ion: 62 Resp: 282332
Ion Ratio Lower Upper
62 100
98 11.1 0.0 21.6





#43

Isopropyl Acetate

Concen: 50.548 ug/l

RT: 8.195 min Scan# 1062

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 43 Resp: 27988

Ion Ratio Lower Upper

43 100

61 22.8 18.1 27.1





87 16.7 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#44

Trichloroethene

Concen: 49.120 ug/l

RT: 8.866 min Scan# 1172

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion:130 Resp: 319131

Ion Ratio Lower Upper

130 100

95 91.9 0.0 180.6





#45

1,2-Dichloropropane

Concen: 50.355 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 63 Resp: 24688

Ion Ratio Lower Upper

63 100





65 31.0 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#46

Dibromomethane

Concen: 50.885 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion: 93 Resp: 157758


Ion Ratio Lower Upper

93 100

95 83.6 65.4 98.0

174 113.9 92.9 139.3





#47

Bromodichloromethane

Concen: 50.450 ug/l

RT: 9.420 min Scan# 1263

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :





ICVVY100725

Tgt Ion:	83	Resp:	393918
Ion	Ratio	Lower	Upper
83	100		
85	64.5	50.8	76.2
127	9.6	7.4	11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#48

Methyl methacrylate

Concen: 49.440 ug/l




RT: 9.219 min Scan# 1230


Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Tgt Ion:	41	Resp:	128606
Ion	Ratio	Lower	Upper
41	100		
69	106.6	81.0	121.4
39	63.5	48.2	72.4





#49

1,4-Dioxane

Concen: 1039.252 ug/l

RT: 9.231 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :





ICVVY100725

Tgt Ion:	88	Resp:	37350
Ion	Ratio	Lower	Upper
88	100		
43	27.4	21.8	32.8
58	61.4	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#50

Toluene-d8

Concen: 49.228 ug/l




RT: 10.103 min Scan# 1375


Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Tgt Ion:	98	Resp:	970614
Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 253.421 ug/l
RT: 10.000 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Tgt Ion: 43 Resp: 724992
Ion Ratio Lower Upper
43 100
58 44.4 35.3 52.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#52
Toluene
Concen: 51.556 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 92 Resp: 743055
Ion Ratio Lower Upper
92 100
91 168.9 136.4 204.6









#53
t-1,3-Dichloropropene
Concen: 51.160 ug/l
RT: 10.390 min Scan# 1422
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Tgt Ion: 75 Resp: 347330
Ion Ratio Lower Upper
75 100
77 31.9 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#54
cis-1,3-Dichloropropene
Concen: 51.649 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion: 75 Resp: 413860
Ion Ratio Lower Upper
75 100
77 31.5 25.4 38.0
39 42.1 32.8 49.2





#55
1,1,2-Trichloroethane
Concen: 50.430 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 97 Resp: 210249
Ion Ratio Lower Upper
97 100
83 81.4 61.8 92.6
85 53.5 41.3 61.9
99 61.8 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#56
Ethyl methacrylate
Concen: 53.385 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 69 Resp: 256731
Ion Ratio Lower Upper
69 100
41 56.1 44.6 67.0
39 34.2 27.0 40.4





#57

1,3-Dichloropropane

Concen: 51.113 ug/l

RT: 10.713 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 76 Resp: 33767

Ion Ratio Lower Upper

76 100





78 32.8 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#58

2-Chloroethyl Vinyl ether

Concen: 272.133 ug/l

RT: 9.713 min Scan# 1311

Delta R.T. 0.006 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion: 63 Resp: 586266

Ion Ratio Lower Upper

63 100


106 33.2 26.6 39.8





#59
2-Hexanone
Concen: 245.480 ug/l
RT: 10.762 min Scan# 1483
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 43 Resp: 50187
Ion Ratio Lower Upper
43 100
58 61.7 30.5 91.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#60
Dibromochloromethane
Concen: 50.974 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion:129 Resp: 297946
Ion Ratio Lower Upper
129 100
127 76.2 38.6 115.7





#61
1,2-Dibromoethane
Concen: 50.658 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:107 Resp: 201415
Ion Ratio Lower Upper
107 100
109 95.4 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#62
4-Bromofluorobenzene
Concen: 49.881 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion: 95 Resp: 324700
Ion Ratio Lower Upper
95 100
174 96.3 0.0 192.8
176 93.9 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVY100725




Tgt Ion:117 Resp: 771372
Ion Ratio Lower Upper
117 100
82 51.2 41.0 61.4
119 32.5 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025




#64
Tetrachloroethene
Concen: 48.087 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Tgt Ion:164 Resp: 382998
Ion Ratio Lower Upper
164 100
166 126.4 99.4 149.0
129 86.7 69.4 104.2
131 83.8 67.5 101.3





#65
Chlorobenzene
Concen: 49.381 ug/l
RT: 11.438 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:112 Resp: 828460
Ion Ratio Lower Upper
112 100
114 32.2 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.537 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion:131 Resp: 296157
Ion Ratio Lower Upper
131 100
133 95.1 48.4 145.2
119 65.2 32.4 97.2





#67

Ethyl Benzene

Concen: 50.841 ug/l

RT: 11.518 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725

Tgt Ion: 91 Resp: 138069

Ion Ratio Lower Upper

91 100





106 32.4 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#68

m/p-Xylenes

Concen: 101.757 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion:106 Resp: 1108386

Ion Ratio Lower Upper

106 100


91 190.7 152.6 229.0





#69
o-Xylene
Concen: 51.577 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:106 Resp: 526954
Ion Ratio Lower Upper
106 100
91 202.1 101.4 304.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#70
Styrene
Concen: 52.122 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion:104 Resp: 875709
Ion Ratio Lower Upper
104 100
78 45.3 36.1 54.1
103 54.2 43.2 64.8





#71
Bromoform
Concen: 50.070 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:173 Resp: 185682
Ion Ratio Lower Upper
173 100
175 48.4 24.8 74.3
254 0.2 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion:152 Resp: 405636
Ion Ratio Lower Upper
152 100
115 54.1 27.1 81.2
150 173.7 0.0 347.4





#73
Isopropylbenzene
Concen: 50.318 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 105 Resp: 1345105
Ion Ratio Lower Upper
105 100
120 27.7 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#74
N-ethyl acetate
Concen: 51.087 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion: 43 Resp: 245291
Ion Ratio Lower Upper
43 100
70 51.8 41.1 61.7
55 28.3 23.0 34.6
61 27.4 21.4 32.2





#75
1,1,2,2-Tetrachloroethane
Concen: 49.688 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 83 Resp: 217249
Ion Ratio Lower Upper
83 100
131 11.4 5.9 17.6
85 64.4 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#76
1,2,3-Trichloropropane
Concen: 42.874 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion: 75 Resp: 153438
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 49.590 ug/l
RT: 12.530 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:156 Resp: 352420
Ion Ratio Lower Upper
156 100
77 152.8 76.8 230.5
158 96.5 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#78
n-propylbenzene
Concen: 50.694 ug/l
RT: 12.591 min Scan# 1783
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 91 Resp: 1568180
Ion Ratio Lower Upper
91 100
120 24.7 12.3 36.8





#79

2-Chlorotoluene

Concen: 50.047 ug/l

RT: 12.676 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion: 91 Resp: 906610

Ion Ratio Lower Upper

91 100





126 37.1 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#80

1,3,5-Trimethylbenzene

Concen: 51.231 ug/l

RT: 12.737 min Scan# 1807

Delta R.T. 0.006 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion:105 Resp: 1106664

Ion Ratio Lower Upper

105 100


120 51.4 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 46.566 ug/l
RT: 12.304 min Scan# 1735
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 75 Resp: 67959
Ion Ratio Lower Upper
75 100
53 82.6 68.6 102.8
89 46.5 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#82
4-Chlorotoluene
Concen: 50.455 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 91 Resp: 941314
Ion Ratio Lower Upper
91 100
126 36.7 18.4 55.4





#83

tert-Butylbenzene

Concen: 52.019 ug/l

RT: 12.993 min Scan# 1849

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion:119 Resp: 1037510

Ion Ratio Lower Upper

119 100

91 59.5 29.9 89.8





134 26.3 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#84

1,2,4-Trimethylbenzene

Concen: 51.520 ug/l

RT: 13.042 min Scan# 1857

Delta R.T. -0.000 min

Lab File: VY023391.D




Acq: 07 Oct 2025 14:12


Tgt Ion:105 Resp: 1110672

Ion Ratio Lower Upper

105 100


120 47.0 23.8 71.2





#85
sec-Butylbenzene
Concen: 50.848 ug/l
RT: 13.176 min Scan# 1878
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVY100725



Tgt Ion:105 Resp: 1454774
Ion Ratio Lower Upper
105 100
134 21.4 10.7 32.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#86
p-Isopropyltoluene
Concen: 51.864 ug/l
RT: 13.292 min Scan# 1898
Delta R.T. 0.006 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion:119 Resp: 1263377
Ion Ratio Lower Upper
119 100
134 26.9 13.6 40.7
91 21.3 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 49.632 ug/l

RT: 13.286 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :





ICVY100725

Tgt Ion:146	Resp: 686854
Ion Ratio	Lower Upper
146	100
111	37.4 18.7 56.1
148	64.5 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#88

1,4-Dichlorobenzene

Concen: 49.140 ug/l




RT: 13.365 min Scan# 1910


Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12


Tgt Ion:146	Resp: 671644
Ion Ratio	Lower Upper
146	100
111	36.2 18.0 54.0
148	64.2 31.9 95.7





#89
n-Butylbenzene
Concen: 51.829 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion: 91 Resp: 1100080
Ion Ratio Lower Upper
91 100
92 53.4 27.0 80.8
134 28.1 14.1 42.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#90
Hexachloroethane
Concen: 49.116 ug/l
RT: 13.877 min Scan# 1994
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12


Tgt Ion:117 Resp: 257335
Ion Ratio Lower Upper
117 100
201 85.7 43.8 131.4





#91
1,2-Dichlorobenzene
Concen: 49.698 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12




Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725



Tgt Ion:146 Resp: 602989
Ion Ratio Lower Upper
146 100
111 38.7 19.0 57.0
148 64.8 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/10/2025
Supervised By :Mahesh Dadoda 10/10/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 48.687 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023391.D
Acq: 07 Oct 2025 14:12

Tgt Ion: 75 Resp: 34928
Ion Ratio Lower Upper
75 100
155 101.5 50.0 150.0
157 128.5 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 51.138 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVVY100725

Tgt Ion:180 Resp: 386800

Ion Ratio Lower Upper

180 100

182 95.5 47.9 143.6





145 28.8 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#94

Hexachlorobutadiene

Concen: 50.061 ug/l

RT: 15.023 min Scan# 2182

Delta R.T. 0.006 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12




Tgt Ion:225 Resp: 236040


Ion Ratio Lower Upper

225 100

223 62.6 30.8 92.3

227 63.6 31.9 95.7





#95

Naphthalene

Concen: 52.154 ug/l

RT: 15.139 min Scan# 2201

Delta R.T. -0.000 min

Lab File: VY023391.D


Acq: 07 Oct 2025 14:12

Instrument :

MSVOA_Y

ClientSampleId :

ICVY100725



Tgt Ion:128 Resp: 628190

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6



129 11.0 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/10/2025

Supervised By :Mahesh Dadoda 10/10/2025



#96

1,2,3-Trichlorobenzene


Concen: 51.172 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023391.D

Acq: 07 Oct 2025 14:12





Tgt Ion:180 Resp: 335267

Ion Ratio Lower Upper

180 100

182 95.7 47.9 143.8

145 31.1 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	44.916	10.2	98	0.00
3 P	Chloromethane	50.000	52.519	-5.0	112	0.00
4 C	Vinyl Chloride	50.000	49.449	1.1#	103	0.00
5 T	Bromomethane	50.000	48.856	2.3	109	0.00
6 T	Chloroethane	50.000	51.367	-2.7	105	0.00
7 T	Trichlorofluoromethane	50.000	48.582	2.8	100	0.00
8 T	Diethyl Ether	50.000	51.050	-2.1	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.728	2.5	100	0.00
10 T	Methyl Iodide	50.000	51.875	-3.8	96	0.00
11 T	Tert butyl alcohol	250.000	231.079	7.6	112	0.00
12 CM	1,1-Dichloroethene	50.000	49.653	0.7#	101	0.00
13 T	Acrolein	250.000	246.184	1.5	101	0.00
14 T	Allyl chloride	50.000	50.645	-1.3	102	0.00
15 T	Acrylonitrile	250.000	251.575	-0.6	106	0.00
16 T	Acetone	250.000	212.910	14.8	99	0.00
17 T	Carbon Disulfide	50.000	50.021	-0.0	101	0.00
18 T	Methyl Acetate	50.000	50.439	-0.9	104	0.00
19 T	Methyl tert-butyl Ether	50.000	52.230	-4.5	107	0.00
20 T	Methylene Chloride	50.000	52.503	-5.0	120	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.482	-3.0	105	0.00
22 T	Diisopropyl ether	50.000	51.364	-2.7	103	0.00
23 T	Vinyl Acetate	250.000	260.981	-4.4	106	0.00
24 P	1,1-Dichloroethane	50.000	50.742	-1.5	104	0.00
25 T	2-Butanone	250.000	233.083	6.8	105	0.00
26 T	2,2-Dichloropropane	50.000	49.626	0.7	100	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.485	-5.0	107	0.00
28 T	Bromochloromethane	50.000	45.427	9.1	99	0.00
29 T	Tetrahydrofuran	250.000	253.317	-1.3	107	0.00
30 C	Chloroform	50.000	51.137	-2.3#	104	0.00
31 T	Cyclohexane	50.000	47.416	5.2	100	0.00
32 T	1,1,1-Trichloroethane	50.000	49.934	0.1	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.450	1.1	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	49.116	1.8	101	0.00
36 T	1,1-Dichloropropene	50.000	49.761	0.5	102	0.00
37 T	Ethyl Acetate	50.000	49.740	0.5	107	0.00
38 T	Carbon Tetrachloride	50.000	48.671	2.7	100	0.00
39 T	Methylcyclohexane	50.000	49.970	0.1	101	0.00
40 TM	Benzene	50.000	49.906	0.2	102	0.00
41 T	Methacrylonitrile	50.000	48.008	4.0	105	0.00
42 TM	1,2-Dichloroethane	50.000	49.318	1.4	104	0.00
43 T	Isopropyl Acetate	50.000	50.548	-1.1	108	0.00
44 TM	Trichloroethene	50.000	49.120	1.8	101	0.00
45 C	1,2-Dichloropropane	50.000	50.355	-0.7#	105	0.00
46 T	Dibromomethane	50.000	50.885	-1.8	106	0.00
47 T	Bromodichloromethane	50.000	50.450	-0.9	104	0.00
48 T	Methyl methacrylate	50.000	49.440	1.1	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1039.252	-3.9	110	0.00
50 S	Toluene-d8	50.000	49.228	1.5	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	253.421	-1.4	106	0.00
52 CM	Toluene	50.000	51.556	-3.1#	105	0.00
53 T	t-1,3-Dichloropropene	50.000	51.160	-2.3	105	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.649	-3.3	106	0.00
55 T	1,1,2-Trichloroethane	50.000	50.430	-0.9	106	0.00
56 T	Ethyl methacrylate	50.000	53.385	-6.8	107	0.00
57 T	1,3-Dichloropropane	50.000	51.113	-2.2	105	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	272.133	-8.9	106	0.00
59 T	2-Hexanone	250.000	245.480	1.8	105	0.00
60 T	Dibromochloromethane	50.000	50.974	-1.9	106	0.00
61 T	1,2-Dibromoethane	50.000	50.658	-1.3	105	0.00
62 S	4-Bromofluorobenzene	50.000	49.881	0.2	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	103	0.00
64 T	Tetrachloroethene	50.000	48.087	3.8	98	0.00
65 PM	Chlorobenzene	50.000	49.381	1.2	105	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.537	0.9	106	0.00
67 C	Ethyl Benzene	50.000	50.841	-1.7#	104	0.00
68 T	m/p-Xylenes	100.000	101.757	-1.8	103	0.00
69 T	o-Xylene	50.000	51.577	-3.2	105	0.00
70 T	Styrene	50.000	52.122	-4.2	105	0.00
71 P	Bromoform	50.000	50.070	-0.1	108	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	103	0.00
73 T	Isopropylbenzene	50.000	50.318	-0.6	103	0.00
74 T	N-amyl acetate	50.000	51.087	-2.2	108	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.688	0.6	111	0.00
76 T	1,2,3-Trichloropropane	50.000	42.874	14.3	90	0.00
77 T	Bromobenzene	50.000	49.590	0.8	106	0.00
78 T	n-propylbenzene	50.000	50.694	-1.4	103	0.00
79 T	2-Chlorotoluene	50.000	50.047	-0.1	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.231	-2.5	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.566	6.9	103	0.00
82 T	4-Chlorotoluene	50.000	50.455	-0.9	105	0.00
83 T	tert-Butylbenzene	50.000	52.019	-4.0	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.520	-3.0	105	0.00
85 T	sec-Butylbenzene	50.000	50.848	-1.7	104	0.00
86 T	p-Isopropyltoluene	50.000	51.864	-3.7	104	0.00
87 T	1,3-Dichlorobenzene	50.000	49.632	0.7	106	0.00
88 T	1,4-Dichlorobenzene	50.000	49.140	1.7	105	0.00
89 T	n-Butylbenzene	50.000	51.829	-3.7	105	0.00
90 T	Hexachloroethane	50.000	49.116	1.8	105	0.00
91 T	1,2-Dichlorobenzene	50.000	49.698	0.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.687	2.6	107	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.138	-2.3	107	0.00
94 T	Hexachlorobutadiene	50.000	50.061	-0.1	104	0.00
95 T	Naphthalene	50.000	52.154	-4.3	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.172	-2.3	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.360	0.323	10.3	98	0.00
3 P	Chloromethane	0.490	0.515	-5.1	112	0.00
4 C	Vinyl Chloride	0.639	0.632	1.1#	103	0.00
5 T	Bromomethane	0.536	0.524	2.2	109	0.00
6 T	Chloroethane	0.435	0.447	-2.8	105	0.00
7 T	Trichlorofluoromethane	0.887	0.862	2.8	100	0.00
8 T	Diethyl Ether	0.232	0.237	-2.2	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.448	2.6	100	0.00
10 T	Methyl Iodide	0.512	0.531	-3.7	96	0.00
11 T	Tert butyl alcohol	0.031	0.029	6.5	112	0.00
12 CM	1,1-Dichloroethene	0.446	0.443	0.7#	101	0.00
13 T	Acrolein	0.044	0.043	2.3	101	0.00
14 T	Allyl chloride	0.544	0.551	-1.3	102	0.00
15 T	Acrylonitrile	0.089	0.089	0.0	106	0.00
16 T	Acetone	0.100	0.085	15.0	99	0.00
17 T	Carbon Disulfide	1.277	1.277	0.0	101	0.00
18 T	Methyl Acetate	0.261	0.263	-0.8	104	0.00
19 T	Methyl tert-butyl Ether	1.072	1.120	-4.5	107	0.00
20 T	Methylene Chloride	0.538	0.565	-5.0	120	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.506	-3.1	105	0.00
22 T	Diisopropyl ether	1.202	1.235	-2.7	103	0.00
23 T	Vinyl Acetate	0.651	0.679	-4.3	106	0.00
24 P	1,1-Dichloroethane	0.796	0.808	-1.5	104	0.00
25 T	2-Butanone	0.122	0.113	7.4	105	0.00
26 T	2,2-Dichloropropane	0.731	0.725	0.8	100	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.595	-4.9	107	0.00
28 T	Bromochloromethane	0.297	0.270	9.1	99	0.00
29 T	Tetrahydrofuran	0.069	0.070	-1.4	107	0.00
30 C	Chloroform	0.875	0.895	-2.3#	104	0.00
31 T	Cyclohexane	0.696	0.660	5.2	100	0.00
32 T	1,1,1-Trichloroethane	0.795	0.794	0.1	101	0.00
33 S	1,2-Dichloroethane-d4	0.418	0.414	1.0	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.307	0.302	1.6	101	0.00
36 T	1,1-Dichloropropene	0.407	0.405	0.5	102	0.00
37 T	Ethyl Acetate	0.161	0.160	0.6	107	0.00
38 T	Carbon Tetrachloride	0.493	0.480	2.6	100	0.00
39 T	Methylcyclohexane	0.530	0.529	0.2	101	0.00
40 TM	Benzene	1.288	1.286	0.2	102	0.00
41 T	Methacrylonitrile	0.087	0.084	3.4	105	0.00
42 TM	1,2-Dichloroethane	0.332	0.328	1.2	104	0.00
43 T	Isopropyl Acetate	0.321	0.325	-1.2	108	0.00
44 TM	Trichloroethene	0.377	0.370	1.9	101	0.00
45 C	1,2-Dichloropropane	0.285	0.287	-0.7#	105	0.00
46 T	Dibromomethane	0.180	0.183	-1.7	106	0.00
47 T	Bromodichloromethane	0.453	0.457	-0.9	104	0.00
48 T	Methyl methacrylate	0.151	0.149	1.3	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	110	0.00
50 S	Toluene-d8	1.144	1.126	1.6	100	0.00
51 T	4-Methyl-2-Pentanone	0.166	0.168	-1.2	106	0.00
52 CM	Toluene	0.836	0.862	-3.1#	105	0.00
53 T	t-1,3-Dichloropropene	0.394	0.403	-2.3	105	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.480	-3.2	106	0.00
55 T	1,1,2-Trichloroethane	0.242	0.244	-0.8	106	0.00
56 T	Ethyl methacrylate	0.279	0.298	-6.8	107	0.00
57 T	1,3-Dichloropropane	0.383	0.392	-2.3	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.136	-8.8	106	0.00
59 T	2-Hexanone	0.119	0.116	2.5	105	0.00
60 T	Dibromochloromethane	0.339	0.346	-2.1	106	0.00
61 T	1,2-Dibromoethane	0.231	0.234	-1.3	105	0.00
62 S	4-Bromofluorobenzene	0.378	0.377	0.3	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.516	0.497	3.7	98	0.00
65 PM	Chlorobenzene	1.087	1.074	1.2	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.384	1.0	106	0.00
67 C	Ethyl Benzene	1.760	1.790	-1.7#	104	0.00
68 T	m/p-Xylenes	0.706	0.718	-1.7	103	0.00
69 T	o-Xylene	0.662	0.683	-3.2	105	0.00
70 T	Styrene	1.089	1.135	-4.2	105	0.00
71 P	Bromoform	0.240	0.241	-0.4	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.295	3.316	-0.6	103	0.00
74 T	N-amyl acetate	0.592	0.605	-2.2	108	0.00
75 P	1,1,2,2-Tetrachloroethane	0.539	0.536	0.6	111	0.00
76 T	1,2,3-Trichloropropane	0.441	0.378	14.3	90	0.00
77 T	Bromobenzene	0.876	0.869	0.8	106	0.00
78 T	n-propylbenzene	3.813	3.866	-1.4	103	0.00
79 T	2-Chlorotoluene	2.233	2.235	-0.1	104	0.00
80 T	1,3,5-Trimethylbenzene	2.663	2.728	-2.4	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.168	6.7	103	0.00
82 T	4-Chlorotoluene	2.300	2.321	-0.9	105	0.00
83 T	tert-Butylbenzene	2.458	2.558	-4.1	105	0.00
84 T	1,2,4-Trimethylbenzene	2.657	2.738	-3.0	105	0.00
85 T	sec-Butylbenzene	3.527	3.586	-1.7	104	0.00
86 T	p-Isopropyltoluene	3.003	3.115	-3.7	104	0.00
87 T	1,3-Dichlorobenzene	1.706	1.693	0.8	106	0.00
88 T	1,4-Dichlorobenzene	1.685	1.656	1.7	105	0.00
89 T	n-Butylbenzene	2.616	2.712	-3.7	105	0.00
90 T	Hexachloroethane	0.646	0.634	1.9	105	0.00
91 T	1,2-Dichlorobenzene	1.496	1.487	0.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.086	2.3	107	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.954	-2.4	107	0.00
94 T	Hexachlorobutadiene	0.581	0.582	-0.2	104	0.00
95 T	Naphthalene	1.485	1.549	-4.3	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023391.D
Acq On : 07 Oct 2025 14:12
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY100725

Quant Time: Oct 08 05:15:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.808	0.827	-2.4	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3419</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/22/2025</u> <u>09:30</u>
Lab File ID:	<u>VY023543.D</u>	Init. Calib. Date(s):	<u>10/07/2025</u> <u>10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17</u> <u>12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.360	0.308		-14.44	20
Chloromethane	0.490	0.409	0.1	-16.53	20
Vinyl Chloride	0.639	0.552		-13.61	20
Bromomethane	0.536	0.456		-14.93	20
Chloroethane	0.435	0.403		-7.36	20
Trichlorofluoromethane	0.887	0.857		-3.38	20
1,1,2-Trichlorotrifluoroethane	0.460	0.469		1.96	20
1,1-Dichloroethene	0.446	0.456		2.24	20
Acetone	0.100	0.096		-4	20
Carbon Disulfide	1.277	1.206		-5.56	20
Methyl tert-butyl Ether	1.072	1.142		6.53	20
Methyl Acetate	0.261	0.294		12.64	20
Methylene Chloride	0.538	0.496		-7.81	20
trans-1,2-Dichloroethene	0.491	0.510		3.87	20
1,1-Dichloroethane	0.796	0.815	0.1	2.39	20
Cyclohexane	0.696	0.650		-6.61	20
2-Butanone	0.122	0.118		-3.28	20
Carbon Tetrachloride	0.493	0.522		5.88	20
cis-1,2-Dichloroethene	0.567	0.611		7.76	20
Bromochloromethane	0.297	0.257		-13.47	20
Chloroform	0.875	0.928		6.06	20
1,1,1-Trichloroethane	0.795	0.839		5.53	20
Methylcyclohexane	0.530	0.561		5.85	20
Benzene	1.288	1.371		6.44	20
1,2-Dichloroethane	0.332	0.340		2.41	20
Trichloroethene	0.377	0.411		9.02	20
1,2-Dichloropropane	0.285	0.303		6.32	20
Bromodichloromethane	0.453	0.487		7.51	20
4-Methyl-2-Pentanone	0.166	0.175		5.42	20
Toluene	0.836	0.926		10.77	20
t-1,3-Dichloropropene	0.394	0.433		9.9	20
cis-1,3-Dichloropropene	0.465	0.509		9.46	20
1,1,2-Trichloroethane	0.242	0.264		9.09	20
2-Hexanone	0.119	0.129		8.4	20
Dibromochloromethane	0.339	0.377		11.21	20
1,2-Dibromoethane	0.231	0.255		10.39	20
Tetrachloroethene	0.516	0.580		12.4	20
Chlorobenzene	1.087	1.168	0.3	7.45	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3419</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/22/2025</u> <u>09:30</u>
Lab File ID:	<u>VY023543.D</u>	Init. Calib. Date(s):	<u>10/07/2025</u> <u>10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17</u> <u>12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.760	1.922		9.2	20
m/p-Xylenes	0.706	0.784		11.05	20
o-Xylene	0.662	0.739		11.63	20
Styrene	1.089	1.228		12.76	20
Bromoform	0.240	0.264	0.1	10	20
Isopropylbenzene	3.295	3.699		12.26	20
1,1,2,2-Tetrachloroethane	0.539	0.562	0.3	4.27	20
1,3-Dichlorobenzene	1.706	1.845		8.15	20
1,4-Dichlorobenzene	1.685	1.798		6.71	20
1,2-Dichlorobenzene	1.496	1.634		9.23	20
1,2-Dibromo-3-Chloropropane	0.088	0.092		4.55	20
1,2,4-Trichlorobenzene	0.932	1.020		9.44	20
1,2,3-Trichlorobenzene	0.808	0.876		8.42	20
1,2-Dichloroethane-d4	0.418	0.381		-8.85	20
Dibromofluoromethane	0.307	0.307		0	20
Toluene-d8	1.144	1.136		-0.7	20
4-Bromofluorobenzene	0.378	0.382		1.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023543.D
 Acq On : 22 Oct 2025 09:30
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025

Quant Time: Oct 24 02:25:18 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 05:12:50 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	557873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	806690	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	720008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	373234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	212794	45.615	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	91.220%		
35) Dibromofluoromethane	7.634	113	247459	49.924	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	99.840%		
50) Toluene-d8	10.109	98	916029	49.624	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	99.240%		
62) 4-Bromofluorobenzene	12.408	95	307868	50.517	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	101.040%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	172025	42.840	ug/l	99
3) Chloromethane	2.074	50	228434	41.760	ug/l	100
4) Vinyl Chloride	2.208	62	307788	43.199	ug/l	98
5) Bromomethane	2.598	94	254415	42.551	ug/l	99
6) Chloroethane	2.739	64	224891	46.306	ug/l	99
7) Trichlorofluoromethane	3.062	101	477889	48.307	ug/l	100
8) Diethyl Ether	3.458	74	129976	50.281	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.818	101	261411	50.922	ug/l	98
10) Methyl Iodide	4.007	142	341595	59.772	ug/l	99
11) Tert butyl alcohol	4.860	59	86872	250.315	ug/l	99
12) 1,1-Dichloroethene	3.793	96	254223	51.086	ug/l	98
13) Acrolein	3.653	56	38914	79.353	ug/l	99
14) Allyl chloride	4.385	41	291262	48.025	ug/l	97
15) Acrylonitrile	5.061	53	248616	251.312	ug/l	99
16) Acetone	3.873	43	266925	240.345	ug/l	99
17) Carbon Disulfide	4.110	76	672714	47.232	ug/l	100
18) Methyl Acetate	4.391	43	163801	56.225	ug/l	97
19) Methyl tert-butyl Ether	5.122	73	636893	53.251	ug/l	96
20) Methylene Chloride	4.616	84	276483	46.039	ug/l	97
21) trans-1,2-Dichloroethene	5.116	96	284466	51.921	ug/l	98
22) Diisopropyl ether	6.025	45	677959	50.537	ug/l	98
23) Vinyl Acetate	5.964	43	1757567	242.046	ug/l	96
24) 1,1-Dichloroethane	5.921	63	454590	51.156	ug/l	98
25) 2-Butanone	6.896	43	328971	242.629	ug/l	100
26) 2,2-Dichloropropane	6.890	77	428536	52.555	ug/l	98
27) cis-1,2-Dichloroethene	6.896	96	340754	53.843	ug/l	96
28) Bromochloromethane	7.250	49	143454	43.243	ug/l	95
29) Tetrahydrofuran	7.268	42	185249	240.938	ug/l	94
30) Chloroform	7.427	83	517452	53.000	ug/l	99
31) Cyclohexane	7.707	56	362890	46.762	ug/l	96
32) 1,1,1-Trichloroethane	7.622	97	467952	52.741	ug/l	99
36) 1,1-Dichloropropene	7.841	75	344096	52.436	ug/l	99
37) Ethyl Acetate	6.994	43	127273	48.956	ug/l	97
38) Carbon Tetrachloride	7.823	117	421482	52.970	ug/l	99
39) Methylcyclohexane	9.109	83	452406	52.949	ug/l	97
40) Benzene	8.085	78	1105993	53.219	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023543.D
 Acq On : 22 Oct 2025 09:30
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025

Quant Time: Oct 24 02:25:18 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 05:12:50 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	68071m	48.413	ug/l	
42) 1,2-Dichloroethane	8.158	62	274669	51.247	ug/l	98
43) Isopropyl Acetate	8.201	43	259676	50.092	ug/l	96
44) Trichloroethene	8.866	130	331676	54.528	ug/l	99
45) 1,2-Dichloropropane	9.140	63	244273	53.216	ug/l	98
46) Dibromomethane	9.231	93	157780	54.359	ug/l	98
47) Bromodichloromethane	9.426	83	392784	53.731	ug/l	100
48) Methyl methacrylate	9.219	41	125981	51.729	ug/l	95
49) 1,4-Dioxane	9.231	88	35080	1042.410	ug/l	97
51) 4-Methyl-2-Pentanone	9.999	43	705865	263.542	ug/l	98
52) Toluene	10.170	92	747259	55.380	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	348965	54.901	ug/l	99
54) cis-1,3-Dichloropropene	9.859	75	410362	54.701	ug/l	97
55) 1,1,2-Trichloroethane	10.573	97	212638	54.478	ug/l	97
56) Ethyl methacrylate	10.438	69	260989	57.967	ug/l	94
57) 1,3-Dichloropropane	10.719	76	337668	54.594	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	574517	284.845	ug/l	98
59) 2-Hexanone	10.762	43	518755	271.019	ug/l	97
60) Dibromochloromethane	10.914	129	304516	55.647	ug/l	99
61) 1,2-Dibromoethane	11.018	107	205601	55.233	ug/l	100
64) Tetrachloroethene	10.646	164	417270	56.127	ug/l	97
65) Chlorobenzene	11.444	112	841259	53.721	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.517	131	302769	54.255	ug/l	99
67) Ethyl Benzene	11.517	91	1383613	54.584	ug/l	100
68) m/p-Xylenes	11.627	106	1128353	110.980	ug/l	99
69) o-Xylene	11.956	106	532269	55.814	ug/l	99
70) Styrene	11.969	104	884177	56.380	ug/l	99
71) Bromoform	12.133	173	190201	54.947	ug/l #	98
73) Isopropylbenzene	12.255	105	1380580	56.129	ug/l	100
74) N-amyl acetate	12.072	43	226427	51.252	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	209645	52.112	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	165233m	50.178	ug/l	
77) Bromobenzene	12.529	156	355320	54.338	ug/l	99
78) n-propylbenzene	12.597	91	1597439	56.123	ug/l	99
79) 2-Chlorotoluene	12.682	91	910692	54.637	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	1123341	56.518	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	74591	55.547	ug/l	95
82) 4-Chlorotoluene	12.779	91	937415	54.609	ug/l	99
83) tert-Butylbenzene	12.999	119	1037184	56.518	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	1118689	56.397	ug/l	99
85) sec-Butylbenzene	13.176	105	1487514	56.506	ug/l	100
86) p-Isopropyltoluene	13.292	119	1282400	57.215	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	688741	54.089	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	671095	53.362	ug/l	99
89) n-Butylbenzene	13.615	91	1101186	56.385	ug/l	99
90) Hexachloroethane	13.877	117	255425	52.984	ug/l	100
91) 1,2-Dichlorobenzene	13.657	146	609689	54.613	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	34265	51.910	ug/l	96
93) 1,2,4-Trichlorobenzene	14.919	180	380515	54.674	ug/l	100
94) Hexachlorobutadiene	15.023	225	233049	53.717	ug/l	100
95) Naphthalene	15.145	128	636726	57.452	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	326959	54.237	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 982
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 55787
Ion Ratio Lower Upper
168 100
99 47.6 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#2
Dichlorodifluoromethane
Concen: 42.840 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Tgt Ion: 85 Resp: 172025
Ion Ratio Lower Upper
85 100
87 32.7 16.0 47.9









#3
Chloromethane
Concen: 41.760 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 228434
Ion Ratio Lower Upper
50 100
52 32.9 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#4
Vinyl Chloride
Concen: 43.199 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Tgt Ion: 62 Resp: 307788
Ion Ratio Lower Upper
62 100
64 33.1 25.8 38.8









#5
Bromomethane
Concen: 42.551 ug/l
RT: 2.598 min Scan# 143
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Instrument :
MSVOA_Y
ClientSampleId :
VSTDC050

Tgt Ion: 94 Resp: 25441
Ion Ratio Lower Upper
94 100
96 94.3 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#6
Chloroethane
Concen: 46.306 ug/l
RT: 2.739 min Scan# 167
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30


Tgt Ion: 64 Resp: 224891
Ion Ratio Lower Upper
64 100
66 32.2 26.2 39.4





#7
Trichlorofluoromethane
Concen: 48.307 ug/l
RT: 3.062 min Scan# 219
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 101 Resp: 477889
Ion Ratio Lower Upper
101 100
103 64.4 51.3 76.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#8
Diethyl Ether
Concen: 50.281 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30



Tgt Ion: 74 Resp: 129976
Ion Ratio Lower Upper
74 100
45 76.8 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.922 ug/l
RT: 3.818 min Scan# 343
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 26141
Ion Ratio Lower Upper
101 100
85 43.3 34.6 51.8
151 86.3 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#10
Methyl Iodide
Concen: 59.772 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Tgt Ion:142 Resp: 341595
Ion Ratio Lower Upper
142 100
127 42.5 34.8 52.2
141 14.3 11.6 17.4





#11

Tert butyl alcohol

Concen: 250.315 ug/l

RT: 4.860 min Scan# 515

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDC050

Tgt Ion: 59 Resp: 86872

Ion Ratio Lower Upper

59 100





57 10.9 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#12

1,1-Dichloroethene

Concen: 51.086 ug/l

RT: 3.793 min Scan# 340

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion: 96 Resp: 254223


Ion Ratio Lower Upper

96 100

61 136.2 106.7 160.1


98 63.6 50.2 75.4





#13
Acrolein
Concen: 79.353 ug/l
RT: 3.653 min Scan# 316
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 38914
Ion Ratio Lower Upper
56 100
55 68.9 56.1 84.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#14
Allyl chloride
Concen: 48.025 ug/l
RT: 4.385 min Scan# 437
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30



Tgt Ion: 41 Resp: 291262
Ion Ratio Lower Upper
41 100
39 76.0 59.9 89.9
76 51.3 38.0 57.0





#15

Acrylonitrile

Concen: 251.312 ug/l

RT: 5.061 min Scan# 54

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 53 Resp: 248610

Ion Ratio Lower Upper

53 100

52 80.3 64.6 97.0





51 35.8 29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#16

Acetone

Concen: 240.345 ug/l

RT: 3.873 min Scan# 353

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion: 43 Resp: 266925

Ion Ratio Lower Upper

43 100

58 35.1 27.7 41.5





#17

Carbon Disulfide

Concen: 47.232 ug/l

RT: 4.110 min Scan# 391

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 672714

Ion Ratio Lower Upper

76 100





78 9.0 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#18

Methyl Acetate

Concen: 56.225 ug/l

RT: 4.391 min Scan# 438

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion: 43 Resp: 163801

Ion Ratio Lower Upper

43 100

74 32.2 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 53.251 ug/l

RT: 5.122 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 63689

Ion Ratio Lower Upper

73 100





57 19.6 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#20

Methylene Chloride

Concen: 46.039 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Tgt Ion: 84 Resp: 276483




Ion Ratio Lower Upper


84 100

49 104.4 80.2 120.4

51 31.1 25.9 38.9


86 64.6 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 51.921 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 96 Resp: 284460
Ion Ratio Lower Upper
96 100
61 113.0 92.7 139.1
98 62.8 49.7 74.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#22
Diisopropyl ether
Concen: 50.537 ug/l
RT: 6.025 min Scan# 706
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30



Tgt Ion: 45 Resp: 677959
Ion Ratio Lower Upper
45 100
43 57.6 45.6 68.4
87 37.0 28.1 42.1
59 14.1 10.2 15.2





#23

Vinyl Acetate

Concen: 242.046 ug/l

RT: 5.964 min Scan# 695

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 175756

Ion Ratio Lower Upper

43

100

86

14.7

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#24

1,1-Dichloroethane

Concen: 51.156 ug/l

RT: 5.921 min Scan# 689

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Tgt Ion: 63 Resp: 454590

Ion Ratio Lower Upper

63

100

98

8.2

4.3




13.1


100

5.0

2.8

8.4





#25

2-Butanone

Concen: 242.629 ug/l

RT: 6.896 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 32897

Ion Ratio Lower Upper

43 100





72 34.4 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#26

2,2-Dichloropropane

Concen: 52.555 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion: 77 Resp: 428536

Ion Ratio Lower Upper

77 100

97 25.0 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 53.843 ug/l

RT: 6.896 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 340754

Ion Ratio Lower Upper

96 100

61 122.7 0.0 258.2





98 65.1 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#28

Bromochloromethane

Concen: 43.243 ug/l

RT: 7.250 min Scan# 907

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion: 49 Resp: 143454


Ion Ratio Lower Upper

49 100

129 2.7 0.0 5.6

130 107.8 81.9 122.9





#29

Tetrahydrofuran

Concen: 240.938 ug/l

RT: 7.268 min Scan# 910

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 185249

Ion Ratio Lower Upper

42 100

72 58.5 43.0 64.6





71 53.4 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#30

Chloroform

Concen: 53.000 ug/l

RT: 7.427 min Scan# 936

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion: 83 Resp: 517452

Ion Ratio Lower Upper

83 100

85 66.6 52.5 78.7





#31

Cyclohexane

Concen: 46.762 ug/l

RT: 7.707 min Scan# 981

Delta R.T. 0.006 min

Lab File: VY023543.D


Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050



Tgt Ion: 56 Resp: 362890

Ion Ratio Lower Upper

56 100

69 37.0 30.0 45.0




84 106.3 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#32

1,1,1-Trichloroethane

Concen: 52.741 ug/l

RT: 7.622 min Scan# 968

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion: 97 Resp: 467952


Ion Ratio Lower Upper

97 100

99 64.3 51.4 77.2

61 38.7 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 45.615 ug/l

RT: 8.067 min Scan# 1040

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 212794

Ion Ratio Lower Upper

65 100





67 53.7 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion:114 Resp: 806690


Ion Ratio Lower Upper

114 100

63 16.1 0.0 34.2

88 15.0 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.924 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 247459

Ion Ratio Lower Upper

113 100

111 102.5 81.8 122.6





192 22.6 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#36

1,1-Dichloropropene

Concen: 52.436 ug/l

RT: 7.841 min Scan# 1004

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion: 75 Resp: 344096


Ion Ratio Lower Upper

75 100

110 40.7 19.9 59.7

77 31.7 25.0 37.6





#37

Ethyl Acetate

Concen: 48.956 ug/l

RT: 6.994 min Scan# 80

Delta R.T. 0.012 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 12727

Ion Ratio Lower Upper

43 100

61 17.2 11.8 17.8





70 13.2 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#38

Carbon Tetrachloride

Concen: 52.970 ug/l

RT: 7.823 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion:117 Resp: 421482


Ion Ratio Lower Upper

117 100

119 97.0 76.8 115.2

121 32.1 24.5 36.7





#39

Methylcyclohexane

Concen: 52.949 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 452400

Ion Ratio Lower Upper

83 100

55 62.1 51.4 77.2





98 48.1 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#40

Benzene

Concen: 53.219 ug/l

RT: 8.085 min Scan# 1044

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion: 78 Resp: 1105993

Ion Ratio Lower Upper

78 100

77 24.0 18.3 27.5





#41

Methacrylonitrile

Concen: 48.413 ug/l m

RT: 7.219 min Scan# 902

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 41 Resp: 6807

Ion Ratio Lower Upper

41 100

39 64.3 52.2 78.4

67 102.9 74.9 112.3





52 38.7 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#42

1,2-Dichloroethane

Concen: 51.247 ug/l

RT: 8.158 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion: 62 Resp: 274669

Ion Ratio Lower Upper

62 100

98 11.4 0.0 21.6





#43

Isopropyl Acetate

Concen: 50.092 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 259670

Ion Ratio Lower Upper

43 100

61 25.1 18.1 27.1





87 18.0 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#44

Trichloroethene

Concen: 54.528 ug/l

RT: 8.866 min Scan# 1172

Delta R.T. -0.000 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion:130 Resp: 331676

Ion Ratio Lower Upper

130 100

95 91.4 0.0 180.6





#45

1,2-Dichloropropane

Concen: 53.216 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 24427

Ion Ratio Lower Upper

63 100





65 32.0 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#46

Dibromomethane

Concen: 54.359 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion: 93 Resp: 157780


Ion Ratio Lower Upper

93 100

95 83.5 65.4 98.0

174 117.4 92.9 139.3





#47

Bromodichloromethane

Concen: 53.731 ug/l

RT: 9.426 min Scan# 1264

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 392784

Ion	Ratio	Lower	Upper
83	100		
85	63.7	50.8	76.2
127	9.3	7.4	11.2

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#48

Methyl methacrylate

Concen: 51.729 ug/l

RT: 9.219 min Scan# 1230




Delta R.T. -0.000 min


Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Tgt Ion: 41 Resp: 125981

Ion	Ratio	Lower	Upper
41	100		
69	108.7	81.0	121.4
39	60.9	48.2	72.4





#49

1,4-Dioxane

Concen: 1042.410 ug/l

RT: 9.231 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	88	Resp:	35080
Ion	Ratio	Lower	Upper
88	100		
43	26.1	21.8	32.8
58	56.0	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#50

Toluene-d8

Concen: 49.624 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30


Tgt Ion:	98	Resp:	916029
Ion	Ratio	Lower	Upper
98	100		
100	66.0	53.0	79.4





#51
4-Methyl-2-Pentanone
Concen: 263.542 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 70586
Ion Ratio Lower Upper
43 100
58 45.5 35.3 52.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#52
Toluene
Concen: 55.380 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Tgt Ion: 92 Resp: 747259
Ion Ratio Lower Upper
92 100
91 168.7 136.4 204.6





#53
t-1,3-Dichloropropene
Concen: 54.901 ug/l
RT: 10.396 min Scan# 1423
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 34896
Ion Ratio Lower Upper
75 100
77 32.0 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#54
cis-1,3-Dichloropropene
Concen: 54.701 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30


Tgt Ion: 75 Resp: 410362
Ion Ratio Lower Upper
75 100
77 30.9 25.4 38.0
39 38.6 32.8 49.2





#55
1,1,2-Trichloroethane
Concen: 54.478 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 212638
Ion Ratio Lower Upper
97 100
83 80.8 61.8 92.6
85 53.5 41.3 61.9
99 63.2 48.7 73.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#56
Ethyl methacrylate
Concen: 57.967 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Tgt Ion: 69 Resp: 260989
Ion Ratio Lower Upper
69 100
41 51.2 44.6 67.0
39 30.1 27.0 40.4





#57
1,3-Dichloropropane
Concen: 54.594 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 337668
Ion Ratio Lower Upper
76 100
78 32.7 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#58
2-Chloroethyl Vinyl ether
Concen: 284.845 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30



Tgt Ion: 63 Resp: 574517
Ion Ratio Lower Upper
63 100
106 34.1 26.6 39.8





#59

2-Hexanone

Concen: 271.019 ug/l

RT: 10.762 min Scan# 1482

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 51875

Ion Ratio Lower Upper

43 100





58 63.3 30.5 91.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#60

Dibromochloromethane

Concen: 55.647 ug/l

RT: 10.914 min Scan# 1508

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion:129 Resp: 304516

Ion Ratio Lower Upper

129 100


127 76.6 38.6 115.7





#61
1,2-Dibromoethane
Concen: 55.233 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 107 Resp: 20560
Ion Ratio Lower Upper
107 100
109 95.5 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#62
4-Bromofluorobenzene
Concen: 50.517 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30


Tgt Ion: 95 Resp: 307868
Ion Ratio Lower Upper
95 100
174 97.4 0.0 192.8
176 93.9 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 1591
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 720000
Ion Ratio Lower Upper
117 100
82 48.7 41.0 61.4
119 32.8 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#64
Tetrachloroethene
Concen: 56.127 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30



Tgt Ion:164 Resp: 417270
Ion Ratio Lower Upper
164 100
166 129.6 99.4 149.0
129 89.0 69.4 104.2
131 86.8 67.5 101.3





#65

Chlorobenzene

Concen: 53.721 ug/l

RT: 11.444 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:112 Resp: 841259

Ion Ratio Lower Upper

112 100





114 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 54.255 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion:131 Resp: 302769


Ion Ratio Lower Upper

131 100

133 95.7 48.4 145.2


119 65.3 32.4 97.2





#67
Ethyl Benzene
Concen: 54.584 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 138361
Ion Ratio Lower Upper
91 100
106 32.6 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#68
m/p-Xylenes
Concen: 110.980 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30


Tgt Ion:106 Resp: 1128353
Ion Ratio Lower Upper
106 100
91 189.3 152.6 229.0





#69
o-Xylene
Concen: 55.814 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:106 Resp: 532269
Ion Ratio Lower Upper
106 100
91 201.1 101.4 304.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#70
Styrene
Concen: 56.380 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Tgt Ion:104 Resp: 884177
Ion Ratio Lower Upper
104 100
78 44.6 36.1 54.1
103 54.2 43.2 64.8





#71
Bromoform
Concen: 54.947 ug/l
RT: 12.133 min Scan# 1707
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:173 Resp: 19020
Ion Ratio Lower Upper
173 100
175 47.9 24.8 74.3
254 0.1 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Tgt Ion:152 Resp: 373234
Ion Ratio Lower Upper
152 100
115 54.1 27.1 81.2
150 174.4 0.0 347.4





#73

Isopropylbenzene

Concen: 56.129 ug/l

RT: 12.255 min Scan# 1728

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 1380580

Ion Ratio Lower Upper

105 100





120 28.0 13.9 41.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#74

N-ethyl acetate

Concen: 51.252 ug/l

RT: 12.072 min Scan# 1698

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Tgt Ion: 43 Resp: 226427




Ion Ratio Lower Upper


43 100

70 54.6 41.1 61.7

55 29.0 23.0 34.6


61 28.9 21.4 32.2





#75
1,1,2,2-Tetrachloroethane
Concen: 52.112 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 20964
Ion Ratio Lower Upper
83 100
131 11.7 5.9 17.6
85 64.8 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#76
1,2,3-Trichloropropane
Concen: 50.178 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Tgt Ion: 75 Resp: 165233
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 54.338 ug/l

RT: 12.529 min Scan# 1773

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:156 Resp: 355320
Ion Ratio Lower Upper
156 100
77 150.7 76.8 230.5
158 98.3 49.0 147.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#78

n-propylbenzene

Concen: 56.123 ug/l




RT: 12.597 min Scan# 1784


Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Tgt Ion: 91 Resp: 1597439
Ion Ratio Lower Upper
91 100
120 24.8 12.3 36.8





#79

2-Chlorotoluene

Concen: 54.637 ug/l

RT: 12.682 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 910692

Ion Ratio Lower Upper

91 100





126 37.3 18.5 55.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#80

1,3,5-Trimethylbenzene

Concen: 56.518 ug/l

RT: 12.737 min Scan# 1807

Delta R.T. 0.006 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion:105 Resp: 1123341

Ion Ratio Lower Upper

105 100


120 52.0 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 55.547 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 7459
Ion Ratio Lower Upper
75 100
53 80.3 68.6 102.8
89 47.0 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#82
4-Chlorotoluene
Concen: 54.609 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30


Tgt Ion: 91 Resp: 937415
Ion Ratio Lower Upper
91 100
126 37.7 18.4 55.4





#83
tert-Butylbenzene
Concen: 56.518 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:119 Resp: 1037184
Ion Ratio Lower Upper
119 100
91 59.6 29.9 89.8
134 26.5 12.9 38.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#84
1,2,4-Trimethylbenzene
Concen: 56.397 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Tgt Ion:105 Resp: 1118689
Ion Ratio Lower Upper
105 100
120 47.9 23.8 71.2





#85
sec-Butylbenzene
Concen: 56.506 ug/l
RT: 13.176 min Scan# 1878
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:105 Resp: 1487514
Ion Ratio Lower Upper
105 100
134 21.6 10.7 32.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#86
p-Isopropyltoluene
Concen: 57.215 ug/l
RT: 13.292 min Scan# 1898
Delta R.T. 0.006 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30



Tgt Ion:119 Resp: 1282400
Ion Ratio Lower Upper
119 100
134 27.1 13.6 40.7
91 21.6 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 54.089 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 68874

Ion Ratio Lower Upper

146 100

111 37.0 18.7 56.1





148 64.4 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#88

1,4-Dichlorobenzene

Concen: 53.362 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion:146 Resp: 671095


Ion Ratio Lower Upper

146 100

111 36.1 18.0 54.0

148 64.6 31.9 95.7





#89

n-Butylbenzene

Concen: 56.385 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 1101180

Ion Ratio Lower Upper

91 100

92 53.7 27.0 80.8





134 28.7 14.1 42.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#90

Hexachloroethane

Concen: 52.984 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. -0.000 min

Lab File: VY023543.D




Acq: 22 Oct 2025 09:30


Tgt Ion:117 Resp: 255425

Ion Ratio Lower Upper

117 100


201 87.8 43.8 131.4





#91
1,2-Dichlorobenzene
Concen: 54.613 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 609689
Ion Ratio Lower Upper
146 100
111 38.0 19.0 57.0
148 63.6 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.910 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023543.D
Acq: 22 Oct 2025 09:30

Tgt Ion: 75 Resp: 34265
Ion Ratio Lower Upper
75 100
155 103.1 50.0 150.0
157 134.1 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 54.674 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023543.D


Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050



Tgt Ion:180 Resp: 38051

Ion Ratio Lower Upper

180 100

182 95.2 47.9 143.6




145 29.1 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#94

Hexachlorobutadiene


Concen: 53.717 ug/l

RT: 15.023 min Scan# 2182

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30





Tgt Ion:225 Resp: 233049


Ion Ratio Lower Upper

225 100

223 62.2 30.8 92.3

227 63.9 31.9 95.7





#95

Naphthalene

Concen: 57.452 ug/l

RT: 15.145 min Scan# 2201

Delta R.T. 0.006 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:128 Resp: 636720

Ion Ratio Lower Upper

128 100

127 12.8 10.4 15.6





129 10.8 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#96

1,2,3-Trichlorobenzene

Concen: 54.237 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023543.D

Acq: 22 Oct 2025 09:30




Tgt Ion:180 Resp: 326959

Ion Ratio Lower Upper

180 100

182 95.0 47.9 143.8

145 30.8 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.360	0.308	14.4	91	0.00
3 P	Chloromethane	0.490	0.409	16.5	86	0.00
4 C	Vinyl Chloride	0.639	0.552	13.6#	87	0.00
5 T	Bromomethane	0.536	0.456	14.9	92	0.00
6 T	Chloroethane	0.435	0.403	7.4	92	0.00
7 T	Trichlorofluoromethane	0.887	0.857	3.4	96	0.00
8 T	Diethyl Ether	0.232	0.233	-0.4	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.469	-2.0	101	0.00
10 T	Methyl Iodide	0.512	0.612	-19.5	108	0.00
11 T	Tert butyl alcohol	0.031	0.031	0.0	118	0.00
12 CM	1,1-Dichloroethene	0.446	0.456	-2.2#	101	0.00
13 T	Acrolein	0.044	0.014	68.2#	31#	0.00
14 T	Allyl chloride	0.544	0.522	4.0	94	0.00
15 T	Acrylonitrile	0.089	0.089	0.0	103	0.00
16 T	Acetone	0.100	0.096	4.0	108	0.00
17 T	Carbon Disulfide	1.277	1.206	5.6	93	0.00
18 T	Methyl Acetate	0.261	0.294	-12.6	113	0.00
19 T	Methyl tert-butyl Ether	1.072	1.142	-6.5	106	0.00
20 T	Methylene Chloride	0.538	0.496	7.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.510	-3.9	103	0.00
22 T	Diisopropyl ether	1.202	1.215	-1.1	99	0.00
23 T	Vinyl Acetate	0.651	0.630	3.2	95	0.00
24 P	1,1-Dichloroethane	0.796	0.815	-2.4	101	0.00
25 T	2-Butanone	0.122	0.118	3.3	106	0.00
26 T	2,2-Dichloropropane	0.731	0.768	-5.1	103	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.611	-7.8	106	0.00
28 T	Bromochloromethane	0.297	0.257	13.5	91	0.00
29 T	Tetrahydrofuran	0.069	0.066	4.3	99	0.00
30 C	Chloroform	0.875	0.928	-6.1#	105	0.00
31 T	Cyclohexane	0.696	0.650	6.6	96	0.00
32 T	1,1,1-Trichloroethane	0.795	0.839	-5.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.418	0.381	8.9	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.307	0.307	0.0	97	0.00
36 T	1,1-Dichloropropene	0.407	0.427	-4.9	101	0.00
37 T	Ethyl Acetate	0.161	0.158	1.9	99	0.01
38 T	Carbon Tetrachloride	0.493	0.522	-5.9	102	0.00
39 T	Methylcyclohexane	0.530	0.561	-5.8	100	0.00
40 TM	Benzene	1.288	1.371	-6.4	102	0.00
41 T	Methacrylonitrile	0.087	0.084	3.4	99	0.00
42 TM	1,2-Dichloroethane	0.332	0.340	-2.4	102	0.00
43 T	Isopropyl Acetate	0.321	0.322	-0.3	100	0.00
44 TM	Trichloroethene	0.377	0.411	-9.0	105	0.00
45 C	1,2-Dichloropropane	0.285	0.303	-6.3#	104	0.00
46 T	Dibromomethane	0.180	0.196	-8.9	106	0.00
47 T	Bromodichloromethane	0.453	0.487	-7.5	104	0.00
48 T	Methyl methacrylate	0.151	0.156	-3.3	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	103	0.00
50 S	Toluene-d8	1.144	1.136	0.7	94	0.00
51 T	4-Methyl-2-Pentanone	0.166	0.175	-5.4	103	0.00
52 CM	Toluene	0.836	0.926	-10.8#	105	0.00
53 T	t-1,3-Dichloropropene	0.394	0.433	-9.9	106	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.509	-9.5	105	0.00
55 T	1,1,2-Trichloroethane	0.242	0.264	-9.1	107	0.00
56 T	Ethyl methacrylate	0.279	0.324	-16.1	109	0.00
57 T	1,3-Dichloropropane	0.383	0.419	-9.4	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.142	-13.6	104	0.00
59 T	2-Hexanone	0.119	0.129	-8.4	108	0.00
60 T	Dibromochloromethane	0.339	0.377	-11.2	108	0.00
61 T	1,2-Dibromoethane	0.231	0.255	-10.4	107	0.00
62 S	4-Bromofluorobenzene	0.378	0.382	-1.1	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.516	0.580	-12.4	106	0.00
65 PM	Chlorobenzene	1.087	1.168	-7.5	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.421	-8.5	108	0.00
67 C	Ethyl Benzene	1.760	1.922	-9.2#	104	0.00
68 T	m/p-Xylenes	0.706	0.784	-11.0	105	0.00
69 T	o-Xylene	0.662	0.739	-11.6	106	0.00
70 T	Styrene	1.089	1.228	-12.8	106	0.00
71 P	Bromoform	0.240	0.264	-10.0	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.295	3.699	-12.3	106	0.00
74 T	N-amyl acetate	0.592	0.607	-2.5	99	0.00
75 P	1,1,2,2-Tetrachloroethane	0.539	0.562	-4.3	107	0.00
76 T	1,2,3-Trichloropropane	0.441	0.443	-0.5	97	0.00
77 T	Bromobenzene	0.876	0.952	-8.7	107	0.00
78 T	n-propylbenzene	3.813	4.280	-12.2	105	0.00
79 T	2-Chlorotoluene	2.233	2.440	-9.3	104	0.00
80 T	1,3,5-Trimethylbenzene	2.663	3.010	-13.0	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.200	-11.1	113	0.00
82 T	4-Chlorotoluene	2.300	2.512	-9.2	105	0.00
83 T	tert-Butylbenzene	2.458	2.779	-13.1	105	0.00
84 T	1,2,4-Trimethylbenzene	2.657	2.997	-12.8	105	0.00
85 T	sec-Butylbenzene	3.527	3.985	-13.0	106	0.00
86 T	p-Isopropyltoluene	3.003	3.436	-14.4	106	0.00
87 T	1,3-Dichlorobenzene	1.706	1.845	-8.1	106	0.00
88 T	1,4-Dichlorobenzene	1.685	1.798	-6.7	105	0.00
89 T	n-Butylbenzene	2.616	2.950	-12.8	105	0.00
90 T	Hexachloroethane	0.646	0.684	-5.9	104	0.00
91 T	1,2-Dichlorobenzene	1.496	1.634	-9.2	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.092	-4.5	105	0.00
93 T	1,2,4-Trichlorobenzene	0.932	1.020	-9.4	106	0.00
94 T	Hexachlorobutadiene	0.581	0.624	-7.4	103	0.00
95 T	Naphthalene	1.485	1.706	-14.9	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.808	0.876	-8.4	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	42.840	14.3	91	0.00
3 P	Chloromethane	50.000	41.760	16.5	86	0.00
4 C	Vinyl Chloride	50.000	43.199	13.6#	87	0.00
5 T	Bromomethane	50.000	42.551	14.9	92	0.00
6 T	Chloroethane	50.000	46.306	7.4	92	0.00
7 T	Trichlorofluoromethane	50.000	48.307	3.4	96	0.00
8 T	Diethyl Ether	50.000	50.281	-0.6	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.922	-1.8	101	0.00
10 T	Methyl Iodide	50.000	59.772	-19.5	108	0.00
11 T	Tert butyl alcohol	250.000	250.315	-0.1	118	0.00
12 CM	1,1-Dichloroethene	50.000	51.086	-2.2#	101	0.00
13 T	Acrolein	250.000	79.353	68.3#	31	0.00
14 T	Allyl chloride	50.000	48.025	4.0	94	0.00
15 T	Acrylonitrile	250.000	251.312	-0.5	103	0.00
16 T	Acetone	250.000	240.345	3.9	108	0.00
17 T	Carbon Disulfide	50.000	47.232	5.5	93	0.00
18 T	Methyl Acetate	50.000	56.225	-12.5	113	0.00
19 T	Methyl tert-butyl Ether	50.000	53.251	-6.5	106	0.00
20 T	Methylene Chloride	50.000	46.039	7.9	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.921	-3.8	103	0.00
22 T	Diisopropyl ether	50.000	50.537	-1.1	99	0.00
23 T	Vinyl Acetate	250.000	242.046	3.2	95	0.00
24 P	1,1-Dichloroethane	50.000	51.156	-2.3	101	0.00
25 T	2-Butanone	250.000	242.629	2.9	106	0.00
26 T	2,2-Dichloropropane	50.000	52.555	-5.1	103	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.843	-7.7	106	0.00
28 T	Bromochloromethane	50.000	43.243	13.5	91	0.00
29 T	Tetrahydrofuran	250.000	240.938	3.6	99	0.00
30 C	Chloroform	50.000	53.000	-6.0#	105	0.00
31 T	Cyclohexane	50.000	46.762	6.5	96	0.00
32 T	1,1,1-Trichloroethane	50.000	52.741	-5.5	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.615	8.8	92	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	49.924	0.2	97	0.00
36 T	1,1-Dichloropropene	50.000	52.436	-4.9	101	0.00
37 T	Ethyl Acetate	50.000	48.956	2.1	99	0.01
38 T	Carbon Tetrachloride	50.000	52.970	-5.9	102	0.00
39 T	Methylcyclohexane	50.000	52.949	-5.9	100	0.00
40 TM	Benzene	50.000	53.219	-6.4	102	0.00
41 T	Methacrylonitrile	50.000	48.413	3.2	99	0.00
42 TM	1,2-Dichloroethane	50.000	51.247	-2.5	102	0.00
43 T	Isopropyl Acetate	50.000	50.092	-0.2	100	0.00
44 TM	Trichloroethene	50.000	54.528	-9.1	105	0.00
45 C	1,2-Dichloropropane	50.000	53.216	-6.4#	104	0.00
46 T	Dibromomethane	50.000	54.359	-8.7	106	0.00
47 T	Bromodichloromethane	50.000	53.731	-7.5	104	0.00
48 T	Methyl methacrylate	50.000	51.729	-3.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound		Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1042.410	-4.2	103	0.00
50 S	Toluene-d8	50.000	49.624	0.8	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	263.542	-5.4	103	0.00
52 CM	Toluene	50.000	55.380	-10.8#	105	0.00
53 T	t-1,3-Dichloropropene	50.000	54.901	-9.8	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.701	-9.4	105	0.00
55 T	1,1,2-Trichloroethane	50.000	54.478	-9.0	107	0.00
56 T	Ethyl methacrylate	50.000	57.967	-15.9	109	0.00
57 T	1,3-Dichloropropane	50.000	54.594	-9.2	105	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	284.845	-13.9	104	0.00
59 T	2-Hexanone	250.000	271.019	-8.4	108	0.00
60 T	Dibromochloromethane	50.000	55.647	-11.3	108	0.00
61 T	1,2-Dibromoethane	50.000	55.233	-10.5	107	0.00
62 S	4-Bromofluorobenzene	50.000	50.517	-1.0	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	56.127	-12.3	106	0.00
65 PM	Chlorobenzene	50.000	53.721	-7.4	107	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.255	-8.5	108	0.00
67 C	Ethyl Benzene	50.000	54.584	-9.2#	104	0.00
68 T	m/p-Xylenes	100.000	110.980	-11.0	105	0.00
69 T	o-Xylene	50.000	55.814	-11.6	106	0.00
70 T	Styrene	50.000	56.380	-12.8	106	0.00
71 P	Bromoform	50.000	54.947	-9.9	110	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	56.129	-12.3	106	0.00
74 T	N-amyl acetate	50.000	51.252	-2.5	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.112	-4.2	107	0.00
76 T	1,2,3-Trichloropropane	50.000	50.178	-0.4	97	0.00
77 T	Bromobenzene	50.000	54.338	-8.7	107	0.00
78 T	n-propylbenzene	50.000	56.123	-12.2	105	0.00
79 T	2-Chlorotoluene	50.000	54.637	-9.3	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	56.518	-13.0	106	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	55.547	-11.1	113	0.00
82 T	4-Chlorotoluene	50.000	54.609	-9.2	105	0.00
83 T	tert-Butylbenzene	50.000	56.518	-13.0	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	56.397	-12.8	105	0.00
85 T	sec-Butylbenzene	50.000	56.506	-13.0	106	0.00
86 T	p-Isopropyltoluene	50.000	57.215	-14.4	106	0.00
87 T	1,3-Dichlorobenzene	50.000	54.089	-8.2	106	0.00
88 T	1,4-Dichlorobenzene	50.000	53.362	-6.7	105	0.00
89 T	n-Butylbenzene	50.000	56.385	-12.8	105	0.00
90 T	Hexachloroethane	50.000	52.984	-6.0	104	0.00
91 T	1,2-Dichlorobenzene	50.000	54.613	-9.2	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.910	-3.8	105	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.674	-9.3	106	0.00
94 T	Hexachlorobutadiene	50.000	53.717	-7.4	103	0.00
95 T	Naphthalene	50.000	57.452	-14.9	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023543.D
Acq On : 22 Oct 2025 09:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 02:25:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.237	-8.5	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3419</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/24/2025 08:56</u>
Lab File ID:	<u>VY023559.D</u>	Init. Calib. Date(s):	<u>10/07/2025 10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17 12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.360	0.296		-17.78	20
Chloromethane	0.490	0.393	0.1	-19.8	20
Vinyl Chloride	0.639	0.540		-15.49	20
Bromomethane	0.536	0.487		-9.14	20
Chloroethane	0.435	0.389		-10.57	20
Trichlorofluoromethane	0.887	0.844		-4.85	20
1,1,2-Trichlorotrifluoroethane	0.460	0.472		2.61	20
1,1-Dichloroethene	0.446	0.451		1.12	20
Acetone	0.100	0.101		1	20
Carbon Disulfide	1.277	1.159		-9.24	20
Methyl tert-butyl Ether	1.072	1.188		10.82	20
Methyl Acetate	0.261	0.280		7.28	20
Methylene Chloride	0.538	0.506		-5.95	20
trans-1,2-Dichloroethene	0.491	0.511		4.07	20
1,1-Dichloroethane	0.796	0.814	0.1	2.26	20
Cyclohexane	0.696	0.644		-7.47	20
2-Butanone	0.122	0.124		1.64	20
Carbon Tetrachloride	0.493	0.528		7.1	20
cis-1,2-Dichloroethene	0.567	0.623		9.88	20
Bromochloromethane	0.297	0.248		-16.5	20
Chloroform	0.875	0.933		6.63	20
1,1,1-Trichloroethane	0.795	0.836		5.16	20
Methylcyclohexane	0.530	0.561		5.85	20
Benzene	1.288	1.380		7.14	20
1,2-Dichloroethane	0.332	0.338		1.81	20
Trichloroethene	0.377	0.413		9.55	20
1,2-Dichloropropane	0.285	0.304		6.67	20
Bromodichloromethane	0.453	0.497		9.71	20
4-Methyl-2-Pentanone	0.166	0.181		9.04	20
Toluene	0.836	0.928		11.01	20
t-1,3-Dichloropropene	0.394	0.436		10.66	20
cis-1,3-Dichloropropene	0.465	0.520		11.83	20
1,1,2-Trichloroethane	0.242	0.275		13.64	20
2-Hexanone	0.119	0.131		10.08	20
Dibromochloromethane	0.339	0.389		14.75	20
1,2-Dibromoethane	0.231	0.263		13.85	20
Tetrachloroethene	0.516	0.576		11.63	20
Chlorobenzene	1.087	1.224	0.3	12.6	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3419</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/24/2025 08:56</u>
Lab File ID:	<u>VY023559.D</u>	Init. Calib. Date(s):	<u>10/07/2025 10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17 12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.760	1.991		13.13	20
m/p-Xylenes	0.706	0.816		15.58	20
o-Xylene	0.662	0.779		17.67	20
Styrene	1.089	1.285		18	20
Bromoform	0.240	0.278	0.1	15.83	20
Isopropylbenzene	3.295	3.754		13.93	20
1,1,2,2-Tetrachloroethane	0.539	0.600	0.3	11.32	20
1,3-Dichlorobenzene	1.706	1.896		11.14	20
1,4-Dichlorobenzene	1.685	1.876		11.34	20
1,2-Dichlorobenzene	1.496	1.689		12.9	20
1,2-Dibromo-3-Chloropropane	0.088	0.093		5.68	20
1,2,4-Trichlorobenzene	0.932	1.065		14.27	20
1,2,3-Trichlorobenzene	0.808	0.912		12.87	20
1,2-Dichloroethane-d4	0.418	0.374		-10.53	20
Dibromofluoromethane	0.307	0.304		-0.98	20
Toluene-d8	1.144	1.115		-2.54	20
4-Bromofluorobenzene	0.378	0.376		-0.53	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023559.D
 Acq On : 24 Oct 2025 08:56
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:45:30 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 05:12:50 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	558901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	809883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	712621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	380279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	208884	44.694	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	89.380%		
35) Dibromofluoromethane	7.634	113	246096	49.453	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.900%		
50) Toluene-d8	10.109	98	902938	48.722	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	97.440%		
62) 4-Bromofluorobenzene	12.407	95	304706	49.801	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	99.600%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.866	85	165459	41.129	ug/l	96
3) Chloromethane	2.074	50	219496	40.052	ug/l	99
4) Vinyl Chloride	2.208	62	301831	42.285	ug/l	100
5) Bromomethane	2.598	94	272400	45.475	ug/l	99
6) Chloroethane	2.738	64	217301	44.661	ug/l	99
7) Trichlorofluoromethane	3.061	101	471600	47.584	ug/l	100
8) Diethyl Ether	3.458	74	133898	51.703	ug/l	94
9) 1,1,2-Trichlorotrifluo...	3.817	101	263863	51.306	ug/l	98
10) Methyl Iodide	4.006	142	326535	57.032	ug/l	97
11) Tert butyl alcohol	4.854	59	102526	294.878	ug/l	97
12) 1,1-Dichloroethene	3.793	96	251789	50.504	ug/l	100
13) Acrolein	3.653	56	33834	68.867	ug/l	99
14) Allyl chloride	4.384	41	289064	47.575	ug/l	97
15) Acrylonitrile	5.061	53	256147	258.449	ug/l	99
16) Acetone	3.866	43	282132	253.570	ug/l	96
17) Carbon Disulfide	4.110	76	647806	45.399	ug/l	100
18) Methyl Acetate	4.384	43	156234	53.529	ug/l	97
19) Methyl tert-butyl Ether	5.122	73	663717	55.391	ug/l	94
20) Methylene Chloride	4.616	84	282698	46.987	ug/l	97
21) trans-1,2-Dichloroethene	5.116	96	285348	51.986	ug/l	99
22) Diisopropyl ether	6.018	45	690797	51.399	ug/l	96
23) Vinyl Acetate	5.963	43	1822354	250.506	ug/l	95
24) 1,1-Dichloroethane	5.921	63	454990	51.107	ug/l	99
25) 2-Butanone	6.896	43	346329	254.962	ug/l	98
26) 2,2-Dichloropropane	6.890	77	436383	53.419	ug/l	98
27) cis-1,2-Dichloroethene	6.890	96	348158	54.911	ug/l	95
28) Bromochloromethane	7.250	49	138645	41.717	ug/l	89
29) Tetrahydrofuran	7.262	42	193584	251.315	ug/l	93
30) Chloroform	7.420	83	521543	53.320	ug/l	100
31) Cyclohexane	7.707	56	360008	46.305	ug/l	96
32) 1,1,1-Trichloroethane	7.622	97	467412	52.584	ug/l	99
36) 1,1-Dichloropropene	7.835	75	345033	52.372	ug/l	99
37) Ethyl Acetate	6.981	43	131442	50.360	ug/l	96
38) Carbon Tetrachloride	7.823	117	427890	53.563	ug/l	99
39) Methylcyclohexane	9.109	83	454580	52.994	ug/l	99
40) Benzene	8.085	78	1117791	53.575	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023559.D
 Acq On : 24 Oct 2025 08:56
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:45:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	76239	54.008	ug/l	94
42) 1,2-Dichloroethane	8.158	62	273693	50.864	ug/l	97
43) Isopropyl Acetate	8.201	43	271189	52.107	ug/l #	85
44) Trichloroethene	8.865	130	334751	54.817	ug/l	98
45) 1,2-Dichloropropane	9.140	63	246074	53.397	ug/l	98
46) Dibromomethane	9.231	93	160236	54.988	ug/l	98
47) Bromodichloromethane	9.426	83	402796	54.883	ug/l	100
48) Methyl methacrylate	9.219	41	124097	50.755	ug/l	90
49) 1,4-Dioxane	9.231	88	38729	1146.304	ug/l	94
51) 4-Methyl-2-Pentanone	9.999	43	731786	272.142	ug/l	97
52) Toluene	10.170	92	751520	55.476	ug/l	100
53) t-1,3-Dichloropropene	10.395	75	353434	55.385	ug/l	98
54) cis-1,3-Dichloropropene	9.859	75	421103	55.911	ug/l	97
55) 1,1,2-Trichloroethane	10.572	97	222722	56.837	ug/l	98
56) Ethyl methacrylate	10.438	69	271697	60.107	ug/l	93
57) 1,3-Dichloropropane	10.719	76	346039	55.727	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	547857	270.556	ug/l	97
59) 2-Hexanone	10.761	43	528461	275.001	ug/l	95
60) Dibromochloromethane	10.914	129	315067	57.348	ug/l	100
61) 1,2-Dibromoethane	11.017	107	213108	57.024	ug/l	100
64) Tetrachloroethene	10.645	164	410235	55.753	ug/l	98
65) Chlorobenzene	11.444	112	872090	56.267	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.517	131	311946	56.479	ug/l	99
67) Ethyl Benzene	11.517	91	1418642	56.546	ug/l	99
68) m/p-Xylenes	11.627	106	1162639	115.537	ug/l	98
69) o-Xylene	11.956	106	555086	58.810	ug/l	97
70) Styrene	11.968	104	915552	58.986	ug/l	99
71) Bromoform	12.133	173	198287	57.877	ug/l #	100
73) Isopropylbenzene	12.255	105	1427449	56.959	ug/l	100
74) N-amyl acetate	12.072	43	240539	53.438	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	228057	55.639	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	168230m	50.142	ug/l	
77) Bromobenzene	12.529	156	372577	55.922	ug/l	98
78) n-propylbenzene	12.596	91	1644549	56.707	ug/l	99
79) 2-Chlorotoluene	12.682	91	947547	55.795	ug/l	99
80) 1,3,5-Trimethylbenzene	12.736	105	1157792	57.172	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	77712	56.799	ug/l	95
82) 4-Chlorotoluene	12.779	91	966872	55.281	ug/l	98
83) tert-Butylbenzene	12.993	119	1094487	58.535	ug/l	99
84) 1,2,4-Trimethylbenzene	13.041	105	1158715	57.333	ug/l	100
85) sec-Butylbenzene	13.175	105	1549057	57.753	ug/l	100
86) p-Isopropyltoluene	13.291	119	1337700	58.577	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	720888	55.564	ug/l	100
88) 1,4-Dichlorobenzene	13.364	146	713367	55.673	ug/l	100
89) n-Butylbenzene	13.614	91	1156866	58.138	ug/l	99
90) Hexachloroethane	13.877	117	266678	54.294	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	642214	56.461	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	35210	52.353	ug/l	94
93) 1,2,4-Trichlorobenzene	14.919	180	405098	57.127	ug/l	99
94) Hexachlorobutadiene	15.023	225	244033	55.207	ug/l	100
95) Naphthalene	15.145	128	662767	58.694	ug/l	100
96) 1,2,3-Trichlorobenzene	15.327	180	346789	56.460	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023559.D
 Acq On : 24 Oct 2025 08:56
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/soil
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_Y

Client Sample Id :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025


Quant Time: Oct 24 23:45:30 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Quant Title : SW846 8260

QLast Update : Wed Oct 08 05:12:50 2025


Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 55890
Ion Ratio Lower Upper
168 100
99 46.3 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#2
Dichlorodifluoromethane
Concen: 41.129 ug/l
RT: 1.866 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion: 85 Resp: 165459
Ion Ratio Lower Upper
85 100
87 34.0 16.0 47.9









#3
Chloromethane
Concen: 40.052 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 219490
Ion Ratio Lower Upper
50 100
52 33.5 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#4
Vinyl Chloride
Concen: 42.285 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion: 62 Resp: 301831
Ion Ratio Lower Upper
62 100
64 32.5 25.8 38.8









#5
Bromomethane
Concen: 45.475 ug/l
RT: 2.598 min Scan# 144
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 272400
Ion Ratio Lower Upper
94 100
96 93.9 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#6
Chloroethane
Concen: 44.661 ug/l
RT: 2.738 min Scan# 167
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion: 64 Resp: 217301
Ion Ratio Lower Upper
64 100
66 32.1 26.2 39.4









#7
Trichlorofluoromethane
Concen: 47.584 ug/l
RT: 3.061 min Scan# 219
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 471600
Ion Ratio Lower Upper
101 100
103 63.9 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#8
Diethyl Ether
Concen: 51.703 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion: 74 Resp: 133898
Ion Ratio Lower Upper
74 100
45 74.5 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.306 ug/l
RT: 3.817 min Scan# 343
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 26386
Ion Ratio Lower Upper
101 100
85 41.9 34.6 51.8
151 85.0 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#10
Methyl Iodide
Concen: 57.032 ug/l
RT: 4.006 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion:142 Resp: 326535
Ion Ratio Lower Upper
142 100
127 41.1 34.8 52.2
141 14.0 11.6 17.4





#11

Tert butyl alcohol

Concen: 294.878 ug/l

RT: 4.854 min Scan# 515

Delta R.T. -0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 102520

Ion Ratio Lower Upper

59 100





57 9.3 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#12

1,1-Dichloroethene

Concen: 50.504 ug/l

RT: 3.793 min Scan# 340

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 96 Resp: 251789


Ion Ratio Lower Upper

96 100

61 133.6 106.7 160.1

98 63.3 50.2 75.4





#13

Acrolein

Concen: 68.867 ug/l

RT: 3.653 min Scan# 316

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 33834

Ion Ratio Lower Upper

56 100





55 69.0 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#14

Allyl chloride

Concen: 47.575 ug/l

RT: 4.384 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 41 Resp: 289064


Ion Ratio Lower Upper

41 100

39 76.1 59.9 89.9


76 51.5 38.0 57.0





#15
Acrylonitrile
Concen: 258.449 ug/l
RT: 5.061 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 25614
Ion Ratio Lower Upper
53 100
52 81.3 64.6 97.0
51 36.4 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#16
Acetone
Concen: 253.570 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56



Tgt Ion: 43 Resp: 282132
Ion Ratio Lower Upper
43 100
58 37.0 27.7 41.5





#17

Carbon Disulfide

Concen: 45.399 ug/l

RT: 4.110 min Scan# 391

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 647800

Ion Ratio Lower Upper

76 100





78 9.0 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#18

Methyl Acetate

Concen: 53.529 ug/l

RT: 4.384 min Scan# 437

Delta R.T. -0.000 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion: 43 Resp: 156234

Ion Ratio Lower Upper

43 100

74 32.2 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 55.391 ug/l

RT: 5.122 min Scan# 5

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 66371

Ion Ratio Lower Upper

73 100





57 18.8 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#20

Methylene Chloride

Concen: 46.987 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Tgt Ion: 84 Resp: 282698




Ion Ratio Lower Upper


84 100

49 96.0 80.2 120.4

51 30.8 25.9 38.9

86 64.6 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 51.986 ug/l

RT: 5.116 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023559.D


Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050



Tgt Ion: 96 Resp: 28534

Ion Ratio Lower Upper

96 100

61 114.1 92.7 139.1




98 62.8 49.7 74.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#22

Diisopropyl ether


Concen: 51.399 ug/l

RT: 6.018 min Scan# 705

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56



Tgt Ion: 45 Resp: 690797



Ion Ratio Lower Upper


45 100

43 59.1 45.6 68.4

87 38.2 28.1 42.1

59 13.9 10.2 15.2





#23

Vinyl Acetate

Concen: 250.506 ug/l

RT: 5.963 min Scan# 695

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 1822354

Ion Ratio Lower Upper

43

100

86

15.1

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#24

1,1-Dichloroethane

Concen: 51.107 ug/l

RT: 5.921 min Scan# 689

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Tgt Ion: 63 Resp: 454990

Ion Ratio Lower Upper

63

100

98

8.2

4.3




13.1


100

5.7

2.8

8.4





#25

2-Butanone

Concen: 254.962 ug/l

RT: 6.896 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 346329

Ion Ratio Lower Upper

43 100





72 35.7 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#26

2,2-Dichloropropane

Concen: 53.419 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion: 77 Resp: 436383

Ion Ratio Lower Upper

77 100

97 25.0 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 54.911 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 348158

Ion Ratio Lower Upper

96 100

61 121.2 0.0 258.2





98 64.9 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#28

Bromochloromethane

Concen: 41.717 ug/l

RT: 7.250 min Scan# 907

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 49 Resp: 138645


Ion Ratio Lower Upper

49 100

129 2.9 0.0 5.6

130 113.9 81.9 122.9





#29

Tetrahydrofuran

Concen: 251.315 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 193584

Ion Ratio Lower Upper

42 100

72 59.3 43.0 64.6





71 54.8 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#30

Chloroform

Concen: 53.320 ug/l

RT: 7.420 min Scan# 935

Delta R.T. -0.000 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion: 83 Resp: 521543

Ion Ratio Lower Upper

83 100

85 65.5 52.5 78.7





#31

Cyclohexane

Concen: 46.305 ug/l

RT: 7.707 min Scan# 981

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 360000

Ion Ratio Lower Upper

56 100

69 37.4 30.0 45.0





84 106.6 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#32

1,1,1-Trichloroethane

Concen: 52.584 ug/l

RT: 7.622 min Scan# 968

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 97 Resp: 467412


Ion Ratio Lower Upper

97 100

99 64.5 51.4 77.2

61 37.9 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 44.694 ug/l

RT: 8.060 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 208884

Ion Ratio Lower Upper

65 100





67 54.2 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion:114 Resp: 809883


Ion Ratio Lower Upper

114 100

63 16.1 0.0 34.2

88 14.2 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.453 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 246090

Ion Ratio Lower Upper

113 100

111 103.3 81.8 122.6





192 23.7 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#36

1,1-Dichloropropene

Concen: 52.372 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 75 Resp: 345033


Ion Ratio Lower Upper

75 100

110 40.9 19.9 59.7

77 31.6 25.0 37.6





#37

Ethyl Acetate

Concen: 50.360 ug/l

RT: 6.981 min Scan# 80

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 13144

Ion Ratio Lower Upper

43 100

61 17.1 11.8 17.8





70 14.0 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#38

Carbon Tetrachloride

Concen: 53.563 ug/l

RT: 7.823 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion:117 Resp: 427890


Ion Ratio Lower Upper

117 100

119 96.6 76.8 115.2

121 30.1 24.5 36.7





#39

Methylcyclohexane

Concen: 52.994 ug/l

RT: 9.109 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 454580

Ion Ratio Lower Upper

83 100

55 63.7 51.4 77.2





98 50.6 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#40

Benzene

Concen: 53.575 ug/l

RT: 8.085 min Scan# 1044

Delta R.T. 0.006 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion: 78 Resp: 1117791

Ion Ratio Lower Upper

78 100

77 23.5 18.3 27.5





#41

Methacrylonitrile

Concen: 54.008 ug/l

RT: 7.225 min Scan# 902

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 41 Resp: 76239

Ion Ratio Lower Upper

41 100

39 53.8 52.2 78.4

67 95.9 74.9 112.3





52 35.4 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#42

1,2-Dichloroethane

Concen: 50.864 ug/l

RT: 8.158 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion: 62 Resp: 273693

Ion Ratio Lower Upper

62 100

98 11.8 0.0 21.6





#43

Isopropyl Acetate

Concen: 52.107 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 271189

Ion Ratio Lower Upper

43 100

61 34.4 18.1 27.1





87 18.0 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#44

Trichloroethene

Concen: 54.817 ug/l

RT: 8.865 min Scan# 1172

Delta R.T. -0.000 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion:130 Resp: 334751

Ion Ratio Lower Upper

130 100

95 92.6 0.0 180.6





#45

1,2-Dichloropropane

Concen: 53.397 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 246074

Ion Ratio Lower Upper

63 100





65 32.4 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#46

Dibromomethane

Concen: 54.988 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 93 Resp: 160236


Ion Ratio Lower Upper

93 100

95 84.3 65.4 98.0

174 118.2 92.9 139.3





#47

Bromodichloromethane

Concen: 54.883 ug/l

RT: 9.426 min Scan# 1263

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDC050

Tgt Ion: 83 Resp: 402790

Ion Ratio Lower Upper

83 100

85 63.3 50.8 76.2





127 9.1 7.4 11.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#48

Methyl methacrylate

Concen: 50.755 ug/l

RT: 9.219 min Scan# 1230

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion: 41 Resp: 124097


Ion Ratio Lower Upper

41 100

69 113.4 81.0 121.4

39 65.5 48.2 72.4





#49

1,4-Dioxane

Concen: 1146.304 ug/l

RT: 9.231 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	88	Resp:	38729
Ion	Ratio	Lower	Upper
88	100		
43	24.0	21.8	32.8
58	54.3	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#50

Toluene-d8

Concen: 48.722 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Tgt Ion:	98	Resp:	902938
Ion	Ratio	Lower	Upper
98	100		
100	65.4	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 272.142 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 731780
Ion Ratio Lower Upper
43 100
58 46.0 35.3 52.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#52
Toluene
Concen: 55.476 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56


Tgt Ion: 92 Resp: 751520
Ion Ratio Lower Upper
92 100
91 170.6 136.4 204.6





#53
t-1,3-Dichloropropene
Concen: 55.385 ug/l
RT: 10.395 min Scan# 1423
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 353434
Ion Ratio Lower Upper
75 100
77 32.4 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#54
cis-1,3-Dichloropropene
Concen: 55.911 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56


Tgt Ion: 75 Resp: 421103
Ion Ratio Lower Upper
75 100
77 31.7 25.4 38.0
39 37.2 32.8 49.2





#55
1,1,2-Trichloroethane
Concen: 56.837 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 22272
Ion Ratio Lower Upper
97 100
83 80.4 61.8 92.6
85 52.2 41.3 61.9
99 62.2 48.7 73.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#56
Ethyl methacrylate
Concen: 60.107 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56



Tgt Ion: 69 Resp: 271697
Ion Ratio Lower Upper
69 100
41 50.3 44.6 67.0
39 30.8 27.0 40.4





#57

1,3-Dichloropropane

Concen: 55.727 ug/l

RT: 10.719 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 346039

Ion Ratio Lower Upper

76 100





78 33.1 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#58

2-Chloroethyl Vinyl ether

Concen: 270.556 ug/l

RT: 9.713 min Scan# 1311

Delta R.T. 0.006 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion: 63 Resp: 547857

Ion Ratio Lower Upper

63 100

106 34.9 26.6 39.8





#59

2-Hexanone

Concen: 275.001 ug/l

RT: 10.761 min Scan# 1482

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 52846

Ion Ratio Lower Upper

43 100





58 64.6 30.5 91.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#60

Dibromochloromethane

Concen: 57.348 ug/l

RT: 10.914 min Scan# 1508

Delta R.T. 0.006 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion:129 Resp: 315067

Ion Ratio Lower Upper

129 100


127 77.5 38.6 115.7





#61
1,2-Dibromoethane
Concen: 57.024 ug/l
RT: 11.017 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 107 Resp: 213108
Ion Ratio Lower Upper
107 100
109 95.4 76.1 114.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#62
4-Bromofluorobenzene
Concen: 49.801 ug/l
RT: 12.407 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Tgt Ion: 95 Resp: 304706
Ion Ratio Lower Upper
95 100
174 99.5 0.0 192.8
176 96.3 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 71262
Ion Ratio Lower Upper
117 100
82 51.2 41.0 61.4
119 32.3 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#64
Tetrachloroethene
Concen: 55.753 ug/l
RT: 10.645 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56



Tgt Ion:164 Resp: 410235
Ion Ratio Lower Upper
164 100
166 126.5 99.4 149.0
129 89.5 69.4 104.2
131 84.4 67.5 101.3





#65

Chlorobenzene

Concen: 56.267 ug/l

RT: 11.444 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:112 Resp: 872090

Ion Ratio Lower Upper

112 100





114 33.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 56.479 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion:131 Resp: 311946


Ion Ratio Lower Upper

131 100

133 95.3 48.4 145.2

119 65.6 32.4 97.2









#67
Ethyl Benzene
Concen: 56.546 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 141864
Ion Ratio Lower Upper
91 100
106 32.8 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#68
m/p-Xylenes
Concen: 115.537 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion:106 Resp: 1162639
Ion Ratio Lower Upper
106 100
91 188.2 152.6 229.0









#69
o-Xylene
Concen: 58.810 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 555080
Ion Ratio Lower Upper
106 100
91 198.6 101.4 304.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#70
Styrene
Concen: 58.986 ug/l
RT: 11.968 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56


Tgt Ion:104 Resp: 915552
Ion Ratio Lower Upper
104 100
78 44.1 36.1 54.1
103 54.3 43.2 64.8





#71
Bromoform
Concen: 57.877 ug/l
RT: 12.133 min Scan# 1707
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 198285
Ion Ratio Lower Upper
173 100
175 49.5 24.8 74.3
254 0.2 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Tgt Ion:152 Resp: 380279
Ion Ratio Lower Upper
152 100
115 53.8 27.1 81.2
150 177.0 0.0 347.4





#73
Isopropylbenzene
Concen: 56.959 ug/l
RT: 12.255 min Scan# 1727
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 1427449
Ion Ratio Lower Upper
105 100
120 27.9 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#74
N-ethyl acetate
Concen: 53.438 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion: 43 Resp: 240539
Ion Ratio Lower Upper
43 100
70 55.3 41.1 61.7
55 28.4 23.0 34.6
61 29.2 21.4 32.2





#75

1,1,2,2-Tetrachloroethane

Concen: 55.639 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	83	Resp:	22805
Ion Ratio	Lower	Upper	
83	100		
131	11.5	5.9	17.6
85	65.0	31.8	95.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#76

1,2,3-Trichloropropane

Concen: 50.142 ug/l m




RT: 12.554 min Scan# 1777


Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Tgt Ion:	75	Resp:	168230
Ion Ratio	Lower	Upper	
75	100		
77	0.0	0.0	0.0





#77

Bromobenzene

Concen: 55.922 ug/l

RT: 12.529 min Scan# 1773

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:156 Resp: 37257

Ion	Ratio	Lower	Upper
156	100		
77	150.5	76.8	230.5
158	97.8	49.0	147.2

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#78

n-propylbenzene

Concen: 56.707 ug/l

RT: 12.596 min Scan# 1784




Delta R.T. 0.006 min


Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Tgt Ion: 91 Resp: 1644549

Ion	Ratio	Lower	Upper
91	100		
120	25.2	12.3	36.8









#79
2-Chlorotoluene
Concen: 55.795 ug/l
RT: 12.682 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 94754
Ion Ratio Lower Upper
91 100
126 37.4 18.5 55.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#80
1,3,5-Trimethylbenzene
Concen: 57.172 ug/l
RT: 12.736 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56


Tgt Ion:105 Resp: 1157792
Ion Ratio Lower Upper
105 100
120 52.0 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 56.799 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 77712
Ion Ratio Lower Upper
75 100
53 80.3 68.6 102.8
89 47.1 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#82
4-Chlorotoluene
Concen: 55.281 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion: 91 Resp: 966872
Ion Ratio Lower Upper
91 100
126 38.3 18.4 55.4





#83

tert-Butylbenzene

Concen: 58.535 ug/l

RT: 12.993 min Scan# 1849

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 109448

Ion Ratio Lower Upper

119 100

91 59.2 29.9 89.8





134 27.0 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#84

1,2,4-Trimethylbenzene

Concen: 57.333 ug/l

RT: 13.041 min Scan# 1857

Delta R.T. -0.000 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion:105 Resp: 1158715

Ion Ratio Lower Upper

105 100


120 47.6 23.8 71.2





#85
sec-Butylbenzene
Concen: 57.753 ug/l
RT: 13.175 min Scan# 1878
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 154905
Ion Ratio Lower Upper
105 100
134 21.5 10.7 32.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#86
p-Isopropyltoluene
Concen: 58.577 ug/l
RT: 13.291 min Scan# 1898
Delta R.T. 0.006 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56

Tgt Ion:119 Resp: 1337700
Ion Ratio Lower Upper
119 100
134 27.2 13.6 40.7
91 21.2 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 55.564 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 720888

Ion Ratio Lower Upper

146 100

111 37.0 18.7 56.1





148 63.9 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#88

1,4-Dichlorobenzene

Concen: 55.673 ug/l

RT: 13.364 min Scan# 1910

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion:146 Resp: 713367


Ion Ratio Lower Upper

146 100

111 35.6 18.0 54.0

148 63.7 31.9 95.7





#89

n-Butylbenzene

Concen: 58.138 ug/l

RT: 13.614 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 115686

Ion Ratio Lower Upper

91 100

92 53.6 27.0 80.8





134 29.0 14.1 42.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#90

Hexachloroethane

Concen: 54.294 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. -0.000 min

Lab File: VY023559.D




Acq: 24 Oct 2025 08:56


Tgt Ion:117 Resp: 266678

Ion Ratio Lower Upper

117 100


201 89.0 43.8 131.4





#91
1,2-Dichlorobenzene
Concen: 56.461 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 642214
Ion Ratio Lower Upper
146 100
111 37.4 19.0 57.0
148 64.1 32.1 96.3

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 52.353 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023559.D
Acq: 24 Oct 2025 08:56



Tgt Ion: 75 Resp: 35210
Ion Ratio Lower Upper
75 100
155 105.8 50.0 150.0
157 134.6 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 57.127 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 405098

Ion Ratio Lower Upper

180 100

182 94.7 47.9 143.6





145 28.3 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#94

Hexachlorobutadiene

Concen: 55.207 ug/l

RT: 15.023 min Scan# 2182

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion:225 Resp: 244033


Ion Ratio Lower Upper

225 100

223 61.9 30.8 92.3

227 63.7 31.9 95.7





#95

Naphthalene

Concen: 58.694 ug/l

RT: 15.145 min Scan# 2201

Delta R.T. 0.006 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:128 Resp: 66276

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6





129 10.9 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#96

1,2,3-Trichlorobenzene

Concen: 56.460 ug/l

RT: 15.327 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023559.D

Acq: 24 Oct 2025 08:56




Tgt Ion:180 Resp: 346789

Ion Ratio Lower Upper

180 100

182 95.1 47.9 143.8

145 30.1 15.3 45.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.360	0.296	17.8	87	0.00
3 P	Chloromethane	0.490	0.393	19.8	83	0.00
4 C	Vinyl Chloride	0.639	0.540	15.5#	86	0.00
5 T	Bromomethane	0.536	0.487	9.1	98	0.00
6 T	Chloroethane	0.435	0.389	10.6	89	0.00
7 T	Trichlorofluoromethane	0.887	0.844	4.8	95	0.00
8 T	Diethyl Ether	0.232	0.240	-3.4	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.472	-2.6	102	0.00
10 T	Methyl Iodide	0.512	0.584	-14.1	103	0.00
11 T	Tert butyl alcohol	0.031	0.037	-19.4	139	0.00
12 CM	1,1-Dichloroethene	0.446	0.451	-1.1#	100	0.00
13 T	Acrolein	0.044	0.012	72.7#	27#	0.00
14 T	Allyl chloride	0.544	0.517	5.0	93	0.00
15 T	Acrylonitrile	0.089	0.092	-3.4	106	0.00
16 T	Acetone	0.100	0.101	-1.0	114	0.00
17 T	Carbon Disulfide	1.277	1.159	9.2	90	0.00
18 T	Methyl Acetate	0.261	0.280	-7.3	108	0.00
19 T	Methyl tert-butyl Ether	1.072	1.188	-10.8	110	0.00
20 T	Methylene Chloride	0.538	0.506	5.9	104	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.511	-4.1	103	0.00
22 T	Diisopropyl ether	1.202	1.236	-2.8	101	0.00
23 T	Vinyl Acetate	0.651	0.652	-0.2	99	0.00
24 P	1,1-Dichloroethane	0.796	0.814	-2.3	101	0.00
25 T	2-Butanone	0.122	0.124	-1.6	112	0.00
26 T	2,2-Dichloropropane	0.731	0.781	-6.8	105	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.623	-9.9	109	0.00
28 T	Bromochloromethane	0.297	0.248	16.5	88	0.00
29 T	Tetrahydrofuran	0.069	0.069	0.0	103	0.00
30 C	Chloroform	0.875	0.933	-6.6#	106	0.00
31 T	Cyclohexane	0.696	0.644	7.5	95	0.00
32 T	1,1,1-Trichloroethane	0.795	0.836	-5.2	103	0.00
33 S	1,2-Dichloroethane-d4	0.418	0.374	10.5	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.307	0.304	1.0	96	0.00
36 T	1,1-Dichloropropene	0.407	0.426	-4.7	101	0.00
37 T	Ethyl Acetate	0.161	0.162	-0.6	102	0.00
38 T	Carbon Tetrachloride	0.493	0.528	-7.1	104	0.00
39 T	Methylcyclohexane	0.530	0.561	-5.8	100	0.00
40 TM	Benzene	1.288	1.380	-7.1	103	0.00
41 T	Methacrylonitrile	0.087	0.094	-8.0	111	0.00
42 TM	1,2-Dichloroethane	0.332	0.338	-1.8	101	0.00
43 T	Isopropyl Acetate	0.321	0.335	-4.4	105	0.00
44 TM	Trichloroethene	0.377	0.413	-9.5	106	0.00
45 C	1,2-Dichloropropane	0.285	0.304	-6.7#	105	0.00
46 T	Dibromomethane	0.180	0.198	-10.0	108	0.00
47 T	Bromodichloromethane	0.453	0.497	-9.7	107	0.00
48 T	Methyl methacrylate	0.151	0.153	-1.3	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	114	0.00
50 S	Toluene-d8	1.144	1.115	2.5	93	0.00
51 T	4-Methyl-2-Pentanone	0.166	0.181	-9.0	107	0.00
52 CM	Toluene	0.836	0.928	-11.0#	106	0.00
53 T	t-1,3-Dichloropropene	0.394	0.436	-10.7	107	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.520	-11.8	108	0.00
55 T	1,1,2-Trichloroethane	0.242	0.275	-13.6	112	0.00
56 T	Ethyl methacrylate	0.279	0.335	-20.1	113	0.00
57 T	1,3-Dichloropropane	0.383	0.427	-11.5	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.135	-8.0	99	0.00
59 T	2-Hexanone	0.119	0.131	-10.1	110	0.00
60 T	Dibromochloromethane	0.339	0.389	-14.7	112	0.00
61 T	1,2-Dibromoethane	0.231	0.263	-13.9	111	0.00
62 S	4-Bromofluorobenzene	0.378	0.376	0.5	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.516	0.576	-11.6	105	0.00
65 PM	Chlorobenzene	1.087	1.224	-12.6	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.438	-12.9	112	0.00
67 C	Ethyl Benzene	1.760	1.991	-13.1#	106	0.00
68 T	m/p-Xylenes	0.706	0.816	-15.6	108	0.00
69 T	o-Xylene	0.662	0.779	-17.7	110	0.00
70 T	Styrene	1.089	1.285	-18.0	110	0.00
71 P	Bromoform	0.240	0.278	-15.8	115	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.295	3.754	-13.9	110	0.00
74 T	N-amyl acetate	0.592	0.633	-6.9	106	0.00
75 P	1,1,2,2-Tetrachloroethane	0.539	0.600	-11.3	117	0.00
76 T	1,2,3-Trichloropropane	0.441	0.442	-0.2	99	0.00
77 T	Bromobenzene	0.876	0.980	-11.9	112	0.00
78 T	n-propylbenzene	3.813	4.325	-13.4	108	0.00
79 T	2-Chlorotoluene	2.233	2.492	-11.6	108	0.00
80 T	1,3,5-Trimethylbenzene	2.663	3.045	-14.3	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.204	-13.3	117	0.00
82 T	4-Chlorotoluene	2.300	2.543	-10.6	108	0.00
83 T	tert-Butylbenzene	2.458	2.878	-17.1	111	0.00
84 T	1,2,4-Trimethylbenzene	2.657	3.047	-14.7	109	0.00
85 T	sec-Butylbenzene	3.527	4.073	-15.5	110	0.00
86 T	p-Isopropyltoluene	3.003	3.518	-17.1	110	0.00
87 T	1,3-Dichlorobenzene	1.706	1.896	-11.1	111	0.00
88 T	1,4-Dichlorobenzene	1.685	1.876	-11.3	112	0.00
89 T	n-Butylbenzene	2.616	3.042	-16.3	110	0.00
90 T	Hexachloroethane	0.646	0.701	-8.5	109	0.00
91 T	1,2-Dichlorobenzene	1.496	1.689	-12.9	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.093	-5.7	108	0.00
93 T	1,2,4-Trichlorobenzene	0.932	1.065	-14.3	113	0.00
94 T	Hexachlorobutadiene	0.581	0.642	-10.5	108	0.00
95 T	Naphthalene	1.485	1.743	-17.4	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.808	0.912	-12.9	112	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	41.129	17.7	87	0.00
3 P	Chloromethane	50.000	40.052	19.9	83	0.00
4 C	Vinyl Chloride	50.000	42.285	15.4#	86	0.00
5 T	Bromomethane	50.000	45.475	9.0	98	0.00
6 T	Chloroethane	50.000	44.661	10.7	89	0.00
7 T	Trichlorofluoromethane	50.000	47.584	4.8	95	0.00
8 T	Diethyl Ether	50.000	51.703	-3.4	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.306	-2.6	102	0.00
10 T	Methyl Iodide	50.000	57.032	-14.1	103	0.00
11 T	Tert butyl alcohol	250.000	294.878	-18.0	139	0.00
12 CM	1,1-Dichloroethene	50.000	50.504	-1.0#	100	0.00
13 T	Acrolein	250.000	68.867	72.5#	27	0.00
14 T	Allyl chloride	50.000	47.575	4.8	93	0.00
15 T	Acrylonitrile	250.000	258.449	-3.4	106	0.00
16 T	Acetone	250.000	253.570	-1.4	114	0.00
17 T	Carbon Disulfide	50.000	45.399	9.2	90	0.00
18 T	Methyl Acetate	50.000	53.529	-7.1	108	0.00
19 T	Methyl tert-butyl Ether	50.000	55.391	-10.8	110	0.00
20 T	Methylene Chloride	50.000	46.987	6.0	104	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.986	-4.0	103	0.00
22 T	Diisopropyl ether	50.000	51.399	-2.8	101	0.00
23 T	Vinyl Acetate	250.000	250.506	-0.2	99	0.00
24 P	1,1-Dichloroethane	50.000	51.107	-2.2	101	0.00
25 T	2-Butanone	250.000	254.962	-2.0	112	0.00
26 T	2,2-Dichloropropane	50.000	53.419	-6.8	105	0.00
27 T	cis-1,2-Dichloroethene	50.000	54.911	-9.8	109	0.00
28 T	Bromochloromethane	50.000	41.717	16.6	88	0.00
29 T	Tetrahydrofuran	250.000	251.315	-0.5	103	0.00
30 C	Chloroform	50.000	53.320	-6.6#	106	0.00
31 T	Cyclohexane	50.000	46.305	7.4	95	0.00
32 T	1,1,1-Trichloroethane	50.000	52.584	-5.2	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	44.694	10.6	90	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	49.453	1.1	96	0.00
36 T	1,1-Dichloropropene	50.000	52.372	-4.7	101	0.00
37 T	Ethyl Acetate	50.000	50.360	-0.7	102	0.00
38 T	Carbon Tetrachloride	50.000	53.563	-7.1	104	0.00
39 T	Methylcyclohexane	50.000	52.994	-6.0	100	0.00
40 TM	Benzene	50.000	53.575	-7.2	103	0.00
41 T	Methacrylonitrile	50.000	54.008	-8.0	111	0.00
42 TM	1,2-Dichloroethane	50.000	50.864	-1.7	101	0.00
43 T	Isopropyl Acetate	50.000	52.107	-4.2	105	0.00
44 TM	Trichloroethene	50.000	54.817	-9.6	106	0.00
45 C	1,2-Dichloropropane	50.000	53.397	-6.8#	105	0.00
46 T	Dibromomethane	50.000	54.988	-10.0	108	0.00
47 T	Bromodichloromethane	50.000	54.883	-9.8	107	0.00
48 T	Methyl methacrylate	50.000	50.755	-1.5	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1146.304	-14.6	114	0.00
50 S	Toluene-d8	50.000	48.722	2.6	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	272.142	-8.9	107	0.00
52 CM	Toluene	50.000	55.476	-11.0#	106	0.00
53 T	t-1,3-Dichloropropene	50.000	55.385	-10.8	107	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.911	-11.8	108	0.00
55 T	1,1,2-Trichloroethane	50.000	56.837	-13.7	112	0.00
56 T	Ethyl methacrylate	50.000	60.107	-20.2	113	0.00
57 T	1,3-Dichloropropane	50.000	55.727	-11.5	108	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	270.556	-8.2	99	0.00
59 T	2-Hexanone	250.000	275.001	-10.0	110	0.00
60 T	Dibromochloromethane	50.000	57.348	-14.7	112	0.00
61 T	1,2-Dibromoethane	50.000	57.024	-14.0	111	0.00
62 S	4-Bromofluorobenzene	50.000	49.801	0.4	97	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	55.753	-11.5	105	0.00
65 PM	Chlorobenzene	50.000	56.267	-12.5	110	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	56.479	-13.0	112	0.00
67 C	Ethyl Benzene	50.000	56.546	-13.1#	106	0.00
68 T	m/p-Xylenes	100.000	115.537	-15.5	108	0.00
69 T	o-Xylene	50.000	58.810	-17.6	110	0.00
70 T	Styrene	50.000	58.986	-18.0	110	0.00
71 P	Bromoform	50.000	57.877	-15.8	115	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	97	0.00
73 T	Isopropylbenzene	50.000	56.959	-13.9	110	0.00
74 T	N-amyl acetate	50.000	53.438	-6.9	106	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	55.639	-11.3	117	0.00
76 T	1,2,3-Trichloropropane	50.000	50.142	-0.3	99	0.00
77 T	Bromobenzene	50.000	55.922	-11.8	112	0.00
78 T	n-propylbenzene	50.000	56.707	-13.4	108	0.00
79 T	2-Chlorotoluene	50.000	55.795	-11.6	108	0.00
80 T	1,3,5-Trimethylbenzene	50.000	57.172	-14.3	109	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	56.799	-13.6	117	0.00
82 T	4-Chlorotoluene	50.000	55.281	-10.6	108	0.00
83 T	tert-Butylbenzene	50.000	58.535	-17.1	111	0.00
84 T	1,2,4-Trimethylbenzene	50.000	57.333	-14.7	109	0.00
85 T	sec-Butylbenzene	50.000	57.753	-15.5	110	0.00
86 T	p-Isopropyltoluene	50.000	58.577	-17.2	110	0.00
87 T	1,3-Dichlorobenzene	50.000	55.564	-11.1	111	0.00
88 T	1,4-Dichlorobenzene	50.000	55.673	-11.3	112	0.00
89 T	n-Butylbenzene	50.000	58.138	-16.3	110	0.00
90 T	Hexachloroethane	50.000	54.294	-8.6	109	0.00
91 T	1,2-Dichlorobenzene	50.000	56.461	-12.9	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.353	-4.7	108	0.00
93 T	1,2,4-Trichlorobenzene	50.000	57.127	-14.3	113	0.00
94 T	Hexachlorobutadiene	50.000	55.207	-10.4	108	0.00
95 T	Naphthalene	50.000	58.694	-17.4	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023559.D
Acq On : 24 Oct 2025 08:56
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 23:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	56.460	-12.9	112	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100725\
Data File : VY023383.D
Acq On : 07 Oct 2025 08:43
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260
Last Update : Wed Oct 08 05:12:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743



Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.9	34816	PASS
75	95	30	60	48.1	99264	PASS
95	95	100	100	100.0	206507	PASS
96	95	5	9	6.6	13540	PASS
173	174	0.00	2	0.9	1817	PASS
174	95	50	100	99.1	204629	PASS
175	174	5	9	7.4	15059	PASS
176	174	95	101	96.2	196949	PASS
177	176	5	9	6.6	12912	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023542.D
Acq On : 22 Oct 2025 09:00
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260
Last Update : Wed Oct 08 05:12:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743



Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.8	43632	PASS
75	95	30	60	47.5	130976	PASS
95	95	100	100	100.0	275756	PASS
96	95	5	9	6.6	18234	PASS
173	174	0.00	2	1.0	2756	PASS
174	95	50	100	99.3	273941	PASS
175	174	5	9	7.2	19793	PASS
176	174	95	101	98.2	268885	PASS
177	176	5	9	6.4	17309	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023558.D
Acq On : 24 Oct 2025 07:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260
Last Update : Wed Oct 08 05:12:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.7	28941	PASS
75	95	30	60	47.4	87403	PASS
95	95	100	100	100.0	184491	PASS
96	95	5	9	6.5	11916	PASS
173	174	0.00	2	1.1	1956	PASS
174	95	50	100	98.2	181248	PASS
175	174	5	9	7.4	13455	PASS
176	174	95	101	98.0	177600	PASS
177	176	5	9	6.6	11721	PASS

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1022SBL01
 Lab Sample ID: VY1022SBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	5.00	U	1	1.10	5.00	ug/Kg	10/22/25 10:00	VY102225
74-87-3	Chloromethane	5.00	U	1	1.10	5.00	ug/Kg	10/22/25 10:00	VY102225
75-01-4	Vinyl Chloride	5.00	U	1	0.79	5.00	ug/Kg	10/22/25 10:00	VY102225
74-83-9	Bromomethane	5.00	U	1	1.10	5.00	ug/Kg	10/22/25 10:00	VY102225
75-00-3	Chloroethane	5.00	U	1	1.30	5.00	ug/Kg	10/22/25 10:00	VY102225
75-69-4	Trichlorofluoromethane	5.00	U	1	1.20	5.00	ug/Kg	10/22/25 10:00	VY102225
76-13-1	1,1,2-Trichlorotrifluoroethane	5.00	U	1	1.10	5.00	ug/Kg	10/22/25 10:00	VY102225
75-35-4	1,1-Dichloroethene	5.00	U	1	1.00	5.00	ug/Kg	10/22/25 10:00	VY102225
67-64-1	Acetone	25.0	U	1	4.70	25.0	ug/Kg	10/22/25 10:00	VY102225
75-15-0	Carbon Disulfide	5.00	U	1	1.10	5.00	ug/Kg	10/22/25 10:00	VY102225
1634-04-4	Methyl tert-butyl Ether	5.00	U	1	0.73	5.00	ug/Kg	10/22/25 10:00	VY102225
79-20-9	Methyl Acetate	5.00	U	1	1.50	5.00	ug/Kg	10/22/25 10:00	VY102225
75-09-2	Methylene Chloride	10.0	U	1	3.50	10.0	ug/Kg	10/22/25 10:00	VY102225
156-60-5	trans-1,2-Dichloroethene	5.00	U	1	0.86	5.00	ug/Kg	10/22/25 10:00	VY102225
75-34-3	1,1-Dichloroethane	5.00	U	1	0.80	5.00	ug/Kg	10/22/25 10:00	VY102225
110-82-7	Cyclohexane	5.00	U	1	0.79	5.00	ug/Kg	10/22/25 10:00	VY102225
78-93-3	2-Butanone	25.0	U	1	6.50	25.0	ug/Kg	10/22/25 10:00	VY102225
56-23-5	Carbon Tetrachloride	5.00	U	1	0.97	5.00	ug/Kg	10/22/25 10:00	VY102225
156-59-2	cis-1,2-Dichloroethene	5.00	U	1	0.75	5.00	ug/Kg	10/22/25 10:00	VY102225
74-97-5	Bromochloromethane	5.00	U	1	1.20	5.00	ug/Kg	10/22/25 10:00	VY102225
67-66-3	Chloroform	5.00	U	1	0.84	5.00	ug/Kg	10/22/25 10:00	VY102225
71-55-6	1,1,1-Trichloroethane	5.00	U	1	0.93	5.00	ug/Kg	10/22/25 10:00	VY102225
108-87-2	Methylcyclohexane	5.00	U	1	0.91	5.00	ug/Kg	10/22/25 10:00	VY102225
71-43-2	Benzene	5.00	U	1	0.79	5.00	ug/Kg	10/22/25 10:00	VY102225
107-06-2	1,2-Dichloroethane	5.00	U	1	0.79	5.00	ug/Kg	10/22/25 10:00	VY102225
79-01-6	Trichloroethene	5.00	U	1	0.81	5.00	ug/Kg	10/22/25 10:00	VY102225
78-87-5	1,2-Dichloropropane	5.00	U	1	0.91	5.00	ug/Kg	10/22/25 10:00	VY102225
75-27-4	Bromodichloromethane	5.00	U	1	0.78	5.00	ug/Kg	10/22/25 10:00	VY102225
108-10-1	4-Methyl-2-Pentanone	25.0	U	1	3.60	25.0	ug/Kg	10/22/25 10:00	VY102225
108-88-3	Toluene	5.00	U	1	0.78	5.00	ug/Kg	10/22/25 10:00	VY102225
10061-02-6	t-1,3-Dichloropropene	5.00	U	1	0.65	5.00	ug/Kg	10/22/25 10:00	VY102225
10061-01-5	cis-1,3-Dichloropropene	5.00	U	1	0.62	5.00	ug/Kg	10/22/25 10:00	VY102225
79-00-5	1,1,2-Trichloroethane	5.00	U	1	0.92	5.00	ug/Kg	10/22/25 10:00	VY102225
591-78-6	2-Hexanone	25.0	U	1	3.70	25.0	ug/Kg	10/22/25 10:00	VY102225
124-48-1	Dibromochloromethane	5.00	U	1	0.87	5.00	ug/Kg	10/22/25 10:00	VY102225
106-93-4	1,2-Dibromoethane	5.00	U	1	0.88	5.00	ug/Kg	10/22/25 10:00	VY102225
127-18-4	Tetrachloroethene	5.00	U	1	1.10	5.00	ug/Kg	10/22/25 10:00	VY102225
108-90-7	Chlorobenzene	5.00	U	1	0.91	5.00	ug/Kg	10/22/25 10:00	VY102225
100-41-4	Ethyl Benzene	5.00	U	1	0.67	5.00	ug/Kg	10/22/25 10:00	VY102225
179601-23-1	m/p-Xylenes	10.0	U	1	1.20	10.0	ug/Kg	10/22/25 10:00	VY102225
95-47-6	o-Xylene	5.00	U	1	0.82	5.00	ug/Kg	10/22/25 10:00	VY102225
100-42-5	Styrene	5.00	U	1	0.71	5.00	ug/Kg	10/22/25 10:00	VY102225
75-25-2	Bromoform	5.00	U	1	0.86	5.00	ug/Kg	10/22/25 10:00	VY102225

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1022SBL01
 Lab Sample ID: VY1022SBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	5.00	U	1	0.78	5.00	ug/Kg	10/22/25 10:00	VY102225
79-34-5	1,1,2,2-Tetrachloroethane	5.00	U	1	1.20	5.00	ug/Kg	10/22/25 10:00	VY102225
541-73-1	1,3-Dichlorobenzene	5.00	U	1	1.70	5.00	ug/Kg	10/22/25 10:00	VY102225
106-46-7	1,4-Dichlorobenzene	5.00	U	1	1.60	5.00	ug/Kg	10/22/25 10:00	VY102225
95-50-1	1,2-Dichlorobenzene	5.00	U	1	1.50	5.00	ug/Kg	10/22/25 10:00	VY102225
96-12-8	1,2-Dibromo-3-Chloropropane	5.00	U	1	1.80	5.00	ug/Kg	10/22/25 10:00	VY102225
120-82-1	1,2,4-Trichlorobenzene	5.00	U	1	3.00	5.00	ug/Kg	10/22/25 10:00	VY102225
87-61-6	1,2,3-Trichlorobenzene	5.00	U	1	3.20	5.00	ug/Kg	10/22/25 10:00	VY102225
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	48.3			63 - 155	97%	SPK: 50		
1868-53-7	Dibromofluoromethane	49.8			70 - 134	100%	SPK: 50		
2037-26-5	Toluene-d8	48.6			74 - 123	97%	SPK: 50		
460-00-4	4-Bromofluorobenzene	41.1			17 - 146	82%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	890000							
540-36-3	1,4-Difluorobenzene	1430000							
3114-55-4	Chlorobenzene-d5	1180000							
3855-82-1	1,4-Dichlorobenzene-d4	485000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023544.D
Acq On : 22 Oct 2025 10:00
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Quant Time: Oct 24 02:26:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	890059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1434768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1180185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	484557	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	359226	48.265	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.520%
35) Dibromofluoromethane	7.640	113	439107	49.808	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.620%
50) Toluene-d8	10.109	98	1595601	48.600	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.200%
62) 4-Bromofluorobenzene	12.408	95	445110	41.064	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.120%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023544.D
Acq On : 22 Oct 2025 10:00
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Quant Time: Oct 24 02:26:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023544.D
Acq: 22 Oct 2025 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01


Tgt Ion:168 Resp: 890059
Ion Ratio Lower Upper
168 100
99 47.3 39.7 59.5



#33
1,2-Dichloroethane-d4
Concen: 48.265 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY023544.D
Acq: 22 Oct 2025 10:00

Tgt Ion: 65 Resp: 359226
Ion Ratio Lower Upper
65 100
67 53.5 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023544.D

Acq: 22 Oct 2025 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBL01





Tgt Ion:114 Resp: 1434768

Ion Ratio Lower Upper

114 100

63 16.2 0.0 34.2

88 14.5 0.0 28.4



#35

Dibromofluoromethane

Concen: 49.808 ug/l

RT: 7.640 min Scan# 971

Delta R.T. 0.006 min

Lab File: VY023544.D

Acq: 22 Oct 2025 10:00




Tgt Ion:113 Resp: 439107


Ion Ratio Lower Upper

113 100

111 101.6 81.8 122.6

192 22.9 18.0 27.0









#50
Toluene-d8
Concen: 48.600 ug/l
RT: 10.109 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023544.D
Acq: 22 Oct 2025 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01


Tgt Ion: 98 Resp: 1595601
Ion Ratio Lower Upper
98 100
100 65.5 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 41.064 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023544.D
Acq: 22 Oct 2025 10:00

Tgt Ion: 95 Resp: 445110
Ion Ratio Lower Upper
95 100
174 97.9 0.0 192.8
176 94.7 0.0 187.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023544.D

Acq: 22 Oct 2025 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBL01





Tgt Ion:117 Resp: 1180185

Ion Ratio Lower Upper

117 100

82 49.3 41.0 61.4

119 32.0 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY023544.D

Acq: 22 Oct 2025 10:00




Tgt Ion:152 Resp: 484557

Ion Ratio Lower Upper

152 100

115 53.2 27.1 81.2

150 154.9 0.0 347.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023544.D
Acq On : 22 Oct 2025 10:00
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Title : SW846 8260

Signal : TIC: VY023544.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	6.890	836	848	860	rBV	52729	169844	4.22%	0.796%
2	7.640	958	971	976	rBV	613566	1444748	35.92%	6.772%
3	7.713	976	983	1000	rVB	1030611	2407906	59.87%	11.287%
4	8.061	1027	1040	1054	rBV2	508015	1257916	31.28%	5.897%
5	8.616	1122	1131	1151	rBV	1561233	3049586	75.83%	14.295%
6	10.109	1368	1376	1390	rBV	2337515	4021836	100.00%	18.853%
7	10.512	1434	1442	1459	rVB	188628	345760	8.60%	1.621%
8	10.877	1496	1502	1513	rBV	104551	178963	4.45%	0.839%
9	11.420	1582	1591	1603	rBV	2023505	3338390	83.01%	15.649%
10	12.408	1746	1753	1766	rBV	1567478	2435469	60.56%	11.416%
11	13.346	1901	1907	1921	rVB	1791615	2682609	66.70%	12.575%


Sum of corrected areas: 21333027

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023544.D
Acq On : 22 Oct 2025 10:00
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023544.D
Acq On : 22 Oct 2025 10:00
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023544.D
Acq On : 22 Oct 2025 10:00
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1024SBL01
 Lab Sample ID: VY1024SBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	5.00	U	1	1.10	5.00	ug/Kg	10/24/25 09:24	VY102425
74-87-3	Chloromethane	5.00	U	1	1.10	5.00	ug/Kg	10/24/25 09:24	VY102425
75-01-4	Vinyl Chloride	5.00	U	1	0.79	5.00	ug/Kg	10/24/25 09:24	VY102425
74-83-9	Bromomethane	5.00	U	1	1.10	5.00	ug/Kg	10/24/25 09:24	VY102425
75-00-3	Chloroethane	5.00	U	1	1.30	5.00	ug/Kg	10/24/25 09:24	VY102425
75-69-4	Trichlorofluoromethane	5.00	U	1	1.20	5.00	ug/Kg	10/24/25 09:24	VY102425
76-13-1	1,1,2-Trichlorotrifluoroethane	5.00	U	1	1.10	5.00	ug/Kg	10/24/25 09:24	VY102425
75-35-4	1,1-Dichloroethene	5.00	U	1	1.00	5.00	ug/Kg	10/24/25 09:24	VY102425
67-64-1	Acetone	25.0	U	1	4.70	25.0	ug/Kg	10/24/25 09:24	VY102425
75-15-0	Carbon Disulfide	5.00	U	1	1.10	5.00	ug/Kg	10/24/25 09:24	VY102425
1634-04-4	Methyl tert-butyl Ether	5.00	U	1	0.73	5.00	ug/Kg	10/24/25 09:24	VY102425
79-20-9	Methyl Acetate	5.00	U	1	1.50	5.00	ug/Kg	10/24/25 09:24	VY102425
75-09-2	Methylene Chloride	10.0	U	1	3.50	10.0	ug/Kg	10/24/25 09:24	VY102425
156-60-5	trans-1,2-Dichloroethene	5.00	U	1	0.86	5.00	ug/Kg	10/24/25 09:24	VY102425
75-34-3	1,1-Dichloroethane	5.00	U	1	0.80	5.00	ug/Kg	10/24/25 09:24	VY102425
110-82-7	Cyclohexane	5.00	U	1	0.79	5.00	ug/Kg	10/24/25 09:24	VY102425
78-93-3	2-Butanone	25.0	U	1	6.50	25.0	ug/Kg	10/24/25 09:24	VY102425
56-23-5	Carbon Tetrachloride	5.00	U	1	0.97	5.00	ug/Kg	10/24/25 09:24	VY102425
156-59-2	cis-1,2-Dichloroethene	5.00	U	1	0.75	5.00	ug/Kg	10/24/25 09:24	VY102425
74-97-5	Bromochloromethane	5.00	U	1	1.20	5.00	ug/Kg	10/24/25 09:24	VY102425
67-66-3	Chloroform	5.00	U	1	0.84	5.00	ug/Kg	10/24/25 09:24	VY102425
71-55-6	1,1,1-Trichloroethane	5.00	U	1	0.93	5.00	ug/Kg	10/24/25 09:24	VY102425
108-87-2	Methylcyclohexane	5.00	U	1	0.91	5.00	ug/Kg	10/24/25 09:24	VY102425
71-43-2	Benzene	5.00	U	1	0.79	5.00	ug/Kg	10/24/25 09:24	VY102425
107-06-2	1,2-Dichloroethane	5.00	U	1	0.79	5.00	ug/Kg	10/24/25 09:24	VY102425
79-01-6	Trichloroethene	5.00	U	1	0.81	5.00	ug/Kg	10/24/25 09:24	VY102425
78-87-5	1,2-Dichloropropane	5.00	U	1	0.91	5.00	ug/Kg	10/24/25 09:24	VY102425
75-27-4	Bromodichloromethane	5.00	U	1	0.78	5.00	ug/Kg	10/24/25 09:24	VY102425
108-10-1	4-Methyl-2-Pentanone	25.0	U	1	3.60	25.0	ug/Kg	10/24/25 09:24	VY102425
108-88-3	Toluene	5.00	U	1	0.78	5.00	ug/Kg	10/24/25 09:24	VY102425
10061-02-6	t-1,3-Dichloropropene	5.00	U	1	0.65	5.00	ug/Kg	10/24/25 09:24	VY102425
10061-01-5	cis-1,3-Dichloropropene	5.00	U	1	0.62	5.00	ug/Kg	10/24/25 09:24	VY102425
79-00-5	1,1,2-Trichloroethane	5.00	U	1	0.92	5.00	ug/Kg	10/24/25 09:24	VY102425
591-78-6	2-Hexanone	25.0	U	1	3.70	25.0	ug/Kg	10/24/25 09:24	VY102425
124-48-1	Dibromochloromethane	5.00	U	1	0.87	5.00	ug/Kg	10/24/25 09:24	VY102425
106-93-4	1,2-Dibromoethane	5.00	U	1	0.88	5.00	ug/Kg	10/24/25 09:24	VY102425
127-18-4	Tetrachloroethene	5.00	U	1	1.10	5.00	ug/Kg	10/24/25 09:24	VY102425
108-90-7	Chlorobenzene	5.00	U	1	0.91	5.00	ug/Kg	10/24/25 09:24	VY102425
100-41-4	Ethyl Benzene	5.00	U	1	0.67	5.00	ug/Kg	10/24/25 09:24	VY102425
179601-23-1	m/p-Xylenes	10.0	U	1	1.20	10.0	ug/Kg	10/24/25 09:24	VY102425
95-47-6	o-Xylene	5.00	U	1	0.82	5.00	ug/Kg	10/24/25 09:24	VY102425
100-42-5	Styrene	5.00	U	1	0.71	5.00	ug/Kg	10/24/25 09:24	VY102425
75-25-2	Bromoform	5.00	U	1	0.86	5.00	ug/Kg	10/24/25 09:24	VY102425

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1024SBL01
 Lab Sample ID: VY1024SBL01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	5.00	U	1	0.78	5.00	ug/Kg	10/24/25 09:24	VY102425
79-34-5	1,1,2,2-Tetrachloroethane	5.00	U	1	1.20	5.00	ug/Kg	10/24/25 09:24	VY102425
541-73-1	1,3-Dichlorobenzene	5.00	U	1	1.70	5.00	ug/Kg	10/24/25 09:24	VY102425
106-46-7	1,4-Dichlorobenzene	5.00	U	1	1.60	5.00	ug/Kg	10/24/25 09:24	VY102425
95-50-1	1,2-Dichlorobenzene	5.00	U	1	1.50	5.00	ug/Kg	10/24/25 09:24	VY102425
96-12-8	1,2-Dibromo-3-Chloropropane	5.00	U	1	1.80	5.00	ug/Kg	10/24/25 09:24	VY102425
120-82-1	1,2,4-Trichlorobenzene	5.00	U	1	3.00	5.00	ug/Kg	10/24/25 09:24	VY102425
87-61-6	1,2,3-Trichlorobenzene	5.00	U	1	3.20	5.00	ug/Kg	10/24/25 09:24	VY102425
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	44.4			63 - 155	89%	SPK: 50		
1868-53-7	Dibromofluoromethane	48.7			70 - 134	97%	SPK: 50		
2037-26-5	Toluene-d8	47.7			74 - 123	95%	SPK: 50		
460-00-4	4-Bromofluorobenzene	40.6			17 - 146	81%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	899000							
540-36-3	1,4-Difluorobenzene	1420000							
3114-55-4	Chlorobenzene-d5	1160000							
3855-82-1	1,4-Dichlorobenzene-d4	469000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023560.D
Acq On : 24 Oct 2025 09:24
Operator : SY/MD
Sample : VY1024SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Quant Time: Oct 24 23:46:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	898566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1421897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1164716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	469443	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	333317	44.360	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.720%
35) Dibromofluoromethane	7.634	113	425849	48.742	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.480%
50) Toluene-d8	10.109	98	1553516	47.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.500%
62) 4-Bromofluorobenzene	12.408	95	436342	40.620	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.240%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023560.D
Acq On : 24 Oct 2025 09:24
Operator : SY/MD
Sample : VY1024SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Quant Time: Oct 24 23:46:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023560.D
Acq: 24 Oct 2025 09:24




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01


Tgt Ion:168 Resp: 898566
Ion Ratio Lower Upper
168 100
99 45.5 39.7 59.5



#33
1,2-Dichloroethane-d4
Concen: 44.360 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY023560.D
Acq: 24 Oct 2025 09:24

Tgt Ion: 65 Resp: 333317
Ion Ratio Lower Upper
65 100
67 53.2 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023560.D

Acq: 24 Oct 2025 09:24

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBL01





Tgt Ion:114 Resp: 1421897

Ion Ratio Lower Upper

114 100

63 16.2 0.0 34.2

88 14.2 0.0 28.4



#35

Dibromofluoromethane

Concen: 48.742 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.000 min

Lab File: VY023560.D

Acq: 24 Oct 2025 09:24




Tgt Ion:113 Resp: 425849


Ion Ratio Lower Upper

113 100

111 102.6 81.8 122.6

192 23.1 18.0 27.0









#50
Toluene-d8
Concen: 47.746 ug/l
RT: 10.109 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023560.D
Acq: 24 Oct 2025 09:24




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01


Tgt Ion: 98 Resp: 1553516
Ion Ratio Lower Upper
98 100
100 66.0 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 40.620 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023560.D
Acq: 24 Oct 2025 09:24

Tgt Ion: 95 Resp: 436342
Ion Ratio Lower Upper
95 100
174 98.8 0.0 192.8
176 95.5 0.0 187.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY023560.D
Acq: 24 Oct 2025 09:24




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Tgt Ion:117 Resp: 1164716
Ion Ratio Lower Upper
117 100
82 50.4 41.0 61.4
119 32.4 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY023560.D
Acq: 24 Oct 2025 09:24

Tgt Ion:152 Resp: 469443
Ion Ratio Lower Upper
152 100
115 53.2 27.1 81.2
150 155.8 0.0 347.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023560.D
Acq On : 24 Oct 2025 09:24
Operator : SY/MD
Sample : VY1024SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M

Title : SW846 8260

Signal : TIC: VY023560.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.622	465	476	484	rBV	23500	73265	1.87%	0.343%
2	6.890	836	848	860	rBV2	57489	192184	4.91%	0.899%
3	7.634	958	970	976	rBV	582068	1409766	36.01%	6.595%
4	7.707	976	982	999	rVB	1021907	2379711	60.78%	11.132%
5	8.061	1025	1040	1053	rBV2	487402	1304536	33.32%	6.103%
6	8.616	1121	1131	1144	rBV	1543129	3000570	76.64%	14.037%
7	10.109	1367	1376	1384	rBV	2295340	3915079	100.00%	18.315%
8	10.512	1434	1442	1457	rVB	250078	477374	12.19%	2.233%
9	10.877	1495	1502	1511	rBV	112004	193115	4.93%	0.903%
10	11.249	1557	1563	1572	rVB	42891	75455	1.93%	0.353%
11	11.414	1582	1590	1603	rBV	1995280	3266102	83.42%	15.279%
12	12.408	1745	1753	1766	rBV2	1545769	2453272	62.66%	11.477%
13	12.743	1802	1808	1812	rBV4	20819	41230	1.05%	0.193%
14	13.347	1901	1907	1922	rVB	1700901	2594749	66.28%	12.138%




Sum of corrected areas: 21376408

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023560.D
Acq On : 24 Oct 2025 09:24
Operator : SY/MD
Sample : VY1024SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023560.D
Acq On : 24 Oct 2025 09:24
Operator : SY/MD
Sample : VY1024SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023560.D
Acq On : 24 Oct 2025 09:24
Operator : SY/MD
Sample : VY1024SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1022SBS01
 Lab Sample ID: VY1022SBS01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	20.0		1	1.10	5.00	ug/Kg	10/22/25 10:28	VY102225
74-87-3	Chloromethane	18.0		1	1.10	5.00	ug/Kg	10/22/25 10:28	VY102225
75-01-4	Vinyl Chloride	18.7		1	0.79	5.00	ug/Kg	10/22/25 10:28	VY102225
74-83-9	Bromomethane	20.0		1	1.10	5.00	ug/Kg	10/22/25 10:28	VY102225
75-00-3	Chloroethane	19.8		1	1.30	5.00	ug/Kg	10/22/25 10:28	VY102225
75-69-4	Trichlorofluoromethane	20.2		1	1.20	5.00	ug/Kg	10/22/25 10:28	VY102225
76-13-1	1,1,2-Trichlorotrifluoroethane	21.3		1	1.10	5.00	ug/Kg	10/22/25 10:28	VY102225
75-35-4	1,1-Dichloroethene	20.7		1	1.00	5.00	ug/Kg	10/22/25 10:28	VY102225
67-64-1	Acetone	130		1	4.70	25.0	ug/Kg	10/22/25 10:28	VY102225
75-15-0	Carbon Disulfide	19.3		1	1.10	5.00	ug/Kg	10/22/25 10:28	VY102225
1634-04-4	Methyl tert-butyl Ether	19.8		1	0.73	5.00	ug/Kg	10/22/25 10:28	VY102225
79-20-9	Methyl Acetate	17.3		1	1.50	5.00	ug/Kg	10/22/25 10:28	VY102225
75-09-2	Methylene Chloride	18.4		1	3.50	10.0	ug/Kg	10/22/25 10:28	VY102225
156-60-5	trans-1,2-Dichloroethene	20.4		1	0.86	5.00	ug/Kg	10/22/25 10:28	VY102225
75-34-3	1,1-Dichloroethane	20.5		1	0.80	5.00	ug/Kg	10/22/25 10:28	VY102225
110-82-7	Cyclohexane	19.3		1	0.79	5.00	ug/Kg	10/22/25 10:28	VY102225
78-93-3	2-Butanone	110		1	6.50	25.0	ug/Kg	10/22/25 10:28	VY102225
56-23-5	Carbon Tetrachloride	21.4		1	0.97	5.00	ug/Kg	10/22/25 10:28	VY102225
156-59-2	cis-1,2-Dichloroethene	21.2		1	0.75	5.00	ug/Kg	10/22/25 10:28	VY102225
74-97-5	Bromochloromethane	19.3		1	1.20	5.00	ug/Kg	10/22/25 10:28	VY102225
67-66-3	Chloroform	21.2		1	0.84	5.00	ug/Kg	10/22/25 10:28	VY102225
71-55-6	1,1,1-Trichloroethane	21.3		1	0.93	5.00	ug/Kg	10/22/25 10:28	VY102225
108-87-2	Methylcyclohexane	20.7		1	0.91	5.00	ug/Kg	10/22/25 10:28	VY102225
71-43-2	Benzene	21.2		1	0.79	5.00	ug/Kg	10/22/25 10:28	VY102225
107-06-2	1,2-Dichloroethane	19.8		1	0.79	5.00	ug/Kg	10/22/25 10:28	VY102225
79-01-6	Trichloroethene	21.9		1	0.81	5.00	ug/Kg	10/22/25 10:28	VY102225
78-87-5	1,2-Dichloropropane	21.0		1	0.91	5.00	ug/Kg	10/22/25 10:28	VY102225
75-27-4	Bromodichloromethane	21.2		1	0.78	5.00	ug/Kg	10/22/25 10:28	VY102225
108-10-1	4-Methyl-2-Pentanone	96.0		1	3.60	25.0	ug/Kg	10/22/25 10:28	VY102225
108-88-3	Toluene	22.0		1	0.78	5.00	ug/Kg	10/22/25 10:28	VY102225
10061-02-6	t-1,3-Dichloropropene	21.0		1	0.65	5.00	ug/Kg	10/22/25 10:28	VY102225
10061-01-5	cis-1,3-Dichloropropene	20.9		1	0.62	5.00	ug/Kg	10/22/25 10:28	VY102225
79-00-5	1,1,2-Trichloroethane	21.6		1	0.92	5.00	ug/Kg	10/22/25 10:28	VY102225
591-78-6	2-Hexanone	110		1	3.70	25.0	ug/Kg	10/22/25 10:28	VY102225
124-48-1	Dibromochloromethane	21.4		1	0.87	5.00	ug/Kg	10/22/25 10:28	VY102225
106-93-4	1,2-Dibromoethane	21.2		1	0.88	5.00	ug/Kg	10/22/25 10:28	VY102225
127-18-4	Tetrachloroethene	22.9		1	1.10	5.00	ug/Kg	10/22/25 10:28	VY102225
108-90-7	Chlorobenzene	21.4		1	0.91	5.00	ug/Kg	10/22/25 10:28	VY102225
100-41-4	Ethyl Benzene	21.2		1	0.67	5.00	ug/Kg	10/22/25 10:28	VY102225
179601-23-1	m/p-Xylenes	43.3		1	1.20	10.0	ug/Kg	10/22/25 10:28	VY102225
95-47-6	o-Xylene	21.6		1	0.82	5.00	ug/Kg	10/22/25 10:28	VY102225
100-42-5	Styrene	21.3		1	0.71	5.00	ug/Kg	10/22/25 10:28	VY102225
75-25-2	Bromoform	20.5		1	0.86	5.00	ug/Kg	10/22/25 10:28	VY102225

Report of Analysis

Client: AECOM	Date Collected:	
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	
Client Sample ID: VY1022SBS01	SDG No.:	Q3419
Lab Sample ID: VY1022SBS01	Matrix:	SOIL
Analytical Method: 8260D	% Solid:	100
Sample Wt/Vol: 5 g	Test:	VOC-TCLVOA-10
Level: LOW		
Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	21.7		1	0.78	5.00	ug/Kg	10/22/25 10:28	VY102225
79-34-5	1,1,2,2-Tetrachloroethane	19.2		1	1.20	5.00	ug/Kg	10/22/25 10:28	VY102225
541-73-1	1,3-Dichlorobenzene	21.3		1	1.70	5.00	ug/Kg	10/22/25 10:28	VY102225
106-46-7	1,4-Dichlorobenzene	21.1		1	1.60	5.00	ug/Kg	10/22/25 10:28	VY102225
95-50-1	1,2-Dichlorobenzene	21.3		1	1.50	5.00	ug/Kg	10/22/25 10:28	VY102225
96-12-8	1,2-Dibromo-3-Chloropropane	18.8		1	1.80	5.00	ug/Kg	10/22/25 10:28	VY102225
120-82-1	1,2,4-Trichlorobenzene	20.4		1	3.00	5.00	ug/Kg	10/22/25 10:28	VY102225
87-61-6	1,2,3-Trichlorobenzene	19.8		1	3.20	5.00	ug/Kg	10/22/25 10:28	VY102225
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	45.7			63 - 155	91%	SPK: 50		
1868-53-7	Dibromofluoromethane	49.3			70 - 134	99%	SPK: 50		
2037-26-5	Toluene-d8	49.8			74 - 123	100%	SPK: 50		
460-00-4	4-Bromofluorobenzene	49.3			17 - 146	99%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	547000							
540-36-3	1,4-Difluorobenzene	794000							
3114-55-4	Chlorobenzene-d5	700000							
3855-82-1	1,4-Dichlorobenzene-d4	365000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023545.D
 Acq On : 22 Oct 2025 10:28
 Operator : SY/MD
 Sample : VY1022SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1022SBS01

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/24/2025
 Supervised By :Mahesh Dadoda 10/24/2025

Quant Time: Oct 24 02:26:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	546555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	794303	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	700136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	365280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	208863	45.699	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	91.400%		
35) Dibromofluoromethane	7.634	113	240525	49.282	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.560%		
50) Toluene-d8	10.109	98	904633	49.771	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	99.540%		
62) 4-Bromofluorobenzene	12.408	95	295724	49.281	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	98.560%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	78862	20.046	ug/l	99
3) Chloromethane	2.074	50	96728	18.049	ug/l	97
4) Vinyl Chloride	2.208	62	130661	18.718	ug/l	99
5) Bromomethane	2.592	94	117294	20.024	ug/l	99
6) Chloroethane	2.739	64	94090	19.775	ug/l	99
7) Trichlorofluoromethane	3.056	101	195620	20.184	ug/l	99
8) Diethyl Ether	3.458	74	51088	20.172	ug/l	95
9) 1,1,2-Trichlorotrifluo...	3.824	101	107027	21.280	ug/l	98
10) Methyl Iodide	4.007	142	117165	20.926	ug/l	97
11) Tert butyl alcohol	4.860	59	28049	82.495	ug/l	98
12) 1,1-Dichloroethene	3.793	96	100716	20.658	ug/l	96
13) Acrolein	3.653	56	23166	48.218	ug/l	91
14) Allyl chloride	4.391	41	110064	18.524	ug/l	96
15) Acrylonitrile	5.061	53	90176	93.042	ug/l	99
16) Acetone	3.873	43	137184	126.081	ug/l	98
17) Carbon Disulfide	4.110	76	268938	19.273	ug/l	98
18) Methyl Acetate	4.391	43	49297	17.272	ug/l	98
19) Methyl tert-butyl Ether	5.122	73	231882	19.789	ug/l	97
20) Methylene Chloride	4.622	84	108140	18.380	ug/l	98
21) trans-1,2-Dichloroethene	5.116	96	109389	20.379	ug/l	97
22) Diisopropyl ether	6.019	45	262379	19.963	ug/l	96
23) Vinyl Acetate	5.964	43	681501	95.797	ug/l	96
24) 1,1-Dichloroethane	5.915	63	178683	20.524	ug/l	100
25) 2-Butanone	6.896	43	142981	107.638	ug/l	95
26) 2,2-Dichloropropane	6.884	77	167931	21.021	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	131514	21.211	ug/l	94
28) Bromochloromethane	7.250	49	62608	19.264	ug/l	94
29) Tetrahydrofuran	7.262	42	65038	86.341	ug/l	93
30) Chloroform	7.421	83	202767	21.198	ug/l	98
31) Cyclohexane	7.707	56	146973	19.331	ug/l	89
32) 1,1,1-Trichloroethane	7.622	97	185487	21.339	ug/l	98
36) 1,1-Dichloropropene	7.835	75	134867	20.873	ug/l	98
37) Ethyl Acetate	6.988	43	47665	18.620	ug/l #	96
38) Carbon Tetrachloride	7.823	117	167506	21.380	ug/l	99
39) Methylcyclohexane	9.109	83	174367	20.726	ug/l	98
40) Benzene	8.085	78	432933	21.157	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023545.D
 Acq On : 22 Oct 2025 10:28
 Operator : SY/MD
 Sample : VY1022SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1022SBS01

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/24/2025
 Supervised By :Mahesh Dadoda 10/24/2025

Quant Time: Oct 24 02:26:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.226	41	23647	17.080	ug/l #	67
42) 1,2-Dichloroethane	8.158	62	104422	19.787	ug/l	99
43) Isopropyl Acetate	8.201	43	93851	18.386	ug/l	96
44) Trichloroethene	8.866	130	131401	21.940	ug/l	99
45) 1,2-Dichloropropane	9.146	63	94941	21.006	ug/l	99
46) Dibromomethane	9.231	93	59570	20.843	ug/l	98
47) Bromodichloromethane	9.426	83	152859	21.237	ug/l	99
48) Methyl methacrylate	9.225	41	43753	18.246	ug/l	92
49) 1,4-Dioxane	9.225	88	12196	368.059	ug/l	96
51) 4-Methyl-2-Pentanone	9.999	43	253262	96.032	ug/l	99
52) Toluene	10.170	92	292104	21.986	ug/l	98
53) t-1,3-Dichloropropene	10.396	75	131626	21.031	ug/l	100
54) cis-1,3-Dichloropropene	9.859	75	154563	20.924	ug/l	97
55) 1,1,2-Trichloroethane	10.573	97	82907	21.572	ug/l	97
56) Ethyl methacrylate	10.438	69	87648	19.771	ug/l	93
57) 1,3-Dichloropropane	10.719	76	126727	20.809	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	191490	96.421	ug/l	98
59) 2-Hexanone	10.762	43	201768	107.056	ug/l	97
60) Dibromochloromethane	10.908	129	115294	21.397	ug/l	99
61) 1,2-Dibromoethane	11.018	107	77772	21.218	ug/l	99
64) Tetrachloroethene	10.646	164	165184	22.850	ug/l	99
65) Chlorobenzene	11.444	112	325435	21.371	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.518	131	116407	21.452	ug/l	100
67) Ethyl Benzene	11.518	91	522038	21.179	ug/l	99
68) m/p-Xylenes	11.627	106	428308	43.322	ug/l	99
69) o-Xylene	11.956	106	200731	21.646	ug/l	97
70) Styrene	11.969	104	325186	21.324	ug/l	100
71) Bromoform	12.133	173	69048	20.514	ug/l #	99
73) Isopropylbenzene	12.255	105	521665	21.671	ug/l	100
74) N-amyl acetate	12.072	43	76665	17.731	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.505	83	75537	19.185	ug/l	97
76) 1,2,3-Trichloropropane	12.554	75	60391m	18.739	ug/l	
77) Bromobenzene	12.530	156	136360	21.307	ug/l	97
78) n-propylbenzene	12.597	91	603121	21.651	ug/l	99
79) 2-Chlorotoluene	12.676	91	348031	21.335	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	426677	21.935	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	26519	20.178	ug/l	97
82) 4-Chlorotoluene	12.773	91	358213	21.322	ug/l	100
83) tert-Butylbenzene	12.999	119	393129	21.889	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	420267	21.648	ug/l	99
85) sec-Butylbenzene	13.176	105	561763	21.804	ug/l	99
86) p-Isopropyltoluene	13.292	119	485141	22.116	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	265067	21.270	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	260280	21.147	ug/l	99
89) n-Butylbenzene	13.615	91	407439	21.317	ug/l	99
90) Hexachloroethane	13.877	117	100017	21.199	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	232327	21.264	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	12117	18.756	ug/l	97
93) 1,2,4-Trichlorobenzene	14.919	180	138915	20.394	ug/l	100
94) Hexachlorobutadiene	15.023	225	88730	20.897	ug/l	97
95) Naphthalene	15.145	128	204152	18.822	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	116532	19.751	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023545.D
Acq On : 22 Oct 2025 10:28
Operator : SY/MD
Sample : VY1022SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025

Quant Time: Oct 24 02:26:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 982
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion:168 Resp: 54655
Ion Ratio Lower Upper
168 100
99 48.4 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#2
Dichlorodifluoromethane
Concen: 20.046 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 85 Resp: 78862
Ion Ratio Lower Upper
85 100
87 32.6 16.0 47.9









#3
Chloromethane
Concen: 18.049 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 50 Resp: 96728
Ion Ratio Lower Upper
50 100
52 34.4 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#4
Vinyl Chloride
Concen: 18.718 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28


Tgt Ion: 62 Resp: 130661
Ion Ratio Lower Upper
62 100
64 32.0 25.8 38.8





#5
Bromomethane
Concen: 20.024 ug/l
RT: 2.592 min Scan# 143
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01




Tgt Ion: 94 Resp: 117294
Ion Ratio Lower Upper
94 100
96 92.4 74.5 111.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#6
Chloroethane
Concen: 19.775 ug/l
RT: 2.739 min Scan# 167
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28



Tgt Ion: 64 Resp: 94090
Ion Ratio Lower Upper
64 100
66 32.3 26.2 39.4









#7
Trichlorofluoromethane
Concen: 20.184 ug/l
RT: 3.056 min Scan# 219
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion:101 Resp: 195620
Ion Ratio Lower Upper
101 100
103 64.5 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#8
Diethyl Ether
Concen: 20.172 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 74 Resp: 51088
Ion Ratio Lower Upper
74 100
45 75.5 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.280 ug/l
RT: 3.824 min Scan# 343
Delta R.T. 0.012 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion:101 Resp: 10702
Ion Ratio Lower Upper
101 100
85 42.9 34.6 51.8
151 85.6 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#10
Methyl Iodide
Concen: 20.926 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion:142 Resp: 117165
Ion Ratio Lower Upper
142 100
127 41.3 34.8 52.2
141 14.2 11.6 17.4





#11

Tert butyl alcohol

Concen: 82.495 ug/l

RT: 4.860 min Scan# 515

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 59 Resp: 28049

Ion Ratio Lower Upper

59 100





57 11.0 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#12

1,1-Dichloroethene

Concen: 20.658 ug/l

RT: 3.793 min Scan# 340

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion: 96 Resp: 100716


Ion Ratio Lower Upper

96 100

61 137.8 106.7 160.1

98 66.4 50.2 75.4





#13

Acrolein

Concen: 48.218 ug/l

RT: 3.653 min Scan# 316

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 56 Resp: 23160

Ion Ratio Lower Upper

56 100





55 77.1 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#14

Allyl chloride

Concen: 18.524 ug/l

RT: 4.391 min Scan# 438

Delta R.T. 0.012 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion: 41 Resp: 110064


Ion Ratio Lower Upper

41 100

39 77.2 59.9 89.9


76 51.6 38.0 57.0





#15
Acrylonitrile
Concen: 93.042 ug/l
RT: 5.061 min Scan# 54
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01




Tgt Ion: 53 Resp: 90170
Ion Ratio Lower Upper
53 100
52 79.9 64.6 97.0
51 37.2 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#16
Acetone
Concen: 126.081 ug/l
RT: 3.873 min Scan# 353
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28



Tgt Ion: 43 Resp: 137184
Ion Ratio Lower Upper
43 100
58 33.2 27.7 41.5









#17
Carbon Disulfide
Concen: 19.273 ug/l
RT: 4.110 min Scan# 391
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 76 Resp: 268938
Ion Ratio Lower Upper
76 100
78 8.3 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#18
Methyl Acetate
Concen: 17.272 ug/l
RT: 4.391 min Scan# 438
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 43 Resp: 49297
Ion Ratio Lower Upper
43 100
74 31.6 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 19.789 ug/l

RT: 5.122 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 73 Resp: 23188

Ion Ratio Lower Upper

73 100





57 20.2 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#20

Methylene Chloride

Concen: 18.380 ug/l

RT: 4.622 min Scan# 476

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion: 84 Resp: 108140




Ion Ratio Lower Upper


84 100

49 100.4 80.2 120.4

51 31.5 25.9 38.9

86 66.5 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 20.379 ug/l

RT: 5.116 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 96 Resp: 109389

Ion Ratio Lower Upper

96 100

61 117.6 92.7 139.1





98 66.9 49.7 74.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#22

Diisopropyl ether

Concen: 19.963 ug/l

RT: 6.019 min Scan# 705

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion: 45 Resp: 262379




Ion Ratio Lower Upper


45 100

43 62.2 45.6 68.4

87 36.1 28.1 42.1

59 12.5 10.2 15.2





#23

Vinyl Acetate

Concen: 95.797 ug/l

RT: 5.964 min Scan# 695

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 43 Resp: 68150

Ion Ratio Lower Upper

43

100

86

14.9

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#24

1,1-Dichloroethane

Concen: 20.524 ug/l

RT: 5.915 min Scan# 688

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion: 63 Resp: 178683

Ion Ratio Lower Upper

63

100

98




8.6


100

5.5

2.8

8.4





#25

2-Butanone

Concen: 107.638 ug/l

RT: 6.896 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 43 Resp: 14298

Ion Ratio Lower Upper

43 100





72 31.3 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#26

2,2-Dichloropropane

Concen: 21.021 ug/l

RT: 6.884 min Scan# 847

Delta R.T. -0.000 min

Lab File: VY023545.D




Acq: 22 Oct 2025 10:28


Tgt Ion: 77 Resp: 167931

Ion Ratio Lower Upper

77 100


97 24.6 12.1 36.3





#27
cis-1,2-Dichloroethene
Concen: 21.211 ug/l
RT: 6.890 min Scan# 848
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01




Tgt Ion: 96 Resp: 131514
Ion Ratio Lower Upper
96 100
61 120.8 0.0 258.2
98 63.9 0.0 132.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#28
Bromochloromethane
Concen: 19.264 ug/l
RT: 7.250 min Scan# 907
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28



Tgt Ion: 49 Resp: 62608
Ion Ratio Lower Upper
49 100
129 2.7 0.0 5.6
130 108.9 81.9 122.9





#29

Tetrahydrofuran

Concen: 86.341 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 42 Resp: 65038

Ion Ratio Lower Upper

42

100

72

71

58.5

54.7

43.0

40.1

64.6





60.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#30

Chloroform

Concen: 21.198 ug/l

RT: 7.421 min Scan# 935

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion: 83 Resp: 202767

Ion Ratio Lower Upper

83




100


85

63.7

52.5

78.7





#31

Cyclohexane

Concen: 19.331 ug/l

RT: 7.707 min Scan# 981

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 56 Resp: 14697

Ion Ratio Lower Upper

56 100

69 44.9 30.0 45.0





84 112.5 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#32

1,1,1-Trichloroethane

Concen: 21.339 ug/l

RT: 7.622 min Scan# 968

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion: 97 Resp: 185487


Ion Ratio Lower Upper

97 100

99 63.2 51.4 77.2

61 38.4 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 45.699 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 65 Resp: 20886

Ion Ratio Lower Upper

65 100





67 53.8 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion:114 Resp: 794303


Ion Ratio Lower Upper

114 100

63 16.3 0.0 34.2

88 14.2 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.282 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion:113 Resp: 24052

Ion Ratio Lower Upper

113 100

111 103.3 81.8 122.6





192 22.6 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#36

1,1-Dichloropropene

Concen: 20.873 ug/l

RT: 7.835 min Scan# 1003

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion: 75 Resp: 134867


Ion Ratio Lower Upper

75 100

110 41.2 19.9 59.7

77 31.7 25.0 37.6





#37

Ethyl Acetate

Concen: 18.620 ug/l

RT: 6.988 min Scan# 80

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 43 Resp: 4766

Ion Ratio Lower Upper

43 100

61 17.8 11.8 17.8





70 13.5 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#38

Carbon Tetrachloride

Concen: 21.380 ug/l

RT: 7.823 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion:117 Resp: 167506


Ion Ratio Lower Upper

117 100

119 95.5 76.8 115.2

121 31.2 24.5 36.7





#39

Methylcyclohexane

Concen: 20.726 ug/l

RT: 9.109 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 83 Resp: 17436

Ion Ratio Lower Upper

83 100

55 66.0 51.4 77.2





98 51.6 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#40

Benzene

Concen: 21.157 ug/l

RT: 8.085 min Scan# 1044

Delta R.T. 0.006 min

Lab File: VY023545.D




Acq: 22 Oct 2025 10:28


Tgt Ion: 78 Resp: 432933

Ion Ratio Lower Upper

78 100

77 23.9 18.3 27.5





#41

Methacrylonitrile

Concen: 17.080 ug/l

RT: 7.226 min Scan# 902

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 41 Resp: 2364

Ion Ratio Lower Upper

41 100

39 0.0 52.2 78.4

67 102.5 74.9 112.3





52 40.1 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#42

1,2-Dichloroethane

Concen: 19.787 ug/l

RT: 8.158 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VY023545.D




Acq: 22 Oct 2025 10:28


Tgt Ion: 62 Resp: 104422

Ion Ratio Lower Upper

62 100

98 11.2 0.0 21.6





#43

Isopropyl Acetate

Concen: 18.386 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :





VY1022SBS01

Tgt Ion: 43	Resp: 9385
Ion Ratio	Lower Upper
43 100	
61 25.1	18.1 27.1
87 17.9	13.5 20.3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#44

Trichloroethene

Concen: 21.940 ug/l




RT: 8.866 min Scan# 1172


Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion:130	Resp: 131401
Ion Ratio	Lower Upper
130 100	
95 89.3	0.0 180.6





#45

1,2-Dichloropropane

Concen: 21.006 ug/l

RT: 9.146 min Scan# 1217

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 63 Resp: 9494

Ion Ratio Lower Upper

63 100





65 30.7 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#46

Dibromomethane

Concen: 20.843 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion: 93 Resp: 59570


Ion Ratio Lower Upper

93 100

95 85.1 65.4 98.0


174 117.7 92.9 139.3





#47
Bromodichloromethane
Concen: 21.237 ug/l
RT: 9.426 min Scan# 1263
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion: 83 Resp: 152859
Ion Ratio Lower Upper
83 100
85 62.3 50.8 76.2
127 8.9 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#48
Methyl methacrylate
Concen: 18.246 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 41 Resp: 43753
Ion Ratio Lower Upper
41 100
69 111.4 81.0 121.4
39 62.9 48.2 72.4





#49

1,4-Dioxane

Concen: 368.059 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01





Tgt Ion:	88	Resp:	12190
Ion	Ratio	Lower	Upper
88	100		
43	30.0	21.8	32.8
58	60.8	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#50

Toluene-d8

Concen: 49.771 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion:	98	Resp:	904633
Ion	Ratio	Lower	Upper
98	100		
100	66.0	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 96.032 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 43 Resp: 253262
Ion Ratio Lower Upper
43 100
58 45.0 35.3 52.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#52
Toluene
Concen: 21.986 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 92 Resp: 292104
Ion Ratio Lower Upper
92 100
91 168.4 136.4 204.6









#53
t-1,3-Dichloropropene
Concen: 21.031 ug/l
RT: 10.396 min Scan# 1423
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 75 Resp: 131620
Ion Ratio Lower Upper
75 100
77 31.5 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#54
cis-1,3-Dichloropropene
Concen: 20.924 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 75 Resp: 154563
Ion Ratio Lower Upper
75 100
77 33.0 25.4 38.0
39 38.6 32.8 49.2





#55

1,1,2-Trichloroethane

Concen: 21.572 ug/l

RT: 10.573 min Scan# 1452

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 97 Resp: 8290

Ion Ratio Lower Upper

97 100

83 80.1 61.8 92.6

85 55.0 41.3 61.9





99 60.7 48.7 73.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#56

Ethyl methacrylate

Concen: 19.771 ug/l

RT: 10.438 min Scan# 1430

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion: 69 Resp: 87648


Ion Ratio Lower Upper

69 100

41 50.7 44.6 67.0

39 30.3 27.0 40.4









#57
1,3-Dichloropropane
Concen: 20.809 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 76 Resp: 12672
Ion Ratio Lower Upper
76 100
78 32.8 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#58
2-Chloroethyl Vinyl ether
Concen: 96.421 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28


Tgt Ion: 63 Resp: 191490
Ion Ratio Lower Upper
63 100
106 34.4 26.6 39.8





#59
2-Hexanone
Concen: 107.056 ug/l
RT: 10.762 min Scan# 1483
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion: 43 Resp: 201768
Ion Ratio Lower Upper
43 100
58 63.4 30.5 91.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#60
Dibromochloromethane
Concen: 21.397 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28


Tgt Ion:129 Resp: 115294
Ion Ratio Lower Upper
129 100
127 76.1 38.6 115.7





#61
1,2-Dibromoethane
Concen: 21.218 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01




Tgt Ion: 107 Resp: 77772
Ion Ratio Lower Upper
107 100
109 94.1 76.1 114.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025




#62
4-Bromofluorobenzene
Concen: 49.281 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28



Tgt Ion: 95 Resp: 295724
Ion Ratio Lower Upper
95 100
174 98.8 0.0 192.8
176 95.7 0.0 187.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion:117 Resp: 700130

Ion Ratio Lower Upper

117 100

82 50.1 41.0 61.4





119 31.7 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#64

Tetrachloroethene

Concen: 22.850 ug/l

RT: 10.646 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion:164 Resp: 165184




Ion Ratio Lower Upper


164 100

166 125.9 99.4 149.0

129 86.5 69.4 104.2


131 84.4 67.5 101.3





#65
Chlorobenzene
Concen: 21.371 ug/l
RT: 11.444 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion:112 Resp: 32543
Ion Ratio Lower Upper
112 100
114 32.1 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 21.452 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion:131 Resp: 116407
Ion Ratio Lower Upper
131 100
133 96.9 48.4 145.2
119 65.4 32.4 97.2









#67
Ethyl Benzene
Concen: 21.179 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 91 Resp: 522038
Ion Ratio Lower Upper
91 100
106 32.9 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#68
m/p-Xylenes
Concen: 43.322 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion:106 Resp: 428308
Ion Ratio Lower Upper
106 100
91 189.4 152.6 229.0





#69

o-Xylene

Concen: 21.646 ug/l

RT: 11.956 min Scan# 1678

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion:106 Resp: 20073

Ion Ratio Lower Upper

106 100





91 198.5 101.4 304.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#70

Styrene

Concen: 21.324 ug/l

RT: 11.969 min Scan# 1681

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion:104 Resp: 325186


Ion Ratio Lower Upper

104 100

78 44.8 36.1 54.1


103 54.1 43.2 64.8





#71
Bromoform
Concen: 20.514 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion:173 Resp: 69048
Ion Ratio Lower Upper
173 100
175 50.4 24.8 74.3
254 0.3 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28


Tgt Ion:152 Resp: 365280
Ion Ratio Lower Upper
152 100
115 53.8 27.1 81.2
150 163.7 0.0 347.4





#73
Isopropylbenzene
Concen: 21.671 ug/l
RT: 12.255 min Scan# 1727
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion:105 Resp: 521665
Ion Ratio Lower Upper
105 100
120 27.8 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#74
N-ethyl acetate
Concen: 17.731 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 43 Resp: 76665
Ion Ratio Lower Upper
43 100
70 55.8 41.1 61.7
55 30.9 23.0 34.6
61 28.4 21.4 32.2





#75

1,1,2,2-Tetrachloroethane

Concen: 19.185 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01





Tgt Ion:	83	Resp:	7553
Ion	Ratio	Lower	Upper
83	100		
131	12.0	5.9	17.6
85	65.8	31.8	95.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#76

1,2,3-Trichloropropane

Concen: 18.739 ug/l m




RT: 12.554 min Scan# 1777


Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion:	75	Resp:	60391
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0









#77
Bromobenzene
Concen: 21.307 ug/l
RT: 12.530 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion:156 Resp: 136360
Ion Ratio Lower Upper
156 100
77 149.0 76.8 230.5
158 96.9 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#78
n-propylbenzene
Concen: 21.651 ug/l
RT: 12.597 min Scan# 1784
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 91 Resp: 603121
Ion Ratio Lower Upper
91 100
120 25.2 12.3 36.8









#79
2-Chlorotoluene
Concen: 21.335 ug/l
RT: 12.676 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion: 91 Resp: 34803
Ion Ratio Lower Upper
91 100
126 37.8 18.5 55.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#80
1,3,5-Trimethylbenzene
Concen: 21.935 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28


Tgt Ion:105 Resp: 426677
Ion Ratio Lower Upper
105 100
120 51.8 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 20.178 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion: 75 Resp: 26519
Ion Ratio Lower Upper
75 100
53 81.9 68.6 102.8
89 48.4 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#82
4-Chlorotoluene
Concen: 21.322 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28


Tgt Ion: 91 Resp: 358213
Ion Ratio Lower Upper
91 100
126 37.0 18.4 55.4





#83
tert-Butylbenzene
Concen: 21.889 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.006 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28




Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01



Tgt Ion:119 Resp: 393129
Ion Ratio Lower Upper
119 100
91 59.2 29.9 89.8
134 26.1 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#84
1,2,4-Trimethylbenzene
Concen: 21.648 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion:105 Resp: 420267
Ion Ratio Lower Upper
105 100
120 47.1 23.8 71.2





#85

sec-Butylbenzene

Concen: 21.804 ug/l

RT: 13.176 min Scan# 1878

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion:105 Resp: 56176

Ion Ratio Lower Upper

105 100





134 21.7 10.7 32.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#86

p-Isopropyltoluene

Concen: 22.116 ug/l

RT: 13.292 min Scan# 1898

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion:119 Resp: 485141


Ion Ratio Lower Upper

119 100

134 27.1 13.6 40.7

91 21.5 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 21.270 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :





VY1022SBS01

Tgt Ion:	146	Resp:	26506
Ion Ratio	Lower	Upper	
146	100		
111	36.7	18.7	56.1
148	64.2	31.9	95.9

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#88

1,4-Dichlorobenzene

Concen: 21.147 ug/l




RT: 13.365 min Scan# 1910


Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Tgt Ion:	146	Resp:	260280
Ion Ratio	Lower	Upper	
146	100		
111	36.9	18.0	54.0
148	63.7	31.9	95.7





#89

n-Butylbenzene

Concen: 21.317 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion: 91 Resp: 407439

Ion Ratio Lower Upper

91 100

92 52.8 27.0 80.8





134 28.3 14.1 42.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#90

Hexachloroethane

Concen: 21.199 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. -0.000 min

Lab File: VY023545.D




Acq: 22 Oct 2025 10:28


Tgt Ion:117 Resp: 100017

Ion Ratio Lower Upper

117 100

201 88.2 43.8 131.4









#91
1,2-Dichlorobenzene
Concen: 21.264 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBS01

Tgt Ion:146 Resp: 23232
Ion Ratio Lower Upper
146 100
111 37.7 19.0 57.0
148 64.1 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/24/2025
Supervised By :Mahesh Dadoda 10/24/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.756 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023545.D
Acq: 22 Oct 2025 10:28

Tgt Ion: 75 Resp: 12117
Ion Ratio Lower Upper
75 100
155 102.3 50.0 150.0
157 132.5 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 20.394 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023545.D


Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01



Tgt Ion:180 Resp: 138915

Ion Ratio Lower Upper

180 100

182 95.6 47.9 143.6




145 28.9 14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#94

Hexachlorobutadiene


Concen: 20.897 ug/l

RT: 15.023 min Scan# 2182

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28





Tgt Ion:225 Resp: 88730


Ion Ratio Lower Upper

225 100

223 64.5 30.8 92.3

227 66.0 31.9 95.7





#95

Naphthalene

Concen: 18.822 ug/l

RT: 15.145 min Scan# 2201

Delta R.T. 0.006 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28

Instrument :

MSVOA_Y

ClientSampleId :

VY1022SBS01

Tgt Ion:128 Resp: 20415

Ion Ratio Lower Upper

128 100

127 13.1 10.4 15.6





129 10.6 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/24/2025

Supervised By :Mahesh Dadoda 10/24/2025



#96

1,2,3-Trichlorobenzene

Concen: 19.751 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023545.D

Acq: 22 Oct 2025 10:28




Tgt Ion:180 Resp: 116532

Ion Ratio Lower Upper

180 100

182 96.2 47.9 143.8

145 30.1 15.3 45.9



Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1024SBS01
 Lab Sample ID: VY1024SBS01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	18.6		1	1.10	5.00	ug/Kg	10/24/25 09:52	VY102425
74-87-3	Chloromethane	17.3		1	1.10	5.00	ug/Kg	10/24/25 09:52	VY102425
75-01-4	Vinyl Chloride	18.0		1	0.79	5.00	ug/Kg	10/24/25 09:52	VY102425
74-83-9	Bromomethane	20.6		1	1.10	5.00	ug/Kg	10/24/25 09:52	VY102425
75-00-3	Chloroethane	18.5		1	1.30	5.00	ug/Kg	10/24/25 09:52	VY102425
75-69-4	Trichlorofluoromethane	19.0		1	1.20	5.00	ug/Kg	10/24/25 09:52	VY102425
76-13-1	1,1,2-Trichlorotrifluoroethane	20.0		1	1.10	5.00	ug/Kg	10/24/25 09:52	VY102425
75-35-4	1,1-Dichloroethene	19.8		1	1.00	5.00	ug/Kg	10/24/25 09:52	VY102425
67-64-1	Acetone	130		1	4.70	25.0	ug/Kg	10/24/25 09:52	VY102425
75-15-0	Carbon Disulfide	18.4		1	1.10	5.00	ug/Kg	10/24/25 09:52	VY102425
1634-04-4	Methyl tert-butyl Ether	20.0		1	0.73	5.00	ug/Kg	10/24/25 09:52	VY102425
79-20-9	Methyl Acetate	17.2		1	1.50	5.00	ug/Kg	10/24/25 09:52	VY102425
75-09-2	Methylene Chloride	18.5		1	3.50	10.0	ug/Kg	10/24/25 09:52	VY102425
156-60-5	trans-1,2-Dichloroethene	20.1		1	0.86	5.00	ug/Kg	10/24/25 09:52	VY102425
75-34-3	1,1-Dichloroethane	19.8		1	0.80	5.00	ug/Kg	10/24/25 09:52	VY102425
110-82-7	Cyclohexane	18.3		1	0.79	5.00	ug/Kg	10/24/25 09:52	VY102425
78-93-3	2-Butanone	110		1	6.50	25.0	ug/Kg	10/24/25 09:52	VY102425
56-23-5	Carbon Tetrachloride	20.6		1	0.97	5.00	ug/Kg	10/24/25 09:52	VY102425
156-59-2	cis-1,2-Dichloroethene	20.9		1	0.75	5.00	ug/Kg	10/24/25 09:52	VY102425
74-97-5	Bromochloromethane	18.9		1	1.20	5.00	ug/Kg	10/24/25 09:52	VY102425
67-66-3	Chloroform	20.4		1	0.84	5.00	ug/Kg	10/24/25 09:52	VY102425
71-55-6	1,1,1-Trichloroethane	20.5		1	0.93	5.00	ug/Kg	10/24/25 09:52	VY102425
108-87-2	Methylcyclohexane	20.0		1	0.91	5.00	ug/Kg	10/24/25 09:52	VY102425
71-43-2	Benzene	20.7		1	0.79	5.00	ug/Kg	10/24/25 09:52	VY102425
107-06-2	1,2-Dichloroethane	19.6		1	0.79	5.00	ug/Kg	10/24/25 09:52	VY102425
79-01-6	Trichloroethene	21.4		1	0.81	5.00	ug/Kg	10/24/25 09:52	VY102425
78-87-5	1,2-Dichloropropane	20.3		1	0.91	5.00	ug/Kg	10/24/25 09:52	VY102425
75-27-4	Bromodichloromethane	20.8		1	0.78	5.00	ug/Kg	10/24/25 09:52	VY102425
108-10-1	4-Methyl-2-Pentanone	96.4		1	3.60	25.0	ug/Kg	10/24/25 09:52	VY102425
108-88-3	Toluene	21.3		1	0.78	5.00	ug/Kg	10/24/25 09:52	VY102425
10061-02-6	t-1,3-Dichloropropene	20.3		1	0.65	5.00	ug/Kg	10/24/25 09:52	VY102425
10061-01-5	cis-1,3-Dichloropropene	20.6		1	0.62	5.00	ug/Kg	10/24/25 09:52	VY102425
79-00-5	1,1,2-Trichloroethane	21.6		1	0.92	5.00	ug/Kg	10/24/25 09:52	VY102425
591-78-6	2-Hexanone	110		1	3.70	25.0	ug/Kg	10/24/25 09:52	VY102425
124-48-1	Dibromochloromethane	21.3		1	0.87	5.00	ug/Kg	10/24/25 09:52	VY102425
106-93-4	1,2-Dibromoethane	21.4		1	0.88	5.00	ug/Kg	10/24/25 09:52	VY102425
127-18-4	Tetrachloroethene	20.6		1	1.10	5.00	ug/Kg	10/24/25 09:52	VY102425
108-90-7	Chlorobenzene	20.9		1	0.91	5.00	ug/Kg	10/24/25 09:52	VY102425
100-41-4	Ethyl Benzene	20.5		1	0.67	5.00	ug/Kg	10/24/25 09:52	VY102425
179601-23-1	m/p-Xylenes	42.3		1	1.20	10.0	ug/Kg	10/24/25 09:52	VY102425
95-47-6	o-Xylene	21.1		1	0.82	5.00	ug/Kg	10/24/25 09:52	VY102425
100-42-5	Styrene	20.9		1	0.71	5.00	ug/Kg	10/24/25 09:52	VY102425
75-25-2	Bromoform	20.9		1	0.86	5.00	ug/Kg	10/24/25 09:52	VY102425

Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: VY1024SBS01
 Lab Sample ID: VY1024SBS01
 Analytical Method: 8260D
 Sample Wt/Vol: 5 g

Level : LOW
 Final Vol: 5000 uL

Date Collected:
 Date Received:
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 100
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	20.6		1	0.78	5.00	ug/Kg	10/24/25 09:52	VY102425
79-34-5	1,1,2,2-Tetrachloroethane	20.3		1	1.20	5.00	ug/Kg	10/24/25 09:52	VY102425
541-73-1	1,3-Dichlorobenzene	20.4		1	1.70	5.00	ug/Kg	10/24/25 09:52	VY102425
106-46-7	1,4-Dichlorobenzene	20.6		1	1.60	5.00	ug/Kg	10/24/25 09:52	VY102425
95-50-1	1,2-Dichlorobenzene	20.4		1	1.50	5.00	ug/Kg	10/24/25 09:52	VY102425
96-12-8	1,2-Dibromo-3-Chloropropane	19.1		1	1.80	5.00	ug/Kg	10/24/25 09:52	VY102425
120-82-1	1,2,4-Trichlorobenzene	19.9		1	3.00	5.00	ug/Kg	10/24/25 09:52	VY102425
87-61-6	1,2,3-Trichlorobenzene	19.6		1	3.20	5.00	ug/Kg	10/24/25 09:52	VY102425
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	44.2			63 - 155	88%	SPK: 50		
1868-53-7	Dibromofluoromethane	49.5			70 - 134	99%	SPK: 50		
2037-26-5	Toluene-d8	48.4			74 - 123	97%	SPK: 50		
460-00-4	4-Bromofluorobenzene	49.3			17 - 146	99%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	552000							
540-36-3	1,4-Difluorobenzene	799000							
3114-55-4	Chlorobenzene-d5	710000							
3855-82-1	1,4-Dichlorobenzene-d4	381000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023561.D
 Acq On : 24 Oct 2025 09:52
 Operator : SY/MD
 Sample : VY1024SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1024SBS01

Quant Time: Oct 24 23:47:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	551537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	799010	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.413	117	709717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	380713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	203935	44.218	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	88.440%		
35) Dibromofluoromethane	7.634	113	243016	49.499	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	99.000%		
50) Toluene-d8	10.109	98	884395	48.371	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	96.740%		
62) 4-Bromofluorobenzene	12.407	95	297693	49.317	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	98.640%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.866	85	73940	18.625	ug/l	99
3) Chloromethane	2.074	50	93760	17.337	ug/l	97
4) Vinyl Chloride	2.208	62	126898	18.015	ug/l	100
5) Bromomethane	2.592	94	121542	20.561	ug/l	100
6) Chloroethane	2.732	64	88904	18.516	ug/l	99
7) Trichlorofluoromethane	3.055	101	186017	19.019	ug/l	99
8) Diethyl Ether	3.458	74	50315	19.688	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.811	101	101494	19.998	ug/l	97
10) Methyl Iodide	4.006	142	106029	18.766	ug/l	97
11) Tert butyl alcohol	4.848	59	31189	90.901	ug/l	97
12) 1,1-Dichloroethene	3.787	96	97421	19.802	ug/l	97
13) Acrolein	3.659	56	23585	48.647	ug/l	98
14) Allyl chloride	4.384	41	109431	18.251	ug/l	97
15) Acrylonitrile	5.061	53	93885	95.993	ug/l	98
16) Acetone	3.866	43	140610	128.062	ug/l	99
17) Carbon Disulfide	4.104	76	258594	18.365	ug/l	99
18) Methyl Acetate	4.390	43	49637	17.234	ug/l	94
19) Methyl tert-butyl Ether	5.116	73	236371	19.990	ug/l	96
20) Methylene Chloride	4.616	84	109899	18.510	ug/l	98
21) trans-1,2-Dichloroethene	5.116	96	108914	20.107	ug/l	97
22) Diisopropyl ether	6.024	45	256039	19.305	ug/l	96
23) Vinyl Acetate	5.963	43	676295	94.207	ug/l	97
24) 1,1-Dichloroethane	5.921	63	174246	19.834	ug/l	98
25) 2-Butanone	6.896	43	151403	112.949	ug/l	100
26) 2,2-Dichloropropane	6.890	77	163903	20.332	ug/l	98
27) cis-1,2-Dichloroethene	6.896	96	130666	20.884	ug/l	95
28) Bromochloromethane	7.250	49	61891	18.871	ug/l	90
29) Tetrahydrofuran	7.268	42	69086	90.887	ug/l	93
30) Chloroform	7.420	83	197077	20.417	ug/l	99
31) Cyclohexane	7.701	56	140190	18.272	ug/l	98
32) 1,1,1-Trichloroethane	7.615	97	179638	20.479	ug/l	98
36) 1,1-Dichloropropene	7.841	75	130575	20.089	ug/l	99
37) Ethyl Acetate	6.987	43	49292	19.142	ug/l	97
38) Carbon Tetrachloride	7.823	117	162485	20.617	ug/l	99
39) Methylcyclohexane	9.109	83	169117	19.984	ug/l	99
40) Benzene	8.079	78	426341	20.712	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023561.D
 Acq On : 24 Oct 2025 09:52
 Operator : SY/MD
 Sample : VY1024SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1024SBS01

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:47:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	26399m	18.956	ug/l	
42) 1,2-Dichloroethane	8.158	62	104042	19.599	ug/l	97
43) Isopropyl Acetate	8.201	43	95722	18.643	ug/l	95
44) Trichloroethene	8.865	130	128825	21.383	ug/l	95
45) 1,2-Dichloropropane	9.146	63	92312	20.304	ug/l	100
46) Dibromomethane	9.231	93	59886	20.831	ug/l	96
47) Bromodichloromethane	9.426	83	150883	20.839	ug/l	99
48) Methyl methacrylate	9.225	41	43337	17.966	ug/l	89
49) 1,4-Dioxane	9.231	88	12987	389.621	ug/l	98
51) 4-Methyl-2-Pentanone	9.999	43	255637	96.362	ug/l	96
52) Toluene	10.170	92	284413	21.281	ug/l	99
53) t-1,3-Dichloropropene	10.395	75	127577	20.264	ug/l	100
54) cis-1,3-Dichloropropene	9.859	75	153131	20.608	ug/l	97
55) 1,1,2-Trichloroethane	10.572	97	83454	21.587	ug/l	97
56) Ethyl methacrylate	10.438	69	88085	19.752	ug/l	94
57) 1,3-Dichloropropane	10.718	76	126908	20.716	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	198103	99.163	ug/l	97
59) 2-Hexanone	10.761	43	211229	111.416	ug/l	96
60) Dibromochloromethane	10.914	129	115677	21.342	ug/l	100
61) 1,2-Dibromoethane	11.017	107	78817	21.377	ug/l	99
64) Tetrachloroethene	10.645	164	151245	20.639	ug/l	97
65) Chlorobenzene	11.444	112	322004	20.861	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.517	131	116589	21.195	ug/l	99
67) Ethyl Benzene	11.517	91	512753	20.521	ug/l	99
68) m/p-Xylenes	11.627	106	423494	42.257	ug/l	99
69) o-Xylene	11.956	106	198083	21.072	ug/l	97
70) Styrene	11.968	104	323830	20.949	ug/l	99
71) Bromoform	12.133	173	71371	20.917	ug/l #	99
73) Isopropylbenzene	12.255	105	516359	20.581	ug/l	100
74) N-amyl acetate	12.072	43	79625	17.669	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.505	83	83375	20.318	ug/l	99
76) 1,2,3-Trichloropropane	12.553	75	61116m	18.195	ug/l	
77) Bromobenzene	12.529	156	137052	20.547	ug/l	98
78) n-propylbenzene	12.596	91	595772	20.520	ug/l	99
79) 2-Chlorotoluene	12.675	91	344916	20.287	ug/l	98
80) 1,3,5-Trimethylbenzene	12.736	105	419394	20.686	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	26936	19.665	ug/l	96
82) 4-Chlorotoluene	12.773	91	358669	20.484	ug/l	99
83) tert-Butylbenzene	12.999	119	392010	20.942	ug/l	98
84) 1,2,4-Trimethylbenzene	13.041	105	420704	20.792	ug/l	99
85) sec-Butylbenzene	13.175	105	560392	20.869	ug/l	100
86) p-Isopropyltoluene	13.291	119	476750	20.853	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	264896	20.394	ug/l	100
88) 1,4-Dichlorobenzene	13.364	146	264053	20.584	ug/l	100
89) n-Butylbenzene	13.614	91	401494	20.154	ug/l	99
90) Hexachloroethane	13.876	117	99943	20.324	ug/l	98
91) 1,2-Dichlorobenzene	13.657	146	232521	20.419	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	12890	19.144	ug/l	96
93) 1,2,4-Trichlorobenzene	14.919	180	141537	19.937	ug/l	100
94) Hexachlorobutadiene	15.023	225	90220	20.387	ug/l	100
95) Naphthalene	15.144	128	210092	18.584	ug/l	99
96) 1,2,3-Trichlorobenzene	15.327	180	120577	19.609	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023561.D
Acq On : 24 Oct 2025 09:52
Operator : SY/MD
Sample : VY1024SBS01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:47:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 90
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion:168 Resp: 55153
Ion Ratio Lower Upper
168 100
99 47.3 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#2
Dichlorodifluoromethane
Concen: 18.625 ug/l
RT: 1.866 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion: 85 Resp: 73940
Ion Ratio Lower Upper
85 100
87 32.4 16.0 47.9









#3
Chloromethane
Concen: 17.337 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion: 50 Resp: 93760
Ion Ratio Lower Upper
50 100
52 34.4 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#4
Vinyl Chloride
Concen: 18.015 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion: 62 Resp: 126898
Ion Ratio Lower Upper
62 100
64 32.5 25.8 38.8





#5
Bromomethane
Concen: 20.561 ug/l
RT: 2.592 min Scan# 143
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion: 94 Resp: 12154
Ion Ratio Lower Upper
94 100
96 93.2 74.5 111.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#6
Chloroethane
Concen: 18.516 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52



Tgt Ion: 64 Resp: 88904
Ion Ratio Lower Upper
64 100
66 33.1 26.2 39.4









#7
Trichlorofluoromethane
Concen: 19.019 ug/l
RT: 3.055 min Scan# 219
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion:101 Resp: 18601
Ion Ratio Lower Upper
101 100
103 63.2 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#8
Diethyl Ether
Concen: 19.688 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion: 74 Resp: 50315
Ion Ratio Lower Upper
74 100
45 77.0 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.998 ug/l
RT: 3.811 min Scan# 343
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion:101 Resp: 101494
Ion Ratio Lower Upper
101 100
85 42.1 34.6 51.8
151 87.2 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#10
Methyl Iodide
Concen: 18.766 ug/l
RT: 4.006 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion:142 Resp: 106029
Ion Ratio Lower Upper
142 100
127 41.3 34.8 52.2
141 14.8 11.6 17.4





#11

Tert butyl alcohol

Concen: 90.901 ug/l

RT: 4.848 min Scan# 515

Delta R.T. -0.013 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 59 Resp: 31189

Ion Ratio Lower Upper

59 100





57 9.4 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#12

1,1-Dichloroethene

Concen: 19.802 ug/l

RT: 3.787 min Scan# 339

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion: 96 Resp: 97421


Ion Ratio Lower Upper

96 100

61 138.7 106.7 160.1

98 63.2 50.2 75.4





#13

Acrolein

Concen: 48.647 ug/l

RT: 3.659 min Scan# 316

Delta R.T. 0.012 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 56 Resp: 2358

Ion Ratio Lower Upper

56 100





55 72.0 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#14

Allyl chloride

Concen: 18.251 ug/l

RT: 4.384 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion: 41 Resp: 109431


Ion Ratio Lower Upper

41 100

39 75.3 59.9 89.9


76 52.3 38.0 57.0





#15
Acrylonitrile
Concen: 95.993 ug/l
RT: 5.061 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion: 53 Resp: 9388
Ion Ratio Lower Upper
53 100
52 79.3 64.6 97.0
51 35.6 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#16
Acetone
Concen: 128.062 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52



Tgt Ion: 43 Resp: 140610
Ion Ratio Lower Upper
43 100
58 34.3 27.7 41.5





#17

Carbon Disulfide

Concen: 18.365 ug/l

RT: 4.104 min Scan# 391

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion: 76 Resp: 258594

Ion Ratio Lower Upper

76 100

78 8.8 7.3 10.9

Instrument :

MSVOA_Y

ClientSampleId :





VY1024SBS01

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#18

Methyl Acetate

Concen: 17.234 ug/l

RT: 4.390 min Scan# 438

Delta R.T. 0.006 min

Lab File: VY023561.D




Acq: 24 Oct 2025 09:52


Tgt Ion: 43 Resp: 49637

Ion Ratio Lower Upper

43 100

74 33.4 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 19.990 ug/l

RT: 5.116 min Scan# 51

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 73 Resp: 23637

Ion Ratio Lower Upper

73 100





57 19.9 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#20

Methylene Chloride

Concen: 18.510 ug/l

RT: 4.616 min Scan# 475

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion: 84 Resp: 109899




Ion Ratio Lower Upper


84 100

49 98.7 80.2 120.4

51 30.3 25.9 38.9

86 63.8 50.6 75.8





#21

trans-1,2-Dichloroethene

Concen: 20.107 ug/l

RT: 5.116 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023561.D


Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01



Tgt Ion: 96 Resp: 108914

Ion Ratio Lower Upper

96 100

61 112.9 92.7 139.1




98 64.7 49.7 74.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#22

Diisopropyl ether


Concen: 19.305 ug/l

RT: 6.024 min Scan# 706

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52



Tgt Ion: 45 Resp: 256039



Ion Ratio Lower Upper


45 100

43 59.1 45.6 68.4

87 39.0 28.1 42.1

59 12.8 10.2 15.2





#23

Vinyl Acetate

Concen: 94.207 ug/l

RT: 5.963 min Scan# 695

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 43 Resp: 67629

Ion Ratio Lower Upper

43

100

86

14.4

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#24

1,1-Dichloroethane

Concen: 19.834 ug/l

RT: 5.921 min Scan# 689

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion: 63 Resp: 174246

Ion Ratio Lower Upper

63

100

98

7.7

4.3




13.1


100

5.1

2.8

8.4





#25

2-Butanone

Concen: 112.949 ug/l

RT: 6.896 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 43 Resp: 15140

Ion Ratio Lower Upper

43 100





72 34.5 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#26

2,2-Dichloropropane

Concen: 20.332 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023561.D




Acq: 24 Oct 2025 09:52


Tgt Ion: 77 Resp: 163903

Ion Ratio Lower Upper

77 100

97 25.2 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 20.884 ug/l

RT: 6.896 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 96 Resp: 130660

Ion Ratio Lower Upper

96 100

61 121.2 0.0 258.2





98 65.3 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#28

Bromochloromethane

Concen: 18.871 ug/l

RT: 7.250 min Scan# 907

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion: 49 Resp: 61891


Ion Ratio Lower Upper

49 100

129 2.9 0.0 5.6

130 113.1 81.9 122.9





#29

Tetrahydrofuran

Concen: 90.887 ug/l

RT: 7.268 min Scan# 910

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :





VY1024SBS01

Tgt Ion:	42	Resp:	6908
Ion	Ratio	Lower	Upper
42	100		
72	59.0	43.0	64.6
71	55.3	40.1	60.1

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#30

Chloroform

Concen: 20.417 ug/l




RT: 7.420 min Scan# 935


Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52


Tgt Ion:	83	Resp:	197077
Ion	Ratio	Lower	Upper
83	100		
85	65.0	52.5	78.7





#31
Cyclohexane
Concen: 18.272 ug/l
RT: 7.701 min Scan# 981
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion: 56 Resp: 140190
Ion Ratio Lower Upper
56 100
69 41.3 30.0 45.0
84 101.5 81.2 121.8

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#32
1,1,1-Trichloroethane
Concen: 20.479 ug/l
RT: 7.615 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52



Tgt Ion: 97 Resp: 179638
Ion Ratio Lower Upper
97 100
99 63.8 51.4 77.2
61 38.0 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 44.218 ug/l

RT: 8.060 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 65 Resp: 20393

Ion Ratio Lower Upper

65 100





67 54.0 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion:114 Resp: 799010


Ion Ratio Lower Upper

114 100

63 16.0 0.0 34.2

88 14.6 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.499 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion:113 Resp: 243010

Ion Ratio Lower Upper

113 100

111 102.4 81.8 122.6





192 22.8 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#36

1,1-Dichloropropene

Concen: 20.089 ug/l

RT: 7.841 min Scan# 1004

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion: 75 Resp: 130575


Ion Ratio Lower Upper

75 100

110 40.3 19.9 59.7

77 32.0 25.0 37.6





#37

Ethyl Acetate

Concen: 19.142 ug/l

RT: 6.987 min Scan# 80

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :





VY1024SBS01

Tgt Ion:	43	Resp:	4929
Ion	Ratio	Lower	Upper
43	100		
61	16.7	11.8	17.8
70	13.3	10.5	15.7

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#38

Carbon Tetrachloride

Concen: 20.617 ug/l




RT: 7.823 min Scan# 1001


Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion:	117	Resp:	162485
Ion	Ratio	Lower	Upper
117	100		
119	95.2	76.8	115.2
121	30.7	24.5	36.7





#39

Methylcyclohexane

Concen: 19.984 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 83 Resp: 16911

Ion Ratio Lower Upper

83 100

55 65.3 51.4 77.2





98 49.7 40.4 60.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#40

Benzene

Concen: 20.712 ug/l

RT: 8.079 min Scan# 1043

Delta R.T. -0.000 min

Lab File: VY023561.D




Acq: 24 Oct 2025 09:52


Tgt Ion: 78 Resp: 426341

Ion Ratio Lower Upper

78 100

77 24.5 18.3 27.5





#41

Methacrylonitrile

Concen: 18.956 ug/l m

RT: 7.225 min Scan# 902

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 41 Resp: 26399

Ion Ratio Lower Upper

41 100

39 48.1 52.2 78.4

67 98.3 74.9 112.3





52 37.0 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#42

1,2-Dichloroethane

Concen: 19.599 ug/l

RT: 8.158 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VY023561.D




Acq: 24 Oct 2025 09:52


Tgt Ion: 62 Resp: 104042

Ion Ratio Lower Upper

62 100

98 11.8 0.0 21.6





#43

Isopropyl Acetate

Concen: 18.643 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023561.D


Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01






Tgt Ion:	43	Resp:	9572
Ion Ratio	100	Lower	Upper
43	100		
61	25.4	18.1	27.1
87	18.4	13.5	20.3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#44

Trichloroethene


Concen: 21.383 ug/l

RT: 8.865 min Scan# 1172



Delta R.T. -0.000 min


Lab File: VY023561.D

Acq: 24 Oct 2025 09:52



Tgt Ion:	130	Resp:	128825
Ion Ratio	100	Lower	Upper
130	100		
95	85.1	0.0	180.6





#45

1,2-Dichloropropane

Concen: 20.304 ug/l

RT: 9.146 min Scan# 1217

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 63 Resp: 9231

Ion Ratio Lower Upper

63 100





65 31.2 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#46

Dibromomethane

Concen: 20.831 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion: 93 Resp: 59886


Ion Ratio Lower Upper

93 100

95 83.6 65.4 98.0

174 121.2 92.9 139.3





#47

Bromodichloromethane

Concen: 20.839 ug/l

RT: 9.426 min Scan# 1263

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 83 Resp: 15088

Ion Ratio Lower Upper

83 100

85 64.7 50.8 76.2





127 9.2 7.4 11.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#48

Methyl methacrylate

Concen: 17.966 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion: 41 Resp: 43337


Ion Ratio Lower Upper

41 100

69 116.4 81.0 121.4

39 63.2 48.2 72.4





#49

1,4-Dioxane

Concen: 389.621 ug/l

RT: 9.231 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion: 88 Resp: 1298

Ion Ratio Lower Upper

88 100

43 27.4 21.8 32.8





58 55.9 46.8 70.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#50

Toluene-d8

Concen: 48.371 ug/l

RT: 10.109 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VY023561.D




Acq: 24 Oct 2025 09:52


Tgt Ion: 98 Resp: 884395

Ion Ratio Lower Upper

98 100

100 66.3 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 96.362 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion: 43 Resp: 25563
Ion Ratio Lower Upper
43 100
58 46.6 35.3 52.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#52
Toluene
Concen: 21.281 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion: 92 Resp: 284413
Ion Ratio Lower Upper
92 100
91 168.8 136.4 204.6









#53
t-1,3-Dichloropropene
Concen: 20.264 ug/l
RT: 10.395 min Scan# 1422
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion: 75 Resp: 12757
Ion Ratio Lower Upper
75 100
77 31.7 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#54
cis-1,3-Dichloropropene
Concen: 20.608 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion: 75 Resp: 153131
Ion Ratio Lower Upper
75 100
77 31.1 25.4 38.0
39 38.3 32.8 49.2





#55
1,1,2-Trichloroethane
Concen: 21.587 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion: 97 Resp: 83454
Ion Ratio Lower Upper
97 100
83 80.2 61.8 92.6
85 53.9 41.3 61.9
99 62.3 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#56
Ethyl methacrylate
Concen: 19.752 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion: 69 Resp: 88085
Ion Ratio Lower Upper
69 100
41 50.5 44.6 67.0
39 31.1 27.0 40.4





#57
1,3-Dichloropropane
Concen: 20.716 ug/l
RT: 10.718 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion: 76 Resp: 126908
Ion Ratio Lower Upper
76 100
78 33.4 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#58
2-Chloroethyl Vinyl ether
Concen: 99.163 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion: 63 Resp: 198103
Ion Ratio Lower Upper
63 100
106 34.9 26.6 39.8





#59
2-Hexanone
Concen: 111.416 ug/l
RT: 10.761 min Scan# 1482
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion: 43 Resp: 211229
Ion Ratio Lower Upper
43 100
58 64.1 30.5 91.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#60
Dibromochloromethane
Concen: 21.342 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Tgt Ion:129 Resp: 115677
Ion Ratio Lower Upper
129 100
127 77.4 38.6 115.7





#61
1,2-Dibromoethane
Concen: 21.377 ug/l
RT: 11.017 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion:107 Resp: 7881
Ion Ratio Lower Upper
107 100
109 94.3 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#62
4-Bromofluorobenzene
Concen: 49.317 ug/l
RT: 12.407 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion: 95 Resp: 297693
Ion Ratio Lower Upper
95 100
174 97.7 0.0 192.8
176 94.3 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.413 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion:117 Resp: 70971
Ion Ratio Lower Upper
117 100
82 48.2 41.0 61.4
119 31.4 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#64
Tetrachloroethene
Concen: 20.639 ug/l
RT: 10.645 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Tgt Ion:164 Resp: 151245
Ion Ratio Lower Upper
164 100
166 128.1 99.4 149.0
129 92.1 69.4 104.2
131 84.7 67.5 101.3





#65
Chlorobenzene
Concen: 20.861 ug/l
RT: 11.444 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion:112 Resp: 322004
Ion Ratio Lower Upper
112 100
114 31.7 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 21.195 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion:131 Resp: 116589
Ion Ratio Lower Upper
131 100
133 95.4 48.4 145.2
119 64.8 32.4 97.2





#67
Ethyl Benzene
Concen: 20.521 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion: 91 Resp: 51275
Ion Ratio Lower Upper
91 100
106 32.9 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#68
m/p-Xylenes
Concen: 42.257 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Tgt Ion:106 Resp: 423494
Ion Ratio Lower Upper
106 100
91 189.9 152.6 229.0





#69
o-Xylene
Concen: 21.072 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion:106 Resp: 198083
Ion Ratio Lower Upper
106 100
91 198.1 101.4 304.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#70
Styrene
Concen: 20.949 ug/l
RT: 11.968 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Tgt Ion:104 Resp: 323830
Ion Ratio Lower Upper
104 100
78 43.4 36.1 54.1
103 53.7 43.2 64.8





#71
Bromoform
Concen: 20.917 ug/l
RT: 12.133 min Scan# 1707
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion:173 Resp: 71372
Ion Ratio Lower Upper
173 100
175 48.9 24.8 74.3
254 0.3 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion:152 Resp: 380713
Ion Ratio Lower Upper
152 100
115 52.6 27.1 81.2
150 163.0 0.0 347.4





#73
Isopropylbenzene
Concen: 20.581 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion:105 Resp: 516359
Ion Ratio Lower Upper
105 100
120 27.9 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#74
N-ethyl acetate
Concen: 17.669 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion: 43 Resp: 79625
Ion Ratio Lower Upper
43 100
70 56.3 41.1 61.7
55 30.4 23.0 34.6
61 28.0 21.4 32.2





#75

1,1,2,2-Tetrachloroethane

Concen: 20.318 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01





Tgt Ion: 83	Resp: 8337
Ion Ratio	Lower Upper
83 100	
131 11.8	5.9 17.6
85 64.5	31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#76

1,2,3-Trichloropropane

Concen: 18.195 ug/l m




RT: 12.553 min Scan# 1777


Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion: 75	Resp: 61116
Ion Ratio	Lower Upper
75 100	
77 0.0	0.0 0.0









#77
Bromobenzene
Concen: 20.547 ug/l
RT: 12.529 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion:156 Resp: 137052
Ion Ratio Lower Upper
156 100
77 149.4 76.8 230.5
158 98.3 49.0 147.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#78
n-propylbenzene
Concen: 20.520 ug/l
RT: 12.596 min Scan# 1784
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion: 91 Resp: 595772
Ion Ratio Lower Upper
91 100
120 25.1 12.3 36.8









#79
2-Chlorotoluene
Concen: 20.287 ug/l
RT: 12.675 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01

Tgt Ion: 91 Resp: 344910
Ion Ratio Lower Upper
91 100
126 38.3 18.5 55.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#80
1,3,5-Trimethylbenzene
Concen: 20.686 ug/l
RT: 12.736 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion:105 Resp: 419394
Ion Ratio Lower Upper
105 100
120 52.0 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 19.665 ug/l
RT: 12.304 min Scan# 1735
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion: 75 Resp: 26930
Ion Ratio Lower Upper
75 100
53 80.9 68.6 102.8
89 46.9 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#82
4-Chlorotoluene
Concen: 20.484 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52


Tgt Ion: 91 Resp: 358669
Ion Ratio Lower Upper
91 100
126 37.5 18.4 55.4





#83
tert-Butylbenzene
Concen: 20.942 ug/l
RT: 12.999 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01



Tgt Ion:119 Resp: 392010
Ion Ratio Lower Upper
119 100
91 58.8 29.9 89.8
134 26.9 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.792 ug/l
RT: 13.041 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52

Tgt Ion:105 Resp: 420704
Ion Ratio Lower Upper
105 100
120 48.4 23.8 71.2





#85

sec-Butylbenzene

Concen: 20.869 ug/l

RT: 13.175 min Scan# 1878

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01

Tgt Ion:105 Resp: 560392

Ion Ratio Lower Upper

105 100





134 21.6 10.7 32.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#86

p-Isopropyltoluene

Concen: 20.853 ug/l

RT: 13.291 min Scan# 1898

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52




Tgt Ion:119 Resp: 476750


Ion Ratio Lower Upper

119 100

134 27.7 13.6 40.7

91 21.7 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 20.394 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01





Tgt Ion:	146	Resp:	264890
Ion	Ratio	Lower	Upper
146	100		
111	36.6	18.7	56.1
148	63.9	31.9	95.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#88

1,4-Dichlorobenzene

Concen: 20.584 ug/l




RT: 13.364 min Scan# 1910


Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion:	146	Resp:	264053
Ion	Ratio	Lower	Upper
146	100		
111	36.4	18.0	54.0
148	63.9	31.9	95.7





#89

n-Butylbenzene

Concen: 20.154 ug/l

RT: 13.614 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01





Tgt Ion:	91	Resp:	401494
Ion	Ratio	Lower	Upper
91	100		
92	53.3	27.0	80.8
134	29.0	14.1	42.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#90

Hexachloroethane

Concen: 20.324 ug/l




RT: 13.876 min Scan# 1994


Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52


Tgt Ion:	117	Resp:	99943
Ion	Ratio	Lower	Upper
117	100		
201	86.1	43.8	131.4





#91
1,2-Dichlorobenzene
Concen: 20.419 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52




Instrument :
MSVOA_Y
ClientSampleId :
VY1024SBS01




Tgt Ion:146 Resp: 23252
Ion Ratio Lower Upper
146 100
111 38.0 19.0 57.0
148 63.7 32.1 96.3

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 19.144 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023561.D
Acq: 24 Oct 2025 09:52



Tgt Ion: 75 Resp: 12890
Ion Ratio Lower Upper
75 100
155 103.4 50.0 150.0
157 132.1 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 19.937 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :





VY1024SBS01

Tgt Ion:180	Resp:	14153
Ion Ratio	Lower	Upper
180	100	
182	95.7	47.9 143.6
145	28.5	14.7 44.1

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#94

Hexachlorobutadiene

Concen: 20.387 ug/l




RT: 15.023 min Scan# 2182


Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion:225	Resp:	90220
Ion Ratio	Lower	Upper
225	100	
223	61.7	30.8 92.3
227	64.1	31.9 95.7





#95

Naphthalene

Concen: 18.584 ug/l

RT: 15.144 min Scan# 2109

Delta R.T. 0.006 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Instrument :

MSVOA_Y

ClientSampleId :

VY1024SBS01





Tgt Ion:	128	Resp:	2109
Ion Ratio	Lower	Upper	
128	100		
127	13.2	10.4	15.6
129	11.3	8.9	13.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#96

1,2,3-Trichlorobenzene

Concen: 19.609 ug/l




RT: 15.327 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023561.D

Acq: 24 Oct 2025 09:52

Tgt Ion:	180	Resp:	120577
Ion Ratio	Lower	Upper	
180	100		
182	96.3	47.9	143.8
145	30.6	15.3	45.9





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: AECOM
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
Client Sample ID: PR-132-S16-015094-20251C
Lab Sample ID: Q3419-02MS
Analytical Method: 8260D
Sample Wt/Vol: 4.51 g

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/21/25
Date Received: 10/21/25
SDG No.: Q3419
Matrix: SOIL
% Solid: 54.4
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	130		1	2.30	10.2	ug/Kg	10/24/25 17:58	VY102425
74-87-3	Chloromethane	110		1	2.30	10.2	ug/Kg	10/24/25 17:58	VY102425
75-01-4	Vinyl Chloride	120		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
74-83-9	Bromomethane	76.6		1	2.20	10.2	ug/Kg	10/24/25 17:58	VY102425
75-00-3	Chloroethane	130		1	2.60	10.2	ug/Kg	10/24/25 17:58	VY102425
75-69-4	Trichlorofluoromethane	130		1	2.50	10.2	ug/Kg	10/24/25 17:58	VY102425
76-13-1	1,1,2-Trichlorotrifluoroethane	120		1	2.20	10.2	ug/Kg	10/24/25 17:58	VY102425
75-35-4	1,1-Dichloroethene	130		1	2.00	10.2	ug/Kg	10/24/25 17:58	VY102425
67-64-1	Acetone	950		1	9.70	50.9	ug/Kg	10/24/25 17:58	VY102425
75-15-0	Carbon Disulfide	97.7		1	2.20	10.2	ug/Kg	10/24/25 17:58	VY102425
1634-04-4	Methyl tert-butyl Ether	150		1	1.50	10.2	ug/Kg	10/24/25 17:58	VY102425
79-20-9	Methyl Acetate	360	E	1	3.10	10.2	ug/Kg	10/24/25 17:58	VY102425
75-09-2	Methylene Chloride	130		1	7.20	20.4	ug/Kg	10/24/25 17:58	VY102425
156-60-5	trans-1,2-Dichloroethene	110		1	1.80	10.2	ug/Kg	10/24/25 17:58	VY102425
75-34-3	1,1-Dichloroethane	130		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
110-82-7	Cyclohexane	62.0		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
78-93-3	2-Butanone	870		1	13.3	50.9	ug/Kg	10/24/25 17:58	VY102425
56-23-5	Carbon Tetrachloride	70.9		1	2.00	10.2	ug/Kg	10/24/25 17:58	VY102425
156-59-2	cis-1,2-Dichloroethene	120		1	1.50	10.2	ug/Kg	10/24/25 17:58	VY102425
74-97-5	Bromochloromethane	87.2		1	2.30	10.2	ug/Kg	10/24/25 17:58	VY102425
67-66-3	Chloroform	130		1	1.70	10.2	ug/Kg	10/24/25 17:58	VY102425
71-55-6	1,1,1-Trichloroethane	130		1	1.90	10.2	ug/Kg	10/24/25 17:58	VY102425
108-87-2	Methylcyclohexane	33.3		1	1.90	10.2	ug/Kg	10/24/25 17:58	VY102425
71-43-2	Benzene	130		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
107-06-2	1,2-Dichloroethane	140		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
79-01-6	Trichloroethene	130		1	1.70	10.2	ug/Kg	10/24/25 17:58	VY102425
78-87-5	1,2-Dichloropropane	130		1	1.90	10.2	ug/Kg	10/24/25 17:58	VY102425
75-27-4	Bromodichloromethane	120		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
108-10-1	4-Methyl-2-Pentanone	890		1	7.30	50.9	ug/Kg	10/24/25 17:58	VY102425
108-88-3	Toluene	110		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
10061-02-6	t-1,3-Dichloropropene	63.4		1	1.30	10.2	ug/Kg	10/24/25 17:58	VY102425
10061-01-5	cis-1,3-Dichloropropene	52.7		1	1.30	10.2	ug/Kg	10/24/25 17:58	VY102425
79-00-5	1,1,2-Trichloroethane	150		1	1.90	10.2	ug/Kg	10/24/25 17:58	VY102425
591-78-6	2-Hexanone	830		1	7.50	50.9	ug/Kg	10/24/25 17:58	VY102425
124-48-1	Dibromochloromethane	120		1	1.80	10.2	ug/Kg	10/24/25 17:58	VY102425
106-93-4	1,2-Dibromoethane	130		1	1.80	10.2	ug/Kg	10/24/25 17:58	VY102425
127-18-4	Tetrachloroethene	100		1	2.10	10.2	ug/Kg	10/24/25 17:58	VY102425
108-90-7	Chlorobenzene	120		1	1.90	10.2	ug/Kg	10/24/25 17:58	VY102425
100-41-4	Ethyl Benzene	96.1		1	1.40	10.2	ug/Kg	10/24/25 17:58	VY102425
179601-23-1	m/p-Xylenes	190		1	2.50	20.4	ug/Kg	10/24/25 17:58	VY102425
95-47-6	o-Xylene	95.7		1	1.70	10.2	ug/Kg	10/24/25 17:58	VY102425
100-42-5	Styrene	87.0		1	1.40	10.2	ug/Kg	10/24/25 17:58	VY102425
75-25-2	Bromoform	150		1	1.80	10.2	ug/Kg	10/24/25 17:58	VY102425



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM	Date Collected:	10/21/25
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received:	10/21/25
Client Sample ID:	PR-132-S16-015094-202510	SDG No.:	Q3419
Lab Sample ID:	Q3419-02MS	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	54.4
Sample Wt/Vol:	4.51 g	Level :	LOW
		Final Vol:	5000 uL
		Test:	VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	100		1	1.60	10.2	ug/Kg	10/24/25 17:58	VY102425
79-34-5	1,1,2,2-Tetrachloroethane	290		1	2.50	10.2	ug/Kg	10/24/25 17:58	VY102425
541-73-1	1,3-Dichlorobenzene	77.7		1	3.50	10.2	ug/Kg	10/24/25 17:58	VY102425
106-46-7	1,4-Dichlorobenzene	82.3		1	3.20	10.2	ug/Kg	10/24/25 17:58	VY102425
95-50-1	1,2-Dichlorobenzene	86.7		1	3.00	10.2	ug/Kg	10/24/25 17:58	VY102425
96-12-8	1,2-Dibromo-3-Chloropropane	240		1	3.70	10.2	ug/Kg	10/24/25 17:58	VY102425
120-82-1	1,2,4-Trichlorobenzene	26.5		1	6.10	10.2	ug/Kg	10/24/25 17:58	VY102425
87-61-6	1,2,3-Trichlorobenzene	28.0		1	6.50	10.2	ug/Kg	10/24/25 17:58	VY102425
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	53.2			63 - 155	106%	SPK: 50		
1868-53-7	Dibromofluoromethane	52.5			70 - 134	105%	SPK: 50		
2037-26-5	Toluene-d8	42.6			74 - 123	85%	SPK: 50		
460-00-4	4-Bromofluorobenzene	28.2			17 - 146	56%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	272000							
540-36-3	1,4-Difluorobenzene	386000							
3114-55-4	Chlorobenzene-d5	250000							
3855-82-1	1,4-Dichlorobenzene-d4	80100							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023580.D
 Acq On : 24 Oct 2025 17:58
 Operator : SY/MD
 Sample : Q3419-02MS
 Misc : 4.51g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PR-132-S16-015094-20251021MS

Quant Time: Oct 24 23:55:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	271890	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	386012	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	249668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	80067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	120955	53.200	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 106.400%			
35) Dibromofluoromethane	7.640	113	124451	52.470	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.940%			
50) Toluene-d8	10.109	98	375970	42.564	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 85.120%			
62) 4-Bromofluorobenzene	12.407	95	82232	28.198	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 56.400%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	123303	63.005	ug/l	99
3) Chloromethane	2.074	50	142826	53.573	ug/l	97
4) Vinyl Chloride	2.208	62	208355	60.002	ug/l	97
5) Bromomethane	2.598	94	109512	37.581	ug/l	100
6) Chloroethane	2.738	64	148217	62.618	ug/l	100
7) Trichlorofluoromethane	3.062	101	316598	65.665	ug/l	99
8) Diethyl Ether	3.458	74	89273	70.860	ug/l	90
9) 1,1,2-Trichlorotrifluo...	3.818	101	141888	56.712	ug/l	98
10) Methyl Iodide	4.013	142	66058	23.717	ug/l	99
11) Tert butyl alcohol	4.860	59	72888	430.929	ug/l	99
12) 1,1-Dichloroethene	3.793	96	160081	66.004	ug/l	96
13) Acrolein	3.665	56	1782	7.456	ug/l #	78
14) Allyl chloride	4.391	41	135719	45.916	ug/l	99
15) Acrylonitrile	5.061	53	194034	402.443	ug/l	99
16) Acetone	3.872	43	251825	465.250	ug/l	96
17) Carbon Disulfide	4.110	76	332900	47.958	ug/l	98
18) Methyl Acetate	4.391	43	249670	175.841	ug/l	98
19) Methyl tert-butyl Ether	5.122	73	437575	75.068	ug/l	97
20) Methylene Chloride	4.622	84	187168	63.948	ug/l	97
21) trans-1,2-Dichloroethene	5.122	96	147659	55.299	ug/l	98
22) Diisopropyl ether	6.024	45	402936	61.629	ug/l	95
23) Vinyl Acetate	5.970	43	176253	49.804	ug/l	95
24) 1,1-Dichloroethane	5.921	63	279950	64.640	ug/l	99
25) 2-Butanone	6.896	43	283296	428.715	ug/l	98
26) 2,2-Dichloropropane	6.890	77	245108	61.678	ug/l	99
27) cis-1,2-Dichloroethene	6.896	96	186769	60.553	ug/l	96
28) Bromochloromethane	7.250	49	69155	42.773	ug/l	94
29) Tetrahydrofuran	7.268	42	158444	422.831	ug/l	94
30) Chloroform	7.427	83	313842	65.956	ug/l	97
31) Cyclohexane	7.707	56	115020	30.411	ug/l	98
32) 1,1,1-Trichloroethane	7.622	97	272172	62.941	ug/l	99
36) 1,1-Dichloropropene	7.841	75	179266	57.089	ug/l	99
37) Ethyl Acetate	6.988	43	78675	63.243	ug/l	96
38) Carbon Tetrachloride	7.817	117	132420	34.778	ug/l	94
39) Methylcyclohexane	9.109	83	66748	16.326	ug/l	97
40) Benzene	8.085	78	632004	63.553	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023580.D
 Acq On : 24 Oct 2025 17:58
 Operator : SY/MD
 Sample : Q3419-02MS
 Misc : 4.51g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PR-132-S16-015094-20251021MS

Quant Time: Oct 24 23:55:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.225	41	55231	82.089	ug/l #	89
42) 1,2-Dichloroethane	8.158	62	179164	69.858	ug/l	99
43) Isopropyl Acetate	8.201	43	176116	70.997	ug/l	99
44) Trichloroethene	8.872	130	180969	62.175	ug/l	100
45) 1,2-Dichloropropane	9.146	63	140373	63.908	ug/l	97
46) Dibromomethane	9.231	93	98831	71.157	ug/l	98
47) Bromodichloromethane	9.426	83	210037	60.045	ug/l	99
48) Methyl methacrylate	9.225	41	97244	83.445	ug/l	91
49) 1,4-Dioxane	9.231	88	29028	1802.611	ug/l	93
51) 4-Methyl-2-Pentanone	9.999	43	558124	435.476	ug/l	98
52) Toluene	10.170	92	340114	52.676	ug/l	100
53) t-1,3-Dichloropropene	10.396	75	94649	31.118	ug/l	99
54) cis-1,3-Dichloropropene	9.859	75	92845	25.864	ug/l	97
55) 1,1,2-Trichloroethane	10.572	97	139759	74.829	ug/l	98
56) Ethyl methacrylate	10.438	69	135479	62.883	ug/l	96
57) 1,3-Dichloropropane	10.719	76	207123	69.983	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	350506	363.168	ug/l	99
59) 2-Hexanone	10.761	43	375076	409.508	ug/l	96
60) Dibromochloromethane	10.914	129	151716	57.938	ug/l	99
61) 1,2-Dibromoethane	11.017	107	117891	66.185	ug/l	99
64) Tetrachloroethene	10.646	164	129361	50.180	ug/l	99
65) Chlorobenzene	11.444	112	306560	56.455	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.517	131	114295	59.065	ug/l	99
67) Ethyl Benzene	11.517	91	414287	47.133	ug/l	97
68) m/p-Xylenes	11.627	106	320246	90.836	ug/l	100
69) o-Xylene	11.956	106	155319	46.969	ug/l	99
70) Styrene	11.969	104	232107	42.682	ug/l	99
71) Bromoform	12.133	173	87609	72.989	ug/l #	100
73) Isopropylbenzene	12.255	105	262027	49.659	ug/l	99
74) N-amyl acetate	12.072	43	33094	34.919	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	121863	141.207	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	77567m	109.805	ug/l	
77) Bromobenzene	12.529	156	101958	72.684	ug/l	93
78) n-propylbenzene	12.596	91	237385	38.877	ug/l	99
79) 2-Chlorotoluene	12.682	91	171658	48.007	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	151905	35.627	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	7609	26.414	ug/l	93
82) 4-Chlorotoluene	12.773	91	175602	47.685	ug/l	99
83) tert-Butylbenzene	12.999	119	123521	31.376	ug/l	100
84) 1,2,4-Trimethylbenzene	13.041	105	158575	37.266	ug/l	99
85) sec-Butylbenzene	13.176	105	129955	23.012	ug/l	99
86) p-Isopropyltoluene	13.291	119	103491	21.524	ug/l	98
87) 1,3-Dichlorobenzene	13.285	146	104101	38.110	ug/l	98
88) 1,4-Dichlorobenzene	13.365	146	108993	40.400	ug/l	98
89) n-Butylbenzene	13.615	91	64229	15.331	ug/l	99
90) Hexachloroethane	13.877	117	5641	5.455	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	101845	42.526	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.273	75	16707	117.984	ug/l	91
93) 1,2,4-Trichlorobenzene	14.919	180	19425	13.011	ug/l	92
94) Hexachlorobutadiene	15.023	225	5601	6.018	ug/l	96
95) Naphthalene	15.139	128	79755	33.546	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	17795	13.760	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023580.D
Acq On : 24 Oct 2025 17:58
Operator : SY/MD
Sample : Q3419-02MS
Misc : 4.51g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 24 Sample Multiplier: 1

Quant Time: Oct 24 23:55:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS

Manual Integrations


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 99
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion:168 Resp: 271890
Ion Ratio Lower Upper
168 100
99 44.9 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#2
Dichlorodifluoromethane
Concen: 63.005 ug/l
RT: 1.867 min Scan# 24
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion: 85 Resp: 123303
Ion Ratio Lower Upper
85 100
87 31.1 16.0 47.9





#3
Chloromethane
Concen: 53.573 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion: 50 Resp: 142820
Ion Ratio Lower Upper
50 100
52 34.2 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#4
Vinyl Chloride
Concen: 60.002 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion: 62 Resp: 208355
Ion Ratio Lower Upper
62 100
64 33.8 25.8 38.8









#5
Bromomethane
Concen: 37.581 ug/l
RT: 2.598 min Scan# 143
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS

Tgt Ion: 94 Resp: 109512
Ion Ratio Lower Upper
94 100
96 93.4 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#6
Chloroethane
Concen: 62.618 ug/l
RT: 2.738 min Scan# 167
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion: 64 Resp: 148217
Ion Ratio Lower Upper
64 100
66 32.7 26.2 39.4





#7
Trichlorofluoromethane
Concen: 65.665 ug/l
RT: 3.062 min Scan# 219
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion: 101 Resp: 316598
Ion Ratio Lower Upper
101 100
103 63.3 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#8
Diethyl Ether
Concen: 70.860 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion: 74 Resp: 89273
Ion Ratio Lower Upper
74 100
45 70.8 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 56.712 ug/l
RT: 3.818 min Scan# 343
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS

Tgt Ion:101 Resp: 14188
Ion Ratio Lower Upper
101 100
85 43.0 34.6 51.8
151 86.3 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#10
Methyl Iodide
Concen: 23.717 ug/l
RT: 4.013 min Scan# 376
Delta R.T. 0.012 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion:142 Resp: 66058
Ion Ratio Lower Upper
142 100
127 43.0 34.8 52.2
141 14.4 11.6 17.4





#11

Tert butyl alcohol

Concen: 430.929 ug/l

RT: 4.860 min Scan# 515

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 59 Resp: 7288

Ion Ratio Lower Upper

59 100





57 10.0 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#12

1,1-Dichloroethene

Concen: 66.004 ug/l

RT: 3.793 min Scan# 340

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion: 96 Resp: 160081


Ion Ratio Lower Upper

96 100

61 140.8 106.7 160.1

98 63.3 50.2 75.4





#13

Acrolein

Concen: 7.456 ug/l

RT: 3.665 min Scan# 316

Delta R.T. 0.018 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion: 56 Resp: 178

Ion Ratio Lower Upper

56 100




55 88.1 56.1 84.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#14

Allyl chloride

Concen: 45.916 ug/l

RT: 4.391 min Scan# 438

Delta R.T. 0.012 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion: 41 Resp: 135719


Ion Ratio Lower Upper

41 100

39 73.7 59.9 89.9


76 47.7 38.0 57.0





#15
Acrylonitrile
Concen: 402.443 ug/l
RT: 5.061 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion: 53 Resp: 194034
Ion Ratio Lower Upper
53 100
52 80.1 64.6 97.0
51 36.1 29.5 44.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#16
Acetone
Concen: 465.250 ug/l
RT: 3.872 min Scan# 353
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion: 43 Resp: 251825
Ion Ratio Lower Upper
43 100
58 36.9 27.7 41.5





#17
Carbon Disulfide
Concen: 47.958 ug/l
RT: 4.110 min Scan# 391
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS




Tgt Ion: 76 Resp: 332900
Ion Ratio Lower Upper
76 100
78 9.7 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#18
Methyl Acetate
Concen: 175.841 ug/l
RT: 4.391 min Scan# 438
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58



Tgt Ion: 43 Resp: 249670
Ion Ratio Lower Upper
43 100
74 31.3 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 75.068 ug/l

RT: 5.122 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 73 Resp: 43757

Ion Ratio Lower Upper

73 100





57 20.2 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#20

Methylene Chloride

Concen: 63.948 ug/l

RT: 4.622 min Scan# 476

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion: 84 Resp: 187168




Ion Ratio Lower Upper


84 100

49 95.8 80.2 120.4

51 30.5 25.9 38.9


86 64.3 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 55.299 ug/l
RT: 5.122 min Scan# 51
Delta R.T. 0.012 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS




Tgt Ion: 96 Resp: 147659
Ion Ratio Lower Upper
96 100
61 116.8 92.7 139.1
98 66.1 49.7 74.5

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#22
Diisopropyl ether
Concen: 61.629 ug/l
RT: 6.024 min Scan# 706
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58



Tgt Ion: 45 Resp: 402936
Ion Ratio Lower Upper
45 100
43 52.8 45.6 68.4
87 37.4 28.1 42.1
59 13.3 10.2 15.2





#23

Vinyl Acetate

Concen: 49.804 ug/l

RT: 5.970 min Scan# 695

Delta R.T. 0.012 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 43 Resp: 17625

Ion Ratio Lower Upper

43

100

86

15.3

10.6





15.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#24

1,1-Dichloroethane

Concen: 64.640 ug/l

RT: 5.921 min Scan# 689

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion: 63 Resp: 279950

Ion Ratio Lower Upper

63

100

98

8.0

4.3




13.1


100

5.7

2.8

8.4





#25

2-Butanone

Concen: 428.715 ug/l

RT: 6.896 min Scan# 84

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 43 Resp: 283290

Ion Ratio Lower Upper

43 100





72 33.0 27.4 41.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#26

2,2-Dichloropropane

Concen: 61.678 ug/l

RT: 6.890 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion: 77 Resp: 245108

Ion Ratio Lower Upper

77 100

97 24.8 12.1 36.3





#27

cis-1,2-Dichloroethene

Concen: 60.553 ug/l

RT: 6.896 min Scan# 848

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 96 Resp: 186769

Ion Ratio Lower Upper

96 100

61 122.8 0.0 258.2





98 66.1 0.0 132.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#28

Bromochloromethane

Concen: 42.773 ug/l

RT: 7.250 min Scan# 907

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion: 49 Resp: 69155


Ion Ratio Lower Upper

49 100

129 2.9 0.0 5.6

130 108.2 81.9 122.9





#29

Tetrahydrofuran

Concen: 422.831 ug/l

RT: 7.268 min Scan# 910

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 42 Resp: 15844

Ion Ratio Lower Upper

42 100

72 58.3 43.0 64.6





71 53.6 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#30

Chloroform

Concen: 65.956 ug/l

RT: 7.427 min Scan# 936

Delta R.T. 0.006 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion: 83 Resp: 313842

Ion Ratio Lower Upper

83 100

85 63.1 52.5 78.7





#31

Cyclohexane

Concen: 30.411 ug/l

RT: 7.707 min Scan# 981

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 56 Resp: 115020

Ion Ratio Lower Upper

56 100

69 38.8 30.0 45.0





84 103.3 81.2 121.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#32

1,1,1-Trichloroethane

Concen: 62.941 ug/l

RT: 7.622 min Scan# 968

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion: 97 Resp: 272172


Ion Ratio Lower Upper

97 100

99 64.2 51.4 77.2

61 38.8 31.9 47.9





#33

1,2-Dichloroethane-d4

Concen: 53.200 ug/l

RT: 8.067 min Scan# 1040

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 65 Resp: 12095

Ion Ratio Lower Upper

65 100





67 53.6 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion:114 Resp: 386012


Ion Ratio Lower Upper

114 100

63 16.7 0.0 34.2

88 14.5 0.0 28.4





#35

Dibromofluoromethane

Concen: 52.470 ug/l

RT: 7.640 min Scan# 91

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion:113 Resp: 12445

Ion Ratio Lower Upper

113 100

111 103.2 81.8 122.6





192 22.8 18.0 27.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#36

1,1-Dichloropropene

Concen: 57.089 ug/l

RT: 7.841 min Scan# 1004

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion: 75 Resp: 179266


Ion Ratio Lower Upper

75 100

110 40.2 19.9 59.7

77 32.2 25.0 37.6





#37

Ethyl Acetate

Concen: 63.243 ug/l

RT: 6.988 min Scan# 80

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:	43	Resp:	7867
Ion	Ratio	Lower	Upper
43	100		
61	16.9	11.8	17.8
70	13.9	10.5	15.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#38

Carbon Tetrachloride

Concen: 34.778 ug/l




RT: 7.817 min Scan# 1000


Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:	117	Resp:	132420
Ion	Ratio	Lower	Upper
117	100		
119	90.1	76.8	115.2
121	28.5	24.5	36.7





#39

Methylcyclohexane

Concen: 16.326 ug/l

RT: 9.109 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:	83	Resp:	66748
Ion	Ratio	Lower	Upper
83	100		
55	61.8	51.4	77.2
98	52.8	40.4	60.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#40

Benzene

Concen: 63.553 ug/l




RT: 8.085 min Scan# 1044


Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:	78	Resp:	632004
Ion	Ratio	Lower	Upper
78	100		
77	23.4	18.3	27.5





#41

Methacrylonitrile

Concen: 82.089 ug/l

RT: 7.225 min Scan# 902

Delta R.T. 0.006 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion: 41 Resp: 5523

Ion Ratio Lower Upper

41 100

39 45.4 52.2 78.4

67 97.8 74.9 112.3




52 37.4 29.2 43.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#42

1,2-Dichloroethane

Concen: 69.858 ug/l

RT: 8.158 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion: 62 Resp: 179164

Ion Ratio Lower Upper

62 100

98 11.3 0.0 21.6





#43

Isopropyl Acetate

Concen: 70.997 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion: 43 Resp: 176110

Ion Ratio Lower Upper

43 100

61 22.9

87 17.8

18.1

27.1




20.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#44

Trichloroethene

Concen: 62.175 ug/l

RT: 8.872 min Scan# 1173

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:130 Resp: 180969




Ion Ratio Lower Upper


130 100

95 90.2

0.0

180.6





#45

1,2-Dichloropropane

Concen: 63.908 ug/l

RT: 9.146 min Scan# 1217

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 63 Resp: 14037

Ion Ratio Lower Upper

63 100





65 32.5 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#46

Dibromomethane

Concen: 71.157 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion: 93 Resp: 98831


Ion Ratio Lower Upper

93 100

95 82.7 65.4 98.0

174 118.2 92.9 139.3





#47

Bromodichloromethane

Concen: 60.045 ug/l

RT: 9.426 min Scan# 1264

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:	83	Resp:	21003
Ion	Ratio	Lower	Upper
83	100		
85	64.8	50.8	76.2
127	9.3	7.4	11.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#48

Methyl methacrylate

Concen: 83.445 ug/l




RT: 9.225 min Scan# 1231


Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:	41	Resp:	97244
Ion	Ratio	Lower	Upper
41	100		
69	112.0	81.0	121.4
39	64.0	48.2	72.4





#49

1,4-Dioxane

Concen: 1802.611 ug/l

RT: 9.231 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:	88	Resp:	29028
Ion	Ratio	Lower	Upper
88	100		
43	27.8	21.8	32.8
58	65.9	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#50

Toluene-d8

Concen: 42.564 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58


Tgt Ion:	98	Resp:	375970
Ion	Ratio	Lower	Upper
98	100		
100	65.6	53.0	79.4





#51
4-Methyl-2-Pentanone
Concen: 435.476 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion: 43 Resp: 558124
Ion Ratio Lower Upper
43 100
58 45.3 35.3 52.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#52
Toluene
Concen: 52.676 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion: 92 Resp: 340114
Ion Ratio Lower Upper
92 100
91 170.9 136.4 204.6





#53
t-1,3-Dichloropropene
Concen: 31.118 ug/l
RT: 10.396 min Scan# 1422
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MS




Tgt Ion: 75 Resp: 94649
Ion Ratio Lower Upper
75 100
77 31.1 25.1 37.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#54
cis-1,3-Dichloropropene
Concen: 25.864 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Tgt Ion: 75 Resp: 92845
Ion Ratio Lower Upper
75 100
77 31.0 25.4 38.0
39 38.0 32.8 49.2





#55
1,1,2-Trichloroethane
Concen: 74.829 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MS



Tgt Ion: 97 Resp: 139759
Ion Ratio Lower Upper
97 100
83 79.8 61.8 92.6
85 52.7 41.3 61.9
99 59.3 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#56
Ethyl methacrylate
Concen: 62.883 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion: 69 Resp: 135479
Ion Ratio Lower Upper
69 100
41 51.7 44.6 67.0
39 32.5 27.0 40.4





#57

1,3-Dichloropropane

Concen: 69.983 ug/l

RT: 10.719 min Scan# 1475

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 76 Resp: 20712

Ion Ratio Lower Upper

76 100





78 32.7 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#58

2-Chloroethyl Vinyl ether

Concen: 363.168 ug/l

RT: 9.713 min Scan# 1311

Delta R.T. 0.006 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion: 63 Resp: 350506

Ion Ratio Lower Upper

63 100


106 33.8 26.6 39.8





#59
2-Hexanone
Concen: 409.508 ug/l
RT: 10.761 min Scan# 1482
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion: 43 Resp: 375070
Ion Ratio Lower Upper
43 100
58 63.7 30.5 91.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#60
Dibromochloromethane
Concen: 57.938 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion:129 Resp: 151716
Ion Ratio Lower Upper
129 100
127 76.7 38.6 115.7





#61
1,2-Dibromoethane
Concen: 66.185 ug/l
RT: 11.017 min Scan# 1525
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion:107 Resp: 11789
Ion Ratio Lower Upper
107 100
109 95.8 76.1 114.1




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#62
4-Bromofluorobenzene
Concen: 28.198 ug/l
RT: 12.407 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion: 95 Resp: 82232
Ion Ratio Lower Upper
95 100
174 98.0 0.0 192.8
176 95.1 0.0 187.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11591
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MS




Tgt Ion:117 Resp: 249668
Ion Ratio Lower Upper
117 100
82 48.1 41.0 61.4
119 32.4 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#64
Tetrachloroethene
Concen: 50.180 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58



Tgt Ion:164 Resp: 129361
Ion Ratio Lower Upper
164 100
166 125.7 99.4 149.0
129 88.6 69.4 104.2
131 85.5 67.5 101.3





#65

Chlorobenzene

Concen: 56.455 ug/l

RT: 11.444 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion:112 Resp: 306560

Ion Ratio Lower Upper

112 100




114 31.9 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 59.065 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion:131 Resp: 114295


Ion Ratio Lower Upper

131 100

133 97.3 48.4 145.2

119 65.4 32.4 97.2





#67

Ethyl Benzene

Concen: 47.133 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 91 Resp: 41428

Ion Ratio Lower Upper

91 100





106 34.2 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#68

m/p-Xylenes

Concen: 90.836 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion:106 Resp: 320246

Ion Ratio Lower Upper

106 100


91 191.0 152.6 229.0





#69
o-Xylene
Concen: 46.969 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS




Tgt Ion:106 Resp: 155319
Ion Ratio Lower Upper
106 100
91 201.6 101.4 304.1

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#70
Styrene
Concen: 42.682 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58



Tgt Ion:104 Resp: 232107
Ion Ratio Lower Upper
104 100
78 45.2 36.1 54.1
103 55.7 43.2 64.8





#71

Bromoform

Concen: 72.989 ug/l

RT: 12.133 min Scan# 1707

Delta R.T. 0.006 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion:173 Resp: 87609

Ion Ratio Lower Upper

173 100

175 49.4 24.8 74.3




254 0.2 0.1 0.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion:152 Resp: 80067


Ion Ratio Lower Upper

152 100

115 52.8 27.1 81.2


150 166.8 0.0 347.4





#73
Isopropylbenzene
Concen: 49.659 ug/l
RT: 12.255 min Scan# 1727
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MS



Tgt Ion:105 Resp: 26202
Ion Ratio Lower Upper
105 100
120 27.5 13.9 41.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#74
N-ethyl acetate
Concen: 34.919 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion: 43 Resp: 33094
Ion Ratio Lower Upper
43 100
70 52.6 41.1 61.7
55 32.1 23.0 34.6
61 27.7 21.4 32.2





#75

1,1,2,2-Tetrachloroethane

Concen: 141.207 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion: 83 Resp: 12186

Ion Ratio Lower Upper

83 100

131 11.5 5.9 17.6





85 65.2 31.8 95.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#76

1,2,3-Trichloropropane

Concen: 109.805 ug/l m

RT: 12.554 min Scan# 1777

Delta R.T. -0.000 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion: 75 Resp: 77567

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 72.684 ug/l

RT: 12.529 min Scan# 1773

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:	156	Resp:	101958
Ion Ratio	Lower	Upper	
156	100		
77	168.3	76.8	230.5
158	99.0	49.0	147.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#78

n-propylbenzene

Concen: 38.877 ug/l




RT: 12.596 min Scan# 1784


Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:	91	Resp:	237385
Ion Ratio	Lower	Upper	
91	100		
120	25.0	12.3	36.8









#79
2-Chlorotoluene
Concen: 48.007 ug/l
RT: 12.682 min Scan# 1798
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MS

Tgt Ion: 91 Resp: 171658
Ion Ratio Lower Upper
91 100
126 37.8 18.5 55.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#80
1,3,5-Trimethylbenzene
Concen: 35.627 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58


Tgt Ion:105 Resp: 151905
Ion Ratio Lower Upper
105 100
120 51.3 25.4 76.4





#81
trans-1,4-Dichloro-2-butene
Concen: 26.414 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.006 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MS



Tgt Ion: 75 Resp: 7609
Ion Ratio Lower Upper
75 100
53 91.5 68.6 102.8
89 43.5 39.0 58.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#82
4-Chlorotoluene
Concen: 47.685 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion: 91 Resp: 175602
Ion Ratio Lower Upper
91 100
126 37.4 18.4 55.4





#83

tert-Butylbenzene

Concen: 31.376 ug/l

RT: 12.999 min Scan# 1850

Delta R.T. 0.006 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion:119 Resp: 12352

Ion Ratio Lower Upper

119 100

91 59.7 29.9 89.8




134 26.2 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#84

1,2,4-Trimethylbenzene

Concen: 37.266 ug/l

RT: 13.041 min Scan# 1857

Delta R.T. -0.000 min

Lab File: VY023580.D




Acq: 24 Oct 2025 17:58


Tgt Ion:105 Resp: 158575

Ion Ratio Lower Upper

105 100

120 48.3 23.8 71.2





#85

sec-Butylbenzene

Concen: 23.012 ug/l

RT: 13.176 min Scan# 1878

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS

Tgt Ion:105 Resp: 12995

Ion Ratio Lower Upper

105 100





134 22.0 10.7 32.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#86

p-Isopropyltoluene

Concen: 21.524 ug/l

RT: 13.291 min Scan# 1898

Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion:119 Resp: 103491


Ion Ratio Lower Upper

119 100

134 25.9 13.6 40.7

91 21.5 10.9 32.9





#87

1,3-Dichlorobenzene

Concen: 38.110 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.000 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS



Tgt Ion:146 Resp: 10410

Ion Ratio Lower Upper

146 100

111 39.6 18.7 56.1




148 63.4 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#88

1,4-Dichlorobenzene


Concen: 40.400 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58





Tgt Ion:146 Resp: 108993


Ion Ratio Lower Upper

146 100

111 36.5 18.0 54.0

148 62.2 31.9 95.7





#89

n-Butylbenzene

Concen: 15.331 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.000 min

Lab File: VY023580.D


Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS






Tgt Ion:	91	Resp:	64229
Ion Ratio	100	Lower	Upper
91	100		
92	53.9	27.0	80.8
134	29.7	14.1	42.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#90

Hexachloroethane


Concen: 5.455 ug/l

RT: 13.877 min Scan# 1994



Delta R.T. -0.000 min


Lab File: VY023580.D

Acq: 24 Oct 2025 17:58




Tgt Ion:	117	Resp:	5641
Ion Ratio	100	Lower	Upper
117	100		
201	88.7	43.8	131.4





#91
1,2-Dichlorobenzene
Concen: 42.526 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58




Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MS



Tgt Ion:146 Resp: 10184
Ion Ratio Lower Upper
146 100
111 41.9 19.0 57.0
148 62.8 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 117.984 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY023580.D
Acq: 24 Oct 2025 17:58

Tgt Ion: 75 Resp: 16707
Ion Ratio Lower Upper
75 100
155 105.7 50.0 150.0
157 140.8 63.8 191.5





#93

1,2,4-Trichlorobenzene

Concen: 13.011 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:180	Resp:	1942
Ion Ratio	Lower	Upper
180	100	
182	101.0	47.9 143.6
145	37.7	14.7 44.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#94

Hexachlorobutadiene

Concen: 6.018 ug/l




RT: 15.023 min Scan# 2182


Delta R.T. 0.006 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:225	Resp:	5601
Ion Ratio	Lower	Upper
225	100	
223	67.2	30.8 92.3
227	64.0	31.9 95.7





#95

Naphthalene

Concen: 33.546 ug/l

RT: 15.139 min Scan# 2201

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MS





Tgt Ion:	128	Resp:	7975
Ion	Ratio	Lower	Upper
128	100		
127	13.8	10.4	15.6
129	11.3	8.9	13.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#96

1,2,3-Trichlorobenzene

Concen: 13.760 ug/l




RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY023580.D

Acq: 24 Oct 2025 17:58

Tgt Ion:	180	Resp:	17795
Ion	Ratio	Lower	Upper
180	100		
182	97.3	47.9	143.8
145	33.9	15.3	45.9



Report of Analysis

Client: AECOM
 Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206
 Client Sample ID: PR-132-S16-015094-20251C
 Lab Sample ID: Q3419-03MSD
 Analytical Method: 8260D
 Sample Wt/Vol: 5.82 g

Level : LOW
 Final Vol: 5000 uL

Date Collected: 10/21/25
 Date Received: 10/21/25
 SDG No.: Q3419
 Matrix: SOIL
 % Solid: 54.4
 Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	99.9		1	1.80	7.90	ug/Kg	10/24/25 18:20	VY102425
74-87-3	Chloromethane	89.3		1	1.80	7.90	ug/Kg	10/24/25 18:20	VY102425
75-01-4	Vinyl Chloride	95.4		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
74-83-9	Bromomethane	45.2		1	1.70	7.90	ug/Kg	10/24/25 18:20	VY102425
75-00-3	Chloroethane	100		1	2.00	7.90	ug/Kg	10/24/25 18:20	VY102425
75-69-4	Trichlorofluoromethane	100		1	1.90	7.90	ug/Kg	10/24/25 18:20	VY102425
76-13-1	1,1,2-Trichlorotrifluoroethane	83.0		1	1.70	7.90	ug/Kg	10/24/25 18:20	VY102425
75-35-4	1,1-Dichloroethene	100		1	1.60	7.90	ug/Kg	10/24/25 18:20	VY102425
67-64-1	Acetone	1200		1	7.50	39.5	ug/Kg	10/24/25 18:20	VY102425
75-15-0	Carbon Disulfide	120		1	1.70	7.90	ug/Kg	10/24/25 18:20	VY102425
1634-04-4	Methyl tert-butyl Ether	150		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
79-20-9	Methyl Acetate	430	E	1	2.40	7.90	ug/Kg	10/24/25 18:20	VY102425
75-09-2	Methylene Chloride	100		1	5.60	15.8	ug/Kg	10/24/25 18:20	VY102425
156-60-5	trans-1,2-Dichloroethene	82.6		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425
75-34-3	1,1-Dichloroethane	100		1	1.30	7.90	ug/Kg	10/24/25 18:20	VY102425
110-82-7	Cyclohexane	45.2		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
78-93-3	2-Butanone	970		1	10.3	39.5	ug/Kg	10/24/25 18:20	VY102425
56-23-5	Carbon Tetrachloride	58.9		1	1.50	7.90	ug/Kg	10/24/25 18:20	VY102425
156-59-2	cis-1,2-Dichloroethene	95.3		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
74-97-5	Bromochloromethane	96.1		1	1.80	7.90	ug/Kg	10/24/25 18:20	VY102425
67-66-3	Chloroform	110		1	1.30	7.90	ug/Kg	10/24/25 18:20	VY102425
71-55-6	1,1,1-Trichloroethane	98.5		1	1.50	7.90	ug/Kg	10/24/25 18:20	VY102425
108-87-2	Methylcyclohexane	29.4		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425
71-43-2	Benzene	100		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
107-06-2	1,2-Dichloroethane	120		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
79-01-6	Trichloroethene	89.8		1	1.30	7.90	ug/Kg	10/24/25 18:20	VY102425
78-87-5	1,2-Dichloropropane	110		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425
75-27-4	Bromodichloromethane	100		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
108-10-1	4-Methyl-2-Pentanone	980		1	5.70	39.5	ug/Kg	10/24/25 18:20	VY102425
108-88-3	Toluene	80.6		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
10061-02-6	t-1,3-Dichloropropene	33.6		1	1.00	7.90	ug/Kg	10/24/25 18:20	VY102425
10061-01-5	cis-1,3-Dichloropropene	24.0		1	0.98	7.90	ug/Kg	10/24/25 18:20	VY102425
79-00-5	1,1,2-Trichloroethane	140		1	1.50	7.90	ug/Kg	10/24/25 18:20	VY102425
591-78-6	2-Hexanone	930		1	5.80	39.5	ug/Kg	10/24/25 18:20	VY102425
124-48-1	Dibromochloromethane	110		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425
106-93-4	1,2-Dibromoethane	120		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425
127-18-4	Tetrachloroethene	85.4		1	1.70	7.90	ug/Kg	10/24/25 18:20	VY102425
108-90-7	Chlorobenzene	87.9		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425
100-41-4	Ethyl Benzene	69.7		1	1.10	7.90	ug/Kg	10/24/25 18:20	VY102425
179601-23-1	m/p-Xylenes	130		1	2.00	15.8	ug/Kg	10/24/25 18:20	VY102425
95-47-6	o-Xylene	70.9		1	1.30	7.90	ug/Kg	10/24/25 18:20	VY102425
100-42-5	Styrene	56.7		1	1.10	7.90	ug/Kg	10/24/25 18:20	VY102425
75-25-2	Bromoform	150		1	1.40	7.90	ug/Kg	10/24/25 18:20	VY102425

Report of Analysis

Client: AECOM	Date Collected: 10/21/25	
Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Date Received: 10/21/25	
Client Sample ID: PR-132-S16-015094-202510	SDG No.: Q3419	
Lab Sample ID: Q3419-03MSD	Matrix: SOIL	
Analytical Method: 8260D	% Solid: 54.4	
Sample Wt/Vol: 5.82 g	Test: VOC-TCLVOA-10	
Level: LOW		
Final Vol: 5000 uL		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	77.2		1	1.20	7.90	ug/Kg	10/24/25 18:20	VY102425
79-34-5	1,1,2,2-Tetrachloroethane	280	E	1	1.90	7.90	ug/Kg	10/24/25 18:20	VY102425
541-73-1	1,3-Dichlorobenzene	62.5		1	2.70	7.90	ug/Kg	10/24/25 18:20	VY102425
106-46-7	1,4-Dichlorobenzene	65.6		1	2.50	7.90	ug/Kg	10/24/25 18:20	VY102425
95-50-1	1,2-Dichlorobenzene	72.3		1	2.30	7.90	ug/Kg	10/24/25 18:20	VY102425
96-12-8	1,2-Dibromo-3-Chloropropane	250	E	1	2.90	7.90	ug/Kg	10/24/25 18:20	VY102425
120-82-1	1,2,4-Trichlorobenzene	26.2		1	4.70	7.90	ug/Kg	10/24/25 18:20	VY102425
87-61-6	1,2,3-Trichlorobenzene	29.6		1	5.00	7.90	ug/Kg	10/24/25 18:20	VY102425
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	54.7			63 - 155	109%	SPK: 50		
1868-53-7	Dibromofluoromethane	46.4			70 - 134	93%	SPK: 50		
2037-26-5	Toluene-d8	34.7	*		74 - 123	69%	SPK: 50		
460-00-4	4-Bromofluorobenzene	25.0			17 - 146	50%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	174000							
540-36-3	1,4-Difluorobenzene	246000							
3114-55-4	Chlorobenzene-d5	158000							
3855-82-1	1,4-Dichlorobenzene-d4	48900							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023581.D
 Acq On : 24 Oct 2025 18:20
 Operator : SY/MD
 Sample : Q3419-03MSD
 Misc : 5.82g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PR-132-S16-015094-20251021MSD

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:57:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.713	168	173647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	245581	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	157959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	48940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	79397	54.679	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 109.360%			
35) Dibromofluoromethane	7.640	113	69999	46.388	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 92.780%			
50) Toluene-d8	10.109	98	194718	34.650	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 69.300%			
62) 4-Bromofluorobenzene	12.408	95	46316	24.964	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 49.920%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	79103	63.288	ug/l	99
3) Chloromethane	2.074	50	96244	56.525	ug/l	98
4) Vinyl Chloride	2.208	62	134030	60.435	ug/l	99
5) Bromomethane	2.605	94	53316	28.648	ug/l	97
6) Chloroethane	2.739	64	95948	63.470	ug/l	98
7) Trichlorofluoromethane	3.062	101	198070	64.324	ug/l	95
8) Diethyl Ether	3.458	74	66810	83.032	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.818	101	84002	52.571	ug/l	99
10) Methyl Iodide	4.013	142	19632	11.036	ug/l	99
11) Tert butyl alcohol	4.866	59	64307	595.297	ug/l	100
12) 1,1-Dichloroethene	3.793	96	99787	64.421	ug/l	99
13) Acrolein	3.659	56	2104	13.784	ug/l #	74
14) Allyl chloride	4.391	41	60022	31.795	ug/l	91
15) Acrylonitrile	5.061	53	172997	561.812	ug/l	98
16) Acetone	3.873	43	256528	742.075	ug/l	98
17) Carbon Disulfide	4.110	76	337360	76.096	ug/l	99
18) Methyl Acetate	4.385	43	249309	274.928	ug/l	100
19) Methyl tert-butyl Ether	5.122	73	346226	93.001	ug/l	97
20) Methylene Chloride	4.622	84	123905	66.285	ug/l	98
21) trans-1,2-Dichloroethene	5.122	96	89197	52.303	ug/l	99
22) Diisopropyl ether	6.025	45	270915	64.879	ug/l	97
24) 1,1-Dichloroethane	5.921	63	183785	66.444	ug/l	99
25) 2-Butanone	6.896	43	259730	615.426	ug/l	100
26) 2,2-Dichloropropane	6.890	77	156385	61.616	ug/l	99
27) cis-1,2-Dichloroethene	6.903	96	118912	60.364	ug/l	97
28) Bromochloromethane	7.250	49	62830	60.847	ug/l	96
29) Tetrahydrofuran	7.268	42	146863	613.662	ug/l	95
30) Chloroform	7.427	83	207694	68.343	ug/l	97
31) Cyclohexane	7.707	56	69141	28.623	ug/l	95
32) 1,1,1-Trichloroethane	7.622	97	172234	62.365	ug/l	98
36) 1,1-Dichloropropene	7.841	75	107322	53.722	ug/l	99
37) Ethyl Acetate	6.988	43	60616	76.589	ug/l	97
38) Carbon Tetrachloride	7.823	117	90291	37.274	ug/l	97
39) Methylcyclohexane	9.109	83	48427	18.618	ug/l	96
40) Benzene	8.085	78	408926	64.635	ug/l	100
41) Methacrylonitrile	7.226	41	47920	111.951	ug/l	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023581.D
 Acq On : 24 Oct 2025 18:20
 Operator : SY/MD
 Sample : Q3419-03MSD
 Misc : 5.82g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PR-132-S16-015094-20251021MSD

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
 Supervised By :Mahesh Dadoda 10/27/2025

Quant Time: Oct 24 23:57:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 1,2-Dichloroethane	8.158	62	126898	77.773	ug/l	98
43) Isopropyl Acetate	8.201	43	135161	85.645	ug/l	97
44) Trichloroethene	8.872	130	105286	56.858	ug/l	99
45) 1,2-Dichloropropane	9.146	63	93324	66.784	ug/l	99
46) Dibromomethane	9.231	93	72267	81.785	ug/l	98
47) Bromodichloromethane	9.427	83	146364	65.769	ug/l	100
48) Methyl methacrylate	9.225	41	86669	116.898	ug/l	97
49) 1,4-Dioxane	9.231	88	26544	2590.939	ug/l	96
51) 4-Methyl-2-Pentanone	10.000	43	507478	622.382	ug/l	98
52) Toluene	10.170	92	209716	51.053	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	41168	21.275	ug/l	97
54) cis-1,3-Dichloropropene	9.859	75	34656	15.175	ug/l	98
55) 1,1,2-Trichloroethane	10.573	97	104391	87.853	ug/l	96
56) Ethyl methacrylate	10.439	69	100718	73.481	ug/l	96
57) 1,3-Dichloropropane	10.719	76	154240	81.916	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.713	63	281024	457.679	ug/l	98
59) 2-Hexanone	10.762	43	341891	586.729	ug/l	97
60) Dibromochloromethane	10.914	129	114167	68.530	ug/l	99
61) 1,2-Dibromoethane	11.018	107	85947	75.843	ug/l	99
64) Tetrachloroethene	10.652	164	88229	54.095	ug/l	94
65) Chlorobenzene	11.444	112	191312	55.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.518	131	83504	68.207	ug/l	99
67) Ethyl Benzene	11.518	91	245521	44.150	ug/l	99
68) m/p-Xylenes	11.627	106	190171	85.258	ug/l	98
69) o-Xylene	11.950	106	93972	44.916	ug/l	98
70) Styrene	11.969	104	123591	35.922	ug/l	99
71) Bromoform	12.133	173	69787	91.897	ug/l #	100
73) Isopropylbenzene	12.255	105	157585	48.860	ug/l	99
74) N-amyl acetate	12.072	43	8440	14.569	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.505	83	93161	176.607	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	63376m	146.778	ug/l	
77) Bromobenzene	12.530	156	64793	75.567	ug/l	86
78) n-propylbenzene	12.591	91	142211	38.103	ug/l	98
79) 2-Chlorotoluene	12.682	91	106207	48.594	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	92656	35.552	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.304	75	1014m	5.759	ug/l	
82) 4-Chlorotoluene	12.780	91	104792	46.556	ug/l	97
83) tert-Butylbenzene	12.999	119	80984	33.655	ug/l	98
84) 1,2,4-Trimethylbenzene	13.042	105	99548	38.273	ug/l	98
85) sec-Butylbenzene	13.176	105	91190	26.418	ug/l	98
86) p-Isopropyltoluene	13.292	119	70459	23.974	ug/l	98
87) 1,3-Dichlorobenzene	13.286	146	66124	39.603	ug/l	96
88) 1,4-Dichlorobenzene	13.365	146	68453	41.511	ug/l	99
89) n-Butylbenzene	13.615	91	47002	18.354	ug/l	98
90) Hexachloroethane	13.877	117	7634	12.077	ug/l	98
91) 1,2-Dichlorobenzene	13.657	146	67020	45.784	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.273	75	13660	157.821	ug/l	93
93) 1,2,4-Trichlorobenzene	14.919	180	15146	16.597	ug/l	96
94) Hexachlorobutadiene	15.023	225	5609	9.860	ug/l	95
95) Naphthalene	15.145	128	68160	46.903	ug/l	97
96) 1,2,3-Trichlorobenzene	15.328	180	14795	18.717	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023581.D
Acq On : 24 Oct 2025 18:20
Operator : SY/MD
Sample : Q3419-03MSD
Misc : 5.82g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 25 Sample Multiplier: 1

Quant Time: Oct 24 23:57:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD

Manual Integrations APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD

Tgt Ion:168 Resp: 17364
Ion Ratio Lower Upper
168 100
99 45.9 39.7 59.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#2
Dichlorodifluoromethane
Concen: 63.288 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion: 85 Resp: 79103
Ion Ratio Lower Upper
85 100
87 32.2 16.0 47.9









#3
Chloromethane
Concen: 56.525 ug/l
RT: 2.074 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD

Tgt Ion: 50 Resp: 96244
Ion Ratio Lower Upper
50 100
52 31.5 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#4
Vinyl Chloride
Concen: 60.435 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion: 62 Resp: 134030
Ion Ratio Lower Upper
62 100
64 32.6 25.8 38.8









#5
Bromomethane
Concen: 28.648 ug/l
RT: 2.605 min Scan# 143
Delta R.T. 0.012 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD

Tgt Ion: 94 Resp: 53310
Ion Ratio Lower Upper
94 100
96 95.5 74.5 111.7




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#6
Chloroethane
Concen: 63.470 ug/l
RT: 2.739 min Scan# 167
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20


Tgt Ion: 64 Resp: 95948
Ion Ratio Lower Upper
64 100
66 33.9 26.2 39.4





#7
Trichlorofluoromethane
Concen: 64.324 ug/l
RT: 3.062 min Scan# 219
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion: 101 Resp: 198070
Ion Ratio Lower Upper
101 100
103 68.2 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#8
Diethyl Ether
Concen: 83.032 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion: 74 Resp: 66810
Ion Ratio Lower Upper
74 100
45 76.7 39.9 119.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.571 ug/l
RT: 3.818 min Scan# 343
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD

Tgt Ion:101 Resp: 8400
Ion Ratio Lower Upper
101 100
85 42.9 34.6 51.8
151 85.5 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#10
Methyl Iodide
Concen: 11.036 ug/l
RT: 4.013 min Scan# 376
Delta R.T. 0.012 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion:142 Resp: 19632
Ion Ratio Lower Upper
142 100
127 43.1 34.8 52.2
141 13.2 11.6 17.4





#11

Tert butyl alcohol

Concen: 595.297 ug/l

RT: 4.866 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 59 Resp: 6430

Ion Ratio Lower Upper

59 100





57 10.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#12

1,1-Dichloroethene

Concen: 64.421 ug/l

RT: 3.793 min Scan# 340

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 96 Resp: 99787


Ion Ratio Lower Upper

96 100

61 135.1 106.7 160.1

98 63.7 50.2 75.4





#13

Acrolein

Concen: 13.784 ug/l

RT: 3.659 min Scan# 316

Delta R.T. 0.012 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD






Tgt Ion:	56	Resp:	210
Ion Ratio	100	Lower	Upper
56	100		
55	91.7	56.1	84.1

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#14

Allyl chloride


Concen: 31.795 ug/l

RT: 4.391 min Scan# 438



Delta R.T. 0.012 min


Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion:	41	Resp:	60022
Ion Ratio	100	Lower	Upper
41	100		
39	66.4	59.9	89.9
76	43.1	38.0	57.0





#15
Acrylonitrile
Concen: 561.812 ug/l
RT: 5.061 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion: 53 Resp: 172997
Ion Ratio Lower Upper
53 100
52 79.6 64.6 97.0
51 35.5 29.5 44.3

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#16
Acetone
Concen: 742.075 ug/l
RT: 3.873 min Scan# 353
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Tgt Ion: 43 Resp: 256528
Ion Ratio Lower Upper
43 100
58 35.9 27.7 41.5





#17
Carbon Disulfide
Concen: 76.096 ug/l
RT: 4.110 min Scan# 391
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion: 76 Resp: 337360
Ion Ratio Lower Upper
76 100
78 8.8 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#18
Methyl Acetate
Concen: 274.928 ug/l
RT: 4.385 min Scan# 437
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion: 43 Resp: 249309
Ion Ratio Lower Upper
43 100
74 30.4 24.2 36.4





#19

Methyl tert-butyl Ether

Concen: 93.001 ug/l

RT: 5.122 min Scan# 51

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 73 Resp: 346220

Ion Ratio Lower Upper

73 100





57 20.3 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#20

Methylene Chloride

Concen: 66.285 ug/l

RT: 4.622 min Scan# 476

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Tgt Ion: 84 Resp: 123905




Ion Ratio Lower Upper


84 100

49 99.0 80.2 120.4

51 30.6 25.9 38.9


86 61.6 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 52.303 ug/l
RT: 5.122 min Scan# 51
Delta R.T. 0.012 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion: 96 Resp: 8919
Ion Ratio Lower Upper
96 100
61 114.7 92.7 139.1
98 63.5 49.7 74.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#22
Diisopropyl ether
Concen: 64.879 ug/l
RT: 6.025 min Scan# 706
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion: 45 Resp: 270915
Ion Ratio Lower Upper
45 100
43 53.5 45.6 68.4
87 36.0 28.1 42.1
59 14.0 10.2 15.2





#24

1,1-Dichloroethane

Concen: 66.444 ug/l

RT: 5.921 min Scan# 689

Delta R.T. 0.006 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD



Tgt Ion: 63 Resp: 18378

Ion Ratio Lower Upper

63 100

98 8.5 4.3 13.1




100 5.1 2.8 8.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#25

2-Butanone

Concen: 615.426 ug/l

RT: 6.896 min Scan# 849

Delta R.T. 0.006 min

Lab File: VY023581.D




Acq: 24 Oct 2025 18:20


Tgt Ion: 43 Resp: 259730

Ion Ratio Lower Upper

43 100

72 34.4 27.4 41.2





#26

2,2-Dichloropropane

Concen: 61.616 ug/l

RT: 6.890 min Scan# 847

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 77 Resp: 15638

Ion Ratio Lower Upper

77 100





97 24.5 12.1 36.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#27

cis-1,2-Dichloroethene

Concen: 60.364 ug/l

RT: 6.903 min Scan# 850

Delta R.T. 0.012 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 96 Resp: 118912


Ion Ratio Lower Upper

96 100

61 123.5 0.0 258.2


98 65.7 0.0 132.0





#28
Bromochloromethane
Concen: 60.847 ug/l
RT: 7.250 min Scan# 906
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion: 49 Resp: 62830
Ion Ratio Lower Upper
49 100
129 2.7 0.0 5.6
130 106.2 81.9 122.9

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#29
Tetrahydrofuran
Concen: 613.662 ug/l
RT: 7.268 min Scan# 910
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20



Tgt Ion: 42 Resp: 146863
Ion Ratio Lower Upper
42 100
72 57.7 43.0 64.6
71 53.6 40.1 60.1





#30

Chloroform

Concen: 68.343 ug/l

RT: 7.427 min Scan# 91

Delta R.T. 0.006 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD



Tgt Ion: 83 Resp: 207694

Ion Ratio Lower Upper

83 100




85 63.2 52.5 78.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#31

Cyclohexane


Concen: 28.623 ug/l

RT: 7.707 min Scan# 982

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20





Tgt Ion: 56 Resp: 69141


Ion Ratio Lower Upper

56 100

69 38.7 30.0 45.0

84 107.1 81.2 121.8





#32

1,1,1-Trichloroethane

Concen: 62.365 ug/l

RT: 7.622 min Scan# 90

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 97 Resp: 172234

Ion Ratio Lower Upper

97 100

99 63.3 51.4 77.2





61 38.7 31.9 47.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#33

1,2-Dichloroethane-d4

Concen: 54.679 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VY023581.D




Acq: 24 Oct 2025 18:20


Tgt Ion: 65 Resp: 79397

Ion Ratio Lower Upper

65 100

67 53.3 0.0 105.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion:114 Resp: 24558

Ion Ratio Lower Upper

114 100

63 16.4 0.0 34.2





88 14.6 0.0 28.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#35

Dibromofluoromethane

Concen: 46.388 ug/l

RT: 7.640 min Scan# 971

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion:113 Resp: 69999


Ion Ratio Lower Upper

113 100

111 102.2 81.8 122.6

192 22.3 18.0 27.0





#36

1,1-Dichloropropene

Concen: 53.722 ug/l

RT: 7.841 min Scan# 1003

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 75 Resp: 10732

Ion Ratio Lower Upper

75 100

110 40.8 19.9 59.7





77 31.0 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#37

Ethyl Acetate

Concen: 76.589 ug/l

RT: 6.988 min Scan# 864

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 43 Resp: 60616


Ion Ratio Lower Upper

43 100

61 16.4 11.8 17.8

70 12.6 10.5 15.7





#38

Carbon Tetrachloride

Concen: 37.274 ug/l

RT: 7.823 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD



Tgt Ion:117 Resp: 9029

Ion Ratio Lower Upper

117 100

119 100.0 76.8 115.2




121 31.5 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#39

Methylcyclohexane

Concen: 18.618 ug/l

RT: 9.109 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 83 Resp: 48427


Ion Ratio Lower Upper

83 100

55 68.5 51.4 77.2


98 52.5 40.4 60.6





#40
Benzene
Concen: 64.635 ug/l
RT: 8.085 min Scan# 1044
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion: 78 Resp: 408920
Ion Ratio Lower Upper
78 100
77 22.8 18.3 27.5




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#41
Methacrylonitrile
Concen: 111.951 ug/l
RT: 7.226 min Scan# 903
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion: 41 Resp: 47920
Ion Ratio Lower Upper
41 100
39 55.8 52.2 78.4
67 102.7 74.9 112.3
52 37.4 29.2 43.8





#42

1,2-Dichloroethane

Concen: 77.773 ug/l

RT: 8.158 min Scan# 1055

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 62 Resp: 126898

Ion Ratio Lower Upper

62 100





98 11.6 0.0 21.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#43

Isopropyl Acetate

Concen: 85.645 ug/l

RT: 8.201 min Scan# 1063

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 43 Resp: 135161


Ion Ratio Lower Upper

43 100

61 24.3 18.1 27.1

87 18.0 13.5 20.3





#44

Trichloroethene

Concen: 56.858 ug/l

RT: 8.872 min Scan# 1173

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion:130 Resp: 105280

Ion Ratio Lower Upper

130 100





95 89.1 0.0 180.6

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#45

1,2-Dichloropropane

Concen: 66.784 ug/l

RT: 9.146 min Scan# 1218

Delta R.T. 0.006 min

Lab File: VY023581.D




Acq: 24 Oct 2025 18:20


Tgt Ion: 63 Resp: 93324

Ion Ratio Lower Upper

63 100

65 30.4 24.9 37.3





#46

Dibromomethane

Concen: 81.785 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :





PR-132-S16-015094-20251021MSD

Tgt Ion: 93	Resp: 7226
Ion Ratio	Lower Upper
93 100	
95 85.2	65.4 98.0
174 116.1	92.9 139.3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#47

Bromodichloromethane

Concen: 65.769 ug/l




RT: 9.427 min Scan# 1264


Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Tgt Ion: 83	Resp: 146364
Ion Ratio	Lower Upper
83 100	
85 63.8	50.8 76.2
127 8.7	7.4 11.2





#48

Methyl methacrylate

Concen: 116.898 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 41 Resp: 86669

Ion Ratio Lower Upper

41 100

69 105.7 81.0 121.4





39 61.2 48.2 72.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#49

1,4-Dioxane

Concen: 2590.939 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 88 Resp: 26544


Ion Ratio Lower Upper

88 100

43 27.8 21.8 32.8

58 63.0 46.8 70.2









#50
Toluene-d8
Concen: 34.650 ug/l
RT: 10.109 min Scan# 1375
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MSD

Tgt Ion: 98 Resp: 194718
Ion Ratio Lower Upper
98 100
100 65.3 53.0 79.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#51
4-Methyl-2-Pentanone
Concen: 622.382 ug/l
RT: 10.000 min Scan# 1358
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20


Tgt Ion: 43 Resp: 507478
Ion Ratio Lower Upper
43 100
58 45.2 35.3 52.9





#52
Toluene
Concen: 51.053 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion: 92 Resp: 209710
Ion Ratio Lower Upper
92 100
91 172.1 136.4 204.6

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#53
t-1,3-Dichloropropene
Concen: 21.275 ug/l
RT: 10.396 min Scan# 1423
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20



Tgt Ion: 75 Resp: 41168
Ion Ratio Lower Upper
75 100
77 32.8 25.1 37.7





#54

cis-1,3-Dichloropropene

Concen: 15.175 ug/l

RT: 9.859 min Scan# 1334

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD





Tgt Ion: 75	Resp: 34650
Ion Ratio	Lower Upper
75 100	
77 33.6	25.4 38.0
39 41.4	32.8 49.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#55

1,1,2-Trichloroethane

Concen: 87.853 ug/l




RT: 10.573 min Scan# 1452


Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20


Tgt Ion: 97	Resp: 104391
Ion Ratio	Lower Upper
97 100	
83 81.8	61.8 92.6
85 53.0	41.3 61.9
99 63.0	48.7 73.1





#56
Ethyl methacrylate
Concen: 73.481 ug/l
RT: 10.439 min Scan# 1430
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion: 69 Resp: 100718
Ion Ratio Lower Upper
69 100
41 51.4 44.6 67.0
39 33.1 27.0 40.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#57
1,3-Dichloropropane
Concen: 81.916 ug/l
RT: 10.719 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20


Tgt Ion: 76 Resp: 154240
Ion Ratio Lower Upper
76 100
78 32.7 25.9 38.9





#58
2-Chloroethyl Vinyl ether
Concen: 457.679 ug/l
RT: 9.713 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion: 63 Resp: 281024
Ion Ratio Lower Upper
63 100
106 34.3 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#59
2-Hexanone
Concen: 586.729 ug/l
RT: 10.762 min Scan# 1483
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Tgt Ion: 43 Resp: 341891
Ion Ratio Lower Upper
43 100
58 63.0 30.5 91.5





#60
Dibromochloromethane
Concen: 68.530 ug/l
RT: 10.914 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion:129 Resp: 11416
Ion Ratio Lower Upper
129 100
127 75.9 38.6 115.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#61
1,2-Dibromoethane
Concen: 75.843 ug/l
RT: 11.018 min Scan# 1525
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Tgt Ion:107 Resp: 85947
Ion Ratio Lower Upper
107 100
109 93.8 76.1 114.1





#62
4-Bromofluorobenzene
Concen: 24.964 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument : MSVOA_Y
ClientSampleId : PR-132-S16-015094-20251021MSD



Tgt Ion: 95 Resp: 46310
Ion Ratio Lower Upper
95 100
174 95.9 0.0 192.8
176 93.0 0.0 187.6




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion:117 Resp: 157959
Ion Ratio Lower Upper
117 100
82 48.9 41.0 61.4
119 31.3 25.6 38.4





#64

Tetrachloroethene

Concen: 54.095 ug/l

RT: 10.652 min Scan# 1465

Delta R.T. 0.006 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD



Tgt Ion:164 Resp: 88229

Ion Ratio Lower Upper

164 100

166 120.2 99.4 149.0

129 79.9 69.4 104.2




131 78.6 67.5 101.3

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#65

Chlorobenzene

Concen: 55.686 ug/l

RT: 11.444 min Scan# 1595

Delta R.T. 0.006 min

Lab File: VY023581.D




Acq: 24 Oct 2025 18:20


Tgt Ion:112 Resp: 191312

Ion Ratio Lower Upper

112 100


114 31.9 25.6 38.4





#66
1,1,1,2-Tetrachloroethane
Concen: 68.207 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion:131 Resp: 83504
Ion Ratio Lower Upper
131 100
133 96.0 48.4 145.2
119 66.2 32.4 97.2




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#67
Ethyl Benzene
Concen: 44.150 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20


Tgt Ion: 91 Resp: 245521
Ion Ratio Lower Upper
91 100
106 32.8 25.8 38.8





#68
m/p-Xylenes
Concen: 85.258 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion:106 Resp: 19017
Ion Ratio Lower Upper
106 100
91 187.4 152.6 229.0




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#69
o-Xylene
Concen: 44.916 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion:106 Resp: 93972
Ion Ratio Lower Upper
106 100
91 199.7 101.4 304.1





#70

Styrene

Concen: 35.922 ug/l

RT: 11.969 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD






Tgt Ion:	104	Resp:	12359
Ion Ratio	Lower	Upper	
104	100		
78	45.2	36.1	54.1
103	55.2	43.2	64.8

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#71

Bromoform


Concen: 91.897 ug/l

RT: 12.133 min Scan# 1708



Delta R.T. 0.006 min


Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion:	173	Resp:	69787
Ion Ratio	Lower	Upper	
173	100		
175	49.7	24.8	74.3
254	0.1	0.1	0.1





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion:152 Resp: 48940
Ion Ratio Lower Upper
152 100
115 54.2 27.1 81.2
150 166.0 0.0 347.4

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#73
Isopropylbenzene
Concen: 48.860 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Tgt Ion:105 Resp: 157585
Ion Ratio Lower Upper
105 100
120 27.4 13.9 41.7





#74
N-amy1 acetate
Concen: 14.569 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion: 43 Resp: 8440
Ion Ratio Lower Upper
43 100
70 58.8 41.1 61.7
55 29.5 23.0 34.6
61 27.8 21.4 32.2

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#75
1,1,2,2-Tetrachloroethane
Concen: 176.607 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Tgt Ion: 83 Resp: 93161
Ion Ratio Lower Upper
83 100
131 11.3 5.9 17.6
85 64.5 31.8 95.3





#76
1,2,3-Trichloropropane
Concen: 146.778 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion: 75 Resp: 63370
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#77
Bromobenzene
Concen: 75.567 ug/l
RT: 12.530 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20



Tgt Ion:156 Resp: 64793
Ion Ratio Lower Upper
156 100
77 180.1 76.8 230.5
158 95.2 49.0 147.2





#78

n-propylbenzene

Concen: 38.103 ug/l

RT: 12.591 min Scan# 1783

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion: 91 Resp: 14221

Ion Ratio Lower Upper

91 100





120 25.5 12.3 36.8

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#79

2-Chlorotoluene

Concen: 48.594 ug/l

RT: 12.682 min Scan# 1798

Delta R.T. 0.006 min

Lab File: VY023581.D




Acq: 24 Oct 2025 18:20


Tgt Ion: 91 Resp: 106207

Ion Ratio Lower Upper

91 100

126 37.7 18.5 55.5





#80

1,3,5-Trimethylbenzene

Concen: 35.552 ug/l

RT: 12.737 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion:105 Resp: 92650

Ion Ratio Lower Upper

105 100





120 52.5 25.4 76.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#81

trans-1,4-Dichloro-2-butene

Concen: 5.759 ug/l m

RT: 12.304 min Scan# 1736

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 75 Resp: 1014


Ion Ratio Lower Upper

75 100

53 0.0 68.6 102.8#


89 31.5 39.0 58.4#





#82
4-Chlorotoluene
Concen: 46.556 ug/l
RT: 12.780 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion: 91 Resp: 104792
Ion Ratio Lower Upper
91 100
126 38.8 18.4 55.4




Manual Integrations
APPROVED


Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



#83
tert-Butylbenzene
Concen: 33.655 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20

Tgt Ion:119 Resp: 80984
Ion Ratio Lower Upper
119 100
91 58.9 29.9 89.8
134 27.3 12.9 38.7





#84

1,2,4-Trimethylbenzene

Concen: 38.273 ug/l

RT: 13.042 min Scan# 1857

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion:105 Resp: 9954

Ion Ratio Lower Upper

105 100





120 46.5 23.8 71.2

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#85

sec-Butylbenzene

Concen: 26.418 ug/l

RT: 13.176 min Scan# 1879

Delta R.T. 0.006 min

Lab File: VY023581.D




Acq: 24 Oct 2025 18:20


Tgt Ion:105 Resp: 91190

Ion Ratio Lower Upper

105 100

134 22.2 10.7 32.1





#86

p-Isopropyltoluene

Concen: 23.974 ug/l

RT: 13.292 min Scan# 1898

Delta R.T. 0.006 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD





Tgt Ion:	119	Resp:	7045
Ion Ratio	Lower	Upper	
119	100		
134	28.4	13.6	40.7
91	22.3	10.9	32.9

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#87

1,3-Dichlorobenzene

Concen: 39.603 ug/l




RT: 13.286 min Scan# 1897


Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Tgt Ion:	146	Resp:	66124
Ion Ratio	Lower	Upper	
146	100		
111	40.4	18.7	56.1
148	61.4	31.9	95.9





#88

1,4-Dichlorobenzene

Concen: 41.511 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. 0.000 min

Lab File: VY023581.D


Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD



Tgt Ion:146 Resp: 6845

Ion Ratio Lower Upper

146 100

111 37.6 18.0 54.0




148 63.6 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#89

n-Butylbenzene

Concen: 18.354 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion: 91 Resp: 47002


Ion Ratio Lower Upper

91 100

92 53.2 27.0 80.8

134 29.9 14.1 42.2





#90

Hexachloroethane

Concen: 12.077 ug/l

RT: 13.877 min Scan# 1994

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD

Tgt Ion:117 Resp: 7634

Ion Ratio Lower Upper

117 100





201 85.9 43.8 131.4

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#91

1,2-Dichlorobenzene

Concen: 45.784 ug/l

RT: 13.657 min Scan# 1958

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20




Tgt Ion:146 Resp: 67020


Ion Ratio Lower Upper

146 100

111 43.4 19.0 57.0

148 64.1 32.1 96.3





#92

1,2-Dibromo-3-Chloropropane

Concen: 157.821 ug/l

RT: 14.273 min Scan# 2059

Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20

Instrument :

MSVOA_Y

ClientSampleId :

PR-132-S16-015094-20251021MSD





Tgt Ion:	75	Resp:	1366
Ion	Ratio	Lower	Upper
75	100		
155	109.8	50.0	150.0
157	133.4	63.8	191.5

Manual Integrations

APPROVED

Reviewed By :Amit Patel 10/27/2025

Supervised By :Mahesh Dadoda 10/27/2025



#93

1,2,4-Trichlorobenzene

Concen: 16.597 ug/l




RT: 14.919 min Scan# 2165


Delta R.T. 0.000 min

Lab File: VY023581.D

Acq: 24 Oct 2025 18:20


Tgt Ion:	180	Resp:	15146
Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.9	143.6
145	35.7	14.7	44.1





#94
Hexachlorobutadiene
Concen: 9.860 ug/l
RT: 15.023 min Scan# 2181
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD




Tgt Ion:225 Resp: 5609
Ion Ratio Lower Upper
225 100
223 59.1 30.8 92.3
227 58.4 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025




#95
Naphthalene
Concen: 46.903 ug/l
RT: 15.145 min Scan# 2202
Delta R.T. 0.006 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20




Tgt Ion:128 Resp: 68160
Ion Ratio Lower Upper
128 100
127 14.6 10.4 15.6
129 11.9 8.9 13.3





#96
1,2,3-Trichlorobenzene
Concen: 18.717 ug/l
RT: 15.328 min Scan# 21
Delta R.T. 0.000 min
Lab File: VY023581.D
Acq: 24 Oct 2025 18:20



Instrument :
MSVOA_Y
ClientSampleId :
PR-132-S16-015094-20251021MSD



Tgt Ion	Ratio	Lower	Upper
180	100		
182	97.2	47.9	143.8
145	36.5	15.3	45.9

Manual Integrations
APPROVED

Reviewed By :Amit Patel 10/27/2025
Supervised By :Mahesh Dadoda 10/27/2025



Manual Integration Report

Sequence:	VY100725	Instrument	MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC005	VY023384.D	1,2,3-Trichloropropane	sam	10/10/2025 2:59:48 AM	MMdadoda	10/10/2025 4:37:28 AM	Peak Integrated by Software
VSTDICC005	VY023384.D	2-Chloroethyl Vinyl ether	sam	10/10/2025 2:59:48 AM	MMdadoda	10/10/2025 4:37:28 AM	Peak Integrated by Software
VSTDICC005	VY023384.D	Methacrylonitrile	sam	10/10/2025 2:59:48 AM	MMdadoda	10/10/2025 4:37:28 AM	Peak Integrated by Software
VSTDICC010	VY023385.D	1,2,3-Trichloropropane	sam	10/10/2025 2:59:52 AM	MMdadoda	10/10/2025 4:37:33 AM	Peak Integrated by Software
VSTDICC010	VY023385.D	2-Chloroethyl Vinyl ether	sam	10/10/2025 2:59:52 AM	MMdadoda	10/10/2025 4:37:33 AM	Peak Integrated by Software
VSTDICC020	VY023386.D	1,2,3-Trichloropropane	sam	10/10/2025 2:59:56 AM	MMdadoda	10/10/2025 4:37:37 AM	Peak Integrated by Software
VSTDICC020	VY023386.D	2-Chloroethyl Vinyl ether	sam	10/10/2025 2:59:56 AM	MMdadoda	10/10/2025 4:37:37 AM	Peak Integrated by Software
VSTDICC020	VY023386.D	Methacrylonitrile	sam	10/10/2025 2:59:56 AM	MMdadoda	10/10/2025 4:37:37 AM	Peak Integrated by Software
VSTDICCC050	VY023387.D	1,2,3-Trichloropropane	sam	10/10/2025 3:00:01 AM	MMdadoda	10/10/2025 4:37:41 AM	Peak Integrated by Software
VSTDICC100	VY023388.D	1,2,3-Trichloropropane	sam	10/10/2025 3:00:05 AM	MMdadoda	10/10/2025 4:37:46 AM	Peak Integrated by Software
VSTDICC150	VY023389.D	1,2,3-Trichloropropane	sam	10/10/2025 3:00:09 AM	MMdadoda	10/10/2025 4:37:51 AM	Peak Integrated by Software
VSTDICV050	VY023391.D	1,2,3-Trichloropropane	sam	10/10/2025 3:00:14 AM	MMdadoda	10/10/2025 4:37:55 AM	Peak Integrated by Software



Manual Integration Report

Sequence:	VY100725	Instrument	MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

Manual Integration Report

Sequence:	VY102225	Instrument	MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY023543.D	1,2,3-Trichloropropane	Amit	10/24/2025 11:56:25 AM	MMDadoda	10/24/2025 2:54:05 PM	Peak Integrated by Software
VSTDCCC050	VY023543.D	Methacrylonitrile	Amit	10/24/2025 11:56:25 AM	MMDadoda	10/24/2025 2:54:05 PM	Peak Integrated by Software
VY1022SBS01	VY023545.D	1,2,3-Trichloropropane	Amit	10/24/2025 11:56:29 AM	MMDadoda	10/24/2025 2:54:10 PM	Peak Integrated by Software

Manual Integration Report

Sequence:	Vy102425	Instrument	MSVOA_y

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VY023559.D	1,2,3-Trichloropropane	Amit	10/27/2025 9:01:50 AM	MMDadoda	10/27/2025 2:51:01 PM	Peak Integrated by Software
VY1024SBS01	VY023561.D	1,2,3-Trichloropropane	Amit	10/27/2025 9:01:57 AM	MMDadoda	10/27/2025 2:51:05 PM	Peak Integrated by Software
VY1024SBS01	VY023561.D	Methacrylonitrile	Amit	10/27/2025 9:01:57 AM	MMDadoda	10/27/2025 2:51:05 PM	Peak Integrated by Software
Q3419-02MS	VY023580.D	1,2,3-Trichloropropane	Amit	10/27/2025 9:02:12 AM	MMDadoda	10/27/2025 2:51:12 PM	Peak Integrated by Software
Q3419-03MSD	VY023581.D	1,2,3-Trichloropropane	Amit	10/27/2025 9:02:22 AM	MMDadoda	10/27/2025 2:51:15 PM	Peak Integrated by Software
Q3419-03MSD	VY023581.D	trans-1,4-Dichloro-2-but ene	Amit	10/27/2025 9:02:22 AM	MMDadoda	10/27/2025 2:51:15 PM	Peak Integrated by Software
VSTDCCC050	VY023583.D	1,2,3-Trichloropropane	Amit	10/27/2025 9:02:32 AM	MMDadoda	10/27/2025 2:51:20 PM	Peak Integrated by Software

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY100725

Review By	Semsettin Yesilyurt	Review On	10/10/2025 3:00:22 AM		
Supervise By	Maresh Dadoda	Supervise On	10/10/2025 4:38:02 AM		
SubDirectory	VY100725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP135766			
Initial Calibration Stds		VP135767,VP135768,VP135769,VP135770,VP135771,VP135772			
CCC					
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP135773			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023383.D	07 Oct 2025 08:43	SY/MD	Ok
2	VSTDIC005	VY023384.D	07 Oct 2025 09:17	SY/MD	Ok,M
3	VSTDIC010	VY023385.D	07 Oct 2025 10:02	SY/MD	Ok,M
4	VSTDIC020	VY023386.D	07 Oct 2025 11:24	SY/MD	Ok,M
5	VSTDIC050	VY023387.D	07 Oct 2025 11:47	SY/MD	Ok,M
6	VSTDIC100	VY023388.D	07 Oct 2025 12:09	SY/MD	Ok,M
7	VSTDIC150	VY023389.D	07 Oct 2025 12:32	SY/MD	Ok,M
8	VIBLK	VY023390.D	07 Oct 2025 13:02	SY/MD	Ok
9	VSTDICV050	VY023391.D	07 Oct 2025 14:12	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY102225

Review By	John Carlone	Review On	10/24/2025 2:07:41 PM		
Supervise By	Mahesh Dadoda	Supervise On	10/24/2025 2:54:27 PM		
SubDirectory	VY102225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135879			
CCC		VP135880,VP135881			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023542.D	22 Oct 2025 09:00	SY/MD	Ok
2	VSTDCCC050	VY023543.D	22 Oct 2025 09:30	SY/MD	Ok,M
3	VY1022SBL01	VY023544.D	22 Oct 2025 10:00	SY/MD	Ok
4	VY1022SBS01	VY023545.D	22 Oct 2025 10:28	SY/MD	Ok,M
5	VY1022SBSD01	VY023546.D	22 Oct 2025 10:51	SY/MD	Ok,M
6	Q3395-04MSD	VY023547.D	22 Oct 2025 11:28	SY/MD	Not Ok
7	Q3416-01	VY023548.D	22 Oct 2025 12:03	SY/MD	Ok
8	Q3404-01RE	VY023549.D	22 Oct 2025 12:52	SY/MD	Confirms
9	Q3427-03	VY023550.D	22 Oct 2025 13:43	SY/MD	Ok
10	Q3419-01	VY023551.D	22 Oct 2025 14:07	SY/MD	Ok
11	Q3419-04	VY023552.D	22 Oct 2025 15:34	SY/MD	Ok
12	Q3419-05	VY023553.D	22 Oct 2025 15:58	SY/MD	Ok
13	Q3422-01	VY023554.D	22 Oct 2025 16:22	SY/MD	Ok
14	Q3424-01	VY023555.D	22 Oct 2025 16:46	SY/MD	Ok
15	Q3424-02	VY023556.D	22 Oct 2025 17:10	SY/MD	Ok
16	Q3424-03	VY023557.D	22 Oct 2025 17:35	SY/MD	Ok

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY102425

Review By	Amit Patel	Review On	10/27/2025 2:54:16 PM		
Supervise By	Maresh Dadoda	Supervise On	10/27/2025 2:54:33 PM		
SubDirectory	VY102425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135928			
CCC		VP135929,VP135930			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY023558.D	24 Oct 2025 07:47	SY/MD	Ok
2	VSTDCCC050	VY023559.D	24 Oct 2025 08:56	SY/MD	Ok,M
3	VY1024SBL01	VY023560.D	24 Oct 2025 09:24	SY/MD	Ok
4	VY1024SBS01	VY023561.D	24 Oct 2025 09:52	SY/MD	Ok,M
5	VY1024SBSD01	VY023562.D	24 Oct 2025 10:14	SY/MD	Ok,M
6	Q3431-01	VY023563.D	24 Oct 2025 10:58	SY/MD	Not Ok
7	Q3413-01	VY023564.D	24 Oct 2025 11:21	SY/MD	Ok
8	VIBLK	VY023565.D	24 Oct 2025 11:44	SY/MD	Ok
9	Q3431-01	VY023566.D	24 Oct 2025 12:30	SY/MD	Ok
10	Q3445-01	VY023567.D	24 Oct 2025 12:54	SY/MD	Ok
11	Q3439-03	VY023568.D	24 Oct 2025 13:17	SY/MD	Ok
12	Q3439-07	VY023569.D	24 Oct 2025 13:41	SY/MD	Ok
13	Q3440-03	VY023570.D	24 Oct 2025 14:04	SY/MD	Ok
14	Q3434-01	VY023571.D	24 Oct 2025 14:28	SY/MD	Ok
15	Q3449-02	VY023572.D	24 Oct 2025 14:51	SY/MD	Ok
16	Q3451-03	VY023573.D	24 Oct 2025 15:15	SY/MD	Ok
17	Q3451-07	VY023574.D	24 Oct 2025 15:38	SY/MD	Ok
18	Q3452-03	VY023575.D	24 Oct 2025 16:01	SY/MD	Ok
19	Q3460-03	VY023576.D	24 Oct 2025 16:25	SY/MD	Ok
20	Q3460-07	VY023577.D	24 Oct 2025 16:48	SY/MD	Ok
21	Q3460-11	VY023578.D	24 Oct 2025 17:12	SY/MD	Ok

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY102425

Review By	Amit Patel	Review On	10/27/2025 2:54:16 PM		
Supervise By	Mahesh Dadoda	Supervise On	10/27/2025 2:54:33 PM		
SubDirectory	VY102425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135928			
CCC		VP135929,VP135930			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q3460-15	VY023579.D	24 Oct 2025 17:35	SY/MD	Ok
23	Q3419-02MS	VY023580.D	24 Oct 2025 17:58	SY/MD	Ok,M
24	Q3419-03MSD	VY023581.D	24 Oct 2025 18:20	SY/MD	Ok,M
25	VIBLK	VY023582.D	24 Oct 2025 18:44	SY/MD	Ok
26	VSTDCCC050	VY023583.D	24 Oct 2025 19:06	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100725

Review By	Semsettin Yesilyurt	Review On	10/10/2025 3:00:22 AM
Supervise By	Maresh Dadoda	Supervise On	10/10/2025 4:38:02 AM
SubDirectory	VY100725	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135766
Initial Calibration Stds	VP135767,VP135768,VP135769,VP135770,VP135771,VP135772
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135773
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023383.D	07 Oct 2025 08:43		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023384.D	07 Oct 2025 09:17		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023385.D	07 Oct 2025 10:02		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023386.D	07 Oct 2025 11:24		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023387.D	07 Oct 2025 11:47		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023388.D	07 Oct 2025 12:09		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023389.D	07 Oct 2025 12:32		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023390.D	07 Oct 2025 13:02		SY/MD	Ok
9	VSTDICV050	ICVVY100725	VY023391.D	07 Oct 2025 14:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102225

Review By	John Carlone	Review On	10/24/2025 2:07:41 PM
Supervise By	Mahesh Dadoda	Supervise On	10/24/2025 2:54:27 PM
SubDirectory	VY102225	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135879
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135880,VP135881 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023542.D	22 Oct 2025 09:00		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023543.D	22 Oct 2025 09:30		SY/MD	Ok,M
3	VY1022SBL01	VY1022SBL01	VY023544.D	22 Oct 2025 10:00		SY/MD	Ok
4	VY1022SBS01	VY1022SBS01	VY023545.D	22 Oct 2025 10:28		SY/MD	Ok,M
5	VY1022SBSD01	VY1022SBSD01	VY023546.D	22 Oct 2025 10:51		SY/MD	Ok,M
6	Q3395-04MSD	PR132-S07-000102-20	VY023547.D	22 Oct 2025 11:28	vial-B Internal Standard Fail	SY/MD	Not Ok
7	Q3416-01	NB-08-10212025	VY023548.D	22 Oct 2025 12:03	vial-A	SY/MD	Ok
8	Q3404-01RE	MOO-25-0285-0288RE	VY023549.D	22 Oct 2025 12:52	vial-B Internal standard fail	SY/MD	Confirms
9	Q3427-03	TP-8-VOC	VY023550.D	22 Oct 2025 13:43	vial-A	SY/MD	Ok
10	Q3419-01	PR-132-S16-015094-20	VY023551.D	22 Oct 2025 14:07	vial-A	SY/MD	Ok
11	Q3419-04	PR132-S16-119133-20	VY023552.D	22 Oct 2025 15:34	vial-A	SY/MD	Ok
12	Q3419-05	PR132-S16-094119-20	VY023553.D	22 Oct 2025 15:58	vial-A	SY/MD	Ok
13	Q3422-01	PR132-S15-000008-20	VY023554.D	22 Oct 2025 16:22	vial-A	SY/MD	Ok
14	Q3424-01	PR132-DP3-20251021	VY023555.D	22 Oct 2025 16:46	vial-A	SY/MD	Ok
15	Q3424-02	PR132-S11-010080-20	VY023556.D	22 Oct 2025 17:10	vial-A	SY/MD	Ok
16	Q3424-03	PR132-S14-000094-20	VY023557.D	22 Oct 2025 17:35	vial-A	SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102425

Review By	Amit Patel	Review On	10/27/2025 2:54:16 PM
Supervise By	Mahesh Dadoda	Supervise On	10/27/2025 2:54:33 PM
SubDirectory	VY102425	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135928
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135929,VP135930 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023558.D	24 Oct 2025 07:47		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023559.D	24 Oct 2025 08:56		SY/MD	Ok,M
3	VY1024SBL01	VY1024SBL01	VY023560.D	24 Oct 2025 09:24		SY/MD	Ok
4	VY1024SBS01	VY1024SBS01	VY023561.D	24 Oct 2025 09:52		SY/MD	Ok,M
5	VY1024SBSD01	VY1024SBSD01	VY023562.D	24 Oct 2025 10:14		SY/MD	Ok,M
6	Q3431-01	EO-02-10222025	VY023563.D	24 Oct 2025 10:58	vial-A	SY/MD	Not Ok
7	Q3413-01	BP-VPB-184-GW-980-S	VY023564.D	24 Oct 2025 11:21	vial-B	SY/MD	Ok
8	VIBLK	VIBLK	VY023565.D	24 Oct 2025 11:44		SY/MD	Ok
9	Q3431-01	EO-02-10222025	VY023566.D	24 Oct 2025 12:30	vial-B	SY/MD	Ok
10	Q3445-01	EO-01-102325	VY023567.D	24 Oct 2025 12:54	vial-A	SY/MD	Ok
11	Q3439-03	TP-9-VOC	VY023568.D	24 Oct 2025 13:17	vial-A	SY/MD	Ok
12	Q3439-07	TP-10-VOC	VY023569.D	24 Oct 2025 13:41	vial-A	SY/MD	Ok
13	Q3440-03	JB-2-VOC	VY023570.D	24 Oct 2025 14:04	vial-A	SY/MD	Ok
14	Q3434-01	PR132-WC1-20251022	VY023571.D	24 Oct 2025 14:28	vial-A	SY/MD	Ok
15	Q3449-02	CF-620-VOC-41	VY023572.D	24 Oct 2025 14:51	vial-A	SY/MD	Ok
16	Q3451-03	TP-11-VOC	VY023573.D	24 Oct 2025 15:15	vial-A	SY/MD	Ok
17	Q3451-07	TP-12-VOC	VY023574.D	24 Oct 2025 15:38	vial-A	SY/MD	Ok
18	Q3452-03	JB-1-VOC	VY023575.D	24 Oct 2025 16:01	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102425

Review By	Amit Patel	Review On	10/27/2025 2:54:16 PM		
Supervise By	Mahesh Dadoda	Supervise On	10/27/2025 2:54:33 PM		
SubDirectory	VY102425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y100725s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP135928				
CCC	VP135929,VP135930				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3460-03	WC-1-VOC	VY023576.D	24 Oct 2025 16:25	vial-A	SY/MD	Ok
20	Q3460-07	WC-2-VOC	VY023577.D	24 Oct 2025 16:48	vial-A	SY/MD	Ok
21	Q3460-11	WC-3-VOC	VY023578.D	24 Oct 2025 17:12	vial-A	SY/MD	Ok
22	Q3460-15	WC-4-VOC	VY023579.D	24 Oct 2025 17:35	vial-A	SY/MD	Ok
23	Q3419-02MS	PR-132-S16-015094-20	VY023580.D	24 Oct 2025 17:58	vial-A	SY/MD	Ok,M
24	Q3419-03MSD	PR-132-S16-015094-20	VY023581.D	24 Oct 2025 18:20	vial-A	SY/MD	Ok,M
25	VIBLK	VIBLK	VY023582.D	24 Oct 2025 18:44		SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY023583.D	24 Oct 2025 19:06		SY/MD	Ok,M

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/23/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/23/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137614

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3419-01	PR-132-S16-015094-20251021	1	1.15	11.01	12.16	7.14	54.4	
Q3419-02	Q3419-01MS	2	1.15	11.01	12.16	7.14	54.4	
Q3419-03	Q3419-01MSD	3	1.15	11.01	12.16	7.14	54.4	
Q3419-04	PR-132-S16-119133-20251021	4	1.16	10.07	11.23	6.37	51.7	
Q3419-05	PR-132-S16-094119-20251021	5	1.19	11.07	12.26	6.38	46.9	
Q3422-01	PR132-S15-000008-20251017	6	1.18	11.02	12.2	7.58	58.1	
Q3424-01	PR132-DP3-20251021	7	1.15	10.36	11.51	5.91	45.9	
Q3424-02	PR132-S11-010080-20251021	8	1.12	10.72	11.84	6.49	50.1	
Q3424-03	PR132-S14-00094-20251021	9	1.14	11.34	12.48	6.82	50.1	
Q3427-01	TP-8	10	1.11	10.77	11.88	10.84	90.3	
Q3427-02	TP-8-EPH	11	1.18	10.97	12.15	10.98	89.3	
Q3427-03	TP-8-VOC	12	1.19	10.26	11.45	10.52	90.9	
Q3429-01	40309	13	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3429-02	40310	14	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3430-01	BUR-25-0095	15	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3431-01	EO-02-10222025	16	1.11	10.55	11.66	10.64	90.3	
Q3431-02	EO-02-10222025-E2	17	1.17	10.59	11.76	10.59	89.0	
Q3434-01	PR132-WC1-20251022	18	1.13	10.43	11.56	7.17	57.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

UB 137614

WorkList Name : %1-102225

WorkList ID : 192597

Department : Wet-Chemistry

Date : 10-22-2025 09:03:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3419-01	PR-132-S16-015094-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-02	Q3419-01MS	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-03	Q3419-01MSD	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-04	PR-132-S16-119133-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3419-05	PR-132-S16-094119-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3422-01	PR132-S15-000008-20251017	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/21/2025	Chemtech -SO
Q3424-01	PR132-DP3-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/21/2025	Chemtech -SO
Q3424-02	PR132-S11-010080-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/21/2025	Chemtech -SO
Q3424-03	PR132-S14-00094-20251021	Solid	Percent Solids	Cool 4 deg C	AECO02	D31	10/21/2025	Chemtech -SO
Q3427-01	TP-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3427-02	TP-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3427-03	TP-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3429-01	40309	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	10/22/2025	Chemtech -SO
Q3429-02	40310	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	10/22/2025	Chemtech -SO
Q3430-01	BUR-25-0095	Solid	Percent Solids	Cool 4 deg C	PSEG03	D51	10/22/2025	Chemtech -SO
Q3431-01	EO-02-10222025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	10/22/2025	Chemtech -SO
Q3431-02	EO-02-10222025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D41	10/22/2025	Chemtech -SO
Q3434-01	PR132-WC1-20251022	Solid	Percent Solids	Cool 4 deg C	AECO02	D41	10/22/2025	Chemtech -SO

Date/Time

10/22/25 15:00

Raw Sample Received by:

SWC

Raw Sample Relinquished by:

AS

Q3419-VOC-TCLVOA-10

Date/Time

10/22/25

Raw Sample Received by:

AS

Raw Sample Relinquished by:

SWC

Prep Standard - Chemical Standard Summary

Order ID : Q3419

Test : VOC-TCLVOA-10

Prepbatch ID :

Sequence ID/Qc Batch ID: VY102225,Vy102425,

Standard ID :

VP133934,VP134147,VP134149,VP134150,VP134934,VP134957,VP135552,VP135553,VP135834,VP135835,VP135879,VP135880,VP135881,VP135928,VP135929,VP135930,

Chemical ID :

V13391,V14200,V14204,V14205,V14290,V14509,V14510,V14531,V14532,V14625,V14636,V14637,V14638,V14639,V14673,V14727,V14728,V14803,V14815,V14844,V14906,V14921,V14928,V14929,V14951,V14952,V15073,V15074,V15075,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1917	8260 Internal standard 50 ppm	VP133934	05/16/2025	11/12/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.10000ml of V14290 + 49.90000ml of V14921 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
252	8260 Working STD (BCM)-First source, 100PPM	VP134147	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/10/2025

FROM 1.00000ml of V14673 + 19.00000ml of V14929 = Final Quantity: 20.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP134149	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/10/2025

FROM 1.00000ml of V14636 + 1.00000ml of V14637 + 1.00000ml of V14638 + 1.00000ml of V14639 + 46.00000ml of V14929 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1811	8260 Working Std(2-CVE)-500ppm	VP134150	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/10/2025

FROM 7.50000ml of V14929 + 12.50000ml of VP134149 = Final Quantity: 20.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
249	8260 Surrogate, 100PPM	VP134934	07/29/2025	01/29/2026	Semsettin Yesilyurt	None	None	Maresh Dadoda
08/06/2025								

FROM 0.10000ml of V14906 + 24.90000ml of V14625 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP134957	08/01/2025	11/22/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
08/06/2025								

FROM 0.50000ml of V13391 + 49.50000ml of V14625 = Final Quantity: 50.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP135552	09/22/2025	11/03/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda 09/26/2025

FROM 0.40000ml of V14844 + 1.00000ml of V14200 + 1.00000ml of V14509 + 1.00000ml of V14510 + 1.00000ml of V14531 + 1.00000ml of V14532 + 1.00000ml of V14727 + 1.00000ml of V14728 + 1.00000ml of V14803 + 1.00000ml of V14815 + 1.00000ml of V14951 + 1.00000ml of V14952 + 1.50000ml of V14204 + 1.50000ml of V14205 + 10.60000ml of V14928 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
244	8260 Calibration Working STD Mix-First source, 100PPM	VP135553	09/22/2025	11/03/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda 09/26/2025

FROM 5.62500ml of V14928 + 9.37500ml of VP135552 = Final Quantity: 15.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP135834	10/17/2025	11/16/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								10/17/2025

FROM 1.00000ml of V15073 + 1.50000ml of V15074 + 1.50000ml of V15075 + 21.00000ml of V14928 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
56	8260 Working STD (Acrolein) -first source, 500PPM	VP135835	10/17/2025	11/16/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								10/17/2025

FROM 7.50000ml of V14928 + 12.50000ml of VP135834 = Final Quantity: 20.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
732	BFB TUNE CHECK - SOIL	VP135879	10/22/2025	10/23/2025	Amit Patel	None	None	Maresh Dadoda
								10/23/2025

FROM 4.99800ml of W3112 + 0.00200ml of VP134957 = Final Quantity: 5.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
773	50 PPB CCC, 8260-SOIL	VP135880	10/22/2025	10/23/2025	Amit Patel	None	None	Maresh Dadoda
								10/23/2025

FROM 4.98000ml of W3112 + 0.00250ml of VP134147 + 0.00250ml of VP134150 + 0.00250ml of VP134934 + 0.00250ml of VP135553
+ 0.00250ml of VP135835 + 0.00500ml of VP133934 = Final Quantity: 5.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
773	50 PPB CCC, 8260-SOIL	VP135881	10/22/2025	10/23/2025	Amit Patel	None	None	Maresh Dadoda
								10/23/2025

FROM 4.98000ml of W3112 + 0.00250ml of VP134147 + 0.00250ml of VP134150 + 0.00250ml of VP134934 + 0.00250ml of VP135553 + 0.00250ml of VP135835 + 0.00500ml of VP133934 = Final Quantity: 5.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
732	BFB TUNE CHECK - SOIL	VP135928	10/24/2025	10/25/2025	Amit Patel	None	None	Maresh Dadoda
								10/29/2025

FROM 4.99800ml of W3112 + 0.00200ml of VP134957 = Final Quantity: 5.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
773	50 PPB CCC, 8260-SOIL	VP135929	10/24/2025	10/25/2025	Amit Patel	None	None	Maresh Dadoda
FROM 4.98000ml of W3112 + 0.00250ml of VP134147 + 0.00250ml of VP134150 + 0.00250ml of VP134934 + 0.00250ml of VP135553 + 0.00250ml of VP135835 + 0.00500ml of VP133934 = Final Quantity: 5.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
773	50 PPB CCC, 8260-SOIL	VP135930	10/24/2025	10/25/2025	Amit Patel	None	None	Maresh Dadoda
FROM 4.98000ml of W3112 + 0.00250ml of VP134147 + 0.00250ml of VP134150 + 0.00250ml of VP134934 + 0.00250ml of VP135553 + 0.00250ml of VP135835 + 0.00500ml of VP133934 = Final Quantity: 5.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	03/22/2026	09/22/2025 / sam	02/28/2024 / SAM	V14200

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	03/22/2026	09/22/2025 / sam	02/28/2024 / SAM	V14204

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	03/22/2026	09/22/2025 / sam	02/28/2024 / SAM	V14205

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	03/22/2026	09/22/2025 / sam	09/17/2024 / SAM	V14509

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	03/22/2026	09/22/2025 / sam	09/17/2024 / SAM	V14510

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	03/22/2026	09/22/2025 / sam	09/18/2024 / SAM	V14531

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	03/22/2026	09/22/2025 / sam	09/18/2024 / SAM	V14532

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	2310762004	01/29/2026	07/29/2025 / SAM	11/26/2024 / SAM	V14625

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14636

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14637

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14638

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14639

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	03/22/2026	09/22/2025 / sam	12/17/2024 / SAM	V14727

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	03/22/2026	09/22/2025 / sam	12/17/2024 / SAM	V14728

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	03/22/2026	09/22/2025 / sam	01/08/2025 / SAM	V14803

LOTS

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	03/22/2026	09/22/2025 / sam	01/08/2025 / SAM	V14815

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	03/22/2026	09/22/2025 / sam	01/21/2025 / SAM	V14844

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0223904	07/29/2026	07/29/2025 / SAM	03/24/2025 / SAM	V14906

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	11/12/2025	05/12/2025 / SAM	05/09/2025 / SAM	V14921

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	03/22/2026	09/22/2025 / sam	05/09/2025 / SAM	V14928

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	12/06/2025	06/06/2025 / SAM	05/09/2025 / SAM	V14929

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0222076	03/22/2026	09/22/2025 / sam	05/19/2025 / SAM	V14951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0222076	03/22/2026	09/22/2025 / sam	05/19/2025 / SAM	V14952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	101625	11/16/2025	10/17/2025 / sam	10/17/2025 / sam	V15073

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	101625	11/16/2025	10/17/2025 / sam	10/17/2025 / sam	V15074

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	101625	11/16/2025	10/17/2025 / sam	10/17/2025 / sam	V15075

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112


Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC



Ken Koehnlein
Sr. Manager, Quality Assurance



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12

Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KMe)	Lot	Dir.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information			
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(+/-) (µg/mL)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102366	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	N/A	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A	
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A	
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg	
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 75mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg	
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg	
11. Methyl acrylate	(0755)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 787mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	H4002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg	
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg	
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	158-59-2	N/A	or-rat 848mg/kg	
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	158-60-5	N/A	or-rat 1235mg/kg	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg	
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.6	20.4	58-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	1999.6	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.6	20.5	74-85-3	100 ppm	or-rat 108mg/kg	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.6	20.5	75-34-3	100 ppm	or-rat 725mg/kg	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.6	20.4	594-20-7	N/A	N/A	
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	1999.6	20.5	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.6	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg	
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40015.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg	
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	563-58-6	N/A	N/A	
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10081-01-5	N/A	N/A	
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10081-02-6	N/A	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40008.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg	
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-88-1	1 ppm	or-rat 4894mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	104-51-8	N/A	or-rat 2699mg/kg	
47. n-Butyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg	
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	87-61-6	N/A	or-rat 1390mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	95-63-6	N/A	or-rat 5g/kg	
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-6	N/A	or-rat 5000mg/kg	
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-06-6	N/A	N/A	
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg	
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg	
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg	
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg	
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 900mg/kg	
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1063mg/kg	
64. 1,3-Dichlorobenzene</																	




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Hexamethylcyclotriphosphorotriazine/Hexyl acetate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentopropane	27.07
22	cis-1,2-Dichloroethane	27.07
23	Toluene	28.03
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrahydrofuran/1,2-Dichloropropane	30.94
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	32.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrafluoroethane	33.40
31	m-Xylene/p-Xylene	33.85
32	o-Xylene	35.53
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	35.78
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,3,5-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12

Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KMe)	Lot	Dil.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information			
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(+/-) (µg/mL)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	NA	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	NA	N/A	
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	NA	N/A	
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	NA	N/A	
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 14800mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg	
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg	
11. Methyl acrylate	(0755)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKBW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	H4002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg	
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	NA	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	N/A	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	NA	N/A	
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	NA	N/A	
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	NA	N/A	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 1235mg/kg	
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 106mg/kg	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	NA	N/A	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg	
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.5	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg	
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40015.0	2000	NA	NA	0.017	NA	NA	2000.2	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	NA	or-mus 3600mg/kg	
36. 1,1-Dichloropropane	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	563-59-6	NA	N/A	
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	NA	N/A	
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	NA	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	NA	or-rat 670mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg	
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-98-1	NA	or-rat 2699mg/kg	
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	NA	N/A	
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	NA	or-rat 4750mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-98-3	200 ppm	or-rat 5000mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	NA	or-mus 1390mg/kg	
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	NA	or-rat 5g/kg	
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	NA	or-rat 5000mg/kg	
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	109-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	NA	N/A	
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-6	NA	N/A	
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2240mg/kg	
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 2290mg/kg	
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 3900mg/kg	
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 2100mg/kg	
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA										




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetone	12.00
5	Isodimethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Hexamethylcyclotriphosphazene/Hexyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentopropane	27.07
22	cis-1,2-Dichloroethane	27.07
23	Toluene	28.03
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethane/1,2-Dichloropropane	30.94
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	32.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.85
32	o-Xylene	35.53
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	35.78
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,3,5-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water

			Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s): **Water**
Lot#: **072324Q**

Formulated By:	Justin Dippold
DATE: 091424	
Reviewed By:	Pedro L. Rentas
DATE: 091424	

Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2 = 200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

27

60000

50000

40000

30000

20000

10000

37

44

65

75

85

119

158 169

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vlog



ANAB ISO 17034 Accredited
AR-1539 Certificate Number
<https://Absolutestandards.com>

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 091424
Description: Acrolein

Solvent(s): Water
Lot# 072324Q


Expiration Date:	101424
Recommended Storage:	Refrigerate (4 °C)
Concentration (µg/mL):	5000
NIST Test ID#:	6UTB

Weight(s) shown below were combined and diluted to (mL):	10.0	0.001	Flask Uncertainty

		091424
Formulated By:	Justin Dippold	DATE
		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05186	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.). Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1 > 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97
	CAS#: 67-56-1	

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of 11-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : October 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pabon at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Padden at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial
Dec 03/31/25
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
gravimetric

V14904 to V14913



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0223904
Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-33313	99%	25,108.0 µg/mL	+/- 1,421.9957
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	0000268853	99%	25,108.0 µg/mL	+/- 1,421.9957
3	Dibromofluoromethane	1868-53-7	VENKAT02	99%	25,232.0 µg/mL	+/- 1,429.0184
4	Toluene-d8	2037-26-5	PR-34141	99%	25,156.0 µg/mL	+/- 1,424.7141

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Brittany Federlino - Operations Tech I

Date Mixed: 27-Mar-2025

Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2ee 05/19/25
CERTIFIED REFERENCE MATERIAL
20 vial

Certificate of Analysis

chromatographic plus

✓ 14951 to ✓ 14970



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0222076
Description : 8260B Acetates Mix
8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBR1889	99%	2,004.0 µg/mL	+/- 69.2674
2	Vinyl acetate	108-05-4	RD240423RSR	99%	2,010.0 µg/mL	+/- 69.4748
3	Ethyl acetate	141-78-6	SHBS3323	99%	2,019.3 µg/mL	+/- 69.7974
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,008.0 µg/mL	+/- 69.4057
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBR2024	99%	2,008.0 µg/mL	+/- 69.4057
7	Amyl acetate	628-63-7	BCBT7442	99%	2,006.7 µg/mL	+/- 69.3596

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2ee 05/19/25
CERTIFIED REFERENCE MATERIAL
20 vial

Certificate of Analysis

chromatographic plus

✓ 14951 to ✓ 14970



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0222076
Description : 8260B Acetates Mix
8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBR1889	99%	2,004.0 µg/mL	+/- 69.2674
2	Vinyl acetate	108-05-4	RD240423RSR	99%	2,010.0 µg/mL	+/- 69.4748
3	Ethyl acetate	141-78-6	SHBS3323	99%	2,019.3 µg/mL	+/- 69.7974
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,008.0 µg/mL	+/- 69.4057
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBR2024	99%	2,008.0 µg/mL	+/- 69.4057
7	Amyl acetate	628-63-7	BCBT7442	99%	2,006.7 µg/mL	+/- 69.3596

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this



Rec 10/16/25 5:14p



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 101625
Description: AcroleinSolvent(s):
Water Lot# 041725Q

		101625
Formulated By:	Prashant Chauhan	DATE
		101625
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 111625

Recommended Storage: Refrigerate (2°C to 8°C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB


5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

V15071-V15075

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05186	0.05180	5013.7	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Rec 10/16/25 5:14p



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 101625
Description: AcroleinSolvent(s):
Water Lot# 041725Q

		101625
Formulated By:	Prashant Chauhan	DATE
		101625
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 111625

Recommended Storage: Refrigerate (2°C to 8°C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB



5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

V15071-V15075

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05180	5013.7	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Rec 10/16/25 5:14p



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 101625
Description: Acrolein

Solvent(s):
Water Lot# 041725Q

		101625
Formulated By:	Prashant Chauhan	DATE
		101625
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 111625

Recommended Storage: Refrigerate (2°C to 8°C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB



5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

V15071-V15075

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05180	5013.7	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025

Preservation Log

BalanceID: VOA-SC-2

Review By: Amit

Supervise By: MMDadoda

Seq	LabID	Vial A Weight(g)= Total Wt-Tare Wt	Vial A Time	Vial B Weight(g)= Total Wt-Tare Wt	Vial B Time	Vial C Weight(g) Preserve with MeOH	Vial C Time	Methanol ID	Preservation Date	Comments
1	Q3419-01	38.70-32.94=5.76	14:01	38.38-32.60=5.78	14:02	36.05-27.88=8.17	14:03	V14921	10/22/2025	TERRACORE SAMPLE 8260
2	Q3419-02	37.53-33.02=4.51	14:03	38.61-32.82=5.79	14:04	34.08-28.93=5.15	14:05	V14921	10/22/2025	
3	Q3419-03	38.59-32.77=5.82	14:06	36.24-32.90=3.34	14:07	35.15-27.98=7.17	14:08	V14921	10/22/2025	
4	Q3419-04	38.49-33.24=5.25	14:09	37.59-32.48=5.11	14:10	34.05-28.32=5.73	14:11	V14921	10/22/2025	
5	Q3419-05	37.82-32.84=4.98	14:12	38.15-32.87=5.28	14:14	33.05-28.00=5.05	14:15	V14921	10/22/2025	

Instructions : For medium Level analysis, 5ml MeOH added for CLP(SOM/SFAM) method and 10ml for regular 8260.

If the samples are not to be analyzed within 48hrs of sampling, preserve samples immediately, Vials A and B are stored in the VOA-FRZ-2.

Vial C - MeOH preserved vials are kept in VOA-REF #6.

QA Contorl # A3041177



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: AECOM
ADDRESS: 605 3rd Ave
CITY: New York STATE: NY ZIP: 10158
ATTENTION: Rob Forstner
PHONE: 212-377-8721 FAX:

PROJECT NAME: Port Richmond
PROJECT NO.: 60693795 LOCATION: Staten Isl
PROJECT MANAGER: Rob Forstner
e-mail: Robert.Forstner@aecom.com
PHONE: 11 FAX:

BILL TO: AECOM PO#: 11
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: Rob Forstner PHONE: 11

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☒ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. SVOCs 2. Pesticides TCL 3. Hg Metals 4. PCBs Hexachlorine 5. Dioxin 6. Grain size 7. TOCs 8. VOCs 9.

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	PR-132-S16-015094-20251021	S		X	10/21	9:31	11	X	X	X	X	X	X	X	X		
2.	PR-132-S16-015094-20251021-HS	S		X	10/21	9:31	10	X	X	X	X	X		X	X		
3.	PR132-S16-015094-20251021-MSD	S		X	10/21	9:31	10	X	X	X	X	X		X	X		
4.	PR132-S16-119133-20281021	S		X	10/21	9:33	11	X	X	X	X	X	X	X	X		
5.	PR132-S16-094119-20281021	S		X	10/21	9:33	11	X	X	X	X	X	X	X	X		
6.	Temp Blank	A			10/21		1	X	X	X	X	X			X		
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>Jim Alderson</u>	DATE/TIME: <u>10/21/25 430</u>	RECEIVED BY: <u>[Signature]</u> <u>1635</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY:	Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>10-21-25</u>	RECEIVED BY: <u>[Signature]</u>	

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3419	AECO02	Order Date : 10/22/2025 11:01:00 AM	Project Mgr :
Client Name : AECOM		Project Name : AE1-CTY 3.2.Z-PR-DOCK	Report Type : NYS ASP <i>B</i>
Client Contact : Rob Forstner		Receive DateTime : 10/21/2025 5:40:00 PM	EDD Type : EQUIS
Invoice Name : AECOM		Purchase Order :	Hard Copy Date :
Invoice Contact : Rob Forstner			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3419-01	PR-132-S16-015094-20251021	Solid	10/21/2025	09:31					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3419-02	Q3419-01MS	Solid	10/21/2025	09:31					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3419-03	Q3419-01MSD	Solid	10/21/2025	09:31					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3419-04	<i>PR132</i> ~~PR-132~~ -S16-119133-20251021	Solid	10/21/2025	09:35					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3419-05	<i>PR132</i> ~~PR-132~~ -S16-094119-20251021	Solid	10/21/2025	09:33					
					VOC-TCLVOA-10		8260D	10 Bus. Days	

Relinquished By : *[Signature]*

Date / Time : *10/22/25 1245*

Received By : *JC*

Date / Time : *10/22/25*

1245

Storage Area : VOA Refridgerator Room