

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3419

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137657

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	10/27/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	10/27/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/27/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	10/27/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/27/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3419

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137664

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	1080	1000	108	90-110	10/23/2025
Sample ID: CCV1 TOC	mg/L	998	1000	100	90-110	10/28/2025
Sample ID: CCV2 TOC	mg/L	1070	1000	107	90-110	10/28/2025
Sample ID: CCV3 TOC	mg/L	1000	1000	100	90-110	10/28/2025