

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV85	Mercury	4.26	4.0	106	90 - 110	CV	10/23/2025	13:51	LB137626

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV79	Mercury	4.94	5.0	99	90 - 110	CV	10/23/2025	14:01	LB137626
CCV80	Mercury	4.97	5.0	99	90 - 110	CV	10/23/2025	14:35	LB137626
CCV81	Mercury	4.86	5.0	97	90 - 110	CV	10/23/2025	15:09	LB137626
CCV82	Mercury	5.04	5.0	101	90 - 110	CV	10/23/2025	15:29	LB137626
CCV83	Mercury	4.68	5.0	94	90 - 110	CV	10/23/2025	15:59	LB137626
CCV84	Mercury	4.71	5.0	94	90 - 110	CV	10/23/2025	16:14	LB137626

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	1500	1600	94	90 - 110	P	10/23/2025	15:44	LB137649
	Antimony	439	400	110	90 - 110	P	10/23/2025	15:44	LB137649
	Arsenic	416	400	104	90 - 110	P	10/23/2025	15:44	LB137649
	Barium	1610	1600	100	90 - 110	P	10/23/2025	15:44	LB137649
	Beryllium	43.2	40.0	108	90 - 110	P	10/23/2025	15:44	LB137649
	Cadmium	217	200	109	90 - 110	P	10/23/2025	15:44	LB137649
	Calcium	4160	4000	104	90 - 110	P	10/23/2025	15:44	LB137649
	Chromium	160	160	100	90 - 110	P	10/23/2025	15:44	LB137649
	Cobalt	403	400	101	90 - 110	P	10/23/2025	15:44	LB137649
	Copper	187	200	94	90 - 110	P	10/23/2025	15:44	LB137649
	Iron	878	800	110	90 - 110	P	10/23/2025	15:44	LB137649
	Lead	387	400	97	90 - 110	P	10/23/2025	15:44	LB137649
	Magnesium	3670	4000	92	90 - 110	P	10/23/2025	15:44	LB137649
	Manganese	394	400	98	90 - 110	P	10/23/2025	15:44	LB137649
	Nickel	404	400	101	90 - 110	P	10/23/2025	15:44	LB137649
	Potassium	4210	4000	105	90 - 110	P	10/23/2025	15:44	LB137649
	Selenium	429	400	107	90 - 110	P	10/23/2025	15:44	LB137649
	Silver	213	200	106	90 - 110	P	10/23/2025	15:44	LB137649
	Sodium	3910	4000	98	90 - 110	P	10/23/2025	15:44	LB137649
	Thallium	391	400	98	90 - 110	P	10/23/2025	15:44	LB137649
	Vanadium	403	400	101	90 - 110	P	10/23/2025	15:44	LB137649
	Zinc	379	400	95	90 - 110	P	10/23/2025	15:44	LB137649

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	20.5	20.0	103	80 - 120	P	10/23/2025	15:53	LB137649
	Antimony	2.03	2.0	102	80 - 120	P	10/23/2025	15:53	LB137649
	Arsenic	1.16	1.0	116	80 - 120	P	10/23/2025	15:53	LB137649
	Barium	9.39	10.0	94	80 - 120	P	10/23/2025	15:53	LB137649
	Beryllium	1.07	1.0	107	80 - 120	P	10/23/2025	15:53	LB137649
	Cadmium	0.86	1.0	86	80 - 120	P	10/23/2025	15:53	LB137649
	Calcium	512	500	102	80 - 120	P	10/23/2025	15:53	LB137649
	Chromium	1.94	2.0	97	80 - 120	P	10/23/2025	15:53	LB137649
	Cobalt	1.05	1.0	105	80 - 120	P	10/23/2025	15:53	LB137649
	Copper	1.94	2.0	97	80 - 120	P	10/23/2025	15:53	LB137649
	Iron	52.8	50.0	106	80 - 120	P	10/23/2025	15:53	LB137649
	Lead	0.83	1.0	83	80 - 120	P	10/23/2025	15:53	LB137649
	Magnesium	479	500	96	80 - 120	P	10/23/2025	15:53	LB137649
	Manganese	0.94	1.0	94	80 - 120	P	10/23/2025	15:53	LB137649
	Nickel	1.00	1.0	100	80 - 120	P	10/23/2025	15:53	LB137649
	Potassium	506	500	101	80 - 120	P	10/23/2025	15:53	LB137649
	Selenium	4.77	5.0	95	80 - 120	P	10/23/2025	15:53	LB137649
	Silver	1.11	1.0	111	80 - 120	P	10/23/2025	15:53	LB137649
	Sodium	484	500	97	80 - 120	P	10/23/2025	15:53	LB137649
	Thallium	0.97	1.0	97	80 - 120	P	10/23/2025	15:53	LB137649
	Vanadium	4.89	5.0	98	80 - 120	P	10/23/2025	15:53	LB137649
	Zinc	5.02	5.0	100	80 - 120	P	10/23/2025	15:53	LB137649

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	48700	50000	98	90 - 110	P	10/23/2025	16:10	LB137649
	Antimony	524	500	105	90 - 110	P	10/23/2025	16:10	LB137649
	Arsenic	515	500	103	90 - 110	P	10/23/2025	16:10	LB137649
	Barium	2570	2500	103	90 - 110	P	10/23/2025	16:10	LB137649
	Beryllium	524	500	105	90 - 110	P	10/23/2025	16:10	LB137649
	Cadmium	503	500	101	90 - 110	P	10/23/2025	16:10	LB137649
	Calcium	251000	250000	100	90 - 110	P	10/23/2025	16:10	LB137649
	Chromium	508	500	102	90 - 110	P	10/23/2025	16:10	LB137649
	Cobalt	503	500	100	90 - 110	P	10/23/2025	16:10	LB137649
	Copper	4980	5000	100	90 - 110	P	10/23/2025	16:10	LB137649
	Iron	128000	125000	103	90 - 110	P	10/23/2025	16:10	LB137649
	Lead	2590	2500	104	90 - 110	P	10/23/2025	16:10	LB137649
	Magnesium	250000	250000	100	90 - 110	P	10/23/2025	16:10	LB137649
	Manganese	5020	5000	100	90 - 110	P	10/23/2025	16:10	LB137649
	Nickel	498	500	100	90 - 110	P	10/23/2025	16:10	LB137649
	Potassium	126000	125000	101	90 - 110	P	10/23/2025	16:10	LB137649
	Selenium	498	500	100	90 - 110	P	10/23/2025	16:10	LB137649
	Silver	510	500	102	90 - 110	P	10/23/2025	16:10	LB137649
	Sodium	247000	250000	99	90 - 110	P	10/23/2025	16:10	LB137649
	Thallium	514	500	103	90 - 110	P	10/23/2025	16:10	LB137649
	Vanadium	522	500	104	90 - 110	P	10/23/2025	16:10	LB137649
	Zinc	4790	5000	96	90 - 110	P	10/23/2025	16:10	LB137649
CCV02	Aluminum	51000	50000	102	90 - 110	P	10/23/2025	16:54	LB137649
	Antimony	536	500	107	90 - 110	P	10/23/2025	16:54	LB137649
	Arsenic	530	500	106	90 - 110	P	10/23/2025	16:54	LB137649
	Barium	2670	2500	107	90 - 110	P	10/23/2025	16:54	LB137649
	Beryllium	530	500	106	90 - 110	P	10/23/2025	16:54	LB137649
	Cadmium	521	500	104	90 - 110	P	10/23/2025	16:54	LB137649
	Calcium	259000	250000	104	90 - 110	P	10/23/2025	16:54	LB137649
	Chromium	538	500	108	90 - 110	P	10/23/2025	16:54	LB137649
	Cobalt	525	500	105	90 - 110	P	10/23/2025	16:54	LB137649
	Copper	5180	5000	104	90 - 110	P	10/23/2025	16:54	LB137649
	Iron	131000	125000	105	90 - 110	P	10/23/2025	16:54	LB137649
	Lead	2690	2500	108	90 - 110	P	10/23/2025	16:54	LB137649

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	260000	250000	104	90 - 110	P	10/23/2025	16:54	LB137649
	Manganese	5300	5000	106	90 - 110	P	10/23/2025	16:54	LB137649
	Nickel	520	500	104	90 - 110	P	10/23/2025	16:54	LB137649
	Potassium	130000	125000	104	90 - 110	P	10/23/2025	16:54	LB137649
	Selenium	511	500	102	90 - 110	P	10/23/2025	16:54	LB137649
	Silver	526	500	105	90 - 110	P	10/23/2025	16:54	LB137649
	Sodium	257000	250000	103	90 - 110	P	10/23/2025	16:54	LB137649
	Thallium	539	500	108	90 - 110	P	10/23/2025	16:54	LB137649
	Vanadium	540	500	108	90 - 110	P	10/23/2025	16:54	LB137649
	Zinc	4990	5000	100	90 - 110	P	10/23/2025	16:54	LB137649
CCV03	Aluminum	50800	50000	102	90 - 110	P	10/23/2025	17:35	LB137649
	Antimony	532	500	106	90 - 110	P	10/23/2025	17:35	LB137649
	Arsenic	531	500	106	90 - 110	P	10/23/2025	17:35	LB137649
	Barium	2650	2500	106	90 - 110	P	10/23/2025	17:35	LB137649
	Beryllium	528	500	106	90 - 110	P	10/23/2025	17:35	LB137649
	Cadmium	514	500	103	90 - 110	P	10/23/2025	17:35	LB137649
	Calcium	259000	250000	104	90 - 110	P	10/23/2025	17:35	LB137649
	Chromium	524	500	105	90 - 110	P	10/23/2025	17:35	LB137649
	Cobalt	536	500	107	90 - 110	P	10/23/2025	17:35	LB137649
	Copper	5230	5000	104	90 - 110	P	10/23/2025	17:35	LB137649
	Iron	133000	125000	106	90 - 110	P	10/23/2025	17:35	LB137649
	Lead	2690	2500	108	90 - 110	P	10/23/2025	17:35	LB137649
	Magnesium	264000	250000	106	90 - 110	P	10/23/2025	17:35	LB137649
	Manganese	5310	5000	106	90 - 110	P	10/23/2025	17:35	LB137649
	Nickel	517	500	104	90 - 110	P	10/23/2025	17:35	LB137649
	Potassium	129000	125000	104	90 - 110	P	10/23/2025	17:35	LB137649
	Selenium	508	500	102	90 - 110	P	10/23/2025	17:35	LB137649
	Silver	528	500	106	90 - 110	P	10/23/2025	17:35	LB137649
	Sodium	260000	250000	104	90 - 110	P	10/23/2025	17:35	LB137649
	Thallium	544	500	109	90 - 110	P	10/23/2025	17:35	LB137649
	Vanadium	540	500	108	90 - 110	P	10/23/2025	17:35	LB137649
	Zinc	5050	5000	101	90 - 110	P	10/23/2025	17:35	LB137649
CCV04	Aluminum	50900	50000	102	90 - 110	P	10/23/2025	18:19	LB137649
	Antimony	542	500	108	90 - 110	P	10/23/2025	18:19	LB137649

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3419

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	528	500	106	90 - 110	P	10/23/2025	18:19	LB137649
	Barium	2660	2500	106	90 - 110	P	10/23/2025	18:19	LB137649
	Beryllium	529	500	106	90 - 110	P	10/23/2025	18:19	LB137649
	Cadmium	517	500	103	90 - 110	P	10/23/2025	18:19	LB137649
	Calcium	263000	250000	105	90 - 110	P	10/23/2025	18:19	LB137649
	Chromium	534	500	107	90 - 110	P	10/23/2025	18:19	LB137649
	Cobalt	533	500	107	90 - 110	P	10/23/2025	18:19	LB137649
	Copper	5160	5000	103	90 - 110	P	10/23/2025	18:19	LB137649
	Iron	133000	125000	106	90 - 110	P	10/23/2025	18:19	LB137649
	Lead	2670	2500	107	90 - 110	P	10/23/2025	18:19	LB137649
	Magnesium	262000	250000	105	90 - 110	P	10/23/2025	18:19	LB137649
	Manganese	5280	5000	106	90 - 110	P	10/23/2025	18:19	LB137649
	Nickel	516	500	103	90 - 110	P	10/23/2025	18:19	LB137649
	Potassium	131000	125000	104	90 - 110	P	10/23/2025	18:19	LB137649
	Selenium	510	500	102	90 - 110	P	10/23/2025	18:19	LB137649
	Silver	528	500	106	90 - 110	P	10/23/2025	18:19	LB137649
	Sodium	260000	250000	104	90 - 110	P	10/23/2025	18:19	LB137649
	Thallium	546	500	109	90 - 110	P	10/23/2025	18:19	LB137649
	Vanadium	542	500	108	90 - 110	P	10/23/2025	18:19	LB137649
	Zinc	5020	5000	100	90 - 110	P	10/23/2025	18:19	LB137649