

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: AECO02
SDG NO.: Q3419

Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PQ070973.D</u>		CF 750 = <u>PQ070974.D</u>			
CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		CF 050 = <u>PQ070977.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	152632258	161381460	169566884	175492100	161647220	164143984
Aroclor-1016-2	(2)	235274296	256838307	264605432	271518004	253235720	256294352
Aroclor-1016-3	(3)	145675938	156782704	164416996	169596776	159119400	159118363
Aroclor-1016-4	(4)	120162951	129378400	134848252	138303760	126953020	129929277
Aroclor-1016-5	(5)	117231883	128545740	133336348	136306768	128593380	128802824
Aroclor-1260-1	(1)	208928914	234195365	241877518	247859876	239493660	234471067
Aroclor-1260-2	(2)	243695794	275266251	287710272	296720404	279652200	276608984
Aroclor-1260-3	(3)	178017013	196424459	204653606	209590628	207855920	199308325
Aroclor-1260-4	(4)	191556315	214263107	223690164	228814336	216092480	214883280
Aroclor-1260-5	(5)	400986941	436898331	447507618	451939420	431553500	433777162
Decachlorobiphenyl		2603997920	2888430733	3017037660	3097201720	2913248400	2903983287
Tetrachloro-m-xylene		4048991050	4420325093	4447782420	4497118040	4337178200	4350278961
Aroclor-1242-1	(1)	117929900	120501580	121917368	124177888	110233240	118951995
Aroclor-1242-2	(2)	179238730	188762288	194471730	197719720	193371880	190712870
Aroclor-1242-3	(3)	111298009	117265485	120728326	123754708	121434700	118896246
Aroclor-1242-4	(4)	91702396	97376287	98944784	101196952	108558560	99555796
Aroclor-1242-5	(5)	63322787	67097277	82718820	102665812	77804800	78721899
Decachlorobiphenyl		3082568640	3243406933	3352805760	3359769360	3286790800	3265068299
Tetrachloro-m-xylene		4441893010	4577290613	4623065280	4605855840	4702516200	4590124189
Aroclor-1248-1	(1)	89951061	92802913	99278242	97399904	67891380	89464700
Aroclor-1248-2	(2)	126849870	131764987	138900528	142257364	142462060	136446962
Aroclor-1248-3	(3)	152980867	157544657	166146850	167481096	162764480	161383590
Aroclor-1248-4	(4)	167077290	173126060	182827440	186713196	157721500	173493097
Aroclor-1248-5	(5)	163667698	168136284	132211082	121389384	159185140	148917918
Decachlorobiphenyl		3177348440	3248518907	3402766540	3434790520	3353916400	3323468161
Tetrachloro-m-xylene		4546508980	4599154480	4739413180	4685828360	4518221600	4617825320
Aroclor-1254-1	(1)	186317660	198212097	200308250	222421948	174131700	196278331
Aroclor-1254-2	(2)	272465681	301191412	300558232	318683640	259516480	290483089
Aroclor-1254-3	(3)	294493285	320137152	318530722	335107472	267769020	307207530
Aroclor-1254-4	(4)	212636724	232647637	229798130	237628972	184991000	219540493
Aroclor-1254-5	(5)	229095618	241104957	242576814	249415576	197285560	231895705
Decachlorobiphenyl		3135993520	3237057933	3300114540	3432442200	2724290800	3165979799
Tetrachloro-m-xylene		4449551480	4584042107	4604497820	4625575480	3701892800	4393111937
Aroclor-1268-1	(1)	555593291	572438855	583365466	598015136	603747960	582632142

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	510325576	524057337	534075228	546100468	539884140	530888550	3
Aroclor-1268-3	(3)	416538407	427532203	436646098	443325192	442459640	433300308	3
Aroclor-1268-4	(4)	179373252	185531188	189624920	193669728	190155160	187670850	3
Aroclor-1268-5	(5)	1283167757	1308010297	1321380362	1326514116	1318957680	1311606042	1
Decachlorobiphenyl		5757514290	5913646120	6037365900	6192444040	6223042400	6024802550	3
Tetrachloro-m-xylene		4734874420	4871269707	4916225920	4964521920	4932484000	4883875193	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: AECO02
SDG NO.: Q3419

Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PQ070973.D</u>		CF 750 = <u>PQ070974.D</u>			
CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		CF 050 = <u>PQ070977.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	198094739	217218485	228189888	237783008	238723760	224001976
Aroclor-1016-2	(2)	306974778	336928535	345611618	359075648	352024660	340123048
Aroclor-1016-3	(3)	164197912	180535471	187035142	195263012	182262860	181858879
Aroclor-1016-4	(4)	139678855	154088441	161162796	170890484	161242760	157412667
Aroclor-1016-5	(5)	171310856	187991799	196963118	205633060	197310680	191841903
Aroclor-1260-1	(1)	322104344	352779140	365742242	382158164	379245720	360405922
Aroclor-1260-2	(2)	392921251	432078265	450339936	466790948	457838380	439993756
Aroclor-1260-3	(3)	374441100	409997701	423595070	436050024	430913160	414999411
Aroclor-1260-4	(4)	324169048	356082777	368301826	380117112	370143240	359762801
Aroclor-1260-5	(5)	773923407	822481391	873089514	880740576	851175200	840282018
Decachlorobiphenyl		4511855410	4959881947	5165429800	5359918520	5250729800	5049563095
Tetrachloro-m-xylene		4790824740	5183285133	5313707940	5302586120	4809696600	5080020107
Aroclor-1242-1	(1)	149120136	160366853	160219200	165386980	179219740	162862582
Aroclor-1242-2	(2)	228406508	237737935	244055020	248772624	247903320	241375081
Aroclor-1242-3	(3)	123267302	131780304	132089372	135429308	136535900	131820437
Aroclor-1242-4	(4)	107929318	113409847	120761922	128344628	102401880	114569519
Aroclor-1242-5	(5)	175313049	196019013	175387368	167275268	156704980	174139936
Decachlorobiphenyl		5194557400	5483323560	5616478880	5698756520	6104877800	5619598832
Tetrachloro-m-xylene		5511103540	5712989200	5725543240	5446452960	5436162600	5566450308
Aroclor-1248-1	(1)	118147108	120641916	127443132	132200888	134031720	126492953
Aroclor-1248-2	(2)	172958749	177705395	189325858	196101732	206725040	188563355
Aroclor-1248-3	(3)	167593672	172934769	178216680	182062792	197467520	179655087
Aroclor-1248-4	(4)	207432918	214191484	230649658	232116576	237264740	224331075
Aroclor-1248-5	(5)	211078475	216720032	243430792	250826296	242157620	232842643
Decachlorobiphenyl		5357284330	5460780160	5730809660	5844375240	6130897000	5704829278
Tetrachloro-m-xylene		5499268270	5571294053	5722936200	5664492400	5408993600	5573396905
Aroclor-1254-1	(1)	326099704	341297420	345505280	377987212	295169840	337211891
Aroclor-1254-2	(2)	283781194	302336180	304849560	326229636	268575420	297154398
Aroclor-1254-3	(3)	455841488	477545104	484627186	501619228	415693040	467065209
Aroclor-1254-4	(4)	291449906	303039177	308578712	317073776	256115840	295251482
Aroclor-1254-5	(5)	417470468	427079459	439495106	452352324	365350180	420349507
Decachlorobiphenyl		5360763820	5577066440	5716894620	5972893640	5256420800	5576807864
Tetrachloro-m-xylene		5311856950	5451428933	5484489360	5522984680	4238156400	5201783265
Aroclor-1268-1	(1)	931455367	961712979	973652708	974423320	1030986200	974446115

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	858872879	891158693	900594916	906701084	936258500	898717214	3
Aroclor-1268-3	(3)	746016914	769745548	779615614	791820108	821422560	781724149	4
Aroclor-1268-4	(4)	330162118	339473573	345636662	356393996	350955500	344524370	3
Aroclor-1268-5	(5)	2374681378	2422263161	2432472106	2427996872	2444443520	2420371407	1
Decachlorobiphenyl		10401085910	10736618840	10941521580	11269357960	11896046400	11048926138	5
Tetrachloro-m-xylene		5581659480	5729850200	5733063640	5828463760	5669436400	5708494696	2