

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114652.D **Time Analyzed:** 09:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.491	5.392	5.592	595.410	500.000	19.1
Aroclor-1016-2	5.513	5.414	5.614	563.270	500.000	12.7
Aroclor-1016-3	5.574	5.476	5.676	552.740	500.000	10.5
Aroclor-1016-4	5.670	5.571	5.771	576.670	500.000	15.3
Aroclor-1016-5	5.961	5.863	6.063	598.910	500.000	19.8
Aroclor-1260-1	7.074	6.975	7.175	582.490	500.000	16.5
Aroclor-1260-2	7.328	7.229	7.429	570.150	500.000	14.0
Aroclor-1260-3	7.685	7.585	7.785	513.160	500.000	2.6
Aroclor-1260-4	7.907	7.807	8.007	564.370	500.000	12.9
Aroclor-1260-5	8.214	8.113	8.313	523.570	500.000	4.7
Decachlorobiphenyl	9.935	9.834	10.034	53.270	50.000	6.5
Tetrachloro-m-xylene	4.352	4.255	4.455	59.030	50.000	18.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114652.D **Time Analyzed:** 09:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.643	4.843	522.020	500.000	4.4
Aroclor-1016-2	4.753	4.662	4.862	519.170	500.000	3.8
Aroclor-1016-3	4.928	4.836	5.036	543.990	500.000	8.8
Aroclor-1016-4	4.970	4.879	5.079	550.990	500.000	10.2
Aroclor-1016-5	5.182	5.092	5.292	545.410	500.000	9.1
Aroclor-1260-1	6.213	6.125	6.325	572.900	500.000	14.6
Aroclor-1260-2	6.401	6.313	6.513	588.880	500.000	17.8
Aroclor-1260-3	6.554	6.466	6.666	546.890	500.000	9.4
Aroclor-1260-4	7.025	6.939	7.139	594.800	500.000	19.0
Aroclor-1260-5	7.265	7.179	7.379	596.500	500.000	19.3
Decachlorobiphenyl	8.651	8.571	8.771	50.870	50.000	1.7
Tetrachloro-m-xylene	3.655	3.560	3.760	51.330	50.000	2.7