

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114667.D **Time Analyzed:** 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.490	5.392	5.592	579.100	500.000	15.8
Aroclor-1016-2	5.512	5.414	5.614	550.560	500.000	10.1
Aroclor-1016-3	5.574	5.476	5.676	547.180	500.000	9.4
Aroclor-1016-4	5.669	5.571	5.771	580.280	500.000	16.1
Aroclor-1016-5	5.961	5.863	6.063	583.840	500.000	16.8
Aroclor-1260-1	7.073	6.975	7.175	571.610	500.000	14.3
Aroclor-1260-2	7.328	7.229	7.429	552.890	500.000	10.6
Aroclor-1260-3	7.685	7.585	7.785	496.660	500.000	-0.7
Aroclor-1260-4	7.907	7.807	8.007	538.630	500.000	7.7
Aroclor-1260-5	8.213	8.113	8.313	519.000	500.000	3.8
Decachlorobiphenyl	9.935	9.834	10.034	51.810	50.000	3.6
Tetrachloro-m-xylene	4.352	4.255	4.455	57.770	50.000	15.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114667.D **Time Analyzed:** 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.643	4.843	511.520	500.000	2.3
Aroclor-1016-2	4.753	4.662	4.862	508.980	500.000	1.8
Aroclor-1016-3	4.928	4.836	5.036	530.730	500.000	6.1
Aroclor-1016-4	4.970	4.879	5.079	535.250	500.000	7.1
Aroclor-1016-5	5.182	5.092	5.292	534.300	500.000	6.9
Aroclor-1260-1	6.213	6.125	6.325	553.070	500.000	10.6
Aroclor-1260-2	6.400	6.313	6.513	571.400	500.000	14.3
Aroclor-1260-3	6.553	6.466	6.666	539.470	500.000	7.9
Aroclor-1260-4	7.024	6.939	7.139	607.610	500.000	21.5
Aroclor-1260-5	7.265	7.179	7.379	585.210	500.000	17.0
Decachlorobiphenyl	8.650	8.571	8.771	50.070	50.000	0.1
Tetrachloro-m-xylene	3.655	3.560	3.760	49.980	50.000	0.0