

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071149.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	582.660	500.000	16.5
Aroclor-1016-2	4.499	4.399	4.599	562.340	500.000	12.5
Aroclor-1016-3	4.556	4.456	4.656	584.520	500.000	16.9
Aroclor-1016-4	4.646	4.546	4.746	585.390	500.000	17.1
Aroclor-1016-5	4.931	4.831	5.031	576.280	500.000	15.3
Aroclor-1260-1	6.020	5.919	6.119	558.050	500.000	11.6
Aroclor-1260-2	6.276	6.175	6.375	557.500	500.000	11.5
Aroclor-1260-3	6.626	6.526	6.726	583.750	500.000	16.8
Aroclor-1260-4	6.841	6.740	6.940	592.250	500.000	18.5
Aroclor-1260-5	7.146	7.045	7.245	579.210	500.000	15.8
Decachlorobiphenyl	8.497	8.396	8.596	59.340	50.000	18.7
Tetrachloro-m-xylene	3.400	3.300	3.500	57.140	50.000	14.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071149.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	461.700	500.000	-7.7
Aroclor-1016-2	3.791	3.693	3.893	465.260	500.000	-6.9
Aroclor-1016-3	3.951	3.853	4.053	480.210	500.000	-4.0
Aroclor-1016-4	4.000	3.902	4.102	488.300	500.000	-2.3
Aroclor-1016-5	4.194	4.097	4.297	480.390	500.000	-3.9
Aroclor-1260-1	5.187	5.089	5.289	477.270	500.000	-4.5
Aroclor-1260-2	5.376	5.278	5.478	476.660	500.000	-4.7
Aroclor-1260-3	5.517	5.419	5.619	480.230	500.000	-4.0
Aroclor-1260-4	5.977	5.879	6.079	481.030	500.000	-3.8
Aroclor-1260-5	6.222	6.124	6.324	480.940	500.000	-3.8
Decachlorobiphenyl	7.531	7.433	7.633	56.210	50.000	12.4
Tetrachloro-m-xylene	2.776	2.678	2.878	49.690	50.000	-0.6