

Instrument ID: ECD\_O

**Daily Analysis Runlog For Sequence/QC Batch ID # PO100925**

Review By	yogesh	Review On	10/9/2025 1:00:25 PM
Supervise By	mohammad	Supervise On	10/11/2025 1:35:39 AM
SubDirectory	PO100925	HP Acquire Method	HP Processing Method PO100925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114311.D	09 Oct 2025 13:37		YP/AJ	Ok
2	I.BLK	I.BLK	PO114312.D	09 Oct 2025 13:55		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114313.D	09 Oct 2025 14:14		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114314.D	09 Oct 2025 14:32		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114315.D	09 Oct 2025 14:50		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114316.D	09 Oct 2025 15:09		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO114317.D	09 Oct 2025 15:27		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114318.D	09 Oct 2025 15:45		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114319.D	09 Oct 2025 16:04		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114320.D	09 Oct 2025 16:22		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114321.D	09 Oct 2025 16:40		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114322.D	09 Oct 2025 16:59		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114323.D	09 Oct 2025 17:17		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114324.D	09 Oct 2025 17:35		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114325.D	09 Oct 2025 17:54		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114326.D	09 Oct 2025 18:12		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114327.D	09 Oct 2025 18:30		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114328.D	09 Oct 2025 18:49		YP/AJ	Ok,M

Instrument ID: ECD\_O

**Daily Analysis Runlog For Sequence/QC Batch ID # PO100925**

Review By	yogesh	Review On	10/9/2025 1:00:25 PM
Supervise By	mohammad	Supervise On	10/11/2025 1:35:39 AM
SubDirectory	PO100925	HP Acquire Method	HP Processing Method PO100925
<b>STD. NAME</b>	<b>STD REF.#</b>		
Tune/Reschk Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114329.D	09 Oct 2025 19:07		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO114330.D	09 Oct 2025 19:26		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114331.D	09 Oct 2025 19:44		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114332.D	09 Oct 2025 20:02		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114333.D	09 Oct 2025 20:21		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114334.D	09 Oct 2025 20:39		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO114335.D	09 Oct 2025 20:57		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114336.D	09 Oct 2025 21:16		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114337.D	09 Oct 2025 21:34		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114338.D	09 Oct 2025 21:52		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114339.D	09 Oct 2025 22:11		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114340.D	09 Oct 2025 22:29		YP/AJ	Ok,M
31	PO100925ICV500	ICVPO100925	PO114341.D	09 Oct 2025 22:47		YP/AJ	Ok
32	AR1242ICV500	ICVPO100925AR1242	PO114342.D	09 Oct 2025 23:24		YP/AJ	Ok
33	AR1248ICV500	ICVPO100925AR1248	PO114343.D	10 Oct 2025 00:01		YP/AJ	Ok
34	AR1254ICV500	ICVPO100925AR1254	PO114344.D	10 Oct 2025 00:38		YP/AJ	Ok
35	AR1268ICV500	ICVPO100925AR1268	PO114345.D	10 Oct 2025 01:14		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114346.D	10 Oct 2025 01:51		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD\_O

**Daily Analysis Runlog For Sequence/QC Batch ID # PO102425**

Review By	yogesh	Review On	10/24/2025 11:08:58 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:25 AM
SubDirectory	PO102425	HP Acquire Method	HP Processing Method PO100925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114651.D	24 Oct 2025 08:45		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO114652.D	24 Oct 2025 09:03		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO114653.D	24 Oct 2025 09:29		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO114654.D	24 Oct 2025 09:47		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO114655.D	24 Oct 2025 10:05		YP/AJ	Ok
6	I.BLK	I.BLK	PO114656.D	24 Oct 2025 10:24		YP/AJ	Ok
7	Q3437-14	SS-25(0-0.5)	PO114657.D	24 Oct 2025 10:42	AR1254 hits	YP/AJ	Dilution
8	Q3437-16	SS-23(0-0.5)	PO114658.D	24 Oct 2025 11:00	AR1254 hits, carry over	YP/AJ	Not Ok
9	Q3419-01	PR-132-S16-015094-20	PO114659.D	24 Oct 2025 11:19	AR1242 & AR1254 hits	YP/AJ	Not Ok
10	Q3437-14DL	SS-25(0-0.5)DL	PO114660.D	24 Oct 2025 11:37	AR1254 hits	YP/AJ	Ok
11	Q3419-03MSD	PR-132-S16-015094-20	PO114661.D	24 Oct 2025 11:55	need to run MS	YP/AJ	Not Ok
12	Q3419-04	PR132-S16-119133-20	PO114662.D	24 Oct 2025 12:14		YP/AJ	Ok
13	Q3419-05	PR132-S16-094119-20	PO114663.D	24 Oct 2025 12:32	AR1242 & AR1254 hits	YP/AJ	Ok,M
14	Q3431-01	EO-02-10222025	PO114664.D	24 Oct 2025 12:49	TCMX high in first column	YP/AJ	Ok
15	Q3439-05	TP-10	PO114665.D	24 Oct 2025 13:08	DCB high in 1st column & TCMX high in both column	YP/AJ	ReRun
16	Q3440-01	JB-2	PO114666.D	24 Oct 2025 13:26	AR1254+1260 Hit	YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO114667.D	24 Oct 2025 14:35		YP/AJ	Ok,M

Instrument ID: ECD\_O

**Daily Analysis Runlog For Sequence/QC Batch ID # PO102425**

Review By	yogesh	Review On	10/24/2025 11:08:58 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:25 AM
SubDirectory	PO102425	HP Acquire Method	HP Processing Method PO100925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Run #	Sample Name	Reference	Method	Time	Operator	Status
18	AR1242CCC500	AR1242CCC500	PO114668.D	24 Oct 2025 14:54	YP/AJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PO114669.D	24 Oct 2025 15:12	YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO114670.D	24 Oct 2025 15:31	YP/AJ	Ok
21	I.BLK	I.BLK	PO114671.D	24 Oct 2025 15:49	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP102125**

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YPIAJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YPIAJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YPIAJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YPIAJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YPIAJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YPIAJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YPIAJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YPIAJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YPIAJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YPIAJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YPIAJ	Ok,M

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP102125**

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
<b>STD. NAME</b>	<b>STD REF.#</b>		
Tune/Reschk Initial Calibration Stds  CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854  PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP102425**

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076011.D	24 Oct 2025 08:41		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076012.D	24 Oct 2025 08:57		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076013.D	24 Oct 2025 09:29		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076014.D	24 Oct 2025 09:45		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076015.D	24 Oct 2025 10:02		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076016.D	24 Oct 2025 10:18		YPIAJ	Ok
7	PB170234BL	PB170234BL	PP076017.D	24 Oct 2025 10:34		YPIAJ	Ok
8	PB170234BS	PB170234BS	PP076018.D	24 Oct 2025 10:51		YPIAJ	Ok
9	Q3437-03	SS-15R(2.0-2.5)	PP076019.D	24 Oct 2025 11:07	AR1254 hits	YPIAJ	Ok
10	Q3437-05	SS-26(0-0.5)	PP076020.D	24 Oct 2025 11:23	AR1254 hits	YPIAJ	Ok,M
11	Q3437-11	SS-27(0-0.5)	PP076021.D	24 Oct 2025 11:39	AR1254 hits	YPIAJ	Ok,M
12	Q3445-01	EO-01-102325	PP076022.D	24 Oct 2025 11:56		YPIAJ	Ok
13	Q3446-01	SB-14	PP076023.D	24 Oct 2025 12:12	F Flag in TCMX in 1st column, All surrogate fail, AR1242 hits, Need 20X	YPIAJ	Dilution
14	Q3446-02	SB-15	PP076024.D	24 Oct 2025 12:28	AR1242 hits, Need 20X	YPIAJ	Dilution
15	Q3414-12DL2	SB-12DL2	PP076025.D	24 Oct 2025 12:44	AR1242 hit	YPIAJ	Ok
16	Q3414-13DL2	SB-13DL2	PP076026.D	24 Oct 2025 13:01	AR1242 hit	YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP076027.D	24 Oct 2025 14:06		YPIAJ	Ok

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP102425**

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
<b>STD. NAME</b>	<b>STD REF.#</b>		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076028.D	24 Oct 2025 14:22		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076029.D	24 Oct 2025 14:38		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076030.D	24 Oct 2025 14:54		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076031.D	24 Oct 2025 15:11		YPIAJ	Ok
22	Q3414-07DL2	SB-7DL2	PP076032.D	24 Oct 2025 15:27	AR1242 hit	YPIAJ	Ok
23	Q3446-06	SB-19	PP076033.D	24 Oct 2025 15:43	F Flag in TCMX in 1st column , AR1242+1260 Hit , TCMX high in both column , DCB low in 1st column , Need dilution	YPIAJ	Dilution
24	Q3446-07	SB-20	PP076034.D	24 Oct 2025 16:00	AR1260 Hit , Need dilution , TCMX high in 2nd column	YPIAJ	Dilution
25	Q3446-08	SB-21	PP076035.D	24 Oct 2025 16:16	AR1242+1260 Hit , DCB low in both column , TCMX high in both column , need dilution	YPIAJ	Dilution
26	Q3446-03	SB-16	PP076036.D	24 Oct 2025 16:32	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
27	Q3446-04	SB-17	PP076037.D	24 Oct 2025 16:48	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
28	Q3446-01DL	SB-14DL	PP076038.D	24 Oct 2025 17:05	AR1242 hit	YPIAJ	Ok,M
29	Q3446-02DL	SB-15DL	PP076039.D	24 Oct 2025 17:21	AR1242 hit	YPIAJ	Ok,M
30	Q3446-05DL	SB-18DL	PP076040.D	24 Oct 2025 17:37	AR1242 Hit	YPIAJ	Ok,M
31	Q3446-06DL	SB-19DL	PP076041.D	24 Oct 2025 17:53	AR1242+1260 Hit	YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP076042.D	24 Oct 2025 18:58		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076043.D	24 Oct 2025 19:31		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076044.D	24 Oct 2025 20:03		YPIAJ	Ok,M



Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP102425**

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
<b>STD. NAME</b>	<b>STD REF.#</b>		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	AR1254CCC500	AR1254CCC500	PP076045.D	24 Oct 2025 20:19		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP076046.D	24 Oct 2025 20:36		YPIAJ	Ok
37	Q3450-10	SS-20-(3.5-4.0)	PP076047.D	24 Oct 2025 20:52	AR1254 Hit	YPIAJ	Ok,M
38	Q3450-11	SS-28-(2.0-2.5)	PP076048.D	24 Oct 2025 21:08		YPIAJ	Ok
39	Q3450-20	SS-21-(0.0-0.5)	PP076049.D	24 Oct 2025 21:24	AR1254 Hit , Need 5x dilution	YPIAJ	Dilution
40	Q3450-21	SS-21-(2.0-2.5)	PP076050.D	24 Oct 2025 21:41		YPIAJ	Ok
41	Q3450-22	SS-21-(3.5-4.0)	PP076051.D	24 Oct 2025 21:57	AR1254 Hit , Need 2x dilution	YPIAJ	Dilution
42	Q3427-01	TP-8	PP076052.D	24 Oct 2025 22:13		YPIAJ	Ok,M
43	Q3446-07DL	SB-20DL	PP076053.D	24 Oct 2025 22:29	AR1242 Hit	YPIAJ	Ok,M
44	Q3446-08DL	SB-21DL	PP076054.D	24 Oct 2025 22:45	AR1242+1260 Hit	YPIAJ	Ok,M
45	Q3446-03DL	SB-16DL	PP076055.D	24 Oct 2025 23:01	AR1260 Hit	YPIAJ	Ok,M
46	Q3446-04DL	SB-17DL	PP076056.D	24 Oct 2025 23:17	AR1260 Hit	YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP076057.D	25 Oct 2025 00:22		YPIAJ	Ok
48	AR1242CCC500	AR1242CCC500	PP076058.D	25 Oct 2025 01:11		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP076059.D	25 Oct 2025 01:43		YPIAJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP076060.D	25 Oct 2025 01:59		YPIAJ	Ok,M
51	I.BLK	I.BLK	PP076061.D	25 Oct 2025 02:16		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025**

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ070971.D	10 Oct 2025 08:27		YPIAJ	Ok
2	I.BLK	I.BLK	PQ070972.D	10 Oct 2025 08:42		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35		YPIAJ	Ok,M

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025**

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method	PQ101025	
<b>STD. NAME</b>	<b>STD REF.#</b>				
Tune/Reschk Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854				
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862				

19	AR1248ICC050	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15		YPIAJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43		YPIAJ	Ok
31	PQ101025ICV	ICVPQ101025	PQ071001.D	10 Oct 2025 16:57		YPIAJ	Ok,M
32	AR1242ICV500	ICVPQ101025AR1242	PQ071002.D	10 Oct 2025 17:41		YPIAJ	Ok,M
33	AR1248ICV500	ICVPQ101025AR1248	PQ071003.D	10 Oct 2025 18:25		YPIAJ	Ok,M
34	AR1254ICV500	ICVPQ101025AR1254	PQ071004.D	10 Oct 2025 19:08		YPIAJ	Ok
35	AR1268ICV500	ICVPQ101025AR1268	PQ071005.D	10 Oct 2025 19:52		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725**

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071148.D	27 Oct 2025 08:42		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57		YPIAJ	Ok
3	I.BLK	I.BLK	PQ071150.D	27 Oct 2025 09:11		YPIAJ	Ok
4	Q3450-01	SS-5R-(2.0-2.5)	PQ071151.D	27 Oct 2025 09:26	AR1254 Hit	YPIAJ	Ok
5	Q3450-02	SS-5R-(3.5-4.0)	PQ071152.D	27 Oct 2025 09:41	AR1254 Hit	YPIAJ	Ok
6	Q3450-02MS	SS-5R-(3.5-4.0)MS	PQ071153.D	27 Oct 2025 09:55		YPIAJ	Ok,M
7	Q3450-02MSD	SS-5R-(3.5-4.0)MSD	PQ071154.D	27 Oct 2025 10:10		YPIAJ	Ok,M
8	Q3450-08	SS-20-(0.0-0.5)	PQ071155.D	27 Oct 2025 10:24	AR1254 Hit	YPIAJ	Ok
9	Q3450-09	SS-20-(2.0-2.5)	PQ071156.D	27 Oct 2025 10:39	AR1254 Hit	YPIAJ	Ok,M
10	Q3437-16	SS-23(0-0.5)	PQ071157.D	27 Oct 2025 10:54	AR1254 Hit, TCX high 1st column	YPIAJ	Ok
11	Q3419-01	PR-132-S16-015094-20	PQ071158.D	27 Oct 2025 11:08	AR1242 hits	YPIAJ	Ok,M
12	Q3419-02MS	PR-132-S16-015094-20	PQ071159.D	27 Oct 2025 11:23		YPIAJ	Ok
13	Q3419-03MSD	PR-132-S16-015094-20	PQ071160.D	27 Oct 2025 11:37		YPIAJ	Ok
14	AR1660CCC500	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35		YPIAJ	Ok
15	AR1242CCC500	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34		YPIAJ	Ok
16	AR1248CCC500	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49		YPIAJ	Ok
17	AR1254CCC500	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03		YPIAJ	Ok

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725**

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025
<b>STD. NAME</b>	<b>STD REF.#</b>		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	I.BLK	I.BLK	PQ071165.D	27 Oct 2025 14:18		YPIAJ	Ok
19	PB170255BL	PB170255BL	PQ071166.D	27 Oct 2025 14:32	DCB high 1st column for PSEG	YPIAJ	Not Ok
20	PB170255BS	PB170255BS	PQ071167.D	27 Oct 2025 14:47	DCB high 1st column for PSEG	YPIAJ	Not Ok
21	Q3451-01	TP-11	PQ071168.D	27 Oct 2025 15:01	Tcmx high in 1st column	YPIAJ	Ok
22	Q3451-05	TP-12	PQ071169.D	27 Oct 2025 15:16		YPIAJ	Ok
23	Q3451-05MS	TP-12MS	PQ071170.D	27 Oct 2025 15:31	Tcmx high in 1st column	YPIAJ	Ok
24	Q3451-05MSD	TP-12MSD	PQ071171.D	27 Oct 2025 15:45	Tcmx high in 1st column	YPIAJ	Ok
25	Q3452-01	JB-1	PQ071172.D	27 Oct 2025 16:00		YPIAJ	Ok
26	Q3460-05	WC-2	PQ071173.D	27 Oct 2025 16:14	TCMX & DCB high in 1st column	YPIAJ	ReRun
27	Q3460-09	WC-3	PQ071174.D	27 Oct 2025 16:29	TCMX & DCB high in 1st column	YPIAJ	ReRun
28	Q3460-13	WC-4	PQ071175.D	27 Oct 2025 16:43	DCB high in 1st column, AR1260 hits	YPIAJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42		YPIAJ	Ok,M
30	AR1242CCC500	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40		YPIAJ	Ok
31	AR1248CCC500	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55		YPIAJ	Ok
32	AR1254CCC500	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09		YPIAJ	Ok,M
33	I.BLK	I.BLK	PQ071180.D	27 Oct 2025 19:24		YPIAJ	Ok

M : Manual Integration