

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO102425

Review By	yogesh	Review On	10/24/2025 11:08:58 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:25 AM
SubDirectory	PO102425	HP Acquire Method	HP Processing Method PO100925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114651.D	24 Oct 2025 08:45		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO114652.D	24 Oct 2025 09:03		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO114653.D	24 Oct 2025 09:29		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO114654.D	24 Oct 2025 09:47		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO114655.D	24 Oct 2025 10:05		YP/AJ	Ok
6	I.BLK	I.BLK	PO114656.D	24 Oct 2025 10:24		YP/AJ	Ok
7	Q3437-14	SS-25(0-0.5)	PO114657.D	24 Oct 2025 10:42	AR1254 hits	YP/AJ	Dilution
8	Q3437-16	SS-23(0-0.5)	PO114658.D	24 Oct 2025 11:00	AR1254 hits, carry over	YP/AJ	Not Ok
9	Q3419-01	PR-132-S16-015094-20	PO114659.D	24 Oct 2025 11:19	AR1242 & AR1254 hits	YP/AJ	Not Ok
10	Q3437-14DL	SS-25(0-0.5)DL	PO114660.D	24 Oct 2025 11:37	AR1254 hits	YP/AJ	Ok
11	Q3419-03MSD	PR-132-S16-015094-20	PO114661.D	24 Oct 2025 11:55	need to run MS	YP/AJ	Not Ok
12	Q3419-04	PR132-S16-119133-20	PO114662.D	24 Oct 2025 12:14		YP/AJ	Ok
13	Q3419-05	PR132-S16-094119-20	PO114663.D	24 Oct 2025 12:32	AR1242 & AR1254 hits	YP/AJ	Ok,M
14	Q3431-01	EO-02-10222025	PO114664.D	24 Oct 2025 12:49	TCMX high in first column	YP/AJ	Ok
15	Q3439-05	TP-10	PO114665.D	24 Oct 2025 13:08	DCB high in 1st column & TCMX high in both column	YP/AJ	ReRun
16	Q3440-01	JB-2	PO114666.D	24 Oct 2025 13:26	AR1254+1260 Hit	YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO114667.D	24 Oct 2025 14:35		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO102425

Review By	yogesh	Review On	10/24/2025 11:08:58 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:25 AM
SubDirectory	PO102425	HP Acquire Method	HP Processing Method PO100925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO114668.D	24 Oct 2025 14:54		YP/AJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PO114669.D	24 Oct 2025 15:12		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO114670.D	24 Oct 2025 15:31		YP/AJ	Ok
21	I.BLK	I.BLK	PO114671.D	24 Oct 2025 15:49		YP/AJ	Ok

M : Manual Integration