

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076011.D	24 Oct 2025 08:41		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076012.D	24 Oct 2025 08:57		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076013.D	24 Oct 2025 09:29		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076014.D	24 Oct 2025 09:45		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076015.D	24 Oct 2025 10:02		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076016.D	24 Oct 2025 10:18		YPIAJ	Ok
7	PB170234BL	PB170234BL	PP076017.D	24 Oct 2025 10:34		YPIAJ	Ok
8	PB170234BS	PB170234BS	PP076018.D	24 Oct 2025 10:51		YPIAJ	Ok
9	Q3437-03	SS-15R(2.0-2.5)	PP076019.D	24 Oct 2025 11:07	AR1254 hits	YPIAJ	Ok
10	Q3437-05	SS-26(0-0.5)	PP076020.D	24 Oct 2025 11:23	AR1254 hits	YPIAJ	Ok,M
11	Q3437-11	SS-27(0-0.5)	PP076021.D	24 Oct 2025 11:39	AR1254 hits	YPIAJ	Ok,M
12	Q3445-01	EO-01-102325	PP076022.D	24 Oct 2025 11:56		YPIAJ	Ok
13	Q3446-01	SB-14	PP076023.D	24 Oct 2025 12:12	F Flag in TCMX in 1st column, All surrogate fail, AR1242 hits, Need 20X	YPIAJ	Dilution
14	Q3446-02	SB-15	PP076024.D	24 Oct 2025 12:28	AR1242 hits, Need 20X	YPIAJ	Dilution
15	Q3414-12DL2	SB-12DL2	PP076025.D	24 Oct 2025 12:44	AR1242 hit	YPIAJ	Ok
16	Q3414-13DL2	SB-13DL2	PP076026.D	24 Oct 2025 13:01	AR1242 hit	YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP076027.D	24 Oct 2025 14:06		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076028.D	24 Oct 2025 14:22		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076029.D	24 Oct 2025 14:38		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076030.D	24 Oct 2025 14:54		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076031.D	24 Oct 2025 15:11		YPIAJ	Ok
22	Q3414-07DL2	SB-7DL2	PP076032.D	24 Oct 2025 15:27	AR1242 hit	YPIAJ	Ok
23	Q3446-06	SB-19	PP076033.D	24 Oct 2025 15:43	F Flag in TCMX in 1st column , AR1242+1260 Hit , TCMX high in both column , DCB low in 1st column , Need dilution	YPIAJ	Dilution
24	Q3446-07	SB-20	PP076034.D	24 Oct 2025 16:00	AR1260 Hit , Need dilution , TCMX high in 2nd column	YPIAJ	Dilution
25	Q3446-08	SB-21	PP076035.D	24 Oct 2025 16:16	AR1242+1260 Hit , DCB low in both column , TCMX high in both column , need dilution	YPIAJ	Dilution
26	Q3446-03	SB-16	PP076036.D	24 Oct 2025 16:32	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
27	Q3446-04	SB-17	PP076037.D	24 Oct 2025 16:48	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
28	Q3446-01DL	SB-14DL	PP076038.D	24 Oct 2025 17:05	AR1242 hit	YPIAJ	Ok,M
29	Q3446-02DL	SB-15DL	PP076039.D	24 Oct 2025 17:21	AR1242 hit	YPIAJ	Ok,M
30	Q3446-05DL	SB-18DL	PP076040.D	24 Oct 2025 17:37	AR1242 Hit	YPIAJ	Ok,M
31	Q3446-06DL	SB-19DL	PP076041.D	24 Oct 2025 17:53	AR1242+1260 Hit	YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP076042.D	24 Oct 2025 18:58		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076043.D	24 Oct 2025 19:31		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076044.D	24 Oct 2025 20:03		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	AR1254CCC500	AR1254CCC500	PP076045.D	24 Oct 2025 20:19		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP076046.D	24 Oct 2025 20:36		YPIAJ	Ok
37	Q3450-10	SS-20-(3.5-4.0)	PP076047.D	24 Oct 2025 20:52	AR1254 Hit	YPIAJ	Ok,M
38	Q3450-11	SS-28-(2.0-2.5)	PP076048.D	24 Oct 2025 21:08		YPIAJ	Ok
39	Q3450-20	SS-21-(0.0-0.5)	PP076049.D	24 Oct 2025 21:24	AR1254 Hit , Need 5x dilution	YPIAJ	Dilution
40	Q3450-21	SS-21-(2.0-2.5)	PP076050.D	24 Oct 2025 21:41		YPIAJ	Ok
41	Q3450-22	SS-21-(3.5-4.0)	PP076051.D	24 Oct 2025 21:57	AR1254 Hit , Need 2x dilution	YPIAJ	Dilution
42	Q3427-01	TP-8	PP076052.D	24 Oct 2025 22:13		YPIAJ	Ok,M
43	Q3446-07DL	SB-20DL	PP076053.D	24 Oct 2025 22:29	AR1242 Hit	YPIAJ	Ok,M
44	Q3446-08DL	SB-21DL	PP076054.D	24 Oct 2025 22:45	AR1242+1260 Hit	YPIAJ	Ok,M
45	Q3446-03DL	SB-16DL	PP076055.D	24 Oct 2025 23:01	AR1260 Hit	YPIAJ	Ok,M
46	Q3446-04DL	SB-17DL	PP076056.D	24 Oct 2025 23:17	AR1260 Hit	YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP076057.D	25 Oct 2025 00:22		YPIAJ	Ok
48	AR1242CCC500	AR1242CCC500	PP076058.D	25 Oct 2025 01:11		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP076059.D	25 Oct 2025 01:43		YPIAJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP076060.D	25 Oct 2025 01:59		YPIAJ	Ok,M
51	I.BLK	I.BLK	PP076061.D	25 Oct 2025 02:16		YPIAJ	Ok

M : Manual Integration