

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071148.D	27 Oct 2025 08:42		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57		YP\AJ	Ok
3	I.BLK	I.BLK	PQ071150.D	27 Oct 2025 09:11		YP\AJ	Ok
4	Q3450-01	SS-5R-(2.0-2.5)	PQ071151.D	27 Oct 2025 09:26	AR1254 Hit	YP\AJ	Ok
5	Q3450-02	SS-5R-(3.5-4.0)	PQ071152.D	27 Oct 2025 09:41	AR1254 Hit	YP\AJ	Ok
6	Q3450-02MS	SS-5R-(3.5-4.0)MS	PQ071153.D	27 Oct 2025 09:55		YP\AJ	Ok,M
7	Q3450-02MSD	SS-5R-(3.5-4.0)MSD	PQ071154.D	27 Oct 2025 10:10		YP\AJ	Ok,M
8	Q3450-08	SS-20-(0.0-0.5)	PQ071155.D	27 Oct 2025 10:24	AR1254 Hit	YP\AJ	Ok
9	Q3450-09	SS-20-(2.0-2.5)	PQ071156.D	27 Oct 2025 10:39	AR1254 Hit	YP\AJ	Ok,M
10	Q3437-16	SS-23(0-0.5)	PQ071157.D	27 Oct 2025 10:54	AR1254 Hit, TCX high 1st column	YP\AJ	Ok
11	Q3419-01	PR-132-S16-015094-20	PQ071158.D	27 Oct 2025 11:08	AR1242 hits	YP\AJ	Ok,M
12	Q3419-02MS	PR-132-S16-015094-20	PQ071159.D	27 Oct 2025 11:23		YP\AJ	Ok
13	Q3419-03MSD	PR-132-S16-015094-20	PQ071160.D	27 Oct 2025 11:37		YP\AJ	Ok
14	AR1660CCC500	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35		YP\AJ	Ok
15	AR1242CCC500	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34		YP\AJ	Ok
16	AR1248CCC500	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49		YP\AJ	Ok
17	AR1254CCC500	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03		YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	I.BLK	I.BLK	PQ071165.D	27 Oct 2025 14:18		YPIAJ	Ok
19	PB170255BL	PB170255BL	PQ071166.D	27 Oct 2025 14:32	DCB high 1st column for PSEG	YPIAJ	Not Ok
20	PB170255BS	PB170255BS	PQ071167.D	27 Oct 2025 14:47	DCB high 1st column for PSEG	YPIAJ	Not Ok
21	Q3451-01	TP-11	PQ071168.D	27 Oct 2025 15:01	Tcmx high in 1st column	YPIAJ	Ok
22	Q3451-05	TP-12	PQ071169.D	27 Oct 2025 15:16		YPIAJ	Ok
23	Q3451-05MS	TP-12MS	PQ071170.D	27 Oct 2025 15:31	Tcmx high in 1st column	YPIAJ	Ok
24	Q3451-05MSD	TP-12MSD	PQ071171.D	27 Oct 2025 15:45	Tcmx high in 1st column	YPIAJ	Ok
25	Q3452-01	JB-1	PQ071172.D	27 Oct 2025 16:00		YPIAJ	Ok
26	Q3460-05	WC-2	PQ071173.D	27 Oct 2025 16:14	TCMX & DCB high in 1st column	YPIAJ	ReRun
27	Q3460-09	WC-3	PQ071174.D	27 Oct 2025 16:29	TCMX & DCB high in 1st column	YPIAJ	ReRun
28	Q3460-13	WC-4	PQ071175.D	27 Oct 2025 16:43	DCB high in 1st column, AR1260 hits	YPIAJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42		YPIAJ	Ok,M
30	AR1242CCC500	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40		YPIAJ	Ok
31	AR1248CCC500	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55		YPIAJ	Ok
32	AR1254CCC500	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09		YPIAJ	Ok,M
33	I.BLK	I.BLK	PQ071180.D	27 Oct 2025 19:24		YPIAJ	Ok

M : Manual Integration