

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076027.D **Time Analyzed:** 14:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.795	5.695	5.895	461.630	500.000	-7.7
Aroclor-1016-2	5.817	5.717	5.917	452.760	500.000	-9.4
Aroclor-1016-3	5.879	5.780	5.980	448.070	500.000	-10.4
Aroclor-1016-4	5.976	5.877	6.077	450.900	500.000	-9.8
Aroclor-1016-5	6.269	6.169	6.369	451.320	500.000	-9.7
Aroclor-1260-1	7.386	7.287	7.487	523.490	500.000	4.7
Aroclor-1260-2	7.639	7.540	7.740	452.380	500.000	-9.5
Aroclor-1260-3	7.997	7.899	8.099	482.820	500.000	-3.4
Aroclor-1260-4	8.225	8.127	8.327	505.200	500.000	1.0
Aroclor-1260-5	8.550	8.452	8.652	478.240	500.000	-4.4
Decachlorobiphenyl	10.416	10.319	10.519	47.970	50.000	-4.1
Tetrachloro-m-xylene	4.643	4.542	4.742	46.210	50.000	-7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3419
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076027.D **Time Analyzed:** 14:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.876	4.778	4.978	490.880	500.000	-1.8
Aroclor-1016-2	4.935	4.836	5.036	504.020	500.000	0.8
Aroclor-1016-3	5.054	4.954	5.154	480.420	500.000	-3.9
Aroclor-1016-4	5.095	4.996	5.196	536.130	500.000	7.2
Aroclor-1016-5	5.310	5.210	5.410	494.710	500.000	-1.1
Aroclor-1260-1	6.337	6.238	6.438	447.540	500.000	-10.5
Aroclor-1260-2	6.523	6.425	6.625	449.850	500.000	-10.0
Aroclor-1260-3	6.676	6.578	6.778	441.980	500.000	-11.6
Aroclor-1260-4	7.145	7.047	7.247	483.880	500.000	-3.2
Aroclor-1260-5	7.386	7.287	7.487	459.520	500.000	-8.1
Decachlorobiphenyl	8.779	8.682	8.882	41.690	50.000	-16.6
Tetrachloro-m-xylene	3.782	3.681	3.881	47.660	50.000	-4.7