

Surrogate Summary

SDG No.: Q3419

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		31	149
		Tetrachloro-m-xyl	1	20	19.4	97		41	134
		Decachlorobiphen	2	20	20.3	102		31	149
		Tetrachloro-m-xyl	2	20	20.0	100		41	134
I.BLK-PD090775.D	PIBLK-PD090775.D	Decachlorobiphen	1	20	21.7	109		57	171
		Tetrachloro-m-xyl	1	20	20.3	101		61	148
		Decachlorobiphen	2	20	24.9	124		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
PB170213BL	PB170213BL	Decachlorobiphen	1	20	20.7	103		20	144
		Tetrachloro-m-xyl	1	20	18.7	94		19	148
		Decachlorobiphen	2	20	23.8	119		20	144
		Tetrachloro-m-xyl	2	20	20.1	101		19	148
PB170213BS	PB170213BS	Decachlorobiphen	1	20	22.1	110		20	144
		Tetrachloro-m-xyl	1	20	17.6	88		19	148
		Decachlorobiphen	2	20	26.0	130		20	144
		Tetrachloro-m-xyl	2	20	18.3	91		19	148
Q3395-03MS	PR132-S07-000102-20251017MS	Decachlorobiphen	1	20	15.2	76		20	144
		Tetrachloro-m-xyl	1	20	14.4	72		19	148
		Decachlorobiphen	2	20	15.5	77		20	144
		Tetrachloro-m-xyl	2	20	15.5	78		19	148
Q3395-04MSD	PR132-S07-000102-20251017MSD	Decachlorobiphen	1	20	16.4	82		20	144
		Tetrachloro-m-xyl	1	20	15.3	76		19	148
		Decachlorobiphen	2	20	16.4	82		20	144
		Tetrachloro-m-xyl	2	20	16.4	82		19	148
I.BLK-PD090788.D	PIBLK-PD090788.D	Decachlorobiphen	1	20	20.9	105		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	23.5	117		57	171
		Tetrachloro-m-xyl	2	20	20.8	104		61	148
Q3419-01	PR-132-S16-015094-20251021	Decachlorobiphen	1	20	12.9	65		20	144
		Tetrachloro-m-xyl	1	20	13.3	67		19	148
		Decachlorobiphen	2	20	13.7	68		20	144
		Tetrachloro-m-xyl	2	20	14.4	72		19	148
Q3419-04	PR132-S16-119133-20251021	Decachlorobiphen	1	20	18.2	91		20	144
		Tetrachloro-m-xyl	1	20	15.7	78		19	148
		Decachlorobiphen	2	20	14.9	74		20	144
		Tetrachloro-m-xyl	2	20	16.5	83		19	148
Q3419-05	PR132-S16-094119-20251021	Decachlorobiphen	1	20	17.0	85		20	144
		Tetrachloro-m-xyl	1	20	13.7	69		19	148
		Decachlorobiphen	2	20	15.3	77		20	144
		Tetrachloro-m-xyl	2	20	15.3	76		19	148
I.BLK-PD090799.D	PIBLK-PD090799.D	Decachlorobiphen	1	20	21.1	105		57	171

Surrogate Summary

SDG No.: Q3419

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090799.D	PIBLK-PD090799.D	Tetrachloro-m-xyl	1	20	19.7	99		61	148
		Decachlorobiphen	2	20	23.4	117		57	171
I.BLK-PD090808.D	PIBLK-PD090808.D	Tetrachloro-m-xyl	2	20	20.3	101		61	148
		Decachlorobiphen	1	20	22.0	110		57	171
		Tetrachloro-m-xyl	1	20	20.8	104		61	148
		Decachlorobiphen	2	20	24.1	120		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
Q3419-02MS	PR-132-S16-015094-20251021M	Decachlorobiphen	1	20	12.3	62		20	144
		Tetrachloro-m-xyl	1	20	12.4	62		19	148
		Decachlorobiphen	2	20	12.6	63		20	144
		Tetrachloro-m-xyl	2	20	12.6	63		19	148
Q3419-03MSD	PR-132-S16-015094-20251021M	Decachlorobiphen	1	20	12.2	61		20	144
		Tetrachloro-m-xyl	1	20	12.1	60		19	148
		Decachlorobiphen	2	20	12.7	64		20	144
		Tetrachloro-m-xyl	2	20	12.6	63		19	148
I.BLK-PD090823.D	PIBLK-PD090823.D	Decachlorobiphen	1	20	17.3	87		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	12.3	61		57	171
		Tetrachloro-m-xyl	2	20	19.6	98		61	148