

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102925

Review By	rahul	Review On	10/30/2025 8:49:10 AM
Supervise By	Jagrut	Supervise On	10/30/2025 4:42:33 PM
SubDirectory	BF102925	HP Acquire Method	BNA_F
		HP Processing Method	BF102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144120.D	29 Oct 2025 11:00		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144121.D	29 Oct 2025 11:30		RC/JU	Ok
3	PB170202TB	PB170202TB	BF144122.D	29 Oct 2025 12:00		RC/JU	Ok
4	PB170259BL	PB170259BL	BF144123.D	29 Oct 2025 12:30		RC/JU	Ok
5	PB170259BS	PB170259BS	BF144124.D	29 Oct 2025 13:00		RC/JU	Ok,M
6	Q3460-17	WC-5	BF144125.D	29 Oct 2025 13:33		RC/JU	Ok,M
7	Q3419-04	PR132-S16-119133-20	BF144126.D	29 Oct 2025 14:03		RC/JU	Ok,M
8	Q3427-04MS	TP-8MS	BF144127.D	29 Oct 2025 14:33		RC/JU	Ok,M
9	Q3427-04MSD	TP-8MSD	BF144128.D	29 Oct 2025 15:02		RC/JU	Ok,M
10	Q3469-01	TP-13	BF144129.D	29 Oct 2025 15:32		RC/JU	Ok,M
11	Q3469-01MS	TP-13MS	BF144130.D	29 Oct 2025 16:02		RC/JU	Ok,M
12	Q3469-01MSD	TP-13MSD	BF144131.D	29 Oct 2025 16:32		RC/JU	Ok,M
13	Q3434-02	PR132-WC2-20251022	BF144132.D	29 Oct 2025 17:02		RC/JU	Ok
14	Q3480-01	RT3449	BF144133.D	29 Oct 2025 17:32		RC/JU	Ok
15	Q3476-01	NB-07-102825	BF144134.D	29 Oct 2025 18:01		RC/JU	Ok
16	Q3466-02	B8-0.0-1.0-20251023	BF144135.D	29 Oct 2025 18:31		RC/JU	Ok,M
17	Q3466-07	B7-1.0-2.0-20251023	BF144136.D	29 Oct 2025 19:01	Need straight Run	RC/JU	Not Ok
18	Q3481-01	OR-02-102825	BF144137.D	29 Oct 2025 19:30		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102925

Review By	rahul	Review On	10/30/2025 8:49:10 AM		
Supervise By	Jagrut	Supervise On	10/30/2025 4:42:33 PM		
SubDirectory	BF102925	HP Acquire Method	BNA_F	HP Processing Method	BF102425
STD. NAME	STD REF.#				
Tune/Reschk	SP6875				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13180,10ul/1000ul sample				
ICV/I.BLK	SP6872				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3472-04	COMP-1	BF144138.D	29 Oct 2025 20:00		RC/JU	Ok
20	Q3472-06	COMP-2	BF144139.D	29 Oct 2025 20:29	Internal standard fail	RC/JU	Ok
21	Q3472-08	COMP-3	BF144140.D	29 Oct 2025 20:59	Internal standard fail	RC/JU	Ok
22	Q3472-10	COMP-4	BF144141.D	29 Oct 2025 21:28	Surrogate Fail,	RC/JU	Ok,M
23	Q3472-12	COMP-5	BF144142.D	29 Oct 2025 21:57		RC/JU	Ok
24	Q3472-14	155	BF144143.D	29 Oct 2025 22:27	need further 10X	RC/JU	Dilution

M : Manual Integration