

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3419

Instrument ID: BNA_F

Calibration Date(s): 10/24/2025 10/24/2025

Calibration Time(s): 10:13 16:24

LAB FILE ID:		RRF2.5 = BF144056.D			RRF005 = BF144057.D		RRF010 = BF144058.D	
		RRF020 = BF144059.D			RRF040 = BF144060.D		RRF050 = BF144061.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.324	1.322	1.328	1.303	1.163	1.264	7.8
Benzaldehyde			1.093	0.965	1.043	0.959	1.084	15.3
Phenol-d6		1.548	1.443	1.481	1.408	1.289	1.396	8.6
Phenol		1.896	1.860	1.820	1.736	1.575	1.722	9.4
bis(2-Chloroethyl) ether		1.406	1.278	1.285	1.248	1.128	1.245	8.7
2-Chlorophenol		1.398	1.310	1.306	1.251	1.124	1.245	9.2
2-Methylphenol		1.052	1.010	1.032	0.931	0.874	0.957	8.9
2,2-oxybis(1-Chloropropane)		2.561	2.432	2.505	2.376	2.149	2.339	8.1
Acetophenone		0.473	0.459	0.433	0.437	0.397	0.432	8.4
3+4-Methylphenols			1.270	1.190	1.167	1.036	1.132	9.7
n-Nitroso-di-n-propylamine	0.900	0.971	0.972	0.949	0.876	0.801	0.903	7.6
Nitrobenzene-d5		0.377	0.378	0.374	0.368	0.335	0.365	6.5
Hexachloroethane		0.553	0.544	0.569	0.544	0.506	0.541	7.2
Nitrobenzene		0.395	0.384	0.404	0.392	0.360	0.386	6.0
Isophorone		0.665	0.662	0.645	0.643	0.584	0.638	6.3
2-Nitrophenol		0.190	0.184	0.196	0.191	0.167	0.184	7.0
2,4-Dimethylphenol		0.289	0.279	0.262	0.287	0.254	0.273	6.7
bis(2-Chloroethoxy) methane		0.430	0.408	0.413	0.404	0.365	0.398	7.2
2,4-Dichlorophenol		0.334	0.345	0.325	0.323	0.293	0.317	7.8
Naphthalene		1.054	1.002	0.962	0.965	0.852	0.944	9.7
4-Chloroaniline		0.332	0.356	0.349	0.336	0.298	0.329	8.1
Hexachlorobutadiene		0.279	0.276	0.274	0.270	0.241	0.267	8.2
Caprolactam			0.078	0.082	0.085	0.072	0.080	7.0
4-Chloro-3-methylphenol		0.302	0.289	0.285	0.282	0.257	0.280	7.0
2-Methylnaphthalene		0.718	0.654	0.631	0.644	0.549	0.622	10.6
Hexachlorocyclopentadiene			0.148	0.208	0.218	0.205	0.212	18.9
2,4,6-Trichlorophenol		0.430	0.447	0.480	0.473	0.427	0.447	8.0
2-Fluorobiphenyl		1.654	1.503	1.411	1.451	1.291	1.421	11.0
2,4,5-Trichlorophenol		0.488	0.485	0.484	0.467	0.427	0.469	7.2
1,1-Biphenyl		1.613	1.533	1.441	1.472	1.317	1.432	10.1
2-Chloronaphthalene		1.336	1.264	1.233	1.222	1.101	1.199	9.7
2-Nitroaniline		0.325	0.359	0.368	0.380	0.333	0.356	8.0
Dimethylphthalate		1.398	1.328	1.332	1.366	1.219	1.311	7.1
Acenaphthylene		1.780	1.693	1.664	1.703	1.476	1.617	9.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3419

Instrument ID: BNA_F

Calibration Date(s): 10/24/2025 10/24/2025

Calibration Time(s): 10:13 16:24

LAB FILE ID:		RRF2.5 = BF144056.D			RRF005 = BF144057.D		RRF010 = BF144058.D	
		RRF020 = BF144059.D			RRF040 = BF144060.D		RRF050 = BF144061.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.228	0.257	0.275	0.276	0.252	0.259	8.3
3-Nitroaniline		0.276	0.294	0.293	0.322	0.281	0.288	7.8
Acenaphthene		1.125	1.153	1.134	1.114	0.999	1.080	8.2
2,4-Dinitrophenol			0.161	0.162	0.124	0.116	0.140	13.4
4-Nitrophenol			0.224	0.199	0.187	0.175	0.195	8.9
Dibenzofuran		1.685	1.572	1.512	1.559	1.391	1.499	9.5
2,4-Dinitrotoluene		0.302	0.343	0.370	0.385	0.336	0.350	9.4
Diethylphthalate		1.286	1.264	1.244	1.290	1.145	1.233	7.0
4-Chlorophenyl-phenylether		0.706	0.667	0.656	0.675	0.591	0.643	9.5
Fluorene		1.359	1.169	1.133	1.188	1.041	1.143	11.3
4-Nitroaniline		0.262	0.266	0.275	0.296	0.260	0.273	6.7
4,6-Dinitro-2-methylphenol			0.133	0.151	0.123	0.115	0.128	10.6
n-Nitrosodiphenylamine		0.716	0.666	0.657	0.643	0.590	0.639	8.8
2,4,6-Tribromophenol		0.251	0.249	0.256	0.262	0.241	0.255	6.8
4-Bromophenyl-phenylether		0.274	0.256	0.247	0.260	0.230	0.250	8.5
Hexachlorobenzene		0.311	0.338	0.309	0.321	0.275	0.307	8.6
Atrazine		0.229	0.223	0.218	0.208	0.185	0.210	11.5
Pentachlorophenol			0.123	0.149	0.150	0.143	0.146	9.9
Phenanthrene		1.073	1.040	1.009	1.017	0.918	0.994	7.9
Anthracene		1.171	1.077	1.029	1.053	0.932	1.026	10.0
Carbazole		0.943	0.920	0.883	0.905	0.818	0.874	8.2
Di-n-butylphthalate		1.090	1.091	1.116	1.118	1.002	1.071	6.6
Fluoranthene		1.171	1.151	1.121	1.123	0.997	1.098	7.9
Pyrene		1.603	1.466	1.426	1.521	1.295	1.438	9.1
Terphenyl-d14		1.280	1.134	1.121	1.212	1.002	1.144	9.4
Butylbenzylphthalate		0.508	0.542	0.566	0.601	0.494	0.544	8.2
3,3-Dichlorobenzidine			0.469	0.485	0.510	0.404	0.470	10.2
Benzo (a) anthracene		1.341	1.322	1.359	1.516	1.206	1.352	8.8
Chrysene		1.330	1.263	1.253	1.319	1.179	1.268	5.9
Bis (2-ethylhexyl) phthalate		0.745	0.725	0.761	0.795	0.690	0.743	7.0
Di-n-octyl phthalate			1.228	1.295	1.399	1.179	1.277	7.6
Benzo (b) fluoranthene		1.134	1.108	1.153	1.264	1.055	1.131	6.4
Benzo (k) fluoranthene		1.206	1.077	1.010	1.059	0.906	1.043	11.4
Benzo (a) pyrene		1.049	0.990	1.036	1.089	0.942	1.014	6.9
Indeno (1,2,3-cd) pyrene		1.351	1.331	1.394	1.477	1.295	1.373	6.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3419

Instrument ID: BNA_F

Calibration Date(s): 10/24/2025 10/24/2025

Calibration Time(s): 10:13 16:24

LAB FILE ID:		RRF2.5 = BF144056.D			RRF005 = BF144057.D		RRF010 = BF144058.D	
		RRF020 = BF144059.D			RRF040 = BF144060.D		RRF050 = BF144061.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.058	1.061	1.107	1.166	1.024	1.087	6.7
Benzo (g,h,i) perylene		1.051	1.051	1.124	1.195	1.023	1.093	7.1
1,2,4,5-Tetrachlorobenzene		0.777	0.707	0.693	0.707	0.629	0.687	9.2
1,4-Dioxane		0.723	0.494	0.485	0.550	0.519	0.546	15.6
2,3,4,6-Tetrachlorophenol		0.317	0.318	0.333	0.360	0.322	0.330	7.5

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3419

Instrument ID: BNA_G

Calibration Date(s): 10/24/2025 10/24/2025

Calibration Time(s): 09:00 15:32

LAB FILE ID:		RRF2.5 = BG064569.D			RRF005 = BG064570.D		RRF010 = BG064571.D	
		RRF020 = BG064572.D			RRF040 = BG064573.D		RRF060 = BG064574.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
2-Fluorophenol		1.098	1.207	1.155	1.292	1.318	1.192	7.1
Benzaldehyde			0.942	0.994	0.940	0.981	0.893	14.1
Phenol-d6		1.469	1.654	1.566	1.771	1.864	1.635	8.6
Phenol		1.665	1.820	1.748	1.944	2.027	1.808	7.4
bis(2-Chloroethyl)ether		1.307	1.409	1.292	1.383	1.443	1.329	6.5
2-Chlorophenol		1.051	1.207	1.197	1.373	1.461	1.247	10.6
2-Methylphenol		1.103	1.225	1.178	1.273	1.345	1.201	7.2
2,2-oxybis(1-Chloropropane)		3.421	3.604	3.396	3.733	3.802	3.487	6.7
Acetophenone		0.535	0.565	0.528	0.584	0.590	0.547	6.0
3+4-Methylphenols			1.652	1.581	1.818	1.895	1.688	8.0
n-Nitroso-di-n-propylamine	1.044	1.038	1.141	1.129	1.227	1.289	1.130	7.9
Nitrobenzene-d5		0.239	0.286	0.314	0.377	0.403	0.333	17.0
Hexachloroethane		0.468	0.535	0.505	0.541	0.557	0.512	6.4
Nitrobenzene		0.274	0.312	0.344	0.410	0.438	0.362	15.6
Isophorone		0.751	0.796	0.771	0.869	0.891	0.799	7.2
2-Nitrophenol		0.076	0.095	0.110	0.137	0.152	0.121	23.2
2,4-Dimethylphenol		0.236	0.294	0.285	0.335	0.343	0.297	11.9
bis(2-Chloroethoxy)methane		0.452	0.464	0.443	0.490	0.493	0.455	6.5
2,4-Dichlorophenol		0.243	0.299	0.315	0.360	0.373	0.320	13.3
Naphthalene		1.083	1.144	1.084	1.177	1.201	1.109	6.0
4-Chloroaniline		0.390	0.456	0.411	0.459	0.486	0.431	8.3
Hexachlorobutadiene		0.266	0.291	0.282	0.299	0.306	0.284	5.3
Caprolactam			0.112	0.114	0.134	0.142	0.123	9.9
4-Chloro-3-methylphenol		0.323	0.381	0.361	0.426	0.454	0.384	11.2
2-Methylnaphthalene		0.764	0.858	0.797	0.865	0.888	0.814	6.7
Hexachlorocyclopentadiene			0.268	0.279	0.324	0.340	0.302	9.0
2,4,6-Trichlorophenol		0.294	0.333	0.375	0.423	0.456	0.384	14.4
2-Fluorobiphenyl		1.310	1.389	1.306	1.384	1.384	1.319	5.2
2,4,5-Trichlorophenol		0.339	0.425	0.439	0.484	0.517	0.443	12.5
1,1-Biphenyl		1.456	1.450	1.401	1.489	1.498	1.422	5.0
2-Chloronaphthalene		1.065	1.082	1.053	1.132	1.145	1.073	4.7
2-Nitroaniline		0.201	0.234	0.277	0.348	0.404	0.311	24.1
Dimethylphthalate		1.389	1.570	1.488	1.641	1.699	1.530	7.2
Acenaphthylene		1.667	1.726	1.647	1.786	1.827	1.693	5.3

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3419

Instrument ID: BNA_G

Calibration Date(s): 10/24/2025 10/24/2025

Calibration Time(s): 09:00 15:32

LAB FILE ID:		RRF2.5 = BG064569.D			RRF005 = BG064570.D		RRF010 = BG064571.D	
		RRF020 = BG064572.D			RRF040 = BG064573.D		RRF060 = BG064574.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
2,6-Dinitrotoluene		0.127	0.180	0.193	0.247	0.283	0.218	24.7
3-Nitroaniline		0.180	0.211	0.252	0.294	0.325	0.263	19.7
Acenaphthene		1.126	1.178	1.115	1.223	1.261	1.157	5.6
2,4-Dinitrophenol			0.086	0.110	0.137	0.165	0.132	22.1
4-Nitrophenol			0.191	0.213	0.261	0.283	0.244	14.4
Dibenzofuran		1.881	1.972	1.826	1.963	1.988	1.870	6.0
2,4-Dinitrotoluene		0.223	0.269	0.300	0.385	0.434	0.340	22.7
Diethylphthalate		1.417	1.645	1.514	1.657	1.718	1.563	7.1
4-Chlorophenyl-phenylether		0.800	0.864	0.792	0.866	0.888	0.820	6.4
Fluorene		1.471	1.559	1.426	1.531	1.549	1.460	6.4
4-Nitroaniline		0.194	0.252	0.286	0.327	0.355	0.293	18.6
4,6-Dinitro-2-methylphenol			0.055	0.064	0.083	0.097	0.079	21.0
n-Nitrosodiphenylamine		0.510	0.551	0.525	0.583	0.594	0.540	6.8
2,4,6-Tribromophenol		0.221	0.266	0.272	0.318	0.346	0.289	14.0
4-Bromophenyl-phenylether		0.218	0.236	0.230	0.258	0.264	0.237	7.3
Hexachlorobenzene		0.254	0.266	0.262	0.292	0.302	0.270	7.0
Atrazine		0.215	0.227	0.249	0.245	0.261	0.232	8.3
Pentachlorophenol			0.132	0.152	0.180	0.199	0.169	13.9
Phenanthrene		1.084	1.133	1.073	1.162	1.179	1.094	6.2
Anthracene		1.087	1.144	1.101	1.187	1.204	1.113	6.2
Carbazole		0.982	1.037	0.982	1.031	1.028	0.981	5.9
Di-n-butylphthalate		1.011	1.156	1.179	1.240	1.198	1.135	7.1
Fluoranthene		1.385	1.463	1.380	1.440	1.403	1.370	5.9
Pyrene		1.287	1.331	1.284	1.411	1.406	1.307	6.2
Terphenyl-d14		1.074	1.120	1.066	1.142	1.102	1.060	7.2
Butylbenzylphthalate		0.284	0.359	0.388	0.450	0.467	0.395	15.4
3,3-Dichlorobenzidine			0.418	0.428	0.477	0.486	0.443	6.9
Benzo (a) anthracene		1.294	1.355	1.292	1.416	1.430	1.326	5.7
Chrysene		1.249	1.291	1.229	1.344	1.359	1.262	5.8
Bis (2-ethylhexyl) phthalate		0.484	0.555	0.581	0.671	0.667	0.590	11.0
Di-n-octyl phthalate			0.871	0.932	1.087	1.117	0.995	9.3
Benzo (b) fluoranthene		1.138	1.246	1.169	1.322	1.353	1.226	6.8
Benzo (k) fluoranthene		1.175	1.260	1.206	1.285	1.364	1.227	6.5
Benzo (a) pyrene		1.061	1.144	1.093	1.202	1.246	1.127	6.4
Indeno (1,2,3-cd) pyrene		1.383	1.491	1.410	1.568	1.642	1.471	6.8

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q3419

Instrument ID: BNA_G

Calibration Date(s): 10/24/2025 10/24/2025

Calibration Time(s): 09:00 15:32

LAB FILE ID:		RRF2.5 = BG064569.D		RRF005 = BG064570.D		RRF010 = BG064571.D		
		RRF020 = BG064572.D		RRF040 = BG064573.D		RRF060 = BG064574.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
Dibenzo(a,h)anthracene		1.120	1.223	1.158	1.268	1.324	1.195	6.5
Benzo(g,h,i)perylene		1.080	1.162	1.089	1.213	1.280	1.142	6.9
1,2,4,5-Tetrachlorobenzene		0.657	0.684	0.648	0.702	0.717	0.671	4.6
1,4-Dioxane		0.570	0.581	0.524	0.525	0.559	0.528	8.8
2,3,4,6-Tetrachlorophenol		0.312	0.385	0.402	0.466	0.508	0.424	15.0

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1