

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
 Data File : BG064618.D
 Acq On : 29 Oct 2025 11:22
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 12:00:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 16:40:26 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	42.495	-6.2	97	0.00
3	Pyridine	40.000	40.966	-2.4	89	0.00
4	n-Nitrosodimethylamine	40.000	40.532	-1.3	87	0.00
5 S	2-Fluorophenol	80.000	84.991	-6.2	89	0.00
6	Aniline	40.000	40.860	-2.1	87	0.00
7 S	Phenol-d6	80.000	81.009	-1.3	85	0.00
8	2-Chlorophenol	40.000	44.170	-10.4	91	0.00
9	Benzaldehyde	40.000	37.558	6.1	81	0.00
10 C	Phenol	40.000	40.664	-1.7	86	0.00
11	bis(2-Chloroethyl)ether	40.000	40.860	-2.1	87	0.00
12	1,3-Dichlorobenzene	40.000	43.073	-7.7	93	0.00
13 C	1,4-Dichlorobenzene	40.000	42.552	-6.4	92	0.00
14	1,2-Dichlorobenzene	40.000	43.105	-7.8	93	0.00
15	Benzyl Alcohol	40.000	42.201	-5.5	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.138	-5.3	90	0.00
17	2-Methylphenol	40.000	42.752	-6.9	88	0.00
18	Hexachloroethane	40.000	44.282	-10.7	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.275	-8.2	91	0.00
20	3+4-Methylphenols	40.000	41.279	-3.2	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	42.523	-6.3	90	0.00
23 S	Nitrobenzene-d5	80.000	96.333	-20.4	96	0.00
24	Nitrobenzene	40.000	47.355	-18.4	95	0.00
25	Isophorone	40.000	41.695	-4.2	87	0.00
26 C	2-Nitrophenol	40.000	47.979	-19.9	111	0.00
27	2,4-Dimethylphenol	40.000	41.002	-2.5	84	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.170	-5.4	89	0.00
29 C	2,4-Dichlorophenol	40.000	44.058	-10.1	88	0.00
30	1,2,4-Trichlorobenzene	40.000	43.927	-9.8	92	0.00
31	Naphthalene	40.000	41.819	-4.5	89	0.00
32	Benzoic acid	40.000	38.781	3.0	90	-0.03
33	4-Chloroaniline	40.000	40.373	-0.9	86	0.00
34 C	Hexachlorobutadiene	40.000	45.306	-13.3	97	0.00
35	Caprolactam	40.000	35.362	11.6	73	-0.03
36 C	4-Chloro-3-methylphenol	40.000	40.462	-1.2	83	0.00
37	2-Methylnaphthalene	40.000	41.085	-2.7	87	0.00
38	1-Methylnaphthalene	40.000	41.190	-3.0	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.962	-12.4	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	52.985	-32.5#	103	0.00
42 S	2,4,6-Tribromophenol	80.000	86.403	-8.0	82	0.00
43 C	2,4,6-Trichlorophenol	40.000	47.391	-18.5	90	0.00
44	2,4,5-Trichlorophenol	40.000	44.415	-11.0	85	0.00
45 S	2-Fluorobiphenyl	80.000	89.318	-11.6	89	0.00
46	1,1'-Biphenyl	40.000	43.180	-7.9	86	0.00
47	2-Chloronaphthalene	40.000	43.804	-9.5	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
 Data File : BG064618.D
 Acq On : 29 Oct 2025 11:22
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 12:00:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 24 16:40:26 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.351	-5.9	90	0.00
49	Acenaphthylene	40.000	41.830	-4.6	83	0.00
50	Dimethylphthalate	40.000	40.572	-1.4	79	0.00
51	2,6-Dinitrotoluene	40.000	43.351	-8.4	92	0.00
52 C	Acenaphthene	40.000	42.377	-5.9	84	0.00
53	3-Nitroaniline	40.000	43.773	-9.4	82	0.00
54 P	2,4-Dinitrophenol	40.000	44.922	-12.3	100	0.00
55	Dibenzofuran	40.000	41.281	-3.2	82	0.00
56 P	4-Nitrophenol	40.000	39.540	1.2	77	0.00
57	2,4-Dinitrotoluene	40.000	41.073	-2.7	86	0.00
58	Fluorene	40.000	40.432	-1.1	80	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.298	-10.7	84	0.00
60	Diethylphthalate	40.000	38.582	3.5	76	0.00
61	4-Chlorophenyl-phenylether	40.000	40.979	-2.4	81	0.00
62	4-Nitroaniline	40.000	41.786	-4.5	78	0.00
63	Azobenzene	40.000	39.183	2.0	76	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	77	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.448	-18.6	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.085	-7.7	77	0.00
67	4-Bromophenyl-phenylether	40.000	44.163	-10.4	78	0.00
68	Hexachlorobenzene	40.000	44.415	-11.0	79	0.00
69	Atrazine	40.000	40.031	-0.1	73	0.00
70 C	Pentachlorophenol	40.000	44.939	-12.3	81	0.00
71	Phenanthrene	40.000	41.397	-3.5	75	0.00
72	Anthracene	40.000	40.436	-1.1	73	0.00
73	Carbazole	40.000	39.556	1.1	73	0.00
74	Di-n-butylphthalate	40.000	40.184	-0.5	71	0.00
75 C	Fluoranthene	40.000	39.193	2.0	72	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	78	0.00
77	Benzidine	40.000	45.404	-13.5	81	0.00
78	Pyrene	40.000	40.467	-1.2	73	0.00
79 S	Terphenyl-d14	80.000	81.066	-1.3	74	0.00
80	Butylbenzylphthalate	40.000	46.801	-17.0	80	0.00
81	Benzo(a)anthracene	40.000	42.093	-5.2	77	0.00
82	3,3'-Dichlorobenzidine	40.000	43.737	-9.3	80	0.00
83	Chrysene	40.000	41.154	-2.9	76	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.515	-13.8	78	0.00
85 c	Di-n-octyl phthalate	40.000	48.350	-20.9#	87	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	85	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.308	-3.3	82	-0.02
88	Benzo(b)fluoranthene	40.000	40.302	-0.8	79	0.00
89	Benzo(k)fluoranthene	40.000	40.260	-0.6	82	0.00
90 C	Benzo(a)pyrene	40.000	41.067	-2.7	82	-0.02
91	Dibenzo(a,h)anthracene	40.000	41.529	-3.8	83	-0.03
92	Benzo(g,h,i)perylene	40.000	41.308	-3.3	82	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102925\
Data File : BG064618.D
Acq On : 29 Oct 2025 11:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 12:00:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 24 16:40:26 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1