

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3419

Client ID: SSTDCCC040

Date Analyzed: 10/28/2025

Lab File ID: BG064600.D

Time Analyzed: 13:45

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	216707	17.573	273944	21.85	304830	25.187
UPPER LIMIT	433414	18.073	547888	22.35	609660	25.687
LOWER LIMIT	108354	17.073	136972	21.35	152415	24.687
EPA SAMPLE NO.						
PR132-S16-094119-20251021	222776	17.58	285835	21.85	312112	25.19

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.