

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3419

Client ID: SSTDCCC040

Date Analyzed: 10/29/2025

Lab File ID: BG064618.D

Time Analyzed: 11:22

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	241683	17.577	261016	21.854	298207	25.197
UPPER LIMIT	483366	18.077	522032	22.354	596414	25.697
LOWER LIMIT	120842	17.077	130508	21.354	149104	24.697
EPA SAMPLE NO.						
01 PB170215BL	248117	17.58	318243	21.85	369003	25.19
02 PB170215BS	252758	17.58	286271	21.85	311364	25.19
03 PR-132-S16-015094-20251021	232719	17.58	307543	21.85	344921	25.19
04 PR-132-S16-015094-20251021	229619	17.58	271892	21.85	296767	25.19
05 PR-132-S16-015094-20251021	230638	17.58	282250	21.85	308600	25.19

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.