


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023558.D
Acq On : 24 Oct 2025 07:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260
Last Update : Wed Oct 08 05:12:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	28941	PASS
75	95	30	60	47.4	87403	PASS
95	95	100	100	100.0	184491	PASS
96	95	5	9	6.5	11916	PASS
173	174	0.00	2	1.1	1956	PASS
174	95	50	100	98.2	181248	PASS
175	174	5	9	7.4	13455	PASS
176	174	95	101	98.0	177600	PASS
177	176	5	9	6.6	11721	PASS