

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3419
 Lab File ID: VY023543.D Date Analyzed: 10/22/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	557873	7.71	806690	8.62	720008	11.42
UPPER LIMIT	1115750	8.213	1613380	9.116	1440020	11.92
LOWER LIMIT	278937	7.213	403345	8.116	360004	10.92
EPA SAMPLE NO.						
PR-132-S16-015094-20251021	785988	7.71	1180613	8.62	922887	11.42
PR132-S16-119133-20251021	900343	7.71	1446852	8.62	1218442	11.42
PR132-S16-094119-20251021	942425	7.71	1514027	8.62	1286219	11.41
VY1022SBL01	890059	7.71	1434768	8.62	1180185	11.42
VY1022SBS01	546555	7.71	794303	8.62	700136	11.41
VY1022SBSD01	542616	7.71	794200	8.62	691787	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.