

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3419
 Lab File ID: VY023559.D Date Analyzed: 10/24/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	380279	13.346				
UPPER LIMIT	760558	13.846				
LOWER LIMIT	190140	12.846				
EPA SAMPLE NO.						
PR-132-S16-015094-20251021MS	80067 *	13.35				
PR-132-S16-015094-20251021MSI	48940 *	13.35				
VY1024SBL01	469443	13.35				
VY1024SBS01	380713	13.35				
VY1024SBSD01	360721	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.