

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3419

Lab File ID: VY023383.D

BFB Injection Date: 10/07/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:43

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.9
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.9 (0.9) 1
174	50.0 - 100.0% of mass 95	99.1
175	5.0 - 9.0% of mass 174	7.3 (7.4) 1
176	95.0 - 101.0% of mass 174	95.4 (96.2) 1
177	5.0 - 9.0% of mass 176	6.3 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY023384.D	10/07/2025	09:17
VSTDIC010	VSTDIC010	VY023385.D	10/07/2025	10:02
VSTDIC020	VSTDIC020	VY023386.D	10/07/2025	11:24
VSTDIC050	VSTDIC050	VY023387.D	10/07/2025	11:47
VSTDIC100	VSTDIC100	VY023388.D	10/07/2025	12:09
VSTDIC150	VSTDIC150	VY023389.D	10/07/2025	12:32