

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3419

Lab File ID: VY023558.D

BFB Injection Date: 10/24/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 07:47

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.7
75	30.0 - 60.0% of mass 95	47.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.1 (1.1) 1
174	50.0 - 100.0% of mass 95	98.2
175	5.0 - 9.0% of mass 174	7.3 (7.4) 1
176	95.0 - 101.0% of mass 174	96.3 (98) 1
177	5.0 - 9.0% of mass 176	6.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023559.D	10/24/2025	08:56
VY1024SBL01	VY1024SBL01	VY023560.D	10/24/2025	09:24
VY1024SBS01	VY1024SBS01	VY023561.D	10/24/2025	09:52
VY1024SBSD01	VY1024SBSD01	VY023562.D	10/24/2025	10:14
PR-132-S16-015094-20251021MS	Q3419-02MS	VY023580.D	10/24/2025	17:58
PR-132-S16-015094-20251021MSD	Q3419-03MSD	VY023581.D	10/24/2025	18:20