

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102225

Review By	John Carlone	Review On	10/24/2025 2:07:41 PM
Supervise By	Maresh Dadoda	Supervise On	10/24/2025 2:54:27 PM
SubDirectory	VY102225	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135879
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135880,VP135881 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023542.D	22 Oct 2025 09:00		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023543.D	22 Oct 2025 09:30		SY/MD	Ok,M
3	VY1022SBL01	VY1022SBL01	VY023544.D	22 Oct 2025 10:00		SY/MD	Ok
4	VY1022SBS01	VY1022SBS01	VY023545.D	22 Oct 2025 10:28		SY/MD	Ok,M
5	VY1022SBSD01	VY1022SBSD01	VY023546.D	22 Oct 2025 10:51		SY/MD	Ok,M
6	Q3395-04MSD	PR132-S07-000102-20	VY023547.D	22 Oct 2025 11:28	vial-B Internal Standard Fail	SY/MD	Not Ok
7	Q3416-01	NB-08-10212025	VY023548.D	22 Oct 2025 12:03	vial-A	SY/MD	Ok
8	Q3404-01RE	MOO-25-0285-0288RE	VY023549.D	22 Oct 2025 12:52	vial-B Internal standard fail	SY/MD	Confirms
9	Q3427-03	TP-8-VOC	VY023550.D	22 Oct 2025 13:43	vial-A	SY/MD	Ok
10	Q3419-01	PR-132-S16-015094-20	VY023551.D	22 Oct 2025 14:07	vial-A	SY/MD	Ok
11	Q3419-04	PR132-S16-119133-20	VY023552.D	22 Oct 2025 15:34	vial-A	SY/MD	Ok
12	Q3419-05	PR132-S16-094119-20	VY023553.D	22 Oct 2025 15:58	vial-A	SY/MD	Ok
13	Q3422-01	PR132-S15-000008-20	VY023554.D	22 Oct 2025 16:22	vial-A	SY/MD	Ok
14	Q3424-01	PR132-DP3-20251021	VY023555.D	22 Oct 2025 16:46	vial-A	SY/MD	Ok
15	Q3424-02	PR132-S11-010080-20	VY023556.D	22 Oct 2025 17:10	vial-A	SY/MD	Ok
16	Q3424-03	PR132-S14-000094-20	VY023557.D	22 Oct 2025 17:35	vial-A	SY/MD	Ok

M : Manual Integration