

LAB CHRONICLE

OrderID:	Q3420	OrderDate:	10/22/2025 11:07:00 AM
Client:	Dal-Tile	Project:	Waste Water
Contact:	Michel Gil	Location:	J22

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3420-01	Oil and Grease #1	WATER			10/21/25 13:36			10/22/25
			Oil and Grease	1664A			10/27/25 10:00	
Q3420-02	Oil and Grease #2	WATER			10/21/25 14:53			10/22/25
			Oil and Grease	1664A			10/27/25 10:00	
Q3420-03	Oil and Grease #3	WATER			10/21/25 14:59			10/22/25
			Oil and Grease	1664A			10/27/25 10:00	
Q3420-04	Cyanide	WATER			10/21/25 15:04			10/22/25
			Cyanide	SM4500-CN C,E		10/24/25	10/24/25 11:11	
Q3420-05	COMPOSITE	WATER			10/21/25 15:16			10/22/25
			Ammonia	SM4500-NH3		10/22/25	10/22/25 14:56	
			BOD5	SM5210 B			10/23/25 15:00	
			Hexavalent Chromium	7196A			10/22/25 13:07	
			Phosphorus-Total	365.3		10/31/25	10/31/25 13:36	
			TSS	SM2540 D			10/24/25 12:30	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client: Dal-Tile
Project: Waste Water
Client Sample ID: Oil and Grease #1
Lab Sample ID: Q3420-01

Date Collected: 10/21/25 13:36
Date Received: 10/22/25
SDG No.: Q3420
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	3.20	J	1	0.29	5.00	mg/L		10/27/25 10:00	1664A

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Dal-Tile
Project: Waste Water
Client Sample ID: Oil and Grease #2
Lab Sample ID: Q3420-02

Date Collected: 10/21/25 14:53
Date Received: 10/22/25
SDG No.: Q3420
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	4.00	J	1	0.29	5.00	mg/L		10/27/25 10:00	1664A

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client: Dal-Tile
Project: Waste Water
Client Sample ID: Oil and Grease #3
Lab Sample ID: Q3420-03

Date Collected: 10/21/25 14:59
Date Received: 10/22/25
SDG No.: Q3420
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	1.20	J	1	0.29	5.00	mg/L		10/27/25 10:00	1664A

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client: Dal-Tile
Project: Waste Water
Client Sample ID: Cyanide
Lab Sample ID: Q3420-04

Date Collected: 10/21/25 15:04
Date Received: 10/22/25
SDG No.: Q3420
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Cyanide	0.19		1	0.0012	0.0050	mg/L	10/24/25 08:10	10/24/25 11:11	SM 4500-CN C-21 plus E-21

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Dal-Tile
Project: Waste Water
Client Sample ID: COMPOSITE
Lab Sample ID: Q3420-05

Date Collected: 10/21/25 15:16
Date Received: 10/22/25
SDG No.: Q3420
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	0.45		1	0.030	0.10	mg/L	10/22/25 12:45	10/22/25 14:56	SM 4500-NH3 B plus G-21
BOD5	230		1	0.20	2.00	mg/L		10/23/25 15:00	SM 5210 B-16
Dissolved Hexavalent Chromium	0.0030	U	1	0.0030	0.010	mg/L		10/22/25 13:07	7196A
Phosphorus, Total	0.18		1	0.0050	0.050	mg/L	10/31/25 10:00	10/31/25 13:36	365.3
TSS	14.1		1	1.00	4.00	mg/L		10/24/25 12:30	SM 2540 D-20

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
RunNo.: LB137615

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	10/22/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	10/22/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	10/22/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3420

Project: Waste Water

RunNo.: LB137617

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.97	1	97	90-110	10/22/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	10/22/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	10/22/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.96	1	96	90-110	10/22/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	10/22/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.98	1	98	90-110	10/22/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3420

Project: Waste Water

RunNo.: LB137634

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.094	0.099	95	85-115	10/24/2025
Sample ID:	CCV1						
Cyanide		mg/L	0.24	0.25	96	90-110	10/24/2025
Sample ID:	CCV2						
Cyanide		mg/L	0.24	0.25	96	90-110	10/24/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile

SDG No.: Q3420

Project: Waste Water

RunNo.: LB137729

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.517	0.50	103	90-110	10/31/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.502	0.50	100	90-110	10/31/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.493	0.50	99	90-110	10/31/2025

Initial and Continuing Calibration Blank Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
RunNo.: LB137615

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/22/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/22/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/22/2025

Initial and Continuing Calibration Blank Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
RunNo.: LB137617

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	10/22/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	10/22/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	10/22/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	10/22/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	10/22/2025
Sample ID: CCB5 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	10/22/2025

Initial and Continuing Calibration Blank Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
RunNo.: LB137634

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	10/24/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	10/24/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	10/24/2025

Initial and Continuing Calibration Blank Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
RunNo.: LB137729

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.007	0.0250	J	0.0045	0.05	10/31/2025
Sample ID: CCB1 Phosphorus, Total	mg/L	0.010	0.0250	J	0.0045	0.05	10/31/2025
Sample ID: CCB2 Phosphorus, Total	mg/L	0.007	0.0250	J	0.0045	0.05	10/31/2025

Preparation Blank Summary

Client: Dal-Tile

SDG No.: Q3420

Project: Waste Water

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137615BL Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.003	0.01	10/22/2025
Sample ID: LB137623BL BOD5	mg/L	< 0.2000	0.2000	U	0.20	2.0	10/23/2025
Sample ID: LB137648BL TSS	mg/L	1	2.0000	J	1	4	10/24/2025
Sample ID: LB137655BL Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	10/27/2025
Sample ID: PB170208BL Ammonia as N	mg/L	< 0.0500	0.0500	U	0.03	0.1	10/22/2025
Sample ID: PB170237BL Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	10/24/2025
Sample ID: PB170350BL Phosphorus, Total	mg/L	0.01	0.0250	J	0.005	0.05	10/31/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.92		0.0030	U	1.0	2	92		10/22/2025
Ammonia as N	mg/L	75-125	1.50		0.45		1	1	105		10/22/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.93		0.0030	U	1.0	2	93		10/22/2025
Ammonia as N	mg/L	75-125	1.50		0.45		1	1	105		10/22/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3442-01
Client ID:	SW-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.50		0.025	J	0.5	1	94		10/31/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3442-01
Client ID:	SW-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.50		0.025	J	0.5	1	96		10/31/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.040		0.0035	J	0.04	1	91		10/24/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.040		0.0035	J	0.04	1	91		10/24/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.4		2.30	J	20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.6		2.30	J	20.0	1	102		10/27/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3455-01
Client ID:	Q3455 -01MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3455-01
Client ID:	Q3455 -01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	102		10/27/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		10/22/2025
Ammonia as N	mg/L	+/-20	0.45		0.45		1	0		10/22/2025
BOD5	mg/L	+/-20	230		223		1	3.18		10/23/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.92		0.93		2	0.22		10/22/2025
Ammonia as N	mg/L	+/-20	1.50		1.50		1	0		10/22/2025

Duplicate Sample Summary

Client: Dal-Tile Project: Waste Water Client ID: SW-1DUP	SDG No.: Q3420 Sample ID: Q3442-01 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.025	J	0.024	J	1	4.08		10/31/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3442-01
Client ID:	SW-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.50		0.50		1	1.2		10/31/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-01
Client ID:	DSN001DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	27.5		27.1		1	1.47		10/24/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0035	J	0.0034	J	1	3		10/24/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.040		0.040		1	0		10/24/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	22.4		22.6		1	0.89		10/27/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3455-01
Client ID:	Q3455 -01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	287		287		1	0.03		10/27/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137615

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137615BS							
Hexavalent Chromium	mg/L	0.5	0.52		103	1	90-111	10/22/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137623

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137623BS							
BOD5	mg/L	198	186		94	1	84.6-115.4	10/23/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137648

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137648BS							
TSS	mg/L	550	531		96	1	90-110	10/24/2025

Laboratory Control Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Run No.:	LB137655

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137655BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	10/27/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137617

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170208BS							
Ammonia as N	mg/L	1	0.98		98	1	90-110	10/22/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137634

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170237BS							
Cyanide	mg/L	0.1	0.096		96	1	85-115	10/24/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137729

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170350BS							
Phosphorus, Total	mg/L	0.50	0.52		103	1	90-110	10/31/2025