

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITES	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.92		0.0030	U	1.0	2	92		10/22/2025
Ammonia as N	mg/L	75-125	1.50		0.45		1	1	105		10/22/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.93		0.0030	U	1.0	2	93		10/22/2025
Ammonia as N	mg/L	75-125	1.50		0.45		1	1	105		10/22/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3442-01
Client ID:	SW-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.50		0.025	J	0.5	1	94		10/31/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3442-01
Client ID:	SW-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.50		0.025	J	0.5	1	96		10/31/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.040		0.0035	J	0.04	1	91		10/24/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.040		0.0035	J	0.04	1	91		10/24/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.4		2.30	J	20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.6		2.30	J	20.0	1	102		10/27/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3455-01
Client ID:	Q3455 -01MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3455-01
Client ID:	Q3455 -01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	102		10/27/2025