

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3420-05
Client ID:	COMPOSITEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		10/22/2025
Ammonia as N	mg/L	+/-20	0.45		0.45		1	0		10/22/2025
BOD5	mg/L	+/-20	230		223		1	3.18		10/23/2025

Duplicate Sample Summary

Client: Dal-Tile	SDG No.: Q3420
Project: Waste Water	Sample ID: Q3420-05
Client ID: COMPOSITMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.92		0.93		2	0.22		10/22/2025
Ammonia as N	mg/L	+/-20	1.50		1.50		1	0		10/22/2025

Duplicate Sample Summary

Client: Dal-Tile	SDG No.: Q3420
Project: Waste Water	Sample ID: Q3442-01
Client ID: SW-1DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.025	J	0.024	J	1	4.08		10/31/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3442-01
Client ID:	SW-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/L	+/-20	0.50		0.50		1	1.2		10/31/2025

Duplicate Sample Summary

Client: Dal-Tile Project: Waste Water Client ID: DSN001DUP	SDG No.: Q3420 Sample ID: Q3448-01 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	27.5		27.1		1	1.47		10/24/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0035	J	0.0034	J	1	3		10/24/2025

Duplicate Sample Summary

Client:	Dal-Tile	SDG No.:	Q3420
Project:	Waste Water	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.040		0.040		1	0		10/24/2025

Duplicate Sample Summary

Client: Dal-Tile	SDG No.: Q3420
Project: Waste Water	Sample ID: Q3448-03
Client ID: DSN001MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	22.4		22.6		1	0.89		10/27/2025

Duplicate Sample Summary

Client: Dal-Tile Project: Waste Water Client ID: Q3455 -01MSD	SDG No.: Q3420 Sample ID: Q3455-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	287		287		1	0.03		10/27/2025