

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137615

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137615BS							
Hexavalent Chromium	mg/L	0.5	0.52		103	1	90-111	10/22/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137623

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137623BS							
BOD5	mg/L	198	186		94	1	84.6-115.4	10/23/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137648

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137648BS							
TSS	mg/L	550	531		96	1	90-110	10/24/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137655

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137655BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	10/27/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137617

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170208BS							
Ammonia as N	mg/L	1	0.98		98	1	90-110	10/22/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137634

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170237BS							
Cyanide	mg/L	0.1	0.096		96	1	85-115	10/24/2025

Laboratory Control Sample Summary

Client: Dal-Tile
Project: Waste Water

SDG No.: Q3420
Run No.: LB137729

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170350BS							
Phosphorus, Total	mg/L	0.50	0.52		103	1	90-110	10/31/2025