



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB86	Mercury	0.076	+/-0.2	U	0.20	CV	10/24/2025	09:23	LB137632

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB83	Mercury	0.076	+/-0.2	U	0.20	CV	10/24/2025	09:30	LB137632
CCB84	Mercury	0.076	+/-0.2	U	0.20	CV	10/24/2025	10:11	LB137632
CCB85	Mercury	0.076	+/-0.2	U	0.20	CV	10/24/2025	10:43	LB137632
CCB86	Mercury	0.076	+/-0.2	U	0.20	CV	10/24/2025	11:14	LB137632
CCB87	Mercury	0.076	+/-0.2	U	0.20	CV	10/24/2025	11:38	LB137632

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/24/2025	14:43	LB137662
	Cadmium	0.50	+/-3	U	6.00	P	10/24/2025	14:43	LB137662
	Chromium	2.12	+/-5	U	10.0	P	10/24/2025	14:43	LB137662
	Copper	4.60	+/-10	U	20.0	P	10/24/2025	14:43	LB137662
	Lead	2.30	+/-6	U	12.0	P	10/24/2025	14:43	LB137662
	Nickel	3.06	+/-20	U	40.0	P	10/24/2025	14:43	LB137662
	Selenium	9.64	+/-10	U	20.0	P	10/24/2025	14:43	LB137662
	Zinc	16.7	+/-20	U	40.0	P	10/24/2025	14:43	LB137662

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/24/2025	15:16	LB137662
	Cadmium	0.50	+/-3	U	6.00	P	10/24/2025	15:16	LB137662
	Chromium	2.12	+/-5	U	10.0	P	10/24/2025	15:16	LB137662
	Copper	4.60	+/-10	U	20.0	P	10/24/2025	15:16	LB137662
	Lead	2.30	+/-6	U	12.0	P	10/24/2025	15:16	LB137662
	Nickel	3.06	+/-20	U	40.0	P	10/24/2025	15:16	LB137662
	Selenium	9.64	+/-10	U	20.0	P	10/24/2025	15:16	LB137662
	Zinc	16.7	+/-20	U	40.0	P	10/24/2025	15:16	LB137662
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/24/2025	16:47	LB137662
	Cadmium	0.50	+/-3	U	6.00	P	10/24/2025	16:47	LB137662
	Chromium	2.12	+/-5	U	10.0	P	10/24/2025	16:47	LB137662
	Copper	4.60	+/-10	U	20.0	P	10/24/2025	16:47	LB137662
	Lead	2.30	+/-6	U	12.0	P	10/24/2025	16:47	LB137662
	Nickel	3.06	+/-20	U	40.0	P	10/24/2025	16:47	LB137662
	Selenium	9.64	+/-10	U	20.0	P	10/24/2025	16:47	LB137662
	Zinc	16.7	+/-20	U	40.0	P	10/24/2025	16:47	LB137662
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/24/2025	18:33	LB137662
	Cadmium	0.50	+/-3	U	6.00	P	10/24/2025	18:33	LB137662
	Chromium	2.12	+/-5	U	10.0	P	10/24/2025	18:33	LB137662
	Copper	4.60	+/-10	U	20.0	P	10/24/2025	18:33	LB137662
	Lead	2.30	+/-6	U	12.0	P	10/24/2025	18:33	LB137662
	Nickel	3.06	+/-20	U	40.0	P	10/24/2025	18:33	LB137662
	Selenium	9.64	+/-10	U	20.0	P	10/24/2025	18:33	LB137662
	Zinc	16.7	+/-20	U	40.0	P	10/24/2025	18:33	LB137662
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/24/2025	19:23	LB137662
	Cadmium	0.50	+/-3	U	6.00	P	10/24/2025	19:23	LB137662
	Chromium	2.12	+/-5	U	10.0	P	10/24/2025	19:23	LB137662
	Copper	4.60	+/-10	U	20.0	P	10/24/2025	19:23	LB137662
	Lead	2.30	+/-6	U	12.0	P	10/24/2025	19:23	LB137662
	Nickel	3.06	+/-20	U	40.0	P	10/24/2025	19:23	LB137662
	Selenium	9.64	+/-10	U	20.0	P	10/24/2025	19:23	LB137662
	Zinc	16.7	+/-20	U	40.0	P	10/24/2025	19:23	LB137662
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/24/2025	20:10	LB137662
	Cadmium	0.50	+/-3	U	6.00	P	10/24/2025	20:10	LB137662
	Chromium	2.12	+/-5	U	10.0	P	10/24/2025	20:10	LB137662
	Copper	4.82	+/-10	J	20.0	P	10/24/2025	20:10	LB137662
	Lead	2.30	+/-6	U	12.0	P	10/24/2025	20:10	LB137662
	Nickel	3.06	+/-20	U	40.0	P	10/24/2025	20:10	LB137662
	Selenium	9.64	+/-10	U	20.0	P	10/24/2025	20:10	LB137662

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Zinc	16.7	+/-20	U	40.0	P	10/24/2025	20:10	LB137662

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	14:12	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	14:12	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	14:12	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	14:12	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	14:12	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	14:41	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	14:41	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	14:41	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	14:41	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	14:41	LB137697
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	15:48	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	15:48	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	15:48	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	15:48	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	15:48	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	15:48	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	15:48	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	15:48	LB137697
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	16:41	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	16:41	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	16:41	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	16:41	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	16:41	LB137697
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	17:31	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	17:31	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	17:31	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	17:31	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	17:31	LB137697
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	18:07	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	18:07	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	18:07	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	18:07	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	18:07	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	18:07	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	18:07	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	18:07	LB137697
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	19:04	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	19:04	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	19:04	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	19:04	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	19:04	LB137697