

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	8.44			-20	20	10/24/2025	14:52	LB137662
	Cadmium	2.28	1.0	228	-5	7	10/24/2025	14:52	LB137662
	Chromium	50.0	52.0	96	42	62	10/24/2025	14:52	LB137662
	Copper	9.81	2.0	490	-18	22	10/24/2025	14:52	LB137662
	Lead	3.20			-12	12	10/24/2025	14:52	LB137662
	Nickel	4.35	2.0	218	-38	42	10/24/2025	14:52	LB137662
	Selenium	7.23			-20	20	10/24/2025	14:52	LB137662
	Zinc	4.78			-40	40	10/24/2025	14:52	LB137662
ICSAB01	Arsenic	110	104	106	88.4	120	10/24/2025	14:56	LB137662
	Cadmium	1030	972	106	826	1120	10/24/2025	14:56	LB137662
	Chromium	566	542	104	460	624	10/24/2025	14:56	LB137662
	Copper	499	511	98	434	588	10/24/2025	14:56	LB137662
	Lead	51.7	49.0	106	37	61	10/24/2025	14:56	LB137662
	Nickel	1030	954	108	810	1100	10/24/2025	14:56	LB137662
	Selenium	54.0	46.0	117	26	66	10/24/2025	14:56	LB137662
	Zinc	1050	952	110	809	1095	10/24/2025	14:56	LB137662
ICSA01	Arsenic	4.77			-20	20	10/29/2025	14:21	LB137697
	Cadmium	0.66	1.0	66	-5	7	10/29/2025	14:21	LB137697
	Chromium	48.9	52.0	94	42	62	10/29/2025	14:21	LB137697
	Copper	14.1	2.0	705	-18	22	10/29/2025	14:21	LB137697
	Lead	0.038			-12	12	10/29/2025	14:21	LB137697
	Nickel	4.85	2.0	242	-38	42	10/29/2025	14:21	LB137697
	Selenium	1.76			-20	20	10/29/2025	14:21	LB137697
	Zinc	3.72			-40	40	10/29/2025	14:21	LB137697
ICSAB01	Arsenic	105	104	101	88.4	120	10/29/2025	14:25	LB137697
	Cadmium	963	972	99	826	1120	10/29/2025	14:25	LB137697
	Chromium	538	542	99	460	624	10/29/2025	14:25	LB137697
	Copper	474	511	93	434	588	10/29/2025	14:25	LB137697
	Lead	47.4	49.0	97	37	61	10/29/2025	14:25	LB137697
	Nickel	968	954	102	810	1100	10/29/2025	14:25	LB137697
	Selenium	51.1	46.0	111	26	66	10/29/2025	14:25	LB137697
	Zinc	999	952	105	809	1095	10/29/2025	14:25	LB137697