

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Dal-Tile

SDG No.: Q3420

Contract: DALT01

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: COMPOSITEA

Sample ID: Q3420-05

Spiked ID: Q3420-05A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	505		51.5		400	113		P
Selenium	ug/L	75 - 125	1690		492		1000	120		P
Zinc	ug/L	75 - 125	376		266		100	110		P