

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3422

MATRIX: /Water

METHOD: 7196A, Lloyd Kahn,

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all compounds.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS: Lab has used least representable sample weight for the sample Q3422-01 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

As per method, aqueous sample for Hexavalent Chromium analysis should be filtered within 15 minutes of collection time. However, samples were not filtered as per requirement therefore Lab has filtered in the Lab within the same day of sample receipt and analyzed within 24 hours from collection.

QA REVIEW

Date