

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV86	Mercury	4.16	4.0	104	90 - 110	CV	10/24/2025	09:20	LB137632

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV83	Mercury	4.66	5.0	93	90 - 110	CV	10/24/2025	09:25	LB137632
CCV84	Mercury	5.06	5.0	101	90 - 110	CV	10/24/2025	10:09	LB137632
CCV85	Mercury	5.05	5.0	101	90 - 110	CV	10/24/2025	10:40	LB137632
CCV86	Mercury	5.01	5.0	100	90 - 110	CV	10/24/2025	11:12	LB137632
CCV87	Mercury	4.70	5.0	94	90 - 110	CV	10/24/2025	11:33	LB137632

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7820	8000	98	90 - 110	P	10/24/2025	14:33	LB137662
	Antimony	4010	4000	100	90 - 110	P	10/24/2025	14:33	LB137662
	Arsenic	3950	4000	99	90 - 110	P	10/24/2025	14:33	LB137662
	Barium	7700	8000	96	90 - 110	P	10/24/2025	14:33	LB137662
	Beryllium	198	200	99	90 - 110	P	10/24/2025	14:33	LB137662
	Cadmium	1950	2000	98	90 - 110	P	10/24/2025	14:33	LB137662
	Calcium	19400	20000	97	90 - 110	P	10/24/2025	14:33	LB137662
	Chromium	812	800	101	90 - 110	P	10/24/2025	14:33	LB137662
	Cobalt	1960	2000	98	90 - 110	P	10/24/2025	14:33	LB137662
	Copper	1010	1000	101	90 - 110	P	10/24/2025	14:33	LB137662
	Iron	3910	4000	98	90 - 110	P	10/24/2025	14:33	LB137662
	Lead	3830	4000	96	90 - 110	P	10/24/2025	14:33	LB137662
	Magnesium	19500	20000	97	90 - 110	P	10/24/2025	14:33	LB137662
	Manganese	1920	2000	96	90 - 110	P	10/24/2025	14:33	LB137662
	Nickel	1970	2000	98	90 - 110	P	10/24/2025	14:33	LB137662
	Potassium	19100	20000	96	90 - 110	P	10/24/2025	14:33	LB137662
	Selenium	3970	4000	99	90 - 110	P	10/24/2025	14:33	LB137662
	Silver	986	1000	99	90 - 110	P	10/24/2025	14:33	LB137662
	Sodium	18600	20000	93	90 - 110	P	10/24/2025	14:33	LB137662
	Thallium	4010	4000	100	90 - 110	P	10/24/2025	14:33	LB137662
	Vanadium	1950	2000	97	90 - 110	P	10/24/2025	14:33	LB137662
	Zinc	1980	2000	99	90 - 110	P	10/24/2025	14:33	LB137662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	108	100	108	80 - 120	P	10/24/2025	14:38	LB137662
	Antimony	52.6	50.0	105	80 - 120	P	10/24/2025	14:38	LB137662
	Arsenic	23.0	20.0	115	80 - 120	P	10/24/2025	14:38	LB137662
	Barium	100	100	100	80 - 120	P	10/24/2025	14:38	LB137662
	Beryllium	6.40	6.0	107	80 - 120	P	10/24/2025	14:38	LB137662
	Cadmium	6.05	6.0	101	80 - 120	P	10/24/2025	14:38	LB137662
	Calcium	2100	2000	105	80 - 120	P	10/24/2025	14:38	LB137662
	Chromium	10.4	10.0	104	80 - 120	P	10/24/2025	14:38	LB137662
	Cobalt	30.5	30.0	102	80 - 120	P	10/24/2025	14:38	LB137662
	Copper	22.9	20.0	114	80 - 120	P	10/24/2025	14:38	LB137662
	Iron	104	100	104	80 - 120	P	10/24/2025	14:38	LB137662
	Lead	12.9	12.0	107	80 - 120	P	10/24/2025	14:38	LB137662
	Magnesium	2170	2000	109	80 - 120	P	10/24/2025	14:38	LB137662
	Manganese	22.0	20.0	110	80 - 120	P	10/24/2025	14:38	LB137662
	Nickel	40.7	40.0	102	80 - 120	P	10/24/2025	14:38	LB137662
	Potassium	2010	2000	101	80 - 120	P	10/24/2025	14:38	LB137662
	Selenium	20.8	20.0	104	80 - 120	P	10/24/2025	14:38	LB137662
	Silver	11.2	10.0	112	80 - 120	P	10/24/2025	14:38	LB137662
	Sodium	1900	2000	95	80 - 120	P	10/24/2025	14:38	LB137662
	Thallium	45.9	40.0	115	80 - 120	P	10/24/2025	14:38	LB137662
	Vanadium	40.6	40.0	102	80 - 120	P	10/24/2025	14:38	LB137662
	Zinc	43.9	40.0	110	80 - 120	P	10/24/2025	14:38	LB137662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	102	90 - 110	P	10/24/2025	15:12	LB137662
	Antimony	5110	5000	102	90 - 110	P	10/24/2025	15:12	LB137662
	Arsenic	5130	5000	102	90 - 110	P	10/24/2025	15:12	LB137662
	Barium	9810	10000	98	90 - 110	P	10/24/2025	15:12	LB137662
	Beryllium	253	250	101	90 - 110	P	10/24/2025	15:12	LB137662
	Cadmium	2550	2500	102	90 - 110	P	10/24/2025	15:12	LB137662
	Calcium	25100	25000	100	90 - 110	P	10/24/2025	15:12	LB137662
	Chromium	1030	1000	103	90 - 110	P	10/24/2025	15:12	LB137662
	Cobalt	2530	2500	101	90 - 110	P	10/24/2025	15:12	LB137662
	Copper	1280	1250	102	90 - 110	P	10/24/2025	15:12	LB137662
	Iron	5070	5000	102	90 - 110	P	10/24/2025	15:12	LB137662
	Lead	5060	5000	101	90 - 110	P	10/24/2025	15:12	LB137662
	Magnesium	25100	25000	100	90 - 110	P	10/24/2025	15:12	LB137662
	Manganese	2480	2500	99	90 - 110	P	10/24/2025	15:12	LB137662
	Nickel	2540	2500	102	90 - 110	P	10/24/2025	15:12	LB137662
	Potassium	24900	25000	100	90 - 110	P	10/24/2025	15:12	LB137662
	Selenium	5140	5000	103	90 - 110	P	10/24/2025	15:12	LB137662
	Silver	1270	1250	102	90 - 110	P	10/24/2025	15:12	LB137662
	Sodium	24300	25000	97	90 - 110	P	10/24/2025	15:12	LB137662
	Thallium	5220	5000	104	90 - 110	P	10/24/2025	15:12	LB137662
	Vanadium	2500	2500	100	90 - 110	P	10/24/2025	15:12	LB137662
	Zinc	2580	2500	103	90 - 110	P	10/24/2025	15:12	LB137662
CCV02	Aluminum	10300	10000	103	90 - 110	P	10/24/2025	16:32	LB137662
	Antimony	5070	5000	101	90 - 110	P	10/24/2025	16:32	LB137662
	Arsenic	5130	5000	103	90 - 110	P	10/24/2025	16:32	LB137662
	Barium	10100	10000	101	90 - 110	P	10/24/2025	16:32	LB137662
	Beryllium	268	250	107	90 - 110	P	10/24/2025	16:32	LB137662
	Cadmium	2580	2500	103	90 - 110	P	10/24/2025	16:32	LB137662
	Calcium	25700	25000	103	90 - 110	P	10/24/2025	16:32	LB137662
	Chromium	1050	1000	105	90 - 110	P	10/24/2025	16:32	LB137662
	Cobalt	2560	2500	103	90 - 110	P	10/24/2025	16:32	LB137662
	Copper	1290	1250	103	90 - 110	P	10/24/2025	16:32	LB137662
	Iron	5080	5000	102	90 - 110	P	10/24/2025	16:32	LB137662
	Lead	5110	5000	102	90 - 110	P	10/24/2025	16:32	LB137662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	25900	25000	104	90 - 110	P	10/24/2025	16:32	LB137662
	Manganese	2560	2500	102	90 - 110	P	10/24/2025	16:32	LB137662
	Nickel	2570	2500	103	90 - 110	P	10/24/2025	16:32	LB137662
	Potassium	26000	25000	104	90 - 110	P	10/24/2025	16:32	LB137662
	Selenium	5120	5000	102	90 - 110	P	10/24/2025	16:32	LB137662
	Silver	1300	1250	104	90 - 110	P	10/24/2025	16:32	LB137662
	Sodium	27200	25000	109	90 - 110	P	10/24/2025	16:32	LB137662
	Thallium	5190	5000	104	90 - 110	P	10/24/2025	16:32	LB137662
	Vanadium	2550	2500	102	90 - 110	P	10/24/2025	16:32	LB137662
	Zinc	2610	2500	104	90 - 110	P	10/24/2025	16:32	LB137662
CCV03	Aluminum	10300	10000	103	90 - 110	P	10/24/2025	18:29	LB137662
	Antimony	4880	5000	98	90 - 110	P	10/24/2025	18:29	LB137662
	Arsenic	4980	5000	100	90 - 110	P	10/24/2025	18:29	LB137662
	Barium	10100	10000	101	90 - 110	P	10/24/2025	18:29	LB137662
	Beryllium	274	250	110	90 - 110	P	10/24/2025	18:29	LB137662
	Cadmium	2530	2500	101	90 - 110	P	10/24/2025	18:29	LB137662
	Calcium	25800	25000	103	90 - 110	P	10/24/2025	18:29	LB137662
	Chromium	1040	1000	104	90 - 110	P	10/24/2025	18:29	LB137662
	Cobalt	2500	2500	100	90 - 110	P	10/24/2025	18:29	LB137662
	Copper	1250	1250	100	90 - 110	P	10/24/2025	18:29	LB137662
	Iron	5170	5000	103	90 - 110	P	10/24/2025	18:29	LB137662
	Lead	5010	5000	100	90 - 110	P	10/24/2025	18:29	LB137662
	Magnesium	26100	25000	104	90 - 110	P	10/24/2025	18:29	LB137662
	Manganese	2560	2500	102	90 - 110	P	10/24/2025	18:29	LB137662
	Nickel	2510	2500	100	90 - 110	P	10/24/2025	18:29	LB137662
	Potassium	24500	25000	98	90 - 110	P	10/24/2025	18:29	LB137662
	Selenium	4940	5000	99	90 - 110	P	10/24/2025	18:29	LB137662
	Silver	1270	1250	102	90 - 110	P	10/24/2025	18:29	LB137662
	Sodium	24500	25000	98	90 - 110	P	10/24/2025	18:29	LB137662
	Thallium	5060	5000	101	90 - 110	P	10/24/2025	18:29	LB137662
	Vanadium	2550	2500	102	90 - 110	P	10/24/2025	18:29	LB137662
	Zinc	2490	2500	100	90 - 110	P	10/24/2025	18:29	LB137662
CCV04	Aluminum	9950	10000	100	90 - 110	P	10/24/2025	19:19	LB137662
	Antimony	4810	5000	96	90 - 110	P	10/24/2025	19:19	LB137662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4830	5000	97	90 - 110	P	10/24/2025	19:19	LB137662
	Barium	10200	10000	102	90 - 110	P	10/24/2025	19:19	LB137662
	Beryllium	265	250	106	90 - 110	P	10/24/2025	19:19	LB137662
	Cadmium	2490	2500	100	90 - 110	P	10/24/2025	19:19	LB137662
	Calcium	25200	25000	101	90 - 110	P	10/24/2025	19:19	LB137662
	Chromium	1010	1000	101	90 - 110	P	10/24/2025	19:19	LB137662
	Cobalt	2470	2500	99	90 - 110	P	10/24/2025	19:19	LB137662
	Copper	1240	1250	99	90 - 110	P	10/24/2025	19:19	LB137662
	Iron	4840	5000	97	90 - 110	P	10/24/2025	19:19	LB137662
	Lead	4940	5000	99	90 - 110	P	10/24/2025	19:19	LB137662
	Magnesium	25200	25000	101	90 - 110	P	10/24/2025	19:19	LB137662
	Manganese	2560	2500	102	90 - 110	P	10/24/2025	19:19	LB137662
	Nickel	2480	2500	99	90 - 110	P	10/24/2025	19:19	LB137662
	Potassium	23800	25000	95	90 - 110	P	10/24/2025	19:19	LB137662
	Selenium	4770	5000	95	90 - 110	P	10/24/2025	19:19	LB137662
	Silver	1230	1250	98	90 - 110	P	10/24/2025	19:19	LB137662
	Sodium	24500	25000	98	90 - 110	P	10/24/2025	19:19	LB137662
	Thallium	4750	5000	95	90 - 110	P	10/24/2025	19:19	LB137662
	Vanadium	2480	2500	99	90 - 110	P	10/24/2025	19:19	LB137662
	Zinc	2430	2500	97	90 - 110	P	10/24/2025	19:19	LB137662
CCV05	Aluminum	9880	10000	99	90 - 110	P	10/24/2025	20:06	LB137662
	Antimony	4740	5000	95	90 - 110	P	10/24/2025	20:06	LB137662
	Arsenic	4800	5000	96	90 - 110	P	10/24/2025	20:06	LB137662
	Barium	9900	10000	99	90 - 110	P	10/24/2025	20:06	LB137662
	Beryllium	265	250	106	90 - 110	P	10/24/2025	20:06	LB137662
	Cadmium	2480	2500	99	90 - 110	P	10/24/2025	20:06	LB137662
	Calcium	24900	25000	100	90 - 110	P	10/24/2025	20:06	LB137662
	Chromium	1010	1000	101	90 - 110	P	10/24/2025	20:06	LB137662
	Cobalt	2450	2500	98	90 - 110	P	10/24/2025	20:06	LB137662
	Copper	1230	1250	98	90 - 110	P	10/24/2025	20:06	LB137662
	Iron	4680	5000	94	90 - 110	P	10/24/2025	20:06	LB137662
	Lead	4890	5000	98	90 - 110	P	10/24/2025	20:06	LB137662
	Magnesium	25200	25000	101	90 - 110	P	10/24/2025	20:06	LB137662
	Manganese	2510	2500	100	90 - 110	P	10/24/2025	20:06	LB137662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2460	2500	98	90 - 110	P	10/24/2025	20:06	LB137662
	Potassium	23900	25000	96	90 - 110	P	10/24/2025	20:06	LB137662
	Selenium	4730	5000	94	90 - 110	P	10/24/2025	20:06	LB137662
	Silver	1210	1250	97	90 - 110	P	10/24/2025	20:06	LB137662
	Sodium	24900	25000	100	90 - 110	P	10/24/2025	20:06	LB137662
	Thallium	4910	5000	98	90 - 110	P	10/24/2025	20:06	LB137662
	Vanadium	2450	2500	98	90 - 110	P	10/24/2025	20:06	LB137662
	Zinc	2420	2500	97	90 - 110	P	10/24/2025	20:06	LB137662

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7290	8000	91	90 - 110	P	10/28/2025	12:30	LB137677
	Antimony	3900	4000	97	90 - 110	P	10/28/2025	12:30	LB137677
	Arsenic	3820	4000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Barium	7600	8000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Beryllium	191	200	95	90 - 110	P	10/28/2025	12:30	LB137677
	Cadmium	1900	2000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Calcium	18800	20000	94	90 - 110	P	10/28/2025	12:30	LB137677
	Chromium	783	800	98	90 - 110	P	10/28/2025	12:30	LB137677
	Cobalt	1920	2000	96	90 - 110	P	10/28/2025	12:30	LB137677
	Copper	983	1000	98	90 - 110	P	10/28/2025	12:30	LB137677
	Iron	3830	4000	96	90 - 110	P	10/28/2025	12:30	LB137677
	Lead	3730	4000	93	90 - 110	P	10/28/2025	12:30	LB137677
	Magnesium	18800	20000	94	90 - 110	P	10/28/2025	12:30	LB137677
	Manganese	1880	2000	94	90 - 110	P	10/28/2025	12:30	LB137677
	Nickel	1920	2000	96	90 - 110	P	10/28/2025	12:30	LB137677
	Potassium	19000	20000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Selenium	3820	4000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Silver	963	1000	96	90 - 110	P	10/28/2025	12:30	LB137677
	Sodium	18700	20000	94	90 - 110	P	10/28/2025	12:30	LB137677
	Thallium	3890	4000	97	90 - 110	P	10/28/2025	12:30	LB137677
	Vanadium	1880	2000	94	90 - 110	P	10/28/2025	12:30	LB137677
	Zinc	1930	2000	97	90 - 110	P	10/28/2025	12:30	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	104	100	104	80 - 120	P	10/28/2025	12:34	LB137677
	Antimony	52.9	50.0	106	80 - 120	P	10/28/2025	12:34	LB137677
	Arsenic	20.3	20.0	102	80 - 120	P	10/28/2025	12:34	LB137677
	Barium	93.0	100	93	80 - 120	P	10/28/2025	12:34	LB137677
	Beryllium	6.41	6.0	107	80 - 120	P	10/28/2025	12:34	LB137677
	Cadmium	5.86	6.0	98	80 - 120	P	10/28/2025	12:34	LB137677
	Calcium	1990	2000	100	80 - 120	P	10/28/2025	12:34	LB137677
	Chromium	10.6	10.0	106	80 - 120	P	10/28/2025	12:34	LB137677
	Cobalt	29.8	30.0	100	80 - 120	P	10/28/2025	12:34	LB137677
	Copper	22.2	20.0	111	80 - 120	P	10/28/2025	12:34	LB137677
	Iron	98.8	100	99	80 - 120	P	10/28/2025	12:34	LB137677
	Lead	12.0	12.0	100	80 - 120	P	10/28/2025	12:34	LB137677
	Magnesium	2050	2000	103	80 - 120	P	10/28/2025	12:34	LB137677
	Manganese	20.1	20.0	101	80 - 120	P	10/28/2025	12:34	LB137677
	Nickel	39.9	40.0	100	80 - 120	P	10/28/2025	12:34	LB137677
	Potassium	1840	2000	92	80 - 120	P	10/28/2025	12:34	LB137677
	Selenium	20.7	20.0	104	80 - 120	P	10/28/2025	12:34	LB137677
	Silver	12.0	10.0	120	80 - 120	P	10/28/2025	12:34	LB137677
	Sodium	1730	2000	86	80 - 120	P	10/28/2025	12:34	LB137677
	Thallium	44.4	40.0	111	80 - 120	P	10/28/2025	12:34	LB137677
	Vanadium	37.3	40.0	93	80 - 120	P	10/28/2025	12:34	LB137677
	Zinc	42.5	40.0	106	80 - 120	P	10/28/2025	12:34	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10400	10000	104	90 - 110	P	10/28/2025	13:03	LB137677
	Antimony	4970	5000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Arsenic	4950	5000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Barium	10000	10000	100	90 - 110	P	10/28/2025	13:03	LB137677
	Beryllium	246	250	98	90 - 110	P	10/28/2025	13:03	LB137677
	Cadmium	2480	2500	99	90 - 110	P	10/28/2025	13:03	LB137677
	Calcium	24600	25000	98	90 - 110	P	10/28/2025	13:03	LB137677
	Chromium	989	1000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Cobalt	2490	2500	99	90 - 110	P	10/28/2025	13:03	LB137677
	Copper	1250	1250	100	90 - 110	P	10/28/2025	13:03	LB137677
	Iron	5030	5000	101	90 - 110	P	10/28/2025	13:03	LB137677
	Lead	4910	5000	98	90 - 110	P	10/28/2025	13:03	LB137677
	Magnesium	24200	25000	97	90 - 110	P	10/28/2025	13:03	LB137677
	Manganese	2480	2500	99	90 - 110	P	10/28/2025	13:03	LB137677
	Nickel	2480	2500	99	90 - 110	P	10/28/2025	13:03	LB137677
	Potassium	24700	25000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Selenium	4960	5000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Silver	1240	1250	99	90 - 110	P	10/28/2025	13:03	LB137677
	Sodium	24600	25000	98	90 - 110	P	10/28/2025	13:03	LB137677
	Thallium	5020	5000	100	90 - 110	P	10/28/2025	13:03	LB137677
	Vanadium	2460	2500	98	90 - 110	P	10/28/2025	13:03	LB137677
	Zinc	2500	2500	100	90 - 110	P	10/28/2025	13:03	LB137677
CCV02	Aluminum	11000	10000	110	90 - 110	P	10/28/2025	14:01	LB137677
	Antimony	5220	5000	104	90 - 110	P	10/28/2025	14:01	LB137677
	Arsenic	5220	5000	104	90 - 110	P	10/28/2025	14:01	LB137677
	Barium	10200	10000	102	90 - 110	P	10/28/2025	14:01	LB137677
	Beryllium	262	250	105	90 - 110	P	10/28/2025	14:01	LB137677
	Cadmium	2610	2500	104	90 - 110	P	10/28/2025	14:01	LB137677
	Calcium	25900	25000	104	90 - 110	P	10/28/2025	14:01	LB137677
	Chromium	1060	1000	106	90 - 110	P	10/28/2025	14:01	LB137677
	Cobalt	2600	2500	104	90 - 110	P	10/28/2025	14:01	LB137677
	Copper	1300	1250	104	90 - 110	P	10/28/2025	14:01	LB137677
	Iron	5170	5000	103	90 - 110	P	10/28/2025	14:01	LB137677
	Lead	5150	5000	103	90 - 110	P	10/28/2025	14:01	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	25800	25000	103	90 - 110	P	10/28/2025	14:01	LB137677
	Manganese	2560	2500	102	90 - 110	P	10/28/2025	14:01	LB137677
	Nickel	2600	2500	104	90 - 110	P	10/28/2025	14:01	LB137677
	Potassium	26100	25000	104	90 - 110	P	10/28/2025	14:01	LB137677
	Selenium	5230	5000	105	90 - 110	P	10/28/2025	14:01	LB137677
	Silver	1290	1250	104	90 - 110	P	10/28/2025	14:01	LB137677
	Sodium	26000	25000	104	90 - 110	P	10/28/2025	14:01	LB137677
	Thallium	5450	5000	109	90 - 110	P	10/28/2025	14:01	LB137677
	Vanadium	2590	2500	103	90 - 110	P	10/28/2025	14:01	LB137677
	Zinc	2590	2500	104	90 - 110	P	10/28/2025	14:01	LB137677
CCV03	Aluminum	10000	10000	100	90 - 110	P	10/28/2025	14:53	LB137677
	Antimony	4950	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Arsenic	4950	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Barium	10500	10000	105	90 - 110	P	10/28/2025	14:53	LB137677
	Beryllium	262	250	105	90 - 110	P	10/28/2025	14:53	LB137677
	Cadmium	2510	2500	101	90 - 110	P	10/28/2025	14:53	LB137677
	Calcium	25900	25000	104	90 - 110	P	10/28/2025	14:53	LB137677
	Chromium	1040	1000	104	90 - 110	P	10/28/2025	14:53	LB137677
	Cobalt	2520	2500	101	90 - 110	P	10/28/2025	14:53	LB137677
	Copper	1260	1250	101	90 - 110	P	10/28/2025	14:53	LB137677
	Iron	5130	5000	103	90 - 110	P	10/28/2025	14:53	LB137677
	Lead	4950	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Magnesium	25700	25000	103	90 - 110	P	10/28/2025	14:53	LB137677
	Manganese	2640	2500	106	90 - 110	P	10/28/2025	14:53	LB137677
	Nickel	2510	2500	100	90 - 110	P	10/28/2025	14:53	LB137677
	Potassium	26000	25000	104	90 - 110	P	10/28/2025	14:53	LB137677
	Selenium	4940	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Silver	1280	1250	102	90 - 110	P	10/28/2025	14:53	LB137677
	Sodium	27500	25000	110	90 - 110	P	10/28/2025	14:53	LB137677
	Thallium	5070	5000	101	90 - 110	P	10/28/2025	14:53	LB137677
	Vanadium	2570	2500	103	90 - 110	P	10/28/2025	14:53	LB137677
	Zinc	2560	2500	102	90 - 110	P	10/28/2025	14:53	LB137677
CCV04	Aluminum	9560	10000	96	90 - 110	P	10/28/2025	15:45	LB137677
	Antimony	4910	5000	98	90 - 110	P	10/28/2025	15:45	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4930	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Barium	9920	10000	99	90 - 110	P	10/28/2025	15:45	LB137677
	Beryllium	256	250	102	90 - 110	P	10/28/2025	15:45	LB137677
	Cadmium	2500	2500	100	90 - 110	P	10/28/2025	15:45	LB137677
	Calcium	24700	25000	99	90 - 110	P	10/28/2025	15:45	LB137677
	Chromium	1000	1000	100	90 - 110	P	10/28/2025	15:45	LB137677
	Cobalt	2490	2500	100	90 - 110	P	10/28/2025	15:45	LB137677
	Copper	1250	1250	100	90 - 110	P	10/28/2025	15:45	LB137677
	Iron	4890	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Lead	4920	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Magnesium	24700	25000	99	90 - 110	P	10/28/2025	15:45	LB137677
	Manganese	2500	2500	100	90 - 110	P	10/28/2025	15:45	LB137677
	Nickel	2490	2500	99	90 - 110	P	10/28/2025	15:45	LB137677
	Potassium	24800	25000	99	90 - 110	P	10/28/2025	15:45	LB137677
	Selenium	4930	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Silver	1240	1250	99	90 - 110	P	10/28/2025	15:45	LB137677
	Sodium	25700	25000	103	90 - 110	P	10/28/2025	15:45	LB137677
	Thallium	5210	5000	104	90 - 110	P	10/28/2025	15:45	LB137677
	Vanadium	2460	2500	98	90 - 110	P	10/28/2025	15:45	LB137677
	Zinc	2480	2500	99	90 - 110	P	10/28/2025	15:45	LB137677
CCV05	Aluminum	9900	10000	99	90 - 110	P	10/28/2025	16:35	LB137677
	Antimony	5120	5000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Arsenic	5100	5000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Barium	10300	10000	103	90 - 110	P	10/28/2025	16:35	LB137677
	Beryllium	262	250	105	90 - 110	P	10/28/2025	16:35	LB137677
	Cadmium	2560	2500	102	90 - 110	P	10/28/2025	16:35	LB137677
	Calcium	25500	25000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Chromium	1030	1000	103	90 - 110	P	10/28/2025	16:35	LB137677
	Cobalt	2580	2500	103	90 - 110	P	10/28/2025	16:35	LB137677
	Copper	1290	1250	103	90 - 110	P	10/28/2025	16:35	LB137677
	Iron	5070	5000	101	90 - 110	P	10/28/2025	16:35	LB137677
	Lead	5060	5000	101	90 - 110	P	10/28/2025	16:35	LB137677
	Magnesium	25400	25000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Manganese	2590	2500	104	90 - 110	P	10/28/2025	16:35	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2560	2500	102	90 - 110	P	10/28/2025	16:35	LB137677
	Potassium	25500	25000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Selenium	5090	5000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Silver	1280	1250	102	90 - 110	P	10/28/2025	16:35	LB137677
	Sodium	26400	25000	106	90 - 110	P	10/28/2025	16:35	LB137677
	Thallium	5040	5000	101	90 - 110	P	10/28/2025	16:35	LB137677
	Vanadium	2540	2500	102	90 - 110	P	10/28/2025	16:35	LB137677
	Zinc	2560	2500	102	90 - 110	P	10/28/2025	16:35	LB137677
CCV06	Aluminum	9680	10000	97	90 - 110	P	10/28/2025	17:27	LB137677
	Antimony	4940	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Arsenic	4960	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Barium	10000	10000	100	90 - 110	P	10/28/2025	17:27	LB137677
	Beryllium	264	250	106	90 - 110	P	10/28/2025	17:27	LB137677
	Cadmium	2520	2500	101	90 - 110	P	10/28/2025	17:27	LB137677
	Calcium	25000	25000	100	90 - 110	P	10/28/2025	17:27	LB137677
	Chromium	1010	1000	101	90 - 110	P	10/28/2025	17:27	LB137677
	Cobalt	2520	2500	101	90 - 110	P	10/28/2025	17:27	LB137677
	Copper	1260	1250	101	90 - 110	P	10/28/2025	17:27	LB137677
	Iron	4880	5000	98	90 - 110	P	10/28/2025	17:27	LB137677
	Lead	4960	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Magnesium	25000	25000	100	90 - 110	P	10/28/2025	17:27	LB137677
	Manganese	2540	2500	102	90 - 110	P	10/28/2025	17:27	LB137677
	Nickel	2510	2500	100	90 - 110	P	10/28/2025	17:27	LB137677
	Potassium	24500	25000	98	90 - 110	P	10/28/2025	17:27	LB137677
	Selenium	4970	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Silver	1260	1250	101	90 - 110	P	10/28/2025	17:27	LB137677
	Sodium	25300	25000	101	90 - 110	P	10/28/2025	17:27	LB137677
	Thallium	4770	5000	96	90 - 110	P	10/28/2025	17:27	LB137677
	Vanadium	2480	2500	99	90 - 110	P	10/28/2025	17:27	LB137677
	Zinc	2520	2500	101	90 - 110	P	10/28/2025	17:27	LB137677
CCV07	Aluminum	9740	10000	97	90 - 110	P	10/28/2025	18:17	LB137677
	Antimony	4970	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Arsenic	4970	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Barium	10200	10000	102	90 - 110	P	10/28/2025	18:17	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	258	250	103	90 - 110	P	10/28/2025	18:17	LB137677
	Cadmium	2520	2500	101	90 - 110	P	10/28/2025	18:17	LB137677
	Calcium	25200	25000	101	90 - 110	P	10/28/2025	18:17	LB137677
	Chromium	1010	1000	101	90 - 110	P	10/28/2025	18:17	LB137677
	Cobalt	2530	2500	101	90 - 110	P	10/28/2025	18:17	LB137677
	Copper	1270	1250	101	90 - 110	P	10/28/2025	18:17	LB137677
	Iron	4990	5000	100	90 - 110	P	10/28/2025	18:17	LB137677
	Lead	4950	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Magnesium	25100	25000	100	90 - 110	P	10/28/2025	18:17	LB137677
	Manganese	2570	2500	103	90 - 110	P	10/28/2025	18:17	LB137677
	Nickel	2530	2500	101	90 - 110	P	10/28/2025	18:17	LB137677
	Potassium	24500	25000	98	90 - 110	P	10/28/2025	18:17	LB137677
	Selenium	4970	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Silver	1260	1250	101	90 - 110	P	10/28/2025	18:17	LB137677
	Sodium	25200	25000	101	90 - 110	P	10/28/2025	18:17	LB137677
	Thallium	4740	5000	95	90 - 110	P	10/28/2025	18:17	LB137677
	Vanadium	2490	2500	100	90 - 110	P	10/28/2025	18:17	LB137677
	Zinc	2520	2500	101	90 - 110	P	10/28/2025	18:17	LB137677
CCV08	Aluminum	10800	10000	108	90 - 110	P	10/28/2025	19:06	LB137677
	Antimony	5100	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Arsenic	5110	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Barium	10500	10000	105	90 - 110	P	10/28/2025	19:06	LB137677
	Beryllium	269	250	108	90 - 110	P	10/28/2025	19:06	LB137677
	Cadmium	2600	2500	104	90 - 110	P	10/28/2025	19:06	LB137677
	Calcium	26100	25000	104	90 - 110	P	10/28/2025	19:06	LB137677
	Chromium	1030	1000	103	90 - 110	P	10/28/2025	19:06	LB137677
	Cobalt	2610	2500	104	90 - 110	P	10/28/2025	19:06	LB137677
	Copper	1300	1250	104	90 - 110	P	10/28/2025	19:06	LB137677
	Iron	5140	5000	103	90 - 110	P	10/28/2025	19:06	LB137677
	Lead	5090	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Magnesium	25900	25000	104	90 - 110	P	10/28/2025	19:06	LB137677
	Manganese	2670	2500	107	90 - 110	P	10/28/2025	19:06	LB137677
	Nickel	2600	2500	104	90 - 110	P	10/28/2025	19:06	LB137677
	Potassium	25200	25000	101	90 - 110	P	10/28/2025	19:06	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5120	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Silver	1290	1250	104	90 - 110	P	10/28/2025	19:06	LB137677
	Sodium	26400	25000	106	90 - 110	P	10/28/2025	19:06	LB137677
	Thallium	4880	5000	98	90 - 110	P	10/28/2025	19:06	LB137677
	Vanadium	2590	2500	104	90 - 110	P	10/28/2025	19:06	LB137677
	Zinc	2590	2500	104	90 - 110	P	10/28/2025	19:06	LB137677
CCV09	Aluminum	9840	10000	98	90 - 110	P	10/28/2025	19:58	LB137677
	Antimony	4980	5000	100	90 - 110	P	10/28/2025	19:58	LB137677
	Arsenic	4990	5000	100	90 - 110	P	10/28/2025	19:58	LB137677
	Barium	10300	10000	104	90 - 110	P	10/28/2025	19:58	LB137677
	Beryllium	262	250	105	90 - 110	P	10/28/2025	19:58	LB137677
	Cadmium	2530	2500	101	90 - 110	P	10/28/2025	19:58	LB137677
	Calcium	25500	25000	102	90 - 110	P	10/28/2025	19:58	LB137677
	Chromium	1010	1000	101	90 - 110	P	10/28/2025	19:58	LB137677
	Cobalt	2540	2500	102	90 - 110	P	10/28/2025	19:58	LB137677
	Copper	1270	1250	102	90 - 110	P	10/28/2025	19:58	LB137677
	Iron	5120	5000	102	90 - 110	P	10/28/2025	19:58	LB137677
	Lead	4960	5000	99	90 - 110	P	10/28/2025	19:58	LB137677
	Magnesium	25300	25000	101	90 - 110	P	10/28/2025	19:58	LB137677
	Manganese	2620	2500	105	90 - 110	P	10/28/2025	19:58	LB137677
	Nickel	2530	2500	101	90 - 110	P	10/28/2025	19:58	LB137677
	Potassium	25600	25000	102	90 - 110	P	10/28/2025	19:58	LB137677
	Selenium	5000	5000	100	90 - 110	P	10/28/2025	19:58	LB137677
	Silver	1280	1250	102	90 - 110	P	10/28/2025	19:58	LB137677
	Sodium	27900	25000	112	90 - 110	P	10/28/2025	19:58	LB137677
	Thallium	4670	5000	93	90 - 110	P	10/28/2025	19:58	LB137677
	Vanadium	2520	2500	101	90 - 110	P	10/28/2025	19:58	LB137677
	Zinc	2560	2500	103	90 - 110	P	10/28/2025	19:58	LB137677
CCV10	Aluminum	9480	10000	95	90 - 110	P	10/28/2025	20:49	LB137677
	Antimony	4890	5000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Arsenic	4880	5000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Barium	10100	10000	101	90 - 110	P	10/28/2025	20:49	LB137677
	Beryllium	251	250	100	90 - 110	P	10/28/2025	20:49	LB137677
	Cadmium	2480	2500	99	90 - 110	P	10/28/2025	20:49	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	24800	25000	99	90 - 110	P	10/28/2025	20:49	LB137677
	Chromium	987	1000	99	90 - 110	P	10/28/2025	20:49	LB137677
	Cobalt	2490	2500	100	90 - 110	P	10/28/2025	20:49	LB137677
	Copper	1250	1250	100	90 - 110	P	10/28/2025	20:49	LB137677
	Iron	4970	5000	99	90 - 110	P	10/28/2025	20:49	LB137677
	Lead	4850	5000	97	90 - 110	P	10/28/2025	20:49	LB137677
	Magnesium	24400	25000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Manganese	2550	2500	102	90 - 110	P	10/28/2025	20:49	LB137677
	Nickel	2480	2500	99	90 - 110	P	10/28/2025	20:49	LB137677
	Potassium	24600	25000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Selenium	4880	5000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Silver	1230	1250	99	90 - 110	P	10/28/2025	20:49	LB137677
	Sodium	25800	25000	103	90 - 110	P	10/28/2025	20:49	LB137677
	Thallium	4470	5000	90	90 - 110	P	10/28/2025	20:49	LB137677
	Vanadium	2450	2500	98	90 - 110	P	10/28/2025	20:49	LB137677
	Zinc	2450	2500	98	90 - 110	P	10/28/2025	20:49	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7260	8000	91	90 - 110	P	10/29/2025	14:01	LB137697
	Antimony	3850	4000	96	90 - 110	P	10/29/2025	14:01	LB137697
	Arsenic	3760	4000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Barium	7490	8000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Beryllium	188	200	94	90 - 110	P	10/29/2025	14:01	LB137697
	Cadmium	1860	2000	93	90 - 110	P	10/29/2025	14:01	LB137697
	Calcium	18700	20000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Chromium	770	800	96	90 - 110	P	10/29/2025	14:01	LB137697
	Cobalt	1880	2000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Copper	961	1000	96	90 - 110	P	10/29/2025	14:01	LB137697
	Iron	3790	4000	95	90 - 110	P	10/29/2025	14:01	LB137697
	Lead	3650	4000	91	90 - 110	P	10/29/2025	14:01	LB137697
	Magnesium	18700	20000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Manganese	1870	2000	93	90 - 110	P	10/29/2025	14:01	LB137697
	Nickel	1880	2000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Potassium	18700	20000	93	90 - 110	P	10/29/2025	14:01	LB137697
	Selenium	3780	4000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Silver	969	1000	97	90 - 110	P	10/29/2025	14:01	LB137697
	Sodium	18800	20000	94	90 - 110	P	10/29/2025	14:01	LB137697
	Thallium	3970	4000	99	90 - 110	P	10/29/2025	14:01	LB137697
	Vanadium	1860	2000	93	90 - 110	P	10/29/2025	14:01	LB137697
	Zinc	1900	2000	95	90 - 110	P	10/29/2025	14:01	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	103	100	103	80 - 120	P	10/29/2025	14:07	LB137697
	Antimony	51.5	50.0	103	80 - 120	P	10/29/2025	14:07	LB137697
	Arsenic	21.1	20.0	105	80 - 120	P	10/29/2025	14:07	LB137697
	Barium	105	100	105	80 - 120	P	10/29/2025	14:07	LB137697
	Beryllium	6.45	6.0	108	80 - 120	P	10/29/2025	14:07	LB137697
	Cadmium	5.83	6.0	97	80 - 120	P	10/29/2025	14:07	LB137697
	Calcium	2080	2000	104	80 - 120	P	10/29/2025	14:07	LB137697
	Chromium	10.3	10.0	102	80 - 120	P	10/29/2025	14:07	LB137697
	Cobalt	30.0	30.0	100	80 - 120	P	10/29/2025	14:07	LB137697
	Copper	22.0	20.0	110	80 - 120	P	10/29/2025	14:07	LB137697
	Iron	107	100	107	80 - 120	P	10/29/2025	14:07	LB137697
	Lead	11.7	12.0	98	80 - 120	P	10/29/2025	14:07	LB137697
	Magnesium	2160	2000	108	80 - 120	P	10/29/2025	14:07	LB137697
	Manganese	21.1	20.0	105	80 - 120	P	10/29/2025	14:07	LB137697
	Nickel	40.4	40.0	101	80 - 120	P	10/29/2025	14:07	LB137697
	Potassium	1880	2000	94	80 - 120	P	10/29/2025	14:07	LB137697
	Selenium	23.1	20.0	115	80 - 120	P	10/29/2025	14:07	LB137697
	Silver	10.6	10.0	106	80 - 120	P	10/29/2025	14:07	LB137697
	Sodium	2010	2000	100	80 - 120	P	10/29/2025	14:07	LB137697
	Thallium	46.3	40.0	116	80 - 120	P	10/29/2025	14:07	LB137697
	Vanadium	41.0	40.0	103	80 - 120	P	10/29/2025	14:07	LB137697
	Zinc	43.3	40.0	108	80 - 120	P	10/29/2025	14:07	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9780	10000	98	90 - 110	P	10/29/2025	14:37	LB137697
	Antimony	4970	5000	99	90 - 110	P	10/29/2025	14:37	LB137697
	Arsenic	4960	5000	99	90 - 110	P	10/29/2025	14:37	LB137697
	Barium	9730	10000	97	90 - 110	P	10/29/2025	14:37	LB137697
	Beryllium	250	250	100	90 - 110	P	10/29/2025	14:37	LB137697
	Cadmium	2460	2500	99	90 - 110	P	10/29/2025	14:37	LB137697
	Calcium	24800	25000	99	90 - 110	P	10/29/2025	14:37	LB137697
	Chromium	1010	1000	100	90 - 110	P	10/29/2025	14:37	LB137697
	Cobalt	2460	2500	98	90 - 110	P	10/29/2025	14:37	LB137697
	Copper	1240	1250	99	90 - 110	P	10/29/2025	14:37	LB137697
	Iron	4870	5000	97	90 - 110	P	10/29/2025	14:37	LB137697
	Lead	4900	5000	98	90 - 110	P	10/29/2025	14:37	LB137697
	Magnesium	24800	25000	99	90 - 110	P	10/29/2025	14:37	LB137697
	Manganese	2460	2500	98	90 - 110	P	10/29/2025	14:37	LB137697
	Nickel	2460	2500	98	90 - 110	P	10/29/2025	14:37	LB137697
	Potassium	24100	25000	97	90 - 110	P	10/29/2025	14:37	LB137697
	Selenium	4930	5000	99	90 - 110	P	10/29/2025	14:37	LB137697
	Silver	1240	1250	99	90 - 110	P	10/29/2025	14:37	LB137697
	Sodium	24400	25000	98	90 - 110	P	10/29/2025	14:37	LB137697
	Thallium	5060	5000	101	90 - 110	P	10/29/2025	14:37	LB137697
	Vanadium	2470	2500	99	90 - 110	P	10/29/2025	14:37	LB137697
	Zinc	2490	2500	100	90 - 110	P	10/29/2025	14:37	LB137697
CCV02	Aluminum	9400	10000	94	90 - 110	P	10/29/2025	15:43	LB137697
	Antimony	4870	5000	97	90 - 110	P	10/29/2025	15:43	LB137697
	Arsenic	4840	5000	97	90 - 110	P	10/29/2025	15:43	LB137697
	Barium	9610	10000	96	90 - 110	P	10/29/2025	15:43	LB137697
	Beryllium	243	250	97	90 - 110	P	10/29/2025	15:43	LB137697
	Cadmium	2420	2500	97	90 - 110	P	10/29/2025	15:43	LB137697
	Calcium	24100	25000	96	90 - 110	P	10/29/2025	15:43	LB137697
	Chromium	988	1000	99	90 - 110	P	10/29/2025	15:43	LB137697
	Cobalt	2420	2500	97	90 - 110	P	10/29/2025	15:43	LB137697
	Copper	1220	1250	98	90 - 110	P	10/29/2025	15:43	LB137697
	Iron	4870	5000	98	90 - 110	P	10/29/2025	15:43	LB137697
	Lead	4820	5000	96	90 - 110	P	10/29/2025	15:43	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24200	25000	97	90 - 110	P	10/29/2025	15:43	LB137697
	Manganese	2390	2500	96	90 - 110	P	10/29/2025	15:43	LB137697
	Nickel	2430	2500	97	90 - 110	P	10/29/2025	15:43	LB137697
	Potassium	24000	25000	96	90 - 110	P	10/29/2025	15:43	LB137697
	Selenium	4790	5000	96	90 - 110	P	10/29/2025	15:43	LB137697
	Silver	1230	1250	98	90 - 110	P	10/29/2025	15:43	LB137697
	Sodium	23900	25000	96	90 - 110	P	10/29/2025	15:43	LB137697
	Thallium	4740	5000	95	90 - 110	P	10/29/2025	15:43	LB137697
	Vanadium	2400	2500	96	90 - 110	P	10/29/2025	15:43	LB137697
	Zinc	2440	2500	98	90 - 110	P	10/29/2025	15:43	LB137697
CCV03	Aluminum	9730	10000	97	90 - 110	P	10/29/2025	16:37	LB137697
	Antimony	5120	5000	102	90 - 110	P	10/29/2025	16:37	LB137697
	Arsenic	5080	5000	102	90 - 110	P	10/29/2025	16:37	LB137697
	Barium	9730	10000	97	90 - 110	P	10/29/2025	16:37	LB137697
	Beryllium	249	250	100	90 - 110	P	10/29/2025	16:37	LB137697
	Cadmium	2510	2500	100	90 - 110	P	10/29/2025	16:37	LB137697
	Calcium	25000	25000	100	90 - 110	P	10/29/2025	16:37	LB137697
	Chromium	1030	1000	103	90 - 110	P	10/29/2025	16:37	LB137697
	Cobalt	2520	2500	101	90 - 110	P	10/29/2025	16:37	LB137697
	Copper	1270	1250	102	90 - 110	P	10/29/2025	16:37	LB137697
	Iron	5080	5000	102	90 - 110	P	10/29/2025	16:37	LB137697
	Lead	5020	5000	100	90 - 110	P	10/29/2025	16:37	LB137697
	Magnesium	25000	25000	100	90 - 110	P	10/29/2025	16:37	LB137697
	Manganese	2470	2500	99	90 - 110	P	10/29/2025	16:37	LB137697
	Nickel	2510	2500	101	90 - 110	P	10/29/2025	16:37	LB137697
	Potassium	25200	25000	101	90 - 110	P	10/29/2025	16:37	LB137697
	Selenium	5070	5000	101	90 - 110	P	10/29/2025	16:37	LB137697
	Silver	1280	1250	102	90 - 110	P	10/29/2025	16:37	LB137697
	Sodium	25300	25000	101	90 - 110	P	10/29/2025	16:37	LB137697
	Thallium	4740	5000	95	90 - 110	P	10/29/2025	16:37	LB137697
	Vanadium	2500	2500	100	90 - 110	P	10/29/2025	16:37	LB137697
	Zinc	2560	2500	102	90 - 110	P	10/29/2025	16:37	LB137697
CCV04	Aluminum	9640	10000	96	90 - 110	P	10/29/2025	17:27	LB137697
	Antimony	4930	5000	99	90 - 110	P	10/29/2025	17:27	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4980	5000	100	90 - 110	P	10/29/2025	17:27	LB137697
	Barium	9640	10000	96	90 - 110	P	10/29/2025	17:27	LB137697
	Beryllium	262	250	105	90 - 110	P	10/29/2025	17:27	LB137697
	Cadmium	2500	2500	100	90 - 110	P	10/29/2025	17:27	LB137697
	Calcium	24600	25000	98	90 - 110	P	10/29/2025	17:27	LB137697
	Chromium	1020	1000	102	90 - 110	P	10/29/2025	17:27	LB137697
	Cobalt	2480	2500	99	90 - 110	P	10/29/2025	17:27	LB137697
	Copper	1250	1250	100	90 - 110	P	10/29/2025	17:27	LB137697
	Iron	4700	5000	94	90 - 110	P	10/29/2025	17:27	LB137697
	Lead	4950	5000	99	90 - 110	P	10/29/2025	17:27	LB137697
	Magnesium	25100	25000	101	90 - 110	P	10/29/2025	17:27	LB137697
	Manganese	2460	2500	99	90 - 110	P	10/29/2025	17:27	LB137697
	Nickel	2490	2500	100	90 - 110	P	10/29/2025	17:27	LB137697
	Potassium	23500	25000	94	90 - 110	P	10/29/2025	17:27	LB137697
	Selenium	4940	5000	99	90 - 110	P	10/29/2025	17:27	LB137697
	Silver	1250	1250	100	90 - 110	P	10/29/2025	17:27	LB137697
	Sodium	23700	25000	95	90 - 110	P	10/29/2025	17:27	LB137697
	Thallium	5310	5000	106	90 - 110	P	10/29/2025	17:27	LB137697
	Vanadium	2460	2500	98	90 - 110	P	10/29/2025	17:27	LB137697
	Zinc	2490	2500	100	90 - 110	P	10/29/2025	17:27	LB137697
CCV05	Aluminum	9470	10000	95	90 - 110	P	10/29/2025	18:03	LB137697
	Antimony	4830	5000	97	90 - 110	P	10/29/2025	18:03	LB137697
	Arsenic	4850	5000	97	90 - 110	P	10/29/2025	18:03	LB137697
	Barium	9570	10000	96	90 - 110	P	10/29/2025	18:03	LB137697
	Beryllium	252	250	101	90 - 110	P	10/29/2025	18:03	LB137697
	Cadmium	2440	2500	98	90 - 110	P	10/29/2025	18:03	LB137697
	Calcium	24400	25000	98	90 - 110	P	10/29/2025	18:03	LB137697
	Chromium	998	1000	100	90 - 110	P	10/29/2025	18:03	LB137697
	Cobalt	2430	2500	97	90 - 110	P	10/29/2025	18:03	LB137697
	Copper	1220	1250	98	90 - 110	P	10/29/2025	18:03	LB137697
	Iron	4780	5000	96	90 - 110	P	10/29/2025	18:03	LB137697
	Lead	4840	5000	97	90 - 110	P	10/29/2025	18:03	LB137697
	Magnesium	24700	25000	99	90 - 110	P	10/29/2025	18:03	LB137697
	Manganese	2440	2500	98	90 - 110	P	10/29/2025	18:03	LB137697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: AECOM

SDG No.: Q3422

Contract: AECO02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2430	2500	97	90 - 110	P	10/29/2025	18:03	LB137697
	Potassium	23100	25000	92	90 - 110	P	10/29/2025	18:03	LB137697
	Selenium	4830	5000	96	90 - 110	P	10/29/2025	18:03	LB137697
	Silver	1220	1250	98	90 - 110	P	10/29/2025	18:03	LB137697
	Sodium	23200	25000	93	90 - 110	P	10/29/2025	18:03	LB137697
	Thallium	5070	5000	102	90 - 110	P	10/29/2025	18:03	LB137697
	Vanadium	2430	2500	97	90 - 110	P	10/29/2025	18:03	LB137697
	Zinc	2450	2500	98	90 - 110	P	10/29/2025	18:03	LB137697
CCV06	Aluminum	9450	10000	94	90 - 110	P	10/29/2025	18:55	LB137697
	Antimony	4740	5000	95	90 - 110	P	10/29/2025	18:55	LB137697
	Arsenic	4790	5000	96	90 - 110	P	10/29/2025	18:55	LB137697
	Barium	9690	10000	97	90 - 110	P	10/29/2025	18:55	LB137697
	Beryllium	253	250	101	90 - 110	P	10/29/2025	18:55	LB137697
	Cadmium	2470	2500	99	90 - 110	P	10/29/2025	18:55	LB137697
	Calcium	24500	25000	98	90 - 110	P	10/29/2025	18:55	LB137697
	Chromium	1010	1000	101	90 - 110	P	10/29/2025	18:55	LB137697
	Cobalt	2460	2500	98	90 - 110	P	10/29/2025	18:55	LB137697
	Copper	1190	1250	95	90 - 110	P	10/29/2025	18:55	LB137697
	Iron	4860	5000	97	90 - 110	P	10/29/2025	18:55	LB137697
	Lead	4920	5000	98	90 - 110	P	10/29/2025	18:55	LB137697
	Magnesium	24800	25000	99	90 - 110	P	10/29/2025	18:55	LB137697
	Manganese	2460	2500	98	90 - 110	P	10/29/2025	18:55	LB137697
	Nickel	2470	2500	99	90 - 110	P	10/29/2025	18:55	LB137697
	Potassium	26000	25000	104	90 - 110	P	10/29/2025	18:55	LB137697
	Selenium	4770	5000	95	90 - 110	P	10/29/2025	18:55	LB137697
	Silver	1240	1250	99	90 - 110	P	10/29/2025	18:55	LB137697
	Sodium	26300	25000	105	90 - 110	P	10/29/2025	18:55	LB137697
	Thallium	5490	5000	110	90 - 110	P	10/29/2025	18:55	LB137697
	Vanadium	2450	2500	98	90 - 110	P	10/29/2025	18:55	LB137697
	Zinc	2250	2500	90	90 - 110	P	10/29/2025	18:55	LB137697